



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1870	OrderDate:	4/24/2025 11:31:00 AM
Client:	Alliance Technical Group, LLC - Newark	Project:	NJ Waste Water PT
Contact:	Mohammad Ahmed	Location:	QA Office

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1870-03	38073-100124	Water	SVOCMS Group2	8270E	04/22/25	04/24/25	04/29/25	04/24/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1870
Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID	Parameter	Concentration	C	MDL	RDL	Units
Client ID :	38073-100124						
Q1870-03	38073-100124	WATER	2,4-Dimethylphenol	38.600	1.9	5	ug/L
			Total Svoc :	38.60			
			Total Concentration:	38.60			



SAMPLE DATA

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	04/22/25
Project:	NJ Waste Water PT		Date Received:	04/24/25
Client Sample ID:	38073-100124		SDG No.:	Q1870
Lab Sample ID:	Q1870-03		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group2
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050038.D	1	04/24/25 12:05	04/29/25 11:39	PB167729

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
105-67-9	2,4-Dimethylphenol	38.6		1.90	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	145		10 - 139	96%	SPK: 150
13127-88-3	Phenol-d6	149		10 - 134	100%	SPK: 150
118-79-6	2,4,6-Tribromophenol	154		44 - 137	103%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	245000	7.763			
1146-65-2	Naphthalene-d8	915000	10.557			
15067-26-2	Acenaphthene-d10	618000	14.41			
1517-22-2	Phenanthrene-d10	1210000	17.157			
1719-03-5	Chrysene-d12	1060000	21.403			
1520-96-3	Perylene-d12	1120000	24.403			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1870

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167729BL	PB167729BL	2-Fluorophenol	150	134	89		10	139
		Phenol-d6	150	129	86		10	134
		2,4,6-Tribromophenol	150	126	84		44	137
PB167729BS	PB167729BS	2-Fluorophenol	150	133	89		10	139
		Phenol-d6	150	131	87		10	134
		2,4,6-Tribromophenol	150	133	89		44	137
Q1870-03	38073-100124	2-Fluorophenol	150	145	96		10	139
		Phenol-d6	150	149	100		10	134
		2,4,6-Tribromophenol	150	154	103		44	137

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1870

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E DataFile: BM050040.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167729BS	2,4-Dimethylphenol	50	43.5	ug/L	87				42	142	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167729BL

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM Case No.: Q1870

SAS No.: Q1870 SDG NO.: Q1870

Lab File ID: BM050036.D

Lab Sample ID: PB167729BL

Instrument ID: BNA_M

Date Extracted: 04/24/2025

Matrix: (soil/water) Water

Date Analyzed: 04/29/2025

Level: (low/med) LOW

Time Analyzed: 09:52

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167729BS	PB167729BS	BM050040.D	04/29/2025
38073-100124	Q1870-03	BM050038.D	04/29/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM

SAS No.: Q1870 SDG NO.: Q1870

Lab File ID: BM050023.D

DFTPP Injection Date: 04/28/2025

Instrument ID: BNA_M

DFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.8
68	Less than 2.0% of mass 69	0.4 (1.3) 1
69	Mass 69 relative abundance	28.5
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	35.7
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.4 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM050024.D	04/28/2025	12:30
SSTDICC005	SSTDICC005	BM050025.D	04/28/2025	13:09
SSTDICC010	SSTDICC010	BM050026.D	04/28/2025	13:48
SSTDICC020	SSTDICC020	BM050027.D	04/28/2025	14:27
SSTDICCC040	SSTDICCC040	BM050028.D	04/28/2025	15:06
SSTDICC050	SSTDICC050	BM050029.D	04/28/2025	15:45
SSTDICC060	SSTDICC060	BM050030.D	04/28/2025	16:24
SSTDICC080	SSTDICC080	BM050031.D	04/28/2025	17:04

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM

SAS No.: Q1870

SDG NO.: Q1870

Lab File ID: BM050034.D

DFTPP Injection Date: 04/29/2025

Instrument ID: BNA_M

DFTPP Injection Time: 08:34

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.3
68	Less than 2.0% of mass 69	0.4 (1.4) 1
69	Mass 69 relative abundance	28.2
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	35
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.5 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050035.D	04/29/2025	09:13
PB167729BL	PB167729BL	BM050036.D	04/29/2025	09:52
38073-100124	Q1870-03	BM050038.D	04/29/2025	11:39
PB167729BS	PB167729BS	BM050040.D	04/29/2025	12:58

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1870 SAS No.: Q1870 SDG NO.: Q1870

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/29/2025

Lab File ID: BM050035.D Time Analyzed: 09:13

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	251798	7.769	911122	10.56	601329	14.41
UPPER LIMIT	503596	8.269	1822240	11.057	1202660	14.91
LOWER LIMIT	125899	7.269	455561	10.057	300665	13.91
EPA SAMPLE NO.						
01 PB167729BL	267473	7.77	926574	10.56	611197	14.41
02 38073-100124	245413	7.76	914618	10.56	618114	14.41
03 PB167729BS	269893	7.76	961152	10.56	608153	14.41

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1870 SAS No.: Q1870 SDG NO.: Q1870

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/29/2025

Lab File ID: BM050035.D Time Analyzed: 09:13

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1168790	17.156	1131300	21.397	1083540	24.397
UPPER LIMIT	2337580	17.656	2262600	21.897	2167080	24.897
LOWER LIMIT	584395	16.656	565650	20.897	541770	23.897
EPA SAMPLE NO.						
01 PB167729BL	1171850	17.16	971539	21.40	1012330	24.40
02 38073-100124	1211510	17.16	1057120	21.40	1117520	24.40
03 PB167729BS	1141690	17.16	1074280	21.40	1054030	24.40

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	
Project:	NJ Waste Water PT		Date Received:	
Client Sample ID:	PB167729BL		SDG No.:	Q1870
Lab Sample ID:	PB167729BL		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group2
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050036.D	1	04/24/25 12:05	04/29/25 09:52	PB167729

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
105-67-9	2,4-Dimethylphenol	1.90	U	1.90	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	134		10 - 139	89%	SPK: 150
13127-88-3	Phenol-d6	129		10 - 134	86%	SPK: 150
118-79-6	2,4,6-Tribromophenol	126		44 - 137	84%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	267000	7.769			
1146-65-2	Naphthalene-d8	927000	10.563			
15067-26-2	Acenaphthene-d10	611000	14.41			
1517-22-2	Phenanthrene-d10	1170000	17.162			
1719-03-5	Chrysene-d12	972000	21.403			
1520-96-3	Perylene-d12	1010000	24.403			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	
Project:	NJ Waste Water PT		Date Received:	
Client Sample ID:	PB167729BS		SDG No.:	Q1870
Lab Sample ID:	PB167729BS		Matrix:	Water
Analytical Method:	SW8270		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group2
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050040.D	1	04/24/25 12:05	04/29/25 12:58	PB167729

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
105-67-9	2,4-Dimethylphenol	43.5		1.90	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	133		10 - 139	89%	SPK: 150
13127-88-3	Phenol-d6	131		10 - 134	87%	SPK: 150
118-79-6	2,4,6-Tribromophenol	133		44 - 137	89%	SPK: 150
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	270000	7.763			
1146-65-2	Naphthalene-d8	961000	10.557			
15067-26-2	Acenaphthene-d10	608000	14.41			
1517-22-2	Phenanthrene-d10	1140000	17.156			
1719-03-5	Chrysene-d12	1070000	21.397			
1520-96-3	Perylene-d12	1050000	24.397			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM042825.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 28 18:09:16 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BM050024.D 5 =BM050025.D 10 =BM050026.D 20 =BM050027.D 40 =BM050028.D 50 =BM050029.D 60 =BM050030.D 80 =BM050031.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane	0.506	0.464	0.478	0.511	0.504	0.491	0.473	0.490		3.75
3) Pyridine	1.201	1.166	1.229	1.331	1.328	1.292	1.261	1.258		5.03
4) n-Nitrosodimet...	0.487	0.459	0.475	0.522	0.518	0.504	0.490	0.493		4.62
5) S 2-Fluorophenol	1.074	1.079	1.107	1.208	1.211	1.173	1.142	1.142		5.04
6) Aniline	1.657	1.689	1.768	1.942	1.922	1.875	1.835	1.813		6.16
7) S Phenol-d6	1.290	1.318	1.398	1.543	1.537	1.494	1.469	1.436		7.13
8) 2-Chlorophenol	1.169	1.161	1.198	1.312	1.301	1.268	1.235	1.235		4.96
9) Benzaldehyde	0.940	0.920	0.971	1.031	1.013	0.956	0.894	0.961		5.08
10) C Phenol	1.349	1.344	1.420	1.539	1.537	1.501	1.483	1.453		5.73
11) bis(2-Chloroet...	1.213	1.105	1.176	1.279	1.271	1.233	1.218	1.213		4.90
12) 1,3-Dichlorobe...	1.483	1.428	1.453	1.566	1.552	1.508	1.453	1.492		3.51
13) C 1,4-Dichlorobe...	1.518	1.418	1.463	1.561	1.557	1.516	1.458	1.499		3.60
14) 1,2-Dichlorobe...	1.420	1.393	1.423	1.522	1.506	1.464	1.405	1.448		3.50
15) Benzyl Alcohol	0.866	0.881	0.966	1.080	1.074	1.050	1.050	0.996		9.16
16) 2,2'-oxybis(1-...	1.487	1.422	1.445	1.552	1.513	1.473	1.433	1.475		3.17
17) 2-Methylphenol	0.857	0.893	0.931	1.033	1.026	0.988	0.975	0.957		6.96
18) Hexachloroethane	0.552	0.508	0.521	0.556	0.548	0.533	0.510	0.533		3.79
19) P n-Nitroso-di-n...	0.802	0.889	0.885	0.921	1.003	0.988	0.949	0.928	0.921	6.95
20) 3+4-Methylphenols	1.137	1.170	1.267	1.393	1.392	1.349	1.341	1.293		8.09
21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone	0.483	0.464	0.485	0.531	0.525	0.506	0.487	0.497		4.94
23) S Nitrobenzene-d5	0.386	0.374	0.397	0.434	0.433	0.417	0.400	0.406		5.68
24) Nitrobenzene	0.337	0.325	0.345	0.374	0.372	0.358	0.346	0.351		5.16
25) Isophorone	0.666	0.648	0.680	0.729	0.729	0.704	0.687	0.692		4.44
26) C 2-Nitrophenol	0.148	0.157	0.172	0.196	0.198	0.192	0.191	0.179		11.33
27) 2,4-Dimethylph...	0.263	0.266	0.289	0.324	0.322	0.312	0.305	0.297		8.42
28) bis(2-Chloroet...	0.407	0.399	0.422	0.455	0.452	0.434	0.425	0.428		4.92
29) C 2,4-Dichloroph...	0.294	0.298	0.326	0.362	0.359	0.351	0.342	0.333		8.47
30) 1,2,4-Trichlor...	0.400	0.384	0.394	0.427	0.426	0.412	0.399	0.406		4.02
31) Naphthalene	1.001	0.959	0.993	1.066	1.057	1.023	0.989	1.013		3.80
32) Benzoic acid		0.138	0.172	0.197	0.206	0.212	0.215	0.190		15.67
33) 4-Chloroaniline	0.361	0.393	0.404	0.441	0.448	0.437	0.423	0.415		7.51
34) C Hexachlorobuta...	0.253	0.240	0.249	0.270	0.270	0.263	0.258	0.258		4.34
35) Caprolactam	0.090	0.085	0.095	0.104	0.104	0.102	0.100	0.097		7.73
36) C 4-Chloro-3-met...	0.289	0.283	0.302	0.325	0.329	0.317	0.309	0.308		5.72
37) 2-Methylnaphth...	0.656	0.639	0.662	0.713	0.714	0.692	0.678	0.679		4.22
38) 1-Methylnaphth...	0.703	0.675	0.708	0.753	0.751	0.731	0.710	0.719		3.92

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM042825.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.624	0.622	0.655	0.742	0.745	0.725	0.733	0.692	8.17	
41) P	Hexachlorocycl...		0.303	0.362	0.455	0.468	0.467	0.483	0.423	17.30	
42) S	2,4,6-Tribromo...	0.253	0.249	0.281	0.318	0.329	0.318	0.323	0.296	11.61	
43) C	2,4,6-Trichlor...	0.372	0.381	0.423	0.476	0.485	0.467	0.469	0.439	10.76	
44)	2,4,5-Trichlor...	0.410	0.417	0.453	0.517	0.519	0.501	0.497	0.474	9.83	
45) S	2-Fluorobiphenyl	1.588	1.555	1.632	1.807	1.822	1.744	1.688	1.691	6.21	
46)	1,1'-Biphenyl	1.427	1.395	1.452	1.573	1.581	1.507	1.473	1.487	4.77	
47)	2-Chloronaphth...	1.126	1.112	1.155	1.250	1.248	1.192	1.158	1.177	4.69	
48)	2-Nitroaniline	0.238	0.242	0.269	0.302	0.304	0.294	0.286	0.276	9.96	
49)	Acenaphthylene	1.794	1.715	1.810	1.972	1.971	1.889	1.845	1.857	5.10	
50)	Dimethylphthalate	1.444	1.377	1.437	1.532	1.543	1.470	1.429	1.462	4.02	
51)	2,6-Dinitrotol...	0.266	0.271	0.299	0.327	0.331	0.316	0.309	0.303	8.50	
52) C	Acenaphthene	1.038	0.998	1.046	1.133	1.145	1.094	1.067	1.075	4.93	
53)	3-Nitroaniline	0.233	0.247	0.284	0.321	0.325	0.311	0.302	0.289	12.47	
54) P	2,4-Dinitrophenol		0.100	0.132	0.178	0.194	0.188	0.191	0.164	23.65	
55)	Dibenzofuran	1.749	1.681	1.749	1.876	1.888	1.810	1.763	1.788	4.17	
56) P	4-Nitrophenol		0.128	0.180	0.224	0.237	0.231	0.231	0.205	20.96	
57)	2,4-Dinitrotol...	0.338	0.366	0.412	0.456	0.468	0.447	0.440	0.418	11.72	
58)	Fluorene	1.387	1.343	1.431	1.558	1.573	1.500	1.453	1.463	5.84	
59)	2,3,4,6-Tetrac...	0.366	0.362	0.398	0.435	0.445	0.432	0.431	0.410	8.44	
60)	Diethylphthalate	1.372	1.301	1.380	1.433	1.451	1.376	1.328	1.377	3.86	
61)	4-Chlorophenyl...	0.741	0.716	0.770	0.857	0.871	0.836	0.836	0.804	7.58	
62)	4-Nitroaniline	0.208	0.234	0.277	0.313	0.321	0.301	0.295	0.279	15.17	
63)	Azobenzene	1.118	1.091	1.150	1.216	1.229	1.165	1.110	1.154	4.59	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.088	0.110	0.135	0.142	0.138	0.137	0.125	17.16	
66) c	n-Nitrosodiphe...	0.580	0.572	0.596	0.655	0.657	0.629	0.610	0.614	5.56	
67)	4-Bromophenyl-...	0.223	0.216	0.229	0.258	0.258	0.256	0.253	0.242	7.60	
68)	Hexachlorobenzene	0.265	0.251	0.264	0.294	0.295	0.291	0.291	0.279	6.49	
69)	Atrazine	0.196	0.198	0.214	0.236	0.242	0.234	0.230	0.222	8.47	
70) C	Pentachlorophenol		0.117	0.141	0.167	0.172	0.170	0.174	0.157	14.52	
71)	Phenanthrene	1.075	1.035	1.083	1.187	1.212	1.173	1.132	1.128	5.84	
72)	Anthracene	1.062	1.038	1.096	1.214	1.236	1.191	1.154	1.142	6.78	
73)	Carbazole	0.909	0.900	0.960	1.054	1.075	1.038	0.999	0.991	7.05	
74)	Di-n-butylphth...	1.106	1.082	1.151	1.246	1.277	1.231	1.165	1.180	6.25	
75) C	Fluoranthene	1.178	1.142	1.250	1.410	1.455	1.425	1.411	1.324	9.86	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine		0.560	0.646	0.772	0.769	0.765	0.751	0.710	12.36	
78)	Pyrene	1.290	1.255	1.332	1.518	1.520	1.465	1.453	1.404	7.83	
79) S	Terphenyl-d14	1.176	1.189	1.332	1.524	1.541	1.458	1.239	1.352	11.58	
80)	Butylbenzylphth...	0.489	0.479	0.515	0.562	0.568	0.545	0.524	0.526	6.53	
81)	Benzo(a)anthra...	1.280	1.231	1.321	1.478	1.479	1.442	1.408	1.377	7.24	
82)	3,3'-Dichlorob...	0.434	0.432	0.483	0.580	0.597	0.583	0.581	0.527	14.14	
83)	Chrysene	1.208	1.170	1.214	1.342	1.369	1.325	1.296	1.275	6.03	
84)	Bis(2-ethylhex...	0.726	0.725	0.775	0.843	0.852	0.814	0.765	0.786	6.61	
85) c	Di-n-octyl pht...	1.221	1.197	1.267	1.366	1.404	1.341	1.275	1.296	5.92	

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM042825.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.301	1.293	1.404	1.602	1.646	1.591	1.595	1.490	10.27
88)	Benzo(b)fluora...	1.128	1.149	1.209	1.406	1.415	1.408	1.390	1.301	10.16
89)	Benzo(k)fluora...	1.219	1.150	1.251	1.387	1.449	1.383	1.372	1.316	8.29
90) C	Benzo(a)pyrene	1.094	1.066	1.147	1.295	1.330	1.298	1.297	1.218	9.15
91)	Dibenzo(a,h)an...	1.056	1.044	1.142	1.300	1.347	1.305	1.310	1.215	10.72
92)	Benzo(g,h,i)pe...	1.066	1.030	1.105	1.236	1.268	1.224	1.212	1.163	8.07

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1870 SAS No.: Q1870 SDG No.: Q1870

Instrument ID: BNA_M Calibration Date/Time: 04/29/2025 09:13

Lab File ID: BM050035.D Init. Calib. Date(s): 04/28/2025 04/28/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:30 17:04

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.142	1.236		8.2	
Phenol-d6	1.436	1.548		7.8	
Nitrobenzene-d5	0.406	0.443		9.1	
2,4-Dimethylphenol	0.297	0.320		7.7	
2-Fluorobiphenyl	1.691	1.844		9.0	
2,4,6-Tribromophenol	0.296	0.317		7.1	
Terphenyl-d14	1.352	1.558		15.2	

All other compounds must meet a minimum RRF of 0.010.