

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1870

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270-Modified **DataFile:** BN036939.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167728BS	Acenaphthylene	0.4	0.37	ug/L	93				60	119	