

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM

SAS No.: Q1870 SDG NO.: Q1870

Lab File ID: BN036932.D

DFTPP Injection Date: 04/29/2025

Instrument ID: BNA_N

DFTPP Injection Time: 08:38

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	70.2
68	Less than 2.0% of mass 69	0.8 (1.3) 1
69	Mass 69 relative abundance	59.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	54.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	23.1
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	7.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.9 (21.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036933.D	04/29/2025	09:17
PB167728BL	PB167728BL	BN036934.D	04/29/2025	09:53
38072-010925	Q1870-01	BN036936.D	04/29/2025	13:05
38072-010925DL	Q1870-01DL	BN036937.D	04/29/2025	13:41
38072-010925DL2	Q1870-01DL2	BN036938.D	04/29/2025	14:17
PB167728BS	PB167728BS	BN036939.D	04/29/2025	14:54