

DATA PACKAGE

GC SEMI-VOLATILES

PROJECT NAME : NJ SOIL PT

ALLIANCE TECHNICAL GROUP, LLC - NEWARK

284 Sheffiled Stree

Suite 1

Mountainside, NJ - 07092

Phone No: 908-789-8900

ORDER ID : Q1872

ATTENTION : Mohammad Ahmed



Laboratory Certification ID # 20012



1) PESTICIDE GROUP2 Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	13
9) QC Data Summary For PESTICIDE Group2	14
9.1) Deuterated Monitoring Compound Summary	15
9.2) LCS/LCSD Summary	16
9.3) Method Blank Summary	17
10) Sample Data	18
10.1) HW0425-PT-CHLR-SOIL	19
11) Calibration Data Summary	26
11.1) Initial Calibration Data	27
11.1.1) PL052125	27
11.2) Continued Calibration Data	67
11.2.1) PL095793.D	67
11.2.2) PL095799.D	77
11.3) Analytical Seq	87
12) Compound Detection Summary	89
13) QC Sample Data	91
13.1) Method Blank Data	92
13.2) PIBLK Data	96
13.3) LCS Data	108
14) Manual Integration	115
15) Analytical Runlogs	119
16) Percent Solid	127
17) Extraction Logs	137
17.1) PB168152.pdf	137
17.2) PB168152IC.pdf	139
18) Standard Prep Logs	140
19) Shipping Document	197
19.1) Chain Of Custody	198

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Cover Page

Order ID : Q1872

Project ID : NJ Soil PT

Client : Alliance Technical Group, LLC - Newark

Lab Sample Number

Q1872-01
Q1872-02
Q1872-03
Q1872-04
Q1872-05
Q1872-06
Q1872-07
Q1872-08
Q1872-09
Q1872-10
Q1872-11
Q1872-12
Q1872-13
Q1872-14
Q1872-15
Q1872-16
Q1872-17
Q1872-18
Q1872-19
Q1872-20
Q1872-21
Q1872-22
Q1872-23
Q1872-24
Q1872-25

Client Sample Number

HW0425-PT-AN-SOIL
HW0425-PT-CORR-SOIL
HW0425-PT-CN-SOIL
HW0425-PT-CN-SOIL
HW0425-PT-FP-SOIL
HW0425-PT-CR6-SOIL
HW0425-PT-NUT-SOIL
HW0425-PT-NUT-SOIL
HW0425-PT-OGR-SOIL
HW0425-PT-MET-SOIL
HW0425-PT-BNA-SOIL
HW0425-PT-TRIAZINE-SOIL
HW0425-PT-PAH-SOIL
HW0425-PT-DIES-SOIL
HW0425-PT-GAS-SOIL
HW0425-PT-NJEPH-SOIL
HW0425-PT-HERB-SOIL
HW0425-PT-PCB-SOIL
HW0425-PT-PCBO-SOIL
HW0425-PT-PEST-SOIL
HW0425-PT-CHLR-SOIL
HW0425-PT-TXP-SOIL
HW0425-PT-VOA-SOIL
HW0425-PT-SOL-SOIL
HW0425-PT-NO2-SOIL

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 6/24/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Alliance Technical Group, LLC - Newark

Project Name: NJ Soil PT

Project # N/A

Order ID # Q1872

Test Name: PESTICIDE Group2

A. Number of Samples and Date of Receipt:

24 Solid samples were received on 04/24/2025.

1 Solid sample was received on 04/28/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:

Ammonia, Anions Group1, Anions Group2, Corrosivity, Cyanide, Diesel Range Organics, EPH, Flash Point, Gasoline Range Organics, Herbicide Group1, Hexavalent Chromium, Mercury, Metals Group3, Metals ICP-Group1, Oil and Grease, PCB, PESTICIDE Group1, PESTICIDE Group2, PESTICIDE Group3, Phosphorus, Total, SVOCMS Group1, SVOCMS Group2, SVOCMS Group3, SVOCMS Group4, SVOCMS Group5, TKN, TOC, TS and VOCMS Group1. This data package contains results for PESTICIDE Group2.

C. Analytical Techniques:

The analysis was performed on instrument ECD_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df.; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of PESTICIDE Group2s was based on method 8081B and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q1872

MATRIX: Solid

METHOD: 8081B/3541

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. The Initial Calibration met the requirements . The Continuous Calibration met the requirements .			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all samples .			✓
7. Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8. Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓
9. Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

The soil samples results are based on a dry weight basis.

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1872

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/24/2025

LAB CHRONICLE

OrderID:	Q1872	OrderDate:	4/24/2025 1:26:50 PM
Client:	Alliance Technical Group, LLC - Newark	Project:	NJ Soil PT
Contact:	Mohammad Ahmed	Location:	QA Office,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1872-14	HW0425-PT-DIES-SOIL	SOIL			04/25/25			04/28/25
			Diesel Range Organics	8015D		05/13/25	05/13/25	
Q1872-15	HW0425-PT-GAS-SOIL	SOIL			04/21/25			04/24/25
			Gasoline Range Organics	8015D			04/29/25	
Q1872-16	HW0425-PT-NJEPH-SOIL	Solid			04/21/25			04/24/25
			EPH	NJEPH		05/07/25	05/08/25	
Q1872-17	HW0425-PT-HERB-SOIL	SOIL			04/21/25			04/24/25
			Herbicide Group1	8151A		05/30/25	06/04/25	
Q1872-17DL	HW0425-PT-HERB-SOILDL	SOIL			04/21/25			04/24/25
			Herbicide Group1	8151A		05/30/25	06/04/25	
Q1872-18	HW0425-PT-PCB-SOIL	SOIL			04/21/25			04/24/25
			PCB	8082A		05/23/25	05/24/25	
Q1872-18DL	HW0425-PT-PCB-SOILDL	SOIL			04/21/25			04/24/25
			PCB	8082A		05/23/25	05/24/25	
Q1872-19	HW0425-PT-PCBO-SOIL	SOIL			04/21/25			04/24/25
			PCB	8082A		05/28/25	05/28/25	
Q1872-20	HW0425-PT-PEST-SOIL	SOIL			04/21/25			04/24/25
			PESTICIDE Group1	8081B		05/23/25	05/27/25	

LAB CHRONICLE

Q1872-20DL	HW0425-PT-PEST-SOI LDL	SOIL			04/21/25		04/24/25
			PESTICIDE Group1	8081B		05/23/25	05/27/25
Q1872-21	HW0425-PT-CHLR-SO IL	SOIL			04/21/25		04/24/25
			PESTICIDE Group2	8081B		05/23/25	05/27/25
Q1872-22	HW0425-PT-TXP-SOIL	SOIL			04/21/25		04/24/25
			PESTICIDE Group3	8081B		05/23/25	05/29/25

Hit Summary Sheet
SW-846

SDG No.: Q1872

Order ID: Q1872

Client: Alliance Technical Group, LLC - Newark

Project ID: NJ Soil PT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : HW0425-PT-CHLR-SOIL								
Q1872-21	HW0425-PT-CHLR-S SOIL		Chlordane	149		3.10	16.9	ug/kg
Total Concentration:				149.000				



QC SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Surrogate Summary

SDG No.: **Q1872**

Client: **Alliance Technical Group, LLC - Newark**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL095732.D	PIBLK-PL095732.D	Decachlorobiphenyl	1	20	18.2	91		57	171
		Tetrachloro-m-xylene	1	20	16.4	82		61	148
		Decachlorobiphenyl	2	20	17.5	88		57	171
		Tetrachloro-m-xylene	2	20	16.0	80		61	148
I.BLK-PL095791.D	PIBLK-PL095791.D	Decachlorobiphenyl	1	20	18.5	92		57	171
		Tetrachloro-m-xylene	1	20	20.3	101		61	148
		Decachlorobiphenyl	2	20	15.2	76		57	171
		Tetrachloro-m-xylene	2	20	19.6	98		61	148
PB168152BL	PB168152BL	Decachlorobiphenyl	1	20	18.9	95		20	144
		Tetrachloro-m-xylene	1	20	20.1	101		19	148
		Decachlorobiphenyl	2	20	18.0	90		20	144
		Tetrachloro-m-xylene	2	20	18.6	93		19	148
PB168152BS	PB168152BS	Decachlorobiphenyl	1	20	18.5	93		20	144
		Tetrachloro-m-xylene	1	20	19.7	99		19	148
		Decachlorobiphenyl	2	20	18.5	92		20	144
		Tetrachloro-m-xylene	2	20	22.6	113		19	148
Q1872-21	HW0425-PT-CHLR-SOIL	Decachlorobiphenyl	1	20	17.3	86		20	144
		Tetrachloro-m-xylene	1	20	17.6	88		19	148
		Decachlorobiphenyl	2	20	16.3	81		20	144
		Tetrachloro-m-xylene	2	20	16.8	84		19	148
I.BLK-PL095797.D	PIBLK-PL095797.D	Decachlorobiphenyl	1	20	20.2	101		57	171
		Tetrachloro-m-xylene	1	20	20.3	101		61	148
		Decachlorobiphenyl	2	20	18.4	92		57	171
		Tetrachloro-m-xylene	2	20	18.8	94		61	148

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q1872</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Alliance Technical Group, LLC - Newark</u>	Datafile :	<u>PL095795.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168152BS (Column 1)	Chlordane	66.6	64.8	ug/kg	97				80	120	
PB168152BS (Column 2)	Chlordane	66.6	64.1	ug/kg	96				80	120	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168152BL

Lab Name: CHEMTECH

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872

SAS No.: Q1872 SDG NO.: Q1872

Lab Sample ID: PB168152BL

Lab File ID: PL095794.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/23/2025

Date Analyzed (1): 05/27/2025

Date Analyzed (2): 05/27/2025

Time Analyzed (1): 15:19

Time Analyzed (2): 15:19

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB168152BS	PB168152BS	PL095795.D	05/27/2025	05/27/2025
HW0425-PT-CHLR-SOIL	Q1872-21	PL095796.D	05/27/2025	05/27/2025

COMMENTS: _____



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	04/21/25
Project:	NJ Soil PT	Date Received:	04/24/25
Client Sample ID:	HW0425-PT-CHLR-SOIL	SDG No.:	Q1872
Lab Sample ID:	Q1872-21	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	100 Decanted:
Sample Wt/Vol:	30.18 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095796.D	1	05/23/25 09:20	05/27/25 16:03	PB168152

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
57-74-9	Chlordane	149		3.10	16.9	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	17.3		20 - 144	86%	SPK: 20
877-09-8	Tetrachloro-m-xylene	17.6		19 - 148	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
 Data File : PL095796.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2025 16:03
 Operator : AR\AJ
 Sample : Q1872-21
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 HW0425-PT-CHLR-SOIL

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
 Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 05:23:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 06:29:30 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.576	2.880	55677219	65926751	17.646m	16.844
28) SA Decachlor...	9.106	8.059	40677841	71300467	17.263	16.299
Target Compounds						
23) Chlordane-1	4.743	3.898	83572115	87871696	469.094m	450.667m
24) Chlordane-2	5.270	4.482	84959482	93665161	442.100m	420.489
25) Chlordane-3	5.978	5.117	301.2E6	330.6E6	407.499	501.772m
26) Chlordane-4	6.061	5.182	347.2E6	271.1E6	386.289	474.939
27) Chlordane-5	6.901	6.078	60615596	103.9E6	442.631m	407.406

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095796.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 16:03
Operator : AR\AJ
Sample : Q1872-21
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
HW0425-PT-CHLR-SOIL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 05:23:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.576 min
Delta R.T.: -0.001 min
Response: 55677219
Conc: 17.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
HW0425-PT-CHLR-SOIL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.880 min
Delta R.T.: -0.003 min
Response: 65926751
Conc: 16.84 ng/ml




#23 Chlordane-1

R.T.: 4.743 min
Delta R.T.: 0.003 min
Response: 83572115
Conc: 469.09 ng/ml m



#23 Chlordane-1

R.T.: 3.898 min
Delta R.T.: -0.007 min
Response: 87871696
Conc: 450.67 ng/ml m




#24 Chlordane-2

R.T.: 5.270 min
Delta R.T.: 0.003 min
Response: 84959482
Conc: 442.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
HW0425-PT-CHLR-SOIL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#24 Chlordane-2

R.T.: 4.482 min
Delta R.T.: -0.006 min
Response: 93665161
Conc: 420.49 ng/ml




#25 Chlordane-3

R.T.: 5.978 min
Delta R.T.: 0.005 min
Response: 301153201
Conc: 407.50 ng/ml



#25 Chlordane-3

R.T.: 5.117 min
Delta R.T.: -0.006 min
Response: 330619441
Conc: 501.77 ng/ml m



#26 Chlordane-4

R.T.: 6.061 min
Delta R.T.: 0.003 min
Response: 347192639
Conc: 386.29 ng/ml

Instrument :


ECD_L

ClientSampleId :

HW0425-PT-CHLR-SOIL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#26 Chlordane-4

R.T.: 5.182 min
Delta R.T.: -0.005 min
Response: 271146691
Conc: 474.94 ng/ml




#27 Chlordane-5

R.T.: 6.901 min
Delta R.T.: 0.004 min
Response: 60615596
Conc: 442.63 ng/ml m



#27 Chlordane-5

R.T.: 6.078 min
Delta R.T.: -0.005 min
Response: 103861473
Conc: 407.41 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.106 min
Delta R.T.: -0.003 min
Response: 40677841
Conc: 17.26 ng/ml

Instrument :

ECD_L

ClientSampleId :


HW0425-PT-CHLR-SOIL

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.005 min
Response: 71300467
Conc: 16.30 ng/ml



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1872 **SAS No.:** Q1872 **SDG NO.:** Q1872
Instrument ID: ECD_L **Calibration Date(s):** 05/21/2025 05/21/2025
Calibration Times: 12:42 13:37

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL095740.D</u>	RT 750 = <u>PL095741.D</u>
RT 500 = <u>PL095742.D</u>	RT 250 = <u>PL095743.D</u>	RT 050 = <u>PL095744.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Chlordane-1	(1)	4.74	4.74	4.74	4.74	4.74	4.74	4.64	4.84
Chlordane-2	(2)	5.27	5.27	5.27	5.27	5.27	5.27	5.17	5.37
Chlordane-3	(3)	5.97	5.97	5.97	5.97	5.97	5.97	5.87	6.07
Chlordane-4	(4)	6.06	6.06	6.06	6.06	6.06	6.06	5.96	6.16
Chlordane-5	(5)	6.90	6.90	6.90	6.90	6.90	6.90	6.80	7.00
Decachlorobiphenyl		9.10	9.10	9.10	9.10	9.10	9.10	9.00	9.20
Tetrachloro-m-xylene		3.57	3.57	3.57	3.57	3.57	3.57	3.47	3.67

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1872 **SAS No.:** Q1872 **SDG NO.:** Q1872
Instrument ID: ECD_L **Calibration Date(s):** 05/21/2025 05/21/2025
Calibration Times: 12:42 13:37

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL095740.D</u>	RT 750 = <u>PL095741.D</u>
RT 500 = <u>PL095742.D</u>	RT 250 = <u>PL095743.D</u>	RT 050 = <u>PL095744.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Chlordane-1	(1)	3.91	3.91	3.91	3.91	3.91	3.91	3.81	4.01
Chlordane-2	(2)	4.49	4.49	4.49	4.49	4.49	4.49	4.39	4.59
Chlordane-3	(3)	5.12	5.12	5.12	5.12	5.12	5.12	5.02	5.22
Chlordane-4	(4)	5.19	5.19	5.19	5.19	5.19	5.19	5.09	5.29
Chlordane-5	(5)	6.08	6.08	6.08	6.08	6.08	6.08	5.98	6.18
Decachlorobiphenyl		8.06	8.06	8.06	8.06	8.06	8.06	7.96	8.16
Tetrachloro-m-xylene		2.89	2.89	2.89	2.89	2.89	2.89	2.79	2.99

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1872 **SAS No.:** Q1872 **SDG NO.:** Q1872
Instrument ID: ECD_L
Calibration Date(s): 05/21/2025 05/21/2025
Calibration Times: 12:42 13:37
GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:							CF 1000 =	<u>PL095740.D</u>	CF 750 =	<u>PL095741.D</u>
CF 500 =		<u>PL095742.D</u>	CF 250 =		<u>PL095743.D</u>	CF 050 =		<u>PL095744.D</u>		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD		
Chlordane-1	(1)	176910000	176577000	175163000	172758000	189374000	178156000	4		
Chlordane-2	(2)	184769000	182015000	186705000	194433000	212941000	192172000	7		
Chlordane-3	(3)	765698000	734986000	729383000	718726000	746347000	739028000	2		
Chlordane-4	(4)	924047000	889161000	887825000	885823000	907091000	898789000	2		
Chlordane-5	(5)	136544000	132699000	133478000	134176000	147822000	136944000	5		
Decachlorobiphenyl		229216000	222071000	225835000	232559000	249544000	231845000	5		
Tetrachloro-m-xylene		311690000	301590000	302575000	306250000	343868000	313195000	6		

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ALLI03
Lab Code: CHEM **Case No.:** Q1872 **SAS No.:** Q1872 **SDG NO.:** Q1872
Instrument ID: ECD_L **Calibration Date(s):** 05/21/2025 05/21/2025
Calibration Times: 12:42 13:37
GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PL095740.D</u>		CF 750 = <u>PL095741.D</u>				
CF 500 = <u>PL095742.D</u>		CF 250 = <u>PL095743.D</u>		CF 050 = <u>PL095744.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Chlordane-1	(1)	197013000	191132000	190212000	190463000	206086000	194981000	3
Chlordane-2	(2)	216789000	213122000	216767000	220165000	246920000	222753000	6
Chlordane-3	(3)	672511000	657526000	652504000	643072000	668907000	658904000	2
Chlordane-4	(4)	576338000	566663000	564939000	560162000	586441000	570909000	2
Chlordane-5	(5)	258188000	250497000	249456000	248463000	268064000	254934000	3
Decachlorobiphenyl		447807000	437656000	442922000	461550000	440042000	445995000	2
Tetrachloro-m-xylene		513344000	499514000	498617000	495411000	528313000	507040000	3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 12:42
Operator : AR\AJ
Sample : PCHLORICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:53:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:52:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.885	311.7E6	513.3E6	103.013	102.954
28) SA Decachlor...	9.102	8.063	229.2E6	447.8E6	101.497	101.103
Target Compounds						
23) Chlordane-1	4.740	3.906	176.9E6	197.0E6	1009.976	1035.759
24) Chlordane-2	5.267	4.487	184.8E6	216.8E6	989.631	1000.100
25) Chlordane-3	5.974	5.123	765.7E6	672.5E6	1049.789	1030.661
26) Chlordane-4	6.058	5.187	924.0E6	576.3E6	1040.798	1020.178
27) Chlordane-5	6.899	6.082	136.5E6	258.2E6	1022.973	1035.006



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 12:42
Operator : AR\AJ
Sample : PCHLORICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:53:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:52:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.573 min
Delta R.T.: 0.000 min
Response: 311690463
Conc: 103.01 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#1 Tetrachloro-m-xylene

R.T.: 2.885 min
Delta R.T.: 0.000 min
Response: 513343917
Conc: 102.95 ng/ml




#23 Chlordane-1

R.T.: 4.740 min
Delta R.T.: 0.000 min
Response: 176910062
Conc: 1009.98 ng/ml



#23 Chlordane-1


R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 197013425
Conc: 1035.76 ng/ml



#24 Chlordane-2


R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 184768939
Conc: 989.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#24 Chlordane-2

R.T.: 4.487 min
Delta R.T.: 0.000 min
Response: 216789108
Conc: 1000.10 ng/ml




#25 Chlordane-3

R.T.: 5.974 min
Delta R.T.: 0.000 min
Response: 765697599
Conc: 1049.79 ng/ml



#25 Chlordane-3


R.T.: 5.123 min
Delta R.T.: 0.000 min
Response: 672510771
Conc: 1030.66 ng/ml



#26 Chlordane-4


R.T.: 6.058 min
Delta R.T.: 0.000 min
Response: 924046648
Conc: 1040.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000




#26 Chlordane-4

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 576338299
Conc: 1020.18 ng/ml




#27 Chlordane-5

R.T.: 6.899 min
Delta R.T.: 0.000 min
Response: 136544196
Conc: 1022.97 ng/ml



#27 Chlordane-5


R.T.: 6.082 min
Delta R.T.: 0.000 min
Response: 258188478
Conc: 1035.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.102 min
Delta R.T.: -0.002 min
Response: 229215739
Conc: 101.50 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC1000



#28 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 447806781
Conc: 101.10 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
 Data File : PL095741.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2025 12:56
 Operator : AR\AJ
 Sample : PCHLORICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 05:53:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 05:52:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.886	226.2E6	374.6E6	74.756	75.135
28) SA Decachlor...	9.103	8.063	166.6E6	328.2E6	73.750	74.108
Target Compounds						
23) Chlordane-1	4.739	3.906	132.4E6	143.3E6	756.054	753.631
24) Chlordane-2	5.267	4.488	136.5E6	159.8E6	731.159	737.389
25) Chlordane-3	5.974	5.124	551.2E6	493.1E6	755.762	755.772
26) Chlordane-4	6.059	5.187	666.9E6	425.0E6	751.128	752.289
27) Chlordane-5	6.899	6.082	99524547	187.9E6	745.626	753.131



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 12:56
Operator : AR\AJ
Sample : PCHLORICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:53:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:52:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.573 min
Delta R.T.: 0.000 min
Response: 226192140
Conc: 74.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#1 Tetrachloro-m-xylene

R.T.: 2.886 min
Delta R.T.: 0.000 min
Response: 374635193
Conc: 75.13 ng/ml




#23 Chlordane-1

R.T.: 4.739 min
Delta R.T.: 0.000 min
Response: 132432401
Conc: 756.05 ng/ml



#23 Chlordane-1


R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 143349350
Conc: 753.63 ng/ml



#24 Chlordane-2


R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 136510975
Conc: 731.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#24 Chlordane-2

R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 159841840
Conc: 737.39 ng/ml




#25 Chlordane-3

R.T.: 5.974 min
Delta R.T.: 0.000 min
Response: 551239706
Conc: 755.76 ng/ml



#25 Chlordane-3


R.T.: 5.124 min
Delta R.T.: 0.000 min
Response: 493144713
Conc: 755.77 ng/ml



#26 Chlordane-4


R.T.: 6.059 min
Delta R.T.: 0.001 min
Response: 666870453
Conc: 751.13 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750




#26 Chlordane-4

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 424997475
Conc: 752.29 ng/ml




#27 Chlordane-5

R.T.: 6.899 min
Delta R.T.: 0.000 min
Response: 99524547
Conc: 745.63 ng/ml



#27 Chlordane-5


R.T.: 6.082 min
Delta R.T.: 0.000 min
Response: 187872907
Conc: 753.13 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.103 min
Delta R.T.: 0.000 min
Response: 166553523
Conc: 73.75 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC750



#28 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 328241961
Conc: 74.11 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095742.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 13:10
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:53:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:52:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.885	151.3E6	249.3E6	50.000	50.000
28) SA Decachlor...	9.104	8.063	112.9E6	221.5E6	50.000	50.000
Target Compounds						
23) Chlordane-1	4.739	3.906	87581330	95105811	500.000	500.000
24) Chlordane-2	5.267	4.487	93352484	108.4E6	500.000	500.000
25) Chlordane-3	5.973	5.123	364.7E6	326.3E6	500.000	500.000
26) Chlordane-4	6.058	5.187	443.9E6	282.5E6	500.000	500.000
27) Chlordane-5	6.899	6.082	66738900	124.7E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095742.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 13:10
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:53:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:52:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.573 min
Delta R.T.: 0.000 min
Response: 151287484
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#1 Tetrachloro-m-xylene

R.T.: 2.885 min
Delta R.T.: 0.000 min
Response: 249308265
Conc: 50.00 ng/ml




#23 Chlordane-1

R.T.: 4.739 min
Delta R.T.: 0.000 min
Response: 87581330
Conc: 500.00 ng/ml



#23 Chlordane-1


R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 95105811
Conc: 500.00 ng/ml



#24 Chlordane-2


R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 93352484
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#24 Chlordane-2

R.T.: 4.487 min
Delta R.T.: 0.000 min
Response: 108383676
Conc: 500.00 ng/ml




#25 Chlordane-3

R.T.: 5.973 min
Delta R.T.: 0.000 min
Response: 364691357
Conc: 500.00 ng/ml



#25 Chlordane-3


R.T.: 5.123 min
Delta R.T.: 0.000 min
Response: 326252241
Conc: 500.00 ng/ml



#26 Chlordane-4


R.T.: 6.058 min
Delta R.T.: 0.000 min
Response: 443912706
Conc: 500.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500




#26 Chlordane-4

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 282469516
Conc: 500.00 ng/ml




#27 Chlordane-5

R.T.: 6.899 min
Delta R.T.: 0.000 min
Response: 66738900
Conc: 500.00 ng/ml



#27 Chlordane-5


R.T.: 6.082 min
Delta R.T.: 0.000 min
Response: 124727963
Conc: 500.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.104 min
Delta R.T.: 0.000 min
Response: 112917337
Conc: 50.00 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 221461141
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
 Data File : PL095743.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2025 13:23
 Operator : AR\AJ
 Sample : PCHLORICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 05:53:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 05:52:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.886	76562532	123.9E6	25.304	24.839
28) SA Decachlor...	9.104	8.064	58139637	115.4E6	25.744	26.051
Target Compounds						
23) Chlordane-1	4.740	3.906	43189581	47615813	246.568	250.331
24) Chlordane-2	5.267	4.488	48608220	55041238	260.348	253.918
25) Chlordane-3	5.974	5.124	179.7E6	160.8E6	246.348	246.386
26) Chlordane-4	6.059	5.187	221.5E6	140.0E6	249.436	247.886
27) Chlordane-5	6.900	6.083	33544019	62115796	251.308	249.005



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095743.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 13:23
Operator : AR\AJ
Sample : PCHLORICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:53:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:52:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.573 min
Delta R.T.: 0.000 min
Response: 76562532
Conc: 25.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#1 Tetrachloro-m-xylene

R.T.: 2.886 min
Delta R.T.: 0.000 min
Response: 123852854
Conc: 24.84 ng/ml




#23 Chlordane-1

R.T.: 4.740 min
Delta R.T.: 0.000 min
Response: 43189581
Conc: 246.57 ng/ml



#23 Chlordane-1


R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 47615813
Conc: 250.33 ng/ml



#24 Chlordane-2


R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 48608220
Conc: 260.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#24 Chlordane-2

R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 55041238
Conc: 253.92 ng/ml




#25 Chlordane-3

R.T.: 5.974 min
Delta R.T.: 0.000 min
Response: 179681621
Conc: 246.35 ng/ml



#25 Chlordane-3


R.T.: 5.124 min
Delta R.T.: 0.000 min
Response: 160767908
Conc: 246.39 ng/ml



#26 Chlordane-4


R.T.: 6.059 min
Delta R.T.: 0.001 min
Response: 221455802
Conc: 249.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250




#26 Chlordane-4

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 140040395
Conc: 247.89 ng/ml




#27 Chlordane-5

R.T.: 6.900 min
Delta R.T.: 0.000 min
Response: 33544019
Conc: 251.31 ng/ml



#27 Chlordane-5


R.T.: 6.083 min
Delta R.T.: 0.000 min
Response: 62115796
Conc: 249.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.104 min
Delta R.T.: 0.000 min
Response: 58139637
Conc: 25.74 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC250



#28 Decachlorobiphenyl

R.T.: 8.064 min
Delta R.T.: 0.000 min
Response: 115387604
Conc: 26.05 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
 Data File : PL095744.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2025 13:37
 Operator : AR\AJ
 Sample : PCHLORICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORICC050

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 05/22/2025

Supervised By :mohammad ahmed 05/23/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 05:54:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 05:52:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.885	17193396	26415668	5.682	5.298
28) SA Decachlor...	9.104	8.064	12477209	22002090	5.525	4.967
Target Compounds						
23) Chlordane-1	4.739	3.906	9468705	10304291	54.057	54.173
24) Chlordane-2	5.267	4.487	10647031	12345982	57.026	56.955
25) Chlordane-3	5.973	5.123	37317354	33445326	51.163	51.257
26) Chlordane-4	6.058	5.187	45354568	29322072	51.085	51.903
27) Chlordane-5	6.897	6.082	7391106	13403185	55.373m	53.730

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095744.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 13:37
Operator : AR\AJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

Client SampleId :

PCHLORICC050

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 05/22/2025

Supervised By :mohammad ahmed 05/23/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:54:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:52:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.573 min
Delta R.T.: 0.000 min
Response: 17193396
Conc: 5.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/22/2025
Supervised By :mohammad ahmed 05/23/2025




#1 Tetrachloro-m-xylene

R.T.: 2.885 min
Delta R.T.: 0.000 min
Response: 26415668
Conc: 5.30 ng/ml




#23 Chlordane-1

R.T.: 4.739 min
Delta R.T.: 0.000 min
Response: 9468705
Conc: 54.06 ng/ml



#23 Chlordane-1

R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 10304291
Conc: 54.17 ng/ml



#24 Chlordane-2

R.T.: 5.267 min
Delta R.T.: 0.000 min
Response: 10647031
Conc: 57.03 ng/ml

Instrument :


ECD_L

ClientSampleId :

PCHLORICC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/22/2025
Supervised By :mohammad ahmed 05/23/2025




#24 Chlordane-2

R.T.: 4.487 min
Delta R.T.: 0.000 min
Response: 12345982
Conc: 56.95 ng/ml




#25 Chlordane-3

R.T.: 5.973 min
Delta R.T.: 0.000 min
Response: 37317354
Conc: 51.16 ng/ml



#25 Chlordane-3

R.T.: 5.123 min
Delta R.T.: 0.000 min
Response: 33445326
Conc: 51.26 ng/ml




#26 Chlordane-4

R.T.: 6.058 min
Delta R.T.: 0.000 min
Response: 45354568
Conc: 51.09 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORICC050


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/22/2025
Supervised By :mohammad ahmed 05/23/2025




#26 Chlordane-4

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 29322072
Conc: 51.90 ng/ml




#27 Chlordane-5

R.T.: 6.897 min
Delta R.T.: -0.002 min
Response: 7391106
Conc: 55.37 ng/ml m



#27 Chlordane-5

R.T.: 6.082 min
Delta R.T.: 0.000 min
Response: 13403185
Conc: 53.73 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.104 min
Delta R.T.: 0.000 min
Response: 12477209
Conc: 5.52 ng/ml

Instrument :

ECD_L

ClientSampleId :


PCHLORICC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 05/22/2025

Supervised By :mohammad ahmed 05/23/2025



#28 Decachlorobiphenyl

R.T.: 8.064 min
Delta R.T.: 0.000 min
Response: 22002090
Conc: 4.97 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 15:12
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL052125CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:59:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:59:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.885	156.9E6	252.7E6	50.100	49.840
28) SA Decachlor...	9.104	8.064	115.9E6	229.0E6	50.010	51.351
Target Compounds						
23) Chlordane-1	4.740	3.906	88597152	96171041	497.300	493.232
24) Chlordane-2	5.268	4.488	96121246	109.3E6	500.182	490.592
25) Chlordane-3	5.974	5.124	377.3E6	331.4E6	510.592	502.970
26) Chlordane-4	6.059	5.187	457.2E6	287.0E6	508.677	502.673
27) Chlordane-5	6.900	6.082	69143587	126.8E6	504.904	497.202



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 15:12
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL052125CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 05:59:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 05:59:09 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.573 min
Delta R.T.: 0.000 min
Response: 156911848
Conc: 50.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL052125CHLOR




#1 Tetrachloro-m-xylene

R.T.: 2.885 min
Delta R.T.: 0.000 min
Response: 252709314
Conc: 49.84 ng/ml




#23 Chlordane-1

R.T.: 4.740 min
Delta R.T.: 0.000 min
Response: 88597152
Conc: 497.30 ng/ml



#23 Chlordane-1


R.T.: 3.906 min
Delta R.T.: 0.000 min
Response: 96171041
Conc: 493.23 ng/ml



#24 Chlordane-2


R.T.: 5.268 min
Delta R.T.: 0.001 min
Response: 96121246
Conc: 500.18 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL052125CHLOR




#24 Chlordane-2

R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 109280725
Conc: 490.59 ng/ml




#25 Chlordane-3

R.T.: 5.974 min
Delta R.T.: 0.002 min
Response: 377342003
Conc: 510.59 ng/ml



#25 Chlordane-3


R.T.: 5.124 min
Delta R.T.: 0.000 min
Response: 331408650
Conc: 502.97 ng/ml



#26 Chlordane-4


R.T.: 6.059 min
Delta R.T.: 0.002 min
Response: 457193182
Conc: 508.68 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL052125CHLOR




#26 Chlordane-4

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 286980441
Conc: 502.67 ng/ml




#27 Chlordane-5

R.T.: 6.900 min
Delta R.T.: 0.003 min
Response: 69143587
Conc: 504.90 ng/ml



#27 Chlordane-5


R.T.: 6.082 min
Delta R.T.: 0.000 min
Response: 126753637
Conc: 497.20 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.104 min
Delta R.T.: 0.000 min
Response: 115945743
Conc: 50.01 ng/ml

Instrument :
ECD_L
ClientSampleId :
ICVPL052125CHLOR



#28 Decachlorobiphenyl

R.T.: 8.064 min
Delta R.T.: 0.000 min
Response: 229025179
Conc: 51.35 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/27/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 14:52 Initial Calibration Time(s): 12:42 13:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.74	4.74	4.64	4.84	0.00
Chlordane-2	(2)	5.27	5.27	5.17	5.37	0.00
Chlordane-3	(3)	5.98	5.97	5.87	6.07	-0.01
Chlordane-4	(4)	6.06	6.06	5.96	6.16	0.00
Chlordane-5	(5)	6.90	6.90	6.80	7.00	0.00
Decachlorobiphenyl		9.11	9.10	9.00	9.20	-0.01
Tetrachloro-m-xylene		3.58	3.57	3.47	3.67	-0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/27/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 14:52 Initial Calibration Time(s): 12:42 13:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.90	3.91	3.81	4.01	0.01
Chlordane-2	(2)	4.48	4.49	4.39	4.59	0.01
Chlordane-3	(3)	5.11	5.12	5.02	5.22	0.01
Chlordane-4	(4)	5.18	5.19	5.09	5.29	0.01
Chlordane-5	(5)	6.08	6.08	5.98	6.18	0.01
Decachlorobiphenyl		8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene		2.88	2.89	2.79	2.99	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095793.D Time Analyzed: 14:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.742	4.639	4.839	496.170	500.000	-0.8
Chlordane-2	5.270	5.167	5.367	459.390	500.000	-8.1
Chlordane-3	5.977	5.873	6.073	477.620	500.000	-4.5
Chlordane-4	6.062	5.958	6.158	460.750	500.000	-7.9
Chlordane-5	6.900	6.799	6.999	487.780	500.000	-2.4
Decachlorobiphenyl	9.106	9.004	9.204	44.320	50.000	-11.4
Tetrachloro-m-xylene	3.576	3.473	3.673	49.910	50.000	-0.2



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095793.D Time Analyzed: 14:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.898	3.806	4.006	465.820	500.000	-6.8
Chlordane-2	4.480	4.387	4.587	452.380	500.000	-9.5
Chlordane-3	5.114	5.023	5.223	494.210	500.000	-1.2
Chlordane-4	5.180	5.087	5.287	449.500	500.000	-10.1
Chlordane-5	6.075	5.982	6.182	436.950	500.000	-12.6
Decachlorobiphenyl	8.058	7.963	8.163	40.610	50.000	-18.8
Tetrachloro-m-xylene	2.878	2.785	2.985	59.730	50.000	19.5

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
 Data File : PL095793.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2025 14:52
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PCHLORCCC500

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 05/29/2025
 Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 15:13:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 06:29:30 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.576	2.878	157.5E6	233.8E6	49.913	59.732
28) SA Decachlor...	9.106	8.058	104.4E6	177.6E6	44.323	40.608
Target Compounds						
23) Chlordane-1	4.742	3.898	88395157	90825351	496.166m	465.816
24) Chlordane-2	5.270	4.480	88281335	100.8E6	459.386m	452.383
25) Chlordane-3	5.977	5.114	353.0E6	325.6E6	477.615	494.208m
26) Chlordane-4	6.062	5.180	414.1E6	256.6E6	460.752	449.495
27) Chlordane-5	6.900	6.075	66797948	111.4E6	487.776m	436.947

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095793.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 14:52
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORCCC500

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 15:13:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.576 min
Delta R.T.: -0.001 min
Response: 157488396
Conc: 49.91 ng/ml

Instrument :

ECD_L

ClientSampleId :


PCHLORCCC500

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.878 min
Delta R.T.: -0.005 min
Response: 233786438
Conc: 59.73 ng/ml




#23 Chlordane-1

R.T.: 4.742 min
Delta R.T.: 0.002 min
Response: 88395157
Conc: 496.17 ng/ml m



#23 Chlordane-1

R.T.: 3.898 min
Delta R.T.: -0.007 min
Response: 90825351
Conc: 465.82 ng/ml




#24 Chlordane-2

R.T.: 5.270 min
Delta R.T.: 0.003 min
Response: 88281335
Conc: 459.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#24 Chlordane-2

R.T.: 4.480 min
Delta R.T.: -0.007 min
Response: 100769646
Conc: 452.38 ng/ml




#25 Chlordane-3

R.T.: 5.977 min
Delta R.T.: 0.004 min
Response: 352970939
Conc: 477.62 ng/ml



#25 Chlordane-3

R.T.: 5.114 min
Delta R.T.: -0.009 min
Response: 325635663
Conc: 494.21 ng/ml m



#26 Chlordane-4

R.T.: 6.062 min
Delta R.T.: 0.004 min
Response: 414119347
Conc: 460.75 ng/ml

Instrument :


ECD_L

ClientSampleId :

PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#26 Chlordane-4

R.T.: 5.180 min
Delta R.T.: -0.007 min
Response: 256620899
Conc: 449.50 ng/ml




#27 Chlordane-5

R.T.: 6.900 min
Delta R.T.: 0.003 min
Response: 66797948
Conc: 487.78 ng/ml m



#27 Chlordane-5

R.T.: 6.075 min
Delta R.T.: -0.007 min
Response: 111392392
Conc: 436.95 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.106 min
Delta R.T.: -0.003 min
Response: 104439753
Conc: 44.32 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025



#28 Decachlorobiphenyl

R.T.: 8.058 min
Delta R.T.: -0.006 min
Response: 177637522
Conc: 40.61 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/27/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 17:00 Initial Calibration Time(s): 12:42 13:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.74	4.74	4.64	4.84	0.00
Chlordane-2	(2)	5.27	5.27	5.17	5.37	0.00
Chlordane-3	(3)	5.97	5.97	5.87	6.07	0.00
Chlordane-4	(4)	6.06	6.06	5.96	6.16	0.00
Chlordane-5	(5)	6.90	6.90	6.80	7.00	0.00
Decachlorobiphenyl		9.10	9.10	9.00	9.20	0.00
Tetrachloro-m-xylene		3.57	3.57	3.47	3.67	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/27/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 17:00 Initial Calibration Time(s): 12:42 13:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Chlordane-1 (1)	3.90	3.91	3.81	4.01	0.01
Chlordane-2 (2)	4.48	4.49	4.39	4.59	0.01
Chlordane-3 (3)	5.12	5.12	5.02	5.22	0.00
Chlordane-4 (4)	5.18	5.19	5.09	5.29	0.01
Chlordane-5 (5)	6.08	6.08	5.98	6.18	0.00
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.88	2.89	2.79	2.99	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095799.D Time Analyzed: 17:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.739	4.639	4.839	480.010	500.000	-4.0
Chlordane-2	5.266	5.167	5.367	474.300	500.000	-5.1
Chlordane-3	5.974	5.873	6.073	464.240	500.000	-7.2
Chlordane-4	6.058	5.958	6.158	448.160	500.000	-10.4
Chlordane-5	6.898	6.799	6.999	499.310	500.000	-0.1
Decachlorobiphenyl	9.101	9.004	9.204	43.590	50.000	-12.8
Tetrachloro-m-xylene	3.574	3.473	3.673	47.850	50.000	-4.3



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095799.D Time Analyzed: 17:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.902	3.806	4.006	471.780	500.000	-5.6
Chlordane-2	4.484	4.387	4.587	469.300	500.000	-6.1
Chlordane-3	5.118	5.023	5.223	471.260	500.000	-5.7
Chlordane-4	5.183	5.087	5.287	465.600	500.000	-6.9
Chlordane-5	6.078	5.982	6.182	452.350	500.000	-9.5
Decachlorobiphenyl	8.059	7.963	8.163	43.740	50.000	-12.5
Tetrachloro-m-xylene	2.882	2.785	2.985	59.960	50.000	19.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
 Data File : PL095799.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2025 17:00
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORCCC500

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 05:27:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 06:29:30 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.574	2.882	151.0E6	234.7E6	47.845	59.957 #
28) SA Decachlor...	9.101	8.059	102.7E6	191.4E6	43.591	43.745
Target Compounds						
23) Chlordane-1	4.739	3.902	85517490	91988367	480.014m	471.780
24) Chlordane-2	5.266	4.484	91147675	104.5E6	474.302m	469.297
25) Chlordane-3	5.974	5.118	343.1E6	310.5E6	464.242	471.259m
26) Chlordane-4	6.058	5.183	402.8E6	265.8E6	448.161	465.605
27) Chlordane-5	6.898	6.078	68377121	115.3E6	499.307m	452.350

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095799.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 17:00
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PCHLORCCC500

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 05:27:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.574 min
Delta R.T.: -0.003 min
Response: 150965109
Conc: 47.85 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.882 min
Delta R.T.: -0.001 min
Response: 234667739
Conc: 59.96 ng/ml




#23 Chlordane-1

R.T.: 4.739 min
Delta R.T.: 0.000 min
Response: 85517490
Conc: 480.01 ng/ml m



#23 Chlordane-1

R.T.: 3.902 min
Delta R.T.: -0.004 min
Response: 91988367
Conc: 471.78 ng/ml




#24 Chlordane-2

R.T.: 5.266 min
Delta R.T.: 0.000 min
Response: 91147675
Conc: 474.30 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#24 Chlordane-2

R.T.: 4.484 min
Delta R.T.: -0.003 min
Response: 104537158
Conc: 469.30 ng/ml




#25 Chlordane-3

R.T.: 5.974 min
Delta R.T.: 0.002 min
Response: 343088094
Conc: 464.24 ng/ml



#25 Chlordane-3

R.T.: 5.118 min
Delta R.T.: -0.005 min
Response: 310514430
Conc: 471.26 ng/ml m




#26 Chlordane-4

R.T.: 6.058 min
Delta R.T.: 0.000 min
Response: 402802651
Conc: 448.16 ng/ml

Instrument :
ECD_L
ClientSampleId :
PCHLORCCC500


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#26 Chlordane-4

R.T.: 5.183 min
Delta R.T.: -0.004 min
Response: 265817712
Conc: 465.60 ng/ml




#27 Chlordane-5

R.T.: 6.898 min
Delta R.T.: 0.000 min
Response: 68377121
Conc: 499.31 ng/ml m



#27 Chlordane-5

R.T.: 6.078 min
Delta R.T.: -0.004 min
Response: 115319383
Conc: 452.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.101 min
Delta R.T.: -0.008 min
Response: 102713933
Conc: 43.59 ng/ml

Instrument :

ECD_L

ClientSampleId :


PCHLORCCC500

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.004 min
Response: 191360545
Conc: 43.74 ng/ml

Analytical Sequence

Client: Alliance Technical Group, LLC - Newark	SDG No.: Q1872
Project: NJ Soil PT	Instrument ID: ECD_L
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 05/21/2025 05/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	05/21/2025	10:47	PL095732.D	9.10	3.57
PCHLORICC1000	PCHLORICC1000	05/21/2025	12:42	PL095740.D	9.10	3.57
PCHLORICC750	PCHLORICC750	05/21/2025	12:56	PL095741.D	9.10	3.57
PCHLORICC500	PCHLORICC500	05/21/2025	13:10	PL095742.D	9.10	3.57
PCHLORICC250	PCHLORICC250	05/21/2025	13:23	PL095743.D	9.10	3.57
PCHLORICC050	PCHLORICC050	05/21/2025	13:37	PL095744.D	9.10	3.57
IBLK	IBLK	05/27/2025	12:42	PL095791.D	9.10	3.58
PCHLORCCC500	PCHLORCCC500	05/27/2025	14:52	PL095793.D	9.11	3.58
PB168152BL	PB168152BL	05/27/2025	15:19	PL095794.D	9.11	3.58
PB168152BS	PB168152BS	05/27/2025	15:33	PL095795.D	9.10	3.57
HW0425-PT-CHLR-SOIL	Q1872-21	05/27/2025	16:03	PL095796.D	9.11	3.58
IBLK	IBLK	05/27/2025	16:30	PL095797.D	9.11	3.58
PCHLORCCC500	PCHLORCCC500	05/27/2025	17:00	PL095799.D	9.10	3.57

Analytical Sequence

Client: Alliance Technical Group, LLC - Newark	SDG No.: Q1872
Project: NJ Soil PT	Instrument ID: ECD_L
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 05/21/2025 05/21/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	05/21/2025	10:47	PL095732.D	8.06	2.89
PCHLORICC1000	PCHLORICC1000	05/21/2025	12:42	PL095740.D	8.06	2.89
PCHLORICC750	PCHLORICC750	05/21/2025	12:56	PL095741.D	8.06	2.89
PCHLORICC500	PCHLORICC500	05/21/2025	13:10	PL095742.D	8.06	2.89
PCHLORICC250	PCHLORICC250	05/21/2025	13:23	PL095743.D	8.06	2.89
PCHLORICC050	PCHLORICC050	05/21/2025	13:37	PL095744.D	8.06	2.89
IBLK	IBLK	05/27/2025	12:42	PL095791.D	8.06	2.88
PCHLORCCC500	PCHLORCCC500	05/27/2025	14:52	PL095793.D	8.06	2.88
PB168152BL	PB168152BL	05/27/2025	15:19	PL095794.D	8.06	2.88
PB168152BS	PB168152BS	05/27/2025	15:33	PL095795.D	8.06	2.88
HW0425-PT-CHLR-SOIL	Q1872-21	05/27/2025	16:03	PL095796.D	8.06	2.88
IBLK	IBLK	05/27/2025	16:30	PL095797.D	8.06	2.88
PCHLORCCC500	PCHLORCCC500	05/27/2025	17:00	PL095799.D	8.06	2.88

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

HW0425-PT-CHLR-SOIL

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872

SAS No.: Q1872 SDG NO.: Q1872

Lab Sample ID: Q1872-21

Date(s) Analyzed: 05/27/2025 05/27/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column:(2): ZB-MR2 ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Chlordane	1	5.79	5.74	5.84	142	4.8
	2	4.95	4.90	5.00	149	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB168152BS

Contract: ALLI03

Lab Code: CHEM

Case No.: Q1872

SAS No.: Q1872

SDG NO.: Q1872

Lab Sample ID: PB168152BS

Date(s) Analyzed: 05/27/2025

05/27/2025

Instrument ID (1): ECD_L

Instrument ID (2): ECD_L

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Chlordane	1	5.79	5.74	5.84	64.8	1.1
	2	4.95	4.90	5.00	64.1	



QC SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark		Date Collected:	
Project:	NJ Soil PT		Date Received:	
Client Sample ID:	PB168152BL		SDG No.:	Q1872
Lab Sample ID:	PB168152BL		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	100
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group2
Extraction Type:			Injection Volume :	
GPC Factor :	1.0	PH :		
Prep Method :	SW3541B			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095794.D	1	05/23/25 09:20	05/27/25 15:19	PB168152

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
57-74-9	Chlordane	3.10	U	3.10	17.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.9		20 - 144	95%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.1		19 - 148	101%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095794.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 15:19
Operator : AR\AJ
Sample : PB168152BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB168152BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 05:21:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.577	2.881	63413058	72777626	20.098	18.594
28) SA Decachlor...	9.105	8.059	44546178	78722365	18.905	17.996

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095794.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 15:19
Operator : AR\AJ
Sample : PB168152BL
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB168152BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 05:21:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.577 min
Delta R.T.: 0.000 min
Response: 63413058
Conc: 20.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB168152BL




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: -0.003 min
Response: 72777626
Conc: 18.59 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.105 min
Delta R.T.: -0.004 min
Response: 44546178
Conc: 18.90 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.005 min
Response: 78722365
Conc: 18.00 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	05/21/25
Project:	NJ Soil PT	Date Received:	05/21/25
Client Sample ID:	PIBLK-PL095732.D	SDG No.:	Q1872
Lab Sample ID:	I.BLK-PL095732.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095732.D	1		05/21/25	PL052125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.2		57 - 171	91%	SPK: 20
877-09-8	Tetrachloro-m-xylene	16.4		61 - 148	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095732.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 10:47
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 06:31:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.573	2.886	51602531	62467417	16.354	15.960
28) SA Decachlor...	9.102	8.063	42816238	76669642	18.171	17.527

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052125\
Data File : PL095732.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2025 10:47
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 06:31:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.573 min
Delta R.T.: -0.004 min
Response: 51602531
Conc: 16.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.886 min
Delta R.T.: 0.002 min
Response: 62467417
Conc: 15.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.102 min
Delta R.T.: -0.006 min
Response: 42816238
Conc: 18.17 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 76669642
Conc: 17.53 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	05/27/25
Project:	NJ Soil PT	Date Received:	05/27/25
Client Sample ID:	PIBLK-PL095791.D	SDG No.:	Q1872
Lab Sample ID:	I.BLK-PL095791.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0 Decanted:
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095791.D	1		05/27/25	PL052725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.5		57 - 171	92%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.3		61 - 148	101%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
 Data File : PL095791.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2025 12:42
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 13:29:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 06:29:30 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.576	2.881	63910773	76769808	20.255	19.614
28) SA Decachlor...	9.104	8.058	43491430	66494116	18.457	15.201

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095791.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 12:42
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 13:29:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.576 min
Delta R.T.: 0.000 min
Response: 63910773
Conc: 20.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: -0.003 min
Response: 76769808
Conc: 19.61 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.104 min
Delta R.T.: -0.005 min
Response: 43491430
Conc: 18.46 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.058 min
Delta R.T.: -0.006 min
Response: 66494116
Conc: 15.20 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	05/27/25
Project:	NJ Soil PT	Date Received:	05/27/25
Client Sample ID:	PIBLK-PL095797.D	SDG No.:	Q1872
Lab Sample ID:	I.BLK-PL095797.D	Matrix:	WATER
Analytical Method:	8081B	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PESTICIDE Group2
GPC Factor :	1.0	PH :	
Prep Method :	3510C	Decanted:	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095797.D	1		05/27/25	PL052725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
57-74-9	Chlordane	0.088	U	0.088	0.50	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	20.2		57 - 171	101%	SPK: 20
877-09-8	Tetrachloro-m-xylene	20.3		61 - 148	101%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095797.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 16:30
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 05:24:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.576	2.881	63912302	73568311	20.256	18.797
28) SA Decachlor...	9.106	8.059	47678170	80528867	20.234	18.409

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095797.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 16:30
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 05:24:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.576 min
Delta R.T.: 0.000 min
Response: 63912302
Conc: 20.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.881 min
Delta R.T.: -0.003 min
Response: 73568311
Conc: 18.80 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.106 min
Delta R.T.: -0.003 min
Response: 47678170
Conc: 20.23 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.005 min
Response: 80528867
Conc: 18.41 ng/ml

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ Soil PT	Date Received:	
Client Sample ID:	PB168152BS	SDG No.:	Q1872
Lab Sample ID:	PB168152BS	Matrix:	SOIL
Analytical Method:	8081B	% Solid:	100 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PESTICIDE Group2
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL095795.D	1	05/23/25 09:20	05/27/25 15:33	PB168152

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
57-74-9	Chlordane	64.8		3.10	17.0	ug/kg
SURROGATES						
2051-24-3	Decachlorobiphenyl	18.5		20 - 144	93%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.6		19 - 148	113%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
 Data File : PL095795.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2025 15:33
 Operator : AR\AJ
 Sample : PB168152BS
 Misc : CHLOR/BS
 ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PB168152BS

Manual Integrations**APPROVED**

Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 05:21:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 06:29:30 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.572	2.883	62144005	88547959	19.695	22.624
28) SA Decachlor...	9.098	8.059	43665979	80872446	18.531	18.487
Target Compounds						
23) Chlordane-1	4.736	3.902	36200834	37637908	203.197m	193.033
24) Chlordane-2	5.264	4.484	38234657	43769053	198.960m	196.492
25) Chlordane-3	5.972	5.118	139.6E6	138.3E6	188.886	209.864m
26) Chlordane-4	6.057	5.183	166.2E6	104.9E6	184.894	183.754
27) Chlordane-5	6.896	6.078	26967527	45571349	196.924m	178.758

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052725\
Data File : PL095795.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 15:33
Operator : AR\AJ
Sample : PB168152BS
Misc : CHLOR/BS
ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PB168152BS

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 05/29/2025

Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 05:21:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm





#1 Tetrachloro-m-xylene

R.T.: 3.572 min
Delta R.T.: -0.005 min
Response: 62144005
Conc: 19.70 ng/ml

Instrument :


ECD_L

ClientSampleId :

PB168152BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#1 Tetrachloro-m-xylene

R.T.: 2.883 min
Delta R.T.: 0.000 min
Response: 88547959
Conc: 22.62 ng/ml




#23 Chlordane-1

R.T.: 4.736 min
Delta R.T.: -0.003 min
Response: 36200834
Conc: 203.20 ng/ml m



#23 Chlordane-1

R.T.: 3.902 min
Delta R.T.: -0.003 min
Response: 37637908
Conc: 193.03 ng/ml



#24 Chlordane-2

R.T.: 5.264 min
Delta R.T.: -0.003 min
Response: 38234657
Conc: 198.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB168152BS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025

#24 Chlordane-2


R.T.: 4.484 min
Delta R.T.: -0.003 min
Response: 43769053
Conc: 196.49 ng/ml

#25 Chlordane-3

R.T.: 5.972 min
Delta R.T.: 0.000 min
Response: 139592391
Conc: 188.89 ng/ml

#25 Chlordane-3

R.T.: 5.118 min
Delta R.T.: -0.005 min
Response: 138280342
Conc: 209.86 ng/ml m




#26 Chlordane-4

R.T.: 6.057 min
Delta R.T.: 0.000 min
Response: 166180416
Conc: 184.89 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB168152BS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025




#26 Chlordane-4

R.T.: 5.183 min
Delta R.T.: -0.004 min
Response: 104906995
Conc: 183.75 ng/ml




#27 Chlordane-5

R.T.: 6.896 min
Delta R.T.: -0.001 min
Response: 26967527
Conc: 196.92 ng/ml m



#27 Chlordane-5

R.T.: 6.078 min
Delta R.T.: -0.004 min
Response: 45571349
Conc: 178.76 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.098 min
Delta R.T.: -0.011 min
Response: 43665979
Conc: 18.53 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB168152BS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/29/2025
Supervised By :mohammad ahmed 05/30/2025



#28 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.005 min
Response: 80872446
Conc: 18.49 ng/ml

Manual Integration Report

Sequence:	PL052125	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL095733.D	4,4"-DDD	Abdul	5/22/2025 8:03:45 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
PEM	PL095733.D	4,4"-DDD #2	Abdul	5/22/2025 8:03:45 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
PEM	PL095733.D	beta-BHC #2	Abdul	5/22/2025 8:03:45 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
PEM	PL095733.D	Endrin aldehyde	Abdul	5/22/2025 8:03:45 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
PEM	PL095733.D	Endrin aldehyde #2	Abdul	5/22/2025 8:03:45 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
PEM	PL095733.D	Endrin ketone	Abdul	5/22/2025 8:03:45 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
PEM	PL095733.D	Endrin ketone #2	Abdul	5/22/2025 8:03:45 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
RESCHK	PL095734.D	Endosulfan I #2	Abdul	5/22/2025 8:03:50 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
RESCHK	PL095734.D	Endrin ketone	Abdul	5/22/2025 8:03:50 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software
PCHLORICC050	PL095744.D	Chlordane-5	Abdul	5/22/2025 8:04:05 AM	mohammad	5/23/2025 5:17:22	Peak Integrated by Software

Manual Integration Report

Sequence:	PL052725	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PL095780.D	4,4"-DDD	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PEM	PL095780.D	4,4"-DDD #2	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PEM	PL095780.D	beta-BHC #2	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PEM	PL095780.D	Endrin aldehyde	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PEM	PL095780.D	Endrin aldehyde #2	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PEM	PL095780.D	Endrin ketone	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PEM	PL095780.D	Endrin ketone #2	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PEM	PL095780.D	Methoxychlor	Abdul	5/28/2025 8:11:00 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PSTDCCC050	PL095792.D	Endrin ketone #2	Abdul	5/28/2025 8:11:44 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PCHLORCCC500	PL095793.D	Chlordane-1	Abdul	5/29/2025 8:15:08 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PCHLORCCC500	PL095793.D	Chlordane-2	Abdul	5/29/2025 8:15:08 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PCHLORCCC500	PL095793.D	Chlordane-3 #2	Abdul	5/29/2025 8:15:08 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PCHLORCCC500	PL095793.D	Chlordane-5	Abdul	5/29/2025 8:15:08 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software

Manual Integration Report

Sequence:	PL052725	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PB168152BS	PL095795.D	Chlordane-1	Abdul	5/29/2025 8:15:13 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PB168152BS	PL095795.D	Chlordane-2	Abdul	5/29/2025 8:15:13 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PB168152BS	PL095795.D	Chlordane-3 #2	Abdul	5/29/2025 8:15:13 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PB168152BS	PL095795.D	Chlordane-5	Abdul	5/29/2025 8:15:13 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
Q1872-21	PL095796.D	Chlordane-1	Abdul	5/29/2025 8:15:18 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
Q1872-21	PL095796.D	Chlordane-1 #2	Abdul	5/29/2025 8:15:18 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
Q1872-21	PL095796.D	Chlordane-2	Abdul	5/29/2025 8:15:18 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
Q1872-21	PL095796.D	Chlordane-3 #2	Abdul	5/29/2025 8:15:18 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
Q1872-21	PL095796.D	Chlordane-5	Abdul	5/29/2025 8:15:18 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
Q1872-21	PL095796.D	Tetrachloro-m-xylene	Abdul	5/29/2025 8:15:18 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PCHLORCCC500	PL095799.D	Chlordane-1	Abdul	5/29/2025 8:15:23 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PCHLORCCC500	PL095799.D	Chlordane-2	Abdul	5/29/2025 8:15:23 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software
PCHLORCCC500	PL095799.D	Chlordane-3 #2	Abdul	5/29/2025 8:15:23 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software

Manual Integration Report

Sequence:	PL052725	Instrument	ECD_I
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PCHLORCCC500	PL095799.D	Chlordane-5	Abdul	5/29/2025 8:15:23 AM	mohammad	5/30/2025 1:44:22	Peak Integrated by Software

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052125

Review By	Abdul	Review On	5/22/2025 8:04:32 AM
Supervise By	mohammad	Supervise On	5/23/2025 5:17:22 AM
SubDirectory	PL052125	HP Acquire Method	HP Processing Method pl052125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL095731.D	21 May 2025 10:33	AR\AJ	Ok
2	I.BLK	PL095732.D	21 May 2025 10:47	AR\AJ	Ok
3	PEM	PL095733.D	21 May 2025 11:01	AR\AJ	Ok,M
4	RESCHK	PL095734.D	21 May 2025 11:14	AR\AJ	Ok,M
5	PSTDICC100	PL095735.D	21 May 2025 11:35	AR\AJ	Ok
6	PSTDICC075	PL095736.D	21 May 2025 11:48	AR\AJ	Ok
7	PSTDICC050	PL095737.D	21 May 2025 12:02	AR\AJ	Ok
8	PSTDICC025	PL095738.D	21 May 2025 12:15	AR\AJ	Ok,M
9	PSTDICC005	PL095739.D	21 May 2025 12:29	AR\AJ	Ok,M
10	PCHLORICC1000	PL095740.D	21 May 2025 12:42	AR\AJ	Ok
11	PCHLORICC750	PL095741.D	21 May 2025 12:56	AR\AJ	Ok
12	PCHLORICC500	PL095742.D	21 May 2025 13:10	AR\AJ	Ok
13	PCHLORICC250	PL095743.D	21 May 2025 13:23	AR\AJ	Ok
14	PCHLORICC050	PL095744.D	21 May 2025 13:37	AR\AJ	Ok,M
15	PTOXICC1000	PL095745.D	21 May 2025 13:50	AR\AJ	Ok
16	PTOXICC750	PL095746.D	21 May 2025 14:04	AR\AJ	Ok
17	PTOXICC500	PL095747.D	21 May 2025 14:18	AR\AJ	Ok
18	PTOXICC250	PL095748.D	21 May 2025 14:31	AR\AJ	Ok
19	PTOXICC100	PL095749.D	21 May 2025 14:45	AR\AJ	Ok,M
20	PSTDICV050	PL095750.D	21 May 2025 14:58	AR\AJ	Ok
21	PCHLORICV500	PL095751.D	21 May 2025 15:12	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052125

Review By	Abdul	Review On	5/22/2025 8:04:32 AM
Supervise By	mohammad	Supervise On	5/23/2025 5:17:22 AM
SubDirectory	PL052125	HP Acquire Method	HP Processing Method pl052125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL095752.D	21 May 2025 15:26	AR\AJ	Ok
----	------------	------------	-------------------	-------	----

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052725

Review By	Abdul	Review On	5/28/2025 8:12:03 AM
Supervise By	mohammad	Supervise On	5/30/2025 1:44:22 AM
SubDirectory	PL052725	HP Acquire Method	HP Processing Method pl052125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL095778.D	27 May 2025 08:24	AR\AJ	Ok
2	I.BLK	PL095779.D	27 May 2025 08:37	AR\AJ	Ok
3	PEM	PL095780.D	27 May 2025 08:51	AR\AJ	Ok,M
4	PSTDCCC050	PL095781.D	27 May 2025 09:04	AR\AJ	Ok
5	PB168151BL	PL095782.D	27 May 2025 09:24	AR\AJ	Ok
6	PB168151BS	PL095783.D	27 May 2025 10:01	AR\AJ	Ok,M
7	Q1872-20	PL095784.D	27 May 2025 10:16	AR\AJ	Dilution
8	Q1872-20DL	PL095785.D	27 May 2025 10:52	AR\AJ	Ok,M
9	Q2109-01	PL095786.D	27 May 2025 11:18	AR\AJ	Ok,M
10	Q2112-01	PL095787.D	27 May 2025 11:31	AR\AJ	Ok,M
11	Q2113-01	PL095788.D	27 May 2025 11:45	AR\AJ	Ok,M
12	Q2109-01MS	PL095789.D	27 May 2025 11:58	AR\AJ	Ok,M
13	Q2109-01MSD	PL095790.D	27 May 2025 12:12	AR\AJ	Ok,M
14	I.BLK	PL095791.D	27 May 2025 12:42	AR\AJ	Ok
15	PSTDCCC050	PL095792.D	27 May 2025 13:19	AR\AJ	Ok,M
16	PCHLORCCC500	PL095793.D	27 May 2025 14:52	AR\AJ	Ok,M
17	PB168152BL	PL095794.D	27 May 2025 15:19	AR\AJ	Ok
18	PB168152BS	PL095795.D	27 May 2025 15:33	AR\AJ	Ok,M
19	Q1872-21	PL095796.D	27 May 2025 16:03	AR\AJ	Ok,M
20	I.BLK	PL095797.D	27 May 2025 16:30	AR\AJ	Ok
21	PSTDCCC050	PL095798.D	27 May 2025 16:44	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052725

Review By	Abdul	Review On	5/28/2025 8:12:03 AM
Supervise By	mohammad	Supervise On	5/30/2025 1:44:22 AM
SubDirectory	PL052725	HP Acquire Method	HP Processing Method pl052125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PCHLORCCC500	PL095799.D	27 May 2025 17:00	AR\AJ	Ok,M
----	--------------	------------	-------------------	-------	------

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052125

Review By	Abdul	Review On	5/22/2025 8:04:32 AM
Supervise By	mohammad	Supervise On	5/23/2025 5:17:22 AM
SubDirectory	PL052125	HP Acquire Method	HP Processing Method
			pl052125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL095731.D	21 May 2025 10:33		AR\AJ	Ok
2	I.BLK	I.BLK	PL095732.D	21 May 2025 10:47		AR\AJ	Ok
3	PEM	PEM	PL095733.D	21 May 2025 11:01		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL095734.D	21 May 2025 11:14		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL095735.D	21 May 2025 11:35		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL095736.D	21 May 2025 11:48		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL095737.D	21 May 2025 12:02		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL095738.D	21 May 2025 12:15		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PL095739.D	21 May 2025 12:29		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL095740.D	21 May 2025 12:42		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL095741.D	21 May 2025 12:56		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL095742.D	21 May 2025 13:10		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL095743.D	21 May 2025 13:23		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL095744.D	21 May 2025 13:37		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PL095745.D	21 May 2025 13:50		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL095746.D	21 May 2025 14:04		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL095747.D	21 May 2025 14:18		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL095748.D	21 May 2025 14:31		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052125

Review By	Abdul	Review On	5/22/2025 8:04:32 AM
Supervise By	mohammad	Supervise On	5/23/2025 5:17:22 AM
SubDirectory	PL052125	HP Acquire Method	HP Processing Method pl052125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL095749.D	21 May 2025 14:45		AR\AJ	Ok,M
20	PSTDICV050	ICVPL052125	PL095750.D	21 May 2025 14:58		AR\AJ	Ok
21	PCHLORICV500	ICVPL052125CHLOR	PL095751.D	21 May 2025 15:12		AR\AJ	Ok
22	PTOXICV500	ICVPL052125TOX	PL095752.D	21 May 2025 15:26		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052725

Review By	Abdul	Review On	5/28/2025 8:12:03 AM
Supervise By	mohammad	Supervise On	5/30/2025 1:44:22 AM
SubDirectory	PL052725	HP Acquire Method	HP Processing Method
			pl052125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24095
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL095778.D	27 May 2025 08:24		AR\AJ	Ok
2	I.BLK	I.BLK	PL095779.D	27 May 2025 08:37		AR\AJ	Ok
3	PEM	PEM	PL095780.D	27 May 2025 08:51		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL095781.D	27 May 2025 09:04		AR\AJ	Ok
5	PB168151BL	PB168151BL	PL095782.D	27 May 2025 09:24		AR\AJ	Ok
6	PB168151BS	PB168151BS	PL095783.D	27 May 2025 10:01		AR\AJ	Ok,M
7	Q1872-20	HW0425-PT-PEST-SO	PL095784.D	27 May 2025 10:16	need 10x	AR\AJ	Dilution
8	Q1872-20DL	HW0425-PT-PEST-SO	PL095785.D	27 May 2025 10:52		AR\AJ	Ok,M
9	Q2109-01	TP-02-MHF	PL095786.D	27 May 2025 11:18		AR\AJ	Ok,M
10	Q2112-01	EO-03-05222025	PL095787.D	27 May 2025 11:31		AR\AJ	Ok,M
11	Q2113-01	HD-02-05222025	PL095788.D	27 May 2025 11:45		AR\AJ	Ok,M
12	Q2109-01MS	TP-02-MHFMS	PL095789.D	27 May 2025 11:58		AR\AJ	Ok,M
13	Q2109-01MSD	TP-02-MHFMSD	PL095790.D	27 May 2025 12:12		AR\AJ	Ok,M
14	I.BLK	I.BLK	PL095791.D	27 May 2025 12:42		AR\AJ	Ok
15	PSTDCCC050	PSTDCCC050	PL095792.D	27 May 2025 13:19		AR\AJ	Ok,M
16	PCHLORCCC500	PCHLORCCC500	PL095793.D	27 May 2025 14:52		AR\AJ	Ok,M
17	PB168152BL	PB168152BL	PL095794.D	27 May 2025 15:19		AR\AJ	Ok
18	PB168152BS	PB168152BS	PL095795.D	27 May 2025 15:33	CHLOR BS	AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL052725

Review By	Abdul	Review On	5/28/2025 8:12:03 AM
Supervise By	mohammad	Supervise On	5/30/2025 1:44:22 AM
SubDirectory	PL052725	HP Acquire Method	HP Processing Method pl052125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1872-21	HW0425-PT-CHLR-SO	PL095796.D	27 May 2025 16:03		AR\AJ	Ok,M
20	I.BLK	I.BLK	PL095797.D	27 May 2025 16:30		AR\AJ	Ok
21	PSTDCCC050	PSTDCCC050	PL095798.D	27 May 2025 16:44		AR\AJ	Ok
22	PCHLORCCC500	PCHLORCCC500	PL095799.D	27 May 2025 17:00		AR\AJ	Ok,M

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/25/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135545

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1869-01	MH-F	1	1.14	10.43	11.57	10.47	89.5	
Q1869-02	MH-F-EPH	2	1.18	9.96	11.14	10.12	89.8	
Q1869-03	MH-F-VOC	3	1.16	10.28	11.44	10.4	89.9	
Q1871-01	MH-A	4	1.14	9.59	10.73	9.86	90.9	
Q1871-02	MH-A-EPH	5	1.18	9.97	11.15	10.25	91.0	
Q1871-03	MH-A-VOC	6	1.15	10.22	11.37	10.47	91.2	
Q1871-05	MH-B	7	1.18	10.31	11.49	10.58	91.2	
Q1871-06	MH-B-EPH	8	1.16	9.63	10.79	10.05	92.3	
Q1871-07	MH-B-VOC	9	1.18	10.35	11.53	10.75	92.5	
Q1872-01	HW0425-PT-AN-SOIL	31	1.00	1.00	2.00	2.00	100.0	
Q1872-02	HW0425-PT-CORR-SOIL	32	1.00	1.00	2.00	2.00	100.0	
Q1872-03	HW0425-PT-CN-SOIL	33	1.00	1.00	2.00	2.00	100.0	
Q1872-04	HW0425-PT-CN-SOIL	34	1.00	1.00	2.00	2.00	100.0	
Q1872-05	HW0425-PT-FP-SOIL	35	1.00	1.00	2.00	2.00	100.0	
Q1872-06	HW0425-PT-CR6-SOIL	36	1.00	1.00	2.00	2.00	100.0	
Q1872-07	HW0425-PT-NUT-SOIL	37	1.00	1.00	2.00	2.00	100.0	
Q1872-08	HW0425-PT-NUT-SOIL	38	1.00	1.00	2.00	2.00	100.0	
Q1872-09	HW0425-PT-OGF-SOIL	39	1.00	1.00	2.00	2.00	100.0	
Q1872-10	HW0425-PT-MET-SOIL	40	1.00	1.00	2.00	2.00	100.0	
Q1872-11	HW0425-PT-BNA-SOIL	41	1.00	1.00	2.00	2.00	100.0	
Q1872-12	HW0425-PT-TRIAZINE-SOIL	42	1.00	1.00	2.00	2.00	100.0	
Q1872-13	HW0425-PT-PAH-SOIL	43	1.00	1.00	2.00	2.00	100.0	
Q1872-14	HW0425-PT-DIES-SOIL	44	1.00	1.00	2.00	2.00	100.0	
Q1872-15	HW0425-PT-GAS-SOIL	45	1.00	1.00	2.00	2.00	100.0	
Q1872-16	HW0425-PT-NJEPH-SOIL	46	1.00	1.00	2.00	2.00	100.0	
Q1872-17	HW0425-PT-HERB-SOIL	47	1.00	1.00	2.00	2.00	100.0	
Q1872-18	HW0425-PT-PCB-SOIL	48	1.00	1.00	2.00	2.00	100.0	
Q1872-19	HW0425-PT-PCBO-SOIL	49	1.00	1.00	2.00	2.00	100.0	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/25/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135545

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1872-20	HW0425-PT-PEST-SOIL	50	1.00	1.00	2.00	2.00	100.0	
Q1872-21	HW0425-PT-CHLR-SOIL	51	1.00	1.00	2.00	2.00	100.0	
Q1872-22	HW0425-PT-TXP-SOIL	52	1.00	1.00	2.00	2.00	100.0	
Q1872-23	HW0425-PT-VOA-SOIL	53	1.00	1.00	2.00	2.00	100.0	
Q1872-25	HW0425-PT-NO2-SOIL	54	1.00	1.00	2.00	2.00	100.0	
Q1873-01	CAM-40619	10	1.14	10.70	11.84	4.97	35.8	
Q1873-02	CAM-40620	11	1.15	10.42	11.57	6.19	48.4	
Q1873-03	CAM-40619-20	12	1.18	10.21	11.39	4.77	35.2	
Q1874-01	VNJ-236	13	1.19	10.45	11.64	10.89	92.8	
Q1874-03	RT1491	14	1.19	11.16	12.35	11.43	91.8	
Q1874-05	HT3727	15	1.16	10.63	11.79	11.06	93.1	
Q1875-01	AUD-25-0053	16	1.14	10.75	11.89	11.19	93.5	
Q1875-03	AUD-25-0054	17	1.14	10.02	11.16	10.52	93.6	
Q1875-04	AUD-25-0024	18	1.14	10.03	11.17	10.77	96.0	
Q1876-01	AUD-25-0058	19	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-02	AUD-25-0059	20	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-03	AUD-25-0060	21	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-04	AUD-25-0061	22	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-05	AUD-25-0062	23	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-06	AUD-25-0063	24	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-07	AUD-25-0064	25	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-08	AUD-25-0065	26	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-09	AUD-25-0066	27	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1877-01	AU-6-042425	55	1.14	10.25	11.39	10.72	93.5	
Q1877-02	AU-6-042425	28	1.14	10.21	11.35	10.54	92.1	
Q1878-01	TR-4-042425	29	1.14	10.17	11.31	11.2	98.9	
Q1878-02	TR-4-042425-E2	30	1.19	10.28	11.47	10.92	94.6	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042425

WorkList ID : 189122

Department : Wet-Chemistry

VB 135545

Date : 04-24-2025 08:52:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1869-01	MH-F	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1869-02	MH-F-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1869-03	MH-F-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-01	MH-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-02	MH-A-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-03	MH-A-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-05	MH-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-06	MH-B-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-07	MH-B-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1872-01	HW0425-PT-AN-SOIL	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1872-02	HW0425-PT-CORR-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-03	HW0425-PT-CN-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-04	HW0425-PT-CN-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-05	HW0425-PT-FP-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-06	HW0425-PT-CR6-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-07	HW0425-PT-NUT-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-08	HW0425-PT-NUT-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-09	HW0425-PT-OGR-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-10	HW0425-PT-MET-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-11	HW0425-PT-BNA-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-12	HW0425-PT-TRIAZINE-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO

Date/Time 04/24/25 15:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/24/25 17:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

WB 135545

WorkList Name : %1-042425

WorkList ID : 189122

Department : Wet-Chemistry

Date : 04-24-2025 08:52:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1872-13	HW0425-PT-PAH-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-14	HW0425-PT-DIES-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-15	HW0425-PT-GAS-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-16	HW0425-PT-NJEPH-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-17	HW0425-PT-HERB-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-18	HW0425-PT-PCB-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-19	HW0425-PT-PCBO-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-20	HW0425-PT-PEST-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-21	HW0425-PT-CHLR-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-22	HW0425-PT-TXP-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-23	HW0425-PT-VOA-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-25	HW0425-PT-NO2-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1873-01	CAM-40619	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1873-02	CAM-40620	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1873-03	CAM-40619-20	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1874-01	VNJ-236	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1874-03	RT1491	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1874-05	HT3727	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1875-01	AUD-25-0053	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1875-03	AUD-25-0054	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1875-04	AUD-25-0024	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO

Date/Time 04/24/25 15:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/24/25 17:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

135545

WorkList Name : %1-042425

WorkList ID : 189122

Department : Wet-Chemistry

Date : 04-24-2025 08:52:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1876-01	AUD-25-0058	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-02	AUD-25-0059	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-03	AUD-25-0060	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-04	AUD-25-0061	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-05	AUD-25-0062	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-06	AUD-25-0063	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-07	AUD-25-0064	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-08	AUD-25-0065	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-09	AUD-25-0066	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1877-01	AU-6-042425	Solid	Percent Solids	Cool 4 deg C	PSEG05	L41	04/24/2025	Chemtech -SO
Q1877-02	AU-6-042425	Solid	Percent Solids	Cool 4 deg C	PSEG05	L41	04/24/2025	Chemtech -SO
Q1878-01	TR-4-042425	Solid	Percent Solids	Cool 4 deg C	PSEG05	L41	04/24/2025	Chemtech -SO
Q1878-02	TR-4-042425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L41	04/24/2025	Chemtech -SO

Date/Time

04/24/25 15:30

Raw Sample Received by:

220001

Raw Sample Relinquished by:

SH

Q1872-PESTICIDE Group2

Date/Time

04/24/23 17:25

Raw Sample Received by:

28 17:35

Raw Sample Relinquished by:

SH

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/29/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 04/28/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 04/29/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135575

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1872-24	HW0425-PT-SOL-SOIL	8	0.92	10.30	11.22	8.82	76.7	
Q1901-01	B-170-SB00	1	1.14	5.55	6.69	6.28	92.6	
Q1901-02	B-167-SB01	2	1.14	10.22	11.36	9.58	82.6	
Q1901-03	B-170-SB01	3	1.19	10.31	11.5	9.75	83.0	
Q1901-04	B-167-SB02	4	1.15	9.78	10.93	6.35	53.2	
Q1901-05	B-170-SB02	5	1.14	10.16	11.3	8.77	75.1	
Q1902-01	343	6	1.19	10.23	11.42	10.7	93.0	
Q1902-02	343	7	1.13	10.19	11.32	10.33	90.3	
Q1903-01	COMP-4	9	1.18	11.14	12.32	10.46	83.3	
Q1903-02	COMP-5	10	1.16	10.50	11.66	9.44	78.9	
Q1903-03	COMP-6	11	1.17	10.60	11.77	10.06	83.9	
Q1904-01	VNJ-210	12	1.19	10.39	11.58	10.6	90.6	
Q1905-01	MH-G	13	1.15	10.35	11.5	10.38	89.2	
Q1905-02	MH-G-EPH	14	1.16	9.65	10.81	9.71	88.6	
Q1905-03	MH-G-VOC	15	1.16	10.33	11.49	10.36	89.1	
Q1905-05	MH-H	16	1.12	10.03	11.15	10.5	93.5	
Q1905-06	MH-H-EPH	17	1.13	10.30	11.43	10.5	91.0	
Q1905-07	MH-H-VOC	18	1.12	10.03	11.15	10.01	88.6	
Q1906-01	WC-4	19	1.15	9.85	11.00	10.14	91.3	
Q1906-02	WC-4-EPH	20	1.16	9.97	11.13	10.17	90.4	
Q1906-03	WC-4-VOC	21	1.18	9.99	11.17	9.91	87.4	
Q1906-05	WC-5	22	1.16	10.82	11.98	10.19	83.5	
Q1906-06	WC-5-EPH	23	1.13	10.41	11.54	9.94	84.6	
Q1906-07	WC-5-VOC	24	1.18	10.47	11.65	11.63	99.8	
Q1906-09	WC-6	25	1.14	10.04	11.18	10.4	92.2	
Q1906-10	WC-6-EPH	26	1.15	10.77	11.92	10.23	84.3	
Q1906-11	WC-6-VOC	27	1.14	10.47	11.61	10.86	92.8	
Q1906-13	WC-7	28	1.14	10.85	11.99	10.31	84.5	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/29/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 04/28/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 04/29/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135575

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1906-14	WC-7-EPH	29	1.12	9.86	10.98	9.7	87.0	
Q1906-15	WC-7-VOC	30	1.13	10.27	11.4	10.23	88.6	
Q1907-01	CO-8R-WC	31	1.13	10.26	11.39	9.81	84.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

W3 B55F5

WorkList Name : %1-042825

WorkList ID : 189159

Department : Wet-Chemistry

Date : 04-28-2025 07:59:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1872-24	HW0425-PT-SOL-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1903-01	COMP-4	Solid	Percent Solids	Cool 4 deg C	POWE02	L51	04/25/2025	Chemtech -SO
Q1903-02	COMP-5	Solid	Percent Solids	Cool 4 deg C	POWE02	L51	04/25/2025	Chemtech -SO
Q1903-03	COMP-6	Solid	Percent Solids	Cool 4 deg C	POWE02	L51	04/25/2025	Chemtech -SO
Q1901-01	B-170-SB00	Solid	Percent Solids	Cool 4 deg C	PORT06	L51	04/26/2025	Chemtech -SO
Q1901-02	B-167-SB01	Solid	Percent Solids	Cool 4 deg C	PORT06	L51	04/26/2025	Chemtech -SO
Q1901-03	B-170-SB01	Solid	Percent Solids	Cool 4 deg C	PORT06	L51	04/26/2025	Chemtech -SO
Q1901-04	B-167-SB02	Solid	Percent Solids	Cool 4 deg C	PORT06	L51	04/26/2025	Chemtech -SO
Q1901-05	B-170-SB02	Solid	Percent Solids	Cool 4 deg C	PORT06	L51	04/26/2025	Chemtech -SO
Q1902-01	343	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1902-02	343	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1904-01	VNJ-210	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1905-01	MH-G	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/28/2025	Chemtech -SO
Q1905-02	MH-G-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/28/2025	Chemtech -SO
Q1905-03	MH-G-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/28/2025	Chemtech -SO
Q1906-13	WC-7	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-14	WC-7-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-15	WC-7-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-05	WC-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-06	WC-5-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-07	WC-5-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO

Date/Time 04/28/25 16:15

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Q1872-PESTICIDE Group2

Date/Time 04/28/25 17:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

WB 135575

WorkList Name : %1-042825

WorkList ID : 189159

Department : Wet-Chemistry

Date : 04-28-2025 07:59:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1906-09	WC-6	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-10	WC-6-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-11	WC-6-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1905-05	MH-H	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/28/2025	Chemtech -SO
Q1905-06	MH-H-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/28/2025	Chemtech -SO
Q1905-07	MH-H-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/28/2025	Chemtech -SO
Q1906-01	WC-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-02	WC-4-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1906-03	WC-4-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/28/2025	Chemtech -SO
Q1907-01	CO-8R-WC	Solid	Percent Solids	Cool 4 deg C	WALS01	L51	04/28/2025	Chemtech -SO

Date/Time 04/28/25 16:15
 Raw Sample Received by: SO WELC
 Raw Sample Relinquished by: CP

Q1872-PESTICIDE Group2

Date/Time 04/28/25 17:30
 Raw Sample Received by: CP
 Raw Sample Relinquished by: SO WELC
 136 of 200

SOP ID:	M3541-ASE Extraction-14		
Clean Up SOP #:	Florisil	Extraction Start Date :	05/23/2025
Matrix :	Solid	Extraction Start Time :	09:20
Weigh By:	RJ	Extraction By:	RJ
Balance check:	RJ	Filter By:	RS
Balance ID:	EX-SC-2	pH Meter ID:	N/A
pH Strip Lot#:	N/A	Hood ID:	3,7
		Concentration By:	EH
		Supervisor By :	RUPESH

Extraction Method:
☐ Separatory Funnel
 ☐ Continuous Liquid/Liquid
 ☐ Sonication
 ☐ Waste Dilution
 ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	2.0ML	1000 PPB	PP24081
Surrogate	1.0ML	200 PPB	PP24597
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2613
Baked Na2SO4	N/A	EP2614
Sand	N/A	E2865
Hexane	N/A	E3938
Florisil	N/A	E3806
9:1 Hexane:Acetone Mixture	N/A	EP2596
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS723.

KD Bath ID:	N/A	Envap ID:	NEVAP-02
KD Bath Temperature:	N/A	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/23/25	RS (ECLab)	Y. P. P. + R. B.
12:35	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 05/23/2025

Sample ID	Client Sample ID	Test	g/ mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168152BL	PBLK152	PESTICIDE Group2	30.02	N/A	ritesh	RUPESH	10			U2-1
PB168152BS	PLCS152	PESTICIDE Group2	30.03	N/A	ritesh	RUPESH	10			2
Q1872-21	HW0425-PT-CHLR-SOIL	PESTICIDE Group2	30.18	N/A	ritesh	RUPESH	10	A		3

[Large handwritten signature]

RS
5/23

* Extracts relinquished on the same date as received.

168152
9:20 AM

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1872

WorkList ID : 189742

Department : Extraction

Date : 05-23-2025 08:51:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1872-21	HW0425-PT-CHLR-SOIL	Solid	PESTICIDE Group2	Cool 4 deg C	ALLI03	QA Of	04/21/2025	8081B
Q1872-22	HW0425-PT-TXP-SOIL	Solid	PESTICIDE Group3	Cool 4 deg C	ALLI03	QA Of	04/21/2025	8081B

Date/Time 5/23/25 9:15
Raw Sample Received by: RJCEA-lab
Raw Sample Relinquished by: AK Sm

Date/Time 5/23/25 9:35
Raw Sample Received by: AK Sm
Raw Sample Relinquished by: RJCEA-lab

Prep Standard - Chemical Standard Summary

Order ID : Q1872
Test : PESTICIDE Group2
Prepbatch ID : PB168152,
Sequence ID/Qc Batch ID: PL052725,

Standard ID :
EP2613,EP2614,PP24081,PP24095,PP24255,PP24256,PP24257,PP24258,PP24259,PP24260,PP24261,PP24262,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284,PP24329,PP24433,PP24597,

Chemical ID :
E2865,E3551,E3806,E3847,E3877,E3914,E3932,E3933,E3938,P12600,P12603,P12611,P13037,P13040,P13195,P13245,P13356,P13357,P13405,P13785,P13861,P9052,W3177,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2613	05/09/2025	11/05/2025	RUPESHKUMAR SHAH	None	None	Riteshkumar Patel 05/09/2025

FROM 4000.00000ml of E3932 + 4000.00000ml of E3933 = Final Quantity: 8000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2614	05/19/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/19/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1501	1000 ppb CHLORDANE SPIKE (RESTEK)	PP24081	12/16/2024	06/16/2025	Abdul Mirza	None	None	Ankita Jodhani
								12/17/2024

FROM 0.10000ml of P12600 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24095	12/23/2024	06/16/2025	Abdul Mirza	None	None	Ankita Jodhani
								12/30/2024

FROM 1.00000ml of P13245 + 99.00000ml of E3847 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24255	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 1.00000ml of P13785 + 9.00000ml of E3877 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP24256	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 1.00000ml of P13040 + 9.00000ml of E3877 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24257	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 1.00000ml of P13037 + 9.00000ml of E3877 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1273	20 PPM Mirex Stock (Primary Source)	PP24258	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.20000ml of P9052 + 9.80000ml of E3877 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24259	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.20000ml of P13195 + 9.80000ml of E3877 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP24260	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 98.50000ml of E3877 + 0.50000ml of PP24255 + 0.50000ml of PP24256 + 0.50000ml of PP24258 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
80	100/100 PPB Pesticide Working Solution 2nd Source	PP24261	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 98.50000ml of E3877 + 0.50000ml of PP24255 + 0.50000ml of PP24257 + 0.50000ml of PP24259 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP24262	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.10000ml of P12603 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP24266	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.10000ml of P12611 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP24267	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.10000ml of P13405 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP24268	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.10000ml of P13861 + 99.40000ml of E3877 + 0.50000ml of PP24255 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3631	75 PPB ICAL PEST STD(RESTEK)	PP24269	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.75000ml of E3877 + 0.25000ml of PP24260 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP24270	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.50000ml of E3877 + 0.50000ml of PP24260 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3633	25 PPB ICAL PEST STD(RESTEK)	PP24271	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani 03/12/2025

FROM 0.75000ml of E3877 + 0.25000ml of PP24260 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3634	5 PPB ICAL PEST STD(RESTEK)	PP24272	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.90000ml of E3877 + 0.10000ml of PP24270 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3988	50 PPB PEST ICV STD(RESTEK)	PP24273	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.50000ml of E3877 + 0.50000ml of PP24261 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
528	CHLOR 750 PPB STD	PP24274	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.25000ml of E3877 + 0.75000ml of PP24262 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP24275	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.50000ml of E3877 + 0.50000ml of PP24262 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
530	CHLOR 250 PPB STD	PP24277	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.75000ml of E3877 + 0.25000ml of PP24262 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3408	CHLOR 50 PPB STD	PP24278	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
								03/12/2025

FROM 0.90000ml of E3877 + 0.10000ml of PP24275 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP24279	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.50000ml of E3877 + 0.50000ml of PP24266 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
533	TOX 750 PPB STD	PP24280	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.25000ml of E3877 + 0.75000ml of PP24267 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP24281	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.50000ml of E3877 + 0.50000ml of PP24267 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
535	TOX 250 PPB STD	PP24282	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.75000ml of E3877 + 0.25000ml of PP24267 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2217	TOX 100 PPB STD	PP24283	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.90000ml of E3877 + 0.10000ml of PP24267 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP24284	03/11/2025	08/12/2025	Abdul Mirza	None	None	Ankita Jodhani
03/12/2025								

FROM 0.50000ml of E3877 + 0.50000ml of PP24268 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
518	Pest/PCB I.BLK 20 PPB	PP24433	03/31/2025	08/22/2025	Abdul Mirza	None	None	Yogesh Patel
								04/02/2025

FROM 99.90000ml of E3914 + 0.10000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24597	05/20/2025	11/05/2025	Abdul Mirza	None	None	Yogesh Patel
FROM 1.00000ml of P13357 + 999.00000ml of E3932 = Final Quantity: 1000.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	0000243821	06/30/2025	04/30/2020 / RAJESH	04/28/2020 / RAJESH	E2865

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	12/04/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agela Technologies Inc.	FS0006 / Cleanert Florisil cartridge	M06518	09/25/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3806

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	06/16/2025	12/16/2024 / Rajesh	12/13/2024 / Rajesh	E3847

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	09/19/2025	03/19/2025 / RUPESH	03/13/2025 / RUPESH	E3914

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3932

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3933

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	12/05/2025	06/05/2025 / RUPESH	05/14/2025 / RUPESH	E3938

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0193299	06/16/2025	12/16/2024 / Abdul	07/03/2023 / Abdul	P12600

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	09/11/2025	03/10/2025 / Abdul	07/03/2023 / Abdul	P12603

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0193299	09/09/2025	03/10/2025 / Abdul	07/03/2023 / Abdul	P12611

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	09/10/2025	03/10/2025 / Abdul	12/26/2023 / Abdul	P13037

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	09/10/2025	03/10/2025 / Abdul	12/26/2023 / Abdul	P13040

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	09/10/2025	03/10/2025 / Abdul	01/17/2024 / Abdul	P13195

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	06/23/2025	12/23/2024 / Abdul	02/09/2024 / Abdul	P13245

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	11/20/2025	05/20/2025 / Abdul	04/22/2024 / Abdul	P13357

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	09/09/2025	03/10/2025 / Abdul	05/15/2024 / Abdul	P13405

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	09/10/2025	03/10/2025 / Abdul	11/19/2024 / Ankita	P13785

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	09/10/2025	03/10/2025 / Abdul	12/09/2024 / Abdul	P13861

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	112018	09/10/2025	03/10/2025 / Abdul	11/01/2019 / Stephen	P9052

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0193299

Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2029 **Storage:** 10°C or colder
Ship: Ambient

P12596
↓
P12602
⑦
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	---%	1,010.0 µg/mL	+/- 56.0475

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Bryan Snyder
Bryan Snyder - Operations Tech I

Date Mixed: 06-Jan-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

Sand
Purified
Washed and Ignited



Material No.: 3382-05
Batch No.: 0000243821
Manufactured Date: 2018/04/09
Retest Date: 2025/04/07
Revision No: 1


Certificate of Analysis

Test	Specification	Result
Substances Soluble in HCl	$\leq 0.16\%$	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

E 2865


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Cleanert Florisil

1g/6ml 30/pkg

固相萃取产品

LOT#:M06518

MFG#:F04074



Made in China

CAT# FS0006

 Agela Technologies

E 3806



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 12/13/24

E3847

J. Croak

Jamie Croak
Director Quality Operations, Bioscience Production

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

 E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd by RS on 3/14/25



E3914

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

RS

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3932

Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**TM



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3933



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3938

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis
chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0193299
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : April 30, 2029 **Storage:** 10°C or colder
Ship: Ambient

P12616
↓
P12615
Five
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	—%	1,010.0 µg/mL	+/- 56.0475

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Bryan Snyder
Bryan Snyder - Operations Tech I

Date Mixed: 06-Jan-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Jan-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

CRm
P 12611
↓
P 12615
⑤
CRout
7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 Lot No.: A0199099

Description : Organochlorine Pesticide Mix AB #1
Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL Pkg Amt: > 1 mL

Expiration Date : June 30, 2027 Storage: 10°C or colder

Ship: Ambient

P130397 5
↓
P13043 1
✓
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1	µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0	µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1	µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0	µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P 13039
 ↓
 P13043
 1
 5
 12/26/23

Quality Confirmation Test

Column:
 30m x .25mm x .2µm
 Rtx-CLP II (cat.# 11323)

Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)


Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD

Split Vent:
 Split ratio 50:1

Inj. Vol
 1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
 Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023

Balance Serial # 1128360905

Jennifer Pollino
 Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **79136**
Lot Number: **042022**
Description: **Mirex**Solvent(s):
Acetone Lot#
81025

<i>Prashant Chauhan</i>		042022
Formulated By:	Prashant Chauhan	DATE
<i>Pedro L. Rentas</i>		042022
Reviewed By:	Pedro L. Rentas	DATE

Expiration Date: 042027

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration (µg/mL): 1000

NIST Test ID#: 6UTB




5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to (mL): 50.0 0.006 Flask Uncertainty

Expanded
SDS Information
(Solvent Safety Info. On Attached pg.)

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Mirex	437	9492400	1000	99.4	0.5	0.05034	0.05040	1001.1	10.3	2385-85-5	N/A	ori-rat 306mg/kg

Method GC7MSD-1.M: Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B = 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5 % of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

01/17/2024
HARRIS
D
3/3/19
3/3/19
5
5



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

P 12603
↓
P 12605
3
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician


Date Mixed: 11-May-2023

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 12603 } (3)
↓
P 12605

7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023 **Balance Serial #** B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



CERTIFIED WEIGHT REPORT

Part Number: **19161**
 Lot Number: **013124**
 Description: **CLP Pesticides & PCB's Resolution Check Standard**
 9 components
 Expiration Date: **013129**
 Recommended Storage: **Refrigerate (4 °C)**
 Nominal Concentration (µg/mL): **Varied**
 NIST Test ID#: **6UTB**
 Volume(s) shown below were combined and diluted to (mL): **100.0**

Solvent(s):
 Hexane
 Toluene

Lot#
 273615 (50%)
 28508 (50%)

5E-05 Balance Uncertainty
 0.021 Flask Uncertainty

		013124
Formulated By:	Lawrence Barry	DATE
		013124
Reviewed By:	Pedro L. Rentas	DATE

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty (+/-) µg/mL	SDS Information (Solvent Safety Info. On Attached pg.)		
									CAS#	OSHA PEL (TWA)	LD50
1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	ori-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	ori-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	ori-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	ori-rat 38300ug/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	ori-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4'-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	ori-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
 • Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
 • Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
 • All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
 • Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

P13243 } (5)
 ↓
 P13247 }
 ↓
 02/9/2024



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
SAUF
04/25/2025
(10)



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente
Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
SAUF
04/25/2025
(10)



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2028 **Storage:** 10°C or colder
Ship: Ambient

P13402
P13406
5/22/2021

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	---%	1,009.0 µg/mL	+/- 55.9920

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I


Date Mixed: 10-Oct-2023

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13402
↓
P13406 } (5)

5/22/2024



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
↓
P19789 AJ
11/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

heliu-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

A.O.E.
Aaron Enyart - Operations Tech I

Date Mixed: 29-Jul-2024

Balance Serial # B345965662

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:
30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
300 ml/min.

Inj. Vol
0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.



Amanda Miller - Operations Tech III - ARM QC


Date Mixed: 11-Apr-2024 **Balance Serial #** B442140311


Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }
↑

12/9/2024



CERTIFIED WEIGHT REPORT

Part Number: **72072**
Lot Number: **112018**
Description: **n-Tetracosane-d50**

Solvent(s):
Methylene chloride
Lot#: **102669**

Expiration Date: **112028**
Recommended Storage: **Ambient (20 °C)**
Nominal Concentration (µg/mL): **1000**
NIST Test ID#: **2684186**

Weight(s) shown below were combined and diluted to (mL):




200.0
5E-05 Balance Uncertainty
0.058 Flask Uncertainty

Received by
SG on 11/1/19
P9044 - P9053


		112018
Formulated By:	Prashant Chauhan	DATE
		112018
Reviewed By:	Pedro Rentas	DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)		
										CAS#	OSHA PEL (TWA)	LD50
1. n-Tetracosane-d50	2072	PR-17753/09216TC1	1000	98	0.2	0.20411	0.20415	1000.2	4.2	16416-32-3	N/A	N/A

Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 250°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Candice Warren.



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).



Run 40, "P72072 L112018 [1000µg/mL in MeCl2]"

Run Length: 35.00 min, 20999 points at 10 points/second.
Created: Thu, Nov 22, 2018 at 7:23:18 AM.
Sampled: Sequence "112018-GC4M1", Method "GC4-M1".
Analyzed using Method "GC4-M1".

Comments

GC4-M1 Analysis by Melissa Stonier


Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5µm Film Thickness

Flow rates; Total Flow = 300 mL/min, Helium (carrier) = 6.5 mL, Helium (make-up) = 25 mL, Hydrogen (detector) = 30 mL,
Air (detector) = 360 mL

Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.

Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDag Channel 1.

Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 µL, Range = 3



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

W3147
W3147
02/03/2023
JP

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	≥ 99.5 %	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	≥ 95 %	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	≤ 0.05 %	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

J. Croak
Jamie Croak
Director Quality Operations, Bioscience
196 of 200



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19



phenova®
Certified Reference Materials

A Phenomenex®
Company

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

Received by: SJ
4/23/2025 15:50

For terms and conditions of your order, please visit:
www.phenova.com/home/termsandsale

Packing List

Date	Order #
04/21/2025	333293



Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
PO2-1668	Net 30	ZCM-100	1500470	FedEx Collect 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	PT-MET-SOIL	SOIL/HW Trace Metals	HW0425	7100-04
1	1	0	PT-CR6-SOIL	SOIL/HW Hexavalent Chromium	HW0425	7100-05B
1	1	0	PT-CN-SOIL	SOIL/HW Cyanide	HW0425	7100-06
1	1	0	PT-CORR-SOIL	SOIL/HW Corrosivity/pH	HW0425	7100-11
1	1	0	PT-FP-SOIL	SOIL/HW Flash Point	HW0425	7100-10
1	1	0	PT-AN-SOIL	SOIL/HW Anions	HW0425	7100-08
1	1	0	PT-NUT-SOIL	SOIL/HW Nutrients	HW0425	7100-09B
1	1	0	PT-SOL-SOIL	SOIL/HW Solids	HW0425	7100-31
1	1	0	PT-NO2-SOIL	SOIL/HW Nitrite as N	HW0425	7100-71
1	1	0	PT-GAS-SOIL	SOIL/HW Gasoline	HW0425	7100-96
1	1	0	PT-OGR-SOIL	SOIL/HW Oil and Grease	HW0425	7100-94
1	1	0	PT-VOA-SOIL	SOIL/HW Volatiles	HW0425	7100-12
1	1	0	PT-BNA-SOIL	SOIL/HW BNAs	HW0425	7100-13
1	1	0	PT-PEST-SOIL	SOIL/HW Pesticides	HW0425	7100-14
1	1	0	PT-CHLR-SOIL	SOIL/HW Chlordane	HW0425	7100-15
1	1	0	PT-TXP-SOIL	SOIL/HW Toxaphene	HW0425	7100-16
1	1	0	PT-PCB-SOIL	SOIL/HW PCBs	HW0425	7100-17
1	1	0	PT-PCBO-SOIL	SOIL/HW PCBs in Oil	HW0425	7100-88
1	1	0	PT-HERB-SOIL	SOIL/HW Herbicides	HW0425	7100-18
1	1	0	PT-PAH-SOIL	SOIL/HW PAHs	HW0425	7100-22
1	1	0	PT-TRIAZINE-SOIL	SOIL/HW Triazine Pesticides	HW0425	7100-106
1	1	0	PT-NJEPH-SOIL	NJ EPH in SOIL	HW0425	7100-105



phenova
Certified Reference Materials

A Phenomenex[®]
Company

Packing List

Date	Order #
04/25/2025	337220



6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ

4/28/2025 9:40

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
CPR	Net 30	ZCM-100	1500470	FedEx Next Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	PT-DIES-SOIL	SOIL/HW Diesel in Soil	HW0425	7100-100

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488