



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Alliance Technical Group, LLC - Newark

Project Name: NJ Soil PT

Project # N/A

Order ID # Q1872

Test Name: Herbicide Group1

A. Number of Samples and Date of Receipt:

24 Solid samples were received on 04/24/2025.

1 Solid sample was received on 04/28/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group1, Anions Group2, Corrosivity, Cyanide, Diesel Range Organics, EPH, Flash Point, Gasoline Range Organics, Herbicide Group1, Hexavalent Chromium, Mercury, Metals Group3, Metals ICP-Group1, Oil and Grease, PCB, PESTICIDE Group1, PESTICIDE Group2, PESTICIDE Group3, Phosphorus, Total, SVOCMS Group1, SVOCMS Group2, SVOCMS Group3, SVOCMS Group4, SVOCMS Group5, TKN, TOC, TS and VOCMS Group1. This data package contains results for Herbicide Group1.

C. Analytical Techniques:

The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 11324. The analysis of Herbicide Group1s was based on method 8151A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS {Q2130-01MS} with File ID: PS030577.D recoveries met the acceptable requirements except for [MCP(1)- 35%],[MCP(2)- 32%],[Dalapon(1)- 15%],[MCPA(1)- 39%],[MCPA(2)- 41%],[Dinoseb(1)- 9%],[Pentachlorophenol(1)- 50%],[Pentachlorophenol(2)- 45%],[4-Nitrophenol(1)- 41%],[4-Nitrophenol(2)- 43%],[PICLORAM(1)- 33%],[PICLORAM(2)- 33%],[DCPA(1)- 37%],[DCPA(2)- 45%],[3,5-DICHLOROBENZOIC ACID(1)- 48%],[3,5-DICHLOROBENZOIC ACID(2)- 47%] due to matrix interference.

The MSD {Q2130-01MSD} with File ID: PS030578.D recoveries met the acceptable requirements except for [MCP(1)- 37%],[MCP(2)- 32%],[Dalapon(1)-



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

15%],[MCPA(1)- 39%],[MCPA(2)- 40%][Dinoseb(1)- 9%],[Pentachlorophenol(1)- 50%],[Pentachlorophenol(2)- 45%] [4-Nitrophenol(1)- 41%],[4-Nitrophenol(2)- 43%][PICLORAM(1)- 33%],[PICLORAM(2)- 32%][DCPA(1)- 37%],[DCPA(2)- 44%] [3,5-DICHLOROBENZOIC ACID(1)- 47%],[3,5-DICHLOROBENZOIC ACID(2)- 46%] due to matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

Sample HW0425-PT-HERB-SOIL was diluted due to high concentration.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____