ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

MATRIX: Solid

ORDER ID: Q1872

METH	OD: 8270-Modified/3541			
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	NA	NO	YES ✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			✓
	The Initial Calibration met the requirements. The Continuous Calibration met the requirements.			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria except for HW0425-PT-TRIAZINE-SOIL [2-Methylnaphthalene-d10 - 0%, Fluoranthene-d10 - 0%]. Failed surrogate are not associated, Therefor no further corrective action was taken.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			\checkmark
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike met requirements for all samples			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met		✓	
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met		✓	
	If not met, list number of days exceeded for each sample:			
	The Holding Times were met for all analysis.			
ADDIT	TONAL COMMENTS:			
The For	rm 6 is not included in the data package because the Initial Calibration was perform	ned using 7 points	s.	
The soi	l samples results are based on a dry weight basis.			
Please ı	use %D calculated based on Avg RF and CCRF for all compounds using Average F	Response Factor v	vhen the	
%RSD	value for a compound is <20% for the Initial Calibration curve and use %D calcula	ited based on Am	ount add	led
and Cal	culated amount for all compounds using Linear Regression when the %RSD value	for a compound i	is > 20%	for
he Initi	al Calibration curve for SW-846 analysis.			
QA RE	VIEW Date			