

Duplicate Sample Summary

Client: Alliance Technical Group, LLC - Newark	SDG No.: Q1872
Project: NJ Soil PT	Sample ID: Q1889-01
Client ID: COMP-1DUP	Percent Solids for Spike Sample: 82.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	5.50	J	5.60	J	1	2		04/30/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q1872
Project:	NJ Soil PT	Sample ID:	Q1889-01
Client ID:	COMP-1MSD	Percent Solids for Spike Sample:	82.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	63.1		65.5		1	4		04/30/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q1872
Project:	NJ Soil PT	Sample ID:	Q1938-07
Client ID:	LOWER-WALL-PILE-DDUP	Percent Solids for Spike Sample:	95

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.074	J	0.082	J	1	10		05/07/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q1872
Project:	NJ Soil PT	Sample ID:	Q1938-07
Client ID:	LOWER-WALL-PILE-DMSD	Percent Solids for Spike Sample:	95

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	2.20		2.10		1	5		05/07/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q1872
Project:	NJ Soil PT	Sample ID:	Q1964-04
Client ID:	MH-LLDUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Corrosivity	pH	+/-20	6.13		6.14		1	0.16		05/07/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q1872
Project:	NJ Soil PT	Sample ID:	Q1965-01
Client ID:	50623DUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Flash Point	o F	+/-2	>212.0		>212.0		1	0		05/08/2025

Duplicate Sample Summary

Client: Alliance Technical Group, LLC - Newark Project: NJ Soil PT Client ID: WC-A4-03-CDUP	SDG No.: Q1872 Sample ID: Q2001-02 Percent Solids for Spike Sample: 79.1
--	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/Kg	+/-20	732		744		1	1.66		05/14/2025

Duplicate Sample Summary

Client: Alliance Technical Group, LLC - Newark Project: NJ Soil PT Client ID: WC-A4-03-CMSD	SDG No.: Q1872 Sample ID: Q2001-02 Percent Solids for Spike Sample: 79.1
--	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/Kg	+/-20	940		946		1	0.57		05/14/2025

Duplicate Sample Summary

Client: Alliance Technical Group, LLC - Newark Project: NJ Soil PT Client ID: SC-4-SED-051525DUP	SDG No.: Q1872 Sample ID: Q2085-01 Percent Solids for Spike Sample: 75.7
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/Kg	+/-20	62.3		61.9		1	0.64		05/28/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q1872
Project:	NJ Soil PT	Sample ID:	Q2085-01
Client ID:	SC-4-SED-051525MSD	Percent Solids for Spike Sample:	75.7

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Phosphorus, Total	mg/Kg	+/-20	93.2	D	93.0	D	2	0.21		05/28/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q1872
Project:	NJ Soil PT	Sample ID:	Q2130-01
Client ID:	TP-3DUP	Percent Solids for Spike Sample:	90.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.076	U	0.076	U	1	0		05/28/2025