



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872
Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV21	Mercury	3.84	4.0	96	90 - 110	CV	05/29/2025	11:18	LB135943

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872
Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV76	Mercury	4.83	5.0	97	90 - 110	CV	05/29/2025	11:22	LB135943
CCV77	Mercury	5.09	5.0	102	90 - 110	CV	05/29/2025	12:25	LB135943
CCV78	Mercury	5.16	5.0	103	90 - 110	CV	05/29/2025	13:01	LB135943
CCV79	Mercury	5.26	5.0	105	90 - 110	CV	05/29/2025	13:19	LB135943

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872

Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2450	2500	98	90 - 110	P	05/30/2025	15:13	LB135969
	Antimony	1020	1000	102	90 - 110	P	05/30/2025	15:13	LB135969
	Arsenic	1030	1000	103	90 - 110	P	05/30/2025	15:13	LB135969
	Barium	490	520	94	90 - 110	P	05/30/2025	15:13	LB135969
	Beryllium	495	510	97	90 - 110	P	05/30/2025	15:13	LB135969
	Boron	2290	2500	92	90 - 110	P	05/30/2025	15:13	LB135969
	Cadmium	518	510	102	90 - 110	P	05/30/2025	15:13	LB135969
	Calcium	9850	10000	98	90 - 110	P	05/30/2025	15:13	LB135969
	Chromium	539	520	104	90 - 110	P	05/30/2025	15:13	LB135969
	Cobalt	520	520	100	90 - 110	P	05/30/2025	15:13	LB135969
	Copper	535	510	105	90 - 110	P	05/30/2025	15:13	LB135969
	Iron	10500	10000	105	90 - 110	P	05/30/2025	15:13	LB135969
	Lead	1010	1000	101	90 - 110	P	05/30/2025	15:13	LB135969
	Lithium	2460	2500	98	90 - 110	P	05/30/2025	15:13	LB135969
	Magnesium	5760	6000	96	90 - 110	P	05/30/2025	15:13	LB135969
	Manganese	504	520	97	90 - 110	P	05/30/2025	15:13	LB135969
	Molybdenum	2520	2500	101	90 - 110	P	05/30/2025	15:13	LB135969
	Nickel	524	530	99	90 - 110	P	05/30/2025	15:13	LB135969
	Potassium	10100	9900	102	90 - 110	P	05/30/2025	15:13	LB135969
	Selenium	1080	1000	108	90 - 110	P	05/30/2025	15:13	LB135969
	Silver	259	250	104	90 - 110	P	05/30/2025	15:13	LB135969
	Sodium	9750	10000	98	90 - 110	P	05/30/2025	15:13	LB135969
	Strontium	2390	2500	95	90 - 110	P	05/30/2025	15:13	LB135969
	Thallium	1090	1000	109	90 - 110	P	05/30/2025	15:13	LB135969
	Tin	2460	2500	98	90 - 110	P	05/30/2025	15:13	LB135969
	Titanium	2340	2500	94	90 - 110	P	05/30/2025	15:13	LB135969
	Vanadium	484	500	97	90 - 110	P	05/30/2025	15:13	LB135969
	Zinc	1000	1000	100	90 - 110	P	05/30/2025	15:13	LB135969

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872

Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	98.2	100	98	80 - 120	P	05/30/2025	15:21	LB135969
	Antimony	50.0	50.0	100	80 - 120	P	05/30/2025	15:21	LB135969
	Arsenic	20.5	20.0	102	80 - 120	P	05/30/2025	15:21	LB135969
	Barium	91.3	100	91	80 - 120	P	05/30/2025	15:21	LB135969
	Beryllium	5.89	6.0	98	80 - 120	P	05/30/2025	15:21	LB135969
	Boron	97.3	100	97	80 - 120	P	05/30/2025	15:21	LB135969
	Cadmium	6.08	6.0	101	80 - 120	P	05/30/2025	15:21	LB135969
	Calcium	1910	2000	95	80 - 120	P	05/30/2025	15:21	LB135969
	Chromium	10.2	10.0	102	80 - 120	P	05/30/2025	15:21	LB135969
	Cobalt	29.3	30.0	98	80 - 120	P	05/30/2025	15:21	LB135969
	Copper	21.5	20.0	107	80 - 120	P	05/30/2025	15:21	LB135969
	Iron	104	100	104	80 - 120	P	05/30/2025	15:21	LB135969
	Lead	11.3	12.0	94	80 - 120	P	05/30/2025	15:21	LB135969
	Lithium	21.2	20.0	106	80 - 120	P	05/30/2025	15:21	LB135969
	Magnesium	1920	2000	96	80 - 120	P	05/30/2025	15:21	LB135969
	Manganese	19.9	20.0	100	80 - 120	P	05/30/2025	15:21	LB135969
	Molybdenum	190	200	95	80 - 120	P	05/30/2025	15:21	LB135969
	Nickel	40.0	40.0	100	80 - 120	P	05/30/2025	15:21	LB135969
	Potassium	1920	2000	96	80 - 120	P	05/30/2025	15:21	LB135969
	Selenium	21.2	20.0	106	80 - 120	P	05/30/2025	15:21	LB135969
	Silver	10.1	10.0	101	80 - 120	P	05/30/2025	15:21	LB135969
	Sodium	1830	2000	92	80 - 120	P	05/30/2025	15:21	LB135969
	Strontium	19.9	20.0	99	80 - 120	P	05/30/2025	15:21	LB135969
	Thallium	33.2	40.0	83	80 - 120	P	05/30/2025	15:21	LB135969
	Tin	36.4	40.0	91	80 - 120	P	05/30/2025	15:21	LB135969
	Titanium	36.8	40.0	92	80 - 120	P	05/30/2025	15:21	LB135969
	Vanadium	36.9	40.0	92	80 - 120	P	05/30/2025	15:21	LB135969
	Zinc	40.5	40.0	101	80 - 120	P	05/30/2025	15:21	LB135969

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872

Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9760	10000	98	90 - 110	P	05/30/2025	16:02	LB135969
	Antimony	5090	5000	102	90 - 110	P	05/30/2025	16:02	LB135969
	Arsenic	5110	5000	102	90 - 110	P	05/30/2025	16:02	LB135969
	Barium	9890	10000	99	90 - 110	P	05/30/2025	16:02	LB135969
	Beryllium	235	250	94	90 - 110	P	05/30/2025	16:02	LB135969
	Boron	4640	5000	93	90 - 110	P	05/30/2025	16:02	LB135969
	Cadmium	2500	2500	100	90 - 110	P	05/30/2025	16:02	LB135969
	Calcium	24400	25000	98	90 - 110	P	05/30/2025	16:02	LB135969
	Chromium	1020	1000	102	90 - 110	P	05/30/2025	16:02	LB135969
	Cobalt	2500	2500	100	90 - 110	P	05/30/2025	16:02	LB135969
	Copper	1280	1250	102	90 - 110	P	05/30/2025	16:02	LB135969
	Iron	5420	5000	108	90 - 110	P	05/30/2025	16:02	LB135969
	Lead	5010	5000	100	90 - 110	P	05/30/2025	16:02	LB135969
	Lithium	4790	5000	96	90 - 110	P	05/30/2025	16:02	LB135969
	Magnesium	24100	25000	96	90 - 110	P	05/30/2025	16:02	LB135969
	Manganese	2430	2500	97	90 - 110	P	05/30/2025	16:02	LB135969
	Molybdenum	5090	5000	102	90 - 110	P	05/30/2025	16:02	LB135969
	Nickel	2510	2500	100	90 - 110	P	05/30/2025	16:02	LB135969
	Potassium	27000	25000	108	90 - 110	P	05/30/2025	16:02	LB135969
	Selenium	5160	5000	103	90 - 110	P	05/30/2025	16:02	LB135969
	Silver	1280	1250	102	90 - 110	P	05/30/2025	16:02	LB135969
	Sodium	26700	25000	107	90 - 110	P	05/30/2025	16:02	LB135969
	Strontium	4970	5000	99	90 - 110	P	05/30/2025	16:02	LB135969
	Thallium	5190	5000	104	90 - 110	P	05/30/2025	16:02	LB135969
	Tin	5040	5000	101	90 - 110	P	05/30/2025	16:02	LB135969
	Titanium	4880	5000	98	90 - 110	P	05/30/2025	16:02	LB135969
	Vanadium	2460	2500	98	90 - 110	P	05/30/2025	16:02	LB135969
	Zinc	2680	2500	107	90 - 110	P	05/30/2025	16:02	LB135969
CCV02	Aluminum	10000	10000	100	90 - 110	P	05/30/2025	16:52	LB135969
	Antimony	5130	5000	103	90 - 110	P	05/30/2025	16:52	LB135969
	Arsenic	5140	5000	103	90 - 110	P	05/30/2025	16:52	LB135969
	Barium	9860	10000	99	90 - 110	P	05/30/2025	16:52	LB135969
	Beryllium	249	250	100	90 - 110	P	05/30/2025	16:52	LB135969
	Boron	4900	5000	98	90 - 110	P	05/30/2025	16:52	LB135969

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872

Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Cadmium	2510	2500	100	90 - 110	P	05/30/2025	16:52	LB135969
	Calcium	24600	25000	99	90 - 110	P	05/30/2025	16:52	LB135969
	Chromium	1020	1000	102	90 - 110	P	05/30/2025	16:52	LB135969
	Cobalt	2510	2500	100	90 - 110	P	05/30/2025	16:52	LB135969
	Copper	1290	1250	103	90 - 110	P	05/30/2025	16:52	LB135969
	Iron	5250	5000	105	90 - 110	P	05/30/2025	16:52	LB135969
	Lead	5030	5000	101	90 - 110	P	05/30/2025	16:52	LB135969
	Lithium	4950	5000	99	90 - 110	P	05/30/2025	16:52	LB135969
	Magnesium	24500	25000	98	90 - 110	P	05/30/2025	16:52	LB135969
	Manganese	2450	2500	98	90 - 110	P	05/30/2025	16:52	LB135969
	Molybdenum	5120	5000	102	90 - 110	P	05/30/2025	16:52	LB135969
	Nickel	2510	2500	100	90 - 110	P	05/30/2025	16:52	LB135969
	Potassium	26400	25000	106	90 - 110	P	05/30/2025	16:52	LB135969
	Selenium	5210	5000	104	90 - 110	P	05/30/2025	16:52	LB135969
	Silver	1290	1250	103	90 - 110	P	05/30/2025	16:52	LB135969
	Sodium	26000	25000	104	90 - 110	P	05/30/2025	16:52	LB135969
	Strontium	4990	5000	100	90 - 110	P	05/30/2025	16:52	LB135969
	Thallium	5160	5000	103	90 - 110	P	05/30/2025	16:52	LB135969
	Tin	5050	5000	101	90 - 110	P	05/30/2025	16:52	LB135969
	Titanium	4920	5000	98	90 - 110	P	05/30/2025	16:52	LB135969
CCV03	Vanadium	2490	2500	100	90 - 110	P	05/30/2025	16:52	LB135969
	Zinc	2610	2500	104	90 - 110	P	05/30/2025	16:52	LB135969
	Aluminum	10200	10000	102	90 - 110	P	05/30/2025	17:41	LB135969
	Antimony	5110	5000	102	90 - 110	P	05/30/2025	17:41	LB135969
	Arsenic	5080	5000	102	90 - 110	P	05/30/2025	17:41	LB135969
	Barium	9790	10000	98	90 - 110	P	05/30/2025	17:41	LB135969
	Beryllium	272	250	109	90 - 110	P	05/30/2025	17:41	LB135969
	Boron	5390	5000	108	90 - 110	P	05/30/2025	17:41	LB135969
	Cadmium	2510	2500	100	90 - 110	P	05/30/2025	17:41	LB135969
	Calcium	25300	25000	101	90 - 110	P	05/30/2025	17:41	LB135969
	Chromium	1030	1000	103	90 - 110	P	05/30/2025	17:41	LB135969
	Cobalt	2510	2500	100	90 - 110	P	05/30/2025	17:41	LB135969
	Copper	1280	1250	102	90 - 110	P	05/30/2025	17:41	LB135969
	Iron	4910	5000	98	90 - 110	P	05/30/2025	17:41	LB135969

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872

Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV03	Lead	5020	5000	100	90 - 110	P	05/30/2025	17:41	LB135969
	Lithium	5120	5000	102	90 - 110	P	05/30/2025	17:41	LB135969
	Magnesium	25600	25000	102	90 - 110	P	05/30/2025	17:41	LB135969
	Manganese	2510	2500	101	90 - 110	P	05/30/2025	17:41	LB135969
	Molybdenum	5140	5000	103	90 - 110	P	05/30/2025	17:41	LB135969
	Nickel	2510	2500	100	90 - 110	P	05/30/2025	17:41	LB135969
	Potassium	24200	25000	97	90 - 110	P	05/30/2025	17:41	LB135969
	Selenium	5130	5000	102	90 - 110	P	05/30/2025	17:41	LB135969
	Silver	1290	1250	103	90 - 110	P	05/30/2025	17:41	LB135969
	Sodium	24100	25000	96	90 - 110	P	05/30/2025	17:41	LB135969
	Strontium	4960	5000	99	90 - 110	P	05/30/2025	17:41	LB135969
	Thallium	4980	5000	100	90 - 110	P	05/30/2025	17:41	LB135969
	Tin	5020	5000	100	90 - 110	P	05/30/2025	17:41	LB135969
	Titanium	5000	5000	100	90 - 110	P	05/30/2025	17:41	LB135969
	Vanadium	2530	2500	101	90 - 110	P	05/30/2025	17:41	LB135969
	Zinc	2510	2500	100	90 - 110	P	05/30/2025	17:41	LB135969
CCV04	Aluminum	9760	10000	98	90 - 110	P	05/30/2025	18:28	LB135969
	Antimony	5120	5000	102	90 - 110	P	05/30/2025	18:28	LB135969
	Arsenic	5120	5000	102	90 - 110	P	05/30/2025	18:28	LB135969
	Barium	9930	10000	99	90 - 110	P	05/30/2025	18:28	LB135969
	Beryllium	239	250	95	90 - 110	P	05/30/2025	18:28	LB135969
	Boron	4730	5000	95	90 - 110	P	05/30/2025	18:28	LB135969
	Cadmium	2490	2500	100	90 - 110	P	05/30/2025	18:28	LB135969
	Calcium	24300	25000	97	90 - 110	P	05/30/2025	18:28	LB135969
	Chromium	1010	1000	102	90 - 110	P	05/30/2025	18:28	LB135969
	Cobalt	2490	2500	100	90 - 110	P	05/30/2025	18:28	LB135969
	Copper	1280	1250	102	90 - 110	P	05/30/2025	18:28	LB135969
	Iron	5320	5000	106	90 - 110	P	05/30/2025	18:28	LB135969
	Lead	4980	5000	100	90 - 110	P	05/30/2025	18:28	LB135969
	Lithium	4850	5000	97	90 - 110	P	05/30/2025	18:28	LB135969
	Magnesium	24000	25000	96	90 - 110	P	05/30/2025	18:28	LB135969
	Manganese	2430	2500	97	90 - 110	P	05/30/2025	18:28	LB135969
	Molybdenum	5100	5000	102	90 - 110	P	05/30/2025	18:28	LB135969
	Nickel	2490	2500	100	90 - 110	P	05/30/2025	18:28	LB135969

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872

Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Potassium	26200	25000	105	90 - 110	P	05/30/2025	18:28	LB135969
	Selenium	5200	5000	104	90 - 110	P	05/30/2025	18:28	LB135969
	Silver	1280	1250	102	90 - 110	P	05/30/2025	18:28	LB135969
	Sodium	26400	25000	106	90 - 110	P	05/30/2025	18:28	LB135969
	Strontium	4970	5000	99	90 - 110	P	05/30/2025	18:28	LB135969
	Thallium	5110	5000	102	90 - 110	P	05/30/2025	18:28	LB135969
	Tin	5000	5000	100	90 - 110	P	05/30/2025	18:28	LB135969
	Titanium	4880	5000	98	90 - 110	P	05/30/2025	18:28	LB135969
	Vanadium	2460	2500	98	90 - 110	P	05/30/2025	18:28	LB135969
	Zinc	2510	2500	100	90 - 110	P	05/30/2025	18:28	LB135969
CCV05	Aluminum	10000	10000	100	90 - 110	P	05/30/2025	18:58	LB135969
	Antimony	5080	5000	102	90 - 110	P	05/30/2025	18:58	LB135969
	Arsenic	5070	5000	101	90 - 110	P	05/30/2025	18:58	LB135969
	Barium	9520	10000	95	90 - 110	P	05/30/2025	18:58	LB135969
	Beryllium	266	250	106	90 - 110	P	05/30/2025	18:58	LB135969
	Boron	5280	5000	106	90 - 110	P	05/30/2025	18:58	LB135969
	Cadmium	2480	2500	99	90 - 110	P	05/30/2025	18:58	LB135969
	Calcium	24200	25000	97	90 - 110	P	05/30/2025	18:58	LB135969
	Chromium	1000	1000	100	90 - 110	P	05/30/2025	18:58	LB135969
	Cobalt	2480	2500	99	90 - 110	P	05/30/2025	18:58	LB135969
	Copper	1280	1250	102	90 - 110	P	05/30/2025	18:58	LB135969
	Iron	4710	5000	94	90 - 110	P	05/30/2025	18:58	LB135969
	Lead	4960	5000	99	90 - 110	P	05/30/2025	18:58	LB135969
	Lithium	4990	5000	100	90 - 110	P	05/30/2025	18:58	LB135969
	Magnesium	24400	25000	98	90 - 110	P	05/30/2025	18:58	LB135969
	Manganese	2430	2500	97	90 - 110	P	05/30/2025	18:58	LB135969
	Molybdenum	5040	5000	101	90 - 110	P	05/30/2025	18:58	LB135969
	Nickel	2480	2500	99	90 - 110	P	05/30/2025	18:58	LB135969
	Potassium	24000	25000	96	90 - 110	P	05/30/2025	18:58	LB135969
	Selenium	5150	5000	103	90 - 110	P	05/30/2025	18:58	LB135969
	Silver	1260	1250	101	90 - 110	P	05/30/2025	18:58	LB135969
	Sodium	23200	25000	93	90 - 110	P	05/30/2025	18:58	LB135969
	Strontium	4880	5000	98	90 - 110	P	05/30/2025	18:58	LB135969
	Thallium	4950	5000	99	90 - 110	P	05/30/2025	18:58	LB135969

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872

Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Tin	4970	5000	99	90 - 110	P	05/30/2025	18:58	LB135969
	Titanium	4850	5000	97	90 - 110	P	05/30/2025	18:58	LB135969
	Vanadium	2450	2500	98	90 - 110	P	05/30/2025	18:58	LB135969
	Zinc	2570	2500	103	90 - 110	P	05/30/2025	18:58	LB135969
CCV06	Aluminum	9730	10000	97	90 - 110	P	05/30/2025	19:30	LB135969
	Antimony	4980	5000	100	90 - 110	P	05/30/2025	19:30	LB135969
	Arsenic	4950	5000	99	90 - 110	P	05/30/2025	19:30	LB135969
	Barium	9770	10000	98	90 - 110	P	05/30/2025	19:30	LB135969
	Beryllium	246	250	98	90 - 110	P	05/30/2025	19:30	LB135969
	Boron	4890	5000	98	90 - 110	P	05/30/2025	19:30	LB135969
	Cadmium	2460	2500	98	90 - 110	P	05/30/2025	19:30	LB135969
	Calcium	24300	25000	97	90 - 110	P	05/30/2025	19:30	LB135969
	Chromium	985	1000	98	90 - 110	P	05/30/2025	19:30	LB135969
	Cobalt	2460	2500	98	90 - 110	P	05/30/2025	19:30	LB135969
	Copper	1250	1250	100	90 - 110	P	05/30/2025	19:30	LB135969
	Iron	4920	5000	98	90 - 110	P	05/30/2025	19:30	LB135969
	Lead	4930	5000	99	90 - 110	P	05/30/2025	19:30	LB135969
	Lithium	4830	5000	97	90 - 110	P	05/30/2025	19:30	LB135969
	Magnesium	24100	25000	96	90 - 110	P	05/30/2025	19:30	LB135969
	Manganese	2440	2500	97	90 - 110	P	05/30/2025	19:30	LB135969
	Molybdenum	4990	5000	100	90 - 110	P	05/30/2025	19:30	LB135969
	Nickel	2460	2500	98	90 - 110	P	05/30/2025	19:30	LB135969
	Potassium	24800	25000	99	90 - 110	P	05/30/2025	19:30	LB135969
	Selenium	5010	5000	100	90 - 110	P	05/30/2025	19:30	LB135969
	Silver	1230	1250	98	90 - 110	P	05/30/2025	19:30	LB135969
	Sodium	24200	25000	97	90 - 110	P	05/30/2025	19:30	LB135969
	Strontium	4930	5000	99	90 - 110	P	05/30/2025	19:30	LB135969
	Thallium	4850	5000	97	90 - 110	P	05/30/2025	19:30	LB135969
	Tin	4930	5000	99	90 - 110	P	05/30/2025	19:30	LB135969
	Titanium	4850	5000	97	90 - 110	P	05/30/2025	19:30	LB135969
	Vanadium	2440	2500	98	90 - 110	P	05/30/2025	19:30	LB135969
	Zinc	2520	2500	101	90 - 110	P	05/30/2025	19:30	LB135969