



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Alliance Technical Group, LLC - Newark **SDG No.:** Q1872
Contract: ALLI03 **Lab Code:** CHEM **Case No.:** Q1872 **SAS No.:** Q1872
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Aluminum	20.6	20.0	103	70 - 130	P	05/23/2025	15:01	LB135941
	Antimony	1.93	2.0	96	70 - 130	P	05/23/2025	15:01	LB135941
	Arsenic	0.96	1.0	96	70 - 130	P	05/23/2025	15:01	LB135941
	Barium	9.30	10.0	93	70 - 130	P	05/23/2025	15:01	LB135941
	Beryllium	1.10	1.0	110	70 - 130	P	05/23/2025	15:01	LB135941
	Cadmium	0.84	1.0	84	70 - 130	P	05/23/2025	15:01	LB135941
	Calcium	474	500	95	70 - 130	P	05/23/2025	15:01	LB135941
	Chromium	1.95	2.0	98	70 - 130	P	05/23/2025	15:01	LB135941
	Cobalt	1.04	1.0	104	50 - 150	P	05/23/2025	15:01	LB135941
	Copper	2.39	2.0	120	70 - 130	P	05/23/2025	15:01	LB135941
	Iron	52.6	50.0	105	70 - 130	P	05/23/2025	15:01	LB135941
	Lead	0.81	1.0	81	70 - 130	P	05/23/2025	15:01	LB135941
	Magnesium	524	500	105	70 - 130	P	05/23/2025	15:01	LB135941
	Manganese	1.05	1.0	105	50 - 150	P	05/23/2025	15:01	LB135941
	Molybdenum	4.65	5.0	93	70 - 130	P	05/23/2025	15:01	LB135941
	Nickel	1.15	1.0	115	70 - 130	P	05/23/2025	15:01	LB135941
	Potassium	493	500	99	70 - 130	P	05/23/2025	15:01	LB135941
	Selenium	5.09	5.0	102	70 - 130	P	05/23/2025	15:01	LB135941
	Silver	1.00	1.0	100	70 - 130	P	05/23/2025	15:01	LB135941
	Sodium	526	500	105	70 - 130	P	05/23/2025	15:01	LB135941
	Strontium	0.94	1.0	94	70 - 130	P	05/23/2025	15:01	LB135941
	Thallium	0.93	1.0	93	70 - 130	P	05/23/2025	15:01	LB135941
	Tin	4.72	5.0	94	70 - 130	P	05/23/2025	15:01	LB135941
	Vanadium	4.86	5.0	97	70 - 130	P	05/23/2025	15:01	LB135941
	Zinc	5.75	5.0	115	50 - 150	P	05/23/2025	15:01	LB135941