



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark SDG No.: Q1872
Contract: ALLI03 Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	1.94	+/-10	U	20.0	P	05/23/2025	14:37	LB135941
	Antimony	0.11	+/-1	U	2.00	P	05/23/2025	14:37	LB135941
	Arsenic	0.089	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Barium	0.21	+/-5	U	10.0	P	05/23/2025	14:37	LB135941
	Beryllium	0.32	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Cadmium	0.34	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Calcium	45.7	+/-250	U	500	P	05/23/2025	14:37	LB135941
	Chromium	0.21	+/-1	U	2.00	P	05/23/2025	14:37	LB135941
	Cobalt	0.070	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Copper	0.30	+/-1	U	2.00	P	05/23/2025	14:37	LB135941
	Iron	7.81	+/-25	U	50.0	P	05/23/2025	14:37	LB135941
	Lead	0.21	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Magnesium	19.5	+/-250	U	500	P	05/23/2025	14:37	LB135941
	Manganese	0.43	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Nickel	0.27	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Potassium	36.4	+/-250	U	500	P	05/23/2025	14:37	LB135941
	Selenium	2.90	+/-2.5	U	5.00	P	05/23/2025	14:37	LB135941
	Tin	0.31	+/-2.5	U	5.00	P	05/23/2025	14:37	LB135941
	Molybdenum	0.27	+/-2.5	U	5.00	P	05/23/2025	14:37	LB135941
	Silver	0.060	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Sodium	128	+/-250	U	500	P	05/23/2025	14:37	LB135941
	Thallium	0.060	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941
	Vanadium	0.077	+/-2.5	U	5.00	P	05/23/2025	14:37	LB135941
	Zinc	1.25	+/-2.5	U	5.00	P	05/23/2025	14:37	LB135941
	Strontium	0.084	+/-0.5	U	1.00	P	05/23/2025	14:37	LB135941

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q1872

Contract: ALLI03

Lab Code: CHEM

Case No.: Q1872

SAS No.: Q1872

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	1.94	+/-10	U	20.0	P	05/23/2025	14:58	LB135941
	Antimony	0.11	+/-1	U	2.00	P	05/23/2025	14:58	LB135941
	Arsenic	0.089	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Barium	0.21	+/-5	U	10.0	P	05/23/2025	14:58	LB135941
	Beryllium	0.32	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Cadmium	0.34	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Calcium	45.7	+/-250	U	500	P	05/23/2025	14:58	LB135941
	Chromium	0.21	+/-1	U	2.00	P	05/23/2025	14:58	LB135941
	Cobalt	0.070	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Copper	0.30	+/-1	U	2.00	P	05/23/2025	14:58	LB135941
	Iron	7.81	+/-25	U	50.0	P	05/23/2025	14:58	LB135941
	Lead	0.21	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Magnesium	19.5	+/-250	U	500	P	05/23/2025	14:58	LB135941
	Manganese	0.43	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Nickel	0.27	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Potassium	36.4	+/-250	U	500	P	05/23/2025	14:58	LB135941
	Selenium	2.90	+/-2.5	U	5.00	P	05/23/2025	14:58	LB135941
	Molybdenum	0.27	+/-2.5	U	5.00	P	05/23/2025	14:58	LB135941
	Tin	0.31	+/-2.5	U	5.00	P	05/23/2025	14:58	LB135941
	Silver	0.060	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Sodium	128	+/-250	U	500	P	05/23/2025	14:58	LB135941
	Thallium	0.060	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
	Vanadium	0.077	+/-2.5	U	5.00	P	05/23/2025	14:58	LB135941
	Zinc	1.25	+/-2.5	U	5.00	P	05/23/2025	14:58	LB135941
	Strontium	0.084	+/-0.5	U	1.00	P	05/23/2025	14:58	LB135941
CCB02	Aluminum	1.94	+/-10	U	20.0	P	05/23/2025	15:55	LB135941
	Antimony	0.11	+/-1	U	2.00	P	05/23/2025	15:55	LB135941
	Arsenic	0.089	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941
	Barium	0.21	+/-5	U	10.0	P	05/23/2025	15:55	LB135941
	Beryllium	0.32	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941
	Cadmium	0.34	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941
	Calcium	45.7	+/-250	U	500	P	05/23/2025	15:55	LB135941
	Chromium	0.21	+/-1	U	2.00	P	05/23/2025	15:55	LB135941
	Cobalt	0.070	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941
	Copper	0.30	+/-1	U	2.00	P	05/23/2025	15:55	LB135941
	Iron	7.81	+/-25	U	50.0	P	05/23/2025	15:55	LB135941
	Lead	0.21	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941
	Magnesium	19.5	+/-250	U	500	P	05/23/2025	15:55	LB135941
	Manganese	0.43	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q1872

Contract: ALLI03

Lab Code: CHEM

Case No.: Q1872

SAS No.: Q1872

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Nickel	0.27	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941
	Potassium	36.4	+/-250	U	500	P	05/23/2025	15:55	LB135941
	Selenium	2.90	+/-2.5	U	5.00	P	05/23/2025	15:55	LB135941
	Tin	0.31	+/-2.5	U	5.00	P	05/23/2025	15:55	LB135941
	Molybdenum	0.27	+/-2.5	U	5.00	P	05/23/2025	15:55	LB135941
	Silver	0.060	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941
	Sodium	128	+/-250	U	500	P	05/23/2025	15:55	LB135941
	Thallium	0.070	+/-0.5	J	1.00	P	05/23/2025	15:55	LB135941
	Vanadium	0.077	+/-2.5	U	5.00	P	05/23/2025	15:55	LB135941
	Zinc	1.25	+/-2.5	U	5.00	P	05/23/2025	15:55	LB135941
	Strontium	0.084	+/-0.5	U	1.00	P	05/23/2025	15:55	LB135941