

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q1872

Contract: ALLI03

Lab Code: CHEM

Case No.: Q1872

SAS No.: Q1872

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168134BS							
Aluminum	mg/Kg	1000	1010		101	80 - 120	P
Antimony	mg/Kg	50.0	49.8		100	80 - 120	P
Arsenic	mg/Kg	50.0	49.5		99	80 - 120	P
Barium	mg/Kg	250	251		100	80 - 120	P
Beryllium	mg/Kg	50.0	49.6		99	80 - 120	P
Cadmium	mg/Kg	50.0	51.7		103	80 - 120	P
Calcium	mg/Kg	5000	4930		99	80 - 120	P
Chromium	mg/Kg	50.0	51.0		102	80 - 120	P
Cobalt	mg/Kg	50.0	50.9		102	80 - 120	P
Copper	mg/Kg	500	508		102	80 - 120	P
Iron	mg/Kg	2500	2640		106	80 - 120	P
Lead	mg/Kg	250	248		99	80 - 120	P
Magnesium	mg/Kg	5000	5180		104	80 - 120	P
Manganese	mg/Kg	500	507		101	80 - 120	P
Nickel	mg/Kg	50.0	47.4		95	80 - 120	P
Potassium	mg/Kg	2500	2560		102	80 - 120	P
Molybdenum	mg/Kg	500	493		99	80 - 120	P
Tin	mg/Kg	50.0	50.1		100	80 - 120	P
Selenium	mg/Kg	50.0	49.5		99	80 - 120	P
Silver	mg/Kg	50.0	50.8		102	80 - 120	P
Sodium	mg/Kg	5000	5150		103	80 - 120	P
Thallium	mg/Kg	50.0	49.2		98	80 - 120	P
Vanadium	mg/Kg	50.0	50.9		102	80 - 120	P
Zinc	mg/Kg	500	507		101	80 - 120	P
Strontium	mg/Kg	50.0	49.1		98	80 - 120	P