

Continuing Calibration Report for SequenceID : FC050825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068791.D

Aliphatic C9-C12	5774515.000	60.000	3.342	6.642	96241.917	99133.167	2.917
Aliphatic C12-C16	3617334.000	40.000	6.643	10.042	90433.350	90984.404	0.606
Aliphatic C16-C21	5106859.000	60.000	10.043	13.408	85114.317	85760.459	0.753
Aliphatic C21-C28	6727060.000	80.000	13.409	17.072	84088.250	83091.075	-1.200
Aliphatic C28-C40	10872761.000	120.000	17.073	22.071	90606.342	84622.692	-7.071
Aliphatic EPH	32098529.000	360.000	3.342	22.071	89162.581	87597.230	-1.787