

Continuing Calibration Report for SequenceID : FC050825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068799.D

Aliphatic C9-C12	5278125.000	60.000	3.342	6.642	87968.750	99133.167	11.262
Aliphatic C12-C16	3477308.000	40.000	6.643	10.042	86932.700	90984.404	4.453
Aliphatic C16-C21	5056855.000	60.000	10.043	13.408	84280.917	85760.459	1.725
Aliphatic C21-C28	6614239.000	80.000	13.409	17.072	82677.988	83091.075	0.497
Aliphatic C28-C40	10921137.000	120.000	17.073	22.071	91009.475	84622.692	-7.547
Aliphatic EPH	31347664.000	360.000	3.342	22.071	87076.844	87597.230	0.594