

Continuing Calibration Report for SequenceID : FC050825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068808.D

Aliphatic C9-C12	5298914.000	60.000	3.342	6.642	88315.233	99133.167	10.913
Aliphatic C12-C16	3471710.000	40.000	6.643	10.042	86792.750	90984.404	4.607
Aliphatic C16-C21	5061619.000	60.000	10.043	13.408	84360.317	85760.459	1.633
Aliphatic C21-C28	6674646.000	80.000	13.409	17.072	83433.075	83091.075	-0.412
Aliphatic C28-C40	11150242.000	120.000	17.073	22.071	92918.683	84622.692	-9.804
Aliphatic EPH	31657131.000	360.000	3.342	22.071	87936.475	87597.230	-0.387