

Continuing Calibration Report for SequenceID : FD050825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049318.D

Aromatic C10-C12	5420453.000	40.000	4.431	6.176	135511.325	128531.556	-5.430
Aromatic C12-C16	8502870.000	60.000	6.177	8.805	141714.500	135727.245	-4.411
Aromatic C16-C21	10349187.000	80.000	8.806	13.095	129364.838	131120.372	1.339
Aromatic C21-C36	20561899.000	180.000	13.0956	18.515	114232.772	117353.852	2.660
Aromatic EPH	44834409.000	360.000	4.431	18.515	124540.025	124717.278	0.142