

Continuing Calibration Report for SequenceID : FD050825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049325.D

Aromatic C10-C12	5652522.000	40.000	4.431	6.176	141313.050	128531.556	-9.944
Aromatic C12-C16	8979975.000	60.000	6.177	8.805	149666.250	135727.245	-10.270
Aromatic C16-C21	11477164.000	80.000	8.806	13.095	143464.550	131120.372	-9.414
Aromatic C21-C36	22885019.000	180.000	13.0956	18.515	127138.994	117353.852	-8.338
Aromatic EPH	48994680.000	360.000	4.431	18.515	136096.333	124717.278	-9.124