

Continuing Calibration Report for SequenceID : FD050825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049334.D

Aromatic C10-C12	5734262.000	40.000	4.431	6.176	143356.550	128531.556	-11.534
Aromatic C12-C16	9081487.000	60.000	6.177	8.805	151358.117	135727.245	-11.516
Aromatic C16-C21	11630703.000	80.000	8.806	13.095	145383.788	131120.372	-10.878
Aromatic C21-C36	23316285.000	180.000	13.0956	18.515	129534.917	117353.852	-10.380
Aromatic EPH	49762737.000	360.000	4.431	18.515	138229.825	124717.278	-10.835