

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 06/04/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 14:23 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND          | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|-------------------|------------|-----------|-----------|-------|------------|
|                   |            |           | FROM      | TO    |            |
| DICAMBA           | 7.56       | 7.56      | 7.46      | 7.66  | 0.00       |
| MCPD              | 7.74       | 7.74      | 7.64      | 7.84  | 0.00       |
| 2,4-DCAA          | 7.37       | 7.37      | 7.27      | 7.47  | 0.00       |
| Dalapon           | 2.72       | 2.72      | 2.62      | 2.82  | 0.00       |
| MCPA              | 7.89       | 7.90      | 7.80      | 8.00  | 0.01       |
| DICHLORPROP       | 8.28       | 8.28      | 8.18      | 8.38  | 0.00       |
| 2,4-D             | 8.51       | 8.51      | 8.41      | 8.61  | 0.00       |
| 2,4,5-TP(Silvex)  | 9.40       | 9.40      | 9.30      | 9.50  | 0.00       |
| 2,4,5-T           | 9.70       | 9.70      | 9.60      | 9.80  | 0.00       |
| 2,4-DB            | 10.28      | 10.28     | 10.18     | 10.38 | 0.00       |
| Dinoseb           | 11.50      | 11.50     | 11.40     | 11.60 | 0.00       |
| Pentachlorophenol | 8.82       | 8.82      | 8.72      | 8.92  | 0.00       |
| 4-Nitrophenol     | 7.17       | 7.17      | 7.07      | 7.27  | 0.00       |
| PICLORAM          | 11.31      | 11.31     | 11.21     | 11.41 | 0.00       |
| DCPA              | 11.80      | 11.80     | 11.70     | 11.90 | 0.00       |
| 3,5-DICHLOROBENZ  | 6.52       | 6.52      | 6.42      | 6.62  | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 06/04/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 14:23 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND          | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|-------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA           | 7.98       | 7.98      | 7.88                   | 8.08  | 0.00       |
| MCPD              | 8.08       | 8.08      | 7.98                   | 8.18  | 0.00       |
| 2,4-DCAA          | 7.77       | 7.77      | 7.67                   | 7.87  | 0.00       |
| Dalapon           | 2.71       | 2.71      | 2.61                   | 2.81  | 0.00       |
| MCPA              | 8.33       | 8.33      | 8.23                   | 8.43  | 0.00       |
| DICHLORPROP       | 8.70       | 8.70      | 8.60                   | 8.80  | 0.00       |
| 2,4-D             | 9.03       | 9.03      | 8.93                   | 9.13  | 0.00       |
| 2,4,5-TP(Silvex)  | 9.94       | 9.94      | 9.84                   | 10.04 | 0.00       |
| 2,4,5-T           | 10.37      | 10.37     | 10.27                  | 10.47 | 0.00       |
| 2,4-DB            | 10.94      | 10.94     | 10.84                  | 11.04 | 0.00       |
| Dinoseb           | 11.32      | 11.32     | 11.22                  | 11.42 | 0.00       |
| Pentachlorophenol | 9.56       | 9.56      | 9.46                   | 9.66  | 0.00       |
| 4-Nitrophenol     | 7.30       | 7.30      | 7.20                   | 7.40  | 0.00       |
| PICLORAM          | 12.43      | 12.43     | 12.33                  | 12.53 | 0.00       |
| DCPA              | 12.37      | 12.37     | 12.27                  | 12.47 | 0.00       |
| 3,5-DICHLOROBENZ  | 6.72       | 6.72      | 6.62                   | 6.82  | 0.00       |