

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 06/05/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 00:04 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.56	7.56	7.46	7.66	0.00
MCP	7.74	7.74	7.64	7.84	0.00
2,4-DCAA	7.36	7.37	7.27	7.47	0.01
Dalapon	2.72	2.72	2.62	2.82	0.01
MCPA	7.89	7.90	7.80	8.00	0.01
DICHLORPROP	8.27	8.28	8.18	8.38	0.01
2,4-D	8.50	8.51	8.41	8.61	0.01
2,4,5-TP(Silvex)	9.40	9.40	9.30	9.50	0.00
2,4,5-T	9.69	9.70	9.60	9.80	0.01
2,4-DB	10.27	10.28	10.18	10.38	0.01
Dinoseb	11.49	11.50	11.40	11.60	0.01
Pentachlorophenol	8.81	8.82	8.72	8.92	0.01
4-Nitrophenol	7.16	7.17	7.07	7.27	0.01
PICLORAM	11.30	11.31	11.21	11.41	0.01
DCPA	11.79	11.80	11.70	11.90	0.01
3,5-DICHLOROBENZ	6.52	6.52	6.42	6.62	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 06/05/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 00:04 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.98	7.98	7.88	8.08	0.00
MCP	8.08	8.08	7.98	8.18	0.01
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
Dalapon	2.71	2.71	2.61	2.81	0.00
MCPA	8.33	8.33	8.23	8.43	0.01
DICHLORPROP	8.70	8.70	8.60	8.80	0.00
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.94	9.94	9.84	10.04	0.00
2,4,5-T	10.37	10.37	10.27	10.47	0.01
2,4-DB	10.93	10.94	10.84	11.04	0.01
Dinoseb	11.32	11.32	11.22	11.42	0.00
Pentachlorophenol	9.56	9.56	9.46	9.66	0.00
4-Nitrophenol	7.30	7.30	7.20	7.40	0.00
PICLORAM	12.43	12.43	12.33	12.53	0.00
DCPA	12.36	12.37	12.27	12.47	0.01
3,5-DICHLOROBENZ	6.72	6.72	6.62	6.82	0.00