

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030489.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.694	9.597	9.797	707.360	712.500	-0.7
2,4,5-TP(Silvex)	9.400	9.304	9.504	701.500	712.500	-1.5
2,4-D	8.508	8.412	8.612	688.650	705.000	-2.3
2,4-DB	10.275	10.179	10.379	711.140	712.500	-0.2
2,4-DCAA	7.366	7.269	7.469	719.780	750.000	-4.0
3,5-DICHLOROBENZOIC ACID	6.523	6.424	6.624	678.120	697.500	-2.8
4-Nitrophenol	7.165	7.067	7.267	657.710	682.500	-3.6
Dalapon	2.718	2.616	2.816	674.120	682.500	-1.2
DCPA	11.793	11.697	11.897	698.910	720.000	-2.9
DICAMBA	7.558	7.461	7.661	694.210	705.000	-1.5
DICHLORPROP	8.275	8.179	8.379	678.650	705.000	-3.7
Dinoseb	11.498	11.402	11.602	681.220	705.000	-3.4
MCPA	7.892	7.795	7.995	68.550	69.750	-1.7
MCPP	7.741	7.644	7.844	70.190	70.500	-0.4
Pentachlorophenol	8.817	8.720	8.920	748.910	712.500	5.1
PICLORAM	11.302	11.207	11.407	686.930	712.500	-3.6

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030489.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.367	10.268	10.468	697.940	712.500	-2.0
2,4,5-TP(Silvex)	9.940	9.842	10.042	690.320	712.500	-3.1
2,4-D	9.033	8.934	9.134	680.340	705.000	-3.5
2,4-DB	10.934	10.836	11.036	685.670	712.500	-3.8
2,4-DCAA	7.773	7.674	7.874	712.640	750.000	-5.0
3,5-DICHLOROBENZOIC ACID	6.716	6.617	6.817	680.790	697.500	-2.4
4-Nitrophenol	7.302	7.203	7.403	668.820	682.500	-2.0
Dalapon	2.711	2.610	2.810	654.200	682.500	-4.1
DCPA	12.365	12.266	12.466	694.060	720.000	-3.6
DICAMBA	7.977	7.877	8.077	688.840	705.000	-2.3
DICHLORPROP	8.697	8.598	8.798	670.150	705.000	-4.9
Dinoseb	11.320	11.221	11.421	676.820	705.000	-4.0
MCPA	8.326	8.227	8.427	70.810	69.750	1.5
MCPP	8.077	7.977	8.177	69.230	70.500	-1.8
Pentachlorophenol	9.560	9.461	9.661	702.450	712.500	-1.4
PICLORAM	12.427	12.330	12.530	695.810	712.500	-2.3