

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030575.D Time Analyzed: 17:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.685	9.597	9.797	689.350	712.500	-3.2
2,4,5-TP(Silvex)	9.390	9.304	9.504	721.350	712.500	1.2
2,4-D	8.500	8.412	8.612	682.250	705.000	-3.2
2,4-DB	10.264	10.179	10.379	633.500	712.500	-11.1
2,4-DCAA	7.360	7.269	7.469	740.450	750.000	-1.3
3,5-DICHLOROBENZOIC ACID	6.518	6.424	6.624	704.080	697.500	0.9
4-Nitrophenol	7.160	7.067	7.267	673.930	682.500	-1.3
Dalapon	2.714	2.616	2.816	692.400	682.500	1.5
DCPA	11.783	11.697	11.897	730.200	720.000	1.4
DICAMBA	7.552	7.461	7.661	726.990	705.000	3.1
DICHLORPROP	8.267	8.179	8.379	700.270	705.000	-0.7
Dinoseb	11.486	11.402	11.602	705.860	705.000	0.1
MCPA	7.884	7.795	7.995	69.410	69.750	-0.5
MCPP	7.734	7.644	7.844	78.620	70.500	11.5
Pentachlorophenol	8.809	8.720	8.920	773.960	712.500	8.6
PICLORAM	11.292	11.207	11.407	627.940	712.500	-11.9

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030575.D Time Analyzed: 17:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.361	10.268	10.468	708.570	712.500	-0.6
2,4,5-TP(Silvex)	9.935	9.842	10.042	715.680	712.500	0.4
2,4-D	9.028	8.934	9.134	699.240	705.000	-0.8
2,4-DB	10.928	10.836	11.036	687.130	712.500	-3.6
2,4-DCAA	7.769	7.674	7.874	720.580	750.000	-3.9
3,5-DICHLOROBENZOIC ACID	6.714	6.617	6.817	681.770	697.500	-2.3
4-Nitrophenol	7.300	7.203	7.403	696.630	682.500	2.1
Dalapon	2.713	2.610	2.810	662.850	682.500	-2.9
DCPA	12.358	12.266	12.466	738.200	720.000	2.5
DICAMBA	7.973	7.877	8.077	704.220	705.000	-0.1
DICHLORPROP	8.693	8.598	8.798	698.740	705.000	-0.9
Dinoseb	11.315	11.221	11.421	708.580	705.000	0.5
MCPA	8.322	8.227	8.427	65.230	69.750	-6.5
MCPP	8.072	7.977	8.177	71.170	70.500	1.0
Pentachlorophenol	9.555	9.461	9.661	734.100	712.500	3.0
PICLORAM	12.422	12.330	12.530	676.580	712.500	-5.0