

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030580.D Time Analyzed: 19:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.684	9.597	9.797	698.110	712.500	-2.0
2,4,5-TP(Silvex)	9.390	9.304	9.504	724.120	712.500	1.6
2,4-D	8.499	8.412	8.612	689.840	705.000	-2.2
2,4-DB	10.264	10.179	10.379	649.610	712.500	-8.8
2,4-DCAA	7.360	7.269	7.469	732.930	750.000	-2.3
3,5-DICHLOROBENZOIC ACID	6.518	6.424	6.624	707.480	697.500	1.4
4-Nitrophenol	7.160	7.067	7.267	671.670	682.500	-1.6
Dalapon	2.715	2.616	2.816	704.100	682.500	3.2
DCPA	11.781	11.697	11.897	734.030	720.000	1.9
DICAMBA	7.552	7.461	7.661	729.450	705.000	3.5
DICHLORPROP	8.268	8.179	8.379	700.780	705.000	-0.6
Dinoseb	11.486	11.402	11.602	716.830	705.000	1.7
MCPA	7.885	7.795	7.995	69.180	69.750	-0.8
MCPP	7.733	7.644	7.844	74.550	70.500	5.7
Pentachlorophenol	8.809	8.720	8.920	776.630	712.500	9.0
PICLORAM	11.291	11.207	11.407	638.160	712.500	-10.4

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030580.D Time Analyzed: 19:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.361	10.268	10.468	710.000	712.500	-0.4
2,4,5-TP(Silvex)	9.935	9.842	10.042	715.970	712.500	0.5
2,4-D	9.028	8.934	9.134	701.870	705.000	-0.4
2,4-DB	10.928	10.836	11.036	686.990	712.500	-3.6
2,4-DCAA	7.769	7.674	7.874	723.920	750.000	-3.5
3,5-DICHLOROBENZOIC ACID	6.715	6.617	6.817	690.800	697.500	-1.0
4-Nitrophenol	7.300	7.203	7.403	709.880	682.500	4.0
Dalapon	2.712	2.610	2.810	674.870	682.500	-1.1
DCPA	12.358	12.266	12.466	732.010	720.000	1.7
DICAMBA	7.973	7.877	8.077	709.970	705.000	0.7
DICHLORPROP	8.693	8.598	8.798	701.790	705.000	-0.5
Dinoseb	11.314	11.221	11.421	712.940	705.000	1.1
MCPA	8.322	8.227	8.427	65.610	69.750	-5.9
MCPP	8.072	7.977	8.177	63.860	70.500	-9.4
Pentachlorophenol	9.555	9.461	9.661	733.920	712.500	3.0
PICLORAM	12.422	12.330	12.530	689.630	712.500	-3.2