

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	Q1872	Analytical Method:	8151A
Client:	Alliance Technical Group, LLC - Newark	DataFile :	PS030578.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2130-01MSD (Column 1)	TP-3MSD											
	DICAMBA	184.1	0	82.5	ug/Kg	45		0		10	112	20
	MCP	18.41	0	6.90	ug/Kg	37	*	6		70	130	20
	Dalapon	184.1	0	26.8	ug/Kg	15	*	0		70	130	20
	MCPA	18.41	0	7.10	ug/Kg	39	*	0		70	130	20
	DICHLORPROP	184.1	0	92.4	ug/Kg	50		0		10	113	20
	2,4-D	184.1	0	99.3	ug/Kg	54		0		10	144	20
	2,4,5-TP(Silvex)	184.1	0	87.6	ug/Kg	48		0		10	114	20
	2,4,5-T	184.1	0	80.3	ug/Kg	44		2		10	115	20
	2,4-DB	184.1	0	44.9	ug/Kg	24		4		10	140	20
	Dinoseb	184.1	0	16.3	ug/Kg	9	*	0		10	118	20
	Pentachlorophenol	184.1	0	91.8	ug/Kg	50	*	0		70	130	20
	4-Nitrophenol	184.1	0	76.0	ug/Kg	41	*	0		70	130	20
	PICLORAM	184.1	0	60.9	ug/Kg	33	*	0		70	130	20
	DCPA	184.1	0	68.2	ug/Kg	37	*	0		70	130	20
	3,5-DICHLOROBENZO	184.1	0	86.9	ug/Kg	47	*	2		70	130	20
Client Sample ID: Q2130-01MSD (Column 2)	TP-3MSD											
	DICAMBA	184.1	0	78.1	ug/Kg	42		0		10	112	20
	MCP	18.41	0	5.80	ug/Kg	32	*	0		70	130	20
	Dalapon	184.1	0	224	ug/Kg	122		2		70	130	20
	MCPA	18.41	0	7.40	ug/Kg	40	*	2		70	130	20
	DICHLORPROP	184.1	0	91.1	ug/Kg	49		2		10	113	20
	2,4-D	184.1	0	94.6	ug/Kg	51		10		10	144	20
	2,4,5-TP(Silvex)	184.1	0	92.0	ug/Kg	50		0		10	114	20
	2,4,5-T	184.1	0	81.6	ug/Kg	44		2		10	115	20
	2,4-DB	184.1	0	74.6	ug/Kg	41		0		10	140	20
	Dinoseb	184.1	0	20.1	ug/Kg	11		0		10	118	20
	Pentachlorophenol	184.1	0	82.0	ug/Kg	45	*	0		70	130	20
	4-Nitrophenol	184.1	0	78.5	ug/Kg	43	*	0		70	130	20
	PICLORAM	184.1	0	59.7	ug/Kg	32	*	3		70	130	20
	DCPA	184.1	0	81.9	ug/Kg	44	*	2		70	130	20
	3,5-DICHLOROBENZO	184.1	0	85.6	ug/Kg	46	*	2		70	130	20