

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1872

Analytical Method: 8151A

Client: Alliance Technical Group, LLC - Newark

Datafile : PS030485.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168207BS (Column 1)	DICAMBA	166.6	161	ug/Kg	97				72	129	
	MCP	16.66	15.1	ug/Kg	91				70	130	
	Dalapon	166.6	154	ug/Kg	92				70	130	
	MCPA	16.66	15.1	ug/Kg	91				70	130	
	DICHLORPROP	166.6	160	ug/Kg	96				77	135	
	2,4-D	166.6	160	ug/Kg	96				65	144	
	2,4,5-TP(Silvex)	166.6	165	ug/Kg	99				74	146	
	2,4,5-T	166.6	163	ug/Kg	98				77	134	
	2,4-DB	166.6	158	ug/Kg	95				72	122	
	Dinoseb	166.6	162	ug/Kg	97				74	132	
	Pentachlorophenol	166.6	177	ug/Kg	106				70	130	
	4-Nitrophenol	166.6	154	ug/Kg	92				70	130	
	PICLORAM	166.6	150	ug/Kg	90				70	130	
	DCPA	166.6	168	ug/Kg	101				70	130	
	3,5-DICHLOROBENZOIC	166.6	160	ug/Kg	96				70	130	
PB168207BS (Column 2)	DICAMBA	166.6	155	ug/Kg	93				72	129	
	MCP	16.66	15.4	ug/Kg	92				70	130	
	Dalapon	166.6	150	ug/Kg	90				70	130	
	MCPA	16.66	15.7	ug/Kg	94				70	130	
	DICHLORPROP	166.6	156	ug/Kg	94				77	135	
	2,4-D	166.6	157	ug/Kg	94				65	144	
	2,4,5-TP(Silvex)	166.6	159	ug/Kg	95				74	146	
	2,4,5-T	166.6	159	ug/Kg	95				77	134	
	2,4-DB	166.6	158	ug/Kg	95				72	122	
	Dinoseb	166.6	156	ug/Kg	94				74	132	
	Pentachlorophenol	166.6	162	ug/Kg	97				70	130	
	4-Nitrophenol	166.6	149	ug/Kg	89				70	130	
	PICLORAM	166.6	150	ug/Kg	90				70	130	
	DCPA	166.6	161	ug/Kg	97				70	130	
	3,5-DICHLOROBENZOIC	166.6	154	ug/Kg	92				70	130	