

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/24/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 02:37 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1221-1 (1)	4.71	4.70	4.60	4.80	0.00
Aroclor-1221-2 (2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1221-3 (3)	4.87	4.86	4.76	4.96	-0.01
Tetrachloro-m-xylene	4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl	10.21	10.21	10.11	10.31	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/24/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 02:37 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1221-1 (1)	4.01	4.01	3.91	4.11	0.00
Aroclor-1221-2 (2)	4.09	4.09	3.99	4.19	0.00
Aroclor-1221-3 (3)	4.17	4.17	4.07	4.27	0.00
Tetrachloro-m-xylene	3.80	3.79	3.69	3.89	-0.01
Decachlorobiphenyl	8.82	8.82	8.72	8.92	0.00