

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/23/2025

Lab Sample No.: AR1660CCC500 Data File : PP072307.D Time Analyzed: 12:10

| COMPOUND             | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|----------------------|--------|--------------------|-------------------|------|
| Aroclor-1016-1       | 5.658  | 5.552                | 5.752  | 480.810            | 500.000           | -3.8 |
| Aroclor-1016-2       | 5.679  | 5.574                | 5.774  | 507.080            | 500.000           | 1.4  |
| Aroclor-1016-3       | 5.741  | 5.636                | 5.836  | 506.840            | 500.000           | 1.4  |
| Aroclor-1016-4       | 5.839  | 5.734                | 5.934  | 521.080            | 500.000           | 4.2  |
| Aroclor-1016-5       | 6.132  | 6.027                | 6.227  | 525.940            | 500.000           | 5.2  |
| Aroclor-1260-1       | 7.250  | 7.147                | 7.347  | 504.090            | 500.000           | 0.8  |
| Aroclor-1260-2       | 7.503  | 7.401                | 7.601  | 503.000            | 500.000           | 0.6  |
| Aroclor-1260-3       | 7.862  | 7.760                | 7.960  | 528.630            | 500.000           | 5.7  |
| Aroclor-1260-4       | 8.086  | 7.985                | 8.185  | 524.460            | 500.000           | 4.9  |
| Aroclor-1260-5       | 8.405  | 8.303                | 8.503  | 528.920            | 500.000           | 5.8  |
| Decachlorobiphenyl   | 10.214 | 10.113               | 10.313 | 53.660             | 50.000            | 7.3  |
| Tetrachloro-m-xylene | 4.505  | 4.396                | 4.596  | 52.260             | 50.000            | 4.5  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/23/2025

Lab Sample No.: AR1660CCC500 Data File : PP072307.D Time Analyzed: 12:10

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.881 | 4.777                | 4.977 | 519.560            | 500.000           | 3.9  |
| Aroclor-1016-2       | 4.899 | 4.795                | 4.995 | 518.720            | 500.000           | 3.7  |
| Aroclor-1016-3       | 5.076 | 4.973                | 5.173 | 534.070            | 500.000           | 6.8  |
| Aroclor-1016-4       | 5.118 | 5.015                | 5.215 | 541.090            | 500.000           | 8.2  |
| Aroclor-1016-5       | 5.332 | 5.229                | 5.429 | 534.440            | 500.000           | 6.9  |
| Aroclor-1260-1       | 6.365 | 6.264                | 6.464 | 557.400            | 500.000           | 11.5 |
| Aroclor-1260-2       | 6.554 | 6.453                | 6.653 | 527.390            | 500.000           | 5.5  |
| Aroclor-1260-3       | 6.706 | 6.606                | 6.806 | 563.330            | 500.000           | 12.7 |
| Aroclor-1260-4       | 7.177 | 7.077                | 7.277 | 551.400            | 500.000           | 10.3 |
| Aroclor-1260-5       | 7.419 | 7.319                | 7.519 | 540.150            | 500.000           | 8.0  |
| Decachlorobiphenyl   | 8.821 | 8.723                | 8.923 | 57.320             | 50.000            | 14.6 |
| Tetrachloro-m-xylene | 3.798 | 3.691                | 3.891 | 52.740             | 50.000            | 5.5  |