

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP072347.D Time Analyzed: 02:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.653	5.552	5.752	461.720	500.000	-7.7
Aroclor-1016-2	5.674	5.574	5.774	487.490	500.000	-2.5
Aroclor-1016-3	5.736	5.636	5.836	487.020	500.000	-2.6
Aroclor-1016-4	5.834	5.734	5.934	503.070	500.000	0.6
Aroclor-1016-5	6.126	6.027	6.227	496.600	500.000	-0.7
Aroclor-1260-1	7.245	7.147	7.347	482.650	500.000	-3.5
Aroclor-1260-2	7.498	7.401	7.601	476.560	500.000	-4.7
Aroclor-1260-3	7.856	7.760	7.960	494.700	500.000	-1.1
Aroclor-1260-4	8.081	7.985	8.185	482.640	500.000	-3.5
Aroclor-1260-5	8.400	8.303	8.503	489.690	500.000	-2.1
Decachlorobiphenyl	10.205	10.113	10.313	49.450	50.000	-1.1
Tetrachloro-m-xylene	4.500	4.396	4.596	50.360	50.000	0.7

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP072347.D Time Analyzed: 02:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	509.160	500.000	1.8
Aroclor-1016-2	4.896	4.795	4.995	521.050	500.000	4.2
Aroclor-1016-3	5.073	4.973	5.173	535.580	500.000	7.1
Aroclor-1016-4	5.115	5.015	5.215	540.020	500.000	8.0
Aroclor-1016-5	5.329	5.229	5.429	532.910	500.000	6.6
Aroclor-1260-1	6.361	6.264	6.464	537.210	500.000	7.4
Aroclor-1260-2	6.550	6.453	6.653	504.890	500.000	1.0
Aroclor-1260-3	6.703	6.606	6.806	534.550	500.000	6.9
Aroclor-1260-4	7.173	7.077	7.277	531.200	500.000	6.2
Aroclor-1260-5	7.415	7.319	7.519	522.390	500.000	4.5
Decachlorobiphenyl	8.816	8.723	8.923	51.900	50.000	3.8
Tetrachloro-m-xylene	3.796	3.691	3.891	51.320	50.000	2.6