

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/24/2025

Lab Sample No.: AR1221CCC500 Data File : PP072348.D Time Analyzed: 02:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.705	4.603	4.803	500.970	500.000	0.2
Aroclor-1221-2	4.791	4.688	4.888	484.700	500.000	-3.1
Aroclor-1221-3	4.866	4.764	4.964	489.220	500.000	-2.2
Decachlorobiphenyl	10.208	10.114	10.314	50.000	50.000	0.0
Tetrachloro-m-xylene	4.503	4.401	4.601	50.740	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 05/24/2025

Lab Sample No.: AR1221CCC500 Data File : PP072348.D Time Analyzed: 02:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1221-1	4.006	3.905	4.105	492.170	500.000	-1.6
Aroclor-1221-2	4.093	3.991	4.191	485.110	500.000	-3.0
Aroclor-1221-3	4.169	4.067	4.267	485.110	500.000	-3.0
Decachlorobiphenyl	8.818	8.724	8.924	52.540	50.000	5.1
Tetrachloro-m-xylene	3.796	3.694	3.894	52.590	50.000	5.2