

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 05/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP072356.D Time Analyzed: 05:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.653	5.552	5.752	456.330	500.000	-8.7
Aroclor-1016-2	5.675	5.574	5.774	491.300	500.000	-1.7
Aroclor-1016-3	5.737	5.636	5.836	483.500	500.000	-3.3
Aroclor-1016-4	5.834	5.734	5.934	498.060	500.000	-0.4
Aroclor-1016-5	6.126	6.027	6.227	496.810	500.000	-0.6
Aroclor-1260-1	7.244	7.147	7.347	481.870	500.000	-3.6
Aroclor-1260-2	7.498	7.401	7.601	473.250	500.000	-5.4
Aroclor-1260-3	7.856	7.760	7.960	496.030	500.000	-0.8
Aroclor-1260-4	8.080	7.985	8.185	503.040	500.000	0.6
Aroclor-1260-5	8.399	8.303	8.503	500.230	500.000	0.0
Decachlorobiphenyl	10.204	10.113	10.313	50.240	50.000	0.5
Tetrachloro-m-xylene	4.500	4.396	4.596	50.270	50.000	0.5

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL05 Date Analyzed: 05/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP072356.D Time Analyzed: 05:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	521.520	500.000	4.3
Aroclor-1016-2	4.896	4.795	4.995	526.470	500.000	5.3
Aroclor-1016-3	5.073	4.973	5.173	537.350	500.000	7.5
Aroclor-1016-4	5.115	5.015	5.215	537.740	500.000	7.5
Aroclor-1016-5	5.329	5.229	5.429	550.370	500.000	10.1
Aroclor-1260-1	6.362	6.264	6.464	535.920	500.000	7.2
Aroclor-1260-2	6.551	6.453	6.653	510.220	500.000	2.0
Aroclor-1260-3	6.703	6.606	6.806	545.080	500.000	9.0
Aroclor-1260-4	7.173	7.077	7.277	550.870	500.000	10.2
Aroclor-1260-5	7.415	7.319	7.519	553.440	500.000	10.7
Decachlorobiphenyl	8.817	8.723	8.923	54.380	50.000	8.8
Tetrachloro-m-xylene	3.796	3.691	3.891	52.440	50.000	4.9