

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL06 Date Analyzed: 05/24/2025

Lab Sample No.: AR1221CCC500 Data File : PP072357.D Time Analyzed: 06:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.705	4.603	4.803	496.260	500.000	-0.7
Aroclor-1221-2	4.790	4.688	4.888	472.720	500.000	-5.5
Aroclor-1221-3	4.866	4.764	4.964	493.930	500.000	-1.2
Decachlorobiphenyl	10.207	10.114	10.314	49.940	50.000	-0.1
Tetrachloro-m-xylene	4.503	4.401	4.601	50.870	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL06 Date Analyzed: 05/24/2025

Lab Sample No.: AR1221CCC500 Data File : PP072357.D Time Analyzed: 06:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1221-1	4.006	3.905	4.105	480.240	500.000	-4.0
Aroclor-1221-2	4.093	3.991	4.191	475.070	500.000	-5.0
Aroclor-1221-3	4.168	4.067	4.267	478.510	500.000	-4.3
Decachlorobiphenyl	8.817	8.724	8.924	53.360	50.000	6.7
Tetrachloro-m-xylene	3.796	3.694	3.894	52.520	50.000	5.0