

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL07 Date Analyzed: 05/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP072404.D Time Analyzed: 14:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.650	5.552	5.752	436.870	500.000	-12.6
Aroclor-1016-2	5.671	5.574	5.774	466.270	500.000	-6.7
Aroclor-1016-3	5.734	5.636	5.836	463.610	500.000	-7.3
Aroclor-1016-4	5.831	5.734	5.934	481.590	500.000	-3.7
Aroclor-1016-5	6.123	6.027	6.227	483.950	500.000	-3.2
Aroclor-1260-1	7.242	7.147	7.347	495.010	500.000	-1.0
Aroclor-1260-2	7.494	7.401	7.601	478.610	500.000	-4.3
Aroclor-1260-3	7.853	7.760	7.960	495.470	500.000	-0.9
Aroclor-1260-4	8.076	7.985	8.185	498.290	500.000	-0.3
Aroclor-1260-5	8.395	8.303	8.503	515.500	500.000	3.1
Decachlorobiphenyl	10.199	10.113	10.313	49.960	50.000	-0.1
Tetrachloro-m-xylene	4.498	4.396	4.596	48.860	50.000	-2.3

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL07 Date Analyzed: 05/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP072404.D Time Analyzed: 14:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.876	4.777	4.977	511.910	500.000	2.4
Aroclor-1016-2	4.894	4.795	4.995	517.490	500.000	3.5
Aroclor-1016-3	5.072	4.973	5.173	523.990	500.000	4.8
Aroclor-1016-4	5.113	5.015	5.215	521.460	500.000	4.3
Aroclor-1016-5	5.327	5.229	5.429	559.530	500.000	11.9
Aroclor-1260-1	6.359	6.264	6.464	537.480	500.000	7.5
Aroclor-1260-2	6.548	6.453	6.653	507.400	500.000	1.5
Aroclor-1260-3	6.700	6.606	6.806	539.000	500.000	7.8
Aroclor-1260-4	7.170	7.077	7.277	542.090	500.000	8.4
Aroclor-1260-5	7.412	7.319	7.519	539.540	500.000	7.9
Decachlorobiphenyl	8.812	8.723	8.923	56.600	50.000	13.2
Tetrachloro-m-xylene	3.795	3.691	3.891	50.310	50.000	0.6