

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL08 Date Analyzed: 05/28/2025

Lab Sample No.: AR1221CCC500 Data File : PP072405.D Time Analyzed: 14:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.702	4.603	4.803	467.620	500.000	-6.5
Aroclor-1221-2	4.787	4.688	4.888	451.060	500.000	-9.8
Aroclor-1221-3	4.863	4.764	4.964	462.830	500.000	-7.4
Decachlorobiphenyl	10.198	10.114	10.314	50.000	50.000	0.0
Tetrachloro-m-xylene	4.500	4.401	4.601	48.170	50.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL08 Date Analyzed: 05/28/2025

Lab Sample No.: AR1221CCC500 Data File : PP072405.D Time Analyzed: 14:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.003	3.905	4.105	451.030	500.000	-9.8
Aroclor-1221-2	4.090	3.991	4.191	444.280	500.000	-11.1
Aroclor-1221-3	4.166	4.067	4.267	454.410	500.000	-9.1
Decachlorobiphenyl	8.810	8.724	8.924	54.590	50.000	9.2
Tetrachloro-m-xylene	3.794	3.694	3.894	50.390	50.000	0.8