

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL09 Date Analyzed: 05/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP072419.D Time Analyzed: 20:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.650	5.552	5.752	415.340	500.000	-16.9
Aroclor-1016-2	5.672	5.574	5.774	445.940	500.000	-10.8
Aroclor-1016-3	5.735	5.636	5.836	441.760	500.000	-11.6
Aroclor-1016-4	5.832	5.734	5.934	455.900	500.000	-8.8
Aroclor-1016-5	6.124	6.027	6.227	449.770	500.000	-10.0
Aroclor-1260-1	7.242	7.147	7.347	459.860	500.000	-8.0
Aroclor-1260-2	7.495	7.401	7.601	457.080	500.000	-8.6
Aroclor-1260-3	7.854	7.760	7.960	471.580	500.000	-5.7
Aroclor-1260-4	8.078	7.985	8.185	470.120	500.000	-6.0
Aroclor-1260-5	8.396	8.303	8.503	482.010	500.000	-3.6
Decachlorobiphenyl	10.199	10.113	10.313	48.070	50.000	-3.9
Tetrachloro-m-xylene	4.499	4.396	4.596	46.740	50.000	-6.5

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL09 Date Analyzed: 05/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP072419.D Time Analyzed: 20:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.877	4.777	4.977	498.470	500.000	-0.3
Aroclor-1016-2	4.895	4.795	4.995	500.220	500.000	0.0
Aroclor-1016-3	5.072	4.973	5.173	506.910	500.000	1.4
Aroclor-1016-4	5.114	5.015	5.215	500.310	500.000	0.1
Aroclor-1016-5	5.328	5.229	5.429	554.630	500.000	10.9
Aroclor-1260-1	6.360	6.264	6.464	525.560	500.000	5.1
Aroclor-1260-2	6.548	6.453	6.653	490.830	500.000	-1.8
Aroclor-1260-3	6.700	6.606	6.806	512.150	500.000	2.4
Aroclor-1260-4	7.171	7.077	7.277	510.340	500.000	2.1
Aroclor-1260-5	7.413	7.319	7.519	505.310	500.000	1.1
Decachlorobiphenyl	8.812	8.723	8.923	52.010	50.000	4.0
Tetrachloro-m-xylene	3.795	3.691	3.891	49.040	50.000	-1.9