

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL10 Date Analyzed: 05/28/2025

Lab Sample No.: AR1221CCC500 Data File : PP072420.D Time Analyzed: 20:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.701	4.603	4.803	466.550	500.000	-6.7
Aroclor-1221-2	4.787	4.688	4.888	586.510	500.000	17.3
Aroclor-1221-3	4.863	4.764	4.964	529.090	500.000	5.8
Decachlorobiphenyl	10.198	10.114	10.314	48.820	50.000	-2.4
Tetrachloro-m-xylene	4.500	4.401	4.601	47.420	50.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL10 Date Analyzed: 05/28/2025

Lab Sample No.: AR1221CCC500 Data File : PP072420.D Time Analyzed: 20:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1221-1	4.005	3.905	4.105	476.980	500.000	-4.6
Aroclor-1221-2	4.092	3.991	4.191	468.970	500.000	-6.2
Aroclor-1221-3	4.167	4.067	4.267	466.860	500.000	-6.6
Decachlorobiphenyl	8.812	8.724	8.924	50.960	50.000	1.9
Tetrachloro-m-xylene	3.795	3.694	3.894	48.890	50.000	-2.2