

Surrogate Summary

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072163.D	PIBLK-PP072163.D	Tetrachloro-m-xylene	1	20	16.3	82		60	140
		Decachlorobiphenyl	1	20	16.7	84		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	17.0	85		60	140
I.BLK-PP072312.D	PIBLK-PP072312.D	Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	20.3	101		60	140
		Tetrachloro-m-xylene	2	20	19.8	99		60	140
		Decachlorobiphenyl	2	20	21.7	109		60	140
PB168149BL	PB168149BL	Tetrachloro-m-xylene	1	20	21.1	106		32	144
		Decachlorobiphenyl	1	20	22.3	111		32	175
		Tetrachloro-m-xylene	2	20	22.4	112		32	144
		Decachlorobiphenyl	2	20	22.6	113		32	175
PB168149BS	PB168149BS	Tetrachloro-m-xylene	1	20	21.5	108		32	144
		Decachlorobiphenyl	1	20	22.7	113		32	175
		Tetrachloro-m-xylene	2	20	21.2	106		32	144
		Decachlorobiphenyl	2	20	24.4	122		32	175
Q2112-01MS	EO-03-05222025MS	Tetrachloro-m-xylene	1	20	13.9	69		32	144
		Decachlorobiphenyl	1	20	14.3	71		32	175
		Tetrachloro-m-xylene	2	20	17.5	87		32	144
		Decachlorobiphenyl	2	20	17.2	86		32	175
Q2112-01MSD	EO-03-05222025MSD	Tetrachloro-m-xylene	1	20	14.7	74		32	144
		Decachlorobiphenyl	1	20	14.3	72		32	175
		Tetrachloro-m-xylene	2	20	17.6	88		32	144
		Decachlorobiphenyl	2	20	17.0	85		32	175
I.BLK-PP072326.D	PIBLK-PP072326.D	Tetrachloro-m-xylene	1	20	17.8	89		60	140
		Decachlorobiphenyl	1	20	17.0	85		60	140
		Tetrachloro-m-xylene	2	20	19.5	98		60	140
		Decachlorobiphenyl	2	20	19.2	96		60	140
I.BLK-PP072340.D	PIBLK-PP072340.D	Tetrachloro-m-xylene	1	20	17.7	88		60	140
		Decachlorobiphenyl	1	20	17.7	88		60	140
		Tetrachloro-m-xylene	2	20	18.9	94		60	140
		Decachlorobiphenyl	2	20	20.6	103		60	140
I.BLK-PP072349.D	PIBLK-PP072349.D	Tetrachloro-m-xylene	1	20	15.3	76		60	140
		Decachlorobiphenyl	1	20	15.4	77		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	18.6	93		60	140
Q1872-18	HW0425-PT-PCB-SOIL	Tetrachloro-m-xylene	1	20	14.0	70		32	144
		Decachlorobiphenyl	1	20	18.7	93		32	175
		Tetrachloro-m-xylene	2	20	18.5	92		32	144
		Decachlorobiphenyl	2	20	21.7	108		32	175
Q1872-18DL	HW0425-PT-PCB-SOILDL	Tetrachloro-m-xylene	1	20	0	0	*	32	144

Surrogate Summary

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1872-18DL	HW0425-PT-PCB-SOILDL	Decachlorobiphenyl	1	20	0	0	*	32	175
		Tetrachloro-m-xylene	2	20	0	0	*	32	144
		Decachlorobiphenyl	2	20	0	0	*	32	175
I.BLK-PP072358.D	PIBLK-PP072358.D	Tetrachloro-m-xylene	1	20	15.3	77		60	140
		Decachlorobiphenyl	1	20	15.8	79		60	140
		Tetrachloro-m-xylene	2	20	16.5	82		60	140
		Decachlorobiphenyl	2	20	18.3	92		60	140
I.BLK-PP072410.D	PIBLK-PP072410.D	Tetrachloro-m-xylene	1	20	16.7	84		60	140
		Decachlorobiphenyl	1	20	18.1	90		60	140
		Tetrachloro-m-xylene	2	20	18.3	91		60	140
		Decachlorobiphenyl	2	20	20.3	102		60	140
PB168154BL	PB168154BL	Tetrachloro-m-xylene	1	20	18.6	93		32	144
		Decachlorobiphenyl	1	20	19.9	99		32	175
		Tetrachloro-m-xylene	2	20	20.9	105		32	144
		Decachlorobiphenyl	2	20	21.8	109		32	175
PB168154BS	PB168154BS	Tetrachloro-m-xylene	1	20	18.8	94		32	144
		Decachlorobiphenyl	1	20	19.6	98		32	175
		Tetrachloro-m-xylene	2	20	19.2	96		32	144
		Decachlorobiphenyl	2	20	23.0	115		32	175
Q1872-19	HW0425-PT-PCBO-SOIL	Tetrachloro-m-xylene	1	20	11.2	56		32	144
		Decachlorobiphenyl	1	20	14.0	70		32	175
		Tetrachloro-m-xylene	2	20	13.0	65		32	144
		Decachlorobiphenyl	2	20	19.1	96		32	175
I.BLK-PP072423.D	PIBLK-PP072423.D	Tetrachloro-m-xylene	1	20	15.8	79		60	140
		Decachlorobiphenyl	1	20	17.0	85		60	140
		Tetrachloro-m-xylene	2	20	16.0	80		60	140
		Decachlorobiphenyl	2	20	19.0	95		60	140