

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q1872</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Alliance Technical Group, LLC - Newark</u>	DataFile :	<u>PP072320.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2112-01MSD (Column 1)	EO-03-05222025MSD AR1016	174.3	0	98.3	ug/kg	56		2		55	146	20
	AR1260	174.3	0	92.4	ug/kg	53		2		31	146	20
Client Sample ID: Q2112-01MSD (Column 2)	EO-03-05222025MSD AR1016	174.3	0	120	ug/kg	69		1		55	146	20
	AR1260	174.3	0	108	ug/kg	62		2		31	146	20