

Surrogate Summary

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL095732.D	PIBLK-PL095732.D	Decachlorobiphenyl	1	20	18.2	91		43	140
		Tetrachloro-m-xylene	1	20	16.4	82		77	126
		Decachlorobiphenyl	2	20	17.5	88		43	140
		Tetrachloro-m-xylene	2	20	16.0	80		77	126
I.BLK-PL095779.D	PIBLK-PL095779.D	Decachlorobiphenyl	1	20	20.4	102		43	140
		Tetrachloro-m-xylene	1	20	20.5	103		77	126
		Decachlorobiphenyl	2	20	19.2	96		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126
PB168151BL	PB168151BL	Decachlorobiphenyl	1	20	20.3	101		20	144
		Tetrachloro-m-xylene	1	20	19.5	98		19	148
		Decachlorobiphenyl	2	20	20.7	103		20	144
		Tetrachloro-m-xylene	2	20	18.4	92		19	148
PB168151BS	PB168151BS	Decachlorobiphenyl	1	20	21.9	109		20	144
		Tetrachloro-m-xylene	1	20	20.7	104		19	148
		Decachlorobiphenyl	2	20	22.5	113		20	144
		Tetrachloro-m-xylene	2	20	19.5	98		19	148
Q1872-20	HW0425-PT-PEST-SOIL	Decachlorobiphenyl	1	20	20.4	102		20	144
		Tetrachloro-m-xylene	1	20	17.8	89		19	148
		Decachlorobiphenyl	2	20	21.6	108		20	144
		Tetrachloro-m-xylene	2	20	16.9	85		19	148
Q1872-20DL	HW0425-PT-PEST-SOILDL	Decachlorobiphenyl	1	20	23.3	117		20	144
		Tetrachloro-m-xylene	1	20	18.4	92		19	148
		Decachlorobiphenyl	2	20	26.2	131		20	144
		Tetrachloro-m-xylene	2	20	17.1	86		19	148
Q2109-01MS	TP-02-MHFMS	Decachlorobiphenyl	1	20	15.4	77		20	144
		Tetrachloro-m-xylene	1	20	11.6	58		19	148
		Decachlorobiphenyl	2	20	10.3	51		20	144
		Tetrachloro-m-xylene	2	20	12.4	62		19	148
Q2109-01MSD	TP-02-MHFMSD	Decachlorobiphenyl	1	20	14.6	73		20	144
		Tetrachloro-m-xylene	1	20	11.9	60		19	148
		Decachlorobiphenyl	2	20	10.4	52		20	144
		Tetrachloro-m-xylene	2	20	11.8	59		19	148
I.BLK-PL095791.D	PIBLK-PL095791.D	Decachlorobiphenyl	1	20	18.5	92		43	140
		Tetrachloro-m-xylene	1	20	20.3	101		77	126
		Decachlorobiphenyl	2	20	15.2	76		43	140
		Tetrachloro-m-xylene	2	20	19.6	98		77	126