

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/27/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 17:00 Initial Calibration Time(s): 12:42 13:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.74	4.74	4.64	4.84	0.00
Chlordane-2	(2)	5.27	5.27	5.17	5.37	0.00
Chlordane-3	(3)	5.97	5.97	5.87	6.07	0.00
Chlordane-4	(4)	6.06	6.06	5.96	6.16	0.00
Chlordane-5	(5)	6.90	6.90	6.80	7.00	0.00
Decachlorobiphenyl		9.10	9.10	9.00	9.20	0.00
Tetrachloro-m-xylene		3.57	3.57	3.47	3.67	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/27/2025 Initial Calibration Date(s): 05/21/2025 05/21/2025

Continuing Calib Time: 17:00 Initial Calibration Time(s): 12:42 13:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	3.90	3.91	3.81	4.01	0.01
Chlordane-2	(2)	4.48	4.49	4.39	4.59	0.01
Chlordane-3	(3)	5.12	5.12	5.02	5.22	0.00
Chlordane-4	(4)	5.18	5.19	5.09	5.29	0.01
Chlordane-5	(5)	6.08	6.08	5.98	6.18	0.00
Decachlorobiphenyl		8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene		2.88	2.89	2.79	2.99	0.01