

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095793.D Time Analyzed: 14:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	4.742	4.639	4.839	496.170	500.000	-0.8
Chlordane-2	5.270	5.167	5.367	459.390	500.000	-8.1
Chlordane-3	5.977	5.873	6.073	477.620	500.000	-4.5
Chlordane-4	6.062	5.958	6.158	460.750	500.000	-7.9
Chlordane-5	6.900	6.799	6.999	487.780	500.000	-2.4
Decachlorobiphenyl	9.106	9.004	9.204	44.320	50.000	-11.4
Tetrachloro-m-xylene	3.576	3.473	3.673	49.910	50.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095793.D Time Analyzed: 14:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.898	3.806	4.006	465.820	500.000	-6.8
Chlordane-2	4.480	4.387	4.587	452.380	500.000	-9.5
Chlordane-3	5.114	5.023	5.223	494.210	500.000	-1.2
Chlordane-4	5.180	5.087	5.287	449.500	500.000	-10.1
Chlordane-5	6.075	5.982	6.182	436.950	500.000	-12.6
Decachlorobiphenyl	8.058	7.963	8.163	40.610	50.000	-18.8
Tetrachloro-m-xylene	2.878	2.785	2.985	59.730	50.000	19.5