

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095799.D Time Analyzed: 17:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	4.739	4.639	4.839	480.010	500.000	-4.0
Chlordane-2	5.266	5.167	5.367	474.300	500.000	-5.1
Chlordane-3	5.974	5.873	6.073	464.240	500.000	-7.2
Chlordane-4	6.058	5.958	6.158	448.160	500.000	-10.4
Chlordane-5	6.898	6.799	6.999	499.310	500.000	-0.1
Decachlorobiphenyl	9.101	9.004	9.204	43.590	50.000	-12.8
Tetrachloro-m-xylene	3.574	3.473	3.673	47.850	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/27/2025

Lab Sample No.: PCHLORCCC500 Data File : PL095799.D Time Analyzed: 17:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.902	3.806	4.006	471.780	500.000	-5.6
Chlordane-2	4.484	4.387	4.587	469.300	500.000	-6.1
Chlordane-3	5.118	5.023	5.223	471.260	500.000	-5.7
Chlordane-4	5.183	5.087	5.287	465.600	500.000	-6.9
Chlordane-5	6.078	5.982	6.182	452.350	500.000	-9.5
Decachlorobiphenyl	8.059	7.963	8.163	43.740	50.000	-12.5
Tetrachloro-m-xylene	2.882	2.785	2.985	59.960	50.000	19.9