

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/29/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 16:04 Initial Calibration Time(s): 13:47 14:41

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Toxaphene-1	(1)	6.24	6.24	6.14	6.34	0.00
Toxaphene-2	(2)	6.44	6.44	6.34	6.54	0.00
Toxaphene-3	(3)	7.15	7.15	7.05	7.25	0.00
Toxaphene-4	(4)	7.56	7.57	7.47	7.67	0.01
Toxaphene-5	(5)	7.93	7.93	7.83	8.03	0.00
Decachlorobiphenyl		9.07	9.08	8.98	9.18	0.01
Tetrachloro-m-xylene		3.55	3.55	3.45	3.65	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Continuing Calib Date: 05/29/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 16:04 Initial Calibration Time(s): 13:47 14:41

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Toxaphene-1	(1)	5.48	5.48	5.38	5.58	0.01
Toxaphene-2	(2)	5.65	5.65	5.55	5.75	0.00
Toxaphene-3	(3)	6.76	6.76	6.66	6.86	0.00
Toxaphene-4	(4)	7.20	7.20	7.10	7.30	0.00
Toxaphene-5	(5)	7.33	7.33	7.23	7.43	0.00
Decachlorobiphenyl		8.07	8.08	7.98	8.18	0.01
Tetrachloro-m-xylene		2.88	2.88	2.78	2.98	0.00