

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/29/2025

Lab Sample No.: PTOXCCC500 Data File : PD088772.D Time Analyzed: 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.080	8.978	9.178	47.130	50.000	-5.7
Tetrachloro-m-xylene	3.555	3.451	3.651	51.630	50.000	3.3
Toxaphene-1	6.248	6.144	6.344	517.210	500.000	3.4
Toxaphene-2	6.447	6.344	6.544	496.850	500.000	-0.6
Toxaphene-3	7.155	7.051	7.251	487.920	500.000	-2.4
Toxaphene-4	7.569	7.466	7.666	490.390	500.000	-1.9
Toxaphene-5	7.934	7.832	8.032	476.150	500.000	-4.8

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/29/2025

Lab Sample No.: PTOXCCC500 Data File : PD088772.D Time Analyzed: 13:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.074	7.976	8.176	45.340	50.000	-9.3
Tetrachloro-m-xylene	2.880	2.782	2.982	49.190	50.000	-1.6
Toxaphene-1	5.476	5.378	5.578	475.890	500.000	-4.8
Toxaphene-2	5.647	5.549	5.749	474.730	500.000	-5.1
Toxaphene-3	6.757	6.659	6.859	419.530	500.000	-16.1
Toxaphene-4	7.199	7.101	7.301	446.290	500.000	-10.7
Toxaphene-5	7.331	7.232	7.432	447.420	500.000	-10.5