

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/29/2025

Lab Sample No.: PTOXCCC500 Data File : PD088778.D Time Analyzed: 16:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.073	8.978	9.178	50.480	50.000	1.0
Tetrachloro-m-xylene	3.549	3.451	3.651	52.600	50.000	5.2
Toxaphene-1	6.241	6.144	6.344	522.650	500.000	4.5
Toxaphene-2	6.442	6.344	6.544	540.630	500.000	8.1
Toxaphene-3	7.148	7.051	7.251	500.090	500.000	0.0
Toxaphene-4	7.564	7.466	7.666	494.920	500.000	-1.0
Toxaphene-5	7.929	7.832	8.032	494.990	500.000	-1.0

CALIBRATION VERIFICATION SUMMARY

Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/29/2025

Lab Sample No.: PTOXCCC500 Data File : PD088778.D Time Analyzed: 16:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.073	7.976	8.176	47.030	50.000	-5.9
Tetrachloro-m-xylene	2.881	2.782	2.982	50.480	50.000	1.0
Toxaphene-1	5.475	5.378	5.578	500.960	500.000	0.2
Toxaphene-2	5.646	5.549	5.749	492.980	500.000	-1.4
Toxaphene-3	6.756	6.659	6.859	479.250	500.000	-4.2
Toxaphene-4	7.198	7.101	7.301	464.150	500.000	-7.2
Toxaphene-5	7.329	7.232	7.432	464.670	500.000	-7.1