

Surrogate Summary

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088583.D	PIBLK-PD088583.D	Decachlorobiphenyl	1	20	18.4	92		57	171
		Tetrachloro-m-xylene	1	20	16.5	83		61	148
		Decachlorobiphenyl	2	20	18.0	90		57	171
		Tetrachloro-m-xylene	2	20	16.9	84		61	148
I.BLK-PD088769.D	PIBLK-PD088769.D	Decachlorobiphenyl	1	20	21.7	108		57	171
		Tetrachloro-m-xylene	1	20	21.5	108		61	148
		Decachlorobiphenyl	2	20	21.0	105		57	171
		Tetrachloro-m-xylene	2	20	22.2	111		61	148
PB168153BL	PB168153BL	Decachlorobiphenyl	1	20	20.3	102		20	144
		Tetrachloro-m-xylene	1	20	18.9	94		19	148
		Decachlorobiphenyl	2	20	19.4	97		20	144
		Tetrachloro-m-xylene	2	20	19.6	98		19	148
PB168153BS	PB168153BS	Decachlorobiphenyl	1	20	22.2	111		20	144
		Tetrachloro-m-xylene	1	20	22.5	113		19	148
		Decachlorobiphenyl	2	20	21.8	109		20	144
		Tetrachloro-m-xylene	2	20	22.1	111		19	148
Q1872-22	HW0425-PT-TXP-SOIL	Decachlorobiphenyl	1	20	20.7	103		20	144
		Tetrachloro-m-xylene	1	20	21.3	106		19	148
		Decachlorobiphenyl	2	20	20.2	101		20	144
		Tetrachloro-m-xylene	2	20	19.1	96		19	148
I.BLK-PD088776.D	PIBLK-PD088776.D	Decachlorobiphenyl	1	20	22.0	110		57	171
		Tetrachloro-m-xylene	1	20	21.9	109		61	148
		Decachlorobiphenyl	2	20	21.4	107		57	171
		Tetrachloro-m-xylene	2	20	22.8	114		61	148