284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

DATA PACKAGE

SEMI-VOLATILE ORGANICS

PROJECT NAME : NJ SOIL PT

ALLIANCE TECHNICAL GROUP, LLC - NEWARK

284 Sheffiled Stree

Suite 1

Mountainside, NJ - 07092

Phone No: 908-789-8900

ORDER ID : Q1872

ATTENTION : Mohammad Ahmed

Laboratory Certification ID # 20012

Alliance TECHNICAL GROUP

Table Of Contents for Q1872

1) SEMI-VOLATILE DATA 2
2) Signature Page 4
3) Case Narrative 5
4) Qualifier Page 7
5) Conformance/Non Conformance 8
6) QA Checklist 10
7) Chronicle 11
8) Hit Summary 12
9) QC Data Summary For SVOCMS Group4 13
- 9.1) Deuterated Monitoring Compound Summary 14
- 9.2) MS/MSD Summary 15
- 9.3) LCS/LCSD Summary 17
- 9.4) Method Blank Summary 18
- 9.5) GS/MS Tune Summary 19
- 9.6) Internal Standard Area and RT Summary 25

10) Sample Data 31
- 10.1) HW0425-PT-TRIAZINE-SOIL 32

11) Calibration Data Summary 42
- 11.1) Initial Calibration Data 43
  - 11.1.1) BM042825 43
  - 11.1.2) BP041425 459
  - 11.1.3) BP051325 868
- 11.2) Continued Calibration Data 1276
  - 11.2.1) BM050140.D 1276
  - 11.2.2) BP024568.D 1333
  - 11.2.3) BP024625.D 1389

12) QC Sample Data 1445
- 12.1) Tune Raw Data 1446
- 12.2) Method Blank Data 1470
- 12.3) LCS Data 1479
- 12.4) MS Data 1529
- 12.5) MSD Data 1580

13) Manual Integration 1631
14) Analytical Runlogs 1639
15) Percent Solid 1651

Table Of Contents for Q1872

16) Extraction Logs 1657

16.1) PB167904.pdf 1657

16.2) PB167904IC.pdf 1659

17) Standard Prep Logs 1660

18) Shipping Document 1742

18.1) Chain Of Custody 1743

18.2) Lab Certificate 1745

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q1872

Project ID : NJ Soil PT

Client : Alliance Technical Group, LLC - Newark

Lab Sample Number	Client Sample Number
Q1872-01	HW0425-PT-AN-SOIL
Q1872-02	HW0425-PT-CORR-SOIL
Q1872-03	HW0425-PT-CN-SOIL
Q1872-04	HW0425-PT-CN-SOIL
Q1872-05	HW0425-PT-FP-SOIL
Q1872-06	HW0425-PT-CR6-SOIL
Q1872-07	HW0425-PT-NUT-SOIL
Q1872-08	HW0425-PT-NUT-SOIL
Q1872-09	HW0425-PT-OGR-SOIL
Q1872-10	HW0425-PT-MET-SOIL
Q1872-11	HW0425-PT-BNA-SOIL
Q1872-12	HW0425-PT-TRIAZINE-SOIL
Q1872-13	HW0425-PT-PAH-SOIL
Q1872-14	HW0425-PT-DIES-SOIL
Q1872-15	HW0425-PT-GAS-SOIL
Q1872-16	HW0425-PT-NJEPH-SOIL
Q1872-17	HW0425-PT-HERB-SOIL
Q1872-18	HW0425-PT-PCB-SOIL
Q1872-19	HW0425-PT-PCBO-SOIL
Q1872-20	HW0425-PT-PEST-SOIL
Q1872-21	HW0425-PT-CHLR-SOIL
Q1872-22	HW0425-PT-TXP-SOIL
Q1872-23	HW0425-PT-VOA-SOIL
Q1872-24	HW0425-PT-SOL-SOIL
Q1872-25	HW0425-PT-NO2-SOIL

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/28/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Alliance Technical Group, LLC - Newark
Project Name: NJ Soil PT
Project # N/A
Order ID # Q1872
Test Name: SVOCMS Group4

A. Number of Samples and Date of Receipt:
24 Solid samples were received on 04/24/2025.
1 Solid sample was received on 04/28/2025.

B. Parameters
According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group1, Anions Group2, Corrosivity, Cyanide, Diesel Range Organics, EPH, Flash Point, Gasoline Range Organics, Herbicide Group1, Hexavalent Chromium, Mercury, Metals Group3, Metals ICP-Group1, Oil and Grease, PCB, PESTICIDE Group1, PESTICIDE Group2, PESTICIDE Group3, Phosphorus, Total, SVOCMS Group1, SVOCMS Group2, SVOCMS Group3, SVOCMS Group4, SVOCMS Group5, TKN, TOC, TS and VOCMS Group1. This data package contains results for SVOCMS Group4.

C. Analytical Techniques:
The samples were analyzed on instrument BNA_M using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOCMS Group4 was based on method 8270E and extraction was done based on method 3541.

D. QA/ QC Samples:
The Holding Times were met for all analysis.
The Surrogate recoveries met the acceptable criteria.
The Internal Standards Areas met the acceptable requirements.
The Retention Times were acceptable for all samples.
The MS recoveries met the requirements for all compounds.
The MSD recoveries met the acceptable requirements.
The RPD met criteria.
The Blank Spike met requirements for all samples.
The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements.

284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Continuous Calibration File ID BP024568.D met the requirements except for 2,4,6-Tribromophenol, Failed compound is not associated with reporting list, therefor no further corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:
The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.
The soil samples results are based on a dry weight basis.
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:
Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q1872 MATRIX: Solid

METHOD: 8270E/3541

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			✓
	The Initial Calibration met the requirements. The Continuous Calibration File ID BP024568.D met the requirements except for 2,4,6-Tribromophenol, Failed compound is not associated with reporting list, therefor no further corrective action was taken.			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The Blank Spike met requirements for all samples.			

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

		NA	NO	YES
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met		✓	
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met		✓	
	If not met, list number of days exceeded for each sample:			
	The Holding Times were met for all analysis.			

ADDITIONAL COMMENTS:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

_____ _____

QA REVIEW Date

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1872

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: SOHIL JODHANI **Date:** 05/28/2025

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1872	**OrderDate:**	4/24/2025 1:26:50 PM
Client:	Alliance Technical Group, LLC - Newark	**Project:**	NJ Soil PT
Contact:	Mohammad Ahmed	**Location:**	QA Office,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1872-11	**HW0425-PT-BNA-SOIL**	**SOIL**			**04/21/25**			**04/24/25**
			SVOCMS Group1	8270E		05/08/25	05/14/25	
			SVOCMS Group2	8270-Modified		05/08/25	05/13/25	
Q1872-11DL	**HW0425-PT-BNA-SOILDL**	**SOIL**			**04/21/25**			**04/24/25**
			SVOCMS Group1	8270E		05/08/25	05/14/25	
			SVOCMS Group2	8270-Modified		05/08/25	05/13/25	
Q1872-12	**HW0425-PT-TRIAZINE-SOIL**	**SOIL**			**04/21/25**			**04/24/25**
			SVOCMS Group4	8270E		05/08/25	05/14/25	
			SVOCMS Group5	8270-Modified		05/08/25	05/13/25	
Q1872-13	**HW0425-PT-PAH-SOIL**	**SOIL**			**04/21/25**			**04/24/25**
			SVOCMS Group3	8270-Modified		05/08/25	05/14/25	
Q1872-13DL	**HW0425-PT-PAH-SOILDL**	**SOIL**			**04/21/25**			**04/24/25**
			SVOCMS Group3	8270-Modified		05/08/25	05/14/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								
				0.000				
			Total Svoc :			**0.00**		
			Total Concentration:			**0.00**		

QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%) Low	Limits (%) High
PB167904BL	PB167904BL	Nitrobenzene-d5	100	88.3	88		18	107
		2-Fluorobiphenyl	100	95.4	95		20	109
		Terphenyl-d14	100	95.5	96		10	105
PB167904BS	PB167904BS	Nitrobenzene-d5	100	88.0	88		18	107
		2-Fluorobiphenyl	100	91.1	91		20	109
		Terphenyl-d14	100	102	102		10	105
Q1872-12	HW0425-PT-TRIAZINE-SOIL	Nitrobenzene-d5	100	71.2	71		18	107
		2-Fluorobiphenyl	100	73.9	74		20	109
		Terphenyl-d14	100	92.8	93		10	105
Q1964-01MS	MH-LLMS	Nitrobenzene-d5	100	62.7	63		18	107
		2-Fluorobiphenyl	100	55.9	56		20	109
		Terphenyl-d14	100	51.3	51		10	105
Q1964-01MSD	MH-LLMSD	Nitrobenzene-d5	100	60.5	61		18	107
		2-Fluorobiphenyl	100	53.4	53		20	109
		Terphenyl-d14	100	47.9	48		10	105

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: Q1964-01MS		**Client Sample ID:** MH-LLMS						**DataFile:** BM050146.D			
Atrazine	1200	0	1100	ug/Kg	92				79	127	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Limits Low	Limits High	RPD
Lab Sample ID: Q1964-01MSD		**Client Sample ID:** MH-LLMSD						**DataFile:** BM050147.D			
Atrazine	1200	0	1100	ug/Kg	92		0		79	127	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1872

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8270E **DataFile:** BP024570.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB167904BS	Atrazine	1700	2500	ug/Kg	147				47	152	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167904BL

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BP024569.D Lab Sample ID: PB167904BL

Instrument ID: BNA_P Date Extracted: 05/08/2025

Matrix: (soil/water) SOIL Date Analyzed: 05/08/2025

Level: (low/med) LOW Time Analyzed: 15:26

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167904BS	PB167904BS	BP024570.D	05/08/2025
HW0425-PT-TRIAZINE-SOIL	Q1872-12	BP024631.D	05/14/2025
MH-LLMS	Q1964-01MS	BM050146.D	05/08/2025
MH-LLMSD	Q1964-01MSD	BM050147.D	05/08/2025

COMMENTS:

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BM050023.D DFTPP Injection Date: 04/28/2025

Instrument ID: BNA_M DFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.8
68	Less than 2.0% of mass 69	0.4 (1.3) 1
69	Mass 69 relative abundance	28.5
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	35.7
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	100.0
443	15.0 - 24.0% of mass 442	14.4 (19.3) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM050024.D	04/28/2025	12:30
SSTDICC005	SSTDICC005	BM050025.D	04/28/2025	13:09
SSTDICC010	SSTDICC010	BM050026.D	04/28/2025	13:48
SSTDICC020	SSTDICC020	BM050027.D	04/28/2025	14:27
SSTDICCC040	SSTDICCC040	BM050028.D	04/28/2025	15:06
SSTDICC050	SSTDICC050	BM050029.D	04/28/2025	15:45
SSTDICC060	SSTDICC060	BM050030.D	04/28/2025	16:24
SSTDICC080	SSTDICC080	BM050031.D	04/28/2025	17:04

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BM050139.D DFTPP Injection Date: 05/08/2025

Instrument ID: BNA_M DFTPP Injection Time: 09:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	21.7
68	Less than 2.0% of mass 69	0.4 (1.4) 1
69	Mass 69 relative abundance	25.8
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	33.5
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27.3
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	13.6
442	Greater than 50% of mass 198	100.0
443	15.0 - 24.0% of mass 442	16.8 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050140.D	05/08/2025	10:41
MH-LLMS	Q1964-01MS	BM050146.D	05/08/2025	14:42
MH-LLMSD	Q1964-01MSD	BM050147.D	05/08/2025	15:21

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BP024274.D DFTPP Injection Date: 04/14/2025

Instrument ID: BNA_P DFTPP Injection Time: 10:25

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.4
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	48.5
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	29.1
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100.0
443	15.0 - 24.0% of mass 442	19.2 (19.4) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024275.D	04/14/2025	11:06
SSTDICC005	SSTDICC005	BP024276.D	04/14/2025	11:47
SSTDICC010	SSTDICC010	BP024277.D	04/14/2025	12:27
SSTDICC020	SSTDICC020	BP024278.D	04/14/2025	13:08
SSTDICCC040	SSTDICCC040	BP024279.D	04/14/2025	13:49
SSTDICC050	SSTDICC050	BP024280.D	04/14/2025	15:10
SSTDICC060	SSTDICC060	BP024281.D	04/14/2025	16:32
SSTDICC080	SSTDICC080	BP024282.D	04/14/2025	17:13

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BP024567.D DFTPP Injection Date: 05/08/2025

Instrument ID: BNA_P DFTPP Injection Time: 13:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	27.9
68	Less than 2.0% of mass 69	0.3 (1.2) 1
69	Mass 69 relative abundance	30.1
70	Less than 2.0% of mass 69	0.1 (0.6) 1
127	10.0 - 80.0% of mass 198	45.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	33.5
365	Greater than 1% of mass 198	5.2
441	Present, but less than mass 443	20.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024568.D	05/08/2025	14:45
PB167904BL	PB167904BL	BP024569.D	05/08/2025	15:26
PB167904BS	PB167904BS	BP024570.D	05/08/2025	16:07

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BP024613.D DFTPP Injection Date: 05/13/2025

Instrument ID: BNA_P DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.5
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	36.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	49.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	33.1
365	Greater than 1% of mass 198	5.4
441	Present, but less than mass 443	19.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024614.D	05/13/2025	10:41
SSTDICC005	SSTDICC005	BP024615.D	05/13/2025	11:22
SSTDICC010	SSTDICC010	BP024616.D	05/13/2025	12:03
SSTDICC020	SSTDICC020	BP024617.D	05/13/2025	12:43
SSTDICCC040	SSTDICCC040	BP024618.D	05/13/2025	13:24
SSTDICC050	SSTDICC050	BP024619.D	05/13/2025	14:05
SSTDICC060	SSTDICC060	BP024620.D	05/13/2025	14:45
SSTDICC080	SSTDICC080	BP024621.D	05/13/2025	15:26

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK

DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM SAS No.: Q1872 SDG NO.: Q1872

Lab File ID: BP024624.D DFTPP Injection Date: 05/14/2025

Instrument ID: BNA_P DFTPP Injection Time: 10:29

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.4
68	Less than 2.0% of mass 69	0.5 (1.9) 1
69	Mass 69 relative abundance	34.4
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	48.6
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	33.4
365	Greater than 1% of mass 198	5.8
441	Present, but less than mass 443	21.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.5 (19.5) 2

1-Value is % mass 69 2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024625.D	05/14/2025	11:35
HW0425-PT-TRIAZINE-SOIL	Q1872-12	BP024631.D	05/14/2025	18:37

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/08/2025

Lab File ID: BM050140.D Time Analyzed: 10:41

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	370324	7.74	1368860	10.53	879874	14.39
	UPPER LIMIT	740648	8.24	2737720	11.034	1759750	14.892
	LOWER LIMIT	185162	7.24	684430	10.034	439937	13.892
	EPA SAMPLE NO.						
01	MH-LLMS	264341	7.74	906955	10.53	562871	14.39
02	MH-LLMSD	253439	7.74	884473	10.53	556753	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/08/2025

Lab File ID: BM050140.D Time Analyzed: 10:41

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	1654930	17.139	1237710	21.386	1026250	24.38
	UPPER LIMIT	3309860	17.639	2475420	21.886	2052500	24.88
	LOWER LIMIT	827465	16.639	618855	20.886	513125	23.88
	EPA SAMPLE NO.						
01	MH-LLMS	1043030	17.14	1002000	21.39	1081330	24.38
02	MH-LLMSD	1069460	17.14	1112270	21.39	1161580	24.38

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/08/2025

Lab File ID: BP024568.D Time Analyzed: 14:45

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	179226	7.699	763529	10.48	492115	14.33
	UPPER LIMIT	358452	8.199	1527060	10.975	984230	14.834
	LOWER LIMIT	89613	7.199	381765	9.975	246058	13.834
	EPA SAMPLE NO.						
01	PB167904BL	197835	7.70	736878	10.48	410163	14.33
02	PB167904BS	175167	7.71	732511	10.48	497695	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/08/2025

Lab File ID: BP024568.D Time Analyzed: 14:45

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	999951	17.139	1084300	21.58	1119940	24.898
	UPPER LIMIT	1999900	17.639	2168600	22.08	2239880	25.398
	LOWER LIMIT	499976	16.639	542150	21.08	559970	24.398
	EPA SAMPLE NO.						
01	PB167904BL	728812	17.14	697724	21.58	845260	24.92
02	PB167904BS	979666	17.13	1019340	21.58	1024840	24.90

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/14/2025

Lab File ID: BP024625.D Time Analyzed: 11:35

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
	12 HOUR STD	115714	7.628	507416	10.41	336535	14.29
	UPPER LIMIT	231428	8.128	1014830	10.91	673070	14.786
	LOWER LIMIT	57857	7.128	253708	9.91	168268	13.786
	EPA SAMPLE NO.						
01	HW0425-PT-TRIAZINE-SOIL	89688	7.66	382712	10.43	254481	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/14/2025

Lab File ID: BP024625.D Time Analyzed: 11:35

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	705050	17.092	813969	21.533	940710	24.821
	UPPER LIMIT	1410100	17.592	1627940	22.033	1881420	25.321
	LOWER LIMIT	352525	16.592	406985	21.033	470355	24.321
	EPA SAMPLE NO.						
01	HW0425-PT-TRIAZINE-SOIL	547126	17.09	601577	21.52	715674	24.79

IS4 (PHN) = Phenanthrene-d10
IS5 (CRY) = Chrysene-d12
IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

SAMPLE DATA

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	04/21/25
Project:	NJ Soil PT	Date Received:	04/24/25
Client Sample ID:	HW0425-PT-TRIAZINE-SOIL	SDG No.:	Q1872
Lab Sample ID:	Q1872-12	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.37 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group4
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024631.D	1	05/08/25 09:15	05/14/25 18:37	PB167904

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1912-24-9	Atrazine	33.6	U	33.6	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	71.2		18 - 107	71%	SPK: 100
321-60-8	2-Fluorobiphenyl	73.9		20 - 109	74%	SPK: 100
1718-51-0	Terphenyl-d14	92.8		10 - 105	93%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	89700	7.663			
1146-65-2	Naphthalene-d8	383000	10.428			
15067-26-2	Acenaphthene-d10	254000	14.292			
1517-22-2	Phenanthrene-d10	547000	17.086			
1719-03-5	Chrysene-d12	602000	21.516			
1520-96-3	Perylene-d12	716000	24.792			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024631.D
Acq On : 14 May 2025 18:37
Operator : RC/JU
Sample : Q1872-12
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Quant Time: May 15 02:06:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	89688	20.000	ng	0.00
21) Naphthalene-d8	10.428	136	382712	20.000	ng	0.00
39) Acenaphthene-d10	14.292	164	254481	20.000	ng	-0.01
64) Phenanthrene-d10	17.086	188	547126	20.000	ng	-0.01
76) Chrysene-d12	21.516	240	601577	20.000	ng	0.00
86) Perylene-d12	24.792	264	715674	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.287	112	615659	117.407	ng	0.00
7) Phenol-d6	6.852	99	849640	126.954	ng	0.00
23) Nitrobenzene-d5	8.810	82	539220	71.189	ng	0.00
42) 2,4,6-Tribromophenol	15.810	330	641020	148.584	ng	0.00
45) 2-Fluorobiphenyl	12.904	172	1434975	73.876	ng	0.00
79) Terphenyl-d14	19.810	244	3254388	92.761	ng	0.00
Target Compounds						Qvalue
84) Bis(2-ethylhexyl)phtha...	21.433	149	60411	2.564	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024631.D
Acq On : 14 May 2025 18:37
Operator : RC/JU
Sample : Q1872-12
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Quant Time: May 15 02:06:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.663 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Tgt Ion:152 Resp: 89688

Ion	Ratio	Lower	Upper
152	100		
150	153.3	125.9	188.9
115	64.0	50.1	75.1





#5
2-Fluorophenol
Concen: 117.407 ng
RT: 5.287 min Scan# 374
Delta R.T. -0.000 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Tgt Ion:112 Resp: 615659

Ion	Ratio	Lower	Upper
112	100		
64	53.8	44.1	66.1
63	29.3	23.5	35.3





#7
Phenol-d6
Concen: 126.954 ng
RT: 6.852 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Tgt Ion: 99 Resp: 849640

Ion	Ratio	Lower	Upper
99	100		
42	17.4	14.4	21.6
71	35.8	29.2	43.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.428 min Scan# 1248
Delta R.T. -0.006 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Tgt Ion:136 Resp: 382712

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.8	13.2
54	6.8	5.5	8.3
68	5.1	4.2	6.2





#23
Nitrobenzene-d5
Concen: 71.189 ng
RT: 8.810 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Tgt Ion: 82 Resp: 539220

Ion	Ratio	Lower	Upper
82	100		
128	45.5	37.0	55.4
54	46.1	38.7	58.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.292 min Scan# 1905
Delta R.T. -0.012 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Tgt Ion:164 Resp: 254481

Ion	Ratio	Lower	Upper
164	100		
162	101.8	81.8	122.6
160	46.0	36.3	54.5





#42
2,4,6-Tribromophenol
Concen: 148.584 ng
RT: 15.810 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Tgt Ion:330 Resp: 641020

Ion	Ratio	Lower	Upper
330	100		
332	97.7	78.2	117.4
141	29.5	24.3	36.5





#45
2-Fluorobiphenyl
Concen: 73.876 ng
RT: 12.904 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Tgt Ion:172 Resp: 1434975

Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.8	43.2
170	23.4	18.8	28.2





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.086 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Tgt Ion:188 Resp: 547126

Ion	Ratio	Lower	Upper
188	100		
94	8.7	6.5	9.7
80	9.8	7.3	10.9





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.516 min Scan# 3133
Delta R.T. -0.006 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Tgt Ion:240 Resp: 601577

Ion	Ratio	Lower	Upper
240	100		
120	10.3	7.8	11.6
236	25.8	20.6	31.0





#79
Terphenyl-d14
Concen: 92.761 ng
RT: 19.810 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Tgt Ion:244 Resp: 3254388

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.5	8.3
122	8.9	7.4	11.0





#84
Bis(2-ethylhexyl)phthalate
Concen: 2.564 ng
RT: 21.433 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Tgt Ion:149 Resp: 60411

Ion	Ratio	Lower	Upper
149	100		
167	26.2	22.1	33.1
279	4.6	4.6	6.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.792 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024631.D
Acq: 14 May 2025 18:37

Instrument :
BNA_P
ClientSampleId :
HW0425-PT-TRIAZINE-SOIL

Tgt Ion:264 Resp: 715674

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.7	28.1
265	22.0	17.0	25.6



CALIBRATION SUMMARY

Response Factor Report BNA_M

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM042825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Apr 28 18:09:16 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BM050024.D 5 =BM050025.D 10 =BM050026.D 20 =BM050027.D 40 =BM050028.D 50 =BM050029.D 60 =BM050030.D 80 =BM0500
31.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	ISTD									
2)		1,4-Dioxane		0.506	0.464	0.478	0.511	0.504	0.491	0.473	0.490	3.75
3)		Pyridine		1.201	1.166	1.229	1.331	1.328	1.292	1.261	1.258	5.03
4)		n-Nitrosodimet...		0.487	0.459	0.475	0.522	0.518	0.504	0.490	0.493	4.62
5)	S	2-Fluorophenol		1.074	1.079	1.107	1.208	1.211	1.173	1.142	1.142	5.04
6)		Aniline		1.657	1.689	1.768	1.942	1.922	1.875	1.835	1.813	6.16
7)	S	Phenol-d6		1.290	1.318	1.398	1.543	1.537	1.494	1.469	1.436	7.13
8)		2-Chlorophenol		1.169	1.161	1.198	1.312	1.301	1.268	1.235	1.235	4.96
9)		Benzaldehyde		0.940	0.920	0.971	1.031	1.013	0.956	0.894	0.961	5.08
10)	C	Phenol		1.349	1.344	1.420	1.539	1.537	1.501	1.483	1.453	5.73
11)		bis(2-Chloroet...		1.213	1.105	1.176	1.279	1.271	1.233	1.218	1.213	4.90
12)		1,3-Dichlorobe...		1.483	1.428	1.453	1.566	1.552	1.508	1.453	1.492	3.51
13)	C	1,4-Dichlorobe...		1.518	1.418	1.463	1.561	1.557	1.516	1.458	1.499	3.60
14)		1,2-Dichlorobe...		1.420	1.393	1.423	1.522	1.506	1.464	1.405	1.448	3.50
15)		Benzyl Alcohol		0.866	0.881	0.966	1.080	1.074	1.050	1.050	0.996	9.16
16)		2,2'-oxybis(1-...		1.487	1.422	1.445	1.552	1.513	1.473	1.433	1.475	3.17
17)		2-Methylphenol		0.857	0.893	0.931	1.033	1.026	0.988	0.975	0.957	6.96
18)		Hexachloroethane		0.552	0.508	0.521	0.556	0.548	0.533	0.510	0.533	3.79
19)	P	n-Nitroso-di-n...	0.802	0.889	0.885	0.921	1.003	0.988	0.949	0.928	0.921	6.95
20)		3+4-Methylphenols		1.137	1.170	1.267	1.393	1.392	1.349	1.341	1.293	8.09
21)	I	Naphthalene-d8	ISTD									
22)		Acetophenone		0.483	0.464	0.485	0.531	0.525	0.506	0.487	0.497	4.94
23)	S	Nitrobenzene-d5		0.386	0.374	0.397	0.434	0.433	0.417	0.400	0.406	5.68
24)		Nitrobenzene		0.337	0.325	0.345	0.374	0.372	0.358	0.346	0.351	5.16
25)		Isophorone		0.666	0.648	0.680	0.729	0.729	0.704	0.687	0.692	4.44
26)	C	2-Nitrophenol		0.148	0.157	0.172	0.196	0.198	0.192	0.191	0.179	11.33
27)		2,4-Dimethylph...		0.263	0.266	0.289	0.324	0.322	0.312	0.305	0.297	8.42
28)		bis(2-Chloroet...		0.407	0.399	0.422	0.455	0.452	0.434	0.425	0.428	4.92
29)	C	2,4-Dichloroph...		0.294	0.298	0.326	0.362	0.359	0.351	0.342	0.333	8.47
30)		1,2,4-Trichlor...		0.400	0.384	0.394	0.427	0.426	0.412	0.399	0.406	4.02
31)		Naphthalene		1.001	0.959	0.993	1.066	1.057	1.023	0.989	1.013	3.80
32)		Benzoic acid			0.138	0.172	0.197	0.206	0.212	0.215	0.190	15.67
33)		4-Chloroaniline		0.361	0.393	0.404	0.441	0.448	0.437	0.423	0.415	7.51
34)	C	Hexachlorobuta...		0.253	0.240	0.249	0.270	0.270	0.263	0.258	0.258	4.34
35)		Caprolactam		0.090	0.085	0.095	0.104	0.104	0.102	0.100	0.097	7.73
36)	C	4-Chloro-3-met...		0.289	0.283	0.302	0.325	0.329	0.317	0.309	0.308	5.72
37)		2-Methylnaphth...		0.656	0.639	0.662	0.713	0.714	0.692	0.678	0.679	4.22
38)		1-Methylnaphth...		0.703	0.675	0.708	0.753	0.751	0.731	0.710	0.719	3.92

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM042825.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.624	0.622	0.655	0.742	0.745	0.725	0.733	0.692	8.17
41)	P	Hexachlorocycl...		0.303	0.362	0.455	0.468	0.467	0.483	0.423	17.30
42)	S	2,4,6-Tribromo...	0.253	0.249	0.281	0.318	0.329	0.318	0.323	0.296	11.61
43)	C	2,4,6-Trichlor...	0.372	0.381	0.423	0.476	0.485	0.467	0.469	0.439	10.76
44)		2,4,5-Trichlor...	0.410	0.417	0.453	0.517	0.519	0.501	0.497	0.474	9.83
45)	S	2-Fluorobiphenyl	1.588	1.555	1.632	1.807	1.822	1.744	1.688	1.691	6.21
46)		1,1'-Biphenyl	1.427	1.395	1.452	1.573	1.581	1.507	1.473	1.487	4.77
47)		2-Chloronaphth...	1.126	1.112	1.155	1.250	1.248	1.192	1.158	1.177	4.69
48)		2-Nitroaniline	0.238	0.242	0.269	0.302	0.304	0.294	0.286	0.276	9.96
49)		Acenaphthylene	1.794	1.715	1.810	1.972	1.971	1.889	1.845	1.857	5.10
50)		Dimethylphthalate	1.444	1.377	1.437	1.532	1.543	1.470	1.429	1.462	4.02
51)		2,6-Dinitrotol...	0.266	0.271	0.299	0.327	0.331	0.316	0.309	0.303	8.50
52)	C	Acenaphthene	1.038	0.998	1.046	1.133	1.145	1.094	1.067	1.075	4.93
53)		3-Nitroaniline	0.233	0.247	0.284	0.321	0.325	0.311	0.302	0.289	12.47
54)	P	2,4-Dinitrophenol		0.100	0.132	0.178	0.194	0.188	0.191	0.164	23.65
55)		Dibenzofuran	1.749	1.681	1.749	1.876	1.888	1.810	1.763	1.788	4.17
56)	P	4-Nitrophenol		0.128	0.180	0.224	0.237	0.231	0.231	0.205	20.96
57)		2,4-Dinitrotol...	0.338	0.366	0.412	0.456	0.468	0.447	0.440	0.418	11.72
58)		Fluorene	1.387	1.343	1.431	1.558	1.573	1.500	1.453	1.463	5.84
59)		2,3,4,6-Tetrac...	0.366	0.362	0.398	0.435	0.445	0.432	0.431	0.410	8.44
60)		Diethylphthalate	1.372	1.301	1.380	1.433	1.451	1.376	1.328	1.377	3.86
61)		4-Chlorophenyl...	0.741	0.716	0.770	0.857	0.871	0.836	0.836	0.804	7.58
62)		4-Nitroaniline	0.208	0.234	0.277	0.313	0.321	0.301	0.295	0.279	15.17
63)		Azobenzene	1.118	1.091	1.150	1.216	1.229	1.165	1.110	1.154	4.59
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.088	0.110	0.135	0.142	0.138	0.137	0.125	17.16
66)	c	n-Nitrosodiphe...	0.580	0.572	0.596	0.655	0.657	0.629	0.610	0.614	5.56
67)		4-Bromophenyl-...	0.223	0.216	0.229	0.258	0.258	0.256	0.253	0.242	7.60
68)		Hexachlorobenzene	0.265	0.251	0.264	0.294	0.295	0.291	0.291	0.279	6.49
69)		Atrazine	0.196	0.198	0.214	0.236	0.242	0.234	0.230	0.222	8.47
70)	C	Pentachlorophenol		0.117	0.141	0.167	0.172	0.170	0.174	0.157	14.52
71)		Phenanthrene	1.075	1.035	1.083	1.187	1.212	1.173	1.132	1.128	5.84
72)		Anthracene	1.062	1.038	1.096	1.214	1.236	1.191	1.154	1.142	6.78
73)		Carbazole	0.909	0.900	0.960	1.054	1.075	1.038	0.999	0.991	7.05
74)		Di-n-butylphth...	1.106	1.082	1.151	1.246	1.277	1.231	1.165	1.180	6.25
75)	C	Fluoranthene	1.178	1.142	1.250	1.410	1.455	1.425	1.411	1.324	9.86
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine		0.560	0.646	0.772	0.769	0.765	0.751	0.710	12.36
78)		Pyrene	1.290	1.255	1.332	1.518	1.520	1.465	1.453	1.404	7.83
79)	S	Terphenyl-d14	1.176	1.189	1.332	1.524	1.541	1.458	1.239	1.352	11.58
80)		Butylbenzylpht...	0.489	0.479	0.515	0.562	0.568	0.545	0.524	0.526	6.53
81)		Benzo(a)anthra...	1.280	1.231	1.321	1.478	1.479	1.442	1.408	1.377	7.24
82)		3,3'-Dichlorob...	0.434	0.432	0.483	0.580	0.597	0.583	0.581	0.527	14.14
83)		Chrysene	1.208	1.170	1.214	1.342	1.369	1.325	1.296	1.275	6.03
84)		Bis(2-ethylhex...	0.726	0.725	0.775	0.843	0.852	0.814	0.765	0.786	6.61
85)	c	Di-n-octyl pht...	1.221	1.197	1.267	1.366	1.404	1.341	1.275	1.296	5.92

Response Factor Report BNA_M

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM042825.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.301	1.293	1.404	1.602	1.646	1.591	1.595	1.490	10.27
88)		Benzo(b)fluora...	1.128	1.149	1.209	1.406	1.415	1.408	1.390	1.301	10.16
89)		Benzo(k)fluora...	1.219	1.150	1.251	1.387	1.449	1.383	1.372	1.316	8.29
90)	C	Benzo(a)pyrene	1.094	1.066	1.147	1.295	1.330	1.298	1.297	1.218	9.15
91)		Dibenzo(a,h)an...	1.056	1.044	1.142	1.300	1.347	1.305	1.310	1.215	10.72
92)		Benzo(g,h,i)pe...	1.066	1.030	1.105	1.236	1.268	1.224	1.212	1.163	8.07

(#) = Out of Range

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050024.D
Acq On : 28 Apr 2025 12:30
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 28 17:52:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	241974	20.000	ng	0.00
21) Naphthalene-d8	10.563	136	868675	20.000	ng	0.00
39) Acenaphthene-d10	14.416	164	604723	20.000	ng	0.00
64) Phenanthrene-d10	17.157	188	1218544	20.000	ng	0.00
76) Chrysene-d12	21.404	240	1187007	20.000	ng	0.00
86) Perylene-d12	24.409	264	1217038	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	8.581	70	24247	2.177	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050024.D
Acq On : 28 Apr 2025 12:30
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 28 17:52:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050024.D
Acq On : 28 Apr 2025 12:30
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 28 17:52:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 241974

Ion	Ratio	Lower	Upper
152	100		
150	153.0	124.1	186.1
115	53.4	41.2	61.8





#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.36 min

Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	53.3
63	27.8





#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.95 min

Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99

Sig	Exp Ratio
99	100
42	16.2
71	32.0









#19
n-Nitroso-di-n-propylamine
Concen: 2.177 ng
RT: 8.581 min Scan# 951
Delta R.T. 0.006 min
Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Tgt Ion: 70 Resp: 24247

Ion	Ratio	Lower	Upper
70	100		
42	47.1	40.5	60.7
101	8.4	7.8	11.6
130	23.2	17.4	26.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 868675

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	6.4	5.1	7.7
68	5.5	4.4	6.6







#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 8.93 min

Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	44.6
54	44.3



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.416 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 604723

Ion	Ratio	Lower	Upper
164	100		
162	99.3	79.4	119.0
160	45.3	36.2	54.4





#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.90 min

Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	96.6
141	28.1





#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 13.03 min

Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172

Sig	Exp Ratio
172	100
171	34.7
170	23.3





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Tgt Ion:188 Resp: 1218544

Ion	Ratio	Lower	Upper
188	100		
94	7.5	5.7	8.5
80	8.1	6.2	9.4









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.404 min Scan# 3
Delta R.T. 0.007 min
Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 1187007

Ion	Ratio	Lower	Upper
240	100		
120	8.7	6.5	9.7
236	25.9	20.9	31.3





#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 19.78 min

Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.0
122	7.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.409 min Scan# 3
Delta R.T. 0.006 min
Lab File: BM050024.D
Acq: 28 Apr 2025 12:30

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 1217038

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.8	28.2
265	21.9	17.0	25.4



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050025.D
Acq On : 28 Apr 2025 13:09
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Quant Time: Apr 28 17:53:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	222427	20.000	ng		0.00
21) Naphthalene-d8	10.563	136	815362	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	578294	20.000	ng		0.00
64) Phenanthrene-d10	17.157	188	1183325	20.000	ng		0.00
76) Chrysene-d12	21.403	240	1136162	20.000	ng		0.00
86) Perylene-d12	24.403	264	1149910	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.363	112	119473	9.406	ng		0.00
7) Phenol-d6	6.957	99	143453	8.985	ng		0.00
23) Nitrobenzene-d5	8.928	82	157222	9.502	ng		0.00
42) 2,4,6-Tribromophenol	15.910	330	73192	8.560	ng		0.00
45) 2-Fluorobiphenyl	13.033	172	459031	9.390	ng		0.00
79) Terphenyl-d14	19.780	244	668289	8.704	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.257	88	28163	5.171	ng		98
3) Pyridine	3.669	79	66766	4.771	ng		98
4) n-Nitrosodimethylamine	3.575	42	27060	4.930	ng	#	92
6) Aniline	7.098	93	92126	4.570	ng		97
8) 2-Chlorophenol	7.345	128	64999	4.733	ng		99
9) Benzaldehyde	6.916	77	52286m	4.900	ng		
10) Phenol	6.987	94	75011	4.641	ng		99
11) bis(2-Chloroethyl)ether	7.192	93	67447	4.998	ng		98
12) 1,3-Dichlorobenzene	7.657	146	82466	4.971	ng		98
13) 1,4-Dichlorobenzene	7.804	146	84400	5.064	ng		97
14) 1,2-Dichlorobenzene	8.122	146	78966	4.905	ng		99
15) Benzyl Alcohol	8.022	79	48180	4.352	ng		95
16) 2,2'-oxybis(1-Chloropr...	8.298	45	82711	5.042	ng		93
17) 2-Methylphenol	8.234	107	47632	4.474	ng		99
18) Hexachloroethane	8.845	117	30717	5.185	ng		97
19) n-Nitroso-di-n-propyla...	8.575	70	49455	4.830	ng		98
20) 3+4-Methylphenols	8.575	107	63223	4.397	ng	#	91
22) Acetophenone	8.598	105	98358	4.852	ng		98
24) Nitrobenzene	8.975	77	68598	4.795	ng		94
25) Isophorone	9.498	82	135828	4.816	ng		98
26) 2-Nitrophenol	9.692	139	30257	4.140	ng		95
27) 2,4-Dimethylphenol	9.763	122	53660	4.427	ng		99
28) bis(2-Chloroethoxy)met...	9.986	93	82947	4.758	ng		99
29) 2,4-Dichlorophenol	10.263	162	59828	4.406	ng		97
30) 1,2,4-Trichlorobenzene	10.428	180	81624	4.934	ng		97
31) Naphthalene	10.616	128	204133	4.944	ng		98
33) 4-Chloroaniline	10.751	127	73542	4.345	ng		97
34) Hexachlorobutadiene	10.898	225	51529	4.907	ng		98
35) Caprolactam	11.522	113	18349m	4.634	ng		
36) 4-Chloro-3-methylphenol	11.886	107	58843	4.690	ng		97
37) 2-Methylnaphthalene	12.233	142	133679	4.829	ng		97
38) 1-Methylnaphthalene	12.451	142	143332	4.892	ng		99
40) 1,2,4,5-Tetrachloroben...	12.604	216	90168	4.504	ng		98
43) 2,4,6-Trichlorophenol	12.863	196	53748	4.233	ng		96
44) 2,4,5-Trichlorophenol	12.951	196	59224	4.325	ng		98
46) 1,1'-Biphenyl	13.245	154	206296	4.799	ng		98

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050025.D
Acq On : 28 Apr 2025 13:09
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Quant Time: Apr 28 17:53:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
47) 2-Chloronaphthalene	13.286	162	162803	4.782	ng		98
48) 2-Nitroaniline	13.504	65	34348	4.299	ng		91
49) Acenaphthylene	14.133	152	259406	4.832	ng		100
50) Dimethylphthalate	13.869	163	208800	4.941	ng		100
51) 2,6-Dinitrotoluene	13.992	165	38453	4.395	ng		94
52) Acenaphthene	14.474	154	150042	4.829	ng		99
53) 3-Nitroaniline	14.339	138	33713	4.036	ng	#	94
55) Dibenzofuran	14.816	168	252898	4.892	ng		99
57) 2,4-Dinitrotoluene	14.798	165	48901	4.044	ng	#	87
58) Fluorene	15.463	166	200542	4.739	ng		95
59) 2,3,4,6-Tetrachlorophenol	15.057	232	52899	4.463	ng		97
60) Diethylphthalate	15.233	149	198362	4.981	ng		98
61) 4-Chlorophenyl-phenyle...	15.457	204	107081	4.607	ng		94
62) 4-Nitroaniline	15.515	138	30138m	3.737	ng		
63) Azobenzene	15.751	77	161662	4.844	ng		98
66) n-Nitrosodiphenylamine	15.674	169	171608	4.723	ng		97
67) 4-Bromophenyl-phenylether	16.351	248	66000	4.612	ng		95
68) Hexachlorobenzene	16.468	284	78506	4.759	ng		99
69) Atrazine	16.627	200	58110	4.433	ng		99
71) Phenanthrene	17.198	178	317954	4.764	ng		99
72) Anthracene	17.292	178	314149	4.651	ng		99
73) Carbazole	17.574	167	268774	4.586	ng		99
74) Di-n-butylphthalate	18.127	149	327178	4.688	ng		100
75) Fluoranthene	19.221	202	348593	4.449	ng		99
78) Pyrene	19.586	202	366473	4.593	ng		99
80) Butylbenzylphthalate	20.480	149	138827	4.646	ng		97
81) Benzo(a)anthracene	21.380	228	363641	4.648	ng		99
82) 3,3'-Dichlorobenzidine	21.315	252	123375	4.119	ng	#	95
83) Chrysene	21.445	228	343085	4.738	ng		98
84) Bis(2-ethylhexyl)phtha...	21.297	149	206231	4.620	ng		96
85) Di-n-octyl phthalate	22.427	149	346799	4.711	ng		98
87) Indeno(1,2,3-cd)pyrene	27.803	276	373944	4.364	ng	#	87
88) Benzo(b)fluoranthene	23.456	252	324395	4.338	ng		99
89) Benzo(k)fluoranthene	23.515	252	350374m	4.631	ng		
90) Benzo(a)pyrene	24.262	252	314608	4.492	ng		98
91) Dibenzo(a,h)anthracene	27.856	278	303557	4.346	ng		98
92) Benzo(g,h,i)perylene	28.862	276	306571	4.585	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050025.D
Acq On : 28 Apr 2025 13:09
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Apr 28 17:53:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050025.D
Acq On : 28 Apr 2025 13:09
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Apr 28 17:53:22 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:152 Resp: 22242

Ion	Ratio	Lower	Upper
152	100		
150	152.9	124.1	186.1
115	51.5	41.2	61.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#2
1,4-Dioxane
Concen: 5.171 ng
RT: 3.257 min Scan# 46
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 88 Resp: 28163

Ion	Ratio	Lower	Upper
88	100		
58	67.8	55.1	82.7
43	27.3	20.9	31.3





#3
Pyridine
Concen: 4.771 ng
RT: 3.669 min Scan# 1
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 79 Resp: 6676

Ion	Ratio	Lower	Upper
79	100		
52	60.5	48.6	72.8
51	34.0	24.6	37.0

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#4
n-Nitrosodimethylamine
Concen: 4.930 ng
RT: 3.575 min Scan# 100
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 42 Resp: 27060

Ion	Ratio	Lower	Upper
42	100		
74	144.5	120.2	180.2
44	32.5	6.2	9.2#





#5
2-Fluorophenol
Concen: 9.406 ng
RT: 5.363 min Scan# 4
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:112 Resp: 11947

Ion	Ratio	Lower	Upper
112	100		
64	54.4	42.6	64.0
63	30.2	22.2	33.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#6
Aniline
Concen: 4.570 ng
RT: 7.098 min Scan# 699
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 93 Resp: 92126

Ion	Ratio	Lower	Upper
93	100		
66	38.9	29.9	44.9
65	20.5	14.5	21.7





#7
Phenol-d6
Concen: 8.985 ng
RT: 6.957 min Scan# 6
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 99 Resp: 14345

Ion	Ratio	Lower	Upper
99	100		
42	15.6	13.0	19.4
71	30.9	25.6	38.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#8
2-Chlorophenol
Concen: 4.733 ng
RT: 7.345 min Scan# 741
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:128 Resp: 64999

Ion	Ratio	Lower	Upper
128	100		
130	32.3	12.8	52.8
64	42.0	21.2	61.2





#9
Benzaldehyde
Concen: 4.900 ng m
RT: 6.916 min Scan# 6
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 77 Resp: 5228

Ion	Ratio	Lower	Upper
77	100		
105	103.4	80.5	120.5
106	94.6	76.1	116.1



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#10
Phenol
Concen: 4.641 ng
RT: 6.987 min Scan# 680
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 94 Resp: 75011

Ion	Ratio	Lower	Upper
94	100		
65	28.6	7.6	47.6
66	35.3	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 4.998 ng
RT: 7.192 min Scan# 7
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 93 Resp: 6744

Ion	Ratio	Lower	Upper
93	100		
63	69.8	52.3	92.3
95	32.0	12.8	52.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#12
1,3-Dichlorobenzene
Concen: 4.971 ng
RT: 7.657 min Scan# 794
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:146 Resp: 82466

Ion	Ratio	Lower	Upper
146	100		
148	66.8	52.5	78.7
75	29.2	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 5.064 ng
RT: 7.804 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:146 Resp: 8440

Ion	Ratio	Lower	Upper
146	100		
148	61.6	52.2	78.2
111	37.2	29.3	43.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#14
1,2-Dichlorobenzene
Concen: 4.905 ng
RT: 8.122 min Scan# 873
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:146 Resp: 78966

Ion	Ratio	Lower	Upper
146	100		
148	66.0	52.4	78.6
111	39.0	31.0	46.6





#15
Benzyl Alcohol
Concen: 4.352 ng
RT: 8.022 min Scan# 8
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 79 Resp: 4818

Ion	Ratio	Lower	Upper
79	100		
108	75.4	66.0	99.0
77	66.1	51.6	77.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.042 ng
RT: 8.298 min Scan# 903
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 45 Resp: 82711

Ion	Ratio	Lower	Upper
45	100		
77	20.5	0.0	36.4
79	14.5	0.0	32.7





#17
2-Methylphenol
Concen: 4.474 ng
RT: 8.234 min Scan# 8
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:107 Resp: 4763

Ion	Ratio	Lower	Upper
107	100		
108	112.6	91.1	136.7
77	45.6	36.2	54.4
79	45.8	35.8	53.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#18
Hexachloroethane
Concen: 5.185 ng
RT: 8.845 min Scan# 996
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:117 Resp: 30717

Ion	Ratio	Lower	Upper
117	100		
119	97.9	77.3	115.9
201	124.0	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 4.830 ng
RT: 8.575 min Scan# 9
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 70 Resp: 4945

Ion	Ratio	Lower	Upper
70	100		
42	50.7	40.5	60.7
101	8.2	7.8	11.6
130	19.5	17.4	26.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#20
3+4-Methylphenols
Concen: 4.397 ng
RT: 8.575 min Scan# 950
Delta R.T. 0.018 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:107 Resp: 63223

Ion	Ratio	Lower	Upper
107	100		
108	92.4	72.4	112.4
77	55.1	13.6	53.6#
79	32.3	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:136 Resp: 81536

Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.9	13.3
54	6.3	5.1	7.7
68	5.3	4.4	6.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#22
Acetophenone
Concen: 4.852 ng
RT: 8.598 min Scan# 954
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:105 Resp: 98358

Ion	Ratio	Lower	Upper
105	100		
71	2.2	1.7	2.5
51	24.3	18.6	27.8
120	24.0	18.3	27.5





#23
Nitrobenzene-d5
Concen: 9.502 ng
RT: 8.928 min Scan# 1(
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 82 Resp: 15722

Ion	Ratio	Lower	Upper
82	100		
128	42.6	35.7	53.5
54	43.3	35.4	53.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#24
Nitrobenzene
Concen: 4.795 ng
RT: 8.975 min Scan# 1018
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 77 Resp: 68598

Ion	Ratio	Lower	Upper
77	100		
123	42.2	37.9	56.9
65	13.3	10.1	15.1





#25
Isophorone
Concen: 4.816 ng
RT: 9.498 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 82 Resp: 13582

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.7	8.5
138	18.2	15.5	23.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#26
2-Nitrophenol
Concen: 4.140 ng
RT: 9.692 min Scan# 1140
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:139 Resp: 30257

Ion	Ratio	Lower	Upper
139	100		
109	25.8	21.9	32.9
65	48.0	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 4.427 ng
RT: 9.763 min Scan# 11
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:122 Resp: 5366

Ion	Ratio	Lower	Upper
122	100		
107	112.5	89.0	133.6
121	57.0	45.6	68.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#28
bis(2-Chloroethoxy)methane
Concen: 4.758 ng
RT: 9.986 min Scan# 1190
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 93 Resp: 82947

Ion	Ratio	Lower	Upper
93	100		
95	31.5	26.1	39.1
123	11.8	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 4.406 ng
RT: 10.263 min Scan# 1
Delta R.T. 0.035 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:162 Resp: 5982

Ion	Ratio	Lower	Upper
162	100		
164	67.0	45.4	85.4
98	28.1	11.3	51.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#30
1,2,4-Trichlorobenzene
Concen: 4.934 ng
RT: 10.428 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:180 Resp: 81624

Ion	Ratio	Lower	Upper
180	100		
182	99.4	77.0	115.4
145	26.2	20.6	30.8





#31
Naphthalene
Concen: 4.944 ng
RT: 10.616 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:128 Resp: 20413

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.8	13.2
127	12.0	10.4	15.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#33
4-Chloroaniline
Concen: 4.345 ng
RT: 10.751 min Scan# 1320
Delta R.T. 0.023 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:127 Resp: 73542

Ion	Ratio	Lower	Upper
127	100		
129	36.4	26.3	39.5
65	28.6	22.5	33.7
92	18.0	14.5	21.7





#34
Hexachlorobutadiene
Concen: 4.907 ng
RT: 10.898 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:225 Resp: 5152

Ion	Ratio	Lower	Upper
225	100		
223	63.4	50.2	75.2
227	65.9	50.6	76.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#35
Caprolactam
Concen: 4.634 ng m
RT: 11.522 min Scan# 1451
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:113 Resp: 18349

Ion	Ratio	Lower	Upper
113	100		
55	148.1	122.9	162.9
56	111.8	88.1	128.1





#36
4-Chloro-3-methylphenol
Concen: 4.690 ng
RT: 11.886 min Scan# 1
Delta R.T. 0.018 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:107 Resp: 5884

Ion	Ratio	Lower	Upper
107	100		
144	29.8	23.8	35.6
142	86.0	71.9	107.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#37
2-Methylnaphthalene
Concen: 4.829 ng
RT: 12.233 min Scan# 1572
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:142 Resp: 133679

Ion	Ratio	Lower	Upper
142	100		
141	84.9	70.4	105.6
115	30.7	25.4	38.2





#38
1-Methylnaphthalene
Concen: 4.892 ng
RT: 12.451 min Scan# 1(
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:142 Resp: 14333

Ion	Ratio	Lower	Upper
142	100		
141	88.7	72.1	108.1
116	3.5	2.7	4.1



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1942
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:164 Resp: 578294

Ion	Ratio	Lower	Upper
164	100		
162	98.6	79.4	119.0
160	44.6	36.2	54.4





#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.504 ng
RT: 12.604 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:216 Resp: 9016

Ion	Ratio	Lower	Upper
216	100		
214	79.1	62.3	93.5
179	20.2	15.4	23.0
108	15.3	11.7	17.5



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#42
2,4,6-Tribromophenol
Concen: 8.560 ng
RT: 15.910 min Scan# 2197
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:330 Resp: 73192

Ion	Ratio	Lower	Upper
330	100		
332	96.1	77.3	115.9
141	32.6	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 4.233 ng
RT: 12.863 min Scan# 1
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:196 Resp: 5374

Ion	Ratio	Lower	Upper
196	100		
198	102.6	79.0	118.6
200	28.7	25.1	37.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#44
2,4,5-Trichlorophenol
Concen: 4.325 ng
RT: 12.951 min Scan# 1694
Delta R.T. 0.018 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:196 Resp: 59224

Ion	Ratio	Lower	Upper
196	100		
198	97.4	76.7	115.1
97	38.4	30.1	45.1
132	23.0	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 9.390 ng
RT: 13.033 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:172 Resp: 45903

Ion	Ratio	Lower	Upper
172	100		
171	34.4	27.8	41.6
170	23.5	18.6	28.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#46
1,1'-Biphenyl
Concen: 4.799 ng
RT: 13.245 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:154 Resp: 206296

Ion	Ratio	Lower	Upper
154	100		
153	41.2	20.0	60.0
76	11.8	0.0	30.3





#47
2-Chloronaphthalene
Concen: 4.782 ng
RT: 13.286 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:162 Resp: 16280

Ion	Ratio	Lower	Upper
162	100		
127	31.2	25.9	38.9
164	32.6	26.6	40.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#48
2-Nitroaniline
Concen: 4.299 ng
RT: 13.504 min Scan# 1788
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 65 Resp: 34348

Ion	Ratio	Lower	Upper
65	100		
92	67.1	57.1	85.7
138	109.2	97.0	145.4





#49
Acenaphthylene
Concen: 4.832 ng
RT: 14.133 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:152 Resp: 25940

Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.8	23.8
153	12.9	10.6	15.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#50
Dimethylphthalate
Concen: 4.941 ng
RT: 13.869 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:163 Resp: 208800

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.2
164	10.2	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 4.395 ng
RT: 13.992 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:165 Resp: 3845

Ion	Ratio	Lower	Upper
165	100		
63	48.5	35.1	52.7
89	51.9	38.4	57.6

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#52
Acenaphthene
Concen: 4.829 ng
RT: 14.474 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:154 Resp: 150042

Ion	Ratio	Lower	Upper
154	100		
153	112.3	89.0	133.4
152	55.2	43.0	64.6





#53
3-Nitroaniline
Concen: 4.036 ng
RT: 14.339 min Scan# 1
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:138 Resp: 3371

Ion	Ratio	Lower	Upper
138	100		
108	15.1	7.4	11.2
92	110.8	84.0	126.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#55
Dibenzofuran
Concen: 4.892 ng
RT: 14.816 min Scan# 2011
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:168 Resp: 252898

Ion	Ratio	Lower	Upper
168	100		
139	33.1	26.1	39.1
169	13.5	10.8	16.2





#57
2,4-Dinitrotoluene
Concen: 4.044 ng
RT: 14.798 min Scan# 2
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:165 Resp: 4890

Ion	Ratio	Lower	Upper
165	100		
63	46.6	29.0	43.6
89	69.6	49.4	74.2
182	2.5	2.2	3.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#58
Fluorene
Concen: 4.739 ng
RT: 15.463 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:166 Resp: 200542

Ion	Ratio	Lower	Upper
166	100		
165	101.8	76.6	115.0
167	13.4	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 4.463 ng
RT: 15.057 min Scan# 2(
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:232 Resp: 5289

Ion	Ratio	Lower	Upper
232	100		
131	38.0	28.2	42.4
130	1.6	1.5	2.3
166	28.6	22.6	34.0

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#60
Diethylphthalate
Concen: 4.981 ng
RT: 15.233 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:149 Resp: 198362

Ion	Ratio	Lower	Upper
149	100		
177	22.6	19.1	28.7
150	12.0	9.6	14.4



#61
4-Chlorophenyl-phenylether
Concen: 4.607 ng
RT: 15.457 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:204 Resp: 10708

Ion	Ratio	Lower	Upper
204	100		
206	31.7	26.3	39.5
141	48.5	34.0	51.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#62
4-Nitroaniline
Concen: 3.737 ng m
RT: 15.515 min Scan# 2130
Delta R.T. 0.023 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:138 Resp: 30138

Ion	Ratio	Lower	Upper
138	100		
92	57.3	25.9	65.9
108	84.6	52.4	92.4





#63
Azobenzene
Concen: 4.844 ng
RT: 15.751 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion: 77 Resp: 16166

Ion	Ratio	Lower	Upper
77	100		
182	31.6	10.5	50.5
105	21.7	1.1	41.1
51	24.8	6.3	46.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:188 Resp: 1183325

Ion	Ratio	Lower	Upper
188	100		
94	7.3	5.7	8.5
80	7.9	6.2	9.4





#66
n-Nitrosodiphenylamine
Concen: 4.723 ng
RT: 15.674 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:169 Resp: 17160

Ion	Ratio	Lower	Upper
169	100		
168	70.0	53.4	80.2
167	36.7	29.4	44.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#67
4-Bromophenyl-phenylether
Concen: 4.612 ng
RT: 16.351 min Scan# 2272
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:248 Resp: 66000

Ion	Ratio	Lower	Upper
248	100		
250	99.9	76.2	114.2
141	60.8	45.9	68.9





#68
Hexachlorobenzene
Concen: 4.759 ng
RT: 16.468 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:284 Resp: 7850

Ion	Ratio	Lower	Upper
284	100		
142	30.1	23.9	35.9
249	29.9	24.5	36.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#69
Atrazine
Concen: 4.433 ng
RT: 16.627 min Scan# 2319
Delta R.T. 0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:200 Resp: 58110

Ion	Ratio	Lower	Upper
200	100		
173	24.8	4.3	44.3
215	49.5	30.6	70.6





#71
Phenanthrene
Concen: 4.764 ng
RT: 17.198 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:178 Resp: 31795

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.3	22.9
179	14.8	12.4	18.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#72
Anthracene
Concen: 4.651 ng
RT: 17.292 min Scan# 2432
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:178 Resp: 314149

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.0	22.4
179	15.0	12.2	18.4





#73
Carbazole
Concen: 4.586 ng
RT: 17.574 min Scan# 24
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:167 Resp: 26877

Ion	Ratio	Lower	Upper
167	100		
166	21.0	17.2	25.8
139	11.7	9.0	13.6

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#74
Di-n-butylphthalate
Concen: 4.688 ng
RT: 18.127 min Scan# 2574
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:149 Resp: 327178

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.4	3.8	5.6





#75
Fluoranthene
Concen: 4.449 ng
RT: 19.221 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:202 Resp: 34859

Ion	Ratio	Lower	Upper
202	100		
101	8.5	0.0	28.0
203	17.6	0.0	37.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.403 min Scan# 3131
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:240 Resp: 1136162

Ion	Ratio	Lower	Upper
240	100		
120	8.2	6.5	9.7
236	25.6	20.9	31.3





#78
Pyrene
Concen: 4.593 ng
RT: 19.586 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:202 Resp: 36647

Ion	Ratio	Lower	Upper
202	100		
200	20.5	16.7	25.1
203	16.9	14.1	21.1



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#79
Terphenyl-d14
Concen: 8.704 ng
RT: 19.780 min Scan# 2855
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:244 Resp: 668289

Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.6	8.4
122	8.7	5.6	8.4#





#80
Butylbenzylphthalate
Concen: 4.646 ng
RT: 20.480 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:149 Resp: 13882

Ion	Ratio	Lower	Upper
149	100		
91	61.7	47.4	71.2
206	23.4	19.4	29.0

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#81
Benzo(a)anthracene
Concen: 4.648 ng
RT: 21.380 min Scan# 3127
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:228 Resp: 363641

Ion	Ratio	Lower	Upper
228	100		
226	26.9	22.1	33.1
229	20.1	15.7	23.5





#82
3,3'-Dichlorobenzidine
Concen: 4.119 ng
RT: 21.315 min Scan# 3
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:252 Resp: 12337

Ion	Ratio	Lower	Upper
252	100		
254	61.0	51.8	77.6
126	9.9	6.0	9.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#83
Chrysene
Concen: 4.738 ng
RT: 21.445 min Scan# 3138
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:228 Resp: 343085

Ion	Ratio	Lower	Upper
228	100		
226	29.1	24.2	36.2
229	18.7	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 4.620 ng
RT: 21.297 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:149 Resp: 20623

Ion	Ratio	Lower	Upper
149	100		
167	28.0	24.1	36.1
279	4.1	3.9	5.9

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#85
Di-n-octyl phthalate
Concen: 4.711 ng
RT: 22.427 min Scan# 3305
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:149 Resp: 346799

Ion	Ratio	Lower	Upper
149	100		
167	1.8	1.3	1.9
43	8.0	5.8	8.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.403 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:264 Resp: 114991

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.8	28.2
265	22.2	17.0	25.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#87
Indeno(1,2,3-cd)pyrene
Concen: 4.364 ng
RT: 27.803 min Scan# 4219
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:276 Resp: 373944

Ion	Ratio	Lower	Upper
276	100		
138	8.9	16.4	24.6#
277	25.9	19.8	29.8





#88
Benzo(b)fluoranthene
Concen: 4.338 ng
RT: 23.456 min Scan# 34
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:252 Resp: 32439

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	8.7	6.1	9.1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#89
Benzo(k)fluoranthene
Concen: 4.631 ng m
RT: 23.515 min Scan# 3490
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:252 Resp: 350374

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.4	26.2
125	9.5	5.8	8.8#





#90
Benzo(a)pyrene
Concen: 4.492 ng
RT: 24.262 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:252 Resp: 31460

Ion	Ratio	Lower	Upper
252	100		
253	21.0	17.4	26.0
125	9.7	6.6	9.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#91
Dibenzo(a,h)anthracene
Concen: 4.346 ng
RT: 27.856 min Scan# 4228
Delta R.T. 0.006 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:278 Resp: 303557

Ion	Ratio	Lower	Upper
278	100		
139	13.5	10.5	15.7
279	22.8	19.4	29.0





#92
Benzo(g,h,i)perylene
Concen: 4.585 ng
RT: 28.862 min Scan# 4
Delta R.T. 0.012 min
Lab File: BM050025.D
Acq: 28 Apr 2025 13:09

Tgt Ion:276 Resp: 30657

Ion	Ratio	Lower	Upper
276	100		
277	22.1	19.1	28.7
138	18.1	14.3	21.5



Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050026.D
Acq On : 28 Apr 2025 13:48
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Apr 28 17:54:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	250110	20.000	ng		0.00
21) Naphthalene-d8	10.563	136	926512	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	636652	20.000	ng		0.00
64) Phenanthrene-d10	17.157	188	1274608	20.000	ng		0.00
76) Chrysene-d12	21.397	240	1213581	20.000	ng		0.00
86) Perylene-d12	24.403	264	1224339	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.363	112	269870	18.894	ng		0.00
7) Phenol-d6	6.957	99	329701	18.366	ng		0.00
23) Nitrobenzene-d5	8.928	82	346334	18.420	ng		0.00
42) 2,4,6-Tribromophenol	15.910	330	158339	16.820	ng		0.00
45) 2-Fluorobiphenyl	13.027	172	989917	18.394	ng		0.00
79) Terphenyl-d14	19.780	244	1443528	17.601	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.257	88	58048	9.478	ng		98
3) Pyridine	3.663	79	145801	9.266	ng		98
4) n-Nitrosodimethylamine	3.569	42	57404	9.302	ng	#	93
6) Aniline	7.098	93	211256	9.319	ng		100
8) 2-Chlorophenol	7.339	128	145228	9.404	ng		99
9) Benzaldehyde	6.916	77	115030m	9.587	ng		
10) Phenol	6.981	94	168041	9.246	ng		99
11) bis(2-Chloroethyl)ether	7.192	93	138197	9.107	ng		100
12) 1,3-Dichlorobenzene	7.657	146	178565	9.572	ng		96
13) 1,4-Dichlorobenzene	7.804	146	177312	9.461	ng		99
14) 1,2-Dichlorobenzene	8.122	146	174224	9.624	ng		100
15) Benzyl Alcohol	8.022	79	110234	8.854	ng		98
16) 2,2'-oxybis(1-Chloropr...	8.292	45	177776	9.637	ng		97
17) 2-Methylphenol	8.228	107	111639	9.325	ng		99
18) Hexachloroethane	8.845	117	63487	9.530	ng		98
19) n-Nitroso-di-n-propyla...	8.575	70	110641	9.609	ng		98
20) 3+4-Methylphenols	8.563	107	146303	9.049	ng		99
22) Acetophenone	8.592	105	214790	9.325	ng		98
24) Nitrobenzene	8.969	77	150534	9.259	ng		99
25) Isophorone	9.498	82	300179	9.366	ng		99
26) 2-Nitrophenol	9.686	139	72521	8.732	ng		97
27) 2,4-Dimethylphenol	9.757	122	123406	8.960	ng		97
28) bis(2-Chloroethoxy)met...	9.975	93	184837	9.331	ng		99
29) 2,4-Dichlorophenol	10.245	162	137916	8.938	ng		99
30) 1,2,4-Trichlorobenzene	10.428	180	177721	9.453	ng		97
31) Naphthalene	10.610	128	444206	9.467	ng		99
32) Benzoic acid	9.875	122	63971m	7.282	ng		
33) 4-Chloroaniline	10.739	127	181907	9.458	ng		99
34) Hexachlorobutadiene	10.898	225	111307	9.327	ng		99
35) Caprolactam	11.516	113	39344	8.745	ng		94
36) 4-Chloro-3-methylphenol	11.880	107	131076	9.194	ng		98
37) 2-Methylnaphthalene	12.227	142	296053	9.411	ng		97
38) 1-Methylnaphthalene	12.451	142	312512	9.387	ng		97
40) 1,2,4,5-Tetrachloroben...	12.604	216	197943	8.981	ng		99
41) Hexachlorocyclopentadiene	12.580	237	96522	7.166	ng		99
43) 2,4,6-Trichlorophenol	12.857	196	121240	8.673	ng		98

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050026.D
Acq On : 28 Apr 2025 13:48
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Quant Time: Apr 28 17:54:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.945	196	132882	8.815	ng		96
46) 1,1'-Biphenyl	13.245	154	443960	9.381	ng		99
47) 2-Chloronaphthalene	13.286	162	354101	9.447	ng		99
48) 2-Nitroaniline	13.498	65	77101	8.764	ng		94
49) Acenaphthylene	14.133	152	545774	9.235	ng		100
50) Dimethylphthalate	13.868	163	438261	9.419	ng		99
51) 2,6-Dinitrotoluene	13.992	165	86242	8.953	ng		98
52) Acenaphthene	14.474	154	317735	9.288	ng		98
53) 3-Nitroaniline	14.333	138	78720	8.559	ng	#	99
54) 2,4-Dinitrophenol	14.563	184	31979	10.558	ng		96
55) Dibenzofuran	14.810	168	535023	9.401	ng		99
56) 4-Nitrophenol	14.710	139	40722	9.668	ng		96
57) 2,4-Dinitrotoluene	14.792	165	116476	8.750	ng		95
58) Fluorene	15.462	166	427384	9.174	ng		98
59) 2,3,4,6-Tetrachlorophenol	15.051	232	115185	8.828	ng		98
60) Diethylphthalate	15.233	149	414149	9.447	ng		100
61) 4-Chlorophenyl-phenyle...	15.457	204	227926	8.908	ng		94
62) 4-Nitroaniline	15.498	138	74565m	8.397	ng		
63) Azobenzene	15.745	77	347347	9.454	ng		99
65) 4,6-Dinitro-2-methylph...	15.562	198	56182	7.042	ng		93
66) n-Nitrosodiphenylamine	15.668	169	364601	9.315	ng		99
67) 4-Bromophenyl-phenylether	16.351	248	137532	8.923	ng		99
68) Hexachlorobenzene	16.468	284	160195	9.016	ng		99
69) Atrazine	16.621	200	126168	8.937	ng		98
70) Pentachlorophenol	16.827	266	74839	7.483	ng		95
71) Phenanthrene	17.198	178	659374	9.172	ng		100
72) Anthracene	17.292	178	661260	9.089	ng		99
73) Carbazole	17.568	167	573667	9.087	ng		99
74) Di-n-butylphthalate	18.121	149	689248	9.168	ng		99
75) Fluoranthene	19.221	202	727607	8.621	ng		99
77) Benzidine	19.415	184	339970	7.887	ng		99
78) Pyrene	19.586	202	761625	8.937	ng		99
80) Butylbenzylphthalate	20.474	149	290617	9.105	ng		93
81) Benzo(a)anthracene	21.380	228	747128	8.940	ng		100
82) 3,3'-Dichlorobenzidine	21.309	252	262017	8.190	ng		100
83) Chrysene	21.444	228	709862	9.178	ng		99
84) Bis(2-ethylhexyl)phtha...	21.297	149	439949	9.228	ng		98
85) Di-n-octyl phthalate	22.427	149	726208	9.236	ng		99
87) Indeno(1,2,3-cd)pyrene	27.791	276	791344	8.674	ng		99
88) Benzo(b)fluoranthene	23.450	252	703118	8.831	ng		98
89) Benzo(k)fluoranthene	23.515	252	704142	8.741	ng		99
90) Benzo(a)pyrene	24.262	252	652617	8.751	ng		99
91) Dibenzo(a,h)anthracene	27.844	278	638913	8.591	ng		98
92) Benzo(g,h,i)perylene	28.850	276	630261	8.852	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050026.D
Acq On : 28 Apr 2025 13:48
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Apr 28 17:54:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050026.D
Acq On : 28 Apr 2025 13:48
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Apr 28 17:54:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:152 Resp: 25011

Ion	Ratio	Lower	Upper
152	100		
150	152.8	124.1	186.1
115	53.1	41.2	61.8

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#2
1,4-Dioxane
Concen: 9.478 ng
RT: 3.257 min Scan# 46
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 88 Resp: 58048

Ion	Ratio	Lower	Upper
88	100		
58	70.3	55.1	82.7
43	24.7	20.9	31.3



#3
Pyridine
Concen: 9.266 ng
RT: 3.663 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 79 Resp: 14580

Ion	Ratio	Lower	Upper
79	100		
52	61.3	48.6	72.8
51	33.3	24.6	37.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#4
n-Nitrosodimethylamine
Concen: 9.302 ng
RT: 3.569 min Scan# 99
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 42 Resp: 57404

Ion	Ratio	Lower	Upper
42	100		
74	157.9	120.2	180.2
44	16.6	6.2	9.2#





#5
2-Fluorophenol
Concen: 18.894 ng
RT: 5.363 min Scan# 4
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:112 Resp: 26987

Ion	Ratio	Lower	Upper
112	100		
64	52.1	42.6	64.0
63	26.8	22.2	33.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#6
Aniline
Concen: 9.319 ng
RT: 7.098 min Scan# 699
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 93 Resp: 211256

Ion	Ratio	Lower	Upper
93	100		
66	37.3	29.9	44.9
65	18.6	14.5	21.7





#7
Phenol-d6
Concen: 18.366 ng
RT: 6.957 min Scan# 6
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 99 Resp: 32970

Ion	Ratio	Lower	Upper
99	100		
42	16.7	13.0	19.4
71	32.1	25.6	38.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#8
2-Chlorophenol
Concen: 9.404 ng
RT: 7.339 min Scan# 740
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:128 Resp: 145228

Ion	Ratio	Lower	Upper
128	100		
130	32.1	12.8	52.8
64	42.2	21.2	61.2





#9
Benzaldehyde
Concen: 9.587 ng m
RT: 6.916 min Scan# 6
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 77 Resp: 11503

Ion	Ratio	Lower	Upper
77	100		
105	101.6	80.5	120.5
106	100.5	76.1	116.1



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#10
Phenol
Concen: 9.246 ng
RT: 6.981 min Scan# 679
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 94 Resp: 168041

Ion	Ratio	Lower	Upper
94	100		
65	26.8	7.6	47.6
66	35.9	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 9.107 ng
RT: 7.192 min Scan# 7
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 93 Resp: 13819

Ion	Ratio	Lower	Upper
93	100		
63	72.0	52.3	92.3
95	32.3	12.8	52.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#12
1,3-Dichlorobenzene
Concen: 9.572 ng
RT: 7.657 min Scan# 794
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:146 Resp: 178565

Ion	Ratio	Lower	Upper
146	100		
148	61.8	52.5	78.7
75	27.8	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 9.461 ng
RT: 7.804 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:146 Resp: 17731

Ion	Ratio	Lower	Upper
146	100		
148	65.5	52.2	78.2
111	35.5	29.3	43.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#14
1,2-Dichlorobenzene
Concen: 9.624 ng
RT: 8.122 min Scan# 873
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:146 Resp: 174224

Ion	Ratio	Lower	Upper
146	100		
148	65.3	52.4	78.6
111	39.2	31.0	46.6





#15
Benzyl Alcohol
Concen: 8.854 ng
RT: 8.022 min Scan# 8
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 79 Resp: 11023

Ion	Ratio	Lower	Upper
79	100		
108	83.7	66.0	99.0
77	67.5	51.6	77.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 9.637 ng
RT: 8.292 min Scan# 902
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 45 Resp: 177776

Ion	Ratio	Lower	Upper
45	100		
77	17.1	0.0	36.4
79	14.1	0.0	32.7





#17
2-Methylphenol
Concen: 9.325 ng
RT: 8.228 min Scan# 8
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:107 Resp: 11163

Ion	Ratio	Lower	Upper
107	100		
108	115.1	91.1	136.7
77	46.6	36.2	54.4
79	44.5	35.8	53.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#18
Hexachloroethane
Concen: 9.530 ng
RT: 8.845 min Scan# 996
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:117 Resp: 63487

Ion	Ratio	Lower	Upper
117	100		
119	100.3	77.3	115.9
201	129.0	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 9.609 ng
RT: 8.575 min Scan# 9
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 70 Resp: 11064

Ion	Ratio	Lower	Upper
70	100		
42	51.9	40.5	60.7
101	8.8	7.8	11.6
130	22.8	17.4	26.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#20
3+4-Methylphenols
Concen: 9.049 ng
RT: 8.563 min Scan# 948
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:107 Resp: 146303

Ion	Ratio	Lower	Upper
107	100		
108	91.7	72.4	112.4
77	35.2	13.6	53.6
79	28.5	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:136 Resp: 92651

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.9	13.3
54	6.3	5.1	7.7
68	5.4	4.4	6.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#22
Acetophenone
Concen: 9.325 ng
RT: 8.592 min Scan# 953
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:105 Resp: 214790

Ion	Ratio	Lower	Upper
105	100		
71	2.3	1.7	2.5
51	25.1	18.6	27.8
120	23.3	18.3	27.5





#23
Nitrobenzene-d5
Concen: 18.420 ng
RT: 8.928 min Scan# 1(
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 82 Resp: 34633

Ion	Ratio	Lower	Upper
82	100		
128	45.5	35.7	53.5
54	43.9	35.4	53.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#24
Nitrobenzene
Concen: 9.259 ng
RT: 8.969 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 77 Resp: 150534

Ion	Ratio	Lower	Upper
77	100		
123	46.6	37.9	56.9
65	13.0	10.1	15.1





#25
Isophorone
Concen: 9.366 ng
RT: 9.498 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 82 Resp: 300179

Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.7	8.5
138	20.1	15.5	23.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#26
2-Nitrophenol
Concen: 8.732 ng
RT: 9.686 min Scan# 1139
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:139 Resp: 72521

Ion	Ratio	Lower	Upper
139	100		
109	29.6	21.9	32.9
65	45.3	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 8.960 ng
RT: 9.757 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:122 Resp: 12340

Ion	Ratio	Lower	Upper
122	100		
107	115.9	89.0	133.6
121	58.2	45.6	68.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#28
bis(2-Chloroethoxy)methane
Concen: 9.331 ng
RT: 9.975 min Scan# 1188
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 93 Resp: 184837

Ion	Ratio	Lower	Upper
93	100		
95	32.0	26.1	39.1
123	11.7	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 8.938 ng
RT: 10.245 min Scan# 1
Delta R.T. 0.018 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:162 Resp: 13791

Ion	Ratio	Lower	Upper
162	100		
164	64.6	45.4	85.4
98	30.7	11.3	51.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#30
1,2,4-Trichlorobenzene
Concen: 9.453 ng
RT: 10.428 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:180 Resp: 177721

Ion	Ratio	Lower	Upper
180	100		
182	93.2	77.0	115.4
145	25.3	20.6	30.8





#31
Naphthalene
Concen: 9.467 ng
RT: 10.610 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:128 Resp: 44420

Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.8	13.2
127	12.7	10.4	15.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#32
Benzoic acid
Concen: 7.282 ng m
RT: 9.875 min Scan# 1171
Delta R.T. -0.035 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:122 Resp: 63971

Ion	Ratio	Lower	Upper
122	100		
105	123.3	103.6	143.6
77	96.2	73.7	113.7





#33
4-Chloroaniline
Concen: 9.458 ng
RT: 10.739 min Scan# 1
Delta R.T. 0.012 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:127 Resp: 18190

Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.3	39.5
65	28.4	22.5	33.7
92	17.8	14.5	21.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#34
Hexachlorobutadiene
Concen: 9.327 ng
RT: 10.898 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:225 Resp: 111307

Ion	Ratio	Lower	Upper
225	100		
223	61.9	50.2	75.2
227	63.9	50.6	76.0





#35
Caprolactam
Concen: 8.745 ng
RT: 11.516 min Scan# 14
Delta R.T. -0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:113 Resp: 3934

Ion	Ratio	Lower	Upper
113	100		
55	150.2	122.9	162.9
56	115.3	88.1	128.1

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#36
4-Chloro-3-methylphenol
Concen: 9.194 ng
RT: 11.880 min Scan# 1512
Delta R.T. 0.012 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:107 Resp: 131076

Ion	Ratio	Lower	Upper
107	100		
144	29.1	23.8	35.6
142	91.9	71.9	107.9





#37
2-Methylnaphthalene
Concen: 9.411 ng
RT: 12.227 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:142 Resp: 29605

Ion	Ratio	Lower	Upper
142	100		
141	85.0	70.4	105.6
115	31.5	25.4	38.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#38
1-Methylnaphthalene
Concen: 9.387 ng
RT: 12.451 min Scan# 1609
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:142 Resp: 312512

Ion	Ratio	Lower	Upper
142	100		
141	93.3	72.1	108.1
116	3.7	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:164 Resp: 63665

Ion	Ratio	Lower	Upper
164	100		
162	97.3	79.4	119.0
160	43.3	36.2	54.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 8.981 ng
RT: 12.604 min Scan# 1635
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:216 Resp: 197943

Ion	Ratio	Lower	Upper
216	100		
214	77.7	62.3	93.5
179	20.0	15.4	23.0
108	15.3	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 7.166 ng
RT: 12.580 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:237 Resp: 9652

Ion	Ratio	Lower	Upper
237	100		
235	63.6	42.9	82.9
272	14.5	0.0	32.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#42
2,4,6-Tribromophenol
Concen: 16.820 ng
RT: 15.910 min Scan# 2197
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:330 Resp: 158339

Ion	Ratio	Lower	Upper
330	100		
332	97.5	77.3	115.9
141	31.1	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 8.673 ng
RT: 12.857 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:196 Resp: 12124

Ion	Ratio	Lower	Upper
196	100		
198	96.5	79.0	118.6
200	30.4	25.1	37.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#44
2,4,5-Trichlorophenol
Concen: 8.815 ng
RT: 12.945 min Scan# 1693
Delta R.T. 0.012 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:196 Resp: 132882

Ion	Ratio	Lower	Upper
196	100		
198	99.7	76.7	115.1
97	39.6	30.1	45.1
132	22.1	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 18.394 ng
RT: 13.027 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:172 Resp: 98991

Ion	Ratio	Lower	Upper
172	100		
171	34.4	27.8	41.6
170	22.8	18.6	28.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#46
1,1'-Biphenyl
Concen: 9.381 ng
RT: 13.245 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:154 Resp: 443960

Ion	Ratio	Lower	Upper
154	100		
153	39.0	20.0	60.0
76	10.4	0.0	30.3





#47
2-Chloronaphthalene
Concen: 9.447 ng
RT: 13.286 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:162 Resp: 35410

Ion	Ratio	Lower	Upper
162	100		
127	32.3	25.9	38.9
164	32.8	26.6	40.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#48
2-Nitroaniline
Concen: 8.764 ng
RT: 13.498 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 65 Resp: 77101

Ion	Ratio	Lower	Upper
65	100		
92	73.1	57.1	85.7
138	112.5	97.0	145.4





#49
Acenaphthylene
Concen: 9.235 ng
RT: 14.133 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:152 Resp: 54577

Ion	Ratio	Lower	Upper
152	100		
151	20.0	15.8	23.8
153	13.2	10.6	15.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#50
Dimethylphthalate
Concen: 9.419 ng
RT: 13.868 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:163 Resp: 438261

Ion	Ratio	Lower	Upper
163	100		
194	4.6	3.4	5.2
164	10.5	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 8.953 ng
RT: 13.992 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:165 Resp: 8624

Ion	Ratio	Lower	Upper
165	100		
63	45.5	35.1	52.7
89	48.7	38.4	57.6

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#52
Acenaphthene
Concen: 9.288 ng
RT: 14.474 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:154 Resp: 317735

Ion	Ratio	Lower	Upper
154	100		
153	108.5	89.0	133.4
152	54.5	43.0	64.6





#53
3-Nitroaniline
Concen: 8.559 ng
RT: 14.333 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:138 Resp: 7872

Ion	Ratio	Lower	Upper
138	100		
108	11.2	7.4	11.2
92	104.0	84.0	126.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#54
2,4-Dinitrophenol
Concen: 10.558 ng
RT: 14.563 min Scan# 1968
Delta R.T. 0.018 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:184 Resp: 31979

Ion	Ratio	Lower	Upper
184	100		
63	52.1	39.8	59.6
154	60.9	46.1	69.1





#55
Dibenzofuran
Concen: 9.401 ng
RT: 14.810 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:168 Resp: 53502

Ion	Ratio	Lower	Upper
168	100		
139	33.0	26.1	39.1
169	13.4	10.8	16.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#56
4-Nitrophenol
Concen: 9.668 ng
RT: 14.710 min Scan# 1993
Delta R.T. 0.041 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:139 Resp: 40722

Ion	Ratio	Lower	Upper
139	100		
109	70.7	47.0	87.0
65	87.5	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 8.750 ng
RT: 14.792 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:165 Resp: 11647

Ion	Ratio	Lower	Upper
165	100		
63	40.1	29.0	43.6
89	64.7	49.4	74.2
182	2.5	2.2	3.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#58
Fluorene
Concen: 9.174 ng
RT: 15.462 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:166 Resp: 427384

Ion	Ratio	Lower	Upper
166	100		
165	98.0	76.6	115.0
167	13.7	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 8.828 ng
RT: 15.051 min Scan# 2(
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:232 Resp: 11518

Ion	Ratio	Lower	Upper
232	100		
131	36.2	28.2	42.4
130	2.0	1.5	2.3
166	27.1	22.6	34.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#60
Diethylphthalate
Concen: 9.447 ng
RT: 15.233 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:149 Resp: 414149

Ion	Ratio	Lower	Upper
149	100		
177	23.9	19.1	28.7
150	11.9	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 8.908 ng
RT: 15.457 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:204 Resp: 22792

Ion	Ratio	Lower	Upper
204	100		
206	31.4	26.3	39.5
141	47.6	34.0	51.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#62
4-Nitroaniline
Concen: 8.397 ng m
RT: 15.498 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:138 Resp: 74565

Ion	Ratio	Lower	Upper
138	100		
92	51.0	25.9	65.9
108	80.0	52.4	92.4





#63
Azobenzene
Concen: 9.454 ng
RT: 15.745 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion: 77 Resp: 34734

Ion	Ratio	Lower	Upper
77	100		
182	31.0	10.5	50.5
105	21.0	1.1	41.1
51	26.8	6.3	46.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:188 Resp: 1274608

Ion	Ratio	Lower	Upper
188	100		
94	7.2	5.7	8.5
80	7.8	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 7.042 ng
RT: 15.562 min Scan# 2
Delta R.T. 0.005 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:198 Resp: 5618

Ion	Ratio	Lower	Upper
198	100		
51	34.7	10.4	50.4
105	38.2	14.6	54.6

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#66
n-Nitrosodiphenylamine
Concen: 9.315 ng
RT: 15.668 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:169 Resp: 364601

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.4	80.2
167	35.8	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 8.923 ng
RT: 16.351 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:248 Resp: 13753

Ion	Ratio	Lower	Upper
248	100		
250	96.1	76.2	114.2
141	57.1	45.9	68.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#68
Hexachlorobenzene
Concen: 9.016 ng
RT: 16.468 min Scan# 2292
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:284 Resp: 160195

Ion	Ratio	Lower	Upper
284	100		
142	30.1	23.9	35.9
249	30.0	24.5	36.7





#69
Atrazine
Concen: 8.937 ng
RT: 16.621 min Scan# 2
Delta R.T. -0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:200 Resp: 12616

Ion	Ratio	Lower	Upper
200	100		
173	26.3	4.3	44.3
215	49.6	30.6	70.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#70
Pentachlorophenol
Concen: 7.483 ng
RT: 16.827 min Scan# 2353
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:266 Resp: 74839

Ion	Ratio	Lower	Upper
266	100		
268	67.5	51.2	76.8
264	66.3	49.6	74.4





#71
Phenanthrene
Concen: 9.172 ng
RT: 17.198 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:178 Resp: 65937

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.3	12.4	18.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#72
Anthracene
Concen: 9.089 ng
RT: 17.292 min Scan# 2432
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:178 Resp: 661260

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	15.7	12.2	18.4





#73
Carbazole
Concen: 9.087 ng
RT: 17.568 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:167 Resp: 57366

Ion	Ratio	Lower	Upper
167	100		
166	20.9	17.2	25.8
139	11.2	9.0	13.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#74
Di-n-butylphthalate
Concen: 9.168 ng
RT: 18.121 min Scan# 2573
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:149 Resp: 689248

Ion	Ratio	Lower	Upper
149	100		
150	9.6	7.3	10.9
104	4.7	3.8	5.6





#75
Fluoranthene
Concen: 8.621 ng
RT: 19.221 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:202 Resp: 72760

Ion	Ratio	Lower	Upper
202	100		
101	8.7	0.0	28.0
203	18.0	0.0	37.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.397 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:240 Resp: 1213581

Ion	Ratio	Lower	Upper
240	100		
120	8.3	6.5	9.7
236	25.7	20.9	31.3





#77
Benzidine
Concen: 7.887 ng
RT: 19.415 min Scan# 2
Delta R.T. 0.011 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:184 Resp: 33997

Ion	Ratio	Lower	Upper
184	100		
185	13.7	11.5	17.3
183	11.0	8.9	13.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#78
Pyrene
Concen: 8.937 ng
RT: 19.586 min Scan# 2822
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:202 Resp: 761625

Ion	Ratio	Lower	Upper
202	100		
200	21.2	16.7	25.1
203	17.3	14.1	21.1





#79
Terphenyl-d14
Concen: 17.601 ng
RT: 19.780 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:244 Resp: 144352

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.6	8.4
122	8.5	5.6	8.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#80
Butylbenzylphthalate
Concen: 9.105 ng
RT: 20.474 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:149 Resp: 290617

Ion	Ratio	Lower	Upper
149	100		
91	64.9	47.4	71.2
206	21.3	19.4	29.0





#81
Benzo(a)anthracene
Concen: 8.940 ng
RT: 21.380 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:228 Resp: 74712

Ion	Ratio	Lower	Upper
228	100		
226	27.4	22.1	33.1
229	19.6	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#82
3,3'-Dichlorobenzidine
Concen: 8.190 ng
RT: 21.309 min Scan# 3115
Delta R.T. 0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:252 Resp: 262017

Ion	Ratio	Lower	Upper
252	100		
254	64.6	51.8	77.6
126	8.8	6.0	9.0





#83
Chrysene
Concen: 9.178 ng
RT: 21.444 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:228 Resp: 70986

Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.2	36.2
229	19.6	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 9.228 ng
RT: 21.297 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:149 Resp: 439949

Ion	Ratio	Lower	Upper
149	100		
167	29.3	24.1	36.1
279	4.2	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 9.236 ng
RT: 22.427 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:149 Resp: 72620

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.7	5.8	8.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.403 min Scan# 3641
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:264 Resp: 1224339

Ion	Ratio	Lower	Upper
264	100		
260	23.5	18.8	28.2
265	21.7	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 8.674 ng
RT: 27.791 min Scan# 4
Delta R.T. -0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:276 Resp: 79134

Ion	Ratio	Lower	Upper
276	100		
138	21.4	16.4	24.6
277	25.3	19.8	29.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#88
Benzo(b)fluoranthene
Concen: 8.831 ng
RT: 23.450 min Scan# 3479
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:252 Resp: 703118

Ion	Ratio	Lower	Upper
252	100		
253	21.2	17.7	26.5
125	8.8	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 8.741 ng
RT: 23.515 min Scan# 34
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:252 Resp: 70414

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.2
125	8.2	5.8	8.8

Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#90
Benzo(a)pyrene
Concen: 8.751 ng
RT: 24.262 min Scan# 3617
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:252 Resp: 652617

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	9.4	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 8.591 ng
RT: 27.844 min Scan# 4
Delta R.T. -0.006 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:278 Resp: 63891

Ion	Ratio	Lower	Upper
278	100		
139	13.2	10.5	15.7
279	25.8	19.4	29.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC010

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#92
Benzo(g,h,i)perylene
Concen: 8.852 ng
RT: 28.850 min Scan# 4397
Delta R.T. -0.000 min
Lab File: BM050026.D
Acq: 28 Apr 2025 13:48

Tgt Ion:276 Resp: 630261

Ion	Ratio	Lower	Upper
276	100		
277	22.8	19.1	28.7
138	17.7	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050027.D
Acq On : 28 Apr 2025 14:27
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Quant Time: Apr 28 17:55:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	242715	20.000	ng		0.00
21) Naphthalene-d8	10.563	136	896366	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	614688	20.000	ng		0.00
64) Phenanthrene-d10	17.157	188	1244440	20.000	ng		0.00
76) Chrysene-d12	21.398	240	1204355	20.000	ng		0.00
86) Perylene-d12	24.403	264	1191041	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.357	112	537488	38.777	ng		0.00
7) Phenol-d6	6.951	99	678404	38.941	ng		0.00
23) Nitrobenzene-d5	8.928	82	712171	39.151	ng		0.00
42) 2,4,6-Tribromophenol	15.904	330	345020	37.960	ng		0.00
45) 2-Fluorobiphenyl	13.027	172	2005788	38.602	ng		0.00
79) Terphenyl-d14	19.780	244	3208242	39.418	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.258	88	115944	19.509	ng		99
3) Pyridine	3.658	79	298256	19.532	ng		98
4) n-Nitrosodimethylamine	3.569	42	115296	19.251	ng	#	100
6) Aniline	7.098	93	429222	19.511	ng		99
8) 2-Chlorophenol	7.340	128	290876	19.409	ng		99
9) Benzaldehyde	6.910	77	235594	20.233	ng		98
10) Phenol	6.981	94	344567	19.537	ng		98
11) bis(2-Chloroethyl)ether	7.193	93	285330	19.375	ng		99
12) 1,3-Dichlorobenzene	7.657	146	352660	19.480	ng		99
13) 1,4-Dichlorobenzene	7.804	146	354975	19.518	ng		99
14) 1,2-Dichlorobenzene	8.122	146	345385	19.660	ng		99
15) Benzyl Alcohol	8.016	79	234541	19.413	ng		98
16) 2,2'-oxybis(1-Chloropr...	8.293	45	350829	19.598	ng		98
17) 2-Methylphenol	8.228	107	225978	19.450	ng		99
18) Hexachloroethane	8.845	117	126530	19.571	ng		99
19) n-Nitroso-di-n-propyla...	8.575	70	223538	20.006	ng		99
20) 3+4-Methylphenols	8.557	107	307585	19.604	ng		97
22) Acetophenone	8.592	105	434539	19.499	ng		98
24) Nitrobenzene	8.969	77	309341	19.667	ng		99
25) Isophorone	9.498	82	609252	19.648	ng		99
26) 2-Nitrophenol	9.681	139	154364	19.211	ng		97
27) 2,4-Dimethylphenol	9.751	122	259319	19.460	ng		99
28) bis(2-Chloroethoxy)met...	9.975	93	378162	19.732	ng		100
29) 2,4-Dichlorophenol	10.234	162	292336	19.582	ng		99
30) 1,2,4-Trichlorobenzene	10.428	180	352792	19.397	ng		98
31) Naphthalene	10.610	128	890530	19.618	ng		99
32) Benzoic acid	9.881	122	154412m	18.168	ng		
33) 4-Chloroaniline	10.734	127	362366	19.475	ng		99
34) Hexachlorobutadiene	10.898	225	223143	19.327	ng		99
35) Caprolactam	11.516	113	84917	19.509	ng		93
36) 4-Chloro-3-methylphenol	11.875	107	270510	19.613	ng		99
37) 2-Methylnaphthalene	12.228	142	593403	19.498	ng		99
38) 1-Methylnaphthalene	12.445	142	634293	19.694	ng		99
40) 1,2,4,5-Tetrachloroben...	12.604	216	402355	18.909	ng		99
41) Hexachlorocyclopentadiene	12.575	237	222254	17.091	ng		100
43) 2,4,6-Trichlorophenol	12.851	196	260161	19.276	ng		96

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050027.D
Acq On : 28 Apr 2025 14:27
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Quant Time: Apr 28 17:55:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.939	196	278307	19.123	ng	98
46) 1,1'-Biphenyl	13.239	154	892394	19.530	ng	99
47) 2-Chloronaphthalene	13.286	162	709815	19.614	ng	100
48) 2-Nitroaniline	13.498	65	165324	19.465	ng	99
49) Acenaphthylene	14.133	152	1112766	19.502	ng	99
50) Dimethylphthalate	13.869	163	883120	19.659	ng	99
51) 2,6-Dinitrotoluene	13.986	165	183493	19.729	ng	99
52) Acenaphthene	14.474	154	643177	19.474	ng	99
53) 3-Nitroaniline	14.327	138	174626	19.666	ng	97
54) 2,4-Dinitrophenol	14.551	184	80991	18.373	ng	99
55) Dibenzofuran	14.810	168	1074975	19.562	ng	99
56) 4-Nitrophenol	14.686	139	110948	19.021	ng	96
57) 2,4-Dinitrotoluene	14.786	165	252994	19.685	ng	100
58) Fluorene	15.457	166	879470	19.553	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.051	232	244928	19.442	ng	99
60) Diethylphthalate	15.233	149	848021	20.034	ng	99
61) 4-Chlorophenyl-phenyle...	15.451	204	473276	19.157	ng	95
62) 4-Nitroaniline	15.492	138	170438	19.880	ng	95
63) Azobenzene	15.745	77	707026	19.930	ng	99
65) 4,6-Dinitro-2-methylph...	15.557	198	136910	17.576	ng	94
66) n-Nitrosodiphenylamine	15.669	169	741920	19.415	ng	99
67) 4-Bromophenyl-phenylether	16.345	248	285352	18.962	ng	98
68) Hexachlorobenzene	16.468	284	327912	18.903	ng	98
69) Atrazine	16.621	200	266068	19.303	ng	98
70) Pentachlorophenol	16.821	266	175794	18.004	ng	99
71) Phenanthrene	17.198	178	1348190	19.208	ng	100
72) Anthracene	17.292	178	1364462	19.210	ng	100
73) Carbazole	17.568	167	1194316	19.377	ng	99
74) Di-n-butylphthalate	18.121	149	1432688	19.519	ng	99
75) Fluoranthene	19.221	202	1555160	18.872	ng	99
77) Benzidine	19.409	184	777502	18.175	ng	99
78) Pyrene	19.580	202	1603874	18.964	ng	100
80) Butylbenzylphthalate	20.474	149	620751	19.597	ng	94
81) Benzo(a)anthracene	21.380	228	1591290	19.187	ng	100
82) 3,3'-Dichlorobenzidine	21.303	252	582038	18.333	ng	99
83) Chrysene	21.445	228	1461777	19.044	ng	99
84) Bis(2-ethylhexyl)phtha...	21.298	149	933505	19.730	ng	98
85) Di-n-octyl phthalate	22.427	149	1525599	19.552	ng	100
87) Indeno(1,2,3-cd)pyrene	27.791	276	1671967	18.840	ng	99
88) Benzo(b)fluoranthene	23.450	252	1439776	18.588	ng	99
89) Benzo(k)fluoranthene	23.515	252	1489410	19.005	ng	100
90) Benzo(a)pyrene	24.262	252	1365706	18.825	ng	99
91) Dibenzo(a,h)anthracene	27.838	278	1360269	18.803	ng	99
92) Benzo(g,h,i)perylene	28.850	276	1316513	19.008	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050027.D
Acq On : 28 Apr 2025 14:27
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 28 17:55:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050027.D
Acq On : 28 Apr 2025 14:27
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Apr 28 17:55:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 8
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:152 Resp: 24271

Ion	Ratio	Lower	Upper
152	100		
150	155.2	124.1	186.1
115	51.1	41.2	61.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#2
1,4-Dioxane
Concen: 19.509 ng
RT: 3.258 min Scan# 46
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 88 Resp: 115944

Ion	Ratio	Lower	Upper
88	100		
58	67.9	55.1	82.7
43	26.7	20.9	31.3





#3
Pyridine
Concen: 19.532 ng
RT: 3.658 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 79 Resp: 29825

Ion	Ratio	Lower	Upper
79	100		
52	61.8	48.6	72.8
51	32.1	24.6	37.0

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#4
n-Nitrosodimethylamine
Concen: 19.251 ng
RT: 3.569 min Scan# 99
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 42 Resp: 115296

Ion	Ratio	Lower	Upper
42	100		
74	150.3	120.2	180.2
44	10.3	6.2	9.2#





#5
2-Fluorophenol
Concen: 38.777 ng
RT: 5.357 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:112 Resp: 53748

Ion	Ratio	Lower	Upper
112	100		
64	54.3	42.6	64.0
63	28.2	22.2	33.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#6
Aniline
Concen: 19.511 ng
RT: 7.098 min Scan# 699
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 93 Resp: 429222

Ion	Ratio	Lower	Upper
93	100		
66	37.5	29.9	44.9
65	18.7	14.5	21.7





#7
Phenol-d6
Concen: 38.941 ng
RT: 6.951 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 99 Resp: 67840

Ion	Ratio	Lower	Upper
99	100		
42	15.9	13.0	19.4
71	32.6	25.6	38.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#8
2-Chlorophenol
Concen: 19.409 ng
RT: 7.340 min Scan# 740
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:128 Resp: 290876

Ion	Ratio	Lower	Upper
128	100		
130	31.9	12.8	52.8
64	41.9	21.2	61.2





#9
Benzaldehyde
Concen: 20.233 ng
RT: 6.910 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 77 Resp: 23559

Ion	Ratio	Lower	Upper
77	100		
105	97.9	80.5	120.5
106	98.3	76.1	116.1



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#10
Phenol
Concen: 19.537 ng
RT: 6.981 min Scan# 679
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 94 Resp: 344567

Ion	Ratio	Lower	Upper
94	100		
65	27.0	7.6	47.6
66	34.4	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 19.375 ng
RT: 7.193 min Scan# 7
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 93 Resp: 28533

Ion	Ratio	Lower	Upper
93	100		
63	71.9	52.3	92.3
95	32.4	12.8	52.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#12
1,3-Dichlorobenzene
Concen: 19.480 ng
RT: 7.657 min Scan# 794
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:146 Resp: 352660

Ion	Ratio	Lower	Upper
146	100		
148	64.7	52.5	78.7
75	27.8	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 19.518 ng
RT: 7.804 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:146 Resp: 35497

Ion	Ratio	Lower	Upper
146	100		
148	64.4	52.2	78.2
111	37.2	29.3	43.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#14
1,2-Dichlorobenzene
Concen: 19.660 ng
RT: 8.122 min Scan# 873
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:146 Resp: 345385

Ion	Ratio	Lower	Upper
146	100		
148	64.8	52.4	78.6
111	39.4	31.0	46.6





#15
Benzyl Alcohol
Concen: 19.413 ng
RT: 8.016 min Scan# 8
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 79 Resp: 23454

Ion	Ratio	Lower	Upper
79	100		
108	79.9	66.0	99.0
77	64.6	51.6	77.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 19.598 ng
RT: 8.293 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 45 Resp: 350829

Ion	Ratio	Lower	Upper
45	100		
77	17.8	0.0	36.4
79	13.0	0.0	32.7





#17
2-Methylphenol
Concen: 19.450 ng
RT: 8.228 min Scan# 8
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:107 Resp: 22597

Ion	Ratio	Lower	Upper
107	100		
108	113.4	91.1	136.7
77	45.7	36.2	54.4
79	44.0	35.8	53.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#18
Hexachloroethane
Concen: 19.571 ng
RT: 8.845 min Scan# 996
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:117 Resp: 126530

Ion	Ratio	Lower	Upper
117	100		
119	95.9	77.3	115.9
201	127.6	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 20.006 ng
RT: 8.575 min Scan# 9
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 70 Resp: 22353

Ion	Ratio	Lower	Upper
70	100		
42	49.5	40.5	60.7
101	9.5	7.8	11.6
130	22.3	17.4	26.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#20
3+4-Methylphenols
Concen: 19.604 ng
RT: 8.557 min Scan# 947
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:107 Resp: 307585

Ion	Ratio	Lower	Upper
107	100		
108	89.1	72.4	112.4
77	33.1	13.6	53.6
79	28.3	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:136 Resp: 89636

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	6.6	5.1	7.7
68	5.4	4.4	6.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#22
Acetophenone
Concen: 19.499 ng
RT: 8.592 min Scan# 953
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:105 Resp: 434539

Ion	Ratio	Lower	Upper
105	100		
71	1.9	1.7	2.5
51	25.1	18.6	27.8
120	23.2	18.3	27.5





#23
Nitrobenzene-d5
Concen: 39.151 ng
RT: 8.928 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 82 Resp: 71217

Ion	Ratio	Lower	Upper
82	100		
128	45.2	35.7	53.5
54	44.8	35.4	53.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#24
Nitrobenzene
Concen: 19.667 ng
RT: 8.969 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 77 Resp: 309341

Ion	Ratio	Lower	Upper
77	100		
123	48.2	37.9	56.9
65	13.1	10.1	15.1





#25
Isophorone
Concen: 19.648 ng
RT: 9.498 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 82 Resp: 60925

Ion	Ratio	Lower	Upper
82	100		
95	6.7	5.7	8.5
138	20.1	15.5	23.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#26
2-Nitrophenol
Concen: 19.211 ng
RT: 9.681 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:139 Resp: 154364

Ion	Ratio	Lower	Upper
139	100		
109	28.2	21.9	32.9
65	47.3	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 19.460 ng
RT: 9.751 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:122 Resp: 25931

Ion	Ratio	Lower	Upper
122	100		
107	113.1	89.0	133.6
121	56.7	45.6	68.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#28
bis(2-Chloroethoxy)methane
Concen: 19.732 ng
RT: 9.975 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 93 Resp: 378162

Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.1	39.1
123	11.6	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 19.582 ng
RT: 10.234 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:162 Resp: 29233

Ion	Ratio	Lower	Upper
162	100		
164	65.2	45.4	85.4
98	30.5	11.3	51.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#30
1,2,4-Trichlorobenzene
Concen: 19.397 ng
RT: 10.428 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:180 Resp: 352792

Ion	Ratio	Lower	Upper
180	100		
182	97.8	77.0	115.4
145	26.6	20.6	30.8





#31
Naphthalene
Concen: 19.618 ng
RT: 10.610 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:128 Resp: 89053

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	12.6	10.4	15.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#32
Benzoic acid
Concen: 18.168 ng m
RT: 9.881 min Scan# 1172
Delta R.T. -0.029 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:122 Resp: 154412

Ion	Ratio	Lower	Upper
122	100		
105	120.1	103.6	143.6
77	89.7	73.7	113.7





#33
4-Chloroaniline
Concen: 19.475 ng
RT: 10.734 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:127 Resp: 36236

Ion	Ratio	Lower	Upper
127	100		
129	31.7	26.3	39.5
65	28.0	22.5	33.7
92	18.4	14.5	21.7

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#34
Hexachlorobutadiene
Concen: 19.327 ng
RT: 10.898 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:225 Resp: 223143

Ion	Ratio	Lower	Upper
225	100		
223	62.6	50.2	75.2
227	62.4	50.6	76.0



#35
Caprolactam
Concen: 19.509 ng
RT: 11.516 min Scan# 14
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:113 Resp: 8491

Ion	Ratio	Lower	Upper
113	100		
55	154.8	122.9	162.9
56	112.6	88.1	128.1



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#36
4-Chloro-3-methylphenol
Concen: 19.613 ng
RT: 11.875 min Scan# 1511
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:107 Resp: 270510

Ion	Ratio	Lower	Upper
107	100		
144	28.8	23.8	35.6
142	89.0	71.9	107.9





#37
2-Methylnaphthalene
Concen: 19.498 ng
RT: 12.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:142 Resp: 59340

Ion	Ratio	Lower	Upper
142	100		
141	88.7	70.4	105.6
115	31.8	25.4	38.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#38
1-Methylnaphthalene
Concen: 19.694 ng
RT: 12.445 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:142 Resp: 634293

Ion	Ratio	Lower	Upper
142	100		
141	91.1	72.1	108.1
116	3.8	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:164 Resp: 61468

Ion	Ratio	Lower	Upper
164	100		
162	100.4	79.4	119.0
160	45.8	36.2	54.4



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 18.909 ng
RT: 12.604 min Scan# 1635
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:216 Resp: 402355

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.3	93.5
179	19.6	15.4	23.0
108	15.0	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 17.091 ng
RT: 12.575 min Scan# 1
Delta R.T. -0.005 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:237 Resp: 22225

Ion	Ratio	Lower	Upper
237	100		
235	63.1	42.9	82.9
272	12.9	0.0	32.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#42
2,4,6-Tribromophenol
Concen: 37.960 ng
RT: 15.904 min Scan# 2196
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:330 Resp: 345020

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.3	115.9
141	30.1	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 19.276 ng
RT: 12.851 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:196 Resp: 26016

Ion	Ratio	Lower	Upper
196	100		
198	94.8	79.0	118.6
200	30.0	25.1	37.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#44
2,4,5-Trichlorophenol
Concen: 19.123 ng
RT: 12.939 min Scan# 1692
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:196 Resp: 278307

Ion	Ratio	Lower	Upper
196	100		
198	97.2	76.7	115.1
97	40.1	30.1	45.1
132	22.1	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 38.602 ng
RT: 13.027 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:172 Resp: 200578

Ion	Ratio	Lower	Upper
172	100		
171	34.6	27.8	41.6
170	23.2	18.6	28.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#46
1,1'-Biphenyl
Concen: 19.530 ng
RT: 13.239 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:154 Resp: 892394

Ion	Ratio	Lower	Upper
154	100		
153	39.5	20.0	60.0
76	11.3	0.0	30.3





#47
2-Chloronaphthalene
Concen: 19.614 ng
RT: 13.286 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:162 Resp: 70981

Ion	Ratio	Lower	Upper
162	100		
127	32.0	25.9	38.9
164	33.3	26.6	40.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#48
2-Nitroaniline
Concen: 19.465 ng
RT: 13.498 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 65 Resp: 165324

Ion	Ratio	Lower	Upper
65	100		
92	73.4	57.1	85.7
138	120.3	97.0	145.4





#49
Acenaphthylene
Concen: 19.502 ng
RT: 14.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:152 Resp: 111276

Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.8	23.8
153	13.3	10.6	15.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#50
Dimethylphthalate
Concen: 19.659 ng
RT: 13.869 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:163 Resp: 883120

Ion	Ratio	Lower	Upper
163	100		
194	4.5	3.4	5.2
164	9.8	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 19.729 ng
RT: 13.986 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:165 Resp: 18349

Ion	Ratio	Lower	Upper
165	100		
63	45.7	35.1	52.7
89	48.1	38.4	57.6

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#52
Acenaphthene
Concen: 19.474 ng
RT: 14.474 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:154 Resp: 643177

Ion	Ratio	Lower	Upper
154	100		
153	112.0	89.0	133.4
152	53.3	43.0	64.6





#53
3-Nitroaniline
Concen: 19.666 ng
RT: 14.327 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:138 Resp: 17462

Ion	Ratio	Lower	Upper
138	100		
108	10.3	7.4	11.2
92	107.9	84.0	126.0

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#54
2,4-Dinitrophenol
Concen: 18.373 ng
RT: 14.551 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:184 Resp: 80991

Ion	Ratio	Lower	Upper
184	100		
63	50.1	39.8	59.6
154	58.5	46.1	69.1





#55
Dibenzofuran
Concen: 19.562 ng
RT: 14.810 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:168 Resp: 107497

Ion	Ratio	Lower	Upper
168	100		
139	33.2	26.1	39.1
169	13.3	10.8	16.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#56
4-Nitrophenol
Concen: 19.021 ng
RT: 14.686 min Scan# 1989
Delta R.T. 0.017 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:139 Resp: 110948

Ion	Ratio	Lower	Upper
139	100		
109	69.3	47.0	87.0
65	88.3	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 19.685 ng
RT: 14.786 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:165 Resp: 25299

Ion	Ratio	Lower	Upper
165	100		
63	36.9	29.0	43.6
89	61.9	49.4	74.2
182	2.8	2.2	3.2



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#58
Fluorene
Concen: 19.553 ng
RT: 15.457 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:166 Resp: 879470

Ion	Ratio	Lower	Upper
166	100		
165	97.0	76.6	115.0
167	13.7	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 19.442 ng
RT: 15.051 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:232 Resp: 24492

Ion	Ratio	Lower	Upper
232	100		
131	35.9	28.2	42.4
130	1.7	1.5	2.3
166	27.7	22.6	34.0

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#60
Diethylphthalate
Concen: 20.034 ng
RT: 15.233 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:149 Resp: 848021

Ion	Ratio	Lower	Upper
149	100		
177	23.5	19.1	28.7
150	12.1	9.6	14.4



#61
4-Chlorophenyl-phenylether
Concen: 19.157 ng
RT: 15.451 min Scan# 2
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:204 Resp: 47327

Ion	Ratio	Lower	Upper
204	100		
206	32.3	26.3	39.5
141	47.7	34.0	51.0

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#62
4-Nitroaniline
Concen: 19.880 ng
RT: 15.492 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:138 Resp: 170438

Ion	Ratio	Lower	Upper
138	100		
92	50.0	25.9	65.9
108	76.3	52.4	92.4



#63
Azobenzene
Concen: 19.930 ng
RT: 15.745 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion: 77 Resp: 70702

Ion	Ratio	Lower	Upper
77	100		
182	30.2	10.5	50.5
105	21.8	1.1	41.1
51	26.4	6.3	46.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:188 Resp: 1244440

Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.7	8.5
80	7.6	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 17.576 ng
RT: 15.557 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:198 Resp: 13691

Ion	Ratio	Lower	Upper
198	100		
51	34.7	10.4	50.4
105	37.2	14.6	54.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#66
n-Nitrosodiphenylamine
Concen: 19.415 ng
RT: 15.669 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:169 Resp: 741920

Ion	Ratio	Lower	Upper
169	100		
168	68.2	53.4	80.2
167	37.4	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 18.962 ng
RT: 16.345 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:248 Resp: 28535

Ion	Ratio	Lower	Upper
248	100		
250	96.3	76.2	114.2
141	60.0	45.9	68.9



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#68
Hexachlorobenzene
Concen: 18.903 ng
RT: 16.468 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:284 Resp: 327912

Ion	Ratio	Lower	Upper
284	100		
142	29.1	23.9	35.9
249	29.3	24.5	36.7





#69
Atrazine
Concen: 19.303 ng
RT: 16.621 min Scan# 2
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:200 Resp: 26606

Ion	Ratio	Lower	Upper
200	100		
173	24.3	4.3	44.3
215	48.7	30.6	70.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#70
Pentachlorophenol
Concen: 18.004 ng
RT: 16.821 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:266 Resp: 175794

Ion	Ratio	Lower	Upper
266	100		
268	63.0	51.2	76.8
264	60.9	49.6	74.4





#71
Phenanthrene
Concen: 19.208 ng
RT: 17.198 min Scan# 24
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:178 Resp: 134819

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.1	12.4	18.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#72
Anthracene
Concen: 19.210 ng
RT: 17.292 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:178 Resp: 1364462

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	15.3	12.2	18.4





#73
Carbazole
Concen: 19.377 ng
RT: 17.568 min Scan# 24
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:167 Resp: 119431

Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.2	25.8
139	11.4	9.0	13.6



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#74
Di-n-butylphthalate
Concen: 19.519 ng
RT: 18.121 min Scan# 2573
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:149 Resp: 1432688

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.3	10.9
104	4.6	3.8	5.6





#75
Fluoranthene
Concen: 18.872 ng
RT: 19.221 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:202 Resp: 155516

Ion	Ratio	Lower	Upper
202	100		
101	8.0	0.0	28.0
203	17.2	0.0	37.7



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.398 min Scan# 3130
Delta R.T. 0.001 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:240 Resp: 1204355

Ion	Ratio	Lower	Upper
240	100		
120	8.2	6.5	9.7
236	25.9	20.9	31.3





#77
Benzidine
Concen: 18.175 ng
RT: 19.409 min Scan# 2
Delta R.T. 0.005 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:184 Resp: 77750

Ion	Ratio	Lower	Upper
184	100		
185	13.7	11.5	17.3
183	11.2	8.9	13.3



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#78
Pyrene
Concen: 18.964 ng
RT: 19.580 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:202 Resp: 1603874

Ion	Ratio	Lower	Upper
202	100		
200	20.9	16.7	25.1
203	17.8	14.1	21.1





#79
Terphenyl-d14
Concen: 39.418 ng
RT: 19.780 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:244 Resp: 320824

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.6	8.4
122	7.4	5.6	8.4

Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#80
Butylbenzylphthalate
Concen: 19.597 ng
RT: 20.474 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:149 Resp: 620751

Ion	Ratio	Lower	Upper
149	100		
91	64.4	47.4	71.2
206	22.0	19.4	29.0





#81
Benzo(a)anthracene
Concen: 19.187 ng
RT: 21.380 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:228 Resp: 159129

Ion	Ratio	Lower	Upper
228	100		
226	27.3	22.1	33.1
229	19.4	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#82
3,3'-Dichlorobenzidine
Concen: 18.333 ng
RT: 21.303 min Scan# 3114
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:252 Resp: 582038

Ion	Ratio	Lower	Upper
252	100		
254	63.9	51.8	77.6
126	8.9	6.0	9.0





#83
Chrysene
Concen: 19.044 ng
RT: 21.445 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:228 Resp: 146177

Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.2	36.2
229	19.5	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 19.730 ng
RT: 21.298 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:149 Resp: 933505

Ion	Ratio	Lower	Upper
149	100		
167	29.0	24.1	36.1
279	4.6	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 19.552 ng
RT: 22.427 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:149 Resp: 152559

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.4	5.8	8.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.403 min Scan# 3641
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:264 Resp: 1191041

Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.8	28.2
265	22.3	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 18.840 ng
RT: 27.791 min Scan# 4
Delta R.T. -0.006 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:276 Resp: 167196

Ion	Ratio	Lower	Upper
276	100		
138	20.9	16.4	24.6
277	25.0	19.8	29.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#88
Benzo(b)fluoranthene
Concen: 18.588 ng
RT: 23.450 min Scan# 3479
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:252 Resp: 1439776

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	7.9	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 19.005 ng
RT: 23.515 min Scan# 34
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:252 Resp: 148941

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.2
125	7.7	5.8	8.8



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#90
Benzo(a)pyrene
Concen: 18.825 ng
RT: 24.262 min Scan# 3617
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:252 Resp: 1365706

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.4	26.0
125	9.0	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 18.803 ng
RT: 27.838 min Scan# 4
Delta R.T. -0.012 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:278 Resp: 136026

Ion	Ratio	Lower	Upper
278	100		
139	13.1	10.5	15.7
279	23.4	19.4	29.0



Instrument :
BNA_M
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#92
Benzo(g,h,i)perylene
Concen: 19.008 ng
RT: 28.850 min Scan# 4397
Delta R.T. 0.000 min
Lab File: BM050027.D
Acq: 28 Apr 2025 14:27

Tgt Ion:276 Resp: 1316513

Ion	Ratio	Lower	Upper
276	100		
277	23.0	19.1	28.7
138	18.1	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050028.D
Acq On : 28 Apr 2025 15:06
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Quant Time: Apr 28 17:56:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	241109	20.000	ng	0.00
21) Naphthalene-d8	10.563	136	890488	20.000	ng	0.00
39) Acenaphthene-d10	14.410	164	591599	20.000	ng	0.00
64) Phenanthrene-d10	17.157	188	1163791	20.000	ng	0.00
76) Chrysene-d12	21.397	240	1114654	20.000	ng	0.00
86) Perylene-d12	24.403	264	1072526	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	1165032	84.611	ng	0.00
7) Phenol-d6	6.951	99	1487856	85.973	ng	0.00
23) Nitrobenzene-d5	8.928	82	1547270	85.620	ng	0.00
42) 2,4,6-Tribromophenol	15.904	330	751927	85.958	ng	0.00
45) 2-Fluorobiphenyl	13.033	172	4275240	85.489	ng	0.00
79) Terphenyl-d14	19.780	244	6797031	90.233	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.252	88	246624	41.774	ng	100
3) Pyridine	3.657	79	641890	42.316	ng	100
4) n-Nitrosodimethylamine	3.563	42	251749	42.316	ng	100
6) Aniline	7.098	93	936673	42.862	ng	100
8) 2-Chlorophenol	7.340	128	632836	42.508	ng	100
9) Benzaldehyde	6.910	77	497166	42.981	ng	100
10) Phenol	6.981	94	742305	42.370	ng	100
11) bis(2-Chloroethyl)ether	7.193	93	616851	42.166	ng	100
12) 1,3-Dichlorobenzene	7.657	146	754937	41.979	ng	100
13) 1,4-Dichlorobenzene	7.804	146	752959	41.677	ng	100
14) 1,2-Dichlorobenzene	8.122	146	733848	42.050	ng	100
15) Benzyl Alcohol	8.016	79	521025	43.413	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.292	45	748400	42.085	ng	100
17) 2-Methylphenol	8.222	107	498005	43.149	ng	100
18) Hexachloroethane	8.845	117	268328	41.780	ng	100
19) n-Nitroso-di-n-propyla...	8.575	70	483699	43.579	ng	100
20) 3+4-Methylphenols	8.557	107	671942	43.111	ng	100
22) Acetophenone	8.592	105	946106	42.735	ng	100
24) Nitrobenzene	8.969	77	666034	42.625	ng	100
25) Isophorone	9.498	82	1298577	42.156	ng	100
26) 2-Nitrophenol	9.681	139	349855	43.827	ng	100
27) 2,4-Dimethylphenol	9.751	122	576549	43.552	ng	100
28) bis(2-Chloroethoxy)met...	9.975	93	809528	42.518	ng	100
29) 2,4-Dichlorophenol	10.228	162	644546	43.461	ng	100
30) 1,2,4-Trichlorobenzene	10.428	180	759987	42.060	ng	100
31) Naphthalene	10.610	128	1898588	42.100	ng	100
32) Benzoic acid	9.910	122	351345	41.613	ng	100
33) 4-Chloroaniline	10.728	127	784625	42.448	ng	100
34) Hexachlorobutadiene	10.898	225	481510	41.981	ng	100
35) Caprolactam	11.522	113	185984	43.010	ng	100
36) 4-Chloro-3-methylphenol	11.869	107	579492	42.293	ng	100
37) 2-Methylnaphthalene	12.227	142	1269061	41.973	ng	100
38) 1-Methylnaphthalene	12.445	142	1341577	41.928	ng	100
40) 1,2,4,5-Tetrachloroben...	12.598	216	878448	42.894	ng	100
41) Hexachlorocyclopentadiene	12.580	237	538788	43.050	ng	100
43) 2,4,6-Trichlorophenol	12.851	196	563324	43.368	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050028.D
Acq On : 28 Apr 2025 15:06
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Quant Time: Apr 28 17:56:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.933	196	612194	43.706	ng	100
46) 1,1'-Biphenyl	13.245	154	1860581	42.309	ng	100
47) 2-Chloronaphthalene	13.286	162	1479560	42.480	ng	100
48) 2-Nitroaniline	13.498	65	357098	43.684	ng	100
49) Acenaphthylene	14.133	152	2332675	42.478	ng	100
50) Dimethylphthalate	13.869	163	1812335	41.918	ng	100
51) 2,6-Dinitrotoluene	13.992	165	386623	43.191	ng	100
52) Acenaphthene	14.474	154	1340946	42.186	ng	100
53) 3-Nitroaniline	14.327	138	379359	44.389	ng	100
54) 2,4-Dinitrophenol	14.545	184	210115	39.788	ng	100
55) Dibenzofuran	14.810	168	2219206	41.961	ng	100
56) 4-Nitrophenol	14.669	139	264964	40.483	ng	100
57) 2,4-Dinitrotoluene	14.786	165	539867	43.646	ng	100
58) Fluorene	15.463	166	1843261	42.580	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	514644	42.446	ng	100
60) Diethylphthalate	15.233	149	1696013	41.631	ng	100
61) 4-Chlorophenyl-phenyle...	15.457	204	1014134	42.652	ng	100
62) 4-Nitroaniline	15.492	138	370884	44.949	ng	100
63) Azobenzene	15.745	77	1439085	42.150	ng	100
65) 4,6-Dinitro-2-methylph...	15.557	198	315113	43.256	ng	100
66) n-Nitrosodiphenylamine	15.668	169	1525193	42.678	ng	100
67) 4-Bromophenyl-phenylether	16.345	248	600373	42.660	ng	100
68) Hexachlorobenzene	16.468	284	683325	42.122	ng	100
69) Atrazine	16.627	200	550018	42.668	ng	100
70) Pentachlorophenol	16.821	266	387613	42.448	ng	100
71) Phenanthrene	17.198	178	2762249	42.082	ng	100
72) Anthracene	17.292	178	2824805	42.525	ng	100
73) Carbazole	17.568	167	2452474	42.547	ng	100
74) Di-n-butylphthalate	18.121	149	2899479	42.239	ng	100
75) Fluoranthene	19.215	202	3282789	42.598	ng	100
77) Benzidine	19.404	184	1720168	43.446	ng	100
78) Pyrene	19.580	202	3383053	43.220	ng	100
80) Butylbenzylphthalate	20.480	149	1251950	42.704	ng	100
81) Benzo(a)anthracene	21.380	228	3295625	42.935	ng	100
82) 3,3'-Dichlorobenzidine	21.303	252	1292611	43.991	ng	100
83) Chrysene	21.445	228	2992790	42.127	ng	100
84) Bis(2-ethylhexyl)phtha...	21.297	149	1878368	42.895	ng	100
85) Di-n-octyl phthalate	22.427	149	3044578	42.160	ng	100
87) Indeno(1,2,3-cd)pyrene	27.797	276	3435665	42.991	ng	100
88) Benzo(b)fluoranthene	23.450	252	3016197	43.243	ng	100
89) Benzo(k)fluoranthene	23.515	252	2974903	42.155	ng	100
90) Benzo(a)pyrene	24.262	252	2777683	42.519	ng	100
91) Dibenzo(a,h)anthracene	27.850	278	2787948	42.796	ng	100
92) Benzo(g,h,i)perylene	28.850	276	2651145	42.508	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050028.D
Acq On : 28 Apr 2025 15:06
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Quant Time: Apr 28 17:56:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050028.D
Acq On : 28 Apr 2025 15:06
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Quant Time: Apr 28 17:56:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 241109

Ion	Ratio	Lower	Upper
152	100		
150	155.1	124.1	186.1
115	51.5	41.2	61.8





#2
1,4-Dioxane
Concen: 41.774 ng
RT: 3.252 min Scan# 45
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 88 Resp: 246624

Ion	Ratio	Lower	Upper
88	100		
58	68.9	55.1	82.7
43	26.1	20.9	31.3





#3
Pyridine
Concen: 42.316 ng
RT: 3.657 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 641890

Ion	Ratio	Lower	Upper
79	100		
52	60.7	48.6	72.8
51	30.8	24.6	37.0





#4
n-Nitrosodimethylamine
Concen: 42.316 ng
RT: 3.563 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 42 Resp: 251749

Ion	Ratio	Lower	Upper
42	100		
74	150.2	120.2	180.2
44	7.7	6.2	9.2





#5
2-Fluorophenol
Concen: 84.611 ng
RT: 5.357 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:112 Resp: 1165032

Ion	Ratio	Lower	Upper
112	100		
64	53.3	42.6	64.0
63	27.8	22.2	33.4





#6
Aniline
Concen: 42.862 ng
RT: 7.098 min Scan# 699
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 93 Resp: 936673

Ion	Ratio	Lower	Upper
93	100		
66	37.4	29.9	44.9
65	18.1	14.5	21.7





#7
Phenol-d6
Concen: 85.973 ng
RT: 6.951 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 99 Resp: 1487856

Ion	Ratio	Lower	Upper
99	100		
42	16.2	13.0	19.4
71	32.0	25.6	38.4





#8
2-Chlorophenol
Concen: 42.508 ng
RT: 7.340 min Scan# 740
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:128 Resp: 632836

Ion	Ratio	Lower	Upper
128	100		
130	32.8	12.8	52.8
64	41.2	21.2	61.2





#9
Benzaldehyde
Concen: 42.981 ng
RT: 6.910 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 497166

Ion	Ratio	Lower	Upper
77	100		
105	100.5	80.5	120.5
106	96.1	76.1	116.1





#10
Phenol
Concen: 42.370 ng
RT: 6.981 min Scan# 679
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 94 Resp: 742305

Ion	Ratio	Lower	Upper
94	100		
65	27.6	7.6	47.6
66	35.8	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 42.166 ng
RT: 7.193 min Scan# 7
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 93 Resp: 616851

Ion	Ratio	Lower	Upper
93	100		
63	72.3	52.3	92.3
95	32.8	12.8	52.8





#12
1,3-Dichlorobenzene
Concen: 41.979 ng
RT: 7.657 min Scan# 794
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:146 Resp: 754937

Ion	Ratio	Lower	Upper
146	100		
148	65.6	52.5	78.7
75	28.3	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 41.677 ng
RT: 7.804 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:146 Resp: 752959

Ion	Ratio	Lower	Upper
146	100		
148	65.2	52.2	78.2
111	36.6	29.3	43.9





#14
1,2-Dichlorobenzene
Concen: 42.050 ng
RT: 8.122 min Scan# 873
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:146 Resp: 733848

Ion	Ratio	Lower	Upper
146	100		
148	65.5	52.4	78.6
111	38.8	31.0	46.6





#15
Benzyl Alcohol
Concen: 43.413 ng
RT: 8.016 min Scan# 8
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 521025

Ion	Ratio	Lower	Upper
79	100		
108	82.5	66.0	99.0
77	64.5	51.6	77.4





#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.085 ng
RT: 8.292 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 45 Resp: 748400

Ion	Ratio	Lower	Upper
45	100		
77	16.4	0.0	36.4
79	12.7	0.0	32.7





#17
2-Methylphenol
Concen: 43.149 ng
RT: 8.222 min Scan# 8
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:107 Resp: 498005

Ion	Ratio	Lower	Upper
107	100		
108	113.9	91.1	136.7
77	45.3	36.2	54.4
79	44.8	35.8	53.8





#18
Hexachloroethane
Concen: 41.780 ng
RT: 8.845 min Scan# 996
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:117 Resp: 268328

Ion	Ratio	Lower	Upper
117	100		
119	96.6	77.3	115.9
201	128.5	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 43.579 ng
RT: 8.575 min Scan# 9
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 70 Resp: 483699

Ion	Ratio	Lower	Upper
70	100		
42	50.6	40.5	60.7
101	9.7	7.8	11.6
130	21.8	17.4	26.2





#20
3+4-Methylphenols
Concen: 43.111 ng
RT: 8.557 min Scan# 947
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:107 Resp: 671942

Ion	Ratio	Lower	Upper
107	100		
108	92.4	72.4	112.4
77	33.6	13.6	53.6
79	29.2	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:136 Resp: 890488

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.9	13.3
54	6.4	5.1	7.7
68	5.5	4.4	6.6





#22
Acetophenone
Concen: 42.735 ng
RT: 8.592 min Scan# 953
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:105 Resp: 946106

Ion	Ratio	Lower	Upper
105	100		
71	2.1	1.7	2.5
51	23.2	18.6	27.8
120	22.9	18.3	27.5





#23
Nitrobenzene-d5
Concen: 85.620 ng
RT: 8.928 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1547270

Ion	Ratio	Lower	Upper
82	100		
128	44.6	35.7	53.5
54	44.3	35.4	53.2





#24
Nitrobenzene
Concen: 42.625 ng
RT: 8.969 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 77 Resp: 666034

Ion	Ratio	Lower	Upper
77	100		
123	47.4	37.9	56.9
65	12.6	10.1	15.1





#25
Isophorone
Concen: 42.156 ng
RT: 9.498 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1298577

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.7	8.5
138	19.4	15.5	23.3





#26
2-Nitrophenol
Concen: 43.827 ng
RT: 9.681 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:139 Resp: 349855

Ion	Ratio	Lower	Upper
139	100		
109	27.4	21.9	32.9
65	44.3	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 43.552 ng
RT: 9.751 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:122 Resp: 576549

Ion	Ratio	Lower	Upper
122	100		
107	111.3	89.0	133.6
121	57.0	45.6	68.4





#28
bis(2-Chloroethoxy)methane
Concen: 42.518 ng
RT: 9.975 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 93 Resp: 809528

Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.1	39.1
123	11.9	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 43.461 ng
RT: 10.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 644546

Ion	Ratio	Lower	Upper
162	100		
164	65.4	45.4	85.4
98	31.3	11.3	51.3





#30
1,2,4-Trichlorobenzene
Concen: 42.060 ng
RT: 10.428 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:180 Resp: 759987

Ion	Ratio	Lower	Upper
180	100		
182	96.2	77.0	115.4
145	25.7	20.6	30.8





#31
Naphthalene
Concen: 42.100 ng
RT: 10.610 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:128 Resp: 1898588

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.0	10.4	15.6





#32
Benzoic acid
Concen: 41.613 ng
RT: 9.910 min Scan# 1177
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:122 Resp: 351345

Ion	Ratio	Lower	Upper
122	100		
105	123.6	103.6	143.6
77	93.7	73.7	113.7





#33
4-Chloroaniline
Concen: 42.448 ng
RT: 10.728 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:127 Resp: 784625

Ion	Ratio	Lower	Upper
127	100		
129	32.9	26.3	39.5
65	28.1	22.5	33.7
92	18.1	14.5	21.7





#34
Hexachlorobutadiene
Concen: 41.981 ng
RT: 10.898 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:225 Resp: 481510

Ion	Ratio	Lower	Upper
225	100		
223	62.7	50.2	75.2
227	63.3	50.6	76.0





#35
Caprolactam
Concen: 43.010 ng
RT: 11.522 min Scan# 14
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:113 Resp: 185984

Ion	Ratio	Lower	Upper
113	100		
55	142.9	122.9	162.9
56	108.1	88.1	128.1





#36
4-Chloro-3-methylphenol
Concen: 42.293 ng
RT: 11.869 min Scan# 1510
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:107 Resp: 579492

Ion	Ratio	Lower	Upper
107	100		
144	29.7	23.8	35.6
142	89.9	71.9	107.9





#37
2-Methylnaphthalene
Concen: 41.973 ng
RT: 12.227 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:142 Resp: 1269061

Ion	Ratio	Lower	Upper
142	100		
141	88.0	70.4	105.6
115	31.8	25.4	38.2





#38
1-Methylnaphthalene
Concen: 41.928 ng
RT: 12.445 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:142 Resp: 1341577

Ion	Ratio	Lower	Upper
142	100		
141	90.1	72.1	108.1
116	3.4	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:164 Resp: 591599

Ion	Ratio	Lower	Upper
164	100		
162	99.2	79.4	119.0
160	45.3	36.2	54.4





#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.894 ng
RT: 12.598 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:216 Resp: 878448

Ion	Ratio	Lower	Upper
216	100		
214	77.9	62.3	93.5
179	19.2	15.4	23.0
108	14.6	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 43.050 ng
RT: 12.580 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:237 Resp: 538788

Ion	Ratio	Lower	Upper
237	100		
235	62.9	42.9	82.9
272	12.9	0.0	32.9





#42
2,4,6-Tribromophenol
Concen: 85.958 ng
RT: 15.904 min Scan# 2196
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:330 Resp: 751927

Ion	Ratio	Lower	Upper
330	100		
332	96.6	77.3	115.9
141	28.1	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 43.368 ng
RT: 12.851 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:196 Resp: 563324

Ion	Ratio	Lower	Upper
196	100		
198	98.8	79.0	118.6
200	31.4	25.1	37.7





#44
2,4,5-Trichlorophenol
Concen: 43.706 ng
RT: 12.933 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:196 Resp: 612194

Ion	Ratio	Lower	Upper
196	100		
198	95.9	76.7	115.1
97	37.6	30.1	45.1
132	20.8	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 85.489 ng
RT: 13.033 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:172 Resp: 4275240

Ion	Ratio	Lower	Upper
172	100		
171	34.7	27.8	41.6
170	23.3	18.6	28.0





#46
1,1'-Biphenyl
Concen: 42.309 ng
RT: 13.245 min Scan# 1744
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:154 Resp: 1860581

Ion	Ratio	Lower	Upper
154	100		
153	40.0	20.0	60.0
76	10.3	0.0	30.3





#47
2-Chloronaphthalene
Concen: 42.480 ng
RT: 13.286 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 1479560

Ion	Ratio	Lower	Upper
162	100		
127	32.4	25.9	38.9
164	33.3	26.6	40.0





#48
2-Nitroaniline
Concen: 43.684 ng
RT: 13.498 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion: 65 Resp: 357098

Ion	Ratio	Lower	Upper
65	100		
92	71.4	57.1	85.7
138	121.2	97.0	145.4





#49
Acenaphthylene
Concen: 42.478 ng
RT: 14.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 2332675

Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.8	23.8
153	13.2	10.6	15.8





#50
Dimethylphthalate
Concen: 41.918 ng
RT: 13.869 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:163 Resp: 1812335

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.2
164	10.2	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 43.191 ng
RT: 13.992 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 386623

Ion	Ratio	Lower	Upper
165	100		
63	43.9	35.1	52.7
89	48.0	38.4	57.6





#52
Acenaphthene
Concen: 42.186 ng
RT: 14.474 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:154 Resp: 1340946

Ion	Ratio	Lower	Upper
154	100		
153	111.2	89.0	133.4
152	53.8	43.0	64.6





#53
3-Nitroaniline
Concen: 44.389 ng
RT: 14.327 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:138 Resp: 379359

Ion	Ratio	Lower	Upper
138	100		
108	9.3	7.4	11.2
92	105.0	84.0	126.0





#54
2,4-Dinitrophenol
Concen: 39.788 ng
RT: 14.545 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:184 Resp: 210115

Ion	Ratio	Lower	Upper
184	100		
63	49.7	39.8	59.6
154	57.6	46.1	69.1





#55
Dibenzofuran
Concen: 41.961 ng
RT: 14.810 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:168 Resp: 2219206

Ion	Ratio	Lower	Upper
168	100		
139	32.6	26.1	39.1
169	13.5	10.8	16.2





#56
4-Nitrophenol
Concen: 40.483 ng
RT: 14.669 min Scan# 1986
Delta R.T. -0.001 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:139 Resp: 264964

Ion	Ratio	Lower	Upper
139	100		
109	67.0	47.0	87.0
65	84.3	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 43.646 ng
RT: 14.786 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 539867

Ion	Ratio	Lower	Upper
165	100		
63	36.3	29.0	43.6
89	61.8	49.4	74.2
182	2.7	2.2	3.2





#58
Fluorene
Concen: 42.580 ng
RT: 15.463 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:166 Resp: 1843261

Ion	Ratio	Lower	Upper
166	100		
165	95.8	76.6	115.0
167	13.3	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 42.446 ng
RT: 15.045 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:232 Resp: 514644

Ion	Ratio	Lower	Upper
232	100		
131	35.3	28.2	42.4
130	1.9	1.5	2.3
166	28.3	22.6	34.0





#60
Diethylphthalate
Concen: 41.631 ng
RT: 15.233 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:149 Resp: 1696013

Ion	Ratio	Lower	Upper
149	100		
177	23.9	19.1	28.7
150	12.0	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 42.652 ng
RT: 15.457 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:204 Resp: 1014134

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.3	39.5
141	42.5	34.0	51.0





#62
4-Nitroaniline
Concen: 44.949 ng
RT: 15.492 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:138 Resp: 370884

Ion	Ratio	Lower	Upper
138	100		
92	45.9	25.9	65.9
108	72.4	52.4	92.4





#63
Azobenzene
Concen: 42.150 ng
RT: 15.745 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 1439085

Ion	Ratio	Lower	Upper
77	100		
182	30.5	10.5	50.5
105	21.1	1.1	41.1
51	26.3	6.3	46.3





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:188 Resp: 1163791

Ion	Ratio	Lower	Upper
188	100		
94	7.1	5.7	8.5
80	7.8	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 43.256 ng
RT: 15.557 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:198 Resp: 315113

Ion	Ratio	Lower	Upper
198	100		
51	30.4	10.4	50.4
105	34.6	14.6	54.6





#66
n-Nitrosodiphenylamine
Concen: 42.678 ng
RT: 15.668 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:169 Resp: 1525193

Ion	Ratio	Lower	Upper
169	100		
168	66.8	53.4	80.2
167	36.7	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 42.660 ng
RT: 16.345 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:248 Resp: 600373

Ion	Ratio	Lower	Upper
248	100		
250	95.2	76.2	114.2
141	57.4	45.9	68.9





#68
Hexachlorobenzene
Concen: 42.122 ng
RT: 16.468 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:284 Resp: 683325

Ion	Ratio	Lower	Upper
284	100		
142	29.9	23.9	35.9
249	30.6	24.5	36.7





#69
Atrazine
Concen: 42.668 ng
RT: 16.627 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:200 Resp: 550018

Ion	Ratio	Lower	Upper
200	100		
173	24.3	4.3	44.3
215	50.6	30.6	70.6





#70
Pentachlorophenol
Concen: 42.448 ng
RT: 16.821 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:266 Resp: 387613

Ion	Ratio	Lower	Upper
266	100		
268	64.0	51.2	76.8
264	62.0	49.6	74.4





#71
Phenanthrene
Concen: 42.082 ng
RT: 17.198 min Scan# 24
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:178 Resp: 2762249

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.3	22.9
179	15.5	12.4	18.6





#72
Anthracene
Concen: 42.525 ng
RT: 17.292 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:178 Resp: 2824805

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.3	12.2	18.4





#73
Carbazole
Concen: 42.547 ng
RT: 17.568 min Scan# 24
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:167 Resp: 2452474

Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.2	25.8
139	11.3	9.0	13.6





#74
Di-n-butylphthalate
Concen: 42.239 ng
RT: 18.121 min Scan# 2573
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:149 Resp: 2899479

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.7	3.8	5.6





#75
Fluoranthene
Concen: 42.598 ng
RT: 19.215 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:202 Resp: 3282789

Ion	Ratio	Lower	Upper
202	100		
101	8.0	0.0	28.0
203	17.7	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.397 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:240 Resp: 1114654

Ion	Ratio	Lower	Upper
240	100		
120	8.1	6.5	9.7
236	26.1	20.9	31.3





#77
Benzidine
Concen: 43.446 ng
RT: 19.404 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:184 Resp: 1720168

Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.5	17.3
183	11.1	8.9	13.3





#78
Pyrene
Concen: 43.220 ng
RT: 19.580 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:202 Resp: 3383053

Ion	Ratio	Lower	Upper
202	100		
200	20.9	16.7	25.1
203	17.6	14.1	21.1





#79
Terphenyl-d14
Concen: 90.233 ng
RT: 19.780 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:244 Resp: 6797031

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.6	8.4
122	7.0	5.6	8.4





#80
Butylbenzylphthalate
Concen: 42.704 ng
RT: 20.480 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:149 Resp: 1251950

Ion	Ratio	Lower	Upper
149	100		
91	59.3	47.4	71.2
206	24.2	19.4	29.0





#81
Benzo(a)anthracene
Concen: 42.935 ng
RT: 21.380 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 3295625

Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.1	33.1
229	19.6	15.7	23.5





#82
3,3'-Dichlorobenzidine
Concen: 43.991 ng
RT: 21.303 min Scan# 3114
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:252 Resp: 1292611

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.8	77.6
126	7.5	6.0	9.0





#83
Chrysene
Concen: 42.127 ng
RT: 21.445 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 2992790

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.2
229	19.6	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 42.895 ng
RT: 21.297 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:149 Resp: 1878368

Ion	Ratio	Lower	Upper
149	100		
167	30.1	24.1	36.1
279	4.9	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 42.160 ng
RT: 22.427 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:149 Resp: 3044578

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.3	5.8	8.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.403 min Scan# 3641
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:264 Resp: 1072526

Ion	Ratio	Lower	Upper
264	100		
260	23.5	18.8	28.2
265	21.2	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 42.991 ng
RT: 27.797 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:276 Resp: 3435665

Ion	Ratio	Lower	Upper
276	100		
138	20.5	16.4	24.6
277	24.8	19.8	29.8





#88
Benzo(b)fluoranthene
Concen: 43.243 ng
RT: 23.450 min Scan# 3479
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:252 Resp: 3016197

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	7.6	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 42.155 ng
RT: 23.515 min Scan# 34
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:252 Resp: 2974903

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.2
125	7.3	5.8	8.8





#90
Benzo(a)pyrene
Concen: 42.519 ng
RT: 24.262 min Scan# 3617
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:252 Resp: 2777683

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.4	26.0
125	8.2	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 42.796 ng
RT: 27.850 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Instrument :
BNA_M
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 2787948

Ion	Ratio	Lower	Upper
278	100		
139	13.1	10.5	15.7
279	24.2	19.4	29.0





#92
Benzo(g,h,i)perylene
Concen: 42.508 ng
RT: 28.850 min Scan# 4397
Delta R.T. 0.000 min
Lab File: BM050028.D
Acq: 28 Apr 2025 15:06

Tgt Ion:276 Resp: 2651145

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.1	28.7
138	17.9	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050029.D
Acq On : 28 Apr 2025 15:45
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Quant Time: Apr 28 17:57:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	228207	20.000	ng	0.00
21) Naphthalene-d8	10.563	136	837552	20.000	ng	0.00
39) Acenaphthene-d10	14.410	164	554448	20.000	ng	0.00
64) Phenanthrene-d10	17.156	188	1105769	20.000	ng	0.00
76) Chrysene-d12	21.397	240	1097646	20.000	ng	0.00
86) Perylene-d12	24.397	264	1061511	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	1382138	106.054	ng	0.00
7) Phenol-d6	6.951	99	1754123	107.089	ng	0.00
23) Nitrobenzene-d5	8.928	82	1811208	106.560	ng	0.00
42) 2,4,6-Tribromophenol	15.904	330	911669	111.203	ng	0.00
45) 2-Fluorobiphenyl	13.033	172	5051261	107.775	ng	0.00
79) Terphenyl-d14	19.780	244	8457811	114.020	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.251	88	287618	51.472	ng	100
3) Pyridine	3.657	79	757569	52.766	ng	99
4) n-Nitrosodimethylamine	3.563	42	295491	52.476	ng	100
6) Aniline	7.098	93	1096746	53.024	ng	99
8) 2-Chlorophenol	7.339	128	742074	52.664	ng	100
9) Benzaldehyde	6.910	77	577740	52.770	ng	99
10) Phenol	6.981	94	877155	52.897	ng	98
11) bis(2-Chloroethyl)ether	7.192	93	725268	52.379	ng	99
12) 1,3-Dichlorobenzene	7.657	146	885183	52.004	ng	100
13) 1,4-Dichlorobenzene	7.804	146	888114	51.937	ng	100
14) 1,2-Dichlorobenzene	8.122	146	859475	52.033	ng	99
15) Benzyl Alcohol	8.016	79	612811	53.947	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.292	45	863443	51.300	ng	99
17) 2-Methylphenol	8.228	107	585120	53.563	ng	99
18) Hexachloroethane	8.845	117	312710	51.444	ng	98
19) n-Nitroso-di-n-propyla...	8.575	70	563847	53.672	ng	100
20) 3+4-Methylphenols	8.551	107	794422	53.851	ng	98
22) Acetophenone	8.592	105	1100024	52.827	ng	99
24) Nitrobenzene	8.969	77	779215	53.020	ng	99
25) Isophorone	9.498	82	1526436	52.685	ng	100
26) 2-Nitrophenol	9.680	139	415316	55.315	ng	99
27) 2,4-Dimethylphenol	9.751	122	674547	54.175	ng	100
28) bis(2-Chloroethoxy)met...	9.975	93	946285	52.843	ng	100
29) 2,4-Dichlorophenol	10.227	162	751117	53.847	ng	99
30) 1,2,4-Trichlorobenzene	10.427	180	891501	52.457	ng	98
31) Naphthalene	10.610	128	2214025	52.198	ng	100
32) Benzoic acid	9.922	122	431557	54.343	ng	98
33) 4-Chloroaniline	10.727	127	937309	53.913	ng	99
34) Hexachlorobutadiene	10.898	225	566049	52.471	ng	99
35) Caprolactam	11.522	113	218572	53.741	ng	92
36) 4-Chloro-3-methylphenol	11.869	107	688099	53.394	ng	98
37) 2-Methylnaphthalene	12.227	142	1494348	52.548	ng	99
38) 1-Methylnaphthalene	12.445	142	1572248	52.243	ng	99
40) 1,2,4,5-Tetrachloroben...	12.604	216	1033027	53.822	ng	99
41) Hexachlorocyclopentadiene	12.580	237	649060	55.336	ng	99
43) 2,4,6-Trichlorophenol	12.851	196	672901	55.275	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050029.D
Acq On : 28 Apr 2025 15:45
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Quant Time: Apr 28 17:57:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.933	196	719930	54.841	ng	100
46) 1,1'-Biphenyl	13.245	154	2191996	53.185	ng	100
47) 2-Chloronaphthalene	13.286	162	1729678	52.989	ng	100
48) 2-Nitroaniline	13.498	65	421489	55.016	ng	99
49) Acenaphthylene	14.133	152	2731966	53.082	ng	100
50) Dimethylphthalate	13.874	163	2138242	52.770	ng	100
51) 2,6-Dinitrotoluene	13.992	165	458476	54.650	ng	99
52) Acenaphthene	14.474	154	1587472	53.287	ng	99
53) 3-Nitroaniline	14.327	138	450021	56.186	ng	96
54) 2,4-Dinitrophenol	14.545	184	269515	52.338	ng	99
55) Dibenzofuran	14.810	168	2617489	52.809	ng	100
56) 4-Nitrophenol	14.663	139	328078	52.029	ng	100
57) 2,4-Dinitrotoluene	14.786	165	648262	55.921	ng	99
58) Fluorene	15.462	166	2180247	53.740	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	617151	54.311	ng	99
60) Diethylphthalate	15.233	149	2011763	52.691	ng	99
61) 4-Chlorophenyl-phenyle...	15.457	204	1207514	54.188	ng	99
62) 4-Nitroaniline	15.492	138	445520	57.612	ng	99
63) Azobenzene	15.745	77	1703932	53.251	ng	100
65) 4,6-Dinitro-2-methylph...	15.557	198	393191	56.806	ng	98
66) n-Nitrosodiphenylamine	15.668	169	1815723	53.474	ng	99
67) 4-Bromophenyl-phenylether	16.345	248	713498	53.359	ng	98
68) Hexachlorobenzene	16.468	284	815392	52.900	ng	98
69) Atrazine	16.627	200	669330	54.648	ng	99
70) Pentachlorophenol	16.821	266	475717	54.830	ng	99
71) Phenanthrene	17.198	178	3349195	53.702	ng	100
72) Anthracene	17.292	178	3417239	54.143	ng	100
73) Carbazole	17.568	167	2971695	54.260	ng	99
74) Di-n-butylphthalate	18.121	149	3530893	54.137	ng	100
75) Fluoranthene	19.215	202	4021516	54.922	ng	100
77) Benzidine	19.403	184	2110748	54.137	ng	100
78) Pyrene	19.580	202	4169831	54.097	ng	100
80) Butylbenzylphthalate	20.474	149	1557463	53.948	ng	96
81) Benzo(a)anthracene	21.380	228	4058011	53.687	ng	100
82) 3,3'-Dichlorobenzidine	21.303	252	1639448	56.659	ng	99
83) Chrysene	21.444	228	3755543	53.683	ng	99
84) Bis(2-ethylhexyl)phtha...	21.297	149	2336926	54.194	ng	98
85) Di-n-octyl phthalate	22.427	149	3852190	54.170	ng	100
87) Indeno(1,2,3-cd)pyrene	27.797	276	4369253	55.240	ng	99
88) Benzo(b)fluoranthene	23.450	252	3755728	54.405	ng	99
89) Benzo(k)fluoranthene	23.515	252	3844095	55.037	ng	99
90) Benzo(a)pyrene	24.262	252	3529489	54.588	ng	99
91) Dibenzo(a,h)anthracene	27.838	278	3575099	55.448	ng	99
92) Benzo(g,h,i)perylene	28.850	276	3363778	54.494	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050029.D
Acq On : 28 Apr 2025 15:45
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Quant Time: Apr 28 17:57:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050029.D
Acq On : 28 Apr 2025 15:45
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Quant Time: Apr 28 17:57:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 228207

Ion	Ratio	Lower	Upper
152	100		
150	157.9	124.1	186.1
115	51.6	41.2	61.8





#2
1,4-Dioxane
Concen: 51.472 ng
RT: 3.251 min Scan# 45
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 88 Resp: 287618

Ion	Ratio	Lower	Upper
88	100		
58	68.7	55.1	82.7
43	26.0	20.9	31.3





#3
Pyridine
Concen: 52.766 ng
RT: 3.657 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 757569

Ion	Ratio	Lower	Upper
79	100		
52	61.0	48.6	72.8
51	31.7	24.6	37.0





#4
n-Nitrosodimethylamine
Concen: 52.476 ng
RT: 3.563 min Scan# 98
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 42 Resp: 295491

Ion	Ratio	Lower	Upper
42	100		
74	150.4	120.2	180.2
44	7.4	6.2	9.2





#5
2-Fluorophenol
Concen: 106.054 ng
RT: 5.357 min Scan# 4
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:112 Resp: 1382138

Ion	Ratio	Lower	Upper
112	100		
64	53.2	42.6	64.0
63	27.8	22.2	33.4





#6
Aniline
Concen: 53.024 ng
RT: 7.098 min Scan# 699
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 93 Resp: 1096746

Ion	Ratio	Lower	Upper
93	100		
66	36.7	29.9	44.9
65	18.7	14.5	21.7





#7
Phenol-d6
Concen: 107.089 ng
RT: 6.951 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 99 Resp: 1754123

Ion	Ratio	Lower	Upper
99	100		
42	15.9	13.0	19.4
71	31.3	25.6	38.4





#8
2-Chlorophenol
Concen: 52.664 ng
RT: 7.339 min Scan# 740
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:128 Resp: 742074

Ion	Ratio	Lower	Upper
128	100		
130	33.2	12.8	52.8
64	41.4	21.2	61.2





#9
Benzaldehyde
Concen: 52.770 ng
RT: 6.910 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 577740

Ion	Ratio	Lower	Upper
77	100		
105	98.2	80.5	120.5
106	95.9	76.1	116.1





#10
Phenol
Concen: 52.897 ng
RT: 6.981 min Scan# 679
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 94 Resp: 877155

Ion	Ratio	Lower	Upper
94	100		
65	26.6	7.6	47.6
66	34.7	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 52.379 ng
RT: 7.192 min Scan# 7
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 93 Resp: 725268

Ion	Ratio	Lower	Upper
93	100		
63	71.7	52.3	92.3
95	31.8	12.8	52.8





#12
1,3-Dichlorobenzene
Concen: 52.004 ng
RT: 7.657 min Scan# 794
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:146 Resp: 885183

Ion	Ratio	Lower	Upper
146	100		
148	65.4	52.5	78.7
75	27.8	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 51.937 ng
RT: 7.804 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:146 Resp: 888114

Ion	Ratio	Lower	Upper
146	100		
148	65.2	52.2	78.2
111	36.5	29.3	43.9





#14
1,2-Dichlorobenzene
Concen: 52.033 ng
RT: 8.122 min Scan# 873
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:146 Resp: 859475

Ion	Ratio	Lower	Upper
146	100		
148	66.1	52.4	78.6
111	38.2	31.0	46.6





#15
Benzyl Alcohol
Concen: 53.947 ng
RT: 8.016 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 612811

Ion	Ratio	Lower	Upper
79	100		
108	81.6	66.0	99.0
77	64.1	51.6	77.4





#16
2,2'-oxybis(1-Chloropropane)
Concen: 51.300 ng
RT: 8.292 min Scan# 902
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 45 Resp: 863443

Ion	Ratio	Lower	Upper
45	100		
77	16.9	0.0	36.4
79	12.7	0.0	32.7





#17
2-Methylphenol
Concen: 53.563 ng
RT: 8.228 min Scan# 8
Delta R.T. 0.006 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:107 Resp: 585120

Ion	Ratio	Lower	Upper
107	100		
108	114.9	91.1	136.7
77	45.2	36.2	54.4
79	43.1	35.8	53.8





#18
Hexachloroethane
Concen: 51.444 ng
RT: 8.845 min Scan# 996
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:117 Resp: 312710

Ion	Ratio	Lower	Upper
117	100		
119	97.6	77.3	115.9
201	131.4	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 53.672 ng
RT: 8.575 min Scan# 9
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 563847

Ion	Ratio	Lower	Upper
70	100		
42	50.7	40.5	60.7
101	9.8	7.8	11.6
130	21.7	17.4	26.2





#20
3+4-Methylphenols
Concen: 53.851 ng
RT: 8.551 min Scan# 946
Delta R.T. -0.006 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:107 Resp: 794422

Ion	Ratio	Lower	Upper
107	100		
108	90.4	72.4	112.4
77	33.0	13.6	53.6
79	27.5	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:136 Resp: 837552

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.9	13.3
54	6.5	5.1	7.7
68	5.4	4.4	6.6





#22
Acetophenone
Concen: 52.827 ng
RT: 8.592 min Scan# 953
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:105 Resp: 1100024

Ion	Ratio	Lower	Upper
105	100		
71	2.3	1.7	2.5
51	23.2	18.6	27.8
120	23.4	18.3	27.5





#23
Nitrobenzene-d5
Concen: 106.560 ng
RT: 8.928 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 1811208

Ion	Ratio	Lower	Upper
82	100		
128	45.3	35.7	53.5
54	44.3	35.4	53.2





#24
Nitrobenzene
Concen: 53.020 ng
RT: 8.969 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 77 Resp: 779215

Ion	Ratio	Lower	Upper
77	100		
123	48.4	37.9	56.9
65	13.0	10.1	15.1





#25
Isophorone
Concen: 52.685 ng
RT: 9.498 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 1526436

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.7	8.5
138	19.3	15.5	23.3





#26
2-Nitrophenol
Concen: 55.315 ng
RT: 9.680 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:139 Resp: 415316

Ion	Ratio	Lower	Upper
139	100		
109	27.8	21.9	32.9
65	45.0	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 54.175 ng
RT: 9.751 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:122 Resp: 674547

Ion	Ratio	Lower	Upper
122	100		
107	110.9	89.0	133.6
121	56.7	45.6	68.4



#28
bis(2-Chloroethoxy)methane
Concen: 52.843 ng
RT: 9.975 min Scan# 1188
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 93 Resp: 946285

Ion	Ratio	Lower	Upper
93	100		
95	32.6	26.1	39.1
123	11.7	9.5	14.3



#29
2,4-Dichlorophenol
Concen: 53.847 ng
RT: 10.227 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 751117

Ion	Ratio	Lower	Upper
162	100		
164	64.8	45.4	85.4
98	29.7	11.3	51.3





#30
1,2,4-Trichlorobenzene
Concen: 52.457 ng
RT: 10.427 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:180 Resp: 891501

Ion	Ratio	Lower	Upper
180	100		
182	97.9	77.0	115.4
145	25.3	20.6	30.8





#31
Naphthalene
Concen: 52.198 ng
RT: 10.610 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:128 Resp: 2214025

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.0	10.4	15.6





#32
Benzoic acid
Concen: 54.343 ng
RT: 9.922 min Scan# 1179
Delta R.T. 0.012 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:122 Resp: 431557

Ion	Ratio	Lower	Upper
122	100		
105	122.2	103.6	143.6
77	90.1	73.7	113.7





#33
4-Chloroaniline
Concen: 53.913 ng
RT: 10.727 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:127 Resp: 937309

Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.3	39.5
65	27.3	22.5	33.7
92	18.0	14.5	21.7





#34
Hexachlorobutadiene
Concen: 52.471 ng
RT: 10.898 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:225 Resp: 566049

Ion	Ratio	Lower	Upper
225	100		
223	61.8	50.2	75.2
227	63.5	50.6	76.0





#35
Caprolactam
Concen: 53.741 ng
RT: 11.522 min Scan# 14
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:113 Resp: 218572

Ion	Ratio	Lower	Upper
113	100		
55	155.5	122.9	162.9
56	114.3	88.1	128.1





#36
4-Chloro-3-methylphenol
Concen: 53.394 ng
RT: 11.869 min Scan# 1510
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:107 Resp: 688099

Ion	Ratio	Lower	Upper
107	100		
144	29.2	23.8	35.6
142	87.9	71.9	107.9





#37
2-Methylnaphthalene
Concen: 52.548 ng
RT: 12.227 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:142 Resp: 1494348

Ion	Ratio	Lower	Upper
142	100		
141	87.5	70.4	105.6
115	32.4	25.4	38.2





#38
1-Methylnaphthalene
Concen: 52.243 ng
RT: 12.445 min Scan# 1608
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:142 Resp: 1572248

Ion	Ratio	Lower	Upper
142	100		
141	90.9	72.1	108.1
116	3.6	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:164 Resp: 554448

Ion	Ratio	Lower	Upper
164	100		
162	99.5	79.4	119.0
160	44.4	36.2	54.4





#40
1,2,4,5-Tetrachlorobenzene
Concen: 53.822 ng
RT: 12.604 min Scan# 1635
Delta R.T. 0.006 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:216 Resp: 1033027

Ion	Ratio	Lower	Upper
216	100		
214	78.6	62.3	93.5
179	19.1	15.4	23.0
108	14.5	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 55.336 ng
RT: 12.580 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:237 Resp: 649060

Ion	Ratio	Lower	Upper
237	100		
235	62.6	42.9	82.9
272	13.5	0.0	32.9





#42
2,4,6-Tribromophenol
Concen: 111.203 ng
RT: 15.904 min Scan# 2196
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:330 Resp: 911669

Ion	Ratio	Lower	Upper
330	100		
332	97.1	77.3	115.9
141	27.3	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 55.275 ng
RT: 12.851 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:196 Resp: 672901

Ion	Ratio	Lower	Upper
196	100		
198	96.4	79.0	118.6
200	31.4	25.1	37.7





#44
2,4,5-Trichlorophenol
Concen: 54.841 ng
RT: 12.933 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:196 Resp: 719930

Ion	Ratio	Lower	Upper
196	100		
198	96.1	76.7	115.1
97	37.9	30.1	45.1
132	21.3	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 107.775 ng
RT: 13.033 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:172 Resp: 5051261

Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.8	41.6
170	23.3	18.6	28.0





#46
1,1'-Biphenyl
Concen: 53.185 ng
RT: 13.245 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:154 Resp: 2191996

Ion	Ratio	Lower	Upper
154	100		
153	39.7	20.0	60.0
76	10.1	0.0	30.3





#47
2-Chloronaphthalene
Concen: 52.989 ng
RT: 13.286 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 1729678

Ion	Ratio	Lower	Upper
162	100		
127	32.1	25.9	38.9
164	33.1	26.6	40.0





#48
2-Nitroaniline
Concen: 55.016 ng
RT: 13.498 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion: 65 Resp: 421489

Ion	Ratio	Lower	Upper
65	100		
92	71.5	57.1	85.7
138	122.2	97.0	145.4





#49
Acenaphthylene
Concen: 53.082 ng
RT: 14.133 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 2731966

Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.1	10.6	15.8





#50
Dimethylphthalate
Concen: 52.770 ng
RT: 13.874 min Scan# 1851
Delta R.T. 0.006 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:163 Resp: 2138242

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.2
164	10.1	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 54.650 ng
RT: 13.992 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 458476

Ion	Ratio	Lower	Upper
165	100		
63	43.1	35.1	52.7
89	47.1	38.4	57.6





#52
Acenaphthene
Concen: 53.287 ng
RT: 14.474 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:154 Resp: 1587472

Ion	Ratio	Lower	Upper
154	100		
153	112.2	89.0	133.4
152	53.4	43.0	64.6





#53
3-Nitroaniline
Concen: 56.186 ng
RT: 14.327 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:138 Resp: 450021

Ion	Ratio	Lower	Upper
138	100		
108	9.1	7.4	11.2
92	101.0	84.0	126.0





#54
2,4-Dinitrophenol
Concen: 52.338 ng
RT: 14.545 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:184 Resp: 269515

Ion	Ratio	Lower	Upper
184	100		
63	49.3	39.8	59.6
154	58.0	46.1	69.1





#55
Dibenzofuran
Concen: 52.809 ng
RT: 14.810 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:168 Resp: 2617489

Ion	Ratio	Lower	Upper
168	100		
139	32.6	26.1	39.1
169	13.5	10.8	16.2





#56
4-Nitrophenol
Concen: 52.029 ng
RT: 14.663 min Scan# 1985
Delta R.T. -0.007 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:139 Resp: 328078

Ion	Ratio	Lower	Upper
139	100		
109	66.7	47.0	87.0
65	84.3	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 55.921 ng
RT: 14.786 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 648262

Ion	Ratio	Lower	Upper
165	100		
63	35.3	29.0	43.6
89	60.8	49.4	74.2
182	2.8	2.2	3.2





#58
Fluorene
Concen: 53.740 ng
RT: 15.462 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:166 Resp: 2180247

Ion	Ratio	Lower	Upper
166	100		
165	96.1	76.6	115.0
167	13.0	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 54.311 ng
RT: 15.045 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:232 Resp: 617151

Ion	Ratio	Lower	Upper
232	100		
131	34.5	28.2	42.4
130	1.9	1.5	2.3
166	27.8	22.6	34.0





#60
Diethylphthalate
Concen: 52.691 ng
RT: 15.233 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:149 Resp: 2011763

Ion	Ratio	Lower	Upper
149	100		
177	24.1	19.1	28.7
150	12.5	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 54.188 ng
RT: 15.457 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:204 Resp: 1207514

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.3	39.5
141	41.6	34.0	51.0





#62
4-Nitroaniline
Concen: 57.612 ng
RT: 15.492 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:138 Resp: 445520

Ion	Ratio	Lower	Upper
138	100		
92	46.9	25.9	65.9
108	72.7	52.4	92.4





#63
Azobenzene
Concen: 53.251 ng
RT: 15.745 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 1703932

Ion	Ratio	Lower	Upper
77	100		
182	30.7	10.5	50.5
105	21.3	1.1	41.1
51	26.3	6.3	46.3





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.156 min Scan# 2409
Delta R.T. -0.001 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:188 Resp: 1105769

Ion	Ratio	Lower	Upper
188	100		
94	7.0	5.7	8.5
80	7.4	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 56.806 ng
RT: 15.557 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:198 Resp: 393191

Ion	Ratio	Lower	Upper
198	100		
51	29.5	10.4	50.4
105	33.1	14.6	54.6





#66
n-Nitrosodiphenylamine
Concen: 53.474 ng
RT: 15.668 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:169 Resp: 1815723

Ion	Ratio	Lower	Upper
169	100		
168	67.4	53.4	80.2
167	36.5	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 53.359 ng
RT: 16.345 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:248 Resp: 713498

Ion	Ratio	Lower	Upper
248	100		
250	98.1	76.2	114.2
141	57.6	45.9	68.9





#68
Hexachlorobenzene
Concen: 52.900 ng
RT: 16.468 min Scan# 2292
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:284 Resp: 815392

Ion	Ratio	Lower	Upper
284	100		
142	28.1	23.9	35.9
249	30.5	24.5	36.7





#69
Atrazine
Concen: 54.648 ng
RT: 16.627 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:200 Resp: 669330

Ion	Ratio	Lower	Upper
200	100		
173	24.3	4.3	44.3
215	52.1	30.6	70.6





#70
Pentachlorophenol
Concen: 54.830 ng
RT: 16.821 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:266 Resp: 475717

Ion	Ratio	Lower	Upper
266	100		
268	63.2	51.2	76.8
264	62.0	49.6	74.4





#71
Phenanthrene
Concen: 53.702 ng
RT: 17.198 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:178 Resp: 3349195

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.3	22.9
179	15.4	12.4	18.6





#72
Anthracene
Concen: 54.143 ng
RT: 17.292 min Scan# 2432
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:178 Resp: 3417239

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.6	12.2	18.4





#73
Carbazole
Concen: 54.260 ng
RT: 17.568 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:167 Resp: 2971695

Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.2	25.8
139	10.9	9.0	13.6





#74
Di-n-butylphthalate
Concen: 54.137 ng
RT: 18.121 min Scan# 2573
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:149 Resp: 3530893

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	4.6	3.8	5.6





#75
Fluoranthene
Concen: 54.922 ng
RT: 19.215 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:202 Resp: 4021516

Ion	Ratio	Lower	Upper
202	100		
101	8.1	0.0	28.0
203	17.7	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.397 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:240 Resp: 1097646

Ion	Ratio	Lower	Upper
240	100		
120	8.3	6.5	9.7
236	25.5	20.9	31.3





#77
Benzidine
Concen: 54.137 ng
RT: 19.403 min Scan# 2
Delta R.T. -0.001 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 2110748

Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.5	17.3
183	11.2	8.9	13.3





#78
Pyrene
Concen: 54.097 ng
RT: 19.580 min Scan# 2821
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:202 Resp: 4169831

Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.7	25.1
203	17.7	14.1	21.1





#79
Terphenyl-d14
Concen: 114.020 ng
RT: 19.780 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:244 Resp: 8457811

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.6	8.4
122	6.7	5.6	8.4





#80
Butylbenzylphthalate
Concen: 53.948 ng
RT: 20.474 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:149 Resp: 1557463

Ion	Ratio	Lower	Upper
149	100		
91	63.4	47.4	71.2
206	23.9	19.4	29.0





#81
Benzo(a)anthracene
Concen: 53.687 ng
RT: 21.380 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 4058011

Ion	Ratio	Lower	Upper
228	100		
226	27.7	22.1	33.1
229	20.0	15.7	23.5





#82
3,3'-Dichlorobenzidine
Concen: 56.659 ng
RT: 21.303 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:252 Resp: 1639448

Ion	Ratio	Lower	Upper
252	100		
254	65.3	51.8	77.6
126	7.1	6.0	9.0





#83
Chrysene
Concen: 53.683 ng
RT: 21.444 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 3755543

Ion	Ratio	Lower	Upper
228	100		
226	30.7	24.2	36.2
229	19.7	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 54.194 ng
RT: 21.297 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:149 Resp: 2336926

Ion	Ratio	Lower	Upper
149	100		
167	30.9	24.1	36.1
279	5.7	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 54.170 ng
RT: 22.427 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:149 Resp: 3852190

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.2	5.8	8.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.397 min Scan# 3640
Delta R.T. -0.006 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:264 Resp: 1061511

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.8	28.2
265	21.9	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 55.240 ng
RT: 27.797 min Scan# 4
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:276 Resp: 4369253

Ion	Ratio	Lower	Upper
276	100		
138	20.1	16.4	24.6
277	24.9	19.8	29.8





#88
Benzo(b)fluoranthene
Concen: 54.405 ng
RT: 23.450 min Scan# 3479
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:252 Resp: 3755728

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	7.2	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 55.037 ng
RT: 23.515 min Scan# 34
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:252 Resp: 3844095

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.4	26.2
125	6.8	5.8	8.8





#90
Benzo(a)pyrene
Concen: 54.588 ng
RT: 24.262 min Scan# 3617
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:252 Resp: 3529489

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.4	26.0
125	8.1	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 55.448 ng
RT: 27.838 min Scan# 4
Delta R.T. -0.012 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Instrument :
BNA_M
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 3575099

Ion	Ratio	Lower	Upper
278	100		
139	12.3	10.5	15.7
279	23.7	19.4	29.0





#92
Benzo(g,h,i)perylene
Concen: 54.494 ng
RT: 28.850 min Scan# 4397
Delta R.T. -0.000 min
Lab File: BM050029.D
Acq: 28 Apr 2025 15:45

Tgt Ion:276 Resp: 3363778

Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.1	28.7
138	17.3	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050030.D
Acq On : 28 Apr 2025 16:24
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Quant Time: Apr 28 17:58:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	238160	20.000	ng		0.00
21) Naphthalene-d8	10.563	136	873845	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	580977	20.000	ng		0.00
64) Phenanthrene-d10	17.157	188	1144389	20.000	ng		0.00
76) Chrysene-d12	21.398	240	1144881	20.000	ng		0.00
86) Perylene-d12	24.397	264	1099459	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.357	112	1676078	123.234	ng		0.00
7) Phenol-d6	6.951	99	2135178	124.905	ng		0.00
23) Nitrobenzene-d5	8.928	82	2186546	123.300	ng		0.00
42) 2,4,6-Tribromophenol	15.910	330	1107899	128.968	ng		0.00
45) 2-Fluorobiphenyl	13.033	172	6079094	123.782	ng		0.00
79) Terphenyl-d14	19.780	244	10017186	129.471	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.252	88	350690	60.136	ng		99
3) Pyridine	3.658	79	923444	61.631	ng		99
4) n-Nitrosodimethylamine	3.563	42	360090	61.276	ng		98
6) Aniline	7.098	93	1339587	62.058	ng		100
8) 2-Chlorophenol	7.340	128	905849	61.600	ng		99
9) Benzaldehyde	6.910	77	683177	59.793	ng		99
10) Phenol	6.981	94	1072175	61.956	ng		99
11) bis(2-Chloroethyl)ether	7.193	93	880705	60.947	ng		100
12) 1,3-Dichlorobenzene	7.657	146	1077725	60.670	ng		99
13) 1,4-Dichlorobenzene	7.804	146	1083367	60.707	ng		99
14) 1,2-Dichlorobenzene	8.122	146	1046214	60.691	ng		98
15) Benzyl Alcohol	8.016	79	750490	63.306	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.293	45	1052561	59.922	ng		100
17) 2-Methylphenol	8.228	107	705836	61.914	ng		99
18) Hexachloroethane	8.845	117	380742	60.018	ng		99
19) n-Nitroso-di-n-propyla...	8.575	70	678107	61.851	ng		100
20) 3+4-Methylphenols	8.551	107	964063	62.619	ng		97
22) Acetophenone	8.592	105	1325893	61.030	ng	#	99
24) Nitrobenzene	8.969	77	938300	61.193	ng		99
25) Isophorone	9.498	82	1845274	61.044	ng		99
26) 2-Nitrophenol	9.681	139	504441	64.395	ng		99
27) 2,4-Dimethylphenol	9.751	122	816717	62.869	ng		98
28) bis(2-Chloroethoxy)met...	9.975	93	1138921	60.958	ng		100
29) 2,4-Dichlorophenol	10.228	162	920741	63.266	ng		98
30) 1,2,4-Trichlorobenzene	10.428	180	1079970	60.907	ng		100
31) Naphthalene	10.610	128	2682942	60.626	ng		100
32) Benzoic acid	9.928	122	556598	67.178	ng		98
33) 4-Chloroaniline	10.728	127	1146085	63.184	ng		99
34) Hexachlorobutadiene	10.898	225	689159	61.230	ng		99
35) Caprolactam	11.528	113	266678	62.845	ng		98
36) 4-Chloro-3-methylphenol	11.869	107	832050	61.882	ng		100
37) 2-Methylnaphthalene	12.228	142	1815345	61.185	ng		99
38) 1-Methylnaphthalene	12.445	142	1916000	61.021	ng		100
40) 1,2,4,5-Tetrachloroben...	12.598	216	1264090	62.853	ng		100
41) Hexachlorocyclopentadiene	12.580	237	814238	66.248	ng		99
43) 2,4,6-Trichlorophenol	12.851	196	814256	63.832	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050030.D
Acq On : 28 Apr 2025 16:24
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Quant Time: Apr 28 17:58:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.933	196	872921	63.459	ng		99
46) 1,1'-Biphenyl	13.245	154	2626494	60.817	ng		99
47) 2-Chloronaphthalene	13.286	162	2078242	60.760	ng		100
48) 2-Nitroaniline	13.498	65	512544	63.847	ng		99
49) Acenaphthylene	14.133	152	3292902	61.060	ng		100
50) Dimethylphthalate	13.874	163	2562233	60.347	ng		100
51) 2,6-Dinitrotoluene	13.992	165	551466	62.732	ng		99
52) Acenaphthene	14.474	154	1906973	61.089	ng		100
53) 3-Nitroaniline	14.327	138	541960	64.574	ng		99
54) 2,4-Dinitrophenol	14.545	184	327924	59.847	ng		97
55) Dibenzofuran	14.810	168	3154566	60.738	ng		100
56) 4-Nitrophenol	14.663	139	402106	60.088	ng		99
57) 2,4-Dinitrotoluene	14.786	165	779604	64.180	ng		98
58) Fluorene	15.463	166	2614340	61.497	ng		100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	752680	63.213	ng		98
60) Diethylphthalate	15.233	149	2397624	59.930	ng		99
61) 4-Chlorophenyl-phenyle...	15.457	204	1457099	62.403	ng		98
62) 4-Nitroaniline	15.492	138	524887	64.776	ng		98
63) Azobenzene	15.745	77	2030208	60.550	ng		99
65) 4,6-Dinitro-2-methylph...	15.557	198	473987	66.167	ng		98
66) n-Nitrosodiphenylamine	15.668	169	2158446	61.422	ng		100
67) 4-Bromophenyl-phenylether	16.351	248	878886	63.509	ng		95
68) Hexachlorobenzene	16.468	284	1000366	62.711	ng		97
69) Atrazine	16.627	200	802538	63.312	ng		99
70) Pentachlorophenol	16.821	266	584874	65.136	ng		99
71) Phenanthrene	17.198	178	4028280	62.410	ng		99
72) Anthracene	17.292	178	4088295	62.590	ng		100
73) Carbazole	17.562	167	3562573	62.854	ng		100
74) Di-n-butylphthalate	18.121	149	4225922	62.607	ng		100
75) Fluoranthene	19.215	202	4891273	64.546	ng		100
77) Benzidine	19.404	184	2626747	64.592	ng		100
78) Pyrene	19.580	202	5030035	62.565	ng		100
80) Butylbenzylphthalate	20.474	149	1873444	62.216	ng		96
81) Benzo(a)anthracene	21.380	228	4954357	62.841	ng		99
82) 3,3'-Dichlorobenzidine	21.303	252	2003019	66.368	ng		99
83) Chrysene	21.445	228	4549388	62.348	ng		99
84) Bis(2-ethylhexyl)phtha...	21.298	149	2797195	62.192	ng	#	97
85) Di-n-octyl phthalate	22.427	149	4605837	62.095	ng		100
87) Indeno(1,2,3-cd)pyrene	27.797	276	5248662	64.068	ng		99
88) Benzo(b)fluoranthene	23.456	252	4643026	64.936	ng		99
89) Benzo(k)fluoranthene	23.515	252	4561134	63.050	ng		100
90) Benzo(a)pyrene	24.262	252	4281665	63.936	ng		100
91) Dibenzo(a,h)anthracene	27.850	278	4305928	64.478	ng		99
92) Benzo(g,h,i)perylene	28.850	276	4036958	63.142	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050030.D
Acq On : 28 Apr 2025 16:24
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Quant Time: Apr 28 17:58:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050030.D
Acq On : 28 Apr 2025 16:24
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Quant Time: Apr 28 17:58:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 238160

Ion	Ratio	Lower	Upper
152	100		
150	155.4	124.1	186.1
115	52.0	41.2	61.8





#2
1,4-Dioxane
Concen: 60.136 ng
RT: 3.252 min Scan# 45
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 88 Resp: 350690

Ion	Ratio	Lower	Upper
88	100		
58	68.4	55.1	82.7
43	25.4	20.9	31.3





#3
Pyridine
Concen: 61.631 ng
RT: 3.658 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 923444

Ion	Ratio	Lower	Upper
79	100		
52	61.2	48.6	72.8
51	31.7	24.6	37.0





#4
n-Nitrosodimethylamine
Concen: 61.276 ng
RT: 3.563 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 42 Resp: 360090

Ion	Ratio	Lower	Upper
42	100		
74	153.4	120.2	180.2
44	7.1	6.2	9.2





#5
2-Fluorophenol
Concen: 123.234 ng
RT: 5.357 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:112 Resp: 1676078

Ion	Ratio	Lower	Upper
112	100		
64	52.8	42.6	64.0
63	27.7	22.2	33.4





#6
Aniline
Concen: 62.058 ng
RT: 7.098 min Scan# 699
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 93 Resp: 1339587

Ion	Ratio	Lower	Upper
93	100		
66	37.4	29.9	44.9
65	18.5	14.5	21.7





#7
Phenol-d6
Concen: 124.905 ng
RT: 6.951 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 99 Resp: 2135178

Ion	Ratio	Lower	Upper
99	100		
42	15.6	13.0	19.4
71	31.8	25.6	38.4





#8
2-Chlorophenol
Concen: 61.600 ng
RT: 7.340 min Scan# 740
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:128 Resp: 905849

Ion	Ratio	Lower	Upper
128	100		
130	33.5	12.8	52.8
64	41.7	21.2	61.2





#9
Benzaldehyde
Concen: 59.793 ng
RT: 6.910 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 683177

Ion	Ratio	Lower	Upper
77	100		
105	99.6	80.5	120.5
106	97.1	76.1	116.1





#10
Phenol
Concen: 61.956 ng
RT: 6.981 min Scan# 679
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 94 Resp: 1072175

Ion	Ratio	Lower	Upper
94	100		
65	26.9	7.6	47.6
66	35.1	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 60.947 ng
RT: 7.193 min Scan# 7
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 93 Resp: 880705

Ion	Ratio	Lower	Upper
93	100		
63	72.0	52.3	92.3
95	32.7	12.8	52.8





#12
1,3-Dichlorobenzene
Concen: 60.670 ng
RT: 7.657 min Scan# 794
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:146 Resp: 1077725

Ion	Ratio	Lower	Upper
146	100		
148	64.9	52.5	78.7
75	27.7	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 60.707 ng
RT: 7.804 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 1083367

Ion	Ratio	Lower	Upper
146	100		
148	64.2	52.2	78.2
111	35.7	29.3	43.9





#14
1,2-Dichlorobenzene
Concen: 60.691 ng
RT: 8.122 min Scan# 873
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:146 Resp: 1046214

Ion	Ratio	Lower	Upper
146	100		
148	64.4	52.4	78.6
111	37.7	31.0	46.6





#15
Benzyl Alcohol
Concen: 63.306 ng
RT: 8.016 min Scan# 8
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 750490

Ion	Ratio	Lower	Upper
79	100		
108	81.0	66.0	99.0
77	65.3	51.6	77.4





#16
2,2'-oxybis(1-Chloropropane)
Concen: 59.922 ng
RT: 8.293 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 45 Resp: 1052561

Ion	Ratio	Lower	Upper
45	100		
77	16.6	0.0	36.4
79	12.7	0.0	32.7





#17
2-Methylphenol
Concen: 61.914 ng
RT: 8.228 min Scan# 8
Delta R.T. 0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:107 Resp: 705836

Ion	Ratio	Lower	Upper
107	100		
108	114.2	91.1	136.7
77	46.3	36.2	54.4
79	43.7	35.8	53.8





#18
Hexachloroethane
Concen: 60.018 ng
RT: 8.845 min Scan# 996
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:117 Resp: 380742

Ion	Ratio	Lower	Upper
117	100		
119	95.4	77.3	115.9
201	130.1	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 61.851 ng
RT: 8.575 min Scan# 9
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 70 Resp: 678107

Ion	Ratio	Lower	Upper
70	100		
42	50.1	40.5	60.7
101	9.7	7.8	11.6
130	21.9	17.4	26.2





#20
3+4-Methylphenols
Concen: 62.619 ng
RT: 8.551 min Scan# 946
Delta R.T. -0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:107 Resp: 964063

Ion	Ratio	Lower	Upper
107	100		
108	89.8	72.4	112.4
77	32.4	13.6	53.6
79	27.0	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:136 Resp: 873845

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	6.7	5.1	7.7
68	5.6	4.4	6.6





#22
Acetophenone
Concen: 61.030 ng
RT: 8.592 min Scan# 953
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:105 Resp: 1325893

Ion	Ratio	Lower	Upper
105	100		
71	2.8	1.7	2.5#
51	23.7	18.6	27.8
120	23.8	18.3	27.5





#23
Nitrobenzene-d5
Concen: 123.300 ng
RT: 8.928 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 2186546

Ion	Ratio	Lower	Upper
82	100		
128	45.3	35.7	53.5
54	43.6	35.4	53.2





#24
Nitrobenzene
Concen: 61.193 ng
RT: 8.969 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 77 Resp: 938300

Ion	Ratio	Lower	Upper
77	100		
123	47.6	37.9	56.9
65	13.2	10.1	15.1





#25
Isophorone
Concen: 61.044 ng
RT: 9.498 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 1845274

Ion	Ratio	Lower	Upper
82	100		
95	6.7	5.7	8.5
138	19.2	15.5	23.3





#26
2-Nitrophenol
Concen: 64.395 ng
RT: 9.681 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:139 Resp: 504441

Ion	Ratio	Lower	Upper
139	100		
109	26.8	21.9	32.9
65	43.9	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 62.869 ng
RT: 9.751 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:122 Resp: 816717

Ion	Ratio	Lower	Upper
122	100		
107	114.4	89.0	133.6
121	58.0	45.6	68.4





#28
bis(2-Chloroethoxy)methane
Concen: 60.958 ng
RT: 9.975 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 93 Resp: 1138921

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.1	39.1
123	11.8	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 63.266 ng
RT: 10.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 920741

Ion	Ratio	Lower	Upper
162	100		
164	63.9	45.4	85.4
98	29.8	11.3	51.3





#30
1,2,4-Trichlorobenzene
Concen: 60.907 ng
RT: 10.428 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:180 Resp: 1079970

Ion	Ratio	Lower	Upper
180	100		
182	96.5	77.0	115.4
145	25.2	20.6	30.8





#31
Naphthalene
Concen: 60.626 ng
RT: 10.610 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 2682942

Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.8	13.2
127	13.0	10.4	15.6





#32
Benzoic acid
Concen: 67.178 ng
RT: 9.928 min Scan# 1180
Delta R.T. 0.018 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:122 Resp: 556598

Ion	Ratio	Lower	Upper
122	100		
105	120.9	103.6	143.6
77	92.5	73.7	113.7





#33
4-Chloroaniline
Concen: 63.184 ng
RT: 10.728 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 1146085

Ion	Ratio	Lower	Upper
127	100		
129	32.7	26.3	39.5
65	27.6	22.5	33.7
92	17.5	14.5	21.7





#34
Hexachlorobutadiene
Concen: 61.230 ng
RT: 10.898 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:225 Resp: 689159

Ion	Ratio	Lower	Upper
225	100		
223	62.1	50.2	75.2
227	64.5	50.6	76.0





#35
Caprolactam
Concen: 62.845 ng
RT: 11.528 min Scan# 14
Delta R.T. 0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:113 Resp: 266678

Ion	Ratio	Lower	Upper
113	100		
55	145.4	122.9	162.9
56	110.1	88.1	128.1





#36
4-Chloro-3-methylphenol
Concen: 61.882 ng
RT: 11.869 min Scan# 1510
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:107 Resp: 832050

Ion	Ratio	Lower	Upper
107	100		
144	29.4	23.8	35.6
142	89.9	71.9	107.9





#37
2-Methylnaphthalene
Concen: 61.185 ng
RT: 12.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:142 Resp: 1815345

Ion	Ratio	Lower	Upper
142	100		
141	87.4	70.4	105.6
115	31.5	25.4	38.2





#38
1-Methylnaphthalene
Concen: 61.021 ng
RT: 12.445 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:142 Resp: 1916000

Ion	Ratio	Lower	Upper
142	100		
141	90.4	72.1	108.1
116	3.5	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:164 Resp: 580977

Ion	Ratio	Lower	Upper
164	100		
162	98.6	79.4	119.0
160	45.7	36.2	54.4





#40
1,2,4,5-Tetrachlorobenzene
Concen: 62.853 ng
RT: 12.598 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:216 Resp: 1264090

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.3	93.5
179	18.9	15.4	23.0
108	14.4	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 66.248 ng
RT: 12.580 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 814238

Ion	Ratio	Lower	Upper
237	100		
235	62.0	42.9	82.9
272	13.1	0.0	32.9





#42
2,4,6-Tribromophenol
Concen: 128.968 ng
RT: 15.910 min Scan# 2197
Delta R.T. 0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:330 Resp: 1107899

Ion	Ratio	Lower	Upper
330	100		
332	96.8	77.3	115.9
141	26.6	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 63.832 ng
RT: 12.851 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:196 Resp: 814256

Ion	Ratio	Lower	Upper
196	100		
198	97.5	79.0	118.6
200	31.9	25.1	37.7





#44
2,4,5-Trichlorophenol
Concen: 63.459 ng
RT: 12.933 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:196 Resp: 872921

Ion	Ratio	Lower	Upper
196	100		
198	96.7	76.7	115.1
97	37.5	30.1	45.1
132	20.9	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 123.782 ng
RT: 13.033 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:172 Resp: 6079094

Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.8	41.6
170	23.2	18.6	28.0





#46
1,1'-Biphenyl
Concen: 60.817 ng
RT: 13.245 min Scan# 1744
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:154 Resp: 2626494

Ion	Ratio	Lower	Upper
154	100		
153	39.2	20.0	60.0
76	9.9	0.0	30.3





#47
2-Chloronaphthalene
Concen: 60.760 ng
RT: 13.286 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 2078242

Ion	Ratio	Lower	Upper
162	100		
127	32.1	25.9	38.9
164	33.5	26.6	40.0





#48
2-Nitroaniline
Concen: 63.847 ng
RT: 13.498 min Scan# 1787
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion: 65 Resp: 512544

Ion	Ratio	Lower	Upper
65	100		
92	71.9	57.1	85.7
138	122.3	97.0	145.4





#49
Acenaphthylene
Concen: 61.060 ng
RT: 14.133 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 3292902

Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.2	10.6	15.8





#50
Dimethylphthalate
Concen: 60.347 ng
RT: 13.874 min Scan# 1851
Delta R.T. 0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:163 Resp: 2562233

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.2
164	10.2	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 62.732 ng
RT: 13.992 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 551466

Ion	Ratio	Lower	Upper
165	100		
63	45.6	35.1	52.7
89	48.1	38.4	57.6





#52
Acenaphthene
Concen: 61.089 ng
RT: 14.474 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:154 Resp: 1906973

Ion	Ratio	Lower	Upper
154	100		
153	111.1	89.0	133.4
152	53.3	43.0	64.6





#53
3-Nitroaniline
Concen: 64.574 ng
RT: 14.327 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:138 Resp: 541960

Ion	Ratio	Lower	Upper
138	100		
108	9.0	7.4	11.2
92	104.3	84.0	126.0





#54
2,4-Dinitrophenol
Concen: 59.847 ng
RT: 14.545 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:184 Resp: 327924

Ion	Ratio	Lower	Upper
184	100		
63	46.9	39.8	59.6
154	56.0	46.1	69.1





#55
Dibenzofuran
Concen: 60.738 ng
RT: 14.810 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:168 Resp: 3154566

Ion	Ratio	Lower	Upper
168	100		
139	32.6	26.1	39.1
169	13.6	10.8	16.2





#56
4-Nitrophenol
Concen: 60.088 ng
RT: 14.663 min Scan# 1985
Delta R.T. -0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:139 Resp: 402106

Ion	Ratio	Lower	Upper
139	100		
109	67.5	47.0	87.0
65	83.0	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 64.180 ng
RT: 14.786 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 779604

Ion	Ratio	Lower	Upper
165	100		
63	35.1	29.0	43.6
89	60.3	49.4	74.2
182	2.6	2.2	3.2





#58
Fluorene
Concen: 61.497 ng
RT: 15.463 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:166 Resp: 2614340

Ion	Ratio	Lower	Upper
166	100		
165	96.0	76.6	115.0
167	13.3	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 63.213 ng
RT: 15.045 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:232 Resp: 752680

Ion	Ratio	Lower	Upper
232	100		
131	34.1	28.2	42.4
130	1.8	1.5	2.3
166	27.0	22.6	34.0





#60
Diethylphthalate
Concen: 59.930 ng
RT: 15.233 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:149 Resp: 2397624

Ion	Ratio	Lower	Upper
149	100		
177	24.5	19.1	28.7
150	12.5	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 62.403 ng
RT: 15.457 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:204 Resp: 1457099

Ion	Ratio	Lower	Upper
204	100		
206	32.8	26.3	39.5
141	40.0	34.0	51.0





#62
4-Nitroaniline
Concen: 64.776 ng
RT: 15.492 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:138 Resp: 524887

Ion	Ratio	Lower	Upper
138	100		
92	46.5	25.9	65.9
108	74.6	52.4	92.4





#63
Azobenzene
Concen: 60.550 ng
RT: 15.745 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 2030208

Ion	Ratio	Lower	Upper
77	100		
182	31.1	10.5	50.5
105	21.4	1.1	41.1
51	25.9	6.3	46.3





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.157 min Scan# 2409
Delta R.T. -0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:188 Resp: 1144389

Ion	Ratio	Lower	Upper
188	100		
94	6.9	5.7	8.5
80	7.3	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 66.167 ng
RT: 15.557 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 473987

Ion	Ratio	Lower	Upper
198	100		
51	30.1	10.4	50.4
105	33.1	14.6	54.6





#66
n-Nitrosodiphenylamine
Concen: 61.422 ng
RT: 15.668 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:169 Resp: 2158446

Ion	Ratio	Lower	Upper
169	100		
168	66.9	53.4	80.2
167	36.6	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 63.509 ng
RT: 16.351 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 878886

Ion	Ratio	Lower	Upper
248	100		
250	97.8	76.2	114.2
141	50.9	45.9	68.9





#68
Hexachlorobenzene
Concen: 62.711 ng
RT: 16.468 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:284 Resp: 1000366

Ion	Ratio	Lower	Upper
284	100		
142	27.0	23.9	35.9
249	30.2	24.5	36.7





#69
Atrazine
Concen: 63.312 ng
RT: 16.627 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:200 Resp: 802538

Ion	Ratio	Lower	Upper
200	100		
173	24.2	4.3	44.3
215	51.0	30.6	70.6





#70
Pentachlorophenol
Concen: 65.136 ng
RT: 16.821 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:266 Resp: 584874

Ion	Ratio	Lower	Upper
266	100		
268	64.5	51.2	76.8
264	62.6	49.6	74.4





#71
Phenanthrene
Concen: 62.410 ng
RT: 17.198 min Scan# 24
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:178 Resp: 4028280

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.3	22.9
179	15.7	12.4	18.6





#72
Anthracene
Concen: 62.590 ng
RT: 17.292 min Scan# 2432
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:178 Resp: 4088295

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.4
179	15.5	12.2	18.4





#73
Carbazole
Concen: 62.854 ng
RT: 17.562 min Scan# 24
Delta R.T. -0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:167 Resp: 3562573

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.2	25.8
139	11.3	9.0	13.6





#74
Di-n-butylphthalate
Concen: 62.607 ng
RT: 18.121 min Scan# 2573
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:149 Resp: 4225922

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	4.5	3.8	5.6





#75
Fluoranthene
Concen: 64.546 ng
RT: 19.215 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:202 Resp: 4891273

Ion	Ratio	Lower	Upper
202	100		
101	7.6	0.0	28.0
203	17.8	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.398 min Scan# 3130
Delta R.T. 0.001 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:240 Resp: 1144881

Ion	Ratio	Lower	Upper
240	100		
120	7.8	6.5	9.7
236	26.1	20.9	31.3





#77
Benzidine
Concen: 64.592 ng
RT: 19.404 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:184 Resp: 2626747

Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.5	17.3
183	11.2	8.9	13.3





#78
Pyrene
Concen: 62.565 ng
RT: 19.580 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:202 Resp: 5030035

Ion	Ratio	Lower	Upper
202	100		
200	21.2	16.7	25.1
203	17.6	14.1	21.1





#79
Terphenyl-d14
Concen: 129.471 ng
RT: 19.780 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:244 Resp:10017186

Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.6	8.4
122	6.6	5.6	8.4





#80
Butylbenzylphthalate
Concen: 62.216 ng
RT: 20.474 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:149 Resp: 1873444

Ion	Ratio	Lower	Upper
149	100		
91	62.8	47.4	71.2
206	25.0	19.4	29.0





#81
Benzo(a)anthracene
Concen: 62.841 ng
RT: 21.380 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:228 Resp: 4954357

Ion	Ratio	Lower	Upper
228	100		
226	27.9	22.1	33.1
229	20.2	15.7	23.5





#82
3,3'-Dichlorobenzidine
Concen: 66.368 ng
RT: 21.303 min Scan# 3114
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:252 Resp: 2003019

Ion	Ratio	Lower	Upper
252	100		
254	63.6	51.8	77.6
126	6.7	6.0	9.0





#83
Chrysene
Concen: 62.348 ng
RT: 21.445 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:228 Resp: 4549388

Ion	Ratio	Lower	Upper
228	100		
226	31.0	24.2	36.2
229	19.7	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 62.192 ng
RT: 21.298 min Scan# 3113
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:149 Resp: 2797195

Ion	Ratio	Lower	Upper
149	100		
167	31.6	24.1	36.1
279	6.0	3.9	5.9#





#85
Di-n-octyl phthalate
Concen: 62.095 ng
RT: 22.427 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 4605837

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.1	5.8	8.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.397 min Scan# 3640
Delta R.T. -0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:264 Resp: 1099459

Ion	Ratio	Lower	Upper
264	100		
260	24.1	18.8	28.2
265	21.8	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 64.068 ng
RT: 27.797 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:276 Resp: 5248662

Ion	Ratio	Lower	Upper
276	100		
138	19.9	16.4	24.6
277	25.1	19.8	29.8





#88
Benzo(b)fluoranthene
Concen: 64.936 ng
RT: 23.456 min Scan# 3480
Delta R.T. 0.006 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:252 Resp: 4643026

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	6.9	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 63.050 ng
RT: 23.515 min Scan# 34
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 4561134

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.2
125	7.0	5.8	8.8





#90
Benzo(a)pyrene
Concen: 63.936 ng
RT: 24.262 min Scan# 3617
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:252 Resp: 4281665

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	8.0	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 64.478 ng
RT: 27.850 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Instrument :
BNA_M
ClientSampleId :
SSTDICC060

Tgt Ion:278 Resp: 4305928

Ion	Ratio	Lower	Upper
278	100		
139	12.3	10.5	15.7
279	23.6	19.4	29.0





#92
Benzo(g,h,i)perylene
Concen: 63.142 ng
RT: 28.850 min Scan# 4397
Delta R.T. 0.000 min
Lab File: BM050030.D
Acq: 28 Apr 2025 16:24

Tgt Ion:276 Resp: 4036958

Ion	Ratio	Lower	Upper
276	100		
277	24.1	19.1	28.7
138	17.5	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050031.D
Acq On : 28 Apr 2025 17:04
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Quant Time: Apr 28 17:59:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.769	152	251380	20.000	ng		0.00
21) Naphthalene-d8	10.563	136	931435	20.000	ng		0.00
39) Acenaphthene-d10	14.410	164	613490	20.000	ng		0.00
64) Phenanthrene-d10	17.156	188	1212251	20.000	ng		0.00
76) Chrysene-d12	21.397	240	1197780	20.000	ng		0.00
86) Perylene-d12	24.397	264	1141905	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.357	112	2297313	160.027	ng		0.00
7) Phenol-d6	6.957	99	2954036	163.719	ng		0.00
23) Nitrobenzene-d5	8.933	82	2984192	157.875	ng		0.00
42) 2,4,6-Tribromophenol	15.909	330	1586083	174.847	ng		0.00
45) 2-Fluorobiphenyl	13.033	172	8283812	159.736	ng		0.00
79) Terphenyl-d14	19.780	244	11877110	146.730	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.251	88	475912	77.317	ng		99
3) Pyridine	3.657	79	1267903	80.171	ng		99
4) n-Nitrosodimethylamine	3.563	42	492501	79.400	ng		99
6) Aniline	7.098	93	1845033	80.979	ng		99
8) 2-Chlorophenol	7.339	128	1241711	79.998	ng		100
9) Benzaldehyde	6.910	77	898993	74.544	ng		99
10) Phenol	6.981	94	1491194	81.638	ng		98
11) bis(2-Chloroethyl)ether	7.192	93	1224517	80.283	ng		99
12) 1,3-Dichlorobenzene	7.657	146	1460820	77.912	ng		100
13) 1,4-Dichlorobenzene	7.804	146	1465881	77.822	ng		99
14) 1,2-Dichlorobenzene	8.122	146	1412270	77.618	ng		99
15) Benzyl Alcohol	8.016	79	1055328	84.339	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.292	45	1440484	77.694	ng		100
17) 2-Methylphenol	8.228	107	980375	81.473	ng		99
18) Hexachloroethane	8.845	117	513126	76.632	ng		98
19) n-Nitroso-di-n-propyla...	8.581	70	933468	80.665	ng		99
20) 3+4-Methylphenols	8.557	107	1348190	82.965	ng		99
22) Acetophenone	8.592	105	1815657	78.406	ng	#	99
24) Nitrobenzene	8.975	77	1289068	78.871	ng		99
25) Isophorone	9.504	82	2559279	79.430	ng		100
26) 2-Nitrophenol	9.680	139	710174	85.053	ng		99
27) 2,4-Dimethylphenol	9.751	122	1136040	82.043	ng		99
28) bis(2-Chloroethoxy)met...	9.975	93	1581985	79.437	ng		99
29) 2,4-Dichlorophenol	10.222	162	1275806	82.243	ng		99
30) 1,2,4-Trichlorobenzene	10.427	180	1485344	78.590	ng		99
31) Naphthalene	10.610	128	3686066	78.143	ng		100
32) Benzoic acid	9.951	122	800660	90.660	ng		97
33) 4-Chloroaniline	10.727	127	1575909	81.508	ng		99
34) Hexachlorobutadiene	10.898	225	959905	80.012	ng		99
35) Caprolactam	11.539	113	371164	82.060	ng		98
36) 4-Chloro-3-methylphenol	11.869	107	1152624	80.424	ng		99
37) 2-Methylnaphthalene	12.227	142	2525532	79.858	ng		99
38) 1-Methylnaphthalene	12.451	142	2645483	79.045	ng		99
40) 1,2,4,5-Tetrachloroben...	12.604	216	1799698	84.742	ng		99
41) Hexachlorocyclopentadiene	12.580	237	1185418	91.336	ng		99
43) 2,4,6-Trichlorophenol	12.851	196	1151641	85.496	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050031.D
Acq On : 28 Apr 2025 17:04
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Quant Time: Apr 28 17:59:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.933	196	1220306	84.011	ng		98
46) 1,1'-Biphenyl	13.245	154	3613893	79.246	ng		99
47) 2-Chloronaphthalene	13.286	162	2842514	78.700	ng		99
48) 2-Nitroaniline	13.498	65	701256	82.725	ng		99
49) Acenaphthylene	14.133	152	4526874	79.492	ng		99
50) Dimethylphthalate	13.874	163	3507333	78.228	ng		99
51) 2,6-Dinitrotoluene	13.992	165	758296	81.689	ng		99
52) Acenaphthene	14.480	154	2618967	79.452	ng		100
53) 3-Nitroaniline	14.327	138	740202	83.521	ng		100
54) 2,4-Dinitrophenol	14.545	184	469472	79.096	ng		98
55) Dibenzofuran	14.815	168	4325783	78.875	ng		99
56) 4-Nitrophenol	14.662	139	566943	78.710	ng		99
57) 2,4-Dinitrotoluene	14.786	165	1080118	84.207	ng		96
58) Fluorene	15.462	166	3565506	79.426	ng		100
59) 2,3,4,6-Tetrachlorophenol	15.045	232	1057707	84.123	ng		96
60) Diethylphthalate	15.239	149	3257954	77.118	ng		97
61) 4-Chlorophenyl-phenyle...	15.457	204	2051105	83.186	ng		96
62) 4-Nitroaniline	15.498	138	724549	84.678	ng		99
63) Azobenzene	15.751	77	2723186	76.914	ng		95
65) 4,6-Dinitro-2-methylph...	15.557	198	665798	87.741	ng		96
66) n-Nitrosodiphenylamine	15.674	169	2957293	79.443	ng		99
67) 4-Bromophenyl-phenylether	16.351	248	1225531	83.600	ng		94
68) Hexachlorobenzene	16.468	284	1412768	83.606	ng		95
69) Atrazine	16.627	200	1116818	83.174	ng		100
70) Pentachlorophenol	16.821	266	843195	88.648	ng		99
71) Phenanthrene	17.203	178	5487968	80.266	ng		99
72) Anthracene	17.292	178	5597195	80.893	ng		99
73) Carbazole	17.568	167	4845637	80.705	ng		99
74) Di-n-butylphthalate	18.121	149	5649166	79.007	ng		99
75) Fluoranthene	19.221	202	6841999	85.234	ng		99
77) Benzidine	19.403	184	3598340	84.575	ng		100
78) Pyrene	19.580	202	6959279	82.738	ng		99
80) Butylbenzylphthalate	20.474	149	2512705	79.760	ng		97
81) Benzo(a)anthracene	21.380	228	6747718	81.808	ng		99
82) 3,3'-Dichlorobenzidine	21.303	252	2782107	88.111	ng		99
83) Chrysene	21.444	228	6207632	81.316	ng		98
84) Bis(2-ethylhexyl)phtha...	21.297	149	3665836	77.905	ng	#	96
85) Di-n-octyl phthalate	22.427	149	6109632	78.732	ng		99
87) Indeno(1,2,3-cd)pyrene	27.803	276	7285836	85.629	ng		98
88) Benzo(b)fluoranthene	23.456	252	6347901	85.480	ng		99
89) Benzo(k)fluoranthene	23.521	252	6265924	83.396	ng		99
90) Benzo(a)pyrene	24.262	252	5925652	85.195	ng		99
91) Dibenzo(a,h)anthracene	27.850	278	5981478	86.239	ng		99
92) Benzo(g,h,i)perylene	28.862	276	5537637	83.395	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050031.D
Acq On : 28 Apr 2025 17:04
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Quant Time: Apr 28 17:59:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050031.D
Acq On : 28 Apr 2025 17:04
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Quant Time: Apr 28 17:59:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 17:38:08 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 251380

Ion	Ratio	Lower	Upper
152	100		
150	151.7	124.1	186.1
115	50.6	41.2	61.8





#2
1,4-Dioxane
Concen: 77.317 ng
RT: 3.251 min Scan# 45
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 88 Resp: 475912

Ion	Ratio	Lower	Upper
88	100		
58	68.0	55.1	82.7
43	25.3	20.9	31.3





#3
Pyridine
Concen: 80.171 ng
RT: 3.657 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 1267903

Ion	Ratio	Lower	Upper
79	100		
52	61.6	48.6	72.8
51	30.8	24.6	37.0





#4
n-Nitrosodimethylamine
Concen: 79.400 ng
RT: 3.563 min Scan# 98
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 42 Resp: 492501

Ion	Ratio	Lower	Upper
42	100		
74	151.4	120.2	180.2
44	6.8	6.2	9.2





#5
2-Fluorophenol
Concen: 160.027 ng
RT: 5.357 min Scan# 4
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:112 Resp: 2297313

Ion	Ratio	Lower	Upper
112	100		
64	52.3	42.6	64.0
63	27.2	22.2	33.4





#6
Aniline
Concen: 80.979 ng
RT: 7.098 min Scan# 699
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 93 Resp: 1845033

Ion	Ratio	Lower	Upper
93	100		
66	37.2	29.9	44.9
65	19.0	14.5	21.7





#7
Phenol-d6
Concen: 163.719 ng
RT: 6.957 min Scan# 6
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 99 Resp: 2954036

Ion	Ratio	Lower	Upper
99	100		
42	15.2	13.0	19.4
71	31.1	25.6	38.4





#8
2-Chlorophenol
Concen: 79.998 ng
RT: 7.339 min Scan# 740
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:128 Resp: 1241711

Ion	Ratio	Lower	Upper
128	100		
130	32.8	12.8	52.8
64	41.7	21.2	61.2





#9
Benzaldehyde
Concen: 74.544 ng
RT: 6.910 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 898993

Ion	Ratio	Lower	Upper
77	100		
105	100.5	80.5	120.5
106	97.5	76.1	116.1





#10
Phenol
Concen: 81.638 ng
RT: 6.981 min Scan# 679
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 94 Resp: 1491194

Ion	Ratio	Lower	Upper
94	100		
65	26.6	7.6	47.6
66	34.4	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 80.283 ng
RT: 7.192 min Scan# 7
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 93 Resp: 1224517

Ion	Ratio	Lower	Upper
93	100		
63	71.8	52.3	92.3
95	31.6	12.8	52.8





#12
1,3-Dichlorobenzene
Concen: 77.912 ng
RT: 7.657 min Scan# 794
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:146 Resp: 1460820

Ion	Ratio	Lower	Upper
146	100		
148	65.3	52.5	78.7
75	27.9	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 77.822 ng
RT: 7.804 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:146 Resp: 1465881

Ion	Ratio	Lower	Upper
146	100		
148	65.6	52.2	78.2
111	36.2	29.3	43.9





#14
1,2-Dichlorobenzene
Concen: 77.618 ng
RT: 8.122 min Scan# 873
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:146 Resp: 1412270

Ion	Ratio	Lower	Upper
146	100		
148	65.2	52.4	78.6
111	37.5	31.0	46.6





#15
Benzyl Alcohol
Concen: 84.339 ng
RT: 8.016 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 1055328

Ion	Ratio	Lower	Upper
79	100		
108	81.2	66.0	99.0
77	64.3	51.6	77.4





#16
2,2'-oxybis(1-Chloropropane)
Concen: 77.694 ng
RT: 8.292 min Scan# 902
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 45 Resp: 1440484

Ion	Ratio	Lower	Upper
45	100		
77	16.4	0.0	36.4
79	12.7	0.0	32.7





#17
2-Methylphenol
Concen: 81.473 ng
RT: 8.228 min Scan# 8
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:107 Resp: 980375

Ion	Ratio	Lower	Upper
107	100		
108	114.4	91.1	136.7
77	45.1	36.2	54.4
79	43.7	35.8	53.8





#18
Hexachloroethane
Concen: 76.632 ng
RT: 8.845 min Scan# 996
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:117 Resp: 513126

Ion	Ratio	Lower	Upper
117	100		
119	96.5	77.3	115.9
201	132.3	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 80.665 ng
RT: 8.581 min Scan# 9
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 70 Resp: 933468

Ion	Ratio	Lower	Upper
70	100		
42	49.9	40.5	60.7
101	10.1	7.8	11.6
130	22.3	17.4	26.2





#20
3+4-Methylphenols
Concen: 82.965 ng
RT: 8.557 min Scan# 947
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:107 Resp: 1348190

Ion	Ratio	Lower	Upper
107	100		
108	92.3	72.4	112.4
77	32.6	13.6	53.6
79	28.3	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:136 Resp: 931435

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.9	13.3
54	6.8	5.1	7.7
68	5.6	4.4	6.6





#22
Acetophenone
Concen: 78.406 ng
RT: 8.592 min Scan# 953
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:105 Resp: 1815657

Ion	Ratio	Lower	Upper
105	100		
71	4.4	1.7	2.5#
51	23.0	18.6	27.8
120	23.5	18.3	27.5





#23
Nitrobenzene-d5
Concen: 157.875 ng
RT: 8.933 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 2984192

Ion	Ratio	Lower	Upper
82	100		
128	46.8	35.7	53.5
54	43.1	35.4	53.2





#24
Nitrobenzene
Concen: 78.871 ng
RT: 8.975 min Scan# 1018
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 77 Resp: 1289068

Ion	Ratio	Lower	Upper
77	100		
123	48.5	37.9	56.9
65	12.7	10.1	15.1





#25
Isophorone
Concen: 79.430 ng
RT: 9.504 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 2559279

Ion	Ratio	Lower	Upper
82	100		
95	6.7	5.7	8.5
138	19.3	15.5	23.3





#26
2-Nitrophenol
Concen: 85.053 ng
RT: 9.680 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:139 Resp: 710174

Ion	Ratio	Lower	Upper
139	100		
109	26.8	21.9	32.9
65	44.0	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 82.043 ng
RT: 9.751 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:122 Resp: 1136040

Ion	Ratio	Lower	Upper
122	100		
107	112.1	89.0	133.6
121	56.7	45.6	68.4





#28
bis(2-Chloroethoxy)methane
Concen: 79.437 ng
RT: 9.975 min Scan# 1188
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 93 Resp: 1581985

Ion	Ratio	Lower	Upper
93	100		
95	32.2	26.1	39.1
123	12.2	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 82.243 ng
RT: 10.222 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 1275806

Ion	Ratio	Lower	Upper
162	100		
164	64.4	45.4	85.4
98	31.4	11.3	51.3



#30
1,2,4-Trichlorobenzene
Concen: 78.590 ng
RT: 10.427 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:180 Resp: 1485344

Ion	Ratio	Lower	Upper
180	100		
182	97.5	77.0	115.4
145	25.6	20.6	30.8



#31
Naphthalene
Concen: 78.143 ng
RT: 10.610 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:128 Resp: 3686066

Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.8	13.2
127	13.0	10.4	15.6





#32
Benzoic acid
Concen: 90.660 ng
RT: 9.951 min Scan# 1184
Delta R.T. 0.041 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:122 Resp: 800660

Ion	Ratio	Lower	Upper
122	100		
105	119.6	103.6	143.6
77	90.5	73.7	113.7





#33
4-Chloroaniline
Concen: 81.508 ng
RT: 10.727 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:127 Resp: 1575909

Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.3	39.5
65	27.3	22.5	33.7
92	17.9	14.5	21.7





#34
Hexachlorobutadiene
Concen: 80.012 ng
RT: 10.898 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:225 Resp: 959905

Ion	Ratio	Lower	Upper
225	100		
223	62.3	50.2	75.2
227	64.6	50.6	76.0





#35
Caprolactam
Concen: 82.060 ng
RT: 11.539 min Scan# 14
Delta R.T. 0.018 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:113 Resp: 371164

Ion	Ratio	Lower	Upper
113	100		
55	145.4	122.9	162.9
56	110.6	88.1	128.1





#36
4-Chloro-3-methylphenol
Concen: 80.424 ng
RT: 11.869 min Scan# 1510
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:107 Resp: 1152624

Ion	Ratio	Lower	Upper
107	100		
144	28.8	23.8	35.6
142	88.7	71.9	107.9





#37
2-Methylnaphthalene
Concen: 79.858 ng
RT: 12.227 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:142 Resp: 2525532

Ion	Ratio	Lower	Upper
142	100		
141	87.3	70.4	105.6
115	31.2	25.4	38.2





#38
1-Methylnaphthalene
Concen: 79.045 ng
RT: 12.451 min Scan# 1609
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:142 Resp: 2645483

Ion	Ratio	Lower	Upper
142	100		
141	91.3	72.1	108.1
116	3.4	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:164 Resp: 613490

Ion	Ratio	Lower	Upper
164	100		
162	100.1	79.4	119.0
160	45.0	36.2	54.4





#40
1,2,4,5-Tetrachlorobenzene
Concen: 84.742 ng
RT: 12.604 min Scan# 1635
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:216 Resp: 1799698

Ion	Ratio	Lower	Upper
216	100		
214	78.4	62.3	93.5
179	18.4	15.4	23.0
108	13.6	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 91.336 ng
RT: 12.580 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:237 Resp: 1185418

Ion	Ratio	Lower	Upper
237	100		
235	61.8	42.9	82.9
272	13.1	0.0	32.9





#42
2,4,6-Tribromophenol
Concen: 174.847 ng
RT: 15.909 min Scan# 2197
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:330 Resp: 1586083

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.3	115.9
141	24.7	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 85.496 ng
RT: 12.851 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:196 Resp: 1151641

Ion	Ratio	Lower	Upper
196	100		
198	98.2	79.0	118.6
200	31.7	25.1	37.7



#44
2,4,5-Trichlorophenol
Concen: 84.011 ng
RT: 12.933 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:196 Resp: 1220306

Ion	Ratio	Lower	Upper
196	100		
198	97.9	76.7	115.1
97	36.2	30.1	45.1
132	20.2	16.6	25.0



#45
2-Fluorobiphenyl
Concen: 159.736 ng
RT: 13.033 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:172 Resp: 8283812

Ion	Ratio	Lower	Upper
172	100		
171	35.0	27.8	41.6
170	23.5	18.6	28.0





#46
1,1'-Biphenyl
Concen: 79.246 ng
RT: 13.245 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:154 Resp: 3613893

Ion	Ratio	Lower	Upper
154	100		
153	39.2	20.0	60.0
76	9.8	0.0	30.3





#47
2-Chloronaphthalene
Concen: 78.700 ng
RT: 13.286 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 2842514

Ion	Ratio	Lower	Upper
162	100		
127	31.8	25.9	38.9
164	33.4	26.6	40.0





#48
2-Nitroaniline
Concen: 82.725 ng
RT: 13.498 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion: 65 Resp: 701256

Ion	Ratio	Lower	Upper
65	100		
92	71.8	57.1	85.7
138	122.1	97.0	145.4





#49
Acenaphthylene
Concen: 79.492 ng
RT: 14.133 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 4526874

Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.8	23.8
153	13.4	10.6	15.8





#50
Dimethylphthalate
Concen: 78.228 ng
RT: 13.874 min Scan# 1851
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:163 Resp: 3507333

Ion	Ratio	Lower	Upper
163	100		
194	4.6	3.4	5.2
164	10.4	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 81.689 ng
RT: 13.992 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 758296

Ion	Ratio	Lower	Upper
165	100		
63	44.6	35.1	52.7
89	47.4	38.4	57.6





#52
Acenaphthene
Concen: 79.452 ng
RT: 14.480 min Scan# 1954
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:154 Resp: 2618967

Ion	Ratio	Lower	Upper
154	100		
153	111.3	89.0	133.4
152	53.1	43.0	64.6





#53
3-Nitroaniline
Concen: 83.521 ng
RT: 14.327 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:138 Resp: 740202

Ion	Ratio	Lower	Upper
138	100		
108	9.4	7.4	11.2
92	105.4	84.0	126.0





#54
2,4-Dinitrophenol
Concen: 79.096 ng
RT: 14.545 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:184 Resp: 469472

Ion	Ratio	Lower	Upper
184	100		
63	48.1	39.8	59.6
154	57.0	46.1	69.1





#55
Dibenzofuran
Concen: 78.875 ng
RT: 14.815 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:168 Resp: 4325783

Ion	Ratio	Lower	Upper
168	100		
139	31.6	26.1	39.1
169	13.5	10.8	16.2





#56
4-Nitrophenol
Concen: 78.710 ng
RT: 14.662 min Scan# 1985
Delta R.T. -0.007 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:139 Resp: 566943

Ion	Ratio	Lower	Upper
139	100		
109	67.2	47.0	87.0
65	82.5	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 84.207 ng
RT: 14.786 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 1080118

Ion	Ratio	Lower	Upper
165	100		
63	33.9	29.0	43.6
89	59.2	49.4	74.2
182	2.7	2.2	3.2





#58
Fluorene
Concen: 79.426 ng
RT: 15.462 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:166 Resp: 3565506

Ion	Ratio	Lower	Upper
166	100		
165	96.0	76.6	115.0
167	13.6	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 84.123 ng
RT: 15.045 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:232 Resp: 1057707

Ion	Ratio	Lower	Upper
232	100		
131	32.8	28.2	42.4
130	1.7	1.5	2.3
166	26.5	22.6	34.0





#60
Diethylphthalate
Concen: 77.118 ng
RT: 15.239 min Scan# 2083
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:149 Resp: 3257954

Ion	Ratio	Lower	Upper
149	100		
177	25.8	19.1	28.7
150	12.2	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 83.186 ng
RT: 15.457 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:204 Resp: 2051105

Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.3	39.5
141	38.7	34.0	51.0





#62
4-Nitroaniline
Concen: 84.678 ng
RT: 15.498 min Scan# 2127
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:138 Resp: 724549

Ion	Ratio	Lower	Upper
138	100		
92	46.5	25.9	65.9
108	72.8	52.4	92.4





#63
Azobenzene
Concen: 76.914 ng
RT: 15.751 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 2723186

Ion	Ratio	Lower	Upper
77	100		
182	34.9	10.5	50.5
105	21.7	1.1	41.1
51	24.7	6.3	46.3





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.156 min Scan# 2409
Delta R.T. -0.001 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:188 Resp: 1212251

Ion	Ratio	Lower	Upper
188	100		
94	7.0	5.7	8.5
80	7.6	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 87.741 ng
RT: 15.557 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:198 Resp: 665798

Ion	Ratio	Lower	Upper
198	100		
51	28.8	10.4	50.4
105	32.0	14.6	54.6





#66
n-Nitrosodiphenylamine
Concen: 79.443 ng
RT: 15.674 min Scan# 2157
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:169 Resp: 2957293

Ion	Ratio	Lower	Upper
169	100		
168	66.1	53.4	80.2
167	36.3	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 83.600 ng
RT: 16.351 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:248 Resp: 1225531

Ion	Ratio	Lower	Upper
248	100		
250	98.2	76.2	114.2
141	50.1	45.9	68.9





#68
Hexachlorobenzene
Concen: 83.606 ng
RT: 16.468 min Scan# 2292
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:284 Resp: 1412768

Ion	Ratio	Lower	Upper
284	100		
142	25.5	23.9	35.9
249	29.9	24.5	36.7





#69
Atrazine
Concen: 83.174 ng
RT: 16.627 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:200 Resp: 1116818

Ion	Ratio	Lower	Upper
200	100		
173	23.9	4.3	44.3
215	50.8	30.6	70.6





#70
Pentachlorophenol
Concen: 88.648 ng
RT: 16.821 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:266 Resp: 843195

Ion	Ratio	Lower	Upper
266	100		
268	64.2	51.2	76.8
264	63.1	49.6	74.4





#71
Phenanthrene
Concen: 80.266 ng
RT: 17.203 min Scan# 24
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:178 Resp: 5487968

Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.3	22.9
179	15.6	12.4	18.6





#72
Anthracene
Concen: 80.893 ng
RT: 17.292 min Scan# 2432
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:178 Resp: 5597195

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.4
179	15.8	12.2	18.4





#73
Carbazole
Concen: 80.705 ng
RT: 17.568 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:167 Resp: 4845637

Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.2	25.8
139	10.8	9.0	13.6





#74
Di-n-butylphthalate
Concen: 79.007 ng
RT: 18.121 min Scan# 2573
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:149 Resp: 5649166

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.3	10.9
104	4.5	3.8	5.6





#75
Fluoranthene
Concen: 85.234 ng
RT: 19.221 min Scan# 2
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:202 Resp: 6841999

Ion	Ratio	Lower	Upper
202	100		
101	6.7	0.0	28.0
203	17.7	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.397 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:240 Resp: 1197780

Ion	Ratio	Lower	Upper
240	100		
120	8.1	6.5	9.7
236	25.8	20.9	31.3





#77
Benzidine
Concen: 84.575 ng
RT: 19.403 min Scan# 2
Delta R.T. -0.001 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:184 Resp: 3598340

Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.5	17.3
183	11.3	8.9	13.3





#78
Pyrene
Concen: 82.738 ng
RT: 19.580 min Scan# 2821
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:202 Resp: 6959279

Ion	Ratio	Lower	Upper
202	100		
200	21.5	16.7	25.1
203	18.0	14.1	21.1





#79
Terphenyl-d14
Concen: 146.730 ng
RT: 19.780 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:244 Resp:11877110

Ion	Ratio	Lower	Upper
244	100		
212	8.6	5.6	8.4#
122	7.3	5.6	8.4





#80
Butylbenzylphthalate
Concen: 79.760 ng
RT: 20.474 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:149 Resp: 2512705

Ion	Ratio	Lower	Upper
149	100		
91	61.7	47.4	71.2
206	25.7	19.4	29.0





#81
Benzo(a)anthracene
Concen: 81.808 ng
RT: 21.380 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:228 Resp: 6747718

Ion	Ratio	Lower	Upper
228	100		
226	28.3	22.1	33.1
229	20.3	15.7	23.5





#82
3,3'-Dichlorobenzidine
Concen: 88.111 ng
RT: 21.303 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:252 Resp: 2782107

Ion	Ratio	Lower	Upper
252	100		
254	64.6	51.8	77.6
126	6.2	6.0	9.0





#83
Chrysene
Concen: 81.316 ng
RT: 21.444 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:228 Resp: 6207632

Ion	Ratio	Lower	Upper
228	100		
226	31.2	24.2	36.2
229	20.2	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 77.905 ng
RT: 21.297 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:149 Resp: 3665836

Ion	Ratio	Lower	Upper
149	100		
167	32.3	24.1	36.1
279	6.7	3.9	5.9#





#85
Di-n-octyl phthalate
Concen: 78.732 ng
RT: 22.427 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 6109632

Ion	Ratio	Lower	Upper
149	100		
167	1.7	1.3	1.9
43	7.0	5.8	8.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.397 min Scan# 3640
Delta R.T. -0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:264 Resp: 1141905

Ion	Ratio	Lower	Upper
264	100		
260	23.8	18.8	28.2
265	22.1	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 85.629 ng
RT: 27.803 min Scan# 4
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:276 Resp: 7285836

Ion	Ratio	Lower	Upper
276	100		
138	19.2	16.4	24.6
277	25.1	19.8	29.8





#88
Benzo(b)fluoranthene
Concen: 85.480 ng
RT: 23.456 min Scan# 3480
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:252 Resp: 6347901

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	6.5	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 83.396 ng
RT: 23.521 min Scan# 34
Delta R.T. 0.006 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 6265924

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.2
125	6.4	5.8	8.8





#90
Benzo(a)pyrene
Concen: 85.195 ng
RT: 24.262 min Scan# 3617
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:252 Resp: 5925652

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.4	26.0
125	7.4	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 86.239 ng
RT: 27.850 min Scan# 4
Delta R.T. -0.000 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Instrument :
BNA_M
ClientSampleId :
SSTDICC080

Tgt Ion:278 Resp: 5981478

Ion	Ratio	Lower	Upper
278	100		
139	12.0	10.5	15.7
279	23.9	19.4	29.0





#92
Benzo(g,h,i)perylene
Concen: 83.395 ng
RT: 28.862 min Scan# 4399
Delta R.T. 0.012 min
Lab File: BM050031.D
Acq: 28 Apr 2025 17:04

Tgt Ion:276 Resp: 5537637

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.1	28.7
138	16.7	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	230127	20.000	ng	0.00
21) Naphthalene-d8	10.563	136	877692	20.000	ng	0.00
39) Acenaphthene-d10	14.410	164	605832	20.000	ng	0.00
64) Phenanthrene-d10	17.156	188	1251563	20.000	ng	0.00
76) Chrysene-d12	21.397	240	1285998	20.000	ng	0.00
86) Perylene-d12	24.397	264	1266367	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.357	112	1155553	87.928	ng	0.00
7) Phenol-d6	6.951	99	1488919	90.140	ng	0.00
23) Nitrobenzene-d5	8.928	82	1385703	77.798	ng	0.00
42) 2,4,6-Tribromophenol	15.904	330	810732	90.503	ng	0.00
45) 2-Fluorobiphenyl	13.027	172	3905385	76.259	ng	0.00
79) Terphenyl-d14	19.780	244	6863055	78.970	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.257	88	253433	44.976	ng	99
3) Pyridine	3.657	79	673723	46.534	ng	98
4) n-Nitrosodimethylamine	3.569	42	260951	45.956	ng	99
6) Aniline	7.098	93	974989	46.744	ng	98
8) 2-Chlorophenol	7.339	128	649705	45.724	ng	99
9) Benzaldehyde	6.910	77	485673m	43.937	ng	
10) Phenol	6.981	94	784469	46.913	ng	98
11) bis(2-Chloroethyl)ether	7.192	93	635601	45.521	ng	98
12) 1,3-Dichlorobenzene	7.657	146	766772	44.672	ng	99
13) 1,4-Dichlorobenzene	7.804	146	774954	44.941	ng	100
14) 1,2-Dichlorobenzene	8.122	146	752460	45.174	ng	100
15) Benzyl Alcohol	8.016	79	551757	48.167	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.298	45	762583	44.929	ng	99
17) 2-Methylphenol	8.222	107	519426	47.153	ng	99
18) Hexachloroethane	8.845	117	274239	44.738	ng	99
19) n-Nitroso-di-n-propyla...	8.575	70	507362	47.892	ng	99
20) 3+4-Methylphenols	8.551	107	713683	47.974	ng	97
22) Acetophenone	8.592	105	985275	45.153	ng	100
24) Nitrobenzene	8.969	77	694640	45.104	ng	99
25) Isophorone	9.498	82	1384372	45.596	ng	99
26) 2-Nitrophenol	9.681	139	370477	47.087	ng	99
27) 2,4-Dimethylphenol	9.751	122	605475	46.404	ng	100
28) bis(2-Chloroethoxy)met...	9.975	93	842968	44.920	ng	100
29) 2,4-Dichlorophenol	10.228	162	677885	46.375	ng	99
30) 1,2,4-Trichlorobenzene	10.428	180	789816	44.348	ng	100
31) Naphthalene	10.610	128	1984698	44.651	ng	100
32) Benzoic acid	9.916	122	384400	46.063	ng	98
33) 4-Chloroaniline	10.727	127	863111	47.375	ng	99
34) Hexachlorobutadiene	10.898	225	497792	44.034	ng	99
35) Caprolactam	11.522	113	198108	46.481	ng	95
36) 4-Chloro-3-methylphenol	11.869	107	632365	46.825	ng	100
37) 2-Methylnaphthalene	12.227	142	1352170	45.374	ng	100
38) 1-Methylnaphthalene	12.445	142	1425407	45.198	ng	100
40) 1,2,4,5-Tetrachloroben...	12.598	216	925725	44.140	ng	100
41) Hexachlorocyclopentadiene	12.574	237	568717	44.374	ng	99
43) 2,4,6-Trichlorophenol	12.851	196	613142	46.094	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.933	196	663718	46.271	ng	99
46) 1,1'-Biphenyl	13.239	154	1990852	44.207	ng	99
47) 2-Chloronaphthalene	13.286	162	1593769	44.684	ng	100
48) 2-Nitroaniline	13.498	65	398645	47.621	ng	99
49) Acenaphthylene	14.133	152	2548342	45.315	ng	100
50) Dimethylphthalate	13.868	163	2007342	45.338	ng	100
51) 2,6-Dinitrotoluene	13.992	165	435925	47.554	ng	98
52) Acenaphthene	14.474	154	1471672	45.210	ng	100
53) 3-Nitroaniline	14.327	138	427464	48.843	ng	99
54) 2,4-Dinitrophenol	14.545	184	249307	45.189	ng	98
55) Dibenzofuran	14.810	168	2427830	44.828	ng	99
56) 4-Nitrophenol	14.663	139	313065	46.012	ng	99
57) 2,4-Dinitrotoluene	14.786	165	611943	48.311	ng	97
58) Fluorene	15.462	166	2044284	46.115	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.045	232	583198	46.970	ng	98
60) Diethylphthalate	15.233	149	1889827	45.299	ng	99
61) 4-Chlorophenyl-phenyle...	15.457	204	1118837	45.950	ng	99
62) 4-Nitroaniline	15.492	138	427339	50.610	ng	100
63) Azobenzene	15.745	77	1599489	45.747	ng	100
65) 4,6-Dinitro-2-methylph...	15.557	198	367586	46.920	ng	97
66) n-Nitrosodiphenylamine	15.668	169	1709666	44.485	ng	100
67) 4-Bromophenyl-phenylether	16.345	248	665088	43.944	ng	97
68) Hexachlorobenzene	16.468	284	770651	44.174	ng	98
69) Atrazine	16.627	200	616670	44.483	ng	99
70) Pentachlorophenol	16.821	266	452303	46.059	ng	99
71) Phenanthrene	17.198	178	3169143	44.895	ng	99
72) Anthracene	17.292	178	3241829	45.381	ng	100
73) Carbazole	17.562	167	2851516	46.001	ng	100
74) Di-n-butylphthalate	18.121	149	3370238	45.654	ng	100
75) Fluoranthene	19.215	202	3895046	46.998	ng	100
77) Benzidine	19.403	184	2019202	44.204	ng	100
78) Pyrene	19.580	202	3992384	44.209	ng	100
80) Butylbenzylphthalate	20.474	149	1512498	44.718	ng	96
81) Benzo(a)anthracene	21.380	228	4029064	45.497	ng	100
82) 3,3'-Dichlorobenzidine	21.303	252	1487254	43.871	ng	98
83) Chrysene	21.439	228	3685603	44.967	ng	99
84) Bis(2-ethylhexyl)phtha...	21.297	149	2284769	45.224	ng	99
85) Di-n-octyl phthalate	22.427	149	3747091	44.974	ng	100
87) Indeno(1,2,3-cd)pyrene	27.791	276	4364032	46.249	ng	99
88) Benzo(b)fluoranthene	23.450	252	3711451	45.066	ng	100
89) Benzo(k)fluoranthene	23.515	252	3813739	45.780	ng	100
90) Benzo(a)pyrene	24.262	252	3515582	45.577	ng	100
91) Dibenzo(a,h)anthracene	27.838	278	3545295	46.091	ng	100
92) Benzo(g,h,i)perylene	28.850	276	3345997	45.437	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.769 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:152 Resp: 23012

Ion	Ratio	Lower	Upper
152	100		
150	157.2	124.1	186.1
115	52.6	41.2	61.8



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#2
1,4-Dioxane
Concen: 44.976 ng
RT: 3.257 min Scan# 46
Delta R.T. 0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 88 Resp: 253433

Ion	Ratio	Lower	Upper
88	100		
58	67.9	55.1	82.7
43	25.9	20.9	31.3





#3
Pyridine
Concen: 46.534 ng
RT: 3.657 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 79 Resp: 67372

Ion	Ratio	Lower	Upper
79	100		
52	62.0	48.6	72.8
51	31.9	24.6	37.0



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#4
n-Nitrosodimethylamine
Concen: 45.956 ng
RT: 3.569 min Scan# 99
Delta R.T. 0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 42 Resp: 260951

Ion	Ratio	Lower	Upper
42	100		
74	151.5	120.2	180.2
44	7.3	6.2	9.2





#5
2-Fluorophenol
Concen: 87.928 ng
RT: 5.357 min Scan# 4
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:112 Resp: 115555

Ion	Ratio	Lower	Upper
112	100		
64	52.4	42.6	64.0
63	28.0	22.2	33.4



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#6
Aniline
Concen: 46.744 ng
RT: 7.098 min Scan# 699
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 93 Resp: 974989

Ion	Ratio	Lower	Upper
93	100		
66	36.7	29.9	44.9
65	19.1	14.5	21.7





#7
Phenol-d6
Concen: 90.140 ng
RT: 6.951 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 99 Resp: 148891

Ion	Ratio	Lower	Upper
99	100		
42	16.0	13.0	19.4
71	31.9	25.6	38.4



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#8
2-Chlorophenol
Concen: 45.724 ng
RT: 7.339 min Scan# 740
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:128 Resp: 649705

Ion	Ratio	Lower	Upper
128	100		
130	32.0	12.8	52.8
64	41.1	21.2	61.2





#9
Benzaldehyde
Concen: 43.937 ng m
RT: 6.910 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 77 Resp: 48567

Ion	Ratio	Lower	Upper
77	100		
105	99.8	80.5	120.5
106	96.6	76.1	116.1



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#10
Phenol
Concen: 46.913 ng
RT: 6.981 min Scan# 679
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 94 Resp: 784469

Ion	Ratio	Lower	Upper
94	100		
65	26.4	7.6	47.6
66	34.6	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 45.521 ng
RT: 7.192 min Scan# 7
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 93 Resp: 63560

Ion	Ratio	Lower	Upper
93	100		
63	71.0	52.3	92.3
95	31.6	12.8	52.8



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#12
1,3-Dichlorobenzene
Concen: 44.672 ng
RT: 7.657 min Scan# 794
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:146 Resp: 766772

Ion	Ratio	Lower	Upper
146	100		
148	65.4	52.5	78.7
75	27.6	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 44.941 ng
RT: 7.804 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:146 Resp: 77495

Ion	Ratio	Lower	Upper
146	100		
148	65.0	52.2	78.2
111	36.6	29.3	43.9



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#14
1,2-Dichlorobenzene
Concen: 45.174 ng
RT: 8.122 min Scan# 873
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:146 Resp: 752460

Ion	Ratio	Lower	Upper
146	100		
148	65.3	52.4	78.6
111	38.3	31.0	46.6





#15
Benzyl Alcohol
Concen: 48.167 ng
RT: 8.016 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 79 Resp: 55175

Ion	Ratio	Lower	Upper
79	100		
108	79.4	66.0	99.0
77	64.4	51.6	77.4



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 44.929 ng
RT: 8.298 min Scan# 903
Delta R.T. 0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 45 Resp: 762583

Ion	Ratio	Lower	Upper
45	100		
77	16.7	0.0	36.4
79	13.0	0.0	32.7





#17
2-Methylphenol
Concen: 47.153 ng
RT: 8.222 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:107 Resp: 51942

Ion	Ratio	Lower	Upper
107	100		
108	114.1	91.1	136.7
77	46.5	36.2	54.4
79	44.7	35.8	53.8



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#18
Hexachloroethane
Concen: 44.738 ng
RT: 8.845 min Scan# 996
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:117 Resp: 274239

Ion	Ratio	Lower	Upper
117	100		
119	96.5	77.3	115.9
201	131.4	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 47.892 ng
RT: 8.575 min Scan# 9
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 70 Resp: 50736

Ion	Ratio	Lower	Upper
70	100		
42	49.7	40.5	60.7
101	9.7	7.8	11.6
130	21.6	17.4	26.2



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#20
3+4-Methylphenols
Concen: 47.974 ng
RT: 8.551 min Scan# 946
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:107 Resp: 713683

Ion	Ratio	Lower	Upper
107	100		
108	89.1	72.4	112.4
77	33.0	13.6	53.6
79	27.7	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.563 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:136 Resp: 87769

Ion	Ratio	Lower	Upper
136	100		
137	10.5	8.9	13.3
54	6.7	5.1	7.7
68	5.4	4.4	6.6



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#22
Acetophenone
Concen: 45.153 ng
RT: 8.592 min Scan# 953
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:105 Resp: 985275

Ion	Ratio	Lower	Upper
105	100		
71	2.0	1.7	2.5
51	23.3	18.6	27.8
120	23.0	18.3	27.5





#23
Nitrobenzene-d5
Concen: 77.798 ng
RT: 8.928 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 82 Resp: 138570

Ion	Ratio	Lower	Upper
82	100		
128	44.9	35.7	53.5
54	44.4	35.4	53.2



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#24
Nitrobenzene
Concen: 45.104 ng
RT: 8.969 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 77 Resp: 694640

Ion	Ratio	Lower	Upper
77	100		
123	47.1	37.9	56.9
65	13.3	10.1	15.1





#25
Isophorone
Concen: 45.596 ng
RT: 9.498 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 82 Resp: 138437

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.7	8.5
138	19.5	15.5	23.3



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#26
2-Nitrophenol
Concen: 47.087 ng
RT: 9.681 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:139 Resp: 370477

Ion	Ratio	Lower	Upper
139	100		
109	26.1	21.9	32.9
65	44.5	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 46.404 ng
RT: 9.751 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:122 Resp: 60547

Ion	Ratio	Lower	Upper
122	100		
107	111.0	89.0	133.6
121	56.9	45.6	68.4



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#28
bis(2-Chloroethoxy)methane
Concen: 44.920 ng
RT: 9.975 min Scan# 1188
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 93 Resp: 842968

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.1	39.1
123	11.8	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 46.375 ng
RT: 10.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:162 Resp: 67788

Ion	Ratio	Lower	Upper
162	100		
164	66.0	45.4	85.4
98	30.5	11.3	51.3

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#30
1,2,4-Trichlorobenzene
Concen: 44.348 ng
RT: 10.428 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:180 Resp: 789816

Ion	Ratio	Lower	Upper
180	100		
182	95.8	77.0	115.4
145	25.7	20.6	30.8





#31
Naphthalene
Concen: 44.651 ng
RT: 10.610 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:128 Resp: 198469

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.8	13.2
127	12.9	10.4	15.6



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#32
Benzoic acid
Concen: 46.063 ng
RT: 9.916 min Scan# 1178
Delta R.T. 0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:122 Resp: 384400

Ion	Ratio	Lower	Upper
122	100		
105	121.1	103.6	143.6
77	91.7	73.7	113.7





#33
4-Chloroaniline
Concen: 47.375 ng
RT: 10.727 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:127 Resp: 86311

Ion	Ratio	Lower	Upper
127	100		
129	32.1	26.3	39.5
65	27.9	22.5	33.7
92	17.4	14.5	21.7



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#34
Hexachlorobutadiene
Concen: 44.034 ng
RT: 10.898 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:225 Resp: 497792

Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.2	75.2
227	65.2	50.6	76.0





#35
Caprolactam
Concen: 46.481 ng
RT: 11.522 min Scan# 14
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:113 Resp: 19810

Ion	Ratio	Lower	Upper
113	100		
55	148.9	122.9	162.9
56	114.6	88.1	128.1



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#36
4-Chloro-3-methylphenol
Concen: 46.825 ng
RT: 11.869 min Scan# 1510
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:107 Resp: 632365

Ion	Ratio	Lower	Upper
107	100		
144	29.2	23.8	35.6
142	89.6	71.9	107.9





#37
2-Methylnaphthalene
Concen: 45.374 ng
RT: 12.227 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:142 Resp: 135217

Ion	Ratio	Lower	Upper
142	100		
141	88.0	70.4	105.6
115	31.8	25.4	38.2



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#38
1-Methylnaphthalene
Concen: 45.198 ng
RT: 12.445 min Scan# 1608
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:142 Resp: 1425407

Ion	Ratio	Lower	Upper
142	100		
141	90.3	72.1	108.1
116	3.7	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.410 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:164 Resp: 60583

Ion	Ratio	Lower	Upper
164	100		
162	99.4	79.4	119.0
160	44.5	36.2	54.4



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 44.140 ng
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:216 Resp: 925725

Ion	Ratio	Lower	Upper
216	100		
214	78.3	62.3	93.5
179	19.1	15.4	23.0
108	14.6	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 44.374 ng
RT: 12.574 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:237 Resp: 56871

Ion	Ratio	Lower	Upper
237	100		
235	61.6	42.9	82.9
272	12.9	0.0	32.9



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#42
2,4,6-Tribromophenol
Concen: 90.503 ng
RT: 15.904 min Scan# 2196
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:330 Resp: 810732

Ion	Ratio	Lower	Upper
330	100		
332	96.9	77.3	115.9
141	28.0	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 46.094 ng
RT: 12.851 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:196 Resp: 61314

Ion	Ratio	Lower	Upper
196	100		
198	95.9	79.0	118.6
200	32.2	25.1	37.7



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#44
2,4,5-Trichlorophenol
Concen: 46.271 ng
RT: 12.933 min Scan# 1691
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:196 Resp: 663718

Ion	Ratio	Lower	Upper
196	100		
198	96.9	76.7	115.1
97	38.0	30.1	45.1
132	20.6	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 76.259 ng
RT: 13.027 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:172 Resp: 390538

Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.8	41.6
170	23.1	18.6	28.0



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#46
1,1'-Biphenyl
Concen: 44.207 ng
RT: 13.239 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:154 Resp: 1990852

Ion	Ratio	Lower	Upper
154	100		
153	39.5	20.0	60.0
76	11.1	0.0	30.3





#47
2-Chloronaphthalene
Concen: 44.684 ng
RT: 13.286 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:162 Resp: 159376

Ion	Ratio	Lower	Upper
162	100		
127	32.2	25.9	38.9
164	33.0	26.6	40.0



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#48
2-Nitroaniline
Concen: 47.621 ng
RT: 13.498 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 65 Resp: 398645

Ion	Ratio	Lower	Upper
65	100		
92	71.5	57.1	85.7
138	122.1	97.0	145.4





#49
Acenaphthylene
Concen: 45.315 ng
RT: 14.133 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:152 Resp: 254834

Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.8
153	13.0	10.6	15.8



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#50
Dimethylphthalate
Concen: 45.338 ng
RT: 13.868 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:163 Resp: 2007342

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.2
164	10.2	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 47.554 ng
RT: 13.992 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:165 Resp: 43592

Ion	Ratio	Lower	Upper
165	100		
63	42.9	35.1	52.7
89	46.4	38.4	57.6



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#52
Acenaphthene
Concen: 45.210 ng
RT: 14.474 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:154 Resp: 1471672

Ion	Ratio	Lower	Upper
154	100		
153	110.7	89.0	133.4
152	53.5	43.0	64.6





#53
3-Nitroaniline
Concen: 48.843 ng
RT: 14.327 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:138 Resp: 42746

Ion	Ratio	Lower	Upper
138	100		
108	9.5	7.4	11.2
92	104.5	84.0	126.0



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#54
2,4-Dinitrophenol
Concen: 45.189 ng
RT: 14.545 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:184 Resp: 249307

Ion	Ratio	Lower	Upper
184	100		
63	47.3	39.8	59.6
154	56.7	46.1	69.1





#55
Dibenzofuran
Concen: 44.828 ng
RT: 14.810 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:168 Resp: 242783

Ion	Ratio	Lower	Upper
168	100		
139	33.0	26.1	39.1
169	13.5	10.8	16.2



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#56
4-Nitrophenol
Concen: 46.012 ng
RT: 14.663 min Scan# 1985
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:139 Resp: 313065

Ion	Ratio	Lower	Upper
139	100		
109	66.8	47.0	87.0
65	83.4	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 48.311 ng
RT: 14.786 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:165 Resp: 61194

Ion	Ratio	Lower	Upper
165	100		
63	34.9	29.0	43.6
89	59.6	49.4	74.2
182	2.8	2.2	3.2



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#58
Fluorene
Concen: 46.115 ng
RT: 15.462 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:166 Resp: 2044284

Ion	Ratio	Lower	Upper
166	100		
165	96.4	76.6	115.0
167	13.3	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 46.970 ng
RT: 15.045 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:232 Resp: 583198

Ion	Ratio	Lower	Upper
232	100		
131	34.5	28.2	42.4
130	1.8	1.5	2.3
166	27.1	22.6	34.0



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#60
Diethylphthalate
Concen: 45.299 ng
RT: 15.233 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:149 Resp: 1889827

Ion	Ratio	Lower	Upper
149	100		
177	24.0	19.1	28.7
150	12.4	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 45.950 ng
RT: 15.457 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:204 Resp: 111883

Ion	Ratio	Lower	Upper
204	100		
206	33.2	26.3	39.5
141	41.1	34.0	51.0



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#62
4-Nitroaniline
Concen: 50.610 ng
RT: 15.492 min Scan# 2126
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:138 Resp: 427339

Ion	Ratio	Lower	Upper
138	100		
92	45.3	25.9	65.9
108	72.5	52.4	92.4





#63
Azobenzene
Concen: 45.747 ng
RT: 15.745 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion: 77 Resp: 159948

Ion	Ratio	Lower	Upper
77	100		
182	30.5	10.5	50.5
105	21.6	1.1	41.1
51	26.1	6.3	46.3



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.156 min Scan# 2409
Delta R.T. -0.001 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:188 Resp: 1251563

Ion	Ratio	Lower	Upper
188	100		
94	7.0	5.7	8.5
80	7.7	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 46.920 ng
RT: 15.557 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:198 Resp: 36758

Ion	Ratio	Lower	Upper
198	100		
51	29.9	10.4	50.4
105	32.3	14.6	54.6



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#66
n-Nitrosodiphenylamine
Concen: 44.485 ng
RT: 15.668 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:169 Resp: 1709666

Ion	Ratio	Lower	Upper
169	100		
168	66.5	53.4	80.2
167	36.6	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 43.944 ng
RT: 16.345 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:248 Resp: 66508

Ion	Ratio	Lower	Upper
248	100		
250	99.3	76.2	114.2
141	58.1	45.9	68.9



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#68
Hexachlorobenzene
Concen: 44.174 ng
RT: 16.468 min Scan# 2292
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:284 Resp: 770651

Ion	Ratio	Lower	Upper
284	100		
142	27.6	23.9	35.9
249	30.4	24.5	36.7





#69
Atrazine
Concen: 44.483 ng
RT: 16.627 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:200 Resp: 61667

Ion	Ratio	Lower	Upper
200	100		
173	24.7	4.3	44.3
215	51.1	30.6	70.6

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#70
Pentachlorophenol
Concen: 46.059 ng
RT: 16.821 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:266 Resp: 452303

Ion	Ratio	Lower	Upper
266	100		
268	64.9	51.2	76.8
264	61.9	49.6	74.4





#71
Phenanthrene
Concen: 44.895 ng
RT: 17.198 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:178 Resp: 316914

Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.3	22.9
179	15.4	12.4	18.6



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#72
Anthracene
Concen: 45.381 ng
RT: 17.292 min Scan# 2432
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:178 Resp: 3241829

Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.0	22.4
179	15.5	12.2	18.4





#73
Carbazole
Concen: 46.001 ng
RT: 17.562 min Scan# 24
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:167 Resp: 285151

Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.2	25.8
139	11.3	9.0	13.6



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#74
Di-n-butylphthalate
Concen: 45.654 ng
RT: 18.121 min Scan# 2573
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:149 Resp: 3370238

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	4.5	3.8	5.6





#75
Fluoranthene
Concen: 46.998 ng
RT: 19.215 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:202 Resp: 389504

Ion	Ratio	Lower	Upper
202	100		
101	8.0	0.0	28.0
203	17.9	0.0	37.7



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.397 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:240 Resp: 1285998

Ion	Ratio	Lower	Upper
240	100		
120	8.0	6.5	9.7
236	25.8	20.9	31.3





#77
Benzidine
Concen: 44.204 ng
RT: 19.403 min Scan# 2
Delta R.T. -0.001 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:184 Resp: 201920

Ion	Ratio	Lower	Upper
184	100		
185	14.1	11.5	17.3
183	11.0	8.9	13.3



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#78
Pyrene
Concen: 44.209 ng
RT: 19.580 min Scan# 2821
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:202 Resp: 3992384

Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.7	25.1
203	17.7	14.1	21.1





#79
Terphenyl-d14
Concen: 78.970 ng
RT: 19.780 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:244 Resp: 686305

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.6	8.4
122	6.6	5.6	8.4



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#80
Butylbenzylphthalate
Concen: 44.718 ng
RT: 20.474 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:149 Resp: 1512498

Ion	Ratio	Lower	Upper
149	100		
91	63.0	47.4	71.2
206	23.7	19.4	29.0





#81
Benzo(a)anthracene
Concen: 45.497 ng
RT: 21.380 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:228 Resp: 402906

Ion	Ratio	Lower	Upper
228	100		
226	27.7	22.1	33.1
229	19.8	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#82
3,3'-Dichlorobenzidine
Concen: 43.871 ng
RT: 21.303 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:252 Resp: 1487254

Ion	Ratio	Lower	Upper
252	100		
254	63.3	51.8	77.6
126	7.4	6.0	9.0





#83
Chrysene
Concen: 44.967 ng
RT: 21.439 min Scan# 3
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:228 Resp: 368560

Ion	Ratio	Lower	Upper
228	100		
226	30.7	24.2	36.2
229	19.9	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 45.224 ng
RT: 21.297 min Scan# 3113
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:149 Resp: 2284769

Ion	Ratio	Lower	Upper
149	100		
167	30.7	24.1	36.1
279	5.6	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 44.974 ng
RT: 22.427 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:149 Resp: 374709

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.2	5.8	8.8

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.397 min Scan# 3640
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:264 Resp: 1266367

Ion	Ratio	Lower	Upper
264	100		
260	24.0	18.8	28.2
265	21.9	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 46.249 ng
RT: 27.791 min Scan# 4
Delta R.T. -0.006 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:276 Resp: 436403

Ion	Ratio	Lower	Upper
276	100		
138	19.8	16.4	24.6
277	25.0	19.8	29.8



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#88
Benzo(b)fluoranthene
Concen: 45.066 ng
RT: 23.450 min Scan# 3479
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:252 Resp: 3711451

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	7.2	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 45.780 ng
RT: 23.515 min Scan# 34
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:252 Resp: 381373

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.2
125	7.1	5.8	8.8



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#90
Benzo(a)pyrene
Concen: 45.577 ng
RT: 24.262 min Scan# 3617
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:252 Resp: 3515582

Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.4	26.0
125	7.9	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 46.091 ng
RT: 27.838 min Scan# 4
Delta R.T. -0.012 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:278 Resp: 354529

Ion	Ratio	Lower	Upper
278	100		
139	12.7	10.5	15.7
279	24.1	19.4	29.0



Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 04/29/2025
Supervised By :Jagrut Upadhyay 04/29/2025



#92
Benzo(g,h,i)perylene
Concen: 45.437 ng
RT: 28.850 min Scan# 4397
Delta R.T. -0.000 min
Lab File: BM050032.D
Acq: 28 Apr 2025 17:43

Tgt Ion:276 Resp: 3345997

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.1	28.7
138	17.9	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2		1,4-Dioxane	0.490	0.551	-12.4	103	0.00
3		Pyridine	1.258	1.464	-16.4	105	0.00
4		n-Nitrosodimethylamine	0.493	0.567	-15.0	104	0.00
5	S	2-Fluorophenol	1.142	1.255	-9.9	99	0.00
6		Aniline	1.813	2.118	-16.8	104	0.00
7	S	Phenol-d6	1.436	1.617	-12.6	100	0.00
8		2-Chlorophenol	1.235	1.412	-14.3	103	0.00
9		Benzaldehyde	0.961	1.055	-9.8	98	0.00
10	C	Phenol	1.453	1.704	-17.3	106	0.00
11		bis(2-Chloroethyl)ether	1.213	1.381	-13.8	103	0.00
12		1,3-Dichlorobenzene	1.492	1.666	-11.7	102	0.00
13	C	1,4-Dichlorobenzene	1.499	1.684	-12.3	103	0.00
14		1,2-Dichlorobenzene	1.448	1.635	-12.9	103	0.00
15		Benzyl Alcohol	0.996	1.199	-20.4	106	0.00
16		2,2'-oxybis(1-Chloropropane	1.475	1.657	-12.3	102	0.00
17		2-Methylphenol	0.957	1.129	-18.0	104	0.00
18		Hexachloroethane	0.533	0.596	-11.8	102	0.00
19	P	n-Nitroso-di-n-propylamine	0.921	1.102	-19.7	105	0.00
20		3+4-Methylphenols	1.293	1.551	-20.0	106	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22		Acetophenone	0.497	0.561	-12.9	104	0.00
23	S	Nitrobenzene-d5	0.406	0.395	2.7	90	0.00
24		Nitrobenzene	0.351	0.396	-12.8	104	0.00
25		Isophorone	0.692	0.789	-14.0	107	0.00
26	C	2-Nitrophenol	0.179	0.211	-17.9	106	0.00
27		2,4-Dimethylphenol	0.297	0.345	-16.2	105	0.00
28		bis(2-Chloroethoxy)methane	0.428	0.480	-12.1	104	0.00
29	C	2,4-Dichlorophenol	0.333	0.386	-15.9	105	0.00
30		1,2,4-Trichlorobenzene	0.406	0.450	-10.8	104	0.00
31		Naphthalene	1.013	1.131	-11.6	105	0.00
32		Benzoic acid	0.190	0.219	-15.3	109	0.00
33		4-Chloroaniline	0.415	0.492	-18.6	110	0.00
34	C	Hexachlorobutadiene	0.258	0.284	-10.1	103	0.00
35		Caprolactam	0.097	0.113	-16.5	107	0.00
36	C	4-Chloro-3-methylphenol	0.308	0.360	-16.9	109	0.00
37		2-Methylnaphthalene	0.679	0.770	-13.4	107	0.00
38		1-Methylnaphthalene	0.719	0.812	-12.9	106	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40		1,2,4,5-Tetrachlorobenzene	0.692	0.764	-10.4	105	0.00
41	P	Hexachlorocyclopentadiene	0.423	0.469	-10.9	106	0.00
42	S	2,4,6-Tribromophenol	0.296	0.335	-13.2	108	0.00
43	C	2,4,6-Trichlorophenol	0.439	0.506	-15.3	109	0.00
44		2,4,5-Trichlorophenol	0.474	0.548	-15.6	108	0.00
45	S	2-Fluorobiphenyl	1.691	1.612	4.7	91	0.00
46		1,1'-Biphenyl	1.487	1.643	-10.5	107	0.00
47		2-Chloronaphthalene	1.177	1.315	-11.7	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.276	0.329	-19.2	112	0.00
49		Acenaphthylene	1.857	2.103	-13.2	109	0.00
50		Dimethylphthalate	1.462	1.657	-13.3	111	0.00
51		2,6-Dinitrotoluene	0.303	0.360	-18.8	113	0.00
52	C	Acenaphthene	1.075	1.215	-13.0	110	0.00
53		3-Nitroaniline	0.289	0.353	-22.1	113	0.00
54	P	2,4-Dinitrophenol	0.164	0.206	-25.6#	119	0.00
55		Dibenzofuran	1.788	2.004	-12.1	109	0.00
56	P	4-Nitrophenol	0.205	0.258	-25.9#	118	0.00
57		2,4-Dinitrotoluene	0.418	0.505	-20.8	113	0.00
58		Fluorene	1.463	1.687	-15.3	111	0.00
59		2,3,4,6-Tetrachlorophenol	0.410	0.481	-17.3	113	0.00
60		Diethylphthalate	1.377	1.560	-13.3	111	0.00
61		4-Chlorophenyl-phenylether	0.804	0.923	-14.8	110	0.00
62		4-Nitroaniline	0.279	0.353	-26.5#	115	0.00
63		Azobenzene	1.154	1.320	-14.4	111	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
65		4,6-Dinitro-2-methylphenol	0.125	0.147	-17.6	117	0.00
66	c	n-Nitrosodiphenylamine	0.614	0.683	-11.2	112	0.00
67		4-Bromophenyl-phenylether	0.242	0.266	-9.9	111	0.00
68		Hexachlorobenzene	0.279	0.308	-10.4	113	0.00
69		Atrazine	0.222	0.246	-10.8	112	0.00
70	C	Pentachlorophenol	0.157	0.181	-15.3	117	0.00
71		Phenanthrene	1.128	1.266	-12.2	115	0.00
72		Anthracene	1.142	1.295	-13.4	115	0.00
73		Carbazole	0.991	1.139	-14.9	116	0.00
74		Di-n-butylphthalate	1.180	1.346	-14.1	116	0.00
75	C	Fluoranthene	1.324	1.556	-17.5	119	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	115	0.00
77		Benzidine	0.710	0.785	-10.6	117	0.00
78		Pyrene	1.404	1.552	-10.5	118	0.00
79	S	Terphenyl-d14	1.352	1.334	1.3	101	0.00
80		Butylbenzylphthalate	0.526	0.588	-11.8	121	0.00
81		Benzo(a)anthracene	1.377	1.567	-13.8	122	0.00
82		3,3'-Dichlorobenzidine	0.527	0.578	-9.7	115	0.00
83		Chrysene	1.275	1.433	-12.4	123	0.00
84		Bis(2-ethylhexyl)phthalate	0.786	0.888	-13.0	122	0.00
85	c	Di-n-octyl phthalate	1.296	1.457	-12.4	123	0.00
86	I	Perylene-d12	1.000	1.000	0.0	118	0.00
87		Indeno(1,2,3-cd)pyrene	1.490	1.723	-15.6	127	0.00
88		Benzo(b)fluoranthene	1.301	1.465	-12.6	123	0.00
89		Benzo(k)fluoranthene	1.316	1.506	-14.4	128	0.00
90	C	Benzo(a)pyrene	1.218	1.388	-14.0	127	0.00
91		Dibenzo(a,h)anthracene	1.215	1.400	-15.2	127	-0.01
92		Benzo(g,h,i)perylene	1.163	1.321	-13.6	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	95	0.00
2		1,4-Dioxane	40.000	44.976	-12.4	103	0.00
3		Pyridine	40.000	46.534	-16.3	105	0.00
4		n-Nitrosodimethylamine	40.000	45.956	-14.9	104	0.00
5	S	2-Fluorophenol	80.000	87.928	-9.9	99	0.00
6		Aniline	40.000	46.744	-16.9	104	0.00
7	S	Phenol-d6	80.000	90.140	-12.7	100	0.00
8		2-Chlorophenol	40.000	45.724	-14.3	103	0.00
9		Benzaldehyde	40.000	43.937	-9.8	98	0.00
10	C	Phenol	40.000	46.913	-17.3	106	0.00
11		bis(2-Chloroethyl)ether	40.000	45.521	-13.8	103	0.00
12		1,3-Dichlorobenzene	40.000	44.672	-11.7	102	0.00
13	C	1,4-Dichlorobenzene	40.000	44.941	-12.4	103	0.00
14		1,2-Dichlorobenzene	40.000	45.174	-12.9	103	0.00
15		Benzyl Alcohol	40.000	48.167	-20.4	106	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	44.929	-12.3	102	0.00
17		2-Methylphenol	40.000	47.153	-17.9	104	0.00
18		Hexachloroethane	40.000	44.738	-11.8	102	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	47.892	-19.7	105	0.00
20		3+4-Methylphenols	40.000	47.974	-19.9	106	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22		Acetophenone	40.000	45.153	-12.9	104	0.00
23	S	Nitrobenzene-d5	80.000	77.798	2.8	90	0.00
24		Nitrobenzene	40.000	45.104	-12.8	104	0.00
25		Isophorone	40.000	45.596	-14.0	107	0.00
26	C	2-Nitrophenol	40.000	47.087	-17.7	106	0.00
27		2,4-Dimethylphenol	40.000	46.404	-16.0	105	0.00
28		bis(2-Chloroethoxy)methane	40.000	44.920	-12.3	104	0.00
29	C	2,4-Dichlorophenol	40.000	46.375	-15.9	105	0.00
30		1,2,4-Trichlorobenzene	40.000	44.348	-10.9	104	0.00
31		Naphthalene	40.000	44.651	-11.6	105	0.00
32		Benzoic acid	40.000	46.063	-15.2	109	0.00
33		4-Chloroaniline	40.000	47.375	-18.4	110	0.00
34	C	Hexachlorobutadiene	40.000	44.034	-10.1	103	0.00
35		Caprolactam	40.000	46.481	-16.2	107	0.00
36	C	4-Chloro-3-methylphenol	40.000	46.825	-17.1	109	0.00
37		2-Methylnaphthalene	40.000	45.374	-13.4	107	0.00
38		1-Methylnaphthalene	40.000	45.198	-13.0	106	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	44.140	-10.4	105	0.00
41	P	Hexachlorocyclopentadiene	40.000	44.374	-10.9	106	0.00
42	S	2,4,6-Tribromophenol	80.000	90.503	-13.1	108	0.00
43	C	2,4,6-Trichlorophenol	40.000	46.094	-15.2	109	0.00
44		2,4,5-Trichlorophenol	40.000	46.271	-15.7	108	0.00
45	S	2-Fluorobiphenyl	80.000	76.259	4.7	91	0.00
46		1,1'-Biphenyl	40.000	44.207	-10.5	107	0.00
47		2-Chloronaphthalene	40.000	44.684	-11.7	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	47.621	-19.1	112	0.00
49		Acenaphthylene	40.000	45.315	-13.3	109	0.00
50		Dimethylphthalate	40.000	45.338	-13.3	111	0.00
51		2,6-Dinitrotoluene	40.000	47.554	-18.9	113	0.00
52	C	Acenaphthene	40.000	45.210	-13.0	110	0.00
53		3-Nitroaniline	40.000	48.843	-22.1	113	0.00
54	P	2,4-Dinitrophenol	40.000	45.189	-13.0	119	0.00
55		Dibenzofuran	40.000	44.828	-12.1	109	0.00
56	P	4-Nitrophenol	40.000	46.012	-15.0	118	0.00
57		2,4-Dinitrotoluene	40.000	48.311	-20.8	113	0.00
58		Fluorene	40.000	46.115	-15.3	111	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	46.970	-17.4	113	0.00
60		Diethylphthalate	40.000	45.299	-13.2	111	0.00
61		4-Chlorophenyl-phenylether	40.000	45.950	-14.9	110	0.00
62		4-Nitroaniline	40.000	50.610	-26.5#	115	0.00
63		Azobenzene	40.000	45.747	-14.4	111	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65		4,6-Dinitro-2-methylphenol	40.000	46.920	-17.3	117	0.00
66	c	n-Nitrosodiphenylamine	40.000	44.485	-11.2	112	0.00
67		4-Bromophenyl-phenylether	40.000	43.944	-9.9	111	0.00
68		Hexachlorobenzene	40.000	44.174	-10.4	113	0.00
69		Atrazine	40.000	44.483	-11.2	112	0.00
70	C	Pentachlorophenol	40.000	46.059	-15.1	117	0.00
71		Phenanthrene	40.000	44.895	-12.2	115	0.00
72		Anthracene	40.000	45.381	-13.5	115	0.00
73		Carbazole	40.000	46.001	-15.0	116	0.00
74		Di-n-butylphthalate	40.000	45.654	-14.1	116	0.00
75	C	Fluoranthene	40.000	46.998	-17.5	119	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	115	0.00
77		Benzidine	40.000	44.204	-10.5	117	0.00
78		Pyrene	40.000	44.209	-10.5	118	0.00
79	S	Terphenyl-d14	80.000	78.970	1.3	101	0.00
80		Butylbenzylphthalate	40.000	44.718	-11.8	121	0.00
81		Benzo(a)anthracene	40.000	45.497	-13.7	122	0.00
82		3,3'-Dichlorobenzidine	40.000	43.871	-9.7	115	0.00
83		Chrysene	40.000	44.967	-12.4	123	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	45.224	-13.1	122	0.00
85	c	Di-n-octyl phthalate	40.000	44.974	-12.4	123	0.00
86	I	Perylene-d12	20.000	20.000	0.0	118	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	46.249	-15.6	127	0.00
88		Benzo(b)fluoranthene	40.000	45.066	-12.7	123	0.00
89		Benzo(k)fluoranthene	40.000	45.780	-14.5	128	0.00
90	C	Benzo(a)pyrene	40.000	45.577	-13.9	127	0.00
91		Dibenzo(a,h)anthracene	40.000	46.091	-15.2	127	-0.01
92		Benzo(g,h,i)perylene	40.000	45.437	-13.6	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Response Factor Report MSBNA_P

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 15 04:48:42 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BP024275.D 5 =BP024276.D 10 =BP024277.D 20 =BP024278.D 40 =BP024279.D 50 =BP024280.D 60 =BP024281.D 80 =BP0242
82.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.551	0.517	0.552	0.512	0.481	0.490	0.477	0.512	6.10
3)		Pyridine		1.265	1.280	1.422	1.396	1.333	1.383	1.378	1.351	4.43
4)		n-Nitrosodimet...		0.424	0.443	0.469	0.457	0.442	0.460	0.444	0.448	3.29
5)	S	2-Fluorophenol		1.123	1.145	1.262	1.254	1.210	1.277	1.196	1.210	4.93
6)		Aniline		1.439	1.487	1.636	1.587	1.403	1.428	1.380	1.480	6.52
7)	S	Phenol-d6		1.463	1.552	1.718	1.748	1.697	1.769	1.647	1.656	6.75
8)		2-Chlorophenol		1.260	1.297	1.412	1.399	1.359	1.400	1.328	1.350	4.30
9)		Benzaldehyde		0.943	0.934	0.942	0.836	0.732	0.726	0.621	0.819	15.70
10)	C	Phenol		1.475	1.565	1.720	1.750	1.698	1.779	1.643	1.661	6.54
11)		bis(2-Chloroet...		1.278	1.317	1.399	1.398	1.321	1.358	1.300	1.339	3.56
12)		1,3-Dichlorobe...		1.464	1.486	1.534	1.486	1.402	1.433	1.373	1.454	3.79
13)	C	1,4-Dichlorobe...		1.496	1.489	1.550	1.494	1.436	1.463	1.399	1.475	3.28
14)		1,2-Dichlorobe...		1.472	1.441	1.502	1.445	1.362	1.411	1.339	1.424	4.10
15)		Benzyl Alcohol		0.857	0.935	1.106	1.158	1.142	1.178	1.122	1.071	11.58
16)		2,2'-oxybis(1-...		1.452	1.496	1.517	1.514	1.405	1.397	1.348	1.447	4.54
17)		2-Methylphenol		0.915	1.002	1.123	1.140	1.105	1.150	1.087	1.075	7.97
18)		Hexachloroethane		0.532	0.537	0.558	0.542	0.521	0.533	0.510	0.533	2.88
19)	P	n-Nitroso-di-n...	0.941	1.008	1.008	1.088	1.084	1.041	1.038	0.989	1.025	4.76
20)		3+4-Methylphenols		1.224	1.365	1.560	1.592	1.567	1.607	1.525	1.491	9.57
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.470	0.487	0.522	0.519	0.490	0.495	0.483	0.495	3.85
23)	S	Nitrobenzene-d5		0.322	0.337	0.367	0.368	0.357	0.356	0.347	0.351	4.76
24)		Nitrobenzene		0.328	0.334	0.362	0.364	0.349	0.353	0.340	0.347	3.96
25)		Isophorone		0.575	0.601	0.653	0.673	0.650	0.654	0.639	0.635	5.41
26)	C	2-Nitrophenol		0.130	0.143	0.172	0.183	0.184	0.191	0.186	0.170	13.97
27)		2,4-Dimethylph...		0.180	0.194	0.219	0.226	0.225	0.228	0.221	0.213	8.81
28)		bis(2-Chloroet...		0.412	0.426	0.444	0.452	0.430	0.425	0.412	0.429	3.51
29)	C	2,4-Dichloroph...		0.242	0.268	0.297	0.308	0.307	0.311	0.302	0.291	8.94
30)		1,2,4-Trichlor...		0.308	0.305	0.318	0.321	0.309	0.315	0.306	0.311	2.00
31)		Naphthalene		1.044	1.054	1.091	1.079	1.037	1.048	1.023	1.054	2.27
32)		Benzoic acid			0.185	0.226	0.242	0.270	0.285	0.282	0.248	15.56
33)		4-Chloroaniline		0.328	0.349	0.389	0.397	0.378	0.379	0.378	0.371	6.46
34)	C	Hexachlorobuta...		0.181	0.181	0.186	0.189	0.181	0.182	0.182	0.183	1.62
35)		Caprolactam		0.085	0.092	0.111	0.112	0.115	0.121	0.115	0.107	12.53
36)	C	4-Chloro-3-met...		0.284	0.310	0.349	0.357	0.356	0.363	0.351	0.339	8.82
37)		2-Methylnaphth...		0.700	0.720	0.755	0.755	0.733	0.740	0.712	0.731	2.89
38)		1-Methylnaphth...		0.700	0.714	0.737	0.731	0.712	0.718	0.691	0.715	2.25

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.522	0.538	0.568	0.576	0.544	0.549	0.552	0.550	3.33
41)	P	Hexachlorocycl...	0.162	0.185	0.197	0.218	0.199	0.197	0.204	0.195	8.92
42)	S	2,4,6-Tribromo...	0.235	0.256	0.278	0.289	0.289	0.296	0.294	0.277	8.29
43)	C	2,4,6-Trichlor...	0.301	0.335	0.369	0.389	0.386	0.390	0.390	0.366	9.51
44)		2,4,5-Trichlor...	0.332	0.371	0.409	0.427	0.429	0.436	0.430	0.405	9.62
45)	S	2-Fluorobiphenyl	1.332	1.334	1.359	1.366	1.263	1.251	1.240	1.306	4.08
46)		1,1'-Biphenyl	1.454	1.489	1.528	1.529	1.434	1.428	1.429	1.470	3.07
47)		2-Chloronaphth...	1.063	1.100	1.135	1.139	1.085	1.090	1.078	1.099	2.60
48)		2-Nitroaniline	0.250	0.270	0.317	0.330	0.324	0.335	0.331	0.308	10.98
49)		Acenaphthylene	1.604	1.680	1.793	1.814	1.726	1.763	1.730	1.730	4.11
50)		Dimethylphthalate	1.429	1.416	1.495	1.495	1.460	1.460	1.426	1.454	2.22
51)		2,6-Dinitrotol...	0.264	0.290	0.319	0.325	0.319	0.328	0.317	0.309	7.58
52)	C	Acenaphthene	1.087	1.087	1.141	1.126	1.081	1.079	1.076	1.097	2.35
53)		3-Nitroaniline	0.250	0.289	0.338	0.354	0.342	0.340	0.358	0.325	12.25
54)	P	2,4-Dinitrophenol		0.125	0.163	0.186	0.196	0.210	0.211	0.182	18.22
55)		Dibenzofuran	1.801	1.807	1.871	1.833	1.746	1.768	1.723	1.793	2.84
56)	P	4-Nitrophenol	0.190	0.231	0.289	0.305	0.306	0.323	0.317	0.280	17.90
57)		2,4-Dinitrotol...	0.334	0.368	0.439	0.448	0.442	0.463	0.456	0.421	11.82
58)		Fluorene	1.376	1.397	1.454	1.398	1.368	1.366	1.356	1.388	2.40
59)		2,3,4,6-Tetrac...	0.327	0.351	0.380	0.389	0.387	0.390	0.389	0.373	6.57
60)		Diethylphthalate	1.451	1.484	1.528	1.550	1.499	1.499	1.481	1.499	2.16
61)		4-Chlorophenyl...	0.667	0.675	0.699	0.680	0.667	0.670	0.659	0.674	1.95
62)		4-Nitroaniline	0.263	0.300	0.353	0.369	0.364	0.380	0.376	0.344	12.99
63)		Azobenzene	1.269	1.383	1.446	1.431	1.385	1.385	1.344	1.378	4.26
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.095	0.118	0.131	0.134	0.140	0.138	0.126	13.50
66)	c	n-Nitrosodiphe...	0.567	0.585	0.616	0.630	0.591	0.608	0.580	0.597	3.71
67)		4-Bromophenyl-...	0.203	0.212	0.220	0.230	0.220	0.225	0.222	0.219	4.11
68)		Hexachlorobenzene	0.247	0.247	0.260	0.271	0.261	0.269	0.263	0.260	3.67
69)		Atrazine	0.176	0.167	0.134	0.121	0.162			0.152	15.57
70)	C	Pentachlorophenol	0.143	0.155	0.175	0.196	0.197	0.203	0.201	0.181	13.35
71)		Phenanthrene	1.098	1.086	1.117	1.127	1.070	1.082	1.058	1.091	2.27
72)		Anthracene	0.995	1.017	1.084	1.111	1.053	1.071	1.030	1.052	3.85
73)		Carbazole	0.968	0.998	1.060	1.084	1.014	1.037	1.031	1.027	3.78
74)		Di-n-butylphth...	1.147	1.230	1.253	1.398	1.310	1.338	1.284	1.280	6.30
75)	C	Fluoranthene	1.289	1.274	1.316	1.347	1.280	1.304	1.274	1.298	2.07
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine	0.144	0.141	0.109	0.436	0.333	0.302	0.307	0.254	48.39
78)		Pyrene	1.192	1.236	1.345	1.305	1.294	1.305	1.219	1.271	4.38
79)	S	Terphenyl-d14	0.967	0.998	1.055	1.032	1.005	0.964	0.926	0.992	4.42
80)		Butylbenzylpht...	0.434	0.488	0.546	0.596	0.580	0.593	0.574	0.544	11.25
81)		Benzo(a)anthra...	1.195	1.214	1.292	1.293	1.234	1.253	1.221	1.243	3.08
82)		3,3'-Dichlorob...	0.345	0.383	0.431	0.486	0.463	0.485	0.460	0.436	12.30
83)		Chrysene	1.180	1.181	1.229	1.213	1.173	1.193	1.167	1.191	1.88
84)		Bis(2-ethylhex...	0.639	0.735	0.779	0.893	0.853	0.867	0.819	0.798	11.09
85)	c	Di-n-octyl pht...	0.945	1.093	1.222	1.461	1.430	1.486	1.446	1.297	16.46

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.252	1.304	1.407	1.444	1.401	1.445	1.466	1.388	5.77
88)		Benzo(b)fluora...	1.130	1.174	1.217	1.258	1.196	1.205	1.212	1.199	3.30
89)		Benzo(k)fluora...	1.088	1.131	1.217	1.212	1.154	1.166	1.138	1.158	3.94
90)	C	Benzo(a)pyrene	0.939	0.971	1.057	1.091	1.053	1.070	1.058	1.034	5.44
91)		Dibenzo(a,h)an...	1.046	1.089	1.172	1.203	1.169	1.186	1.202	1.152	5.28
92)		Benzo(g,h,i)pe...	1.073	1.110	1.198	1.218	1.180	1.209	1.224	1.173	4.99

(#) = Out of Range

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024275.D
Acq On : 14 Apr 2025 11:06
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 14 17:26:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	229109	20.000	ng	0.00
21) Naphthalene-d8	10.505	136	962244	20.000	ng	0.00
39) Acenaphthene-d10	14.363	164	614109	20.000	ng	0.00
64) Phenanthrene-d10	17.181	188	1238206	20.000	ng	0.02
76) Chrysene-d12	21.616	240	1367665	20.000	ng	0.00
86) Perylene-d12	24.980	264	1473609	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	8.528	70	26959	2.285	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024275.D
Acq On : 14 Apr 2025 11:06
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 14 17:26:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 229109

Ion	Ratio	Lower	Upper
152	100		
150	155.3	124.9	187.3
115	60.4	47.7	71.5





#5
2-Fluorophenol
Concen: 0.000 ng
Expected RT: 5.35 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	51.4
63	25.7





#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.92 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99

Sig	Exp Ratio
99	100
42	13.3
71	29.6



#19
n-Nitroso-di-n-propylamine
Concen: 2.285 ng
RT: 8.528 min Scan# 925
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 70 Resp: 26959

Ion	Ratio	Lower	Upper
70	100		
42	50.7	37.3	55.9
101	11.7	8.2	12.4
130	23.9	18.6	28.0





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.505 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 962244

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	7.2	6.0	9.0
68	6.7	5.5	8.3





#23
Nitrobenzene-d5
Concen: 0.000 ng
Expected RT: 8.88 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	45.6
54	41.8





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 614109

Ion	Ratio	Lower	Upper
164	100		
162	100.1	80.6	120.8
160	44.5	35.3	52.9





#42
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.88 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	96.4
141	30.9





#45
2-Fluorobiphenyl
Concen: 0.000 ng
Expected RT: 12.97 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172

Sig	Exp Ratio
172	100
171	35.2
170	23.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.181 min Scan# 2396
Delta R.T. 0.018 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion:188 Resp: 1238206

Ion	Ratio	Lower	Upper
188	100		
94	10.5	8.6	13.0
80	11.3	9.8	14.6





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 1367665

Ion	Ratio	Lower	Upper
240	100		
120	12.7	9.8	14.8
236	25.9	20.9	31.3





#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 19.90 min

Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	7.0
122	11.7





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.980 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024275.D
Acq: 14 Apr 2025 11:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 1473609

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.9	28.3
265	22.3	17.8	26.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024276.D
Acq On : 14 Apr 2025 11:47
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Apr 14 17:26:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.734	152	223010	20.000	ng		0.00
21) Naphthalene-d8	10.510	136	946483	20.000	ng		0.00
39) Acenaphthene-d10	14.369	164	609743	20.000	ng		0.00
64) Phenanthrene-d10	17.163	188	1234074	20.000	ng		0.00
76) Chrysene-d12	21.610	240	1401979	20.000	ng		-0.01
86) Perylene-d12	24.968	264	1519057	20.000	ng		-0.02
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	125173	9.263	ng		0.00
7) Phenol-d6	6.916	99	163122	8.824	ng		0.00
23) Nitrobenzene-d5	8.875	82	152286	9.160	ng		0.00
42) 2,4,6-Tribromophenol	15.875	330	71560	8.571	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	406060	10.111	ng		0.00
79) Terphenyl-d14	19.886	244	677812	9.638	ng		-0.01
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	30738	5.328	ng		100
3) Pyridine	3.699	79	70539	4.697	ng		97
4) n-Nitrosodimethylamine	3.599	42	23634	4.720	ng	#	90
6) Aniline	7.069	93	80239	4.808	ng		99
8) 2-Chlorophenol	7.305	128	70239	4.651	ng		98
9) Benzaldehyde	6.887	77	52595m	5.493	ng		
10) Phenol	6.940	94	82245	4.431	ng		97
11) bis(2-Chloroethyl)ether	7.163	93	71228	4.749	ng		97
12) 1,3-Dichlorobenzene	7.622	146	81600	4.987	ng		98
13) 1,4-Dichlorobenzene	7.769	146	83408	5.027	ng		99
14) 1,2-Dichlorobenzene	8.081	146	82066	5.115	ng		98
15) Benzyl Alcohol	7.975	79	47777	4.032	ng		98
16) 2,2'-oxybis(1-Chloropr...	8.258	45	80975	4.962	ng		97
17) 2-Methylphenol	8.175	107	51039	4.267	ng		99
18) Hexachloroethane	8.805	117	29643	4.950	ng		99
19) n-Nitroso-di-n-propyla...	8.528	70	56212	4.896	ng		96
20) 3+4-Methylphenols	8.505	107	68264	4.120	ng		99
22) Acetophenone	8.546	105	111159	4.725	ng		99
24) Nitrobenzene	8.922	77	77593	4.709	ng		96
25) Isophorone	9.446	82	136090	4.534	ng		99
26) 2-Nitrophenol	9.628	139	30712	3.882	ng		99
27) 2,4-Dimethylphenol	9.693	122	42510	4.239	ng		94
28) bis(2-Chloroethoxy)met...	9.922	93	97512	4.774	ng		98
29) 2,4-Dichlorophenol	10.169	162	57301	4.193	ng		97
30) 1,2,4-Trichlorobenzene	10.369	180	72789	4.923	ng		98
31) Naphthalene	10.557	128	246999	4.930	ng		99
33) 4-Chloroaniline	10.669	127	77667	4.437	ng		98
34) Hexachlorobutadiene	10.846	225	42723	4.926	ng		99
35) Caprolactam	11.446	113	20082	4.002	ng		94
36) 4-Chloro-3-methylphenol	11.804	107	67116	4.214	ng		100
37) 2-Methylnaphthalene	12.169	142	165617	4.770	ng		99
38) 1-Methylnaphthalene	12.399	142	165656	4.870	ng		99
40) 1,2,4,5-Tetrachloroben...	12.540	216	79501	4.744	ng		100
41) Hexachlorocyclopentadiene	12.528	237	24696	4.198	ng		96
43) 2,4,6-Trichlorophenol	12.787	196	45925	4.165	ng		95
44) 2,4,5-Trichlorophenol	12.863	196	50675	4.147	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024276.D
Acq On : 14 Apr 2025 11:47
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Apr 14 17:26:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46) 1,1'-Biphenyl	13.187	154	221653	4.922	ng		99
47) 2-Chloronaphthalene	13.234	162	162099	4.824	ng		99
48) 2-Nitroaniline	13.440	65	38152	4.111	ng		90
49) Acenaphthylene	14.087	152	244566	4.637	ng		99
50) Dimethylphthalate	13.822	163	217848	4.897	ng		100
51) 2,6-Dinitrotoluene	13.934	165	40182	4.287	ng		94
52) Acenaphthene	14.428	154	165694	4.940	ng		97
53) 3-Nitroaniline	14.275	138	38178	3.926	ng	#	94
55) Dibenzofuran	14.769	168	274506	4.991	ng		99
56) 4-Nitrophenol	14.610	139	29033	3.473	ng		98
57) 2,4-Dinitrotoluene	14.745	165	50843	4.013	ng		97
58) Fluorene	15.428	166	209763	4.939	ng		97
59) 2,3,4,6-Tetrachlorophenol	15.004	232	49911	4.415	ng		98
60) Diethylphthalate	15.204	149	221119	4.830	ng		99
61) 4-Chlorophenyl-phenyle...	15.422	204	101636	4.928	ng		99
62) 4-Nitroaniline	15.445	138	40139	3.893	ng		96
63) Azobenzene	15.722	77	193402	4.586	ng		99
66) n-Nitrosodiphenylamine	15.645	169	174867	4.726	ng		99
67) 4-Bromophenyl-phenylether	16.334	248	62558	4.647	ng		97
68) Hexachlorobenzene	16.451	284	76267	4.766	ng		96
69) Atrazine	16.616	200	54439	5.803	ng		98
70) Pentachlorophenol	16.810	266	44119	4.015	ng		98
71) Phenanthrene	17.204	178	338611	5.005	ng		100
72) Anthracene	17.292	178	307065	4.716	ng		99
73) Carbazole	17.586	167	298495	4.711	ng		100
74) Di-n-butylphthalate	18.175	149	354008	4.484	ng		100
75) Fluoranthene	19.292	202	397601	4.951	ng		99
77) Benzidine	19.498	184	50615m	2.925	ng		
78) Pyrene	19.663	202	417746	4.657	ng		99
80) Butylbenzylphthalate	20.622	149	152196	4.024	ng		99
81) Benzo(a)anthracene	21.592	228	418745	4.791	ng		100
82) 3,3'-Dichlorobenzidine	21.516	252	121047	3.994	ng		99
83) Chrysene	21.657	228	413570	4.937	ng		99
84) Bis(2-ethylhexyl)phtha...	21.533	149	223960	4.021	ng		98
85) Di-n-octyl phthalate	22.816	149	331185	3.712	ng		99
87) Indeno(1,2,3-cd)pyrene	28.762	276	475578	4.552	ng	#	92
88) Benzo(b)fluoranthene	23.904	252	429214	4.723	ng		99
89) Benzo(k)fluoranthene	23.968	252	413302	4.686	ng		99
90) Benzo(a)pyrene	24.810	252	356658	4.559	ng		98
91) Dibenzo(a,h)anthracene	28.845	278	397247	4.571	ng		100
92) Benzo(g,h,i)perylene	29.951	276	407522	4.607	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024276.D
Acq On : 14 Apr 2025 11:47
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Quant Time: Apr 14 17:26:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:152 Resp: 223010

Ion	Ratio	Lower	Upper
152	100		
150	153.5	124.9	187.3
115	59.8	47.7	71.5





#2
1,4-Dioxane
Concen: 5.328 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 88 Resp: 30738

Ion	Ratio	Lower	Upper
88	100		
58	64.3	51.4	77.2
43	22.3	17.1	25.7





#3
Pyridine
Concen: 4.697 ng
RT: 3.699 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 70539

Ion	Ratio	Lower	Upper
79	100		
52	56.9	45.4	68.2
51	32.8	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 4.720 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 42 Resp: 23634

Ion	Ratio	Lower	Upper
42	100		
74	161.0	139.5	209.3
44	17.2	5.8	8.8#





#5
2-Fluorophenol
Concen: 9.263 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:112 Resp: 125173

Ion	Ratio	Lower	Upper
112	100		
64	51.4	41.1	61.7
63	27.1	20.6	30.8





#6
Aniline
Concen: 4.808 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 93 Resp: 80239

Ion	Ratio	Lower	Upper
93	100		
66	36.8	29.1	43.7
65	18.0	14.7	22.1





#7
Phenol-d6
Concen: 8.824 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 99 Resp: 163122

Ion	Ratio	Lower	Upper
99	100		
42	14.7	10.6	16.0
71	30.7	23.7	35.5





#8
2-Chlorophenol
Concen: 4.651 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:128 Resp: 70239

Ion	Ratio	Lower	Upper
128	100		
130	33.5	12.9	52.9
64	44.2	22.7	62.7





#9
Benzaldehyde
Concen: 5.493 ng m
RT: 6.887 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 52595

Ion	Ratio	Lower	Upper
77	100		
105	100.6	79.2	119.2
106	97.4	76.2	116.2





#10
Phenol
Concen: 4.431 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 94 Resp: 82245

Ion	Ratio	Lower	Upper
94	100		
65	26.3	4.5	44.5
66	33.7	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 4.749 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 71228

Ion	Ratio	Lower	Upper
93	100		
63	71.6	48.6	88.6
95	32.9	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 4.987 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:146 Resp: 81600

Ion	Ratio	Lower	Upper
146	100		
148	62.9	51.4	77.0
75	29.6	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 5.027 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:146 Resp: 83408

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.1	76.7
111	42.2	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 5.115 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:146 Resp: 82066

Ion	Ratio	Lower	Upper
146	100		
148	62.6	51.0	76.4
111	41.2	33.8	50.8





#15
Benzyl Alcohol
Concen: 4.032 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 79 Resp: 47777

Ion	Ratio	Lower	Upper
79	100		
108	79.4	65.1	97.7
77	65.4	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 4.962 ng
RT: 8.258 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 45 Resp: 80975

Ion	Ratio	Lower	Upper
45	100		
77	19.4	0.0	37.7
79	14.4	0.0	33.5





#17
2-Methylphenol
Concen: 4.267 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 51039

Ion	Ratio	Lower	Upper
107	100		
108	112.3	89.8	134.6
77	44.7	34.2	51.4
79	43.3	33.8	50.8





#18
Hexachloroethane
Concen: 4.950 ng
RT: 8.805 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:117 Resp: 29643

Ion	Ratio	Lower	Upper
117	100		
119	96.4	77.8	116.6
201	96.8	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 4.896 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 70 Resp: 56212

Ion	Ratio	Lower	Upper
70	100		
42	43.5	37.3	55.9
101	9.4	8.2	12.4
130	22.0	18.6	28.0





#20
3+4-Methylphenols
Concen: 4.120 ng
RT: 8.505 min Scan# 921
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:107 Resp: 68264

Ion	Ratio	Lower	Upper
107	100		
108	89.1	69.8	109.8
77	30.9	11.5	51.5
79	28.1	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.510 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:136 Resp: 946483

Ion	Ratio	Lower	Upper
136	100		
137	10.8	9.2	13.8
54	7.1	6.0	9.0
68	6.5	5.5	8.3





#22
Acetophenone
Concen: 4.725 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:105 Resp: 111159

Ion	Ratio	Lower	Upper
105	100		
71	1.9	1.7	2.5
51	22.3	16.9	25.3
120	23.5	18.6	28.0





#23
Nitrobenzene-d5
Concen: 9.160 ng
RT: 8.875 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 152286

Ion	Ratio	Lower	Upper
82	100		
128	43.4	36.5	54.7
54	41.8	33.4	50.2





#24
Nitrobenzene
Concen: 4.709 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 77 Resp: 77593

Ion	Ratio	Lower	Upper
77	100		
123	45.4	39.2	58.8
65	12.3	9.8	14.6





#25
Isophorone
Concen: 4.534 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 82 Resp: 136090

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.4	8.2
138	17.6	14.4	21.6





#26
2-Nitrophenol
Concen: 3.882 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:139 Resp: 30712

Ion	Ratio	Lower	Upper
139	100		
109	28.1	21.8	32.6
65	45.7	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 4.239 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:122 Resp: 42510

Ion	Ratio	Lower	Upper
122	100		
107	109.1	93.7	140.5
121	56.5	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 4.774 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion: 93 Resp: 97512

Ion	Ratio	Lower	Upper
93	100		
95	34.0	26.0	39.0
123	11.2	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 4.193 ng
RT: 10.169 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:162 Resp: 57301

Ion	Ratio	Lower	Upper
162	100		
164	64.2	43.3	83.3
98	41.0	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 4.923 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:180 Resp: 72789

Ion	Ratio	Lower	Upper
180	100		
182	95.1	77.3	115.9
145	30.3	25.1	37.7





#31
Naphthalene
Concen: 4.930 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:128 Resp: 246999

Ion	Ratio	Lower	Upper
128	100		
129	11.5	8.6	13.0
127	13.1	10.3	15.5





#33
4-Chloroaniline
Concen: 4.437 ng
RT: 10.669 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:127 Resp: 77667

Ion	Ratio	Lower	Upper
127	100		
129	31.4	26.1	39.1
65	28.1	22.1	33.1
92	20.0	15.0	22.4





#34
Hexachlorobutadiene
Concen: 4.926 ng
RT: 10.846 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:225 Resp: 42723

Ion	Ratio	Lower	Upper
225	100		
223	62.1	49.5	74.3
227	61.5	50.9	76.3





#35
Caprolactam
Concen: 4.002 ng
RT: 11.446 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:113 Resp: 20082

Ion	Ratio	Lower	Upper
113	100		
55	146.2	115.9	155.9
56	104.7	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 4.214 ng
RT: 11.804 min Scan# 14
Delta R.T. 0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 67116

Ion	Ratio	Lower	Upper
107	100		
144	25.9	20.5	30.7
142	79.3	63.4	95.2





#37
2-Methylnaphthalene
Concen: 4.770 ng
RT: 12.169 min Scan# 1544
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:142 Resp: 165617

Ion	Ratio	Lower	Upper
142	100		
141	89.1	70.3	105.5
115	33.8	26.9	40.3





#38
1-Methylnaphthalene
Concen: 4.870 ng
RT: 12.399 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:142 Resp: 165656

Ion	Ratio	Lower	Upper
142	100		
141	91.3	73.6	110.4
116	4.0	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.369 min Scan# 1918
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:164 Resp: 609743

Ion	Ratio	Lower	Upper
164	100		
162	100.9	80.6	120.8
160	44.4	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.744 ng
RT: 12.540 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:216 Resp: 79501

Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.6	93.8
179	21.7	17.2	25.8
108	19.0	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 4.198 ng
RT: 12.528 min Scan# 1605
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:237 Resp: 24696

Ion	Ratio	Lower	Upper
237	100		
235	60.2	43.9	83.9
272	14.0	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 8.571 ng
RT: 15.875 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:330 Resp: 71560

Ion	Ratio	Lower	Upper
330	100		
332	99.6	77.1	115.7
141	33.2	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 4.165 ng
RT: 12.787 min Scan# 1649
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:196 Resp: 45925

Ion	Ratio	Lower	Upper
196	100		
198	97.9	74.3	111.5
200	32.2	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 4.147 ng
RT: 12.863 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:196 Resp: 50675

Ion	Ratio	Lower	Upper
196	100		
198	97.0	77.3	115.9
97	48.1	40.4	60.6
132	28.5	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 10.111 ng
RT: 12.975 min Scan# 1681
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:172 Resp: 406060

Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.2	42.2
170	23.1	18.6	27.8





#46
1,1'-Biphenyl
Concen: 4.922 ng
RT: 13.187 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 221653

Ion	Ratio	Lower	Upper
154	100		
153	39.5	20.3	60.3
76	13.7	0.0	33.5





#47
2-Chloronaphthalene
Concen: 4.824 ng
RT: 13.234 min Scan# 1725
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:162 Resp: 162099

Ion	Ratio	Lower	Upper
162	100		
127	39.0	31.9	47.9
164	32.5	26.2	39.4





#48
2-Nitroaniline
Concen: 4.111 ng
RT: 13.440 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 65 Resp: 38152

Ion	Ratio	Lower	Upper
65	100		
92	71.5	62.3	93.5
138	107.8	97.2	145.8





#49
Acenaphthylene
Concen: 4.637 ng
RT: 14.087 min Scan# 1870
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:152 Resp: 244566

Ion	Ratio	Lower	Upper
152	100		
151	20.4	15.9	23.9
153	13.2	10.6	15.8





#50
Dimethylphthalate
Concen: 4.897 ng
RT: 13.822 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:163 Resp: 217848

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	9.8	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 4.287 ng
RT: 13.934 min Scan# 1844
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:165 Resp: 40182

Ion	Ratio	Lower	Upper
165	100		
63	50.3	38.0	57.0
89	57.0	41.7	62.5





#52
Acenaphthene
Concen: 4.940 ng
RT: 14.428 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 165694

Ion	Ratio	Lower	Upper
154	100		
153	115.2	89.8	134.6
152	54.6	42.6	63.8





#53
3-Nitroaniline
Concen: 3.926 ng
RT: 14.275 min Scan# 1902
Delta R.T. 0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:138 Resp: 38178

Ion	Ratio	Lower	Upper
138	100		
108	12.8	8.3	12.5#
92	118.4	89.8	134.8





#55
Dibenzofuran
Concen: 4.991 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:168 Resp: 274506

Ion	Ratio	Lower	Upper
168	100		
139	37.3	30.3	45.5
169	12.9	10.6	16.0





#56
4-Nitrophenol
Concen: 3.473 ng
RT: 14.610 min Scan# 1959
Delta R.T. 0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:139 Resp: 29033

Ion	Ratio	Lower	Upper
139	100		
109	67.8	46.1	86.1
65	83.9	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 4.013 ng
RT: 14.745 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:165 Resp: 50843

Ion	Ratio	Lower	Upper
165	100		
63	40.8	30.4	45.6
89	69.9	54.6	82.0
182	2.8	2.2	3.2





#58
Fluorene
Concen: 4.939 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:166 Resp: 209763

Ion	Ratio	Lower	Upper
166	100		
165	101.0	78.1	117.1
167	13.9	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 4.415 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:232 Resp: 49911

Ion	Ratio	Lower	Upper
232	100		
131	48.4	36.8	55.2
130	2.5	1.9	2.9
166	31.3	25.0	37.6





#60
Diethylphthalate
Concen: 4.830 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:149 Resp: 221119

Ion	Ratio	Lower	Upper
149	100		
177	21.3	17.4	26.2
150	12.7	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 4.928 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:204 Resp: 101636

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.5	39.7
141	62.0	49.4	74.0





#62
4-Nitroaniline
Concen: 3.893 ng
RT: 15.445 min Scan# 2101
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:138 Resp: 40139

Ion	Ratio	Lower	Upper
138	100		
92	51.4	27.8	67.8
108	71.8	48.9	88.9





#63
Azobenzene
Concen: 4.586 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion: 77 Resp: 193402

Ion	Ratio	Lower	Upper
77	100		
182	26.4	7.3	47.3
105	22.3	2.3	42.3
51	23.2	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:188 Resp: 1234074

Ion	Ratio	Lower	Upper
188	100		
94	11.1	8.6	13.0
80	12.1	9.8	14.6





#66
n-Nitrosodiphenylamine
Concen: 4.726 ng
RT: 15.645 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:169 Resp: 174867

Ion	Ratio	Lower	Upper
169	100		
168	66.8	53.0	79.4
167	35.6	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 4.647 ng
RT: 16.334 min Scan# 2252
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:248 Resp: 62558

Ion	Ratio	Lower	Upper
248	100		
250	101.6	78.2	117.4
141	74.2	57.7	86.5





#68
Hexachlorobenzene
Concen: 4.766 ng
RT: 16.451 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:284 Resp: 76267

Ion	Ratio	Lower	Upper
284	100		
142	35.1	24.6	37.0
249	30.1	23.8	35.6





#69
Atrazine
Concen: 5.803 ng
RT: 16.616 min Scan# 2300
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:200 Resp: 54439

Ion	Ratio	Lower	Upper
200	100		
173	25.6	7.5	47.5
215	49.4	28.8	68.8





#70
Pentachlorophenol
Concen: 4.015 ng
RT: 16.810 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:266 Resp: 44119

Ion	Ratio	Lower	Upper
266	100		
268	67.1	51.0	76.6
264	61.9	49.8	74.6





#71
Phenanthrene
Concen: 5.005 ng
RT: 17.204 min Scan# 2400
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:178 Resp: 338611

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.4	23.0
179	15.5	12.5	18.7





#72
Anthracene
Concen: 4.716 ng
RT: 17.292 min Scan# 24
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:178 Resp: 307065

Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.7	22.1
179	14.8	12.4	18.6





#73
Carbazole
Concen: 4.711 ng
RT: 17.586 min Scan# 2465
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:167 Resp: 298495

Ion	Ratio	Lower	Upper
167	100		
166	21.1	17.1	25.7
139	13.3	10.6	16.0





#74
Di-n-butylphthalate
Concen: 4.484 ng
RT: 18.175 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 354008

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	5.1	3.9	5.9





#75
Fluoranthene
Concen: 4.951 ng
RT: 19.292 min Scan# 2755
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:202 Resp: 397601

Ion	Ratio	Lower	Upper
202	100		
101	12.5	0.0	33.2
203	17.3	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.610 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:240 Resp: 1401979

Ion	Ratio	Lower	Upper
240	100		
120	12.2	9.8	14.8
236	25.8	20.9	31.3





#77
Benzidine
Concen: 2.925 ng m
RT: 19.498 min Scan# 2790
Delta R.T. -0.000 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:184 Resp: 50615

Ion	Ratio	Lower	Upper
184	100		
185	15.8	11.6	17.4
183	12.4	9.2	13.8





#78
Pyrene
Concen: 4.657 ng
RT: 19.663 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:202 Resp: 417746

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.5	14.4	21.6





#79
Terphenyl-d14
Concen: 9.638 ng
RT: 19.886 min Scan# 2856
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:244 Resp: 677812

Ion	Ratio	Lower	Upper
244	100		
212	6.4	5.6	8.4
122	11.7	9.4	14.0





#80
Butylbenzylphthalate
Concen: 4.024 ng
RT: 20.622 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 152196

Ion	Ratio	Lower	Upper
149	100		
91	69.9	56.4	84.6
206	21.6	16.3	24.5





#81
Benzo(a)anthracene
Concen: 4.791 ng
RT: 21.592 min Scan# 3146
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:228 Resp: 418745

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	20.0	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 3.994 ng
RT: 21.516 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 121047

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.7	77.5
126	12.8	8.8	13.2





#83
Chrysene
Concen: 4.937 ng
RT: 21.657 min Scan# 3157
Delta R.T. -0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:228 Resp: 413570

Ion	Ratio	Lower	Upper
228	100		
226	28.9	24.2	36.2
229	19.4	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 4.021 ng
RT: 21.533 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:149 Resp: 223960

Ion	Ratio	Lower	Upper
149	100		
167	27.0	22.4	33.6
279	5.0	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 3.712 ng
RT: 22.816 min Scan# 3354
Delta R.T. -0.012 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:149 Resp: 331185

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.9	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.968 min Scan# 3
Delta R.T. -0.018 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:264 Resp: 1519057

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.9	28.3
265	21.6	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 4.552 ng
RT: 28.762 min Scan# 4365
Delta R.T. -0.047 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:276 Resp: 475578

Ion	Ratio	Lower	Upper
276	100		
138	16.7	14.5	21.7
277	19.3	20.2	30.2#





#88
Benzo(b)fluoranthene
Concen: 4.723 ng
RT: 23.904 min Scan# 3
Delta R.T. -0.024 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 429214

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.7	26.5
125	11.9	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 4.686 ng
RT: 23.968 min Scan# 3550
Delta R.T. -0.024 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:252 Resp: 413302

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.7	26.5
125	11.4	9.1	13.7





#90
Benzo(a)pyrene
Concen: 4.559 ng
RT: 24.810 min Scan# 3
Delta R.T. -0.024 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:252 Resp: 356658

Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.6	26.4
125	13.6	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 4.571 ng
RT: 28.845 min Scan# 4379
Delta R.T. -0.059 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Tgt Ion:278 Resp: 397247

Ion	Ratio	Lower	Upper
278	100		
139	16.2	12.6	19.0
279	23.1	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 4.607 ng
RT: 29.951 min Scan# 4
Delta R.T. -0.035 min
Lab File: BP024276.D
Acq: 14 Apr 2025 11:47

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Tgt Ion:276 Resp: 407522

Ion	Ratio	Lower	Upper
276	100		
277	23.9	18.9	28.3
138	23.8	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024277.D
Acq On : 14 Apr 2025 12:27
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Apr 14 17:26:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.734	152	237860	20.000	ng		0.00
21) Naphthalene-d8	10.510	136	1022895	20.000	ng		0.00
39) Acenaphthene-d10	14.363	164	658652	20.000	ng		0.00
64) Phenanthrene-d10	17.163	188	1310701	20.000	ng		0.00
76) Chrysene-d12	21.616	240	1408913	20.000	ng		0.00
86) Perylene-d12	24.980	264	1508235	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	272237	18.889	ng		0.00
7) Phenol-d6	6.910	99	369199	18.725	ng		0.00
23) Nitrobenzene-d5	8.881	82	345094	19.207	ng		0.00
42) 2,4,6-Tribromophenol	15.886	330	168790	18.715	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	878366	20.247	ng		0.00
79) Terphenyl-d14	19.892	244	1405643	19.888	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	61433	9.984	ng		98
3) Pyridine	3.693	79	152243	9.506	ng		99
4) n-Nitrosodimethylamine	3.599	42	52702	9.868	ng	#	98
6) Aniline	7.069	93	176840	9.935	ng		99
8) 2-Chlorophenol	7.305	128	154194	9.573	ng		100
9) Benzaldehyde	6.887	77	111093m	10.879	ng		
10) Phenol	6.940	94	186113	9.401	ng		97
11) bis(2-Chloroethyl)ether	7.157	93	156684	9.794	ng		98
12) 1,3-Dichlorobenzene	7.622	146	176712	10.126	ng		97
13) 1,4-Dichlorobenzene	7.769	146	177067	10.006	ng		99
14) 1,2-Dichlorobenzene	8.081	146	171376	10.016	ng		99
15) Benzyl Alcohol	7.969	79	111176	8.797	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.252	45	177896	10.221	ng		98
17) 2-Methylphenol	8.175	107	119161	9.340	ng		98
18) Hexachloroethane	8.804	117	63912	10.007	ng		98
19) n-Nitroso-di-n-propyla...	8.528	70	119861	9.787	ng		100
20) 3+4-Methylphenols	8.499	107	162332	9.187	ng		99
22) Acetophenone	8.546	105	248823	9.787	ng		99
24) Nitrobenzene	8.922	77	170770	9.590	ng		98
25) Isophorone	9.446	82	307272	9.473	ng		99
26) 2-Nitrophenol	9.628	139	73320	8.575	ng		96
27) 2,4-Dimethylphenol	9.693	122	99186	9.152	ng		97
28) bis(2-Chloroethoxy)met...	9.922	93	217866	9.869	ng		99
29) 2,4-Dichlorophenol	10.163	162	136816	9.263	ng		99
30) 1,2,4-Trichlorobenzene	10.369	180	155965	9.760	ng		100
31) Naphthalene	10.557	128	538892	9.953	ng		99
32) Benzoic acid	9.781	122	94803m	8.095	ng		
33) 4-Chloroaniline	10.669	127	178451	9.433	ng		99
34) Hexachlorobutadiene	10.845	225	92662	9.885	ng		99
35) Caprolactam	11.440	113	47014	8.670	ng		98
36) 4-Chloro-3-methylphenol	11.798	107	158563	9.213	ng		99
37) 2-Methylnaphthalene	12.169	142	368122	9.810	ng		98
38) 1-Methylnaphthalene	12.387	142	365139	9.933	ng		100
40) 1,2,4,5-Tetrachloroben...	12.540	216	177188	9.789	ng		100
41) Hexachlorocyclopentadiene	12.516	237	60914	9.586	ng		96
43) 2,4,6-Trichlorophenol	12.781	196	110203	9.253	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024277.D
Acq On : 14 Apr 2025 12:27
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Apr 14 17:26:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.857	196	122104	9.251	ng		98
46) 1,1'-Biphenyl	13.187	154	490325	10.080	ng		100
47) 2-Chloronaphthalene	13.228	162	362419	9.985	ng		98
48) 2-Nitroaniline	13.434	65	88889	8.866	ng		95
49) Acenaphthylene	14.081	152	553215	9.711	ng		100
50) Dimethylphthalate	13.816	163	466314	9.704	ng		99
51) 2,6-Dinitrotoluene	13.934	165	95596	9.442	ng		96
52) Acenaphthene	14.428	154	357828	9.877	ng		99
53) 3-Nitroaniline	14.269	138	95237	9.067	ng		97
54) 2,4-Dinitrophenol	14.481	184	41132	7.588	ng		97
55) Dibenzofuran	14.769	168	594953	10.013	ng		99
56) 4-Nitrophenol	14.598	139	75995	8.416	ng		95
57) 2,4-Dinitrotoluene	14.734	165	121276	8.862	ng	#	95
58) Fluorene	15.428	166	459952	10.026	ng		100
59) 2,3,4,6-Tetrachlorophenol	15.004	232	115645	9.470	ng		100
60) Diethylphthalate	15.198	149	488847	9.884	ng		100
61) 4-Chlorophenyl-phenyle...	15.428	204	222437	9.985	ng		97
62) 4-Nitroaniline	15.451	138	98638	8.857	ng		93
63) Azobenzene	15.722	77	455459	9.999	ng		96
65) 4,6-Dinitro-2-methylph...	15.516	198	62421	8.151	ng		96
66) n-Nitrosodiphenylamine	15.639	169	383558	9.759	ng		99
67) 4-Bromophenyl-phenylether	16.339	248	138724	9.703	ng		97
68) Hexachlorobenzene	16.457	284	161944	9.528	ng		98
69) Atrazine	16.628	200	109503	10.990	ng		98
70) Pentachlorophenol	16.810	266	101318	8.682	ng		96
71) Phenanthrene	17.210	178	711407	9.900	ng		99
72) Anthracene	17.304	178	666269	9.635	ng		100
73) Carbazole	17.586	167	654205	9.721	ng		99
74) Di-n-butylphthalate	18.180	149	806280	9.616	ng		99
75) Fluoranthene	19.292	202	835153	9.791	ng		98
77) Benzidine	19.498	184	99659	5.730	ng		99
78) Pyrene	19.669	202	870767	9.659	ng		99
80) Butylbenzylphthalate	20.627	149	343670	9.043	ng		99
81) Benzo(a)anthracene	21.598	228	854983	9.734	ng		98
82) 3,3'-Dichlorobenzidine	21.521	252	269772	8.857	ng		98
83) Chrysene	21.668	228	832263	9.887	ng		100
84) Bis(2-ethylhexyl)phtha...	21.539	149	517894	9.253	ng		99
85) Di-n-octyl phthalate	22.827	149	769654	8.584	ng		99
87) Indeno(1,2,3-cd)pyrene	28.792	276	983323	9.480	ng		99
88) Benzo(b)fluoranthene	23.915	252	885212	9.810	ng		99
89) Benzo(k)fluoranthene	23.986	252	852606	9.735	ng		99
90) Benzo(a)pyrene	24.815	252	732209	9.426	ng		99
91) Dibenzo(a,h)anthracene	28.862	278	821160	9.517	ng		99
92) Benzo(g,h,i)perylene	29.962	276	836993	9.530	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024277.D
Acq On : 14 Apr 2025 12:27
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Quant Time: Apr 14 17:26:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 237860

Ion	Ratio	Lower	Upper
152	100		
150	153.1	124.9	187.3
115	59.6	47.7	71.5





#2
1,4-Dioxane
Concen: 9.984 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 88 Resp: 61433

Ion	Ratio	Lower	Upper
88	100		
58	65.6	51.4	77.2
43	22.9	17.1	25.7





#3
Pyridine
Concen: 9.506 ng
RT: 3.693 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 152243

Ion	Ratio	Lower	Upper
79	100		
52	57.2	45.4	68.2
51	29.7	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 9.868 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 42 Resp: 52702

Ion	Ratio	Lower	Upper
42	100		
74	172.0	139.5	209.3
44	10.8	5.8	8.8#





#5
2-Fluorophenol
Concen: 18.889 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:112 Resp: 272237

Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.1	61.7
63	26.0	20.6	30.8





#6
Aniline
Concen: 9.935 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 93 Resp: 176840

Ion	Ratio	Lower	Upper
93	100		
66	36.8	29.1	43.7
65	18.6	14.7	22.1





#7
Phenol-d6
Concen: 18.725 ng
RT: 6.910 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 99 Resp: 369199

Ion	Ratio	Lower	Upper
99	100		
42	13.5	10.6	16.0
71	31.0	23.7	35.5





#8
2-Chlorophenol
Concen: 9.573 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:128 Resp: 154194

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.9	52.9
64	42.8	22.7	62.7





#9
Benzaldehyde
Concen: 10.879 ng m
RT: 6.887 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 111093

Ion	Ratio	Lower	Upper
77	100		
105	98.4	79.2	119.2
106	93.5	76.2	116.2





#10
Phenol
Concen: 9.401 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 94 Resp: 186113

Ion	Ratio	Lower	Upper
94	100		
65	25.7	4.5	44.5
66	34.4	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 9.794 ng
RT: 7.157 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 156684

Ion	Ratio	Lower	Upper
93	100		
63	70.5	48.6	88.6
95	31.6	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 10.126 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:146 Resp: 176712

Ion	Ratio	Lower	Upper
146	100		
148	60.8	51.4	77.0
75	28.4	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 10.006 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:146 Resp: 177067

Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.1	76.7
111	41.1	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 10.016 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:146 Resp: 171376

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.0	76.4
111	44.3	33.8	50.8





#15
Benzyl Alcohol
Concen: 8.797 ng
RT: 7.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 79 Resp: 111176

Ion	Ratio	Lower	Upper
79	100		
108	81.3	65.1	97.7
77	64.5	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.221 ng
RT: 8.252 min Scan# 878
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 45 Resp: 177896

Ion	Ratio	Lower	Upper
45	100		
77	18.4	0.0	37.7
79	14.1	0.0	33.5





#17
2-Methylphenol
Concen: 9.340 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:107 Resp: 119161

Ion	Ratio	Lower	Upper
107	100		
108	109.3	89.8	134.6
77	43.9	34.2	51.4
79	43.8	33.8	50.8





#18
Hexachloroethane
Concen: 10.007 ng
RT: 8.804 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:117 Resp: 63912

Ion	Ratio	Lower	Upper
117	100		
119	95.8	77.8	116.6
201	92.4	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 9.787 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 70 Resp: 119861

Ion	Ratio	Lower	Upper
70	100		
42	46.2	37.3	55.9
101	10.3	8.2	12.4
130	23.3	18.6	28.0





#20
3+4-Methylphenols
Concen: 9.187 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:107 Resp: 162332

Ion	Ratio	Lower	Upper
107	100		
108	90.5	69.8	109.8
77	32.3	11.5	51.5
79	27.4	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.510 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:136 Resp: 1022895

Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.2	13.8
54	6.6	6.0	9.0
68	6.5	5.5	8.3





#22
Acetophenone
Concen: 9.787 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:105 Resp: 248823

Ion	Ratio	Lower	Upper
105	100		
71	2.1	1.7	2.5
51	21.7	16.9	25.3
120	23.2	18.6	28.0





#23
Nitrobenzene-d5
Concen: 19.207 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 345094

Ion	Ratio	Lower	Upper
82	100		
128	43.8	36.5	54.7
54	40.9	33.4	50.2





#24
Nitrobenzene
Concen: 9.590 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 77 Resp: 170770

Ion	Ratio	Lower	Upper
77	100		
123	47.0	39.2	58.8
65	12.1	9.8	14.6





#25
Isophorone
Concen: 9.473 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 82 Resp: 307272

Ion	Ratio	Lower	Upper
82	100		
95	6.6	5.4	8.2
138	18.3	14.4	21.6





#26
2-Nitrophenol
Concen: 8.575 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:139 Resp: 73320

Ion	Ratio	Lower	Upper
139	100		
109	29.2	21.8	32.6
65	48.2	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 9.152 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:122 Resp: 99186

Ion	Ratio	Lower	Upper
122	100		
107	112.4	93.7	140.5
121	59.0	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 9.869 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 93 Resp: 217866

Ion	Ratio	Lower	Upper
93	100		
95	31.8	26.0	39.0
123	11.7	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 9.263 ng
RT: 10.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 136816

Ion	Ratio	Lower	Upper
162	100		
164	62.8	43.3	83.3
98	37.9	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 9.760 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:180 Resp: 155965

Ion	Ratio	Lower	Upper
180	100		
182	96.5	77.3	115.9
145	31.1	25.1	37.7





#31
Naphthalene
Concen: 9.953 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:128 Resp: 538892

Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.6	13.0
127	13.1	10.3	15.5





#32
Benzoic acid
Concen: 8.095 ng m
RT: 9.781 min Scan# 1138
Delta R.T. -0.047 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:122 Resp: 94803

Ion	Ratio	Lower	Upper
122	100		
105	123.3	104.8	144.8
77	92.9	69.2	109.2





#33
4-Chloroaniline
Concen: 9.433 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:127 Resp: 178451

Ion	Ratio	Lower	Upper
127	100		
129	32.9	26.1	39.1
65	28.1	22.1	33.1
92	19.2	15.0	22.4





#34
Hexachlorobutadiene
Concen: 9.885 ng
RT: 10.845 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:225 Resp: 92662

Ion	Ratio	Lower	Upper
225	100		
223	61.4	49.5	74.3
227	65.0	50.9	76.3





#35
Caprolactam
Concen: 8.670 ng
RT: 11.440 min Scan# 14
Delta R.T. -0.012 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:113 Resp: 47014

Ion	Ratio	Lower	Upper
113	100		
55	134.3	115.9	155.9
56	97.4	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 9.213 ng
RT: 11.798 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:107 Resp: 158563

Ion	Ratio	Lower	Upper
107	100		
144	25.8	20.5	30.7
142	80.6	63.4	95.2





#37
2-Methylnaphthalene
Concen: 9.810 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:142 Resp: 368122

Ion	Ratio	Lower	Upper
142	100		
141	89.7	70.3	105.5
115	32.5	26.9	40.3





#38
1-Methylnaphthalene
Concen: 9.933 ng
RT: 12.387 min Scan# 1581
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:142 Resp: 365139

Ion	Ratio	Lower	Upper
142	100		
141	92.1	73.6	110.4
116	3.5	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:164 Resp: 658652

Ion	Ratio	Lower	Upper
164	100		
162	100.6	80.6	120.8
160	45.0	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 9.789 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:216 Resp: 177188

Ion	Ratio	Lower	Upper
216	100		
214	78.3	62.6	93.8
179	21.5	17.2	25.8
108	19.8	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 9.586 ng
RT: 12.516 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:237 Resp: 60914

Ion	Ratio	Lower	Upper
237	100		
235	60.0	43.9	83.9
272	13.3	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 18.715 ng
RT: 15.886 min Scan# 2176
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:330 Resp: 168790

Ion	Ratio	Lower	Upper
330	100		
332	95.5	77.1	115.7
141	31.2	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 9.253 ng
RT: 12.781 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:196 Resp: 110203

Ion	Ratio	Lower	Upper
196	100		
198	93.9	74.3	111.5
200	29.3	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 9.251 ng
RT: 12.857 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:196 Resp: 122104

Ion	Ratio	Lower	Upper
196	100		
198	94.4	77.3	115.9
97	50.8	40.4	60.6
132	29.9	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 20.247 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:172 Resp: 878366

Ion	Ratio	Lower	Upper
172	100		
171	34.7	28.2	42.2
170	23.0	18.6	27.8





#46
1,1'-Biphenyl
Concen: 10.080 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:154 Resp: 490325

Ion	Ratio	Lower	Upper
154	100		
153	40.0	20.3	60.3
76	13.6	0.0	33.5





#47
2-Chloronaphthalene
Concen: 9.985 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:162 Resp: 362419

Ion	Ratio	Lower	Upper
162	100		
127	38.3	31.9	47.9
164	32.5	26.2	39.4





#48
2-Nitroaniline
Concen: 8.866 ng
RT: 13.434 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion: 65 Resp: 88889

Ion	Ratio	Lower	Upper
65	100		
92	77.0	62.3	93.5
138	113.7	97.2	145.8





#49
Acenaphthylene
Concen: 9.711 ng
RT: 14.081 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:152 Resp: 553215

Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.9	23.9
153	13.1	10.6	15.8





#50
Dimethylphthalate
Concen: 9.704 ng
RT: 13.816 min Scan# 1824
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:163 Resp: 466314

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.5	5.3
164	9.8	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 9.442 ng
RT: 13.934 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 95596

Ion	Ratio	Lower	Upper
165	100		
63	50.6	38.0	57.0
89	54.0	41.7	62.5





#52
Acenaphthene
Concen: 9.877 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:154 Resp: 357828

Ion	Ratio	Lower	Upper
154	100		
153	111.6	89.8	134.6
152	53.6	42.6	63.8





#53
3-Nitroaniline
Concen: 9.067 ng
RT: 14.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:138 Resp: 95237

Ion	Ratio	Lower	Upper
138	100		
108	11.7	8.3	12.5
92	116.0	89.8	134.8





#54
2,4-Dinitrophenol
Concen: 7.588 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:184 Resp: 41132

Ion	Ratio	Lower	Upper
184	100		
63	57.2	42.6	64.0
154	58.9	47.2	70.8





#55
Dibenzofuran
Concen: 10.013 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:168 Resp: 594953

Ion	Ratio	Lower	Upper
168	100		
139	37.5	30.3	45.5
169	13.1	10.6	16.0





#56
4-Nitrophenol
Concen: 8.416 ng
RT: 14.598 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:139 Resp: 75995

Ion	Ratio	Lower	Upper
139	100		
109	67.3	46.1	86.1
65	92.4	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 8.862 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:165 Resp: 121276

Ion	Ratio	Lower	Upper
165	100		
63	39.5	30.4	45.6
89	73.4	54.6	82.0
182	3.4	2.2	3.2#





#58
Fluorene
Concen: 10.026 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:166 Resp: 459952

Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.1	117.1
167	13.4	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 9.470 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:232 Resp: 115645

Ion	Ratio	Lower	Upper
232	100		
131	45.6	36.8	55.2
130	2.5	1.9	2.9
166	31.3	25.0	37.6





#60
Diethylphthalate
Concen: 9.884 ng
RT: 15.198 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 488847

Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.4	26.2
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 9.985 ng
RT: 15.428 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:204 Resp: 222437

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.5	39.7
141	58.8	49.4	74.0





#62
4-Nitroaniline
Concen: 8.857 ng
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:138 Resp: 98638

Ion	Ratio	Lower	Upper
138	100		
92	53.3	27.8	67.8
108	74.5	48.9	88.9





#63
Azobenzene
Concen: 9.999 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion: 77 Resp: 455459

Ion	Ratio	Lower	Upper
77	100		
182	22.7	7.3	47.3
105	20.9	2.3	42.3
51	23.4	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:188 Resp: 1310701

Ion	Ratio	Lower	Upper
188	100		
94	10.9	8.6	13.0
80	12.1	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 8.151 ng
RT: 15.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:198 Resp: 62421

Ion	Ratio	Lower	Upper
198	100		
51	41.0	18.4	58.4
105	49.0	26.5	66.5





#66
n-Nitrosodiphenylamine
Concen: 9.759 ng
RT: 15.639 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:169 Resp: 383558

Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.0	79.4
167	37.1	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 9.703 ng
RT: 16.339 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:248 Resp: 138724

Ion	Ratio	Lower	Upper
248	100		
250	97.2	78.2	117.4
141	76.6	57.7	86.5





#68
Hexachlorobenzene
Concen: 9.528 ng
RT: 16.457 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:284 Resp: 161944

Ion	Ratio	Lower	Upper
284	100		
142	32.4	24.6	37.0
249	29.1	23.8	35.6





#69
Atrazine
Concen: 10.990 ng
RT: 16.628 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:200 Resp: 109503

Ion	Ratio	Lower	Upper
200	100		
173	28.8	7.5	47.5
215	49.5	28.8	68.8





#70
Pentachlorophenol
Concen: 8.682 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:266 Resp: 101318

Ion	Ratio	Lower	Upper
266	100		
268	67.1	51.0	76.6
264	65.3	49.8	74.6





#71
Phenanthrene
Concen: 9.900 ng
RT: 17.210 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:178 Resp: 711407

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.4	23.0
179	15.5	12.5	18.7





#72
Anthracene
Concen: 9.635 ng
RT: 17.304 min Scan# 2417
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:178 Resp: 666269

Ion	Ratio	Lower	Upper
178	100		
176	18.5	14.7	22.1
179	15.4	12.4	18.6





#73
Carbazole
Concen: 9.721 ng
RT: 17.586 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:167 Resp: 654205

Ion	Ratio	Lower	Upper
167	100		
166	20.8	17.1	25.7
139	12.5	10.6	16.0





#74
Di-n-butylphthalate
Concen: 9.616 ng
RT: 18.180 min Scan# 2566
Delta R.T. 0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 806280

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.4	11.2
104	4.9	3.9	5.9





#75
Fluoranthene
Concen: 9.791 ng
RT: 19.292 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:202 Resp: 835153

Ion	Ratio	Lower	Upper
202	100		
101	12.3	0.0	33.2
203	17.1	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:240 Resp: 1408913

Ion	Ratio	Lower	Upper
240	100		
120	12.7	9.8	14.8
236	25.0	20.9	31.3





#77
Benzidine
Concen: 5.730 ng
RT: 19.498 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:184 Resp: 99659

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.6	17.4
183	10.6	9.2	13.8





#78
Pyrene
Concen: 9.659 ng
RT: 19.669 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:202 Resp: 870767

Ion	Ratio	Lower	Upper
202	100		
200	21.1	17.2	25.8
203	17.4	14.4	21.6





#79
Terphenyl-d14
Concen: 19.888 ng
RT: 19.892 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:244 Resp: 1405643

Ion	Ratio	Lower	Upper
244	100		
212	6.4	5.6	8.4
122	11.8	9.4	14.0





#80
Butylbenzylphthalate
Concen: 9.043 ng
RT: 20.627 min Scan# 2982
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 343670

Ion	Ratio	Lower	Upper
149	100		
91	69.9	56.4	84.6
206	19.7	16.3	24.5





#81
Benzo(a)anthracene
Concen: 9.734 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 854983

Ion	Ratio	Lower	Upper
228	100		
226	26.1	21.8	32.8
229	20.2	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 8.857 ng
RT: 21.521 min Scan# 3134
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:252 Resp: 269772

Ion	Ratio	Lower	Upper
252	100		
254	63.4	51.7	77.5
126	12.2	8.8	13.2





#83
Chrysene
Concen: 9.887 ng
RT: 21.668 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:228 Resp: 832263

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.2	36.2
229	19.7	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 9.253 ng
RT: 21.539 min Scan# 3137
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:149 Resp: 517894

Ion	Ratio	Lower	Upper
149	100		
167	27.4	22.4	33.6
279	4.4	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 8.584 ng
RT: 22.827 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:149 Resp: 769654

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.9	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.980 min Scan# 3722
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:264 Resp: 1508235

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.9	28.3
265	21.8	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 9.480 ng
RT: 28.792 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:276 Resp: 983323

Ion	Ratio	Lower	Upper
276	100		
138	19.2	14.5	21.7
277	25.4	20.2	30.2





#88
Benzo(b)fluoranthene
Concen: 9.810 ng
RT: 23.915 min Scan# 3541
Delta R.T. -0.012 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:252 Resp: 885212

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	12.2	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 9.735 ng
RT: 23.986 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:252 Resp: 852606

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.7	26.5
125	11.5	9.1	13.7





#90
Benzo(a)pyrene
Concen: 9.426 ng
RT: 24.815 min Scan# 3694
Delta R.T. -0.018 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:252 Resp: 732209

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.6	26.4
125	13.0	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 9.517 ng
RT: 28.862 min Scan# 4
Delta R.T. -0.041 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 821160

Ion	Ratio	Lower	Upper
278	100		
139	16.8	12.6	19.0
279	23.4	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 9.530 ng
RT: 29.962 min Scan# 4569
Delta R.T. -0.024 min
Lab File: BP024277.D
Acq: 14 Apr 2025 12:27

Tgt Ion:276 Resp: 836993

Ion	Ratio	Lower	Upper
276	100		
277	24.1	18.9	28.3
138	22.5	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024278.D
Acq On : 14 Apr 2025 13:08
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Apr 14 17:26:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	251998	20.000	ng	0.00
21) Naphthalene-d8	10.505	136	1081087	20.000	ng	0.00
39) Acenaphthene-d10	14.369	164	688775	20.000	ng	0.00
64) Phenanthrene-d10	17.169	188	1378480	20.000	ng	0.00
76) Chrysene-d12	21.616	240	1415765	20.000	ng	0.00
86) Perylene-d12	24.986	264	1549781	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.352	112	636245	41.668	ng	0.00
7) Phenol-d6	6.911	99	866004	41.458	ng	0.00
23) Nitrobenzene-d5	8.875	82	794485	41.839	ng	0.00
42) 2,4,6-Tribromophenol	15.875	330	383088	40.617	ng	0.00
45) 2-Fluorobiphenyl	12.975	172	1871534	41.254	ng	0.00
79) Terphenyl-d14	19.898	244	2987245	42.062	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	139222	21.358	ng	99
3) Pyridine	3.693	79	358370	21.120	ng	99
4) n-Nitrosodimethylamine	3.599	42	118077	20.869	ng	98
6) Aniline	7.069	93	412168	21.857	ng	99
8) 2-Chlorophenol	7.305	128	355815	20.852	ng	99
9) Benzaldehyde	6.881	77	237371	21.940	ng	100
10) Phenol	6.940	94	433531	20.670	ng	98
11) bis(2-Chloroethyl)ether	7.164	93	352530	20.800	ng	100
12) 1,3-Dichlorobenzene	7.622	146	386470	20.903	ng	100
13) 1,4-Dichlorobenzene	7.763	146	390483	20.828	ng	100
14) 1,2-Dichlorobenzene	8.081	146	378518	20.880	ng	99
15) Benzyl Alcohol	7.969	79	278608	20.809	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.258	45	382308	20.734	ng	99
17) 2-Methylphenol	8.175	107	283057	20.942	ng	99
18) Hexachloroethane	8.805	117	140558	20.773	ng	99
19) n-Nitroso-di-n-propyla...	8.528	70	274118	21.127	ng	100
20) 3+4-Methylphenols	8.499	107	393131	20.999	ng	99
22) Acetophenone	8.546	105	563994	20.991	ng	99
24) Nitrobenzene	8.916	77	390911	20.770	ng	98
25) Isophorone	9.446	82	705771	20.588	ng	98
26) 2-Nitrophenol	9.628	139	185740	20.553	ng	99
27) 2,4-Dimethylphenol	9.687	122	236436	20.643	ng	98
28) bis(2-Chloroethoxy)met...	9.916	93	480036	20.575	ng	100
29) 2,4-Dichlorophenol	10.163	162	320579	20.536	ng	99
30) 1,2,4-Trichlorobenzene	10.369	180	343641	20.348	ng	99
31) Naphthalene	10.557	128	1179848	20.618	ng	100
32) Benzoic acid	9.805	122	244213m	19.731	ng	
33) 4-Chloroaniline	10.669	127	420086	21.010	ng	100
34) Hexachlorobutadiene	10.846	225	200785	20.267	ng	99
35) Caprolactam	11.446	113	119912	20.922	ng	98
36) 4-Chloro-3-methylphenol	11.799	107	377768	20.767	ng	99
37) 2-Methylnaphthalene	12.169	142	816314	20.583	ng	100
38) 1-Methylnaphthalene	12.387	142	796550	20.503	ng	100
40) 1,2,4,5-Tetrachloroben...	12.540	216	391241	20.669	ng	100
41) Hexachlorocyclopentadiene	12.522	237	135654	20.414	ng	97
43) 2,4,6-Trichlorophenol	12.787	196	254189	20.410	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024278.D
Acq On : 14 Apr 2025 13:08
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Apr 14 17:26:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	281887	20.424	ng	99
46) 1,1'-Biphenyl	13.187	154	1052618	20.694	ng	100
47) 2-Chloronaphthalene	13.228	162	782040	20.603	ng	99
48) 2-Nitroaniline	13.440	65	218194	20.811	ng	99
49) Acenaphthylene	14.087	152	1234728	20.726	ng	99
50) Dimethylphthalate	13.822	163	1029599	20.489	ng	100
51) 2,6-Dinitrotoluene	13.940	165	219492	20.731	ng	98
52) Acenaphthene	14.434	154	785789	20.741	ng	99
53) 3-Nitroaniline	14.269	138	233006	21.212	ng	99
54) 2,4-Dinitrophenol	14.481	184	112157	19.785	ng	94
55) Dibenzofuran	14.769	168	1288363	20.735	ng	99
56) 4-Nitrophenol	14.593	139	199315	21.109	ng	99
57) 2,4-Dinitrotoluene	14.734	165	302332	21.126	ng	98
58) Fluorene	15.428	166	1001515	20.876	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.998	232	262077	20.523	ng	99
60) Diethylphthalate	15.198	149	1052183	20.344	ng	100
61) 4-Chlorophenyl-phenyle...	15.422	204	481631	20.675	ng	98
62) 4-Nitroaniline	15.445	138	243257	20.888	ng	96
63) Azobenzene	15.722	77	995912	20.907	ng	100
65) 4,6-Dinitro-2-methylph...	15.510	198	162501	20.176	ng	98
66) n-Nitrosodiphenylamine	15.640	169	849476	20.552	ng	100
67) 4-Bromophenyl-phenylether	16.334	248	302601	20.125	ng	99
68) Hexachlorobenzene	16.451	284	358889	20.077	ng	100
69) Atrazine	16.616	200	184151	17.573	ng	98
70) Pentachlorophenol	16.810	266	241815	19.702	ng	98
71) Phenanthrene	17.210	178	1540112	20.378	ng	100
72) Anthracene	17.298	178	1494006	20.543	ng	99
73) Carbazole	17.581	167	1461729	20.653	ng	100
74) Di-n-butylphthalate	18.175	149	1727074	19.585	ng	100
75) Fluoranthene	19.292	202	1813662	20.218	ng	99
77) Benzidine	19.498	184	154964	8.867	ng	99
78) Pyrene	19.669	202	1904884	21.028	ng	99
80) Butylbenzylphthalate	20.633	149	772776	20.235	ng	96
81) Benzo(a)anthracene	21.598	228	1829798	20.732	ng	99
82) 3,3'-Dichlorobenzidine	21.527	252	610749	19.956	ng	99
83) Chrysene	21.669	228	1739548	20.565	ng	99
84) Bis(2-ethylhexyl)phtha...	21.545	149	1103263	19.616	ng	99
85) Di-n-octyl phthalate	22.827	149	1730436	19.207	ng	100
87) Indeno(1,2,3-cd)pyrene	28.792	276	2180177	20.454	ng	99
88) Benzo(b)fluoranthene	23.916	252	1885654	20.337	ng	99
89) Benzo(k)fluoranthene	23.980	252	1885747	20.955	ng	99
90) Benzo(a)pyrene	24.816	252	1637648	20.518	ng	98
91) Dibenzo(a,h)anthracene	28.868	278	1815743	20.480	ng	98
92) Benzo(g,h,i)perylene	29.951	276	1856380	20.571	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024278.D
Acq On : 14 Apr 2025 13:08
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Quant Time: Apr 14 17:26:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 251998

Ion	Ratio	Lower	Upper
152	100		
150	155.4	124.9	187.3
115	59.4	47.7	71.5





#2
1,4-Dioxane
Concen: 21.358 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 88 Resp: 139222

Ion	Ratio	Lower	Upper
88	100		
58	64.8	51.4	77.2
43	22.2	17.1	25.7





#3
Pyridine
Concen: 21.120 ng
RT: 3.693 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 358370

Ion	Ratio	Lower	Upper
79	100		
52	55.9	45.4	68.2
51	28.3	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 20.869 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 42 Resp: 118077

Ion	Ratio	Lower	Upper
42	100		
74	171.1	139.5	209.3
44	8.3	5.8	8.8





#5
2-Fluorophenol
Concen: 41.668 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:112 Resp: 636245

Ion	Ratio	Lower	Upper
112	100		
64	52.4	41.1	61.7
63	25.4	20.6	30.8





#6
Aniline
Concen: 21.857 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 93 Resp: 412168

Ion	Ratio	Lower	Upper
93	100		
66	35.2	29.1	43.7
65	18.5	14.7	22.1





#7
Phenol-d6
Concen: 41.458 ng
RT: 6.911 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 99 Resp: 866004

Ion	Ratio	Lower	Upper
99	100		
42	13.7	10.6	16.0
71	30.2	23.7	35.5





#8
2-Chlorophenol
Concen: 20.852 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:128 Resp: 355815

Ion	Ratio	Lower	Upper
128	100		
130	32.2	12.9	52.9
64	42.7	22.7	62.7





#9
Benzaldehyde
Concen: 21.940 ng
RT: 6.881 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 237371

Ion	Ratio	Lower	Upper
77	100		
105	99.1	79.2	119.2
106	95.6	76.2	116.2





#10
Phenol
Concen: 20.670 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 94 Resp: 433531

Ion	Ratio	Lower	Upper
94	100		
65	25.8	4.5	44.5
66	32.8	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 20.800 ng
RT: 7.164 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 93 Resp: 352530

Ion	Ratio	Lower	Upper
93	100		
63	68.7	48.6	88.6
95	32.7	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 20.903 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:146 Resp: 386470

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.4	77.0
75	28.4	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 20.828 ng
RT: 7.763 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:146 Resp: 390483

Ion	Ratio	Lower	Upper
146	100		
148	63.6	51.1	76.7
111	41.1	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 20.880 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:146 Resp: 378518

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.0	76.4
111	42.3	33.8	50.8





#15
Benzyl Alcohol
Concen: 20.809 ng
RT: 7.969 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 79 Resp: 278608

Ion	Ratio	Lower	Upper
79	100		
108	79.3	65.1	97.7
77	63.4	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.734 ng
RT: 8.258 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 45 Resp: 382308

Ion	Ratio	Lower	Upper
45	100		
77	18.0	0.0	37.7
79	14.4	0.0	33.5





#17
2-Methylphenol
Concen: 20.942 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:107 Resp: 283057

Ion	Ratio	Lower	Upper
107	100		
108	112.9	89.8	134.6
77	43.0	34.2	51.4
79	41.4	33.8	50.8





#18
Hexachloroethane
Concen: 20.773 ng
RT: 8.805 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:117 Resp: 140558

Ion	Ratio	Lower	Upper
117	100		
119	97.5	77.8	116.6
201	96.0	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 21.127 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 70 Resp: 274118

Ion	Ratio	Lower	Upper
70	100		
42	46.5	37.3	55.9
101	10.3	8.2	12.4
130	22.8	18.6	28.0





#20
3+4-Methylphenols
Concen: 20.999 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:107 Resp: 393131

Ion	Ratio	Lower	Upper
107	100		
108	90.3	69.8	109.8
77	32.3	11.5	51.5
79	26.9	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.505 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:136 Resp: 1081087

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	7.3	6.0	9.0
68	6.8	5.5	8.3





#22
Acetophenone
Concen: 20.991 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:105 Resp: 563994

Ion	Ratio	Lower	Upper
105	100		
71	1.7	1.7	2.5
51	21.2	16.9	25.3
120	23.9	18.6	28.0





#23
Nitrobenzene-d5
Concen: 41.839 ng
RT: 8.875 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 794485

Ion	Ratio	Lower	Upper
82	100		
128	45.1	36.5	54.7
54	41.7	33.4	50.2





#24
Nitrobenzene
Concen: 20.770 ng
RT: 8.916 min Scan# 991
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 77 Resp: 390911

Ion	Ratio	Lower	Upper
77	100		
123	46.9	39.2	58.8
65	12.3	9.8	14.6





#25
Isophorone
Concen: 20.588 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 705771

Ion	Ratio	Lower	Upper
82	100		
95	6.7	5.4	8.2
138	19.2	14.4	21.6





#26
2-Nitrophenol
Concen: 20.553 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:139 Resp: 185740

Ion	Ratio	Lower	Upper
139	100		
109	28.3	21.8	32.6
65	45.4	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 20.643 ng
RT: 9.687 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:122 Resp: 236436

Ion	Ratio	Lower	Upper
122	100		
107	114.9	93.7	140.5
121	59.2	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 20.575 ng
RT: 9.916 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 93 Resp: 480036

Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.0	39.0
123	11.8	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 20.536 ng
RT: 10.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 320579

Ion	Ratio	Lower	Upper
162	100		
164	63.7	43.3	83.3
98	35.6	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 20.348 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:180 Resp: 343641

Ion	Ratio	Lower	Upper
180	100		
182	95.5	77.3	115.9
145	32.3	25.1	37.7





#31
Naphthalene
Concen: 20.618 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:128 Resp: 1179848

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	13.0
127	13.2	10.3	15.5





#32
Benzoic acid
Concen: 19.731 ng m
RT: 9.805 min Scan# 1142
Delta R.T. -0.024 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:122 Resp: 244213

Ion	Ratio	Lower	Upper
122	100		
105	124.1	104.8	144.8
77	89.8	69.2	109.2





#33
4-Chloroaniline
Concen: 21.010 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:127 Resp: 420086

Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.1	39.1
65	27.2	22.1	33.1
92	18.7	15.0	22.4





#34
Hexachlorobutadiene
Concen: 20.267 ng
RT: 10.846 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:225 Resp: 200785

Ion	Ratio	Lower	Upper
225	100		
223	63.4	49.5	74.3
227	63.3	50.9	76.3





#35
Caprolactam
Concen: 20.922 ng
RT: 11.446 min Scan# 14
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:113 Resp: 119912

Ion	Ratio	Lower	Upper
113	100		
55	134.1	115.9	155.9
56	103.0	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 20.767 ng
RT: 11.799 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:107 Resp: 377768

Ion	Ratio	Lower	Upper
107	100		
144	26.5	20.5	30.7
142	80.4	63.4	95.2





#37
2-Methylnaphthalene
Concen: 20.583 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:142 Resp: 816314

Ion	Ratio	Lower	Upper
142	100		
141	87.5	70.3	105.5
115	33.7	26.9	40.3





#38
1-Methylnaphthalene
Concen: 20.503 ng
RT: 12.387 min Scan# 1581
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:142 Resp: 796550

Ion	Ratio	Lower	Upper
142	100		
141	92.2	73.6	110.4
116	3.8	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:164 Resp: 688775

Ion	Ratio	Lower	Upper
164	100		
162	101.2	80.6	120.8
160	45.0	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 20.669 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:216 Resp: 391241

Ion	Ratio	Lower	Upper
216	100		
214	77.8	62.6	93.8
179	21.2	17.2	25.8
108	18.9	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 20.414 ng
RT: 12.522 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:237 Resp: 135654

Ion	Ratio	Lower	Upper
237	100		
235	61.3	43.9	83.9
272	13.5	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 40.617 ng
RT: 15.875 min Scan# 2174
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:330 Resp: 383088

Ion	Ratio	Lower	Upper
330	100		
332	97.5	77.1	115.7
141	30.9	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 20.410 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:196 Resp: 254189

Ion	Ratio	Lower	Upper
196	100		
198	96.6	74.3	111.5
200	31.2	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 20.424 ng
RT: 12.863 min Scan# 1662
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:196 Resp: 281887

Ion	Ratio	Lower	Upper
196	100		
198	96.9	77.3	115.9
97	48.9	40.4	60.6
132	29.0	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 41.254 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:172 Resp: 1871534

Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.2	42.2
170	23.4	18.6	27.8





#46
1,1'-Biphenyl
Concen: 20.694 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:154 Resp: 1052618

Ion	Ratio	Lower	Upper
154	100		
153	40.1	20.3	60.3
76	13.7	0.0	33.5





#47
2-Chloronaphthalene
Concen: 20.603 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:162 Resp: 782040

Ion	Ratio	Lower	Upper
162	100		
127	40.4	31.9	47.9
164	32.6	26.2	39.4





#48
2-Nitroaniline
Concen: 20.811 ng
RT: 13.440 min Scan# 1760
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion: 65 Resp: 218194

Ion	Ratio	Lower	Upper
65	100		
92	74.9	62.3	93.5
138	121.6	97.2	145.8





#49
Acenaphthylene
Concen: 20.726 ng
RT: 14.087 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:152 Resp: 1234728

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	12.7	10.6	15.8





#50
Dimethylphthalate
Concen: 20.489 ng
RT: 13.822 min Scan# 1825
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:163 Resp: 1029599

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.5	5.3
164	10.1	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 20.731 ng
RT: 13.940 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 219492

Ion	Ratio	Lower	Upper
165	100		
63	48.3	38.0	57.0
89	53.4	41.7	62.5





#52
Acenaphthene
Concen: 20.741 ng
RT: 14.434 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:154 Resp: 785789

Ion	Ratio	Lower	Upper
154	100		
153	113.9	89.8	134.6
152	53.8	42.6	63.8





#53
3-Nitroaniline
Concen: 21.212 ng
RT: 14.269 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:138 Resp: 233006

Ion	Ratio	Lower	Upper
138	100		
108	10.5	8.3	12.5
92	111.7	89.8	134.8





#54
2,4-Dinitrophenol
Concen: 19.785 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:184 Resp: 112157

Ion	Ratio	Lower	Upper
184	100		
63	61.0	42.6	64.0
154	60.0	47.2	70.8





#55
Dibenzofuran
Concen: 20.735 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:168 Resp: 1288363

Ion	Ratio	Lower	Upper
168	100		
139	37.3	30.3	45.5
169	13.2	10.6	16.0





#56
4-Nitrophenol
Concen: 21.109 ng
RT: 14.593 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:139 Resp: 199315

Ion	Ratio	Lower	Upper
139	100		
109	67.0	46.1	86.1
65	86.5	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 21.126 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 302332

Ion	Ratio	Lower	Upper
165	100		
63	37.0	30.4	45.6
89	66.2	54.6	82.0
182	3.0	2.2	3.2





#58
Fluorene
Concen: 20.876 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:166 Resp: 1001515

Ion	Ratio	Lower	Upper
166	100		
165	97.3	78.1	117.1
167	13.2	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 20.523 ng
RT: 14.998 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:232 Resp: 262077

Ion	Ratio	Lower	Upper
232	100		
131	45.5	36.8	55.2
130	2.4	1.9	2.9
166	31.1	25.0	37.6





#60
Diethylphthalate
Concen: 20.344 ng
RT: 15.198 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 1052183

Ion	Ratio	Lower	Upper
149	100		
177	22.0	17.4	26.2
150	12.8	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 20.675 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:204 Resp: 481631

Ion	Ratio	Lower	Upper
204	100		
206	32.6	26.5	39.7
141	59.2	49.4	74.0





#62
4-Nitroaniline
Concen: 20.888 ng
RT: 15.445 min Scan# 2101
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:138 Resp: 243257

Ion	Ratio	Lower	Upper
138	100		
92	51.2	27.8	67.8
108	72.0	48.9	88.9





#63
Azobenzene
Concen: 20.907 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion: 77 Resp: 995912

Ion	Ratio	Lower	Upper
77	100		
182	27.3	7.3	47.3
105	22.0	2.3	42.3
51	23.1	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.169 min Scan# 2394
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:188 Resp: 1378480

Ion	Ratio	Lower	Upper
188	100		
94	10.2	8.6	13.0
80	11.1	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 20.176 ng
RT: 15.510 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:198 Resp: 162501

Ion	Ratio	Lower	Upper
198	100		
51	38.2	18.4	58.4
105	43.8	26.5	66.5





#66
n-Nitrosodiphenylamine
Concen: 20.552 ng
RT: 15.640 min Scan# 2134
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:169 Resp: 849476

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.0	79.4
167	36.1	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 20.125 ng
RT: 16.334 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:248 Resp: 302601

Ion	Ratio	Lower	Upper
248	100		
250	96.8	78.2	117.4
141	71.8	57.7	86.5





#68
Hexachlorobenzene
Concen: 20.077 ng
RT: 16.451 min Scan# 2272
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:284 Resp: 358889

Ion	Ratio	Lower	Upper
284	100		
142	31.0	24.6	37.0
249	29.5	23.8	35.6





#69
Atrazine
Concen: 17.573 ng
RT: 16.616 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:200 Resp: 184151

Ion	Ratio	Lower	Upper
200	100		
173	26.9	7.5	47.5
215	47.4	28.8	68.8





#70
Pentachlorophenol
Concen: 19.702 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:266 Resp: 241815

Ion	Ratio	Lower	Upper
266	100		
268	65.1	51.0	76.6
264	63.6	49.8	74.6





#71
Phenanthrene
Concen: 20.378 ng
RT: 17.210 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:178 Resp: 1540112

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.4	23.0
179	15.6	12.5	18.7





#72
Anthracene
Concen: 20.543 ng
RT: 17.298 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:178 Resp: 1494006

Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.7	22.1
179	15.1	12.4	18.6





#73
Carbazole
Concen: 20.653 ng
RT: 17.581 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:167 Resp: 1461729

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	12.8	10.6	16.0





#74
Di-n-butylphthalate
Concen: 19.585 ng
RT: 18.175 min Scan# 2565
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 1727074

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.2
104	4.8	3.9	5.9





#75
Fluoranthene
Concen: 20.218 ng
RT: 19.292 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:202 Resp: 1813662

Ion	Ratio	Lower	Upper
202	100		
101	12.3	0.0	33.2
203	17.6	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:240 Resp: 1415765

Ion	Ratio	Lower	Upper
240	100		
120	13.0	9.8	14.8
236	25.2	20.9	31.3





#77
Benzidine
Concen: 8.867 ng
RT: 19.498 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:184 Resp: 154964

Ion	Ratio	Lower	Upper
184	100		
185	14.9	11.6	17.4
183	11.1	9.2	13.8





#78
Pyrene
Concen: 21.028 ng
RT: 19.669 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:202 Resp: 1904884

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.6	14.4	21.6





#79
Terphenyl-d14
Concen: 42.062 ng
RT: 19.898 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:244 Resp: 2987245

Ion	Ratio	Lower	Upper
244	100		
212	6.6	5.6	8.4
122	11.7	9.4	14.0





#80
Butylbenzylphthalate
Concen: 20.235 ng
RT: 20.633 min Scan# 2983
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 772776

Ion	Ratio	Lower	Upper
149	100		
91	67.1	56.4	84.6
206	21.2	16.3	24.5





#81
Benzo(a)anthracene
Concen: 20.732 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 1829798

Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.8	32.8
229	19.6	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 19.956 ng
RT: 21.527 min Scan# 3135
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:252 Resp: 610749

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.7	77.5
126	12.4	8.8	13.2





#83
Chrysene
Concen: 20.565 ng
RT: 21.669 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:228 Resp: 1739548

Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.2	36.2
229	19.8	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 19.616 ng
RT: 21.545 min Scan# 3138
Delta R.T. 0.006 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:149 Resp: 1103263

Ion	Ratio	Lower	Upper
149	100		
167	27.5	22.4	33.6
279	4.9	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 19.207 ng
RT: 22.827 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:149 Resp: 1730436

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.7	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. -0.000 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:264 Resp: 1549781

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.9	28.3
265	21.7	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 20.454 ng
RT: 28.792 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:276 Resp: 2180177

Ion	Ratio	Lower	Upper
276	100		
138	16.7	14.5	21.7
277	25.2	20.2	30.2





#88
Benzo(b)fluoranthene
Concen: 20.337 ng
RT: 23.916 min Scan# 3541
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:252 Resp: 1885654

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.7	26.5
125	11.7	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 20.955 ng
RT: 23.980 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:252 Resp: 1885747

Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.7	26.5
125	11.4	9.1	13.7





#90
Benzo(a)pyrene
Concen: 20.518 ng
RT: 24.816 min Scan# 3694
Delta R.T. -0.018 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:252 Resp: 1637648

Ion	Ratio	Lower	Upper
252	100		
253	21.3	17.6	26.4
125	13.2	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 20.480 ng
RT: 28.868 min Scan# 4
Delta R.T. -0.035 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Tgt Ion:278 Resp: 1815743

Ion	Ratio	Lower	Upper
278	100		
139	16.4	12.6	19.0
279	24.2	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 20.571 ng
RT: 29.951 min Scan# 4567
Delta R.T. -0.035 min
Lab File: BP024278.D
Acq: 14 Apr 2025 13:08

Tgt Ion:276 Resp: 1856380

Ion	Ratio	Lower	Upper
276	100		
277	24.1	18.9	28.3
138	22.6	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024279.D
Acq On : 14 Apr 2025 13:49
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Apr 14 17:26:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	246363	20.000	ng	0.00
21) Naphthalene-d8	10.505	136	1046876	20.000	ng	0.00
39) Acenaphthene-d10	14.363	164	657357	20.000	ng	0.00
64) Phenanthrene-d10	17.163	188	1277198	20.000	ng	0.00
76) Chrysene-d12	21.622	240	1349860	20.000	ng	0.00
86) Perylene-d12	24.986	264	1485135	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.352	112	1236220	82.813	ng	0.00
7) Phenol-d6	6.916	99	1722202	84.332	ng	0.00
23) Nitrobenzene-d5	8.875	82	1540790	83.791	ng	0.00
42) 2,4,6-Tribromophenol	15.881	330	759528	84.379	ng	0.00
45) 2-Fluorobiphenyl	12.969	172	3591010	82.939	ng	0.00
79) Terphenyl-d14	19.898	244	5570563	82.266	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	252519	39.624	ng	100
3) Pyridine	3.693	79	687730	41.457	ng	100
4) n-Nitrosodimethylamine	3.599	42	225273	40.725	ng	100
6) Aniline	7.069	93	781907	42.413	ng	100
8) 2-Chlorophenol	7.305	128	689257	41.317	ng	100
9) Benzaldehyde	6.881	77	411846	38.937	ng	100
10) Phenol	6.940	94	862244	42.050	ng	100
11) bis(2-Chloroethyl)ether	7.164	93	688989	41.581	ng	100
12) 1,3-Dichlorobenzene	7.622	146	732329	40.516	ng	100
13) 1,4-Dichlorobenzene	7.763	146	736270	40.171	ng	100
14) 1,2-Dichlorobenzene	8.081	146	711951	40.172	ng	100
15) Benzyl Alcohol	7.969	79	570437	43.581	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.252	45	746041	41.386	ng	100
17) 2-Methylphenol	8.169	107	561621	42.503	ng	100
18) Hexachloroethane	8.799	117	266915	40.349	ng	100
19) n-Nitroso-di-n-propyla...	8.528	70	533912	42.092	ng	100
20) 3+4-Methylphenols	8.499	107	784336	42.854	ng	100
22) Acetophenone	8.546	105	1086788	41.770	ng	100
24) Nitrobenzene	8.922	77	762756	41.852	ng	100
25) Isophorone	9.440	82	1408146	42.419	ng	100
26) 2-Nitrophenol	9.628	139	383751	43.852	ng	100
27) 2,4-Dimethylphenol	9.687	122	472488	42.600	ng	100
28) bis(2-Chloroethoxy)met...	9.922	93	947250	41.926	ng	100
29) 2,4-Dichlorophenol	10.157	162	645849	42.725	ng	100
30) 1,2,4-Trichlorobenzene	10.369	180	671264	41.046	ng	100
31) Naphthalene	10.557	128	2258924	40.765	ng	100
32) Benzoic acid	9.828	122	506813	42.286	ng	100
33) 4-Chloroaniline	10.669	127	830870	42.913	ng	100
34) Hexachlorobutadiene	10.846	225	394737	41.146	ng	100
35) Caprolactam	11.452	113	234519	42.256	ng	100
36) 4-Chloro-3-methylphenol	11.793	107	747541	42.438	ng	100
37) 2-Methylnaphthalene	12.163	142	1580497	41.154	ng	100
38) 1-Methylnaphthalene	12.381	142	1530982	40.695	ng	100
40) 1,2,4,5-Tetrachloroben...	12.540	216	757748	41.945	ng	100
41) Hexachlorocyclopentadiene	12.522	237	286135	45.116	ng	100
43) 2,4,6-Trichlorophenol	12.781	196	511566	43.038	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024279.D
Acq On : 14 Apr 2025 13:49
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Apr 14 17:26:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.857	196	561631	42.637	ng	100
46) 1,1'-Biphenyl	13.187	154	2010029	41.404	ng	100
47) 2-Chloronaphthalene	13.228	162	1497152	41.328	ng	100
48) 2-Nitroaniline	13.434	65	434050	43.377	ng	100
49) Acenaphthylene	14.081	152	2384785	41.944	ng	100
50) Dimethylphthalate	13.822	163	1965142	40.976	ng	100
51) 2,6-Dinitrotoluene	13.934	165	427248	42.283	ng	100
52) Acenaphthene	14.434	154	1480526	40.947	ng	100
53) 3-Nitroaniline	14.269	138	465347	44.389	ng	100
54) 2,4-Dinitrophenol	14.481	184	244714	45.231	ng	100
55) Dibenzofuran	14.775	168	2409526	40.633	ng	100
56) 4-Nitrophenol	14.593	139	401591	44.564	ng	100
57) 2,4-Dinitrotoluene	14.734	165	588798	43.110	ng	100
58) Fluorene	15.434	166	1838264	40.150	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.004	232	511714	41.987	ng	100
60) Diethylphthalate	15.204	149	2037863	41.286	ng	100
61) 4-Chlorophenyl-phenyle...	15.428	204	894350	40.226	ng	100
62) 4-Nitroaniline	15.457	138	485622	43.693	ng	100
63) Azobenzene	15.728	77	1881773	41.392	ng	100
65) 4,6-Dinitro-2-methylph...	15.510	198	335347	44.938	ng	100
66) n-Nitrosodiphenylamine	15.651	169	1609251	42.021	ng	100
67) 4-Bromophenyl-phenylether	16.340	248	587416	42.165	ng	100
68) Hexachlorobenzene	16.457	284	692546	41.815	ng	100
69) Atrazine	16.622	200	308618	31.786	ng	100
70) Pentachlorophenol	16.810	266	499858	43.955	ng	100
71) Phenanthrene	17.210	178	2879498	41.121	ng	100
72) Anthracene	17.304	178	2838724	42.130	ng	100
73) Carbazole	17.587	167	2770072	42.242	ng	100
74) Di-n-butylphthalate	18.175	149	3571220	43.708	ng	100
75) Fluoranthene	19.298	202	3441282	41.404	ng	100
77) Benzidine	19.498	184	1178258	70.714	ng	100
78) Pyrene	19.669	202	3523104	40.790	ng	100
80) Butylbenzylphthalate	20.628	149	1607881	44.158	ng	100
81) Benzo(a)anthracene	21.598	228	3489938	41.472	ng	100
82) 3,3'-Dichlorobenzidine	21.522	252	1311770	44.953	ng	100
83) Chrysene	21.675	228	3274153	40.596	ng	100
84) Bis(2-ethylhexyl)phtha...	21.539	149	2411998	44.979	ng	100
85) Di-n-octyl phthalate	22.827	149	3943939	45.913	ng	100
87) Indeno(1,2,3-cd)pyrene	28.809	276	4288630	41.987	ng	100
88) Benzo(b)fluoranthene	23.927	252	3737035	42.058	ng	100
89) Benzo(k)fluoranthene	23.992	252	3600626	41.753	ng	100
90) Benzo(a)pyrene	24.833	252	3240684	42.369	ng	100
91) Dibenzo(a,h)anthracene	28.904	278	3573562	42.061	ng	100
92) Benzo(g,h,i)perylene	29.986	276	3617069	41.826	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024279.D
Acq On : 14 Apr 2025 13:49
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Quant Time: Apr 14 17:26:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 246363

Ion	Ratio	Lower	Upper
152	100		
150	156.1	124.9	187.3
115	59.6	47.7	71.5





#2
1,4-Dioxane
Concen: 39.624 ng
RT: 3.293 min Scan# 35
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 88 Resp: 252519

Ion	Ratio	Lower	Upper
88	100		
58	64.3	51.4	77.2
43	21.4	17.1	25.7





#3
Pyridine
Concen: 41.457 ng
RT: 3.693 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 687730

Ion	Ratio	Lower	Upper
79	100		
52	56.8	45.4	68.2
51	28.2	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 40.725 ng
RT: 3.599 min Scan# 87
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 42 Resp: 225273

Ion	Ratio	Lower	Upper
42	100		
74	174.4	139.5	209.3
44	7.3	5.8	8.8





#5
2-Fluorophenol
Concen: 82.813 ng
RT: 5.352 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:112 Resp: 1236220

Ion	Ratio	Lower	Upper
112	100		
64	51.4	41.1	61.7
63	25.7	20.6	30.8





#6
Aniline
Concen: 42.413 ng
RT: 7.069 min Scan# 677
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 93 Resp: 781907

Ion	Ratio	Lower	Upper
93	100		
66	36.4	29.1	43.7
65	18.4	14.7	22.1





#7
Phenol-d6
Concen: 84.332 ng
RT: 6.916 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 99 Resp: 1722202

Ion	Ratio	Lower	Upper
99	100		
42	13.3	10.6	16.0
71	29.6	23.7	35.5





#8
2-Chlorophenol
Concen: 41.317 ng
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:128 Resp: 689257

Ion	Ratio	Lower	Upper
128	100		
130	32.9	12.9	52.9
64	42.7	22.7	62.7





#9
Benzaldehyde
Concen: 38.937 ng
RT: 6.881 min Scan# 64
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 411846

Ion	Ratio	Lower	Upper
77	100		
105	99.2	79.2	119.2
106	96.2	76.2	116.2





#10
Phenol
Concen: 42.050 ng
RT: 6.940 min Scan# 655
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 94 Resp: 862244

Ion	Ratio	Lower	Upper
94	100		
65	24.5	4.5	44.5
66	32.4	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 41.581 ng
RT: 7.164 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 93 Resp: 688989

Ion	Ratio	Lower	Upper
93	100		
63	68.6	48.6	88.6
95	32.3	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 40.516 ng
RT: 7.622 min Scan# 771
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:146 Resp: 732329

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.0
75	28.6	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 40.171 ng
RT: 7.763 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:146 Resp: 736270

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.1	76.7
111	40.7	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 40.172 ng
RT: 8.081 min Scan# 849
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:146 Resp: 711951

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.4
111	42.3	33.8	50.8





#15
Benzyl Alcohol
Concen: 43.581 ng
RT: 7.969 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 79 Resp: 570437

Ion	Ratio	Lower	Upper
79	100		
108	81.4	65.1	97.7
77	63.4	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.386 ng
RT: 8.252 min Scan# 878
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 45 Resp: 746041

Ion	Ratio	Lower	Upper
45	100		
77	17.7	0.0	37.7
79	13.5	0.0	33.5





#17
2-Methylphenol
Concen: 42.503 ng
RT: 8.169 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:107 Resp: 561621

Ion	Ratio	Lower	Upper
107	100		
108	112.2	89.8	134.6
77	42.8	34.2	51.4
79	42.3	33.8	50.8





#18
Hexachloroethane
Concen: 40.349 ng
RT: 8.799 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:117 Resp: 266915

Ion	Ratio	Lower	Upper
117	100		
119	97.2	77.8	116.6
201	94.9	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 42.092 ng
RT: 8.528 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 70 Resp: 533912

Ion	Ratio	Lower	Upper
70	100		
42	46.6	37.3	55.9
101	10.3	8.2	12.4
130	23.3	18.6	28.0





#20
3+4-Methylphenols
Concen: 42.854 ng
RT: 8.499 min Scan# 920
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:107 Resp: 784336

Ion	Ratio	Lower	Upper
107	100		
108	89.8	69.8	109.8
77	31.5	11.5	51.5
79	27.8	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.505 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:136 Resp: 1046876

Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	7.5	6.0	9.0
68	6.9	5.5	8.3





#22
Acetophenone
Concen: 41.770 ng
RT: 8.546 min Scan# 928
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:105 Resp: 1086788

Ion	Ratio	Lower	Upper
105	100		
71	2.1	1.7	2.5
51	21.1	16.9	25.3
120	23.3	18.6	28.0





#23
Nitrobenzene-d5
Concen: 83.791 ng
RT: 8.875 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1540790

Ion	Ratio	Lower	Upper
82	100		
128	45.6	36.5	54.7
54	41.8	33.4	50.2





#24
Nitrobenzene
Concen: 41.852 ng
RT: 8.922 min Scan# 992
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 77 Resp: 762756

Ion	Ratio	Lower	Upper
77	100		
123	49.0	39.2	58.8
65	12.2	9.8	14.6





#25
Isophorone
Concen: 42.419 ng
RT: 9.440 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 1408146

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.4	8.2
138	18.0	14.4	21.6





#26
2-Nitrophenol
Concen: 43.852 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:139 Resp: 383751

Ion	Ratio	Lower	Upper
139	100		
109	27.2	21.8	32.6
65	45.2	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 42.600 ng
RT: 9.687 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:122 Resp: 472488

Ion	Ratio	Lower	Upper
122	100		
107	117.1	93.7	140.5
121	60.2	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 41.926 ng
RT: 9.922 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 93 Resp: 947250

Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.0	39.0
123	11.4	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 42.725 ng
RT: 10.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 645849

Ion	Ratio	Lower	Upper
162	100		
164	63.3	43.3	83.3
98	36.6	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 41.046 ng
RT: 10.369 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:180 Resp: 671264

Ion	Ratio	Lower	Upper
180	100		
182	96.6	77.3	115.9
145	31.4	25.1	37.7





#31
Naphthalene
Concen: 40.765 ng
RT: 10.557 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:128 Resp: 2258924

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	13.0
127	12.9	10.3	15.5





#32
Benzoic acid
Concen: 42.286 ng
RT: 9.828 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:122 Resp: 506813

Ion	Ratio	Lower	Upper
122	100		
105	124.8	104.8	144.8
77	89.2	69.2	109.2





#33
4-Chloroaniline
Concen: 42.913 ng
RT: 10.669 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:127 Resp: 830870

Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.1	39.1
65	27.6	22.1	33.1
92	18.7	15.0	22.4





#34
Hexachlorobutadiene
Concen: 41.146 ng
RT: 10.846 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:225 Resp: 394737

Ion	Ratio	Lower	Upper
225	100		
223	61.9	49.5	74.3
227	63.6	50.9	76.3





#35
Caprolactam
Concen: 42.256 ng
RT: 11.452 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:113 Resp: 234519

Ion	Ratio	Lower	Upper
113	100		
55	135.9	115.9	155.9
56	101.1	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 42.438 ng
RT: 11.793 min Scan# 1480
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:107 Resp: 747541

Ion	Ratio	Lower	Upper
107	100		
144	25.6	20.5	30.7
142	79.3	63.4	95.2





#37
2-Methylnaphthalene
Concen: 41.154 ng
RT: 12.163 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:142 Resp: 1580497

Ion	Ratio	Lower	Upper
142	100		
141	87.9	70.3	105.5
115	33.6	26.9	40.3





#38
1-Methylnaphthalene
Concen: 40.695 ng
RT: 12.381 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:142 Resp: 1530982

Ion	Ratio	Lower	Upper
142	100		
141	92.0	73.6	110.4
116	3.8	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:164 Resp: 657357

Ion	Ratio	Lower	Upper
164	100		
162	100.7	80.6	120.8
160	44.1	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.945 ng
RT: 12.540 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:216 Resp: 757748

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.6	93.8
179	21.5	17.2	25.8
108	18.8	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 45.116 ng
RT: 12.522 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:237 Resp: 286135

Ion	Ratio	Lower	Upper
237	100		
235	63.9	43.9	83.9
272	13.6	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 84.379 ng
RT: 15.881 min Scan# 2175
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:330 Resp: 759528

Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.1	115.7
141	30.9	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 43.038 ng
RT: 12.781 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:196 Resp: 511566

Ion	Ratio	Lower	Upper
196	100		
198	92.9	74.3	111.5
200	30.0	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 42.637 ng
RT: 12.857 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:196 Resp: 561631

Ion	Ratio	Lower	Upper
196	100		
198	96.6	77.3	115.9
97	50.5	40.4	60.6
132	29.2	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 82.939 ng
RT: 12.969 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:172 Resp: 3591010

Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.2	42.2
170	23.2	18.6	27.8





#46
1,1'-Biphenyl
Concen: 41.404 ng
RT: 13.187 min Scan# 1717
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:154 Resp: 2010029

Ion	Ratio	Lower	Upper
154	100		
153	40.3	20.3	60.3
76	13.5	0.0	33.5





#47
2-Chloronaphthalene
Concen: 41.328 ng
RT: 13.228 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:162 Resp: 1497152

Ion	Ratio	Lower	Upper
162	100		
127	39.9	31.9	47.9
164	32.8	26.2	39.4





#48
2-Nitroaniline
Concen: 43.377 ng
RT: 13.434 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion: 65 Resp: 434050

Ion	Ratio	Lower	Upper
65	100		
92	77.9	62.3	93.5
138	121.5	97.2	145.8





#49
Acenaphthylene
Concen: 41.944 ng
RT: 14.081 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:152 Resp: 2384785

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.2	10.6	15.8





#50
Dimethylphthalate
Concen: 40.976 ng
RT: 13.822 min Scan# 1825
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:163 Resp: 1965142

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.5	5.3
164	10.0	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 42.283 ng
RT: 13.934 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 427248

Ion	Ratio	Lower	Upper
165	100		
63	47.5	38.0	57.0
89	52.1	41.7	62.5





#52
Acenaphthene
Concen: 40.947 ng
RT: 14.434 min Scan# 1929
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:154 Resp: 1480526

Ion	Ratio	Lower	Upper
154	100		
153	112.2	89.8	134.6
152	53.2	42.6	63.8





#53
3-Nitroaniline
Concen: 44.389 ng
RT: 14.269 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:138 Resp: 465347

Ion	Ratio	Lower	Upper
138	100		
108	10.4	8.3	12.5
92	112.3	89.8	134.8





#54
2,4-Dinitrophenol
Concen: 45.231 ng
RT: 14.481 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:184 Resp: 244714

Ion	Ratio	Lower	Upper
184	100		
63	53.3	42.6	64.0
154	59.0	47.2	70.8





#55
Dibenzofuran
Concen: 40.633 ng
RT: 14.775 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:168 Resp: 2409526

Ion	Ratio	Lower	Upper
168	100		
139	37.9	30.3	45.5
169	13.3	10.6	16.0





#56
4-Nitrophenol
Concen: 44.564 ng
RT: 14.593 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:139 Resp: 401591

Ion	Ratio	Lower	Upper
139	100		
109	66.1	46.1	86.1
65	85.0	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 43.110 ng
RT: 14.734 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 588798

Ion	Ratio	Lower	Upper
165	100		
63	38.0	30.4	45.6
89	68.3	54.6	82.0
182	2.7	2.2	3.2





#58
Fluorene
Concen: 40.150 ng
RT: 15.434 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:166 Resp: 1838264

Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.1	117.1
167	13.6	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 41.987 ng
RT: 15.004 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:232 Resp: 511714

Ion	Ratio	Lower	Upper
232	100		
131	46.0	36.8	55.2
130	2.4	1.9	2.9
166	31.3	25.0	37.6





#60
Diethylphthalate
Concen: 41.286 ng
RT: 15.204 min Scan# 2060
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 2037863

Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.4	26.2
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 40.226 ng
RT: 15.428 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:204 Resp: 894350

Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.5	39.7
141	61.7	49.4	74.0





#62
4-Nitroaniline
Concen: 43.693 ng
RT: 15.457 min Scan# 2103
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:138 Resp: 485622

Ion	Ratio	Lower	Upper
138	100		
92	47.8	27.8	67.8
108	68.9	48.9	88.9





#63
Azobenzene
Concen: 41.392 ng
RT: 15.728 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion: 77 Resp: 1881773

Ion	Ratio	Lower	Upper
77	100		
182	27.3	7.3	47.3
105	22.3	2.3	42.3
51	23.4	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:188 Resp: 1277198

Ion	Ratio	Lower	Upper
188	100		
94	10.8	8.6	13.0
80	12.2	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 44.938 ng
RT: 15.510 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:198 Resp: 335347

Ion	Ratio	Lower	Upper
198	100		
51	38.4	18.4	58.4
105	46.5	26.5	66.5





#66
n-Nitrosodiphenylamine
Concen: 42.021 ng
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:169 Resp: 1609251

Ion	Ratio	Lower	Upper
169	100		
168	66.2	53.0	79.4
167	36.1	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 42.165 ng
RT: 16.340 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:248 Resp: 587416

Ion	Ratio	Lower	Upper
248	100		
250	97.8	78.2	117.4
141	72.1	57.7	86.5





#68
Hexachlorobenzene
Concen: 41.815 ng
RT: 16.457 min Scan# 2273
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:284 Resp: 692546

Ion	Ratio	Lower	Upper
284	100		
142	30.8	24.6	37.0
249	29.7	23.8	35.6





#69
Atrazine
Concen: 31.786 ng
RT: 16.622 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:200 Resp: 308618

Ion	Ratio	Lower	Upper
200	100		
173	27.5	7.5	47.5
215	48.8	28.8	68.8





#70
Pentachlorophenol
Concen: 43.955 ng
RT: 16.810 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:266 Resp: 499858

Ion	Ratio	Lower	Upper
266	100		
268	63.8	51.0	76.6
264	62.2	49.8	74.6





#71
Phenanthrene
Concen: 41.121 ng
RT: 17.210 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:178 Resp: 2879498

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.0
179	15.6	12.5	18.7





#72
Anthracene
Concen: 42.130 ng
RT: 17.304 min Scan# 2417
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:178 Resp: 2838724

Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.7	22.1
179	15.5	12.4	18.6





#73
Carbazole
Concen: 42.242 ng
RT: 17.587 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:167 Resp: 2770072

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.1	25.7
139	13.3	10.6	16.0





#74
Di-n-butylphthalate
Concen: 43.708 ng
RT: 18.175 min Scan# 2565
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 3571220

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	4.9	3.9	5.9





#75
Fluoranthene
Concen: 41.404 ng
RT: 19.298 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:202 Resp: 3441282

Ion	Ratio	Lower	Upper
202	100		
101	13.2	0.0	33.2
203	17.7	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.622 min Scan# 3151
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:240 Resp: 1349860

Ion	Ratio	Lower	Upper
240	100		
120	12.3	9.8	14.8
236	26.1	20.9	31.3





#77
Benzidine
Concen: 70.714 ng
RT: 19.498 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:184 Resp: 1178258

Ion	Ratio	Lower	Upper
184	100		
185	14.5	11.6	17.4
183	11.5	9.2	13.8





#78
Pyrene
Concen: 40.790 ng
RT: 19.669 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:202 Resp: 3523104

Ion	Ratio	Lower	Upper
202	100		
200	21.5	17.2	25.8
203	18.0	14.4	21.6





#79
Terphenyl-d14
Concen: 82.266 ng
RT: 19.898 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:244 Resp: 5570563

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.6	8.4
122	11.7	9.4	14.0





#80
Butylbenzylphthalate
Concen: 44.158 ng
RT: 20.628 min Scan# 2982
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 1607881

Ion	Ratio	Lower	Upper
149	100		
91	70.5	56.4	84.6
206	20.4	16.3	24.5





#81
Benzo(a)anthracene
Concen: 41.472 ng
RT: 21.598 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 3489938

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.8	32.8
229	19.9	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 44.953 ng
RT: 21.522 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:252 Resp: 1311770

Ion	Ratio	Lower	Upper
252	100		
254	64.6	51.7	77.5
126	11.0	8.8	13.2





#83
Chrysene
Concen: 40.596 ng
RT: 21.675 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:228 Resp: 3274153

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.2
229	19.5	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 44.979 ng
RT: 21.539 min Scan# 3137
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:149 Resp: 2411998

Ion	Ratio	Lower	Upper
149	100		
167	28.0	22.4	33.6
279	5.0	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 45.913 ng
RT: 22.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:149 Resp: 3943939

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.5	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:264 Resp: 1485135

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.9	28.3
265	22.2	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 41.987 ng
RT: 28.809 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:276 Resp: 4288630

Ion	Ratio	Lower	Upper
276	100		
138	18.1	14.5	21.7
277	25.2	20.2	30.2





#88
Benzo(b)fluoranthene
Concen: 42.058 ng
RT: 23.927 min Scan# 3543
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:252 Resp: 3737035

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.7	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 41.753 ng
RT: 23.992 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:252 Resp: 3600626

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.4	9.1	13.7





#90
Benzo(a)pyrene
Concen: 42.369 ng
RT: 24.833 min Scan# 3697
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:252 Resp: 3240684

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	12.6	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 42.061 ng
RT: 28.904 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Tgt Ion:278 Resp: 3573562

Ion	Ratio	Lower	Upper
278	100		
139	15.8	12.6	19.0
279	23.1	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 41.826 ng
RT: 29.986 min Scan# 4573
Delta R.T. 0.000 min
Lab File: BP024279.D
Acq: 14 Apr 2025 13:49

Tgt Ion:276 Resp: 3617069

Ion	Ratio	Lower	Upper
276	100		
277	23.6	18.9	28.3
138	23.4	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024280.D
Acq On : 14 Apr 2025 15:10
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Apr 14 17:26:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.734	152	277364	20.000	ng		0.00
21) Naphthalene-d8	10.504	136	1176925	20.000	ng		0.00
39) Acenaphthene-d10	14.363	164	760235	20.000	ng		0.00
64) Phenanthrene-d10	17.163	188	1505097	20.000	ng		0.00
76) Chrysene-d12	21.616	240	1527193	20.000	ng		0.00
86) Perylene-d12	24.986	264	1706545	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	1677842	99.834	ng		0.00
7) Phenol-d6	6.916	99	2353543	102.366	ng		0.00
23) Nitrobenzene-d5	8.881	82	2100100	101.588	ng		0.00
42) 2,4,6-Tribromophenol	15.881	330	1098486	105.521	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	4800845	95.877	ng		0.00
79) Terphenyl-d14	19.898	244	7671902	100.142	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	333474	46.479	ng		100
3) Pyridine	3.687	79	924382m	49.495	ng		
4) n-Nitrosodimethylamine	3.593	42	306279	49.180	ng		99
6) Aniline	7.075	93	972988	46.879	ng		100
8) 2-Chlorophenol	7.310	128	942078	50.160	ng		99
9) Benzaldehyde	6.887	77	507606	42.627	ng		97
10) Phenol	6.946	94	1177623	51.012	ng		99
11) bis(2-Chloroethyl)ether	7.163	93	915679	49.085	ng		100
12) 1,3-Dichlorobenzene	7.622	146	972089	47.770	ng		100
13) 1,4-Dichlorobenzene	7.769	146	995872	48.262	ng		99
14) 1,2-Dichlorobenzene	8.087	146	944314	47.327	ng		99
15) Benzyl Alcohol	7.975	79	791998	53.745	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.257	45	973910	47.988	ng		99
17) 2-Methylphenol	8.175	107	766536	51.526	ng		100
18) Hexachloroethane	8.805	117	360999	48.472	ng		98
19) n-Nitroso-di-n-propyla...	8.534	70	722064	50.562	ng		98
20) 3+4-Methylphenols	8.505	107	1086273	52.718	ng		99
22) Acetophenone	8.546	105	1441728	49.289	ng	#	99
24) Nitrobenzene	8.922	77	1025983	50.074	ng		99
25) Isophorone	9.446	82	1913172	51.264	ng		99
26) 2-Nitrophenol	9.622	139	541332	55.024	ng		98
27) 2,4-Dimethylphenol	9.687	122	662183	53.106	ng		99
28) bis(2-Chloroethoxy)met...	9.922	93	1265001	49.803	ng		100
29) 2,4-Dichlorophenol	10.157	162	904519	53.225	ng		100
30) 1,2,4-Trichlorobenzene	10.369	180	908732	49.426	ng		98
31) Naphthalene	10.557	128	3049887	48.957	ng		100
32) Benzoic acid	9.851	122	794832	58.988	ng		99
33) 4-Chloroaniline	10.669	127	1110772	51.030	ng		99
34) Hexachlorobutadiene	10.840	225	533336	49.450	ng		99
35) Caprolactam	11.463	113	339834	54.466	ng		100
36) 4-Chloro-3-methylphenol	11.798	107	1046795	52.860	ng		99
37) 2-Methylnaphthalene	12.169	142	2156247	49.942	ng		99
38) 1-Methylnaphthalene	12.393	142	2095184	49.538	ng		99
40) 1,2,4,5-Tetrachloroben...	12.540	216	1034638	49.522	ng		99
41) Hexachlorocyclopentadiene	12.522	237	378578	51.614	ng		99
43) 2,4,6-Trichlorophenol	12.787	196	732902	53.315	ng		96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024280.D
Acq On : 14 Apr 2025 15:10
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Apr 14 17:26:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	814885	53.491	ng	98
46) 1,1'-Biphenyl	13.187	154	2725840	48.551	ng	99
47) 2-Chloronaphthalene	13.234	162	2062098	49.220	ng	100
48) 2-Nitroaniline	13.434	65	616599	53.282	ng	98
49) Acenaphthylene	14.081	152	3279666	49.877	ng	100
50) Dimethylphthalate	13.822	163	2775759	50.046	ng	100
51) 2,6-Dinitrotoluene	13.934	165	606525	51.903	ng	98
52) Acenaphthene	14.428	154	2054333	49.128	ng	100
53) 3-Nitroaniline	14.275	138	649071	53.536	ng	98
54) 2,4-Dinitrophenol	14.481	184	373165	59.640	ng	99
55) Dibenzofuran	14.769	168	3318798	48.392	ng	100
56) 4-Nitrophenol	14.592	139	581023	55.750	ng	97
57) 2,4-Dinitrotoluene	14.734	165	839701	53.160	ng	99
58) Fluorene	15.428	166	2599105	49.085	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.998	232	734741	52.128	ng	100
60) Diethylphthalate	15.204	149	2849294	49.914	ng	100
61) 4-Chlorophenyl-phenyle...	15.422	204	1267291	49.287	ng	99
62) 4-Nitroaniline	15.451	138	691081	53.764	ng	96
63) Azobenzene	15.722	77	2632843	50.076	ng	99
65) 4,6-Dinitro-2-methylph...	15.510	198	503765	57.285	ng	97
66) n-Nitrosodiphenylamine	15.645	169	2225475	49.312	ng	100
67) 4-Bromophenyl-phenylether	16.334	248	827245	50.389	ng	99
68) Hexachlorobenzene	16.457	284	983768	50.405	ng	98
69) Atrazine	16.622	200	610479	53.356	ng	100
70) Pentachlorophenol	16.810	266	739779	55.203	ng	99
71) Phenanthrene	17.210	178	4024282	48.768	ng	100
72) Anthracene	17.298	178	3960913	49.883	ng	100
73) Carbazole	17.581	167	3814743	49.364	ng	100
74) Di-n-butylphthalate	18.180	149	4930617	51.208	ng	100
75) Fluoranthene	19.298	202	4815021	49.160	ng	99
77) Benzidine	19.498	184	1272924	67.525	ng	99
78) Pyrene	19.675	202	4942328	50.577	ng	100
80) Butylbenzylphthalate	20.627	149	2215843	53.788	ng	97
81) Benzo(a)anthracene	21.598	228	4711301	49.485	ng	99
82) 3,3'-Dichlorobenzidine	21.527	252	1768950	53.582	ng	99
83) Chrysene	21.669	228	4480150	49.099	ng	100
84) Bis(2-ethylhexyl)phtha...	21.539	149	3256722	53.679	ng	99
85) Di-n-octyl phthalate	22.821	149	5458484	56.166	ng	99
87) Indeno(1,2,3-cd)pyrene	28.792	276	5977144m	50.926	ng	
88) Benzo(b)fluoranthene	23.915	252	5102465	49.975	ng	99
89) Benzo(k)fluoranthene	23.986	252	4923203	49.683	ng	99
90) Benzo(a)pyrene	24.821	252	4490517	51.093	ng	100
91) Dibenzo(a,h)anthracene	28.880	278	4987219	51.084	ng	98
92) Benzo(g,h,i)perylene	29.980	276	5033807	50.656	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024280.D
Acq On : 14 Apr 2025 15:10
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: Apr 14 17:26:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 277364

Ion	Ratio	Lower	Upper
152	100		
150	156.9	124.9	187.3
115	58.9	47.7	71.5





#2
1,4-Dioxane
Concen: 46.479 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 88 Resp: 333474

Ion	Ratio	Lower	Upper
88	100		
58	64.0	51.4	77.2
43	21.2	17.1	25.7





#3
Pyridine
Concen: 49.495 ng m
RT: 3.687 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 924382

Ion	Ratio	Lower	Upper
79	100		
52	58.2	45.4	68.2
51	27.7	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 49.180 ng
RT: 3.593 min Scan# 86
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 42 Resp: 306279

Ion	Ratio	Lower	Upper
42	100		
74	172.6	139.5	209.3
44	7.3	5.8	8.8





#5
2-Fluorophenol
Concen: 99.834 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:112 Resp: 1677842

Ion	Ratio	Lower	Upper
112	100		
64	52.2	41.1	61.7
63	25.9	20.6	30.8





#6
Aniline
Concen: 46.879 ng
RT: 7.075 min Scan# 678
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 93 Resp: 972988

Ion	Ratio	Lower	Upper
93	100		
66	36.4	29.1	43.7
65	18.2	14.7	22.1





#7
Phenol-d6
Concen: 102.366 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 99 Resp: 2353543

Ion	Ratio	Lower	Upper
99	100		
42	13.7	10.6	16.0
71	30.4	23.7	35.5





#8
2-Chlorophenol
Concen: 50.160 ng
RT: 7.310 min Scan# 718
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:128 Resp: 942078

Ion	Ratio	Lower	Upper
128	100		
130	32.7	12.9	52.9
64	41.2	22.7	62.7





#9
Benzaldehyde
Concen: 42.627 ng
RT: 6.887 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 507606

Ion	Ratio	Lower	Upper
77	100		
105	101.5	79.2	119.2
106	99.2	76.2	116.2





#10
Phenol
Concen: 51.012 ng
RT: 6.946 min Scan# 656
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 94 Resp: 1177623

Ion	Ratio	Lower	Upper
94	100		
65	24.8	4.5	44.5
66	32.8	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 49.085 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 93 Resp: 915679

Ion	Ratio	Lower	Upper
93	100		
63	68.6	48.6	88.6
95	31.9	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 47.770 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:146 Resp: 972089

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.0
75	28.7	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 48.262 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:146 Resp: 995872

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.1	76.7
111	40.3	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 47.327 ng
RT: 8.087 min Scan# 850
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:146 Resp: 944314

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.0	76.4
111	41.8	33.8	50.8





#15
Benzyl Alcohol
Concen: 53.745 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 791998

Ion	Ratio	Lower	Upper
79	100		
108	81.9	65.1	97.7
77	63.9	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 47.988 ng
RT: 8.257 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 45 Resp: 973910

Ion	Ratio	Lower	Upper
45	100		
77	17.6	0.0	37.7
79	14.2	0.0	33.5





#17
2-Methylphenol
Concen: 51.526 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:107 Resp: 766536

Ion	Ratio	Lower	Upper
107	100		
108	111.5	89.8	134.6
77	42.7	34.2	51.4
79	42.2	33.8	50.8





#18
Hexachloroethane
Concen: 48.472 ng
RT: 8.805 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:117 Resp: 360999

Ion	Ratio	Lower	Upper
117	100		
119	96.9	77.8	116.6
201	98.2	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 50.562 ng
RT: 8.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 722064

Ion	Ratio	Lower	Upper
70	100		
42	45.2	37.3	55.9
101	10.6	8.2	12.4
130	23.7	18.6	28.0





#20
3+4-Methylphenols
Concen: 52.718 ng
RT: 8.505 min Scan# 921
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:107 Resp: 1086273

Ion	Ratio	Lower	Upper
107	100		
108	91.1	69.8	109.8
77	31.0	11.5	51.5
79	28.1	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.504 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:136 Resp: 1176925

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	6.9	6.0	9.0
68	6.4	5.5	8.3





#22
Acetophenone
Concen: 49.289 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:105 Resp: 1441728

Ion	Ratio	Lower	Upper
105	100		
71	3.6	1.7	2.5#
51	21.1	16.9	25.3
120	23.9	18.6	28.0





#23
Nitrobenzene-d5
Concen: 101.588 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 2100100

Ion	Ratio	Lower	Upper
82	100		
128	46.9	36.5	54.7
54	41.8	33.4	50.2





#24
Nitrobenzene
Concen: 50.074 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 77 Resp: 1025983

Ion	Ratio	Lower	Upper
77	100		
123	48.7	39.2	58.8
65	12.7	9.8	14.6





#25
Isophorone
Concen: 51.264 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 1913172

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.4	8.2
138	18.7	14.4	21.6





#26
2-Nitrophenol
Concen: 55.024 ng
RT: 9.622 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:139 Resp: 541332

Ion	Ratio	Lower	Upper
139	100		
109	27.5	21.8	32.6
65	46.7	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 53.106 ng
RT: 9.687 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:122 Resp: 662183

Ion	Ratio	Lower	Upper
122	100		
107	116.6	93.7	140.5
121	58.7	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 49.803 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 93 Resp: 1265001

Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.0	39.0
123	11.8	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 53.225 ng
RT: 10.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 904519

Ion	Ratio	Lower	Upper
162	100		
164	63.5	43.3	83.3
98	36.6	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 49.426 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:180 Resp: 908732

Ion	Ratio	Lower	Upper
180	100		
182	94.5	77.3	115.9
145	31.4	25.1	37.7





#31
Naphthalene
Concen: 48.957 ng
RT: 10.557 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:128 Resp: 3049887

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	13.0
127	13.1	10.3	15.5





#32
Benzoic acid
Concen: 58.988 ng
RT: 9.851 min Scan# 1150
Delta R.T. 0.023 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:122 Resp: 794832

Ion	Ratio	Lower	Upper
122	100		
105	124.7	104.8	144.8
77	90.5	69.2	109.2





#33
4-Chloroaniline
Concen: 51.030 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:127 Resp: 1110772

Ion	Ratio	Lower	Upper
127	100		
129	31.7	26.1	39.1
65	26.8	22.1	33.1
92	18.4	15.0	22.4





#34
Hexachlorobutadiene
Concen: 49.450 ng
RT: 10.840 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:225 Resp: 533336

Ion	Ratio	Lower	Upper
225	100		
223	63.1	49.5	74.3
227	64.1	50.9	76.3





#35
Caprolactam
Concen: 54.466 ng
RT: 11.463 min Scan# 14
Delta R.T. 0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:113 Resp: 339834

Ion	Ratio	Lower	Upper
113	100		
55	136.1	115.9	155.9
56	100.2	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 52.860 ng
RT: 11.798 min Scan# 1481
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:107 Resp: 1046795

Ion	Ratio	Lower	Upper
107	100		
144	25.4	20.5	30.7
142	80.1	63.4	95.2





#37
2-Methylnaphthalene
Concen: 49.942 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:142 Resp: 2156247

Ion	Ratio	Lower	Upper
142	100		
141	88.8	70.3	105.5
115	34.5	26.9	40.3





#38
1-Methylnaphthalene
Concen: 49.538 ng
RT: 12.393 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:142 Resp: 2095184

Ion	Ratio	Lower	Upper
142	100		
141	91.2	73.6	110.4
116	3.7	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:164 Resp: 760235

Ion	Ratio	Lower	Upper
164	100		
162	101.6	80.6	120.8
160	44.8	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 49.522 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:216 Resp: 1034638

Ion	Ratio	Lower	Upper
216	100		
214	78.9	62.6	93.8
179	21.4	17.2	25.8
108	19.0	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 51.614 ng
RT: 12.522 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:237 Resp: 378578

Ion	Ratio	Lower	Upper
237	100		
235	63.4	43.9	83.9
272	13.7	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 105.521 ng
RT: 15.881 min Scan# 2175
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:330 Resp: 1098486

Ion	Ratio	Lower	Upper
330	100		
332	96.8	77.1	115.7
141	29.5	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 53.315 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:196 Resp: 732902

Ion	Ratio	Lower	Upper
196	100		
198	96.7	74.3	111.5
200	31.2	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 53.491 ng
RT: 12.863 min Scan# 1662
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:196 Resp: 814885

Ion	Ratio	Lower	Upper
196	100		
198	94.8	77.3	115.9
97	48.1	40.4	60.6
132	28.3	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 95.877 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:172 Resp: 4800845

Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.2	42.2
170	23.9	18.6	27.8





#46
1,1'-Biphenyl
Concen: 48.551 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:154 Resp: 2725840

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.3	60.3
76	13.9	0.0	33.5





#47
2-Chloronaphthalene
Concen: 49.220 ng
RT: 13.234 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 2062098

Ion	Ratio	Lower	Upper
162	100		
127	39.5	31.9	47.9
164	32.7	26.2	39.4





#48
2-Nitroaniline
Concen: 53.282 ng
RT: 13.434 min Scan# 1759
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion: 65 Resp: 616599

Ion	Ratio	Lower	Upper
65	100		
92	76.6	62.3	93.5
138	123.3	97.2	145.8





#49
Acenaphthylene
Concen: 49.877 ng
RT: 14.081 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 3279666

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.0	10.6	15.8





#50
Dimethylphthalate
Concen: 50.046 ng
RT: 13.822 min Scan# 1825
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:163 Resp: 2775759

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	10.0	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 51.903 ng
RT: 13.934 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 606525

Ion	Ratio	Lower	Upper
165	100		
63	49.3	38.0	57.0
89	53.4	41.7	62.5





#52
Acenaphthene
Concen: 49.128 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:154 Resp: 2054333

Ion	Ratio	Lower	Upper
154	100		
153	112.5	89.8	134.6
152	53.4	42.6	63.8





#53
3-Nitroaniline
Concen: 53.536 ng
RT: 14.275 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:138 Resp: 649071

Ion	Ratio	Lower	Upper
138	100		
108	10.2	8.3	12.5
92	110.0	89.8	134.8





#54
2,4-Dinitrophenol
Concen: 59.640 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:184 Resp: 373165

Ion	Ratio	Lower	Upper
184	100		
63	53.7	42.6	64.0
154	57.8	47.2	70.8





#55
Dibenzofuran
Concen: 48.392 ng
RT: 14.769 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:168 Resp: 3318798

Ion	Ratio	Lower	Upper
168	100		
139	38.1	30.3	45.5
169	13.4	10.6	16.0





#56
4-Nitrophenol
Concen: 55.750 ng
RT: 14.592 min Scan# 1956
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:139 Resp: 581023

Ion	Ratio	Lower	Upper
139	100		
109	68.1	46.1	86.1
65	87.3	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 53.160 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 839701

Ion	Ratio	Lower	Upper
165	100		
63	37.5	30.4	45.6
89	69.1	54.6	82.0
182	2.9	2.2	3.2





#58
Fluorene
Concen: 49.085 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:166 Resp: 2599105

Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.1	117.1
167	13.5	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 52.128 ng
RT: 14.998 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:232 Resp: 734741

Ion	Ratio	Lower	Upper
232	100		
131	46.0	36.8	55.2
130	2.3	1.9	2.9
166	31.2	25.0	37.6





#60
Diethylphthalate
Concen: 49.914 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 2849294

Ion	Ratio	Lower	Upper
149	100		
177	22.1	17.4	26.2
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 49.287 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:204 Resp: 1267291

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.5	39.7
141	62.8	49.4	74.0





#62
4-Nitroaniline
Concen: 53.764 ng
RT: 15.451 min Scan# 2102
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:138 Resp: 691081

Ion	Ratio	Lower	Upper
138	100		
92	50.2	27.8	67.8
108	72.0	48.9	88.9





#63
Azobenzene
Concen: 50.076 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 2632843

Ion	Ratio	Lower	Upper
77	100		
182	26.3	7.3	47.3
105	22.1	2.3	42.3
51	23.9	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:188 Resp: 1505097

Ion	Ratio	Lower	Upper
188	100		
94	10.8	8.6	13.0
80	12.1	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 57.285 ng
RT: 15.510 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:198 Resp: 503765

Ion	Ratio	Lower	Upper
198	100		
51	37.1	18.4	58.4
105	44.5	26.5	66.5





#66
n-Nitrosodiphenylamine
Concen: 49.312 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:169 Resp: 2225475

Ion	Ratio	Lower	Upper
169	100		
168	66.2	53.0	79.4
167	35.9	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 50.389 ng
RT: 16.334 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:248 Resp: 827245

Ion	Ratio	Lower	Upper
248	100		
250	97.5	78.2	117.4
141	73.8	57.7	86.5





#68
Hexachlorobenzene
Concen: 50.405 ng
RT: 16.457 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:284 Resp: 983768

Ion	Ratio	Lower	Upper
284	100		
142	32.4	24.6	37.0
249	29.7	23.8	35.6





#69
Atrazine
Concen: 53.356 ng
RT: 16.622 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:200 Resp: 610479

Ion	Ratio	Lower	Upper
200	100		
173	27.4	7.5	47.5
215	49.1	28.8	68.8





#70
Pentachlorophenol
Concen: 55.203 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:266 Resp: 739779

Ion	Ratio	Lower	Upper
266	100		
268	64.9	51.0	76.6
264	62.9	49.8	74.6





#71
Phenanthrene
Concen: 48.768 ng
RT: 17.210 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:178 Resp: 4024282

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.4	23.0
179	15.5	12.5	18.7





#72
Anthracene
Concen: 49.883 ng
RT: 17.298 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:178 Resp: 3960913

Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.5	12.4	18.6





#73
Carbazole
Concen: 49.364 ng
RT: 17.581 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:167 Resp: 3814743

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.1	25.7
139	13.3	10.6	16.0





#74
Di-n-butylphthalate
Concen: 51.208 ng
RT: 18.180 min Scan# 2566
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 4930617

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.4	11.2
104	5.0	3.9	5.9





#75
Fluoranthene
Concen: 49.160 ng
RT: 19.298 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:202 Resp: 4815021

Ion	Ratio	Lower	Upper
202	100		
101	13.0	0.0	33.2
203	17.4	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:240 Resp: 1527193

Ion	Ratio	Lower	Upper
240	100		
120	11.7	9.8	14.8
236	25.6	20.9	31.3





#77
Benzidine
Concen: 67.525 ng
RT: 19.498 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 1272924

Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.6	17.4
183	11.7	9.2	13.8





#78
Pyrene
Concen: 50.577 ng
RT: 19.675 min Scan# 2820
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:202 Resp: 4942328

Ion	Ratio	Lower	Upper
202	100		
200	21.4	17.2	25.8
203	17.7	14.4	21.6





#79
Terphenyl-d14
Concen: 100.142 ng
RT: 19.898 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:244 Resp: 7671902

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.6	8.4
122	11.2	9.4	14.0





#80
Butylbenzylphthalate
Concen: 53.788 ng
RT: 20.627 min Scan# 2982
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 2215843

Ion	Ratio	Lower	Upper
149	100		
91	67.6	56.4	84.6
206	21.3	16.3	24.5





#81
Benzo(a)anthracene
Concen: 49.485 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 4711301

Ion	Ratio	Lower	Upper
228	100		
226	27.7	21.8	32.8
229	20.0	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 53.582 ng
RT: 21.527 min Scan# 3135
Delta R.T. 0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:252 Resp: 1768950

Ion	Ratio	Lower	Upper
252	100		
254	64.2	51.7	77.5
126	11.3	8.8	13.2





#83
Chrysene
Concen: 49.099 ng
RT: 21.669 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 4480150

Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.2
229	19.7	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 53.679 ng
RT: 21.539 min Scan# 3137
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:149 Resp: 3256722

Ion	Ratio	Lower	Upper
149	100		
167	28.3	22.4	33.6
279	5.1	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 56.166 ng
RT: 22.821 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:149 Resp: 5458484

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.2	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. -0.000 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:264 Resp: 1706545

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.9	28.3
265	21.5	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 50.926 ng m
RT: 28.792 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:276 Resp: 5977144

Ion	Ratio	Lower	Upper
276	100		
138	17.7	14.5	21.7
277	20.8	20.2	30.2





#88
Benzo(b)fluoranthene
Concen: 49.975 ng
RT: 23.915 min Scan# 3541
Delta R.T. -0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:252 Resp: 5102465

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	10.7	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 49.683 ng
RT: 23.986 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:252 Resp: 4923203

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	11.0	9.1	13.7





#90
Benzo(a)pyrene
Concen: 51.093 ng
RT: 24.821 min Scan# 3695
Delta R.T. -0.012 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:252 Resp: 4490517

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	12.5	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 51.084 ng
RT: 28.880 min Scan# 4
Delta R.T. -0.024 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 4987219

Ion	Ratio	Lower	Upper
278	100		
139	17.1	12.6	19.0
279	24.0	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 50.656 ng
RT: 29.980 min Scan# 4572
Delta R.T. -0.006 min
Lab File: BP024280.D
Acq: 14 Apr 2025 15:10

Tgt Ion:276 Resp: 5033807

Ion	Ratio	Lower	Upper
276	100		
277	23.7	18.9	28.3
138	22.4	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024281.D
Acq On : 14 Apr 2025 16:32
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Apr 14 17:27:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.728	152	283549	20.000	ng		0.00
21) Naphthalene-d8	10.510	136	1206231	20.000	ng		0.00
39) Acenaphthene-d10	14.363	164	777459	20.000	ng		0.00
64) Phenanthrene-d10	17.169	188	1536805	20.000	ng		0.00
76) Chrysene-d12	21.621	240	1596148	20.000	ng		0.00
86) Perylene-d12	24.986	264	1858365	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	2173071	126.481	ng		0.00
7) Phenol-d6	6.916	99	3009690	128.049	ng		0.00
23) Nitrobenzene-d5	8.881	82	2579126	121.729	ng		0.00
42) 2,4,6-Tribromophenol	15.881	330	1382037	129.818	ng		0.00
45) 2-Fluorobiphenyl	12.975	172	5835791	113.964	ng		0.00
79) Terphenyl-d14	19.898	244	9228969	115.262	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	417124	56.869	ng		100
3) Pyridine	3.687	79	1176097m	61.599	ng		
4) n-Nitrosodimethylamine	3.599	42	391176	61.443	ng		99
6) Aniline	7.069	93	1214636	57.245	ng		100
8) 2-Chlorophenol	7.305	128	1190812	62.021	ng		99
9) Benzaldehyde	6.881	77	617878	50.755	ng		99
10) Phenol	6.946	94	1513303	64.123	ng		99
11) bis(2-Chloroethyl)ether	7.163	93	1155297	60.579	ng		100
12) 1,3-Dichlorobenzene	7.622	146	1218832	58.589	ng		100
13) 1,4-Dichlorobenzene	7.763	146	1244205	58.981	ng		100
14) 1,2-Dichlorobenzene	8.081	146	1199988	58.830	ng		100
15) Benzyl Alcohol	7.975	79	1002308	66.533	ng		100
16) 2,2'-oxybis(1-Chloropr...	8.257	45	1187992	57.259	ng		99
17) 2-Methylphenol	8.175	107	978531	64.342	ng		99
18) Hexachloroethane	8.804	117	453349	59.545	ng		98
19) n-Nitroso-di-n-propyla...	8.534	70	883042	60.486	ng		99
20) 3+4-Methylphenols	8.499	107	1367005	64.895	ng		99
22) Acetophenone	8.546	105	1792666	59.797	ng	#	99
24) Nitrobenzene	8.922	77	1276229	60.775	ng		100
25) Isophorone	9.446	82	2365423	61.842	ng		99
26) 2-Nitrophenol	9.628	139	690735	68.504	ng		99
27) 2,4-Dimethylphenol	9.693	122	826335	64.660	ng		98
28) bis(2-Chloroethoxy)met...	9.922	93	1539129	59.124	ng		99
29) 2,4-Dichlorophenol	10.157	162	1124170	64.543	ng		99
30) 1,2,4-Trichlorobenzene	10.369	180	1138896	60.440	ng		99
31) Naphthalene	10.557	128	3790924	59.373	ng		99
32) Benzoic acid	9.869	122	1032910m	74.795	ng		
33) 4-Chloroaniline	10.675	127	1372601	61.527	ng		100
34) Hexachlorobutadiene	10.845	225	660129	59.719	ng		99
35) Caprolactam	11.475	113	437691	68.445	ng		99
36) 4-Chloro-3-methylphenol	11.804	107	1314421	64.762	ng		98
37) 2-Methylnaphthalene	12.169	142	2676555	60.487	ng		99
38) 1-Methylnaphthalene	12.387	142	2599119	59.960	ng		100
40) 1,2,4,5-Tetrachloroben...	12.540	216	1281552	59.981	ng		100
41) Hexachlorocyclopentadiene	12.522	237	459486	61.257	ng		99
43) 2,4,6-Trichlorophenol	12.787	196	909999	64.732	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024281.D
Acq On : 14 Apr 2025 16:32
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Apr 14 17:27:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	1017414	65.306	ng	99
46) 1,1'-Biphenyl	13.187	154	3330176	58.001	ng	99
47) 2-Chloronaphthalene	13.228	162	2542212	59.336	ng	99
48) 2-Nitroaniline	13.439	65	781584	66.042	ng	98
49) Acenaphthylene	14.087	152	4111553	61.143	ng	99
50) Dimethylphthalate	13.828	163	3404463	60.022	ng	100
51) 2,6-Dinitrotoluene	13.939	165	764810	63.998	ng	98
52) Acenaphthene	14.434	154	2516981	58.858	ng	99
53) 3-Nitroaniline	14.275	138	793796	64.022	ng	99
54) 2,4-Dinitrophenol	14.486	184	490497	76.655	ng	97
55) Dibenzofuran	14.775	168	4124640	58.810	ng	100
56) 4-Nitrophenol	14.598	139	754038	70.748	ng	99
57) 2,4-Dinitrotoluene	14.739	165	1079633	66.836	ng	98
58) Fluorene	15.434	166	3184960	58.817	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.004	232	909497	63.097	ng	100
60) Diethylphthalate	15.204	149	3495577	59.879	ng	100
61) 4-Chlorophenyl-phenyle...	15.428	204	1563042	59.442	ng	99
62) 4-Nitroaniline	15.457	138	886163	67.414	ng	98
63) Azobenzene	15.728	77	3229873	60.070	ng	98
65) 4,6-Dinitro-2-methylph...	15.516	198	647253	72.083	ng	94
66) n-Nitrosodiphenylamine	15.651	169	2805137	60.874	ng	100
67) 4-Bromophenyl-phenylether	16.339	248	1038071	61.926	ng	98
68) Hexachlorobenzene	16.457	284	1239550	62.200	ng	97
70) Pentachlorophenol	16.816	266	936562	68.445	ng	100
71) Phenanthrene	17.222	178	4988840	59.209	ng	99
72) Anthracene	17.310	178	4938117	60.907	ng	99
73) Carbazole	17.592	167	4780408	60.584	ng	99
74) Di-n-butylphthalate	18.186	149	6167206	62.730	ng	100
75) Fluoranthene	19.304	202	6009748	60.092	ng	98
77) Benzidine	19.504	184	1446166	73.401	ng	100
78) Pyrene	19.680	202	6250781	61.204	ng	100
80) Butylbenzylphthalate	20.621	149	2840001	65.961	ng	98
81) Benzo(a)anthracene	21.604	228	6002059	60.318	ng	99
82) 3,3'-Dichlorobenzidine	21.521	252	2323294	67.333	ng	100
83) Chrysene	21.674	228	5714775	59.924	ng	99
84) Bis(2-ethylhexyl)phtha...	21.545	149	4153128	65.497	ng	99
85) Di-n-octyl phthalate	22.839	149	7116217	70.060	ng	99
87) Indeno(1,2,3-cd)pyrene	28.821	276	8057576m	63.043	ng	
88) Benzo(b)fluoranthene	23.927	252	6715444	60.400	ng	99
89) Benzo(k)fluoranthene	24.004	252	6500782	60.244	ng	99
90) Benzo(a)pyrene	24.827	252	5964163	62.316	ng	99
91) Dibenzo(a,h)anthracene	28.892	278	6614173	62.214	ng	98
92) Benzo(g,h,i)perylene	30.003	276	6740929	62.294	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024281.D
Acq On : 14 Apr 2025 16:32
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: Apr 14 17:27:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.728 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 283549

Ion	Ratio	Lower	Upper
152	100		
150	156.8	124.9	187.3
115	61.7	47.7	71.5





#2
1,4-Dioxane
Concen: 56.869 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 88 Resp: 417124

Ion	Ratio	Lower	Upper
88	100		
58	64.4	51.4	77.2
43	21.4	17.1	25.7





#3
Pyridine
Concen: 61.599 ng m
RT: 3.687 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 1176097

Ion	Ratio	Lower	Upper
79	100		
52	57.7	45.4	68.2
51	28.4	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 61.443 ng
RT: 3.599 min Scan# 87
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 42 Resp: 391176

Ion	Ratio	Lower	Upper
42	100		
74	176.0	139.5	209.3
44	6.7	5.8	8.8





#5
2-Fluorophenol
Concen: 126.481 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:112 Resp: 2173071

Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.1	61.7
63	25.5	20.6	30.8





#6
Aniline
Concen: 57.245 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 93 Resp: 1214636

Ion	Ratio	Lower	Upper
93	100		
66	36.6	29.1	43.7
65	18.1	14.7	22.1





#7
Phenol-d6
Concen: 128.049 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 99 Resp: 3009690

Ion	Ratio	Lower	Upper
99	100		
42	13.8	10.6	16.0
71	30.8	23.7	35.5





#8
2-Chlorophenol
Concen: 62.021 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:128 Resp: 1190812

Ion	Ratio	Lower	Upper
128	100		
130	32.3	12.9	52.9
64	42.3	22.7	62.7





#9
Benzaldehyde
Concen: 50.755 ng
RT: 6.881 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 617878

Ion	Ratio	Lower	Upper
77	100		
105	98.5	79.2	119.2
106	94.7	76.2	116.2





#10
Phenol
Concen: 64.123 ng
RT: 6.946 min Scan# 656
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 94 Resp: 1513303

Ion	Ratio	Lower	Upper
94	100		
65	24.8	4.5	44.5
66	32.8	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 60.579 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 93 Resp: 1155297

Ion	Ratio	Lower	Upper
93	100		
63	68.8	48.6	88.6
95	32.2	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 58.589 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:146 Resp: 1218832

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.4	77.0
75	28.7	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 58.981 ng
RT: 7.763 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 1244205

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.1	76.7
111	41.2	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 58.830 ng
RT: 8.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:146 Resp: 1199988

Ion	Ratio	Lower	Upper
146	100		
148	63.5	51.0	76.4
111	42.0	33.8	50.8





#15
Benzyl Alcohol
Concen: 66.533 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 1002308

Ion	Ratio	Lower	Upper
79	100		
108	81.4	65.1	97.7
77	62.9	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 57.259 ng
RT: 8.257 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 45 Resp: 1187992

Ion	Ratio	Lower	Upper
45	100		
77	17.9	0.0	37.7
79	14.3	0.0	33.5





#17
2-Methylphenol
Concen: 64.342 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:107 Resp: 978531

Ion	Ratio	Lower	Upper
107	100		
108	111.3	89.8	134.6
77	43.4	34.2	51.4
79	42.4	33.8	50.8





#18
Hexachloroethane
Concen: 59.545 ng
RT: 8.804 min Scan# 972
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:117 Resp: 453349

Ion	Ratio	Lower	Upper
117	100		
119	96.7	77.8	116.6
201	99.1	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 60.486 ng
RT: 8.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 70 Resp: 883042

Ion	Ratio	Lower	Upper
70	100		
42	45.8	37.3	55.9
101	10.7	8.2	12.4
130	23.9	18.6	28.0





#20
3+4-Methylphenols
Concen: 64.895 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:107 Resp: 1367005

Ion	Ratio	Lower	Upper
107	100		
108	89.0	69.8	109.8
77	31.1	11.5	51.5
79	26.7	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.510 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:136 Resp: 1206231

Ion	Ratio	Lower	Upper
136	100		
137	10.6	9.2	13.8
54	6.7	6.0	9.0
68	6.0	5.5	8.3





#22
Acetophenone
Concen: 59.797 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:105 Resp: 1792666

Ion	Ratio	Lower	Upper
105	100		
71	3.4	1.7	2.5#
51	21.4	16.9	25.3
120	23.4	18.6	28.0





#23
Nitrobenzene-d5
Concen: 121.729 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 2579126

Ion	Ratio	Lower	Upper
82	100		
128	47.8	36.5	54.7
54	41.9	33.4	50.2





#24
Nitrobenzene
Concen: 60.775 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 77 Resp: 1276229

Ion	Ratio	Lower	Upper
77	100		
123	49.1	39.2	58.8
65	12.4	9.8	14.6





#25
Isophorone
Concen: 61.842 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 2365423

Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.4	8.2
138	18.7	14.4	21.6





#26
2-Nitrophenol
Concen: 68.504 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:139 Resp: 690735

Ion	Ratio	Lower	Upper
139	100		
109	27.1	21.8	32.6
65	44.6	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 64.660 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:122 Resp: 826335

Ion	Ratio	Lower	Upper
122	100		
107	115.3	93.7	140.5
121	59.3	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 59.124 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 93 Resp: 1539129

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.0	39.0
123	11.9	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 64.543 ng
RT: 10.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 1124170

Ion	Ratio	Lower	Upper
162	100		
164	64.5	43.3	83.3
98	37.0	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 60.440 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:180 Resp: 1138896

Ion	Ratio	Lower	Upper
180	100		
182	95.5	77.3	115.9
145	30.8	25.1	37.7





#31
Naphthalene
Concen: 59.373 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 3790924

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.6	13.0
127	13.2	10.3	15.5





#32
Benzoic acid
Concen: 74.795 ng m
RT: 9.869 min Scan# 1153
Delta R.T. 0.041 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:122 Resp: 1032910

Ion	Ratio	Lower	Upper
122	100		
105	123.4	104.8	144.8
77	89.7	69.2	109.2





#33
4-Chloroaniline
Concen: 61.527 ng
RT: 10.675 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 1372601

Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.1	39.1
65	27.2	22.1	33.1
92	18.9	15.0	22.4





#34
Hexachlorobutadiene
Concen: 59.719 ng
RT: 10.845 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:225 Resp: 660129

Ion	Ratio	Lower	Upper
225	100		
223	62.2	49.5	74.3
227	64.1	50.9	76.3





#35
Caprolactam
Concen: 68.445 ng
RT: 11.475 min Scan# 14
Delta R.T. 0.023 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:113 Resp: 437691

Ion	Ratio	Lower	Upper
113	100		
55	134.0	115.9	155.9
56	101.6	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 64.762 ng
RT: 11.804 min Scan# 1482
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:107 Resp: 1314421

Ion	Ratio	Lower	Upper
107	100		
144	26.3	20.5	30.7
142	81.7	63.4	95.2





#37
2-Methylnaphthalene
Concen: 60.487 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:142 Resp: 2676555

Ion	Ratio	Lower	Upper
142	100		
141	89.0	70.3	105.5
115	34.4	26.9	40.3





#38
1-Methylnaphthalene
Concen: 59.960 ng
RT: 12.387 min Scan# 1581
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:142 Resp: 2599119

Ion	Ratio	Lower	Upper
142	100		
141	92.3	73.6	110.4
116	3.7	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:164 Resp: 777459

Ion	Ratio	Lower	Upper
164	100		
162	100.8	80.6	120.8
160	44.6	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 59.981 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:216 Resp: 1281552

Ion	Ratio	Lower	Upper
216	100		
214	78.7	62.6	93.8
179	21.4	17.2	25.8
108	18.7	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 61.257 ng
RT: 12.522 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 459486

Ion	Ratio	Lower	Upper
237	100		
235	62.7	43.9	83.9
272	13.7	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 129.818 ng
RT: 15.881 min Scan# 2175
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:330 Resp: 1382037

Ion	Ratio	Lower	Upper
330	100		
332	96.7	77.1	115.7
141	30.0	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 64.732 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:196 Resp: 909999

Ion	Ratio	Lower	Upper
196	100		
198	95.1	74.3	111.5
200	31.2	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 65.306 ng
RT: 12.863 min Scan# 1662
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:196 Resp: 1017414

Ion	Ratio	Lower	Upper
196	100		
198	96.2	77.3	115.9
97	48.5	40.4	60.6
132	28.9	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 113.964 ng
RT: 12.975 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:172 Resp: 5835791

Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.2	42.2
170	23.8	18.6	27.8





#46
1,1'-Biphenyl
Concen: 58.001 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:154 Resp: 3330176

Ion	Ratio	Lower	Upper
154	100		
153	40.6	20.3	60.3
76	13.8	0.0	33.5





#47
2-Chloronaphthalene
Concen: 59.336 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 2542212

Ion	Ratio	Lower	Upper
162	100		
127	41.3	31.9	47.9
164	32.8	26.2	39.4





#48
2-Nitroaniline
Concen: 66.042 ng
RT: 13.439 min Scan# 1760
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion: 65 Resp: 781584

Ion	Ratio	Lower	Upper
65	100		
92	75.9	62.3	93.5
138	123.3	97.2	145.8





#49
Acenaphthylene
Concen: 61.143 ng
RT: 14.087 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 4111553

Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.9	23.9
153	13.1	10.6	15.8





#50
Dimethylphthalate
Concen: 60.022 ng
RT: 13.828 min Scan# 1826
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:163 Resp: 3404463

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	10.1	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 63.998 ng
RT: 13.939 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 764810

Ion	Ratio	Lower	Upper
165	100		
63	45.9	38.0	57.0
89	51.1	41.7	62.5





#52
Acenaphthene
Concen: 58.858 ng
RT: 14.434 min Scan# 1929
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:154 Resp: 2516981

Ion	Ratio	Lower	Upper
154	100		
153	113.6	89.8	134.6
152	53.7	42.6	63.8





#53
3-Nitroaniline
Concen: 64.022 ng
RT: 14.275 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:138 Resp: 793796

Ion	Ratio	Lower	Upper
138	100		
108	10.7	8.3	12.5
92	111.3	89.8	134.8





#54
2,4-Dinitrophenol
Concen: 76.655 ng
RT: 14.486 min Scan# 1938
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:184 Resp: 490497

Ion	Ratio	Lower	Upper
184	100		
63	50.8	42.6	64.0
154	57.7	47.2	70.8





#55
Dibenzofuran
Concen: 58.810 ng
RT: 14.775 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:168 Resp: 4124640

Ion	Ratio	Lower	Upper
168	100		
139	37.9	30.3	45.5
169	13.3	10.6	16.0





#56
4-Nitrophenol
Concen: 70.748 ng
RT: 14.598 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:139 Resp: 754038

Ion	Ratio	Lower	Upper
139	100		
109	66.1	46.1	86.1
65	82.9	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 66.836 ng
RT: 14.739 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 1079633

Ion	Ratio	Lower	Upper
165	100		
63	36.6	30.4	45.6
89	66.4	54.6	82.0
182	2.9	2.2	3.2





#58
Fluorene
Concen: 58.817 ng
RT: 15.434 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:166 Resp: 3184960

Ion	Ratio	Lower	Upper
166	100		
165	97.2	78.1	117.1
167	13.4	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 63.097 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:232 Resp: 909497

Ion	Ratio	Lower	Upper
232	100		
131	46.4	36.8	55.2
130	2.3	1.9	2.9
166	31.4	25.0	37.6





#60
Diethylphthalate
Concen: 59.879 ng
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:149 Resp: 3495577

Ion	Ratio	Lower	Upper
149	100		
177	21.9	17.4	26.2
150	12.9	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 59.442 ng
RT: 15.428 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:204 Resp: 1563042

Ion	Ratio	Lower	Upper
204	100		
206	33.3	26.5	39.7
141	61.1	49.4	74.0





#62
4-Nitroaniline
Concen: 67.414 ng
RT: 15.457 min Scan# 2103
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:138 Resp: 886163

Ion	Ratio	Lower	Upper
138	100		
92	48.6	27.8	67.8
108	70.7	48.9	88.9





#63
Azobenzene
Concen: 60.070 ng
RT: 15.728 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 3229873

Ion	Ratio	Lower	Upper
77	100		
182	25.5	7.3	47.3
105	22.1	2.3	42.3
51	24.3	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.169 min Scan# 2394
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:188 Resp: 1536805

Ion	Ratio	Lower	Upper
188	100		
94	10.7	8.6	13.0
80	12.0	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 72.083 ng
RT: 15.516 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 647253

Ion	Ratio	Lower	Upper
198	100		
51	34.1	18.4	58.4
105	42.6	26.5	66.5





#66
n-Nitrosodiphenylamine
Concen: 60.874 ng
RT: 15.651 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:169 Resp: 2805137

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.0	79.4
167	36.3	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 61.926 ng
RT: 16.339 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 1038071

Ion	Ratio	Lower	Upper
248	100		
250	96.7	78.2	117.4
141	74.2	57.7	86.5





#68
Hexachlorobenzene
Concen: 62.200 ng
RT: 16.457 min Scan# 2273
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:284 Resp: 1239550

Ion	Ratio	Lower	Upper
284	100		
142	32.9	24.6	37.0
249	30.7	23.8	35.6





#70
Pentachlorophenol
Concen: 68.445 ng
RT: 16.816 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:266 Resp: 936562

Ion	Ratio	Lower	Upper
266	100		
268	64.2	51.0	76.6
264	62.3	49.8	74.6





#71
Phenanthrene
Concen: 59.209 ng
RT: 17.222 min Scan# 2403
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:178 Resp: 4988840

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.4	23.0
179	15.3	12.5	18.7





#72
Anthracene
Concen: 60.907 ng
RT: 17.310 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:178 Resp: 4938117

Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.7	22.1
179	15.7	12.4	18.6





#73
Carbazole
Concen: 60.584 ng
RT: 17.592 min Scan# 2466
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:167 Resp: 4780408

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.1	25.7
139	13.7	10.6	16.0





#74
Di-n-butylphthalate
Concen: 62.730 ng
RT: 18.186 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 6167206

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	5.0	3.9	5.9





#75
Fluoranthene
Concen: 60.092 ng
RT: 19.304 min Scan# 2757
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:202 Resp: 6009748

Ion	Ratio	Lower	Upper
202	100		
101	11.9	0.0	33.2
203	17.6	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.621 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:240 Resp: 1596148

Ion	Ratio	Lower	Upper
240	100		
120	12.5	9.8	14.8
236	25.8	20.9	31.3





#77
Benzidine
Concen: 73.401 ng
RT: 19.504 min Scan# 2791
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:184 Resp: 1446166

Ion	Ratio	Lower	Upper
184	100		
185	14.4	11.6	17.4
183	11.5	9.2	13.8





#78
Pyrene
Concen: 61.204 ng
RT: 19.680 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:202 Resp: 6250781

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.2	25.8
203	17.9	14.4	21.6





#79
Terphenyl-d14
Concen: 115.262 ng
RT: 19.898 min Scan# 2858
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:244 Resp: 9228969

Ion	Ratio	Lower	Upper
244	100		
212	7.6	5.6	8.4
122	10.9	9.4	14.0





#80
Butylbenzylphthalate
Concen: 65.961 ng
RT: 20.621 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 2840001

Ion	Ratio	Lower	Upper
149	100		
91	72.0	56.4	84.6
206	20.7	16.3	24.5





#81
Benzo(a)anthracene
Concen: 60.318 ng
RT: 21.604 min Scan# 3148
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:228 Resp: 6002059

Ion	Ratio	Lower	Upper
228	100		
226	27.9	21.8	32.8
229	20.1	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 67.333 ng
RT: 21.521 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 2323294

Ion	Ratio	Lower	Upper
252	100		
254	64.8	51.7	77.5
126	11.5	8.8	13.2





#83
Chrysene
Concen: 59.924 ng
RT: 21.674 min Scan# 3160
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:228 Resp: 5714775

Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.2	36.2
229	20.2	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 65.497 ng
RT: 21.545 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 4153128

Ion	Ratio	Lower	Upper
149	100		
167	28.6	22.4	33.6
279	5.3	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 70.060 ng
RT: 22.839 min Scan# 3358
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:149 Resp: 7116217

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.2	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:264 Resp: 1858365

Ion	Ratio	Lower	Upper
264	100		
260	23.3	18.9	28.3
265	21.8	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 63.043 ng m
RT: 28.821 min Scan# 4375
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:276 Resp: 8057576

Ion	Ratio	Lower	Upper
276	100		
138	18.2	14.5	21.7
277	20.7	20.2	30.2





#88
Benzo(b)fluoranthene
Concen: 60.400 ng
RT: 23.927 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 6715444

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.0	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 60.244 ng
RT: 24.004 min Scan# 3556
Delta R.T. 0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:252 Resp: 6500782

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	10.4	9.1	13.7





#90
Benzo(a)pyrene
Concen: 62.316 ng
RT: 24.827 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 5964163

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	11.8	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 62.214 ng
RT: 28.892 min Scan# 4387
Delta R.T. -0.012 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Tgt Ion:278 Resp: 6614173

Ion	Ratio	Lower	Upper
278	100		
139	16.6	12.6	19.0
279	24.1	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 62.294 ng
RT: 30.003 min Scan# 4
Delta R.T. 0.017 min
Lab File: BP024281.D
Acq: 14 Apr 2025 16:32

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:276 Resp: 6740929

Ion	Ratio	Lower	Upper
276	100		
277	24.0	18.9	28.3
138	21.5	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024282.D
Acq On : 14 Apr 2025 17:13
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Apr 14 18:14:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.728	152	244732	20.000	ng		0.00
21) Naphthalene-d8	10.504	136	1024316	20.000	ng		0.00
39) Acenaphthene-d10	14.369	164	641442	20.000	ng		0.00
64) Phenanthrene-d10	17.180	188	1279313	20.000	ng		0.02
76) Chrysene-d12	21.621	240	1378105	20.000	ng		0.00
86) Perylene-d12	24.986	264	1583281	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.352	112	2341577	157.905	ng		0.00
7) Phenol-d6	6.916	99	3224074	158.927	ng		0.00
23) Nitrobenzene-d5	8.881	82	2847003	158.236	ng		0.00
42) 2,4,6-Tribromophenol	15.892	330	1507534	171.633	ng		0.01
45) 2-Fluorobiphenyl	12.981	172	6362479	150.596	ng		0.01
79) Terphenyl-d14	19.922	244	10209149	147.678	ng		0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.293	88	467234	73.805	ng		99
3) Pyridine	3.687	79	1349044m	81.864	ng		
4) n-Nitrosodimethylamine	3.593	42	434861	79.138	ng		97
6) Aniline	7.069	93	1351337	73.789	ng		100
8) 2-Chlorophenol	7.305	128	1299769	78.433	ng		99
9) Benzaldehyde	6.881	77	608272	57.891	ng		99
10) Phenol	6.946	94	1608027	78.944	ng		98
11) bis(2-Chloroethyl)ether	7.163	93	1272327	77.297	ng		99
12) 1,3-Dichlorobenzene	7.622	146	1344292	74.869	ng		100
13) 1,4-Dichlorobenzene	7.763	146	1369087	75.195	ng		99
14) 1,2-Dichlorobenzene	8.075	146	1310407	74.432	ng		100
15) Benzyl Alcohol	7.975	79	1098602	84.492	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.257	45	1319830	73.704	ng		98
17) 2-Methylphenol	8.175	107	1064555	81.101	ng		100
18) Hexachloroethane	8.799	117	498897	75.920	ng		99
19) n-Nitroso-di-n-propyla...	8.534	70	968278	76.844	ng		99
20) 3+4-Methylphenols	8.499	107	1492527	82.092	ng		99
22) Acetophenone	8.546	105	1977578	77.680	ng	#	100
24) Nitrobenzene	8.922	77	1392093	78.066	ng		98
25) Isophorone	9.446	82	2617853	80.596	ng		99
26) 2-Nitrophenol	9.628	139	760077	88.768	ng		98
27) 2,4-Dimethylphenol	9.693	122	905891	83.475	ng		99
28) bis(2-Chloroethoxy)met...	9.922	93	1689686	76.434	ng		99
29) 2,4-Dichlorophenol	10.157	162	1238253	83.719	ng		99
30) 1,2,4-Trichlorobenzene	10.369	180	1252476	78.272	ng		99
31) Naphthalene	10.557	128	4190486	77.287	ng		100
32) Benzoic acid	9.875	122	1154243m	98.425	ng		
33) 4-Chloroaniline	10.669	127	1547427	81.682	ng		99
34) Hexachlorobutadiene	10.840	225	743811	79.240	ng		99
35) Caprolactam	11.481	113	469910	86.534	ng		98
36) 4-Chloro-3-methylphenol	11.804	107	1438601	83.468	ng		98
37) 2-Methylnaphthalene	12.169	142	2918685	77.673	ng		99
38) 1-Methylnaphthalene	12.393	142	2831173	76.912	ng		100
40) 1,2,4,5-Tetrachloroben...	12.545	216	1416894	80.378	ng		100
41) Hexachlorocyclopentadiene	12.528	237	523136	84.532	ng		99
43) 2,4,6-Trichlorophenol	12.787	196	1000309	86.244	ng		97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024282.D
Acq On : 14 Apr 2025 17:13
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Apr 14 18:14:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.869	196	1103073	85.819	ng	98
46) 1,1'-Biphenyl	13.198	154	3665607	77.381	ng	99
47) 2-Chloronaphthalene	13.234	162	2766302	78.257	ng	99
48) 2-Nitroaniline	13.445	65	848619	86.912	ng	98
49) Acenaphthylene	14.087	152	4438648	80.004	ng	100
50) Dimethylphthalate	13.834	163	3659579	78.201	ng	99
51) 2,6-Dinitrotoluene	13.945	165	814314	82.589	ng	98
52) Acenaphthene	14.439	154	2761536	78.270	ng	100
53) 3-Nitroaniline	14.281	138	919775	89.913	ng	99
54) 2,4-Dinitrophenol	14.492	184	541483	102.567	ng	99
55) Dibenzofuran	14.781	168	4420987	76.402	ng	99
56) 4-Nitrophenol	14.604	139	814261	92.599	ng	98
57) 2,4-Dinitrotoluene	14.745	165	1169554	87.755	ng	98
58) Fluorene	15.439	166	3479094	77.873	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.004	232	999335	84.031	ng	100
60) Diethylphthalate	15.222	149	3800651	78.910	ng	99
61) 4-Chlorophenyl-phenyle...	15.439	204	1689620	77.881	ng	100
62) 4-Nitroaniline	15.469	138	965926	89.063	ng	98
63) Azobenzene	15.739	77	3448680	77.741	ng	98
65) 4,6-Dinitro-2-methylph...	15.528	198	705506	94.385	ng	92
66) n-Nitrosodiphenylamine	15.657	169	2970516	77.438	ng	99
67) 4-Bromophenyl-phenylether	16.345	248	1137128	81.488	ng	99
68) Hexachlorobenzene	16.469	284	1347567	81.231	ng	97
70) Pentachlorophenol	16.828	266	1027691	90.221	ng	99
71) Phenanthrene	17.228	178	5415046	77.203	ng	100
72) Anthracene	17.316	178	5270987	78.097	ng	99
73) Carbazole	17.604	167	5275449	80.314	ng	100
74) Di-n-butylphthalate	18.192	149	6569078	80.266	ng	100
75) Fluoranthene	19.316	202	6520527	78.322	ng	99
77) Benzidine	19.510	184	1694694m	99.624	ng	
78) Pyrene	19.686	202	6718789	76.195	ng	100
80) Butylbenzylphthalate	20.645	149	3163633	85.104	ng	96
81) Benzo(a)anthracene	21.604	228	6729971	78.335	ng	99
82) 3,3'-Dichlorobenzidine	21.539	252	2536509	85.143	ng	99
83) Chrysene	21.680	228	6433173	78.130	ng	99
84) Bis(2-ethylhexyl)phtha...	21.545	149	4516949	82.506	ng	98
85) Di-n-octyl phthalate	22.833	149	7969779	90.878	ng	99
87) Indeno(1,2,3-cd)pyrene	28.815	276	9286186	85.280	ng	99
88) Benzo(b)fluoranthene	23.933	252	7677613	81.051	ng	98
89) Benzo(k)fluoranthene	23.998	252	7207214	78.395	ng	100
90) Benzo(a)pyrene	24.833	252	6703562	82.211	ng	99
91) Dibenzo(a,h)anthracene	28.886	278	7614901	84.071	ng	98
92) Benzo(g,h,i)perylene	29.997	276	7750941	84.072	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024282.D
Acq On : 14 Apr 2025 17:13
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Quant Time: Apr 14 18:14:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 17:03:26 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.728 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 244732

Ion	Ratio	Lower	Upper
152	100		
150	157.4	124.9	187.3
115	61.0	47.7	71.5





#2
1,4-Dioxane
Concen: 73.805 ng
RT: 3.293 min Scan# 35
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 88 Resp: 467234

Ion	Ratio	Lower	Upper
88	100		
58	64.8	51.4	77.2
43	21.7	17.1	25.7





#3
Pyridine
Concen: 81.864 ng m
RT: 3.687 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 1349044

Ion	Ratio	Lower	Upper
79	100		
52	57.7	45.4	68.2
51	28.3	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 79.138 ng
RT: 3.593 min Scan# 86
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 42 Resp: 434861

Ion	Ratio	Lower	Upper
42	100		
74	170.0	139.5	209.3
44	6.8	5.8	8.8





#5
2-Fluorophenol
Concen: 157.905 ng
RT: 5.352 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:112 Resp: 2341577

Ion	Ratio	Lower	Upper
112	100		
64	51.6	41.1	61.7
63	26.0	20.6	30.8





#6
Aniline
Concen: 73.789 ng
RT: 7.069 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 93 Resp: 1351337

Ion	Ratio	Lower	Upper
93	100		
66	36.1	29.1	43.7
65	18.2	14.7	22.1





#7
Phenol-d6
Concen: 158.927 ng
RT: 6.916 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 99 Resp: 3224074

Ion	Ratio	Lower	Upper
99	100		
42	13.9	10.6	16.0
71	31.1	23.7	35.5





#8
2-Chlorophenol
Concen: 78.433 ng
RT: 7.305 min Scan# 717
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:128 Resp: 1299769

Ion	Ratio	Lower	Upper
128	100		
130	32.6	12.9	52.9
64	42.1	22.7	62.7





#9
Benzaldehyde
Concen: 57.891 ng
RT: 6.881 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 608272

Ion	Ratio	Lower	Upper
77	100		
105	101.2	79.2	119.2
106	96.2	76.2	116.2





#10
Phenol
Concen: 78.944 ng
RT: 6.946 min Scan# 656
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 94 Resp: 1608027

Ion	Ratio	Lower	Upper
94	100		
65	25.1	4.5	44.5
66	33.5	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 77.297 ng
RT: 7.163 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 93 Resp: 1272327

Ion	Ratio	Lower	Upper
93	100		
63	68.2	48.6	88.6
95	31.9	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 74.869 ng
RT: 7.622 min Scan# 771
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:146 Resp: 1344292

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.0
75	28.4	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 75.195 ng
RT: 7.763 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:146 Resp: 1369087

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.1	76.7
111	40.7	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 74.432 ng
RT: 8.075 min Scan# 848
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:146 Resp: 1310407

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.4
111	42.7	33.8	50.8





#15
Benzyl Alcohol
Concen: 84.492 ng
RT: 7.975 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 79 Resp: 1098602

Ion	Ratio	Lower	Upper
79	100		
108	82.4	65.1	97.7
77	62.4	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 73.704 ng
RT: 8.257 min Scan# 879
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 45 Resp: 1319830

Ion	Ratio	Lower	Upper
45	100		
77	18.1	0.0	37.7
79	14.7	0.0	33.5





#17
2-Methylphenol
Concen: 81.101 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:107 Resp: 1064555

Ion	Ratio	Lower	Upper
107	100		
108	111.5	89.8	134.6
77	43.0	34.2	51.4
79	42.3	33.8	50.8





#18
Hexachloroethane
Concen: 75.920 ng
RT: 8.799 min Scan# 971
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:117 Resp: 498897

Ion	Ratio	Lower	Upper
117	100		
119	98.3	77.8	116.6
201	96.5	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 76.844 ng
RT: 8.534 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 70 Resp: 968278

Ion	Ratio	Lower	Upper
70	100		
42	46.3	37.3	55.9
101	11.0	8.2	12.4
130	24.9	18.6	28.0





#20
3+4-Methylphenols
Concen: 82.092 ng
RT: 8.499 min Scan# 920
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:107 Resp: 1492527

Ion	Ratio	Lower	Upper
107	100		
108	89.0	69.8	109.8
77	30.4	11.5	51.5
79	26.5	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.504 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:136 Resp: 1024316

Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.2	13.8
54	6.9	6.0	9.0
68	6.2	5.5	8.3





#22
Acetophenone
Concen: 77.680 ng
RT: 8.546 min Scan# 928
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:105 Resp: 1977578

Ion	Ratio	Lower	Upper
105	100		
71	4.0	1.7	2.5#
51	21.2	16.9	25.3
120	23.3	18.6	28.0





#23
Nitrobenzene-d5
Concen: 158.236 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 2847003

Ion	Ratio	Lower	Upper
82	100		
128	47.0	36.5	54.7
54	42.3	33.4	50.2





#24
Nitrobenzene
Concen: 78.066 ng
RT: 8.922 min Scan# 992
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 77 Resp: 1392093

Ion	Ratio	Lower	Upper
77	100		
123	50.5	39.2	58.8
65	12.9	9.8	14.6





#25
Isophorone
Concen: 80.596 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 82 Resp: 2617853

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.4	8.2
138	18.4	14.4	21.6





#26
2-Nitrophenol
Concen: 88.768 ng
RT: 9.628 min Scan# 1112
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:139 Resp: 760077

Ion	Ratio	Lower	Upper
139	100		
109	27.5	21.8	32.6
65	43.6	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 83.475 ng
RT: 9.693 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:122 Resp: 905891

Ion	Ratio	Lower	Upper
122	100		
107	116.1	93.7	140.5
121	59.3	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 76.434 ng
RT: 9.922 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 93 Resp: 1689686

Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.0	39.0
123	12.2	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 83.719 ng
RT: 10.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 1238253

Ion	Ratio	Lower	Upper
162	100		
164	64.3	43.3	83.3
98	37.2	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 78.272 ng
RT: 10.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:180 Resp: 1252476

Ion	Ratio	Lower	Upper
180	100		
182	95.5	77.3	115.9
145	30.9	25.1	37.7





#31
Naphthalene
Concen: 77.287 ng
RT: 10.557 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:128 Resp: 4190486

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.6	13.0
127	13.0	10.3	15.5





#32
Benzoic acid
Concen: 98.425 ng m
RT: 9.875 min Scan# 1154
Delta R.T. 0.047 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:122 Resp: 1154243

Ion	Ratio	Lower	Upper
122	100		
105	122.9	104.8	144.8
77	89.8	69.2	109.2





#33
4-Chloroaniline
Concen: 81.682 ng
RT: 10.669 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:127 Resp: 1547427

Ion	Ratio	Lower	Upper
127	100		
129	32.0	26.1	39.1
65	26.8	22.1	33.1
92	18.5	15.0	22.4





#34
Hexachlorobutadiene
Concen: 79.240 ng
RT: 10.840 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:225 Resp: 743811

Ion	Ratio	Lower	Upper
225	100		
223	63.0	49.5	74.3
227	64.3	50.9	76.3





#35
Caprolactam
Concen: 86.534 ng
RT: 11.481 min Scan# 14
Delta R.T. 0.029 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:113 Resp: 469910

Ion	Ratio	Lower	Upper
113	100		
55	131.9	115.9	155.9
56	99.8	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 83.468 ng
RT: 11.804 min Scan# 1482
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:107 Resp: 1438601

Ion	Ratio	Lower	Upper
107	100		
144	26.4	20.5	30.7
142	81.0	63.4	95.2





#37
2-Methylnaphthalene
Concen: 77.673 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:142 Resp: 2918685

Ion	Ratio	Lower	Upper
142	100		
141	88.4	70.3	105.5
115	34.9	26.9	40.3





#38
1-Methylnaphthalene
Concen: 76.912 ng
RT: 12.393 min Scan# 1582
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:142 Resp: 2831173

Ion	Ratio	Lower	Upper
142	100		
141	91.8	73.6	110.4
116	3.8	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.369 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:164 Resp: 641442

Ion	Ratio	Lower	Upper
164	100		
162	99.6	80.6	120.8
160	43.9	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 80.378 ng
RT: 12.545 min Scan# 1608
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:216 Resp: 1416894

Ion	Ratio	Lower	Upper
216	100		
214	78.4	62.6	93.8
179	21.5	17.2	25.8
108	18.6	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 84.532 ng
RT: 12.528 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:237 Resp: 523136

Ion	Ratio	Lower	Upper
237	100		
235	62.9	43.9	83.9
272	13.7	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 171.633 ng
RT: 15.892 min Scan# 2177
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:330 Resp: 1507534

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.1	115.7
141	29.7	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 86.244 ng
RT: 12.787 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:196 Resp: 1000309

Ion	Ratio	Lower	Upper
196	100		
198	96.5	74.3	111.5
200	31.2	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 85.819 ng
RT: 12.869 min Scan# 1663
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:196 Resp: 1103073

Ion	Ratio	Lower	Upper
196	100		
198	95.0	77.3	115.9
97	47.7	40.4	60.6
132	28.6	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 150.596 ng
RT: 12.981 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:172 Resp: 6362479

Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.2	42.2
170	23.8	18.6	27.8





#46
1,1'-Biphenyl
Concen: 77.381 ng
RT: 13.198 min Scan# 1719
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:154 Resp: 3665607

Ion	Ratio	Lower	Upper
154	100		
153	40.7	20.3	60.3
76	13.2	0.0	33.5





#47
2-Chloronaphthalene
Concen: 78.257 ng
RT: 13.234 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:162 Resp: 2766302

Ion	Ratio	Lower	Upper
162	100		
127	40.9	31.9	47.9
164	32.8	26.2	39.4





#48
2-Nitroaniline
Concen: 86.912 ng
RT: 13.445 min Scan# 1761
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion: 65 Resp: 848619

Ion	Ratio	Lower	Upper
65	100		
92	76.5	62.3	93.5
138	123.5	97.2	145.8





#49
Acenaphthylene
Concen: 80.004 ng
RT: 14.087 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:152 Resp: 4438648

Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.9	23.9
153	13.2	10.6	15.8





#50
Dimethylphthalate
Concen: 78.201 ng
RT: 13.834 min Scan# 1827
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:163 Resp: 3659579

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.5	5.3
164	10.3	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 82.589 ng
RT: 13.945 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 814314

Ion	Ratio	Lower	Upper
165	100		
63	46.4	38.0	57.0
89	51.0	41.7	62.5





#52
Acenaphthene
Concen: 78.270 ng
RT: 14.439 min Scan# 1930
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:154 Resp: 2761536

Ion	Ratio	Lower	Upper
154	100		
153	112.7	89.8	134.6
152	53.5	42.6	63.8





#53
3-Nitroaniline
Concen: 89.913 ng
RT: 14.281 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:138 Resp: 919775

Ion	Ratio	Lower	Upper
138	100		
108	10.5	8.3	12.5
92	111.0	89.8	134.8





#54
2,4-Dinitrophenol
Concen: 102.567 ng
RT: 14.492 min Scan# 1939
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:184 Resp: 541483

Ion	Ratio	Lower	Upper
184	100		
63	53.6	42.6	64.0
154	58.2	47.2	70.8





#55
Dibenzofuran
Concen: 76.402 ng
RT: 14.781 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:168 Resp: 4420987

Ion	Ratio	Lower	Upper
168	100		
139	38.7	30.3	45.5
169	13.3	10.6	16.0





#56
4-Nitrophenol
Concen: 92.599 ng
RT: 14.604 min Scan# 1958
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:139 Resp: 814261

Ion	Ratio	Lower	Upper
139	100		
109	68.4	46.1	86.1
65	86.6	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 87.755 ng
RT: 14.745 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:165 Resp: 1169554

Ion	Ratio	Lower	Upper
165	100		
63	36.2	30.4	45.6
89	66.4	54.6	82.0
182	2.9	2.2	3.2





#58
Fluorene
Concen: 77.873 ng
RT: 15.439 min Scan# 2100
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:166 Resp: 3479094

Ion	Ratio	Lower	Upper
166	100		
165	96.6	78.1	117.1
167	13.7	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 84.031 ng
RT: 15.004 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:232 Resp: 999335

Ion	Ratio	Lower	Upper
232	100		
131	46.1	36.8	55.2
130	2.4	1.9	2.9
166	31.4	25.0	37.6





#60
Diethylphthalate
Concen: 78.910 ng
RT: 15.222 min Scan# 2063
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:149 Resp: 3800651

Ion	Ratio	Lower	Upper
149	100		
177	22.6	17.4	26.2
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 77.881 ng
RT: 15.439 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:204 Resp: 1689620

Ion	Ratio	Lower	Upper
204	100		
206	33.3	26.5	39.7
141	61.4	49.4	74.0





#62
4-Nitroaniline
Concen: 89.063 ng
RT: 15.469 min Scan# 2105
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:138 Resp: 965926

Ion	Ratio	Lower	Upper
138	100		
92	48.5	27.8	67.8
108	71.7	48.9	88.9





#63
Azobenzene
Concen: 77.741 ng
RT: 15.739 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion: 77 Resp: 3448680

Ion	Ratio	Lower	Upper
77	100		
182	29.0	7.3	47.3
105	22.7	2.3	42.3
51	24.5	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.180 min Scan# 2396
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:188 Resp: 1279313

Ion	Ratio	Lower	Upper
188	100		
94	10.4	8.6	13.0
80	11.2	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 94.385 ng
RT: 15.528 min Scan# 21
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:198 Resp: 705506

Ion	Ratio	Lower	Upper
198	100		
51	34.3	18.4	58.4
105	41.0	26.5	66.5





#66
n-Nitrosodiphenylamine
Concen: 77.438 ng
RT: 15.657 min Scan# 2137
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:169 Resp: 2970516

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.0	79.4
167	36.3	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 81.488 ng
RT: 16.345 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:248 Resp: 1137128

Ion	Ratio	Lower	Upper
248	100		
250	98.5	78.2	117.4
141	73.2	57.7	86.5





#68
Hexachlorobenzene
Concen: 81.231 ng
RT: 16.469 min Scan# 2275
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:284 Resp: 1347567

Ion	Ratio	Lower	Upper
284	100		
142	33.3	24.6	37.0
249	30.5	23.8	35.6





#70
Pentachlorophenol
Concen: 90.221 ng
RT: 16.828 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:266 Resp: 1027691

Ion	Ratio	Lower	Upper
266	100		
268	64.0	51.0	76.6
264	63.1	49.8	74.6





#71
Phenanthrene
Concen: 77.203 ng
RT: 17.228 min Scan# 2404
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:178 Resp: 5415046

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.4	23.0
179	15.6	12.5	18.7





#72
Anthracene
Concen: 78.097 ng
RT: 17.316 min Scan# 24
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:178 Resp: 5270987

Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.7	22.1
179	15.6	12.4	18.6





#73
Carbazole
Concen: 80.314 ng
RT: 17.604 min Scan# 2468
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:167 Resp: 5275449

Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.1	25.7
139	13.2	10.6	16.0





#74
Di-n-butylphthalate
Concen: 80.266 ng
RT: 18.192 min Scan# 2
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 6569078

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.4	11.2
104	4.9	3.9	5.9





#75
Fluoranthene
Concen: 78.322 ng
RT: 19.316 min Scan# 2759
Delta R.T. 0.017 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:202 Resp: 6520527

Ion	Ratio	Lower	Upper
202	100		
101	12.6	0.0	33.2
203	17.9	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.621 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:240 Resp: 1378105

Ion	Ratio	Lower	Upper
240	100		
120	11.8	9.8	14.8
236	25.5	20.9	31.3





#77
Benzidine
Concen: 99.624 ng m
RT: 19.510 min Scan# 2792
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:184 Resp: 1694694

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.6	17.4
183	11.8	9.2	13.8





#78
Pyrene
Concen: 76.195 ng
RT: 19.686 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:202 Resp: 6718789

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.2	25.8
203	18.1	14.4	21.6





#79
Terphenyl-d14
Concen: 147.678 ng
RT: 19.922 min Scan# 2862
Delta R.T. 0.023 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:244 Resp:10209149

Ion	Ratio	Lower	Upper
244	100		
212	7.5	5.6	8.4
122	10.4	9.4	14.0





#80
Butylbenzylphthalate
Concen: 85.104 ng
RT: 20.645 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 3163633

Ion	Ratio	Lower	Upper
149	100		
91	67.1	56.4	84.6
206	22.0	16.3	24.5





#81
Benzo(a)anthracene
Concen: 78.335 ng
RT: 21.604 min Scan# 3148
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:228 Resp: 6729971

Ion	Ratio	Lower	Upper
228	100		
226	27.8	21.8	32.8
229	20.1	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 85.143 ng
RT: 21.539 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 2536509

Ion	Ratio	Lower	Upper
252	100		
254	65.1	51.7	77.5
126	10.0	8.8	13.2





#83
Chrysene
Concen: 78.130 ng
RT: 21.680 min Scan# 3161
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:228 Resp: 6433173

Ion	Ratio	Lower	Upper
228	100		
226	30.5	24.2	36.2
229	20.0	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 82.506 ng
RT: 21.545 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:149 Resp: 4516949

Ion	Ratio	Lower	Upper
149	100		
167	28.8	22.4	33.6
279	5.4	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 90.878 ng
RT: 22.833 min Scan# 3357
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:149 Resp: 7969779

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.3	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:264 Resp: 1583281

Ion	Ratio	Lower	Upper
264	100		
260	23.8	18.9	28.3
265	22.3	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 85.280 ng
RT: 28.815 min Scan# 4374
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:276 Resp: 9286186

Ion	Ratio	Lower	Upper
276	100		
138	17.4	14.5	21.7
277	25.4	20.2	30.2





#88
Benzo(b)fluoranthene
Concen: 81.051 ng
RT: 23.933 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 7677613

Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.7	26.5
125	10.4	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 78.395 ng
RT: 23.998 min Scan# 3555
Delta R.T. 0.006 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:252 Resp: 7207214

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.1	9.1	13.7





#90
Benzo(a)pyrene
Concen: 82.211 ng
RT: 24.833 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:252 Resp: 6703562

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	11.8	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 84.071 ng
RT: 28.886 min Scan# 4386
Delta R.T. -0.018 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Tgt Ion:278 Resp: 7614901

Ion	Ratio	Lower	Upper
278	100		
139	16.2	12.6	19.0
279	24.2	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 84.072 ng
RT: 29.997 min Scan# 4
Delta R.T. 0.012 min
Lab File: BP024282.D
Acq: 14 Apr 2025 17:13

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:276 Resp: 7750941

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.9	28.3
138	22.4	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	232241	20.000	ng	0.00
21) Naphthalene-d8	10.504	136	956448	20.000	ng	0.00
39) Acenaphthene-d10	14.363	164	600031	20.000	ng	0.00
64) Phenanthrene-d10	17.163	188	1178916	20.000	ng	-0.02
76) Chrysene-d12	21.616	240	1288125	20.000	ng	0.00
86) Perylene-d12	24.986	264	1476872	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	1141956	81.302	ng	0.00
7) Phenol-d6	6.911	99	1559050	81.063	ng	0.00
23) Nitrobenzene-d5	8.881	82	1401115	83.531	ng	0.00
42) 2,4,6-Tribromophenol	15.875	330	709703	85.488	ng	-0.02
45) 2-Fluorobiphenyl	12.975	172	3319933	84.715	ng	0.00
79) Terphenyl-d14	19.892	244	5217888	81.649	ng	-0.03
Target Compounds						Qvalue
2) 1,4-Dioxane	3.293	88	240091	40.412	ng	100
3) Pyridine	3.687	79	626813	39.955	ng	99
4) n-Nitrosodimethylamine	3.593	42	209326	40.205	ng	98
6) Aniline	7.069	93	710981	41.370	ng	100
8) 2-Chlorophenol	7.305	128	632515	40.334	ng	99
9) Benzaldehyde	6.881	77	399803	42.024	ng	98
10) Phenol	6.940	94	777631	40.306	ng	100
11) bis(2-Chloroethyl)ether	7.163	93	620261	39.902	ng	99
12) 1,3-Dichlorobenzene	7.622	146	680343	40.298	ng	100
13) 1,4-Dichlorobenzene	7.769	146	694020	40.516	ng	99
14) 1,2-Dichlorobenzene	8.081	146	659947	39.898	ng	99
15) Benzyl Alcohol	7.969	79	517678	41.621	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.258	45	674999	40.173	ng	99
17) 2-Methylphenol	8.175	107	512292	41.046	ng	100
18) Hexachloroethane	8.799	117	250852	40.522	ng	99
19) n-Nitroso-di-n-propyla...	8.528	70	481540	40.471	ng	99
20) 3+4-Methylphenols	8.499	107	706733	40.810	ng	99
22) Acetophenone	8.546	105	986019	41.652	ng	100
24) Nitrobenzene	8.922	77	689231	41.537	ng	100
25) Isophorone	9.446	82	1264480	41.648	ng	99
26) 2-Nitrophenol	9.628	139	352229	43.376	ng	99
27) 2,4-Dimethylphenol	9.687	122	425062	41.689	ng	99
28) bis(2-Chloroethoxy)met...	9.922	93	849161	41.402	ng	99
29) 2,4-Dichlorophenol	10.163	162	590306	42.461	ng	98
30) 1,2,4-Trichlorobenzene	10.369	180	619818	41.612	ng	99
31) Naphthalene	10.557	128	2085090	41.386	ng	99
32) Benzoic acid	9.834	122	486627	40.959	ng	98
33) 4-Chloroaniline	10.669	127	761817	42.937	ng	99
34) Hexachlorobutadiene	10.846	225	367684	42.007	ng	100
35) Caprolactam	11.457	113	216229	42.152	ng	99
36) 4-Chloro-3-methylphenol	11.799	107	675857	41.737	ng	98
37) 2-Methylnaphthalene	12.169	142	1446647	41.403	ng	100
38) 1-Methylnaphthalene	12.387	142	1408633	41.210	ng	99
40) 1,2,4,5-Tetrachloroben...	12.540	216	697580	42.275	ng	100
41) Hexachlorocyclopentadiene	12.516	237	254676	43.639	ng	99
43) 2,4,6-Trichlorophenol	12.781	196	475349	43.329	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.863	196	520298	42.828	ng	97
46) 1,1'-Biphenyl	13.187	154	1834570	41.595	ng	99
47) 2-Chloronaphthalene	13.228	162	1377424	41.786	ng	100
48) 2-Nitroaniline	13.434	65	392742	42.475	ng	98
49) Acenaphthylene	14.081	152	2168474	41.783	ng	99
50) Dimethylphthalate	13.816	163	1809923	41.478	ng	99
51) 2,6-Dinitrotoluene	13.940	165	398345	42.990	ng	97
52) Acenaphthene	14.428	154	1344151	40.853	ng	99
53) 3-Nitroaniline	14.275	138	427828	43.931	ng	98
54) 2,4-Dinitrophenol	14.481	184	231882	42.485	ng	98
55) Dibenzofuran	14.763	168	2187786	40.679	ng	99
56) 4-Nitrophenol	14.592	139	378214	44.968	ng	99
57) 2,4-Dinitrotoluene	14.734	165	550175	43.527	ng	97
58) Fluorene	15.428	166	1738215	41.750	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.998	232	470547	41.995	ng	99
60) Diethylphthalate	15.198	149	1869274	41.570	ng	99
61) 4-Chlorophenyl-phenyle...	15.422	204	844629	41.777	ng	99
62) 4-Nitroaniline	15.451	138	444482	43.114	ng	96
63) Azobenzene	15.722	77	1722900	41.686	ng	98
65) 4,6-Dinitro-2-methylph...	15.510	198	317548	42.724	ng	95
66) n-Nitrosodiphenylamine	15.645	169	1460282	41.500	ng	100
67) 4-Bromophenyl-phenylether	16.334	248	545579	42.314	ng	97
68) Hexachlorobenzene	16.451	284	642095	41.909	ng	96
69) Atrazine	16.622	200	282908	31.567	ng	98
70) Pentachlorophenol	16.810	266	458529	42.900	ng	99
71) Phenanthrene	17.204	178	2612052	40.615	ng	100
72) Anthracene	17.298	178	2606508	42.051	ng	100
73) Carbazole	17.581	167	2564955	42.351	ng	99
74) Di-n-butylphthalate	18.175	149	3360635	44.539	ng	100
75) Fluoranthene	19.298	202	3195045	41.771	ng	99
77) Benzidine	19.492	184	588487	36.041	ng	100
78) Pyrene	19.675	202	3276290	40.022	ng	99
80) Butylbenzylphthalate	20.627	149	1511473	43.107	ng	97
81) Benzo(a)anthracene	21.598	228	3299250	41.207	ng	99
82) 3,3'-Dichlorobenzidine	21.516	252	1212290	43.139	ng	99
83) Chrysene	21.669	228	3136332	40.888	ng	100
84) Bis(2-ethylhexyl)phtha...	21.539	149	2282222	44.400	ng	100
85) Di-n-octyl phthalate	22.827	149	3803266	45.513	ng	100
87) Indeno(1,2,3-cd)pyrene	28.798	276	4313548m	42.071	ng	
88) Benzo(b)fluoranthene	23.933	252	3654984	41.288	ng	99
89) Benzo(k)fluoranthene	23.998	252	3512834	41.081	ng	100
90) Benzo(a)pyrene	24.833	252	3196130	41.855	ng	99
91) Dibenzo(a,h)anthracene	28.880	278	3566000	41.902	ng	98
92) Benzo(g,h,i)perylene	29.992	276	3590938	41.455	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.734 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:152 Resp: 232241

Ion	Ratio	Lower	Upper
152	100		
150	156.8	124.9	187.3
115	58.6	47.7	71.5





#2
1,4-Dioxane
Concen: 40.412 ng
RT: 3.293 min Scan# 35
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 88 Resp: 240091

Ion	Ratio	Lower	Upper
88	100		
58	64.4	51.4	77.2
43	21.7	17.1	25.7





#3
Pyridine
Concen: 39.955 ng
RT: 3.687 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 79 Resp: 626813

Ion	Ratio	Lower	Upper
79	100		
52	58.1	45.4	68.2
51	28.0	22.6	33.8





#4
n-Nitrosodimethylamine
Concen: 40.205 ng
RT: 3.593 min Scan# 86
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 42 Resp: 209326

Ion	Ratio	Lower	Upper
42	100		
74	171.4	139.5	209.3
44	7.3	5.8	8.8





#5
2-Fluorophenol
Concen: 81.302 ng
RT: 5.346 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:112 Resp: 1141956

Ion	Ratio	Lower	Upper
112	100		
64	53.2	41.1	61.7
63	26.8	20.6	30.8





#6
Aniline
Concen: 41.370 ng
RT: 7.069 min Scan# 677
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 93 Resp: 710981

Ion	Ratio	Lower	Upper
93	100		
66	36.2	29.1	43.7
65	18.2	14.7	22.1





#7
Phenol-d6
Concen: 81.063 ng
RT: 6.911 min Scan# 6
Delta R.T. -0.005 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 99 Resp: 1559050

Ion	Ratio	Lower	Upper
99	100		
42	13.6	10.6	16.0
71	30.9	23.7	35.5





#8
2-Chlorophenol
Concen: 40.334 ng
RT: 7.305 min Scan# 717
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:128 Resp: 632515

Ion	Ratio	Lower	Upper
128	100		
130	32.3	12.9	52.9
64	41.9	22.7	62.7





#9
Benzaldehyde
Concen: 42.024 ng
RT: 6.881 min Scan# 64
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 77 Resp: 399803

Ion	Ratio	Lower	Upper
77	100		
105	101.7	79.2	119.2
106	98.0	76.2	116.2





#10
Phenol
Concen: 40.306 ng
RT: 6.940 min Scan# 655
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 94 Resp: 777631

Ion	Ratio	Lower	Upper
94	100		
65	24.7	4.5	44.5
66	32.5	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 39.902 ng
RT: 7.163 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 93 Resp: 620261

Ion	Ratio	Lower	Upper
93	100		
63	69.2	48.6	88.6
95	32.0	12.3	52.3





#12
1,3-Dichlorobenzene
Concen: 40.298 ng
RT: 7.622 min Scan# 771
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:146 Resp: 680343

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.4	77.0
75	28.2	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 40.516 ng
RT: 7.769 min Scan# 7
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:146 Resp: 694020

Ion	Ratio	Lower	Upper
146	100		
148	63.3	51.1	76.7
111	39.9	32.6	48.8





#14
1,2-Dichlorobenzene
Concen: 39.898 ng
RT: 8.081 min Scan# 849
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:146 Resp: 659947

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.0	76.4
111	43.3	33.8	50.8





#15
Benzyl Alcohol
Concen: 41.621 ng
RT: 7.969 min Scan# 8
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 79 Resp: 517678

Ion	Ratio	Lower	Upper
79	100		
108	81.4	65.1	97.7
77	63.1	50.7	76.1





#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.173 ng
RT: 8.258 min Scan# 879
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 45 Resp: 674999

Ion	Ratio	Lower	Upper
45	100		
77	17.8	0.0	37.7
79	13.9	0.0	33.5





#17
2-Methylphenol
Concen: 41.046 ng
RT: 8.175 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:107 Resp: 512292

Ion	Ratio	Lower	Upper
107	100		
108	111.8	89.8	134.6
77	43.2	34.2	51.4
79	42.5	33.8	50.8





#18
Hexachloroethane
Concen: 40.522 ng
RT: 8.799 min Scan# 971
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:117 Resp: 250852

Ion	Ratio	Lower	Upper
117	100		
119	95.2	77.8	116.6
201	95.1	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 40.471 ng
RT: 8.528 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 70 Resp: 481540

Ion	Ratio	Lower	Upper
70	100		
42	47.1	37.3	55.9
101	10.4	8.2	12.4
130	24.3	18.6	28.0





#20
3+4-Methylphenols
Concen: 40.810 ng
RT: 8.499 min Scan# 920
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:107 Resp: 706733

Ion	Ratio	Lower	Upper
107	100		
108	90.2	69.8	109.8
77	30.7	11.5	51.5
79	27.6	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.504 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:136 Resp: 956448

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	6.8	6.0	9.0
68	6.6	5.5	8.3





#22
Acetophenone
Concen: 41.652 ng
RT: 8.546 min Scan# 928
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:105 Resp: 986019

Ion	Ratio	Lower	Upper
105	100		
71	1.8	1.7	2.5
51	20.8	16.9	25.3
120	23.2	18.6	28.0





#23
Nitrobenzene-d5
Concen: 83.531 ng
RT: 8.881 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 82 Resp: 1401115

Ion	Ratio	Lower	Upper
82	100		
128	47.0	36.5	54.7
54	42.5	33.4	50.2





#24
Nitrobenzene
Concen: 41.537 ng
RT: 8.922 min Scan# 992
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 77 Resp: 689231

Ion	Ratio	Lower	Upper
77	100		
123	48.7	39.2	58.8
65	12.2	9.8	14.6





#25
Isophorone
Concen: 41.648 ng
RT: 9.446 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 82 Resp: 1264480

Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.4	8.2
138	18.8	14.4	21.6





#26
2-Nitrophenol
Concen: 43.376 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:139 Resp: 352229

Ion	Ratio	Lower	Upper
139	100		
109	28.2	21.8	32.6
65	45.3	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 41.689 ng
RT: 9.687 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:122 Resp: 425062

Ion	Ratio	Lower	Upper
122	100		
107	116.0	93.7	140.5
121	59.3	48.2	72.2





#28
bis(2-Chloroethoxy)methane
Concen: 41.402 ng
RT: 9.922 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 93 Resp: 849161

Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.0	39.0
123	11.7	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 42.461 ng
RT: 10.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:162 Resp: 590306

Ion	Ratio	Lower	Upper
162	100		
164	64.7	43.3	83.3
98	37.4	16.6	56.6





#30
1,2,4-Trichlorobenzene
Concen: 41.612 ng
RT: 10.369 min Scan# 1238
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:180 Resp: 619818

Ion	Ratio	Lower	Upper
180	100		
182	95.8	77.3	115.9
145	31.6	25.1	37.7





#31
Naphthalene
Concen: 41.386 ng
RT: 10.557 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:128 Resp: 2085090

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.6	13.0
127	13.2	10.3	15.5





#32
Benzoic acid
Concen: 40.959 ng
RT: 9.834 min Scan# 1147
Delta R.T. -0.041 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:122 Resp: 486627

Ion	Ratio	Lower	Upper
122	100		
105	122.5	104.8	144.8
77	90.6	69.2	109.2





#33
4-Chloroaniline
Concen: 42.937 ng
RT: 10.669 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:127 Resp: 761817

Ion	Ratio	Lower	Upper
127	100		
129	32.1	26.1	39.1
65	27.1	22.1	33.1
92	18.4	15.0	22.4





#34
Hexachlorobutadiene
Concen: 42.007 ng
RT: 10.846 min Scan# 1319
Delta R.T. 0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:225 Resp: 367684

Ion	Ratio	Lower	Upper
225	100		
223	61.9	49.5	74.3
227	63.4	50.9	76.3





#35
Caprolactam
Concen: 42.152 ng
RT: 11.457 min Scan# 14
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:113 Resp: 216229

Ion	Ratio	Lower	Upper
113	100		
55	136.8	115.9	155.9
56	102.3	81.1	121.1





#36
4-Chloro-3-methylphenol
Concen: 41.737 ng
RT: 11.799 min Scan# 1481
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:107 Resp: 675857

Ion	Ratio	Lower	Upper
107	100		
144	26.2	20.5	30.7
142	81.6	63.4	95.2





#37
2-Methylnaphthalene
Concen: 41.403 ng
RT: 12.169 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:142 Resp: 1446647

Ion	Ratio	Lower	Upper
142	100		
141	88.1	70.3	105.5
115	34.4	26.9	40.3





#38
1-Methylnaphthalene
Concen: 41.210 ng
RT: 12.387 min Scan# 1581
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:142 Resp: 1408633

Ion	Ratio	Lower	Upper
142	100		
141	92.7	73.6	110.4
116	3.7	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.363 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:164 Resp: 600031

Ion	Ratio	Lower	Upper
164	100		
162	98.5	80.6	120.8
160	43.2	35.3	52.9





#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.275 ng
RT: 12.540 min Scan# 1607
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:216 Resp: 697580

Ion	Ratio	Lower	Upper
216	100		
214	78.6	62.6	93.8
179	21.5	17.2	25.8
108	18.8	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 43.639 ng
RT: 12.516 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:237 Resp: 254676

Ion	Ratio	Lower	Upper
237	100		
235	63.3	43.9	83.9
272	13.9	0.0	33.6





#42
2,4,6-Tribromophenol
Concen: 85.488 ng
RT: 15.875 min Scan# 2174
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:330 Resp: 709703

Ion	Ratio	Lower	Upper
330	100		
332	96.7	77.1	115.7
141	30.7	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 43.329 ng
RT: 12.781 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:196 Resp: 475349

Ion	Ratio	Lower	Upper
196	100		
198	96.3	74.3	111.5
200	31.4	24.0	36.0





#44
2,4,5-Trichlorophenol
Concen: 42.828 ng
RT: 12.863 min Scan# 1662
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:196 Resp: 520298

Ion	Ratio	Lower	Upper
196	100		
198	95.0	77.3	115.9
97	47.4	40.4	60.6
132	28.1	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 84.715 ng
RT: 12.975 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:172 Resp: 3319933

Ion	Ratio	Lower	Upper
172	100		
171	34.9	28.2	42.2
170	23.3	18.6	27.8





#46
1,1'-Biphenyl
Concen: 41.595 ng
RT: 13.187 min Scan# 1717
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:154 Resp: 1834570

Ion	Ratio	Lower	Upper
154	100		
153	39.9	20.3	60.3
76	14.0	0.0	33.5





#47
2-Chloronaphthalene
Concen: 41.786 ng
RT: 13.228 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:162 Resp: 1377424

Ion	Ratio	Lower	Upper
162	100		
127	39.9	31.9	47.9
164	33.0	26.2	39.4





#48
2-Nitroaniline
Concen: 42.475 ng
RT: 13.434 min Scan# 1759
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion: 65 Resp: 392742

Ion	Ratio	Lower	Upper
65	100		
92	77.0	62.3	93.5
138	123.6	97.2	145.8





#49
Acenaphthylene
Concen: 41.783 ng
RT: 14.081 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:152 Resp: 2168474

Ion	Ratio	Lower	Upper
152	100		
151	20.1	15.9	23.9
153	12.9	10.6	15.8





#50
Dimethylphthalate
Concen: 41.478 ng
RT: 13.816 min Scan# 1824
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:163 Resp: 1809923

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.5	5.3
164	10.2	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 42.990 ng
RT: 13.940 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:165 Resp: 398345

Ion	Ratio	Lower	Upper
165	100		
63	45.1	38.0	57.0
89	50.5	41.7	62.5





#52
Acenaphthene
Concen: 40.853 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:154 Resp: 1344151

Ion	Ratio	Lower	Upper
154	100		
153	112.6	89.8	134.6
152	53.9	42.6	63.8





#53
3-Nitroaniline
Concen: 43.931 ng
RT: 14.275 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:138 Resp: 427828

Ion	Ratio	Lower	Upper
138	100		
108	10.1	8.3	12.5
92	110.2	89.8	134.8





#54
2,4-Dinitrophenol
Concen: 42.485 ng
RT: 14.481 min Scan# 1937
Delta R.T. -0.011 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:184 Resp: 231882

Ion	Ratio	Lower	Upper
184	100		
63	56.3	42.6	64.0
154	58.5	47.2	70.8





#55
Dibenzofuran
Concen: 40.679 ng
RT: 14.763 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:168 Resp: 2187786

Ion	Ratio	Lower	Upper
168	100		
139	38.6	30.3	45.5
169	13.3	10.6	16.0





#56
4-Nitrophenol
Concen: 44.968 ng
RT: 14.592 min Scan# 1956
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:139 Resp: 378214

Ion	Ratio	Lower	Upper
139	100		
109	66.9	46.1	86.1
65	86.3	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 43.527 ng
RT: 14.734 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:165 Resp: 550175

Ion	Ratio	Lower	Upper
165	100		
63	36.1	30.4	45.6
89	65.2	54.6	82.0
182	3.0	2.2	3.2





#58
Fluorene
Concen: 41.750 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:166 Resp: 1738215

Ion	Ratio	Lower	Upper
166	100		
165	97.0	78.1	117.1
167	13.2	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 41.995 ng
RT: 14.998 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:232 Resp: 470547

Ion	Ratio	Lower	Upper
232	100		
131	46.8	36.8	55.2
130	2.3	1.9	2.9
166	31.3	25.0	37.6





#60
Diethylphthalate
Concen: 41.570 ng
RT: 15.198 min Scan# 2059
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 1869274

Ion	Ratio	Lower	Upper
149	100		
177	21.5	17.4	26.2
150	12.8	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 41.777 ng
RT: 15.422 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:204 Resp: 844629

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.5	39.7
141	62.6	49.4	74.0





#62
4-Nitroaniline
Concen: 43.114 ng
RT: 15.451 min Scan# 2102
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:138 Resp: 444482

Ion	Ratio	Lower	Upper
138	100		
92	49.9	27.8	67.8
108	72.1	48.9	88.9





#63
Azobenzene
Concen: 41.686 ng
RT: 15.722 min Scan# 21
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion: 77 Resp: 1722900

Ion	Ratio	Lower	Upper
77	100		
182	25.1	7.3	47.3
105	22.1	2.3	42.3
51	24.1	3.4	43.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.163 min Scan# 2393
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:188 Resp: 1178916

Ion	Ratio	Lower	Upper
188	100		
94	9.8	8.6	13.0
80	11.0	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 42.724 ng
RT: 15.510 min Scan# 21
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:198 Resp: 317548

Ion	Ratio	Lower	Upper
198	100		
51	35.0	18.4	58.4
105	43.4	26.5	66.5





#66
n-Nitrosodiphenylamine
Concen: 41.500 ng
RT: 15.645 min Scan# 2135
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:169 Resp: 1460282

Ion	Ratio	Lower	Upper
169	100		
168	66.1	53.0	79.4
167	35.8	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 42.314 ng
RT: 16.334 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:248 Resp: 545579

Ion	Ratio	Lower	Upper
248	100		
250	99.3	78.2	117.4
141	76.1	57.7	86.5





#68
Hexachlorobenzene
Concen: 41.909 ng
RT: 16.451 min Scan# 2272
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:284 Resp: 642095

Ion	Ratio	Lower	Upper
284	100		
142	34.3	24.6	37.0
249	30.1	23.8	35.6





#69
Atrazine
Concen: 31.567 ng
RT: 16.622 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:200 Resp: 282908

Ion	Ratio	Lower	Upper
200	100		
173	25.4	7.5	47.5
215	48.7	28.8	68.8





#70
Pentachlorophenol
Concen: 42.900 ng
RT: 16.810 min Scan# 2333
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:266 Resp: 458529

Ion	Ratio	Lower	Upper
266	100		
268	65.3	51.0	76.6
264	63.0	49.8	74.6





#71
Phenanthrene
Concen: 40.615 ng
RT: 17.204 min Scan# 24
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:178 Resp: 2612052

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.4	23.0
179	15.5	12.5	18.7





#72
Anthracene
Concen: 42.051 ng
RT: 17.298 min Scan# 2416
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:178 Resp: 2606508

Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.7	22.1
179	15.4	12.4	18.6





#73
Carbazole
Concen: 42.351 ng
RT: 17.581 min Scan# 24
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:167 Resp: 2564955

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.1	25.7
139	13.1	10.6	16.0





#74
Di-n-butylphthalate
Concen: 44.539 ng
RT: 18.175 min Scan# 2565
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 3360635

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.4	11.2
104	4.9	3.9	5.9





#75
Fluoranthene
Concen: 41.771 ng
RT: 19.298 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:202 Resp: 3195045

Ion	Ratio	Lower	Upper
202	100		
101	12.6	0.0	33.2
203	17.5	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.616 min Scan# 3150
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:240 Resp: 1288125

Ion	Ratio	Lower	Upper
240	100		
120	12.7	9.8	14.8
236	25.5	20.9	31.3





#77
Benzidine
Concen: 36.041 ng
RT: 19.492 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:184 Resp: 588487

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.6	17.4
183	11.3	9.2	13.8





#78
Pyrene
Concen: 40.022 ng
RT: 19.675 min Scan# 2820
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:202 Resp: 3276290

Ion	Ratio	Lower	Upper
202	100		
200	21.0	17.2	25.8
203	17.6	14.4	21.6





#79
Terphenyl-d14
Concen: 81.649 ng
RT: 19.892 min Scan# 2
Delta R.T. -0.029 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:244 Resp: 5217888

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.6	8.4
122	10.9	9.4	14.0





#80
Butylbenzylphthalate
Concen: 43.107 ng
RT: 20.627 min Scan# 2982
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 1511473

Ion	Ratio	Lower	Upper
149	100		
91	68.1	56.4	84.6
206	21.2	16.3	24.5





#81
Benzo(a)anthracene
Concen: 41.207 ng
RT: 21.598 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:228 Resp: 3299250

Ion	Ratio	Lower	Upper
228	100		
226	26.9	21.8	32.8
229	19.6	15.9	23.9





#82
3,3'-Dichlorobenzidine
Concen: 43.139 ng
RT: 21.516 min Scan# 3133
Delta R.T. -0.023 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:252 Resp: 1212290

Ion	Ratio	Lower	Upper
252	100		
254	64.4	51.7	77.5
126	11.8	8.8	13.2





#83
Chrysene
Concen: 40.888 ng
RT: 21.669 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:228 Resp: 3136332

Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.2	36.2
229	19.8	15.6	23.4





#84
Bis(2-ethylhexyl)phthalate
Concen: 44.400 ng
RT: 21.539 min Scan# 3137
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:149 Resp: 2282222

Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.4	33.6
279	5.2	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 45.513 ng
RT: 22.827 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:149 Resp: 3803266

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.4	6.0	9.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.986 min Scan# 3723
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:264 Resp: 1476872

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.9	28.3
265	22.1	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 42.071 ng m
RT: 28.798 min Scan# 4
Delta R.T. -0.018 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:276 Resp: 4313548

Ion	Ratio	Lower	Upper
276	100		
138	16.8	14.5	21.7
277	20.9	20.2	30.2





#88
Benzo(b)fluoranthene
Concen: 41.288 ng
RT: 23.933 min Scan# 3544
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:252 Resp: 3654984

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	10.8	9.4	14.0





#89
Benzo(k)fluoranthene
Concen: 41.081 ng
RT: 23.998 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:252 Resp: 3512834

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	11.1	9.1	13.7





#90
Benzo(a)pyrene
Concen: 41.855 ng
RT: 24.833 min Scan# 3697
Delta R.T. 0.000 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:252 Resp: 3196130

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	11.9	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 41.902 ng
RT: 28.880 min Scan# 4
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Tgt Ion:278 Resp: 3566000

Ion	Ratio	Lower	Upper
278	100		
139	16.8	12.6	19.0
279	24.0	18.5	27.7





#92
Benzo(g,h,i)perylene
Concen: 41.455 ng
RT: 29.992 min Scan# 4574
Delta R.T. -0.006 min
Lab File: BP024283.D
Acq: 14 Apr 2025 18:35

Tgt Ion:276 Resp: 3590938

Ion	Ratio	Lower	Upper
276	100		
277	23.4	18.9	28.3
138	21.2	18.7	28.1



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2		1,4-Dioxane	0.512	0.517	-1.0	95	0.00
3		Pyridine	1.351	1.349	0.1	91	0.00
4		n-Nitrosodimethylamine	0.448	0.451	-0.7	93	0.00
5	S	2-Fluorophenol	1.210	1.229	-1.6	92	0.00
6		Aniline	1.480	1.531	-3.4	91	0.00
7	S	Phenol-d6	1.656	1.678	-1.3	91	0.00
8		2-Chlorophenol	1.350	1.362	-0.9	92	0.00
9		Benzaldehyde	0.819	0.861	-5.1	97	0.00
10	C	Phenol	1.661	1.674	-0.8	90	0.00
11		bis(2-Chloroethyl)ether	1.339	1.335	0.3	90	0.00
12		1,3-Dichlorobenzene	1.454	1.465	-0.8	93	0.00
13	C	1,4-Dichlorobenzene	1.475	1.494	-1.3	94	0.00
14		1,2-Dichlorobenzene	1.424	1.421	0.2	93	0.00
15		Benzyl Alcohol	1.071	1.115	-4.1	91	0.00
16		2,2'-oxybis(1-Chloropropane	1.447	1.453	-0.4	90	0.00
17		2-Methylphenol	1.075	1.103	-2.6	91	0.00
18		Hexachloroethane	0.533	0.540	-1.3	94	0.00
19	P	n-Nitroso-di-n-propylamine	1.025	1.037	-1.2	90	0.00
20		3+4-Methylphenols	1.491	1.522	-2.1	90	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22		Acetophenone	0.495	0.515	-4.0	91	0.00
23	S	Nitrobenzene-d5	0.351	0.366	-4.3	91	0.00
24		Nitrobenzene	0.347	0.360	-3.7	90	0.00
25		Isophorone	0.635	0.661	-4.1	90	0.00
26	C	2-Nitrophenol	0.170	0.184	-8.2	92	0.00
27		2,4-Dimethylphenol	0.213	0.222	-4.2	90	0.00
28		bis(2-Chloroethoxy)methane	0.429	0.444	-3.5	90	0.00
29	C	2,4-Dichlorophenol	0.291	0.309	-6.2	91	0.00
30		1,2,4-Trichlorobenzene	0.311	0.324	-4.2	92	0.00
31		Naphthalene	1.054	1.090	-3.4	92	0.00
32		Benzoic acid	0.248	0.254	-2.4	96	-0.04
33		4-Chloroaniline	0.371	0.398	-7.3	92	0.00
34	C	Hexachlorobutadiene	0.183	0.192	-4.9	93	0.00
35		Caprolactam	0.107	0.113	-5.6	92	-0.02
36	C	4-Chloro-3-methylphenol	0.339	0.353	-4.1	90	0.00
37		2-Methylnaphthalene	0.731	0.756	-3.4	92	0.00
38		1-Methylnaphthalene	0.715	0.736	-2.9	92	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40		1,2,4,5-Tetrachlorobenzene	0.550	0.581	-5.6	92	0.00
41	P	Hexachlorocyclopentadiene	0.195	0.212	-8.7	89	-0.01
42	S	2,4,6-Tribromophenol	0.277	0.296	-6.9	93	-0.02
43	C	2,4,6-Trichlorophenol	0.366	0.396	-8.2	93	0.00
44		2,4,5-Trichlorophenol	0.405	0.434	-7.2	93	0.00
45	S	2-Fluorobiphenyl	1.306	1.383	-5.9	92	0.00
46		1,1'-Biphenyl	1.470	1.529	-4.0	91	-0.01
47		2-Chloronaphthalene	1.099	1.148	-4.5	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.308	0.327	-6.2	90	-0.01
49		Acenaphthylene	1.730	1.807	-4.5	91	0.00
50		Dimethylphthalate	1.454	1.508	-3.7	92	-0.02
51		2,6-Dinitrotoluene	0.309	0.332	-7.4	93	0.00
52	C	Acenaphthene	1.097	1.120	-2.1	91	-0.01
53		3-Nitroaniline	0.325	0.357	-9.8	92	0.00
54	P	2,4-Dinitrophenol	0.182	0.193	-6.0	95	-0.01
55		Dibenzofuran	1.793	1.823	-1.7	91	-0.02
56	P	4-Nitrophenol	0.280	0.315	-12.5	94	-0.01
57		2,4-Dinitrotoluene	0.421	0.458	-8.8	93	-0.01
58		Fluorene	1.388	1.448	-4.3	95	-0.01
59		2,3,4,6-Tetrachlorophenol	0.373	0.392	-5.1	92	0.00
60		Diethylphthalate	1.499	1.558	-3.9	92	-0.02
61		4-Chlorophenyl-phenylether	0.674	0.704	-4.5	94	-0.02
62		4-Nitroaniline	0.344	0.370	-7.6	92	-0.02
63		Azobenzene	1.378	1.436	-4.2	92	-0.02
64	I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.02
65		4,6-Dinitro-2-methylphenol	0.126	0.135	-7.1	95	-0.02
66	c	n-Nitrosodiphenylamine	0.597	0.619	-3.7	91	-0.01
67		4-Bromophenyl-phenylether	0.219	0.231	-5.5	93	-0.01
68		Hexachlorobenzene	0.260	0.272	-4.6	93	-0.02
69		Atrazine	0.152	0.120	21.1	92	0.00
70	C	Pentachlorophenol	0.181	0.194	-7.2	92	-0.02
71		Phenanthrene	1.091	1.108	-1.6	91	-0.02
72		Anthracene	1.052	1.105	-5.0	92	-0.02
73		Carbazole	1.027	1.088	-5.9	93	-0.02
74		Di-n-butylphthalate	1.280	1.425	-11.3	94	-0.02
75	C	Fluoranthene	1.298	1.355	-4.4	93	-0.02
76	I	Chrysene-d12	1.000	1.000	0.0	95	0.00
77		Benzidine	0.254	0.228	10.2	50#	-0.02
78		Pyrene	1.271	1.272	-0.1	93	-0.01
79	S	Terphenyl-d14	0.992	1.013	-2.1	94	-0.03
80		Butylbenzylphthalate	0.544	0.587	-7.9	94	-0.02
81		Benzo(a)anthracene	1.243	1.281	-3.1	95	0.00
82		3,3'-Dichlorobenzidine	0.436	0.471	-8.0	92	-0.02
83		Chrysene	1.191	1.217	-2.2	96	-0.01
84		Bis(2-ethylhexyl)phthalate	0.798	0.886	-11.0	95	0.00
85	c	Di-n-octyl phthalate	1.297	1.476	-13.8	96	0.00
86	I	Perylene-d12	1.000	1.000	0.0	99	0.00
87		Indeno(1,2,3-cd)pyrene	1.388	1.460	-5.2	101	-0.02
88		Benzo(b)fluoranthene	1.199	1.237	-3.2	98	0.00
89		Benzo(k)fluoranthene	1.158	1.189	-2.7	98	0.00
90	C	Benzo(a)pyrene	1.034	1.082	-4.6	99	0.00
91		Dibenzo(a,h)anthracene	1.152	1.207	-4.8	100	0.00
92		Benzo(g,h,i)perylene	1.173	1.216	-3.7	99	0.00

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2		1,4-Dioxane	40.000	40.412	-1.0	95	0.00
3		Pyridine	40.000	39.955	0.1	91	0.00
4		n-Nitrosodimethylamine	40.000	40.205	-0.5	93	0.00
5	S	2-Fluorophenol	80.000	81.302	-1.6	92	0.00
6		Aniline	40.000	41.370	-3.4	91	0.00
7	S	Phenol-d6	80.000	81.063	-1.3	91	0.00
8		2-Chlorophenol	40.000	40.334	-0.8	92	0.00
9		Benzaldehyde	40.000	42.024	-5.1	97	0.00
10	C	Phenol	40.000	40.306	-0.8	90	0.00
11		bis(2-Chloroethyl)ether	40.000	39.902	0.2	90	0.00
12		1,3-Dichlorobenzene	40.000	40.298	-0.7	93	0.00
13	C	1,4-Dichlorobenzene	40.000	40.516	-1.3	94	0.00
14		1,2-Dichlorobenzene	40.000	39.898	0.3	93	0.00
15		Benzyl Alcohol	40.000	41.621	-4.1	91	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	40.173	-0.4	90	0.00
17		2-Methylphenol	40.000	41.046	-2.6	91	0.00
18		Hexachloroethane	40.000	40.522	-1.3	94	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	40.471	-1.2	90	0.00
20		3+4-Methylphenols	40.000	40.810	-2.0	90	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22		Acetophenone	40.000	41.652	-4.1	91	0.00
23	S	Nitrobenzene-d5	80.000	83.531	-4.4	91	0.00
24		Nitrobenzene	40.000	41.537	-3.8	90	0.00
25		Isophorone	40.000	41.648	-4.1	90	0.00
26	C	2-Nitrophenol	40.000	43.376	-8.4	92	0.00
27		2,4-Dimethylphenol	40.000	41.689	-4.2	90	0.00
28		bis(2-Chloroethoxy)methane	40.000	41.402	-3.5	90	0.00
29	C	2,4-Dichlorophenol	40.000	42.461	-6.2	91	0.00
30		1,2,4-Trichlorobenzene	40.000	41.612	-4.0	92	0.00
31		Naphthalene	40.000	41.386	-3.5	92	0.00
32		Benzoic acid	40.000	40.959	-2.4	96	-0.04
33		4-Chloroaniline	40.000	42.937	-7.3	92	0.00
34	C	Hexachlorobutadiene	40.000	42.007	-5.0	93	0.00
35		Caprolactam	40.000	42.152	-5.4	92	-0.02
36	C	4-Chloro-3-methylphenol	40.000	41.737	-4.3	90	0.00
37		2-Methylnaphthalene	40.000	41.403	-3.5	92	0.00
38		1-Methylnaphthalene	40.000	41.210	-3.0	92	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	91	0.00
40		1,2,4,5-Tetrachlorobenzene	40.000	42.275	-5.7	92	0.00
41	P	Hexachlorocyclopentadiene	40.000	43.639	-9.1	89	-0.01
42	S	2,4,6-Tribromophenol	80.000	85.488	-6.9	93	-0.02
43	C	2,4,6-Trichlorophenol	40.000	43.329	-8.3	93	0.00
44		2,4,5-Trichlorophenol	40.000	42.828	-7.1	93	0.00
45	S	2-Fluorobiphenyl	80.000	84.715	-5.9	92	0.00
46		1,1'-Biphenyl	40.000	41.595	-4.0	91	-0.01
47		2-Chloronaphthalene	40.000	41.786	-4.5	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	42.475	-6.2	90	-0.01
49		Acenaphthylene	40.000	41.783	-4.5	91	0.00
50		Dimethylphthalate	40.000	41.478	-3.7	92	-0.02
51		2,6-Dinitrotoluene	40.000	42.990	-7.5	93	0.00
52	C	Acenaphthene	40.000	40.853	-2.1	91	-0.01
53		3-Nitroaniline	40.000	43.931	-9.8	92	0.00
54	P	2,4-Dinitrophenol	40.000	42.485	-6.2	95	-0.01
55		Dibenzofuran	40.000	40.679	-1.7	91	-0.02
56	P	4-Nitrophenol	40.000	44.968	-12.4	94	-0.01
57		2,4-Dinitrotoluene	40.000	43.527	-8.8	93	-0.01
58		Fluorene	40.000	41.750	-4.4	95	-0.01
59		2,3,4,6-Tetrachlorophenol	40.000	41.995	-5.0	92	0.00
60		Diethylphthalate	40.000	41.570	-3.9	92	-0.02
61		4-Chlorophenyl-phenylether	40.000	41.777	-4.4	94	-0.02
62		4-Nitroaniline	40.000	43.114	-7.8	92	-0.02
63		Azobenzene	40.000	41.686	-4.2	92	-0.02
64	I	Phenanthrene-d10	20.000	20.000	0.0	92	-0.02
65		4,6-Dinitro-2-methylphenol	40.000	42.724	-6.8	95	-0.02
66	c	n-Nitrosodiphenylamine	40.000	41.500	-3.8	91	-0.01
67		4-Bromophenyl-phenylether	40.000	42.314	-5.8	93	-0.01
68		Hexachlorobenzene	40.000	41.909	-4.8	93	-0.02
69		Atrazine	40.000	31.567	21.1	92	0.00
70	C	Pentachlorophenol	40.000	42.900	-7.2	92	-0.02
71		Phenanthrene	40.000	40.615	-1.5	91	-0.02
72		Anthracene	40.000	42.051	-5.1	92	-0.02
73		Carbazole	40.000	42.351	-5.9	93	-0.02
74		Di-n-butylphthalate	40.000	44.539	-11.3	94	-0.02
75	C	Fluoranthene	40.000	41.771	-4.4	93	-0.02
76	I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77		Benzidine	40.000	36.041	9.9	50	-0.02
78		Pyrene	40.000	40.022	-0.1	93	-0.01
79	S	Terphenyl-d14	80.000	81.649	-2.1	94	-0.03
80		Butylbenzylphthalate	40.000	43.107	-7.8	94	-0.02
81		Benzo(a)anthracene	40.000	41.207	-3.0	95	0.00
82		3,3'-Dichlorobenzidine	40.000	43.139	-7.8	92	-0.02
83		Chrysene	40.000	40.888	-2.2	96	-0.01
84		Bis(2-ethylhexyl)phthalate	40.000	44.400	-11.0	95	0.00
85	c	Di-n-octyl phthalate	40.000	45.513	-13.8	96	0.00
86	I	Perylene-d12	20.000	20.000	0.0	99	0.00
87		Indeno(1,2,3-cd)pyrene	40.000	42.071	-5.2	101	-0.02
88		Benzo(b)fluoranthene	40.000	41.288	-3.2	98	0.00
89		Benzo(k)fluoranthene	40.000	41.081	-2.7	98	0.00
90	C	Benzo(a)pyrene	40.000	41.855	-4.6	99	0.00
91		Dibenzo(a,h)anthracene	40.000	41.902	-4.8	100	0.00
92		Benzo(g,h,i)perylene	40.000	41.455	-3.6	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024283.D
Acq On : 14 Apr 2025 18:35
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041425

Quant Time: Apr 15 04:54:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 15 04:48:42 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Response Factor Report MSBNA_P

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP051325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue May 13 16:21:39 2025
Response Via : Initial Calibration

Calibration Files
2.5 =BP024614.D 5 =BP024615.D 10 =BP024616.D 20 =BP024617.D 40 =BP024618.D 50 =BP024619.D 60 =BP024620.D 80 =BP0246
21.D

		Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1)	I	1,4-Dichlorobenzen...	----------ISTD----------									
2)		1,4-Dioxane		0.641	0.557	0.535	0.505	0.562	0.530	0.513	0.549	8.31
3)		Pyridine		1.054	1.135	1.202	1.198	1.363	1.307	1.263	1.218	8.57
4)		n-Nitrosodimet...		0.536	0.507	0.527	0.517	0.580	0.557	0.537	0.538	4.61
5)	S	2-Fluorophenol		1.134	1.096	1.162	1.105	1.272	1.223	1.193	1.169	5.50
6)		Aniline		1.717	1.703	1.932	1.917	2.130	2.091	1.999	1.927	8.67
7)	S	Phenol-d6		1.301	1.344	1.486	1.472	1.671	1.620	1.552	1.492	9.09
8)		2-Chlorophenol		1.200	1.204	1.311	1.303	1.465	1.433	1.376	1.327	7.83
9)		Benzaldehyde		0.991	1.008	1.046	0.955	1.043	0.954	0.766	0.966	9.91
10)	C	Phenol		1.373	1.401	1.526	1.509	1.707	1.670	1.597	1.540	8.24
11)		bis(2-Chloroet...		1.317	1.170	1.257	1.213	1.357	1.299	1.247	1.266	5.04
12)		1,3-Dichlorobe...		1.596	1.487	1.520	1.440	1.622	1.546	1.480	1.527	4.29
13)	C	1,4-Dichlorobe...		1.558	1.505	1.525	1.470	1.637	1.567	1.491	1.536	3.67
14)		1,2-Dichlorobe...		1.530	1.467	1.498	1.426	1.579	1.529	1.449	1.497	3.56
15)		Benzyl Alcohol		0.883	0.950	1.128	1.149	1.306	1.288	1.236	1.134	14.43
16)		2,2'-oxybis(1-...		1.543	1.445	1.520	1.467	1.617	1.564	1.454	1.516	4.22
17)		2-Methylphenol		0.966	0.949	1.105	1.099	1.246	1.213	1.154	1.105	10.30
18)		Hexachloroethane		0.606	0.591	0.582	0.551	0.623	0.597	0.570	0.589	4.04
19)	P	n-Nitroso-di-n...	0.880	0.977	0.967	1.086	1.048	1.142	1.132	1.026	1.032	8.63
20)		3+4-Methylphenols		1.221	1.302	1.503	1.515	1.702	1.692	1.584	1.503	12.21
21)	I	Naphthalene-d8	----------ISTD----------									
22)		Acetophenone		0.486	0.481	0.511	0.483	0.534	0.512	0.491	0.500	3.97
23)	S	Nitrobenzene-d5		0.381	0.376	0.401	0.382	0.429	0.408	0.394	0.396	4.67
24)		Nitrobenzene		0.333	0.340	0.355	0.337	0.373	0.356	0.344	0.348	4.03
25)		Isophorone		0.644	0.640	0.695	0.673	0.753	0.721	0.682	0.687	5.91
26)	C	2-Nitrophenol		0.135	0.143	0.168	0.172	0.196	0.192	0.189	0.171	14.05
27)		2,4-Dimethylph...		0.273	0.279	0.300	0.298	0.329	0.318	0.309	0.301	6.60
28)		bis(2-Chloroet...		0.403	0.391	0.411	0.399	0.443	0.419	0.405	0.410	4.15
29)	C	2,4-Dichloroph...		0.250	0.258	0.290	0.291	0.327	0.316	0.309	0.292	9.96
30)		1,2,4-Trichlor...		0.346	0.322	0.327	0.306	0.347	0.329	0.323	0.329	4.33
31)		Naphthalene		1.064	1.026	1.049	0.980	1.091	1.037	1.007	1.036	3.54
32)		Benzoic acid			0.089	0.146	0.191	0.215	0.220	0.225	0.181	29.55
33)		4-Chloroaniline		0.348	0.371	0.420	0.418	0.465	0.462	0.442	0.418	10.62
34)	C	Hexachlorobuta...		0.231	0.214	0.211	0.198	0.222	0.209	0.206	0.213	5.08
35)		Caprolactam		0.082	0.092	0.104	0.109	0.121	0.121	0.110	0.106	13.80
36)	C	4-Chloro-3-met...		0.316	0.325	0.358	0.356	0.391	0.384	0.359	0.356	7.78
37)		2-Methylnaphth...		0.673	0.654	0.670	0.640	0.713	0.693	0.654	0.671	3.73
38)		1-Methylnaphth...		0.741	0.700	0.720	0.690	0.751	0.733	0.690	0.718	3.47

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP051325.M

39)	I	Acenaphthene-d10	----------ISTD----------								
40)		1,2,4,5-Tetrac...	0.593	0.558	0.574	0.549	0.625	0.580	0.586	0.581	4.26
41)	P	Hexachlorocycl...	0.250	0.276	0.350	0.356	0.420	0.391	0.421	0.352	19.02
42)	S	2,4,6-Tribromo...	0.324	0.321	0.335	0.323	0.369	0.357	0.346	0.339	5.49
43)	C	2,4,6-Trichlor...	0.305	0.341	0.375	0.371	0.428	0.410	0.402	0.376	11.29
44)		2,4,5-Trichlor...	0.376	0.375	0.418	0.404	0.471	0.447	0.440	0.419	8.68
45)	S	2-Fluorobiphenyl	1.601	1.498	1.553	1.432	1.602	1.515	1.485	1.527	4.10
46)		1,1'-Biphenyl	1.524	1.434	1.492	1.383	1.586	1.473	1.456	1.478	4.40
47)		2-Chloronaphth...	1.147	1.106	1.131	1.065	1.196	1.132	1.113	1.127	3.56
48)		2-Nitroaniline	0.279	0.304	0.331	0.335	0.381	0.364	0.343	0.334	10.28
49)		Acenaphthylene	1.845	1.818	1.920	1.804	2.029	1.930	1.856	1.886	4.19
50)		Dimethylphthalate	1.570	1.527	1.518	1.450	1.611	1.535	1.460	1.524	3.73
51)		2,6-Dinitrotol...	0.299	0.308	0.326	0.312	0.354	0.337	0.319	0.322	5.75
52)	C	Acenaphthene	1.105	1.054	1.094	1.028	1.159	1.099	1.052	1.084	4.02
53)		3-Nitroaniline	0.258	0.275	0.329	0.340	0.379	0.370	0.350	0.329	14.04
54)	P	2,4-Dinitrophenol		0.098	0.143	0.172	0.201	0.201	0.201	0.169	24.79
55)		Dibenzofuran	1.810	1.716	1.740	1.633	1.816	1.733	1.658	1.730	4.00
56)	P	4-Nitrophenol		0.159	0.222	0.257	0.299	0.288	0.274	0.250	20.82
57)		2,4-Dinitrotol...	0.400	0.422	0.456	0.454	0.508	0.483	0.454	0.454	7.87
58)		Fluorene	1.466	1.415	1.410	1.336	1.494	1.410	1.341	1.410	4.14
59)		2,3,4,6-Tetrac...	0.367	0.364	0.381	0.379	0.428	0.406	0.394	0.388	5.89
60)		Diethylphthalate	1.593	1.542	1.537	1.444	1.639	1.563	1.468	1.541	4.40
61)		4-Chlorophenyl...	0.752	0.697	0.700	0.656	0.741	0.709	0.668	0.703	5.01
62)		4-Nitroaniline	0.228	0.259	0.326	0.330	0.375	0.363	0.341	0.317	17.06
63)		Azobenzene	1.327	1.299	1.356	1.250	1.407	1.343	1.243	1.318	4.45
64)	I	Phenanthrene-d10	----------ISTD----------								
65)		4,6-Dinitro-2-...		0.094	0.120	0.129	0.149	0.143	0.141	0.129	15.64
66)	c	n-Nitrosodiphe...	0.603	0.589	0.627	0.593	0.667	0.621	0.600	0.614	4.41
67)		4-Bromophenyl-...	0.231	0.223	0.237	0.226	0.252	0.242	0.240	0.236	4.31
68)		Hexachlorobenzene	0.303	0.288	0.298	0.282	0.321	0.304	0.300	0.299	4.25
69)		Atrazine	0.210	0.209	0.227	0.218	0.248	0.237	0.226	0.225	6.32
70)	C	Pentachlorophenol	0.124	0.137	0.163	0.172	0.198	0.191	0.192	0.168	17.12
71)		Phenanthrene	1.144	1.085	1.125	1.051	1.170	1.125	1.083	1.112	3.68
72)		Anthracene	1.097	1.064	1.142	1.075	1.205	1.139	1.108	1.119	4.31
73)		Carbazole	0.923	0.944	1.039	1.001	1.114	1.053	1.013	1.012	6.43
74)		Di-n-butylphth...	1.182	1.218	1.332	1.260	1.447	1.363	1.313	1.302	6.97
75)	C	Fluoranthene	1.271	1.254	1.314	1.250	1.399	1.337	1.272	1.300	4.18
76)	I	Chrysene-d12	----------ISTD----------								
77)		Benzidine		0.362	0.543	0.560	0.657	0.599	0.548	0.545	18.22
78)		Pyrene	1.184	1.137	1.197	1.153	1.273	1.222	1.223	1.198	3.85
79)	S	Terphenyl-d14	1.192	1.113	1.165	1.125	1.226	1.175	1.167	1.166	3.30
80)		Butylbenzylpht...	0.458	0.465	0.537	0.527	0.604	0.571	0.569	0.533	10.28
81)		Benzo(a)anthra...	1.251	1.218	1.253	1.208	1.336	1.275	1.250	1.256	3.36
82)		3,3'-Dichlorob...	0.357	0.392	0.474	0.477	0.539	0.511	0.514	0.466	14.45
83)		Chrysene	1.214	1.180	1.185	1.131	1.255	1.188	1.180	1.190	3.16
84)		Bis(2-ethylhex...	0.673	0.683	0.801	0.766	0.884	0.836	0.841	0.783	10.32
85)	c	Di-n-octyl pht...	1.048	1.116	1.272	1.280	1.458	1.376	1.417	1.281	11.95

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP051325.M

86)	I	Perylene-d12	----------ISTD----------							
87)		Indeno(1,2,3-c...	1.341	1.357	1.447	1.424	1.608	1.549	1.494	1.460	6.70
88)		Benzo(b)fluora...	1.028	1.070	1.152	1.103	1.288	1.184	1.189	1.145	7.58
89)		Benzo(k)fluora...	1.147	1.106	1.182	1.140	1.258	1.238	1.185	1.180	4.59
90)	C	Benzo(a)pyrene	0.995	1.001	1.112	1.064	1.221	1.167	1.136	1.099	7.68
91)		Dibenzo(a,h)an...	1.068	1.066	1.186	1.167	1.321	1.271	1.236	1.188	8.18
92)		Benzo(g,h,i)pe...	1.101	1.106	1.167	1.147	1.294	1.242	1.212	1.181	6.07

(#) = Out of Range

Quantitation Report (QT Reviewed)

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024614.D
Acq On : 13 May 2025 10:41
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: May 13 16:01:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	105461	20.00	ng	0.00
21) Naphthalene-d8	10.428	136	434352	20.00	ng	0.00
39) Acenaphthene-d10	14.287	164	287970	20.00	ng	-0.02
64) Phenanthrene-d10	17.092	188	596044	20.00	ng	0.00
76) Chrysene-d12	21.516	240	663028	20.00	ng	0.00
86) Perylene-d12	24.780	264	769627	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.00	ng	
7) Phenol-d6	0.000	99	0d	0.00	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.00	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.00	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.00	ng	
79) Terphenyl-d14	0.000	244	0d	0.00	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	8.458	70	11601	2.13	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024614.D
Acq On : 13 May 2025 10:41
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: May 13 16:01:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.658 min Scan# 7
Delta R.T. -0.005 min
Lab File: BP024614.D
Acq: 13 May 2025 10:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:152 Resp: 105461

Ion	Ratio	Lower	Upper
152	100		
150	155.0	125.9	188.9
115	65.0	50.1	75.1





#5
2-Fluorophenol
Concen: 0.00 ng
Expected RT: 5.29 min

Lab File: BP024614.D
Acq: 13 May 2025 10:41

Tgt Ion: 112

Sig	Exp Ratio
112	100
64	55.1
63	29.4





#7
Phenol-d6
Concen: 0.00 ng
Expected RT: 6.85 min

Lab File: BP024614.D
Acq: 13 May 2025 10:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 99

Sig	Exp Ratio
99	100
42	18.0
71	36.5





#19
n-Nitroso-di-n-propylamine
Concen: 2.13 ng
RT: 8.458 min Scan# 913
Delta R.T. 0.000 min
Lab File: BP024614.D
Acq: 13 May 2025 10:41

Tgt Ion: 70 Resp: 11601

Ion	Ratio	Lower	Upper
70	100		
42	56.0	43.0	64.6
101	10.0	8.7	13.1
130	27.5	19.4	29.0









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.428 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024614.D
Acq: 13 May 2025 10:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:136 Resp: 434352

Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.8	13.2
54	6.8	5.5	8.3
68	4.7	4.2	6.2





#23
Nitrobenzene-d5
Concen: 0.00 ng
Expected RT: 8.81 min

Lab File: BP024614.D
Acq: 13 May 2025 10:41

Tgt Ion: 82

Sig	Exp Ratio
82	100
128	46.2
54	48.4





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.287 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024614.D
Acq: 13 May 2025 10:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:164 Resp: 287970

Ion	Ratio	Lower	Upper
164	100		
162	102.1	81.8	122.6
160	45.1	36.3	54.5





#42
2,4,6-Tribromophenol
Concen: 0.00 ng
Expected RT: 15.81 min

Lab File: BP024614.D
Acq: 13 May 2025 10:41

Tgt Ion: 330

Sig	Exp Ratio
330	100
332	97.8
141	30.4





#45
2-Fluorobiphenyl
Concen: 0.00 ng
Expected RT: 12.90 min

Lab File: BP024614.D
Acq: 13 May 2025 10:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion: 172

Sig	Exp Ratio
172	100
171	36.0
170	23.5





#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.092 min Scan# 2381
Delta R.T. -0.006 min
Lab File: BP024614.D
Acq: 13 May 2025 10:41

Tgt Ion:188 Resp: 596044

Ion	Ratio	Lower	Upper
188	100		
94	7.7	6.5	9.7
80	8.9	7.3	10.9









#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.516 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024614.D
Acq: 13 May 2025 10:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:240 Resp: 663028

Ion	Ratio	Lower	Upper
240	100		
120	8.9	7.8	11.6
236	26.1	20.6	31.0





#79
Terphenyl-d14
Concen: 0.00 ng
Expected RT: 19.81 min

Lab File: BP024614.D
Acq: 13 May 2025 10:41

Tgt Ion: 244

Sig	Exp Ratio
244	100
212	6.9
122	9.2





#86
Perylene-d12
Concen: 20.00 ng
RT: 24.780 min Scan# 3
Delta R.T. -0.023 min
Lab File: BP024614.D
Acq: 13 May 2025 10:41

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Tgt Ion:264 Resp: 769627

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.7	28.1
265	21.8	17.0	25.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024615.D
Acq On : 13 May 2025 11:22
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:02:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.664	152	90088	20.00	ng		0.00
21) Naphthalene-d8	10.434	136	378055	20.00	ng		0.00
39) Acenaphthene-d10	14.293	164	254561	20.00	ng		-0.01
64) Phenanthrene-d10	17.092	188	531548	20.00	ng		0.00
76) Chrysene-d12	21.522	240	614288	20.00	ng		0.00
86) Perylene-d12	24.810	264	724260	20.00	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.293	112	51093	9.70	ng		0.00
7) Phenol-d6	6.858	99	58619	8.72	ng		0.00
23) Nitrobenzene-d5	8.811	82	72056	9.63	ng		0.00
42) 2,4,6-Tribromophenol	15.810	330	41219	9.55	ng		0.00
45) 2-Fluorobiphenyl	12.904	172	203719	10.48	ng		0.00
79) Terphenyl-d14	19.822	244	366245	10.22	ng		0.01
Target Compounds							Qvalue
2) 1,4-Dioxane	3.228	88	14435	5.84	ng		95
3) Pyridine	3.646	79	23737	4.33	ng	#	91
4) n-Nitrosodimethylamine	3.546	42	12080	4.99	ng	#	96
6) Aniline	7.011	93	38664	4.45	ng		97
8) 2-Chlorophenol	7.240	128	27022	4.52	ng		93
9) Benzaldehyde	6.828	77	22328m	5.14	ng		
10) Phenol	6.887	94	30912	4.46	ng		98
11) bis(2-Chloroethyl)ether	7.099	93	29668	5.20	ng		97
12) 1,3-Dichlorobenzene	7.552	146	35946	5.23	ng		96
13) 1,4-Dichlorobenzene	7.699	146	35091	5.07	ng		95
14) 1,2-Dichlorobenzene	8.011	146	34458	5.11	ng		99
15) Benzyl Alcohol	7.917	79	19888	3.89	ng		97
16) 2,2'-oxybis(1-Chloropr...	8.193	45	34754	5.09	ng		95
17) 2-Methylphenol	8.116	107	21754	4.37	ng		97
18) Hexachloroethane	8.728	117	13650	5.15	ng		96
19) n-Nitroso-di-n-propyla...	8.464	70	22015	4.73	ng		98
20) 3+4-Methylphenols	8.452	107	27500	4.06	ng		97
22) Acetophenone	8.487	105	45968	4.87	ng	#	99
24) Nitrobenzene	8.852	77	31500	4.78	ng		97
25) Isophorone	9.375	82	60855	4.69	ng		98
26) 2-Nitrophenol	9.569	139	12786	3.96	ng		95
27) 2,4-Dimethylphenol	9.628	122	25830	4.54	ng		92
28) bis(2-Chloroethoxy)met...	9.858	93	38071	4.91	ng		97
29) 2,4-Dichlorophenol	10.122	162	23623	4.29	ng		97
30) 1,2,4-Trichlorobenzene	10.299	180	32699	5.26	ng		99
31) Naphthalene	10.481	128	100564	5.13	ng		100
33) 4-Chloroaniline	10.616	127	32908	4.16	ng		99
34) Hexachlorobutadiene	10.763	225	21834	5.43	ng		95
35) Caprolactam	11.399	113	7742m	3.88	ng		
36) 4-Chloro-3-methylphenol	11.746	107	29849	4.44	ng		95
37) 2-Methylnaphthalene	12.099	142	63562	5.01	ng		99
38) 1-Methylnaphthalene	12.316	142	70043	5.16	ng		100
40) 1,2,4,5-Tetrachloroben...	12.469	216	37738	5.11	ng		99
41) Hexachlorocyclopentadiene	12.446	237	15939	3.56	ng		93
43) 2,4,6-Trichlorophenol	12.722	196	19429	4.06	ng		92
44) 2,4,5-Trichlorophenol	12.816	196	23918	4.49	ng		93

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024615.D
Acq On : 13 May 2025 11:22
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:02:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
46) 1,1'-Biphenyl	13.116	154	96956	5.15	ng		98
47) 2-Chloronaphthalene	13.157	162	72977	5.09	ng		99
48) 2-Nitroaniline	13.381	65	17778	4.19	ng		98
49) Acenaphthylene	14.010	152	117448	4.89	ng		99
50) Dimethylphthalate	13.751	163	99940	5.15	ng		100
51) 2,6-Dinitrotoluene	13.869	165	19052	4.65	ng	#	85
52) Acenaphthene	14.357	154	70302	5.09	ng		96
53) 3-Nitroaniline	14.222	138	16400	3.92	ng		97
55) Dibenzofuran	14.698	168	115180	5.23	ng		98
57) 2,4-Dinitrotoluene	14.675	165	25481	4.41	ng		92
58) Fluorene	15.363	166	93303	5.20	ng		98
59) 2,3,4,6-Tetrachlorophenol	14.946	232	23382	4.73	ng		98
60) Diethylphthalate	15.134	149	101366	5.17	ng		99
61) 4-Chlorophenyl-phenyle...	15.357	204	47886	5.35	ng		98
62) 4-Nitroaniline	15.410	138	14482m	3.55	ng		
63) Azobenzene	15.651	77	84462	5.03	ng		99
66) n-Nitrosodiphenylamine	15.575	169	80078	4.90	ng		99
67) 4-Bromophenyl-phenylether	16.275	248	30702	4.90	ng		97
68) Hexachlorobenzene	16.387	284	40234	5.06	ng		99
69) Atrazine	16.551	200	27909	4.66	ng		98
70) Pentachlorophenol	16.757	266	16456	3.68	ng		95
71) Phenanthrene	17.134	178	152030	5.14	ng		99
72) Anthracene	17.234	178	145809	4.90	ng		99
73) Carbazole	17.522	167	122666	4.56	ng		98
74) Di-n-butylphthalate	18.104	149	157014	4.54	ng		99
75) Fluoranthene	19.228	202	168865	4.89	ng		99
78) Pyrene	19.604	202	181817	4.94	ng		99
80) Butylbenzylphthalate	20.551	149	70341	4.30	ng		100
81) Benzo(a)anthracene	21.498	228	192110	4.98	ng		99
82) 3,3'-Dichlorobenzidine	21.445	252	54793	3.82	ng		99
83) Chrysene	21.575	228	186449	5.10	ng		99
84) Bis(2-ethylhexyl)phtha...	21.445	149	103286	4.29	ng		97
85) Di-n-octyl phthalate	22.692	149	160883	4.09	ng		99
87) Indeno(1,2,3-cd)pyrene	28.504	276	242814	4.59	ng	#	87
88) Benzo(b)fluoranthene	23.763	252	186058	4.49	ng		97
89) Benzo(k)fluoranthene	23.839	252	207645	4.86	ng		100
90) Benzo(a)pyrene	24.657	252	180157	4.52	ng		99
91) Dibenzo(a,h)anthracene	28.586	278	193373	4.50	ng		98
92) Benzo(g,h,i)perylene	29.651	276	199361	4.66	ng		98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024615.D
Acq On : 13 May 2025 11:22
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: May 13 16:02:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.664 min Scan# 7
Delta R.T. 0.001 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:152 Resp: 9008

Ion	Ratio	Lower	Upper
152	100		
150	156.0	125.9	188.9
115	62.5	50.1	75.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#2
1,4-Dioxane
Concen: 5.84 ng
RT: 3.228 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 88 Resp: 14435

Ion	Ratio	Lower	Upper
88	100		
58	63.6	54.0	81.0
43	24.2	21.4	32.0





#3
Pyridine
Concen: 4.33 ng
RT: 3.646 min Scan# 9
Delta R.T. 0.024 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 79 Resp: 2373

Ion	Ratio	Lower	Upper
79	100		
52	63.2	50.2	75.2
51	47.4	26.8	40.2#



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#4
n-Nitrosodimethylamine
Concen: 4.99 ng
RT: 3.546 min Scan# 78
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 42 Resp: 12080

Ion	Ratio	Lower	Upper
42	100		
74	129.2	106.2	159.2
44	16.9	5.4	8.2#





#5
2-Fluorophenol
Concen: 9.70 ng
RT: 5.293 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:112 Resp: 5109

Ion	Ratio	Lower	Upper
112	100		
64	54.6	44.1	66.1
63	29.1	23.5	35.3

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#6
Aniline
Concen: 4.45 ng
RT: 7.011 min Scan# 667
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 93 Resp: 38664

Ion	Ratio	Lower	Upper
93	100		
66	36.2	31.0	46.6
65	20.2	15.5	23.3





#7
Phenol-d6
Concen: 8.72 ng
RT: 6.858 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 99 Resp: 5861

Ion	Ratio	Lower	Upper
99	100		
42	18.9	14.4	21.6
71	35.5	29.2	43.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#8
2-Chlorophenol
Concen: 4.52 ng
RT: 7.240 min Scan# 706
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:128 Resp: 27022

Ion	Ratio	Lower	Upper
128	100		
130	30.3	13.0	53.0
64	47.6	22.3	62.3





#9
Benzaldehyde
Concen: 5.14 ng m
RT: 6.828 min Scan# 6
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 77 Resp: 2232

Ion	Ratio	Lower	Upper
77	100		
105	105.0	83.3	123.3
106	86.1	78.0	118.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#10
Phenol
Concen: 4.46 ng
RT: 6.887 min Scan# 646
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 94 Resp: 30912

Ion	Ratio	Lower	Upper
94	100		
65	28.2	8.7	48.7
66	41.5	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 5.20 ng
RT: 7.099 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 93 Resp: 2966

Ion	Ratio	Lower	Upper
93	100		
63	68.3	50.6	90.6
95	31.8	12.9	52.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#12
1,3-Dichlorobenzene
Concen: 5.23 ng
RT: 7.552 min Scan# 759
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:146 Resp: 35946

Ion	Ratio	Lower	Upper
146	100		
148	60.0	51.4	77.2
75	31.0	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 5.07 ng
RT: 7.699 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:146 Resp: 3509

Ion	Ratio	Lower	Upper
146	100		
148	67.6	51.0	76.6
111	44.5	33.0	49.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#14
1,2-Dichlorobenzene
Concen: 5.11 ng
RT: 8.011 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:146 Resp: 34458

Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.0	76.4
111	43.7	35.2	52.8





#15
Benzyl Alcohol
Concen: 3.89 ng
RT: 7.917 min Scan# 8
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 79 Resp: 1988

Ion	Ratio	Lower	Upper
79	100		
108	78.3	60.8	91.2
77	68.2	51.9	77.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.09 ng
RT: 8.193 min Scan# 868
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 45 Resp: 34754

Ion	Ratio	Lower	Upper
45	100		
77	20.2	0.0	37.7
79	15.9	0.0	34.0





#17
2-Methylphenol
Concen: 4.37 ng
RT: 8.116 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:107 Resp: 2175

Ion	Ratio	Lower	Upper
107	100		
108	110.2	88.9	133.3
77	53.0	39.0	58.6
79	45.8	38.4	57.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#18
Hexachloroethane
Concen: 5.15 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:117 Resp: 13650

Ion	Ratio	Lower	Upper
117	100		
119	96.6	79.4	119.0
201	98.3	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 4.73 ng
RT: 8.464 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 70 Resp: 2201

Ion	Ratio	Lower	Upper
70	100		
42	52.6	43.0	64.6
101	11.9	8.7	13.1
130	23.5	19.4	29.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#20
3+4-Methylphenols
Concen: 4.06 ng
RT: 8.452 min Scan# 912
Delta R.T. 0.018 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:107 Resp: 27500

Ion	Ratio	Lower	Upper
107	100		
108	85.4	66.7	106.7
77	37.3	13.1	53.1
79	29.9	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:136 Resp: 37805

Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.8	13.2
54	6.8	5.5	8.3
68	4.7	4.2	6.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#22
Acetophenone
Concen: 4.87 ng
RT: 8.487 min Scan# 918
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:105 Resp: 45968

Ion	Ratio	Lower	Upper
105	100		
71	1.4	2.0	3.0#
51	26.0	20.4	30.6
120	22.4	18.1	27.1





#23
Nitrobenzene-d5
Concen: 9.63 ng
RT: 8.811 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 82 Resp: 7205

Ion	Ratio	Lower	Upper
82	100		
128	43.6	37.0	55.4
54	49.2	38.7	58.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#24
Nitrobenzene
Concen: 4.78 ng
RT: 8.852 min Scan# 980
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 77 Resp: 31500

Ion	Ratio	Lower	Upper
77	100		
123	44.5	37.4	56.0
65	12.7	11.3	16.9





#25
Isophorone
Concen: 4.69 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 82 Resp: 6085

Ion	Ratio	Lower	Upper
82	100		
95	7.8	6.3	9.5
138	17.9	15.3	22.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#26
2-Nitrophenol
Concen: 3.96 ng
RT: 9.569 min Scan# 1102
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:139 Resp: 12786

Ion	Ratio	Lower	Upper
139	100		
109	32.4	25.9	38.9
65	53.2	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 4.54 ng
RT: 9.628 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:122 Resp: 2583

Ion	Ratio	Lower	Upper
122	100		
107	131.6	96.2	144.2
121	61.8	47.6	71.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#28
bis(2-Chloroethoxy)methane
Concen: 4.91 ng
RT: 9.858 min Scan# 1151
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 93 Resp: 38071

Ion	Ratio	Lower	Upper
93	100		
95	33.7	25.5	38.3
123	12.5	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 4.29 ng
RT: 10.122 min Scan# 11
Delta R.T. 0.030 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:162 Resp: 2362

Ion	Ratio	Lower	Upper
162	100		
164	61.3	42.3	82.3
98	34.7	18.4	58.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#30
1,2,4-Trichlorobenzene
Concen: 5.26 ng
RT: 10.299 min Scan# 1226
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:180 Resp: 32699

Ion	Ratio	Lower	Upper
180	100		
182	97.6	79.0	118.4
145	32.5	26.6	39.8





#31
Naphthalene
Concen: 5.13 ng
RT: 10.481 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:128 Resp: 10056

Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.8	13.2
127	13.5	10.7	16.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#33
4-Chloroaniline
Concen: 4.16 ng
RT: 10.616 min Scan# 1280
Delta R.T. 0.018 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:127 Resp: 32908

Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.8	38.6
65	31.0	23.9	35.9
92	19.1	16.1	24.1





#34
Hexachlorobutadiene
Concen: 5.43 ng
RT: 10.763 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:225 Resp: 2183

Ion	Ratio	Lower	Upper
225	100		
223	60.5	51.0	76.6
227	71.7	53.1	79.7



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#35
Caprolactam
Concen: 3.88 ng m
RT: 11.399 min Scan# 1413
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:113 Resp: 7742

Ion	Ratio	Lower	Upper
113	100		
55	118.0	119.0	159.0#
56	82.4	90.3	130.3#





#36
4-Chloro-3-methylphenol
Concen: 4.44 ng
RT: 11.746 min Scan# 14
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:107 Resp: 2984

Ion	Ratio	Lower	Upper
107	100		
144	25.0	20.1	30.1
142	83.9	62.1	93.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#37
2-Methylnaphthalene
Concen: 5.01 ng
RT: 12.099 min Scan# 1532
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:142 Resp: 63562

Ion	Ratio	Lower	Upper
142	100		
141	90.2	71.2	106.8
115	39.4	30.6	46.0





#38
1-Methylnaphthalene
Concen: 5.16 ng
RT: 12.316 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:142 Resp: 7004

Ion	Ratio	Lower	Upper
142	100		
141	93.7	74.7	112.1
116	4.0	3.4	5.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.293 min Scan# 1905
Delta R.T. -0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:164 Resp: 254561

Ion	Ratio	Lower	Upper
164	100		
162	103.0	81.8	122.6
160	45.5	36.3	54.5





#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.11 ng
RT: 12.469 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:216 Resp: 3773

Ion	Ratio	Lower	Upper
216	100		
214	78.4	62.3	93.5
179	22.0	17.7	26.5
108	19.0	16.0	24.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#41
Hexachlorocyclopentadiene
Concen: 3.56 ng
RT: 12.446 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:237 Resp: 15939

Ion	Ratio	Lower	Upper
237	100		
235	66.6	39.9	79.9
272	13.5	0.0	33.2





#42
2,4,6-Tribromophenol
Concen: 9.55 ng
RT: 15.810 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:330 Resp: 4121

Ion	Ratio	Lower	Upper
330	100		
332	95.4	78.2	117.4
141	29.5	24.3	36.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#43
2,4,6-Trichlorophenol
Concen: 4.06 ng
RT: 12.722 min Scan# 1638
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:196 Resp: 19429

Ion	Ratio	Lower	Upper
196	100		
198	104.9	76.9	115.3
200	34.8	25.2	37.8





#44
2,4,5-Trichlorophenol
Concen: 4.49 ng
RT: 12.816 min Scan# 1
Delta R.T. 0.018 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:196 Resp: 2391

Ion	Ratio	Lower	Upper
196	100		
198	90.8	79.2	118.8
97	47.8	42.3	63.5
132	29.0	23.3	34.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#45
2-Fluorobiphenyl
Concen: 10.48 ng
RT: 12.904 min Scan# 1669
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:172 Resp: 203719

Ion	Ratio	Lower	Upper
172	100		
171	36.1	28.8	43.2
170	23.9	18.8	28.2





#46
1,1'-Biphenyl
Concen: 5.15 ng
RT: 13.116 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:154 Resp: 9695

Ion	Ratio	Lower	Upper
154	100		
153	39.8	21.0	61.0
76	11.2	0.0	31.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#47
2-Chloronaphthalene
Concen: 5.09 ng
RT: 13.157 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:162 Resp: 72977

Ion	Ratio	Lower	Upper
162	100		
127	41.6	33.8	50.6
164	33.0	26.0	39.0





#48
2-Nitroaniline
Concen: 4.19 ng
RT: 13.381 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 65 Resp: 1777

Ion	Ratio	Lower	Upper
65	100		
92	72.3	58.7	88.1
138	110.8	90.2	135.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#49
Acenaphthylene
Concen: 4.89 ng
RT: 14.010 min Scan# 1857
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:152 Resp: 117448

Ion	Ratio	Lower	Upper
152	100		
151	20.9	16.2	24.4
153	12.8	10.2	15.4





#50
Dimethylphthalate
Concen: 5.15 ng
RT: 13.751 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:163 Resp: 9994

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	9.9	8.1	12.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#51
2,6-Dinitrotoluene
Concen: 4.65 ng
RT: 13.869 min Scan# 1833
Delta R.T. -0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:165 Resp: 19052

Ion	Ratio	Lower	Upper
165	100		
63	61.7	42.4	63.6
89	68.7	44.8	67.2#





#52
Acenaphthene
Concen: 5.09 ng
RT: 14.357 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:154 Resp: 7030

Ion	Ratio	Lower	Upper
154	100		
153	117.7	91.0	136.4
152	57.3	42.9	64.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#53
3-Nitroaniline
Concen: 3.92 ng
RT: 14.222 min Scan# 1893
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:138 Resp: 16400

Ion	Ratio	Lower	Upper
138	100		
108	13.4	9.0	13.4
92	109.3	90.1	135.1





#55
Dibenzofuran
Concen: 5.23 ng
RT: 14.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:168 Resp: 11518

Ion	Ratio	Lower	Upper
168	100		
139	41.5	34.2	51.4
169	13.2	10.8	16.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#57
2,4-Dinitrotoluene
Concen: 4.41 ng
RT: 14.675 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:165 Resp: 25481

Ion	Ratio	Lower	Upper
165	100		
63	46.2	33.8	50.6
89	76.8	55.6	83.4
182	2.9	2.4	3.6





#58
Fluorene
Concen: 5.20 ng
RT: 15.363 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:166 Resp: 9330

Ion	Ratio	Lower	Upper
166	100		
165	95.4	78.5	117.7
167	13.4	10.7	16.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#59
2,3,4,6-Tetrachlorophenol
Concen: 4.73 ng
RT: 14.946 min Scan# 2016
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:232 Resp: 23382

Ion	Ratio	Lower	Upper
232	100		
131	48.6	39.5	59.3
130	3.0	2.2	3.4
166	30.5	25.7	38.5





#60
Diethylphthalate
Concen: 5.17 ng
RT: 15.134 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:149 Resp: 10136

Ion	Ratio	Lower	Upper
149	100		
177	20.9	16.7	25.1
150	11.6	10.2	15.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#61
4-Chlorophenyl-phenylether
Concen: 5.35 ng
RT: 15.357 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:204 Resp: 47886

Ion	Ratio	Lower	Upper
204	100		
206	31.0	26.0	39.0
141	67.1	52.9	79.3





#62
4-Nitroaniline
Concen: 3.55 ng m
RT: 15.410 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:138 Resp: 1448

Ion	Ratio	Lower	Upper
138	100		
92	51.3	29.2	69.2
108	88.5	69.0	109.0

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#63
Azobenzene
Concen: 5.03 ng
RT: 15.651 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion: 77 Resp: 84462

Ion	Ratio	Lower	Upper
77	100		
182	26.0	5.7	45.7
105	21.2	1.3	41.3
51	32.1	10.6	50.6





#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.092 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:188 Resp: 53154

Ion	Ratio	Lower	Upper
188	100		
94	8.0	6.5	9.7
80	9.1	7.3	10.9

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#66
n-Nitrosodiphenylamine
Concen: 4.90 ng
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:169 Resp: 80078

Ion	Ratio	Lower	Upper
169	100		
168	67.8	53.2	79.8
167	36.5	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 4.90 ng
RT: 16.275 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:248 Resp: 3070

Ion	Ratio	Lower	Upper
248	100		
250	96.2	78.2	117.4
141	75.3	63.0	94.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#68
Hexachlorobenzene
Concen: 5.06 ng
RT: 16.387 min Scan# 2261
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:284 Resp: 40234

Ion	Ratio	Lower	Upper
284	100		
142	32.9	25.9	38.9
249	29.6	24.1	36.1





#69
Atrazine
Concen: 4.66 ng
RT: 16.551 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:200 Resp: 2790

Ion	Ratio	Lower	Upper
200	100		
173	25.2	5.9	45.9
215	46.5	28.3	68.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#70
Pentachlorophenol
Concen: 3.68 ng
RT: 16.757 min Scan# 2324
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:266 Resp: 16456

Ion	Ratio	Lower	Upper
266	100		
268	69.9	52.2	78.4
264	60.1	51.1	76.7





#71
Phenanthrene
Concen: 5.14 ng
RT: 17.134 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:178 Resp: 15203

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.4	23.2
179	15.7	12.2	18.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#72
Anthracene
Concen: 4.90 ng
RT: 17.234 min Scan# 2405
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:178 Resp: 145809

Ion	Ratio	Lower	Upper
178	100		
176	18.5	15.0	22.4
179	14.8	12.2	18.2





#73
Carbazole
Concen: 4.56 ng
RT: 17.522 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:167 Resp: 12266

Ion	Ratio	Lower	Upper
167	100		
166	22.6	17.7	26.5
139	13.1	11.3	16.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#74
Di-n-butylphthalate
Concen: 4.54 ng
RT: 18.104 min Scan# 2553
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:149 Resp: 157014

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.2	10.8
104	6.0	4.3	6.5





#75
Fluoranthene
Concen: 4.89 ng
RT: 19.228 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:202 Resp: 16886

Ion	Ratio	Lower	Upper
202	100		
101	9.5	0.0	30.3
203	17.2	0.0	37.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.522 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:240 Resp: 614288

Ion	Ratio	Lower	Upper
240	100		
120	9.4	7.8	11.6
236	25.3	20.6	31.0





#78
Pyrene
Concen: 4.94 ng
RT: 19.604 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:202 Resp: 18181

Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.9	25.3
203	17.8	13.9	20.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#79
Terphenyl-d14
Concen: 10.22 ng
RT: 19.822 min Scan# 2845
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:244 Resp: 366245

Ion	Ratio	Lower	Upper
244	100		
212	7.2	5.5	8.3
122	9.4	7.4	11.0





#80
Butylbenzylphthalate
Concen: 4.30 ng
RT: 20.551 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:149 Resp: 7034

Ion	Ratio	Lower	Upper
149	100		
91	69.3	55.4	83.0
206	19.5	16.2	24.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#81
Benzo(a)anthracene
Concen: 4.98 ng
RT: 21.498 min Scan# 3130
Delta R.T. -0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:228 Resp: 192110

Ion	Ratio	Lower	Upper
228	100		
226	27.5	21.4	32.0
229	19.5	15.5	23.3





#82
3,3'-Dichlorobenzidine
Concen: 3.82 ng
RT: 21.445 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:252 Resp: 5479

Ion	Ratio	Lower	Upper
252	100		
254	64.0	51.8	77.6
126	11.0	7.7	11.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#83
Chrysene
Concen: 5.10 ng
RT: 21.575 min Scan# 3143
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:228 Resp: 186449

Ion	Ratio	Lower	Upper
228	100		
226	29.5	23.9	35.9
229	18.9	15.5	23.3





#84
Bis(2-ethylhexyl)phthalate
Concen: 4.29 ng
RT: 21.445 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:149 Resp: 10328

Ion	Ratio	Lower	Upper
149	100		
167	25.9	22.1	33.1
279	5.4	4.6	6.8

Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#85
Di-n-octyl phthalate
Concen: 4.09 ng
RT: 22.692 min Scan# 3333
Delta R.T. 0.000 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:149 Resp: 160883

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	8.2	6.3	9.5





#86
Perylene-d12
Concen: 20.00 ng
RT: 24.810 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:264 Resp: 72426

Ion	Ratio	Lower	Upper
264	100		
260	23.2	18.7	28.1
265	22.0	17.0	25.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#87
Indeno(1,2,3-cd)pyrene
Concen: 4.59 ng
RT: 28.504 min Scan# 4321
Delta R.T. -0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:276 Resp: 242814

Ion	Ratio	Lower	Upper
276	100		
138	7.4	11.2	16.8#
277	24.4	15.7	23.5#





#88
Benzo(b)fluoranthene
Concen: 4.49 ng
RT: 23.763 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:252 Resp: 18605

Ion	Ratio	Lower	Upper
252	100		
253	23.6	17.7	26.5
125	9.0	7.5	11.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#89
Benzo(k)fluoranthene
Concen: 4.86 ng
RT: 23.839 min Scan# 3528
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:252 Resp: 207645

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	8.7	7.0	10.4





#90
Benzo(a)pyrene
Concen: 4.52 ng
RT: 24.657 min Scan# 3
Delta R.T. 0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:252 Resp: 18015

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	10.6	8.1	12.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#91
Dibenzo(a,h)anthracene
Concen: 4.50 ng
RT: 28.586 min Scan# 4335
Delta R.T. -0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:278 Resp: 193373

Ion	Ratio	Lower	Upper
278	100		
139	13.9	10.2	15.4
279	24.7	19.0	28.4





#92
Benzo(g,h,i)perylene
Concen: 4.66 ng
RT: 29.651 min Scan# 4
Delta R.T. -0.012 min
Lab File: BP024615.D
Acq: 13 May 2025 11:22

Tgt Ion:276 Resp: 19936

Ion	Ratio	Lower	Upper
276	100		
277	22.7	19.0	28.4
138	17.2	14.4	21.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024616.D
Acq On : 13 May 2025 12:03
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:04:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.663	152	94972	20.00	ng		0.00
21) Naphthalene-d8	10.434	136	398657	20.00	ng		0.00
39) Acenaphthene-d10	14.292	164	268335	20.00	ng		-0.01
64) Phenanthrene-d10	17.098	188	556673	20.00	ng		0.00
76) Chrysene-d12	21.516	240	659092	20.00	ng		0.00
86) Perylene-d12	24.798	264	780142	20.00	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.293	112	104115	18.75	ng		0.00
7) Phenol-d6	6.858	99	127652	18.01	ng		0.00
23) Nitrobenzene-d5	8.810	82	149876	19.00	ng		0.00
42) 2,4,6-Tribromophenol	15.810	330	86099	18.93	ng		0.00
45) 2-Fluorobiphenyl	12.904	172	401986	19.63	ng		0.00
79) Terphenyl-d14	19.822	244	733687	19.09	ng		0.01
Target Compounds							Qvalue
2) 1,4-Dioxane	3.228	88	26472	10.15	ng		99
3) Pyridine	3.634	79	53915	9.33	ng		98
4) n-Nitrosodimethylamine	3.540	42	24093	9.44	ng	#	96
6) Aniline	7.005	93	80863	8.84	ng		99
8) 2-Chlorophenol	7.240	128	57182	9.07	ng		99
9) Benzaldehyde	6.822	77	47848m	10.45	ng		
10) Phenol	6.881	94	66520	9.09	ng		96
11) bis(2-Chloroethyl)ether	7.093	93	55575	9.25	ng		99
12) 1,3-Dichlorobenzene	7.557	146	70590	9.73	ng		100
13) 1,4-Dichlorobenzene	7.699	146	71467	9.80	ng		99
14) 1,2-Dichlorobenzene	8.010	146	69643	9.80	ng		96
15) Benzyl Alcohol	7.916	79	45094	8.37	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.187	45	68599	9.53	ng		98
17) 2-Methylphenol	8.116	107	45071	8.59	ng		96
18) Hexachloroethane	8.728	117	28078	10.05	ng		96
19) n-Nitroso-di-n-propyla...	8.463	70	45927	9.37	ng		96
20) 3+4-Methylphenols	8.446	107	61832	8.66	ng		95
22) Acetophenone	8.481	105	95803	9.62	ng	#	99
24) Nitrobenzene	8.857	77	67735	9.76	ng		98
25) Isophorone	9.375	82	127474	9.31	ng		99
26) 2-Nitrophenol	9.563	139	28470	8.37	ng		94
27) 2,4-Dimethylphenol	9.628	122	55664	9.28	ng		98
28) bis(2-Chloroethoxy)met...	9.851	93	78019	9.54	ng		97
29) 2,4-Dichlorophenol	10.110	162	51386	8.84	ng		98
30) 1,2,4-Trichlorobenzene	10.304	180	64245	9.81	ng		96
31) Naphthalene	10.481	128	204468	9.90	ng		99
32) Benzoic acid	9.746	122	17770m	10.42	ng		
33) 4-Chloroaniline	10.610	127	73990	8.88	ng		99
34) Hexachlorobutadiene	10.769	225	42567	10.03	ng		97
35) Caprolactam	11.381	113	18245m	8.68	ng		
36) 4-Chloro-3-methylphenol	11.746	107	64857	9.15	ng		98
37) 2-Methylnaphthalene	12.098	142	130412	9.75	ng		100
38) 1-Methylnaphthalene	12.316	142	139573	9.75	ng		100
40) 1,2,4,5-Tetrachloroben...	12.475	216	74824	9.60	ng		100
41) Hexachlorocyclopentadiene	12.451	237	37072	7.85	ng		98
43) 2,4,6-Trichlorophenol	12.722	196	45798	9.07	ng		96

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024616.D
Acq On : 13 May 2025 12:03
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:04:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.810	196	50348	8.96	ng		94
46) 1,1'-Biphenyl	13.122	154	192445	9.70	ng		100
47) 2-Chloronaphthalene	13.163	162	148443	9.82	ng		99
48) 2-Nitroaniline	13.375	65	40786	9.11	ng		93
49) Acenaphthylene	14.010	152	243913	9.64	ng		98
50) Dimethylphthalate	13.751	163	204825	10.01	ng		99
51) 2,6-Dinitrotoluene	13.869	165	41257	9.55	ng		92
52) Acenaphthene	14.357	154	141360	9.72	ng		99
53) 3-Nitroaniline	14.216	138	36906	8.37	ng		99
54) 2,4-Dinitrophenol	14.439	184	13128	10.84	ng		95
55) Dibenzofuran	14.704	168	230286	9.92	ng		99
56) 4-Nitrophenol	14.575	139	21349	9.53	ng		93
57) 2,4-Dinitrotoluene	14.675	165	56617	9.30	ng		94
58) Fluorene	15.357	166	189781	10.03	ng		100
59) 2,3,4,6-Tetrachlorophenol	14.939	232	48821	9.37	ng		99
60) Diethylphthalate	15.134	149	206886	10.01	ng		99
61) 4-Chlorophenyl-phenyle...	15.357	204	93517	9.91	ng		97
62) 4-Nitroaniline	15.404	138	34813m	8.11	ng		
63) Azobenzene	15.657	77	174322	9.86	ng		98
65) 4,6-Dinitro-2-methylph...	15.457	198	26174	7.28	ng		96
66) n-Nitrosodiphenylamine	15.575	169	164008	9.59	ng		99
67) 4-Bromophenyl-phenylether	16.269	248	62011	9.45	ng		97
68) Hexachlorobenzene	16.392	284	80088	9.61	ng		98
69) Atrazine	16.551	200	58245	9.30	ng		99
70) Pentachlorophenol	16.757	266	38095m	8.14	ng		
71) Phenanthrene	17.139	178	301863	9.75	ng		100
72) Anthracene	17.239	178	296146	9.51	ng		100
73) Carbazole	17.522	167	262687	9.32	ng		100
74) Di-n-butylphthalate	18.116	149	338893	9.35	ng		100
75) Fluoranthene	19.227	202	349083	9.65	ng		99
77) Benzidine	19.445	184	119229m	6.59	ng		
78) Pyrene	19.604	202	374573	9.48	ng		100
80) Butylbenzylphthalate	20.551	149	153275	8.73	ng		99
81) Benzo(a)anthracene	21.498	228	401322	9.70	ng		99
82) 3,3'-Dichlorobenzidine	21.433	252	129324	8.41	ng		99
83) Chrysene	21.569	228	388846	9.91	ng		98
84) Bis(2-ethylhexyl)phtha...	21.439	149	225047	8.72	ng		100
85) Di-n-octyl phthalate	22.680	149	367934	8.72	ng		100
87) Indeno(1,2,3-cd)pyrene	28.503	276	529260	9.29	ng	#	92
88) Benzo(b)fluoranthene	23.751	252	417343	9.35	ng		99
89) Benzo(k)fluoranthene	23.821	252	431575	9.38	ng		99
90) Benzo(a)pyrene	24.627	252	390288	9.10	ng		100
91) Dibenzo(a,h)anthracene	28.574	278	415768	8.97	ng		99
92) Benzo(g,h,i)perylene	29.650	276	431559	9.37	ng		97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024616.D
Acq On : 13 May 2025 12:03
Operator : RC/JU
Sample : SSTDICC010
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 13 16:04:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.663 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:152 Resp: 9497

Ion	Ratio	Lower	Upper
152	100		
150	158.0	125.9	188.9
115	64.2	50.1	75.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#2
1,4-Dioxane
Concen: 10.15 ng
RT: 3.228 min Scan# 24
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 88 Resp: 26472

Ion	Ratio	Lower	Upper
88	100		
58	67.6	54.0	81.0
43	27.8	21.4	32.0





#3
Pyridine
Concen: 9.33 ng
RT: 3.634 min Scan# 9
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 79 Resp: 5391

Ion	Ratio	Lower	Upper
79	100		
52	62.8	50.2	75.2
51	37.1	26.8	40.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#4
n-Nitrosodimethylamine
Concen: 9.44 ng
RT: 3.540 min Scan# 77
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 42 Resp: 24093

Ion	Ratio	Lower	Upper
42	100		
74	127.8	106.2	159.2
44	9.9	5.4	8.2#





#5
2-Fluorophenol
Concen: 18.75 ng
RT: 5.293 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:112 Resp: 10411

Ion	Ratio	Lower	Upper
112	100		
64	54.6	44.1	66.1
63	29.0	23.5	35.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#6
Aniline
Concen: 8.84 ng
RT: 7.005 min Scan# 666
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 93 Resp: 80863

Ion	Ratio	Lower	Upper
93	100		
66	39.2	31.0	46.6
65	19.8	15.5	23.3





#7
Phenol-d6
Concen: 18.01 ng
RT: 6.858 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 99 Resp: 12765

Ion	Ratio	Lower	Upper
99	100		
42	18.4	14.4	21.6
71	36.3	29.2	43.8

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#8
2-Chlorophenol
Concen: 9.07 ng
RT: 7.240 min Scan# 706
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:128 Resp: 57182

Ion	Ratio	Lower	Upper
128	100		
130	31.6	13.0	53.0
64	42.7	22.3	62.3





#9
Benzaldehyde
Concen: 10.45 ng m
RT: 6.822 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 77 Resp: 4784

Ion	Ratio	Lower	Upper
77	100		
105	106.2	83.3	123.3
106	100.5	78.0	118.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#10
Phenol
Concen: 9.09 ng
RT: 6.881 min Scan# 645
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 94 Resp: 66520

Ion	Ratio	Lower	Upper
94	100		
65	27.8	8.7	48.7
66	42.7	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 9.25 ng
RT: 7.093 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 93 Resp: 5557

Ion	Ratio	Lower	Upper
93	100		
63	69.5	50.6	90.6
95	32.4	12.9	52.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#12
1,3-Dichlorobenzene
Concen: 9.73 ng
RT: 7.557 min Scan# 760
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:146 Resp: 70590

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.4	77.2
75	31.1	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 9.80 ng
RT: 7.699 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:146 Resp: 7146

Ion	Ratio	Lower	Upper
146	100		
148	63.9	51.0	76.6
111	40.6	33.0	49.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#14
1,2-Dichlorobenzene
Concen: 9.80 ng
RT: 8.010 min Scan# 837
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:146 Resp: 69643

Ion	Ratio	Lower	Upper
146	100		
148	67.5	51.0	76.4
111	46.5	35.2	52.8





#15
Benzyl Alcohol
Concen: 8.37 ng
RT: 7.916 min Scan# 8
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 79 Resp: 4509

Ion	Ratio	Lower	Upper
79	100		
108	76.3	60.8	91.2
77	63.5	51.9	77.9

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#16
2,2'-oxybis(1-Chloropropane)
Concen: 9.53 ng
RT: 8.187 min Scan# 867
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 45 Resp: 68599

Ion	Ratio	Lower	Upper
45	100		
77	16.5	0.0	37.7
79	14.1	0.0	34.0





#17
2-Methylphenol
Concen: 8.59 ng
RT: 8.116 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:107 Resp: 4507

Ion	Ratio	Lower	Upper
107	100		
108	114.4	88.9	133.3
77	51.8	39.0	58.6
79	51.1	38.4	57.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#18
Hexachloroethane
Concen: 10.05 ng
RT: 8.728 min Scan# 959
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:117 Resp: 28078

Ion	Ratio	Lower	Upper
117	100		
119	92.3	79.4	119.0
201	101.4	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 9.37 ng
RT: 8.463 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 70 Resp: 4592

Ion	Ratio	Lower	Upper
70	100		
42	58.2	43.0	64.6
101	11.2	8.7	13.1
130	24.1	19.4	29.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#20
3+4-Methylphenols
Concen: 8.66 ng
RT: 8.446 min Scan# 911
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:107 Resp: 61832

Ion	Ratio	Lower	Upper
107	100		
108	91.0	66.7	106.7
77	35.6	13.1	53.1
79	32.7	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:136 Resp: 39865

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	6.7	5.5	8.3
68	5.2	4.2	6.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#22
Acetophenone
Concen: 9.62 ng
RT: 8.481 min Scan# 917
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:105 Resp: 95803

Ion	Ratio	Lower	Upper
105	100		
71	1.9	2.0	3.0#
51	25.8	20.4	30.6
120	22.0	18.1	27.1





#23
Nitrobenzene-d5
Concen: 19.00 ng
RT: 8.810 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 82 Resp: 14987

Ion	Ratio	Lower	Upper
82	100		
128	44.1	37.0	55.4
54	48.5	38.7	58.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#24
Nitrobenzene
Concen: 9.76 ng
RT: 8.857 min Scan# 981
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 77 Resp: 67735

Ion	Ratio	Lower	Upper
77	100		
123	45.2	37.4	56.0
65	14.4	11.3	16.9





#25
Isophorone
Concen: 9.31 ng
RT: 9.375 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 82 Resp: 12747

Ion	Ratio	Lower	Upper
82	100		
95	8.4	6.3	9.5
138	18.7	15.3	22.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#26
2-Nitrophenol
Concen: 8.37 ng
RT: 9.563 min Scan# 1101
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:139 Resp: 28470

Ion	Ratio	Lower	Upper
139	100		
109	36.2	25.9	38.9
65	51.3	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 9.28 ng
RT: 9.628 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:122 Resp: 5566

Ion	Ratio	Lower	Upper
122	100		
107	118.0	96.2	144.2
121	58.4	47.6	71.4

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#28
bis(2-Chloroethoxy)methane
Concen: 9.54 ng
RT: 9.851 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 93 Resp: 78019

Ion	Ratio	Lower	Upper
93	100		
95	33.5	25.5	38.3
123	11.2	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 8.84 ng
RT: 10.110 min Scan# 11
Delta R.T. 0.018 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:162 Resp: 5138

Ion	Ratio	Lower	Upper
162	100		
164	61.7	42.3	82.3
98	36.7	18.4	58.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#30
1,2,4-Trichlorobenzene
Concen: 9.81 ng
RT: 10.304 min Scan# 1227
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:180 Resp: 64245

Ion	Ratio	Lower	Upper
180	100		
182	94.1	79.0	118.4
145	32.8	26.6	39.8





#31
Naphthalene
Concen: 9.90 ng
RT: 10.481 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:128 Resp: 20446

Ion	Ratio	Lower	Upper
128	100		
129	11.5	8.8	13.2
127	13.2	10.7	16.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#32
Benzoic acid
Concen: 10.42 ng m
RT: 9.746 min Scan# 1132
Delta R.T. -0.029 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:122 Resp: 17770

Ion	Ratio	Lower	Upper
122	100		
105	110.3	106.9	146.9
77	95.0	75.1	115.1





#33
4-Chloroaniline
Concen: 8.88 ng
RT: 10.610 min Scan# 1
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:127 Resp: 7399

Ion	Ratio	Lower	Upper
127	100		
129	31.9	25.8	38.6
65	29.5	23.9	35.9
92	19.5	16.1	24.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#34
Hexachlorobutadiene
Concen: 10.03 ng
RT: 10.769 min Scan# 1306
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:225 Resp: 42567

Ion	Ratio	Lower	Upper
225	100		
223	61.0	51.0	76.6
227	64.7	53.1	79.7





#35
Caprolactam
Concen: 8.68 ng m
RT: 11.381 min Scan# 14
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:113 Resp: 1824

Ion	Ratio	Lower	Upper
113	100		
55	134.4	119.0	159.0
56	107.7	90.3	130.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#36
4-Chloro-3-methylphenol
Concen: 9.15 ng
RT: 11.746 min Scan# 1472
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:107 Resp: 64857

Ion	Ratio	Lower	Upper
107	100		
144	25.3	20.1	30.1
142	75.0	62.1	93.1





#37
2-Methylnaphthalene
Concen: 9.75 ng
RT: 12.098 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:142 Resp: 13041

Ion	Ratio	Lower	Upper
142	100		
141	88.9	71.2	106.8
115	38.0	30.6	46.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#38
1-Methylnaphthalene
Concen: 9.75 ng
RT: 12.316 min Scan# 1569
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:142 Resp: 139573

Ion	Ratio	Lower	Upper
142	100		
141	93.1	74.7	112.1
116	4.0	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.292 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:164 Resp: 26833

Ion	Ratio	Lower	Upper
164	100		
162	103.8	81.8	122.6
160	46.2	36.3	54.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 9.60 ng
RT: 12.475 min Scan# 1596
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:216 Resp: 74824

Ion	Ratio	Lower	Upper
216	100		
214	77.9	62.3	93.5
179	22.2	17.7	26.5
108	20.0	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 7.85 ng
RT: 12.451 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:237 Resp: 3707

Ion	Ratio	Lower	Upper
237	100		
235	61.4	39.9	79.9
272	14.2	0.0	33.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#42
2,4,6-Tribromophenol
Concen: 18.93 ng
RT: 15.810 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:330 Resp: 86099

Ion	Ratio	Lower	Upper
330	100		
332	95.1	78.2	117.4
141	30.5	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 9.07 ng
RT: 12.722 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:196 Resp: 4579

Ion	Ratio	Lower	Upper
196	100		
198	91.7	76.9	115.3
200	30.5	25.2	37.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#44
2,4,5-Trichlorophenol
Concen: 8.96 ng
RT: 12.810 min Scan# 1653
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:196 Resp: 50348

Ion	Ratio	Lower	Upper
196	100		
198	92.8	79.2	118.8
97	47.0	42.3	63.5
132	28.1	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 19.63 ng
RT: 12.904 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:172 Resp: 40198

Ion	Ratio	Lower	Upper
172	100		
171	36.0	28.8	43.2
170	23.9	18.8	28.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#46
1,1'-Biphenyl
Concen: 9.70 ng
RT: 13.122 min Scan# 1706
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:154 Resp: 192445

Ion	Ratio	Lower	Upper
154	100		
153	41.1	21.0	61.0
76	11.5	0.0	31.5





#47
2-Chloronaphthalene
Concen: 9.82 ng
RT: 13.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:162 Resp: 14844

Ion	Ratio	Lower	Upper
162	100		
127	41.8	33.8	50.6
164	32.0	26.0	39.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#48
2-Nitroaniline
Concen: 9.11 ng
RT: 13.375 min Scan# 1749
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 65 Resp: 40786

Ion	Ratio	Lower	Upper
65	100		
92	71.8	58.7	88.1
138	101.2	90.2	135.2





#49
Acenaphthylene
Concen: 9.64 ng
RT: 14.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:152 Resp: 24391

Ion	Ratio	Lower	Upper
152	100		
151	21.1	16.2	24.4
153	13.3	10.2	15.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#50
Dimethylphthalate
Concen: 10.01 ng
RT: 13.751 min Scan# 1813
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:163 Resp: 204825

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	10.3	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 9.55 ng
RT: 13.869 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:165 Resp: 4125

Ion	Ratio	Lower	Upper
165	100		
63	57.3	42.4	63.6
89	63.7	44.8	67.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#52
Acenaphthene
Concen: 9.72 ng
RT: 14.357 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:154 Resp: 141360

Ion	Ratio	Lower	Upper
154	100		
153	113.4	91.0	136.4
152	54.6	42.9	64.3





#53
3-Nitroaniline
Concen: 8.37 ng
RT: 14.216 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:138 Resp: 3690

Ion	Ratio	Lower	Upper
138	100		
108	9.9	9.0	13.4
92	112.3	90.1	135.1

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#54
2,4-Dinitrophenol
Concen: 10.84 ng
RT: 14.439 min Scan# 1930
Delta R.T. 0.011 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:184 Resp: 13128

Ion	Ratio	Lower	Upper
184	100		
63	60.3	51.3	76.9
154	62.9	53.5	80.3





#55
Dibenzofuran
Concen: 9.92 ng
RT: 14.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:168 Resp: 23028

Ion	Ratio	Lower	Upper
168	100		
139	43.3	34.2	51.4
169	13.2	10.8	16.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#56
4-Nitrophenol
Concen: 9.53 ng
RT: 14.575 min Scan# 1953
Delta R.T. 0.024 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:139 Resp: 21349

Ion	Ratio	Lower	Upper
139	100		
109	84.4	61.4	101.4
65	101.4	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 9.30 ng
RT: 14.675 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:165 Resp: 5661

Ion	Ratio	Lower	Upper
165	100		
63	47.4	33.8	50.6
89	74.1	55.6	83.4
182	3.0	2.4	3.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#58
Fluorene
Concen: 10.03 ng
RT: 15.357 min Scan# 2086
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:166 Resp: 189781

Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.5	117.7
167	13.5	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 9.37 ng
RT: 14.939 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:232 Resp: 4882

Ion	Ratio	Lower	Upper
232	100		
131	50.6	39.5	59.3
130	2.8	2.2	3.4
166	32.1	25.7	38.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#60
Diethylphthalate
Concen: 10.01 ng
RT: 15.134 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:149 Resp: 206886

Ion	Ratio	Lower	Upper
149	100		
177	20.6	16.7	25.1
150	12.4	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 9.91 ng
RT: 15.357 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:204 Resp: 9351

Ion	Ratio	Lower	Upper
204	100		
206	33.5	26.0	39.0
141	68.7	52.9	79.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#62
4-Nitroaniline
Concen: 8.11 ng m
RT: 15.404 min Scan# 2094
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:138 Resp: 34813

Ion	Ratio	Lower	Upper
138	100		
92	57.0	29.2	69.2
108	95.0	69.0	109.0





#63
Azobenzene
Concen: 9.86 ng
RT: 15.657 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion: 77 Resp: 17432

Ion	Ratio	Lower	Upper
77	100		
182	23.5	5.7	45.7
105	20.9	1.3	41.3
51	30.1	10.6	50.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.098 min Scan# 2382
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:188 Resp: 556673

Ion	Ratio	Lower	Upper
188	100		
94	8.0	6.5	9.7
80	9.0	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 7.28 ng
RT: 15.457 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:198 Resp: 2617

Ion	Ratio	Lower	Upper
198	100		
51	40.9	21.1	61.1
105	47.8	23.6	63.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#66
n-Nitrosodiphenylamine
Concen: 9.59 ng
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:169 Resp: 164008

Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.2	79.8
167	37.2	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 9.45 ng
RT: 16.269 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:248 Resp: 6201

Ion	Ratio	Lower	Upper
248	100		
250	98.7	78.2	117.4
141	84.3	63.0	94.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#68
Hexachlorobenzene
Concen: 9.61 ng
RT: 16.392 min Scan# 2262
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:284 Resp: 80088

Ion	Ratio	Lower	Upper
284	100		
142	34.2	25.9	38.9
249	29.6	24.1	36.1





#69
Atrazine
Concen: 9.30 ng
RT: 16.551 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:200 Resp: 5824

Ion	Ratio	Lower	Upper
200	100		
173	26.8	5.9	45.9
215	47.8	28.3	68.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#70
Pentachlorophenol
Concen: 8.14 ng m
RT: 16.757 min Scan# 2324
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:266 Resp: 38095

Ion	Ratio	Lower	Upper
266	100		
268	66.7	52.2	78.4
264	63.1	51.1	76.7





#71
Phenanthrene
Concen: 9.75 ng
RT: 17.139 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:178 Resp: 30186

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.2
179	15.1	12.2	18.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#72
Anthracene
Concen: 9.51 ng
RT: 17.239 min Scan# 2406
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:178 Resp: 296146

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.5	12.2	18.2





#73
Carbazole
Concen: 9.32 ng
RT: 17.522 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:167 Resp: 26268

Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.7	26.5
139	13.9	11.3	16.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#74
Di-n-butylphthalate
Concen: 9.35 ng
RT: 18.116 min Scan# 2555
Delta R.T. 0.018 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:149 Resp: 338893

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.2	10.8
104	5.5	4.3	6.5





#75
Fluoranthene
Concen: 9.65 ng
RT: 19.227 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:202 Resp: 34908

Ion	Ratio	Lower	Upper
202	100		
101	9.6	0.0	30.3
203	17.1	0.0	37.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.516 min Scan# 3133
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:240 Resp: 659092

Ion	Ratio	Lower	Upper
240	100		
120	9.8	7.8	11.6
236	25.9	20.6	31.0





#77
Benzidine
Concen: 6.59 ng m
RT: 19.445 min Scan# 2
Delta R.T. 0.029 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:184 Resp: 11922

Ion	Ratio	Lower	Upper
184	100		
185	13.7	11.2	16.8
183	12.1	9.8	14.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#78
Pyrene
Concen: 9.48 ng
RT: 19.604 min Scan# 2808
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:202 Resp: 374573

Ion	Ratio	Lower	Upper
202	100		
200	21.3	16.9	25.3
203	17.2	13.9	20.9





#79
Terphenyl-d14
Concen: 19.09 ng
RT: 19.822 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:244 Resp: 73368

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.5	8.3
122	9.5	7.4	11.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#80
Butylbenzylphthalate
Concen: 8.73 ng
RT: 20.551 min Scan# 2969
Delta R.T. 0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:149 Resp: 153275

Ion	Ratio	Lower	Upper
149	100		
91	68.8	55.4	83.0
206	21.3	16.2	24.2





#81
Benzo(a)anthracene
Concen: 9.70 ng
RT: 21.498 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:228 Resp: 40132

Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.4	32.0
229	19.5	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#82
3,3'-Dichlorobenzidine
Concen: 8.41 ng
RT: 21.433 min Scan# 3119
Delta R.T. 0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:252 Resp: 129324

Ion	Ratio	Lower	Upper
252	100		
254	65.2	51.8	77.6
126	11.2	7.7	11.5





#83
Chrysene
Concen: 9.91 ng
RT: 21.569 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:228 Resp: 38884

Ion	Ratio	Lower	Upper
228	100		
226	28.8	23.9	35.9
229	19.0	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 8.72 ng
RT: 21.439 min Scan# 3120
Delta R.T. -0.000 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:149 Resp: 225047

Ion	Ratio	Lower	Upper
149	100		
167	27.6	22.1	33.1
279	5.6	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 8.72 ng
RT: 22.680 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:149 Resp: 36793

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	8.0	6.3	9.5

Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#86
Perylene-d12
Concen: 20.00 ng
RT: 24.798 min Scan# 3691
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:264 Resp: 780142

Ion	Ratio	Lower	Upper
264	100		
260	23.5	18.7	28.1
265	22.3	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 9.29 ng
RT: 28.503 min Scan# 4
Delta R.T. -0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:276 Resp: 52926

Ion	Ratio	Lower	Upper
276	100		
138	15.5	11.2	16.8
277	24.8	15.7	23.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#88
Benzo(b)fluoranthene
Concen: 9.35 ng
RT: 23.751 min Scan# 3513
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:252 Resp: 417343

Ion	Ratio	Lower	Upper
252	100		
253	22.6	17.7	26.5
125	9.1	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 9.38 ng
RT: 23.821 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:252 Resp: 43157

Ion	Ratio	Lower	Upper
252	100		
253	21.4	17.6	26.4
125	8.7	7.0	10.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#90
Benzo(a)pyrene
Concen: 9.10 ng
RT: 24.627 min Scan# 3662
Delta R.T. -0.018 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:252 Resp: 390288

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	10.2	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 8.97 ng
RT: 28.574 min Scan# 4
Delta R.T. -0.024 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:278 Resp: 41576

Ion	Ratio	Lower	Upper
278	100		
139	12.7	10.2	15.4
279	24.7	19.0	28.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC010

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#92
Benzo(g,h,i)perylene
Concen: 9.37 ng
RT: 29.650 min Scan# 4516
Delta R.T. -0.012 min
Lab File: BP024616.D
Acq: 13 May 2025 12:03

Tgt Ion:276 Resp: 431559

Ion	Ratio	Lower	Upper
276	100		
277	25.3	19.0	28.4
138	16.8	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024617.D
Acq On : 13 May 2025 12:43
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:05:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.664	152	112650	20.00	ng		0.00
21) Naphthalene-d8	10.434	136	488721	20.00	ng		0.00
39) Acenaphthene-d10	14.299	164	323928	20.00	ng		0.00
64) Phenanthrene-d10	17.087	188	644367	20.00	ng		-0.01
76) Chrysene-d12	21.522	240	745224	20.00	ng		0.00
86) Perylene-d12	24.792	264	850264	20.00	ng		-0.01
System Monitoring Compounds							
5) 2-Fluorophenol	5.287	112	261821	39.75	ng		0.00
7) Phenol-d6	6.858	99	334811	39.83	ng		0.00
23) Nitrobenzene-d5	8.811	82	391739	40.50	ng		0.00
42) 2,4,6-Tribromophenol	15.810	330	217092	39.53	ng		0.00
45) 2-Fluorobiphenyl	12.904	172	1006150	40.69	ng		0.00
79) Terphenyl-d14	19.810	244	1736879	39.96	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.229	88	60251	19.48	ng		98
3) Pyridine	3.628	79	135451	19.75	ng		99
4) n-Nitrosodimethylamine	3.540	42	59324	19.59	ng		99
6) Aniline	7.005	93	217646	20.05	ng		100
8) 2-Chlorophenol	7.240	128	147726	19.76	ng		99
9) Benzaldehyde	6.822	77	117823m	21.69	ng		
10) Phenol	6.881	94	171949	19.82	ng		99
11) bis(2-Chloroethyl)ether	7.099	93	141599	19.86	ng		99
12) 1,3-Dichlorobenzene	7.552	146	171270	19.91	ng		99
13) 1,4-Dichlorobenzene	7.699	146	171756	19.85	ng		98
14) 1,2-Dichlorobenzene	8.011	146	168795	20.02	ng		99
15) Benzyl Alcohol	7.911	79	127076	19.89	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.187	45	171274	20.06	ng		100
17) 2-Methylphenol	8.111	107	124506	20.01	ng		99
18) Hexachloroethane	8.728	117	65527	19.76	ng		97
19) n-Nitroso-di-n-propyla...	8.458	70	122386	21.05	ng		98
20) 3+4-Methylphenols	8.440	107	169264	20.00	ng		97
22) Acetophenone	8.481	105	249692	20.45	ng	#	100
24) Nitrobenzene	8.858	77	173686	20.40	ng		99
25) Isophorone	9.375	82	339697	20.25	ng		100
26) 2-Nitrophenol	9.563	139	82278	19.73	ng		98
27) 2,4-Dimethylphenol	9.622	122	146819	19.96	ng		97
28) bis(2-Chloroethoxy)met...	9.852	93	200872	20.04	ng		99
29) 2,4-Dichlorophenol	10.099	162	141785	19.90	ng		98
30) 1,2,4-Trichlorobenzene	10.299	180	159998	19.92	ng		95
31) Naphthalene	10.481	128	512888	20.25	ng		100
32) Benzoic acid	9.746	122	71538m	18.73	ng		
33) 4-Chloroaniline	10.605	127	205433	20.10	ng		99
34) Hexachlorobutadiene	10.763	225	103364	19.87	ng		99
35) Caprolactam	11.381	113	50721	19.68	ng		99
36) 4-Chloro-3-methylphenol	11.734	107	174840	20.12	ng		99
37) 2-Methylnaphthalene	12.093	142	327447	19.97	ng		99
38) 1-Methylnaphthalene	12.316	142	352091	20.07	ng		100
40) 1,2,4,5-Tetrachloroben...	12.469	216	185982	19.77	ng		99
41) Hexachlorocyclopentadiene	12.440	237	113499	19.90	ng		98
43) 2,4,6-Trichlorophenol	12.722	196	121520	19.95	ng		100

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024617.D
Acq On : 13 May 2025 12:43
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:05:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.799	196	135306	19.95	ng	98
46) 1,1'-Biphenyl	13.116	154	483352	20.19	ng	100
47) 2-Chloronaphthalene	13.163	162	366247	20.06	ng	100
48) 2-Nitroaniline	13.375	65	107062	19.81	ng	98
49) Acenaphthylene	14.016	152	621997	20.36	ng	100
50) Dimethylphthalate	13.751	163	491623	19.91	ng	99
51) 2,6-Dinitrotoluene	13.869	165	105479	20.23	ng	97
52) Acenaphthene	14.357	154	354519	20.19	ng	99
53) 3-Nitroaniline	14.210	138	106527	20.01	ng	98
54) 2,4-Dinitrophenol	14.428	184	46351	19.37	ng	91
55) Dibenzofuran	14.698	168	563719	20.12	ng	99
56) 4-Nitrophenol	14.551	139	71975	19.02	ng	93
57) 2,4-Dinitrotoluene	14.675	165	147676	20.09	ng	99
58) Fluorene	15.357	166	456792	20.00	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.934	232	123265	19.59	ng	99
60) Diethylphthalate	15.128	149	497916	19.95	ng	99
61) 4-Chlorophenyl-phenyle...	15.351	204	226790	19.91	ng	98
62) 4-Nitroaniline	15.393	138	105617	20.37	ng	99
63) Azobenzene	15.651	77	439183	20.57	ng	99
65) 4,6-Dinitro-2-methylph...	15.451	198	77131	18.53	ng	96
66) n-Nitrosodiphenylamine	15.575	169	404085	20.42	ng	99
67) 4-Bromophenyl-phenylether	16.263	248	152393	20.07	ng	97
68) Hexachlorobenzene	16.381	284	192290	19.93	ng	99
69) Atrazine	16.545	200	146468	20.19	ng	99
70) Pentachlorophenol	16.740	266	104916	19.36	ng	98
71) Phenanthrene	17.128	178	724922	20.24	ng	100
72) Anthracene	17.222	178	735787	20.42	ng	100
73) Carbazole	17.510	167	669730	20.53	ng	99
74) Di-n-butylphthalate	18.092	149	858522	20.46	ng	100
75) Fluoranthene	19.210	202	846684	20.22	ng	100
77) Benzidine	19.422	184	404890	19.79	ng	99
78) Pyrene	19.592	202	892037	19.98	ng	99
80) Butylbenzylphthalate	20.539	149	400050	20.14	ng	97
81) Benzo(a)anthracene	21.498	228	933614	19.95	ng	99
82) 3,3'-Dichlorobenzidine	21.428	252	353000	20.31	ng	100
83) Chrysene	21.569	228	883001	19.91	ng	99
84) Bis(2-ethylhexyl)phtha...	21.439	149	597012	20.45	ng	98
85) Di-n-octyl phthalate	22.686	149	947571	19.85	ng	99
87) Indeno(1,2,3-cd)pyrene	28.486	276	1230125m	19.81	ng	
88) Benzo(b)fluoranthene	23.757	252	979873	20.13	ng	99
89) Benzo(k)fluoranthene	23.822	252	1005349	20.05	ng	100
90) Benzo(a)pyrene	24.627	252	945764	20.23	ng	100
91) Dibenzo(a,h)anthracene	28.568	278	1008288	19.97	ng	99
92) Benzo(g,h,i)perylene	29.627	276	991852	19.75	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024617.D
Acq On : 13 May 2025 12:43
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Quant Time: May 13 16:05:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.664 min Scan# 7
Delta R.T. 0.001 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:152 Resp: 11265

Ion	Ratio	Lower	Upper
152	100		
150	156.4	125.9	188.9
115	63.6	50.1	75.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#2
1,4-Dioxane
Concen: 19.48 ng
RT: 3.229 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 88 Resp: 60251

Ion	Ratio	Lower	Upper
88	100		
58	69.2	54.0	81.0
43	27.3	21.4	32.0





#3
Pyridine
Concen: 19.75 ng
RT: 3.628 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 79 Resp: 13545

Ion	Ratio	Lower	Upper
79	100		
52	62.1	50.2	75.2
51	34.1	26.8	40.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#4
n-Nitrosodimethylamine
Concen: 19.59 ng
RT: 3.540 min Scan# 77
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 42 Resp: 59324

Ion	Ratio	Lower	Upper
42	100		
74	131.7	106.2	159.2
44	7.0	5.4	8.2





#5
2-Fluorophenol
Concen: 39.75 ng
RT: 5.287 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:112 Resp: 26182

Ion	Ratio	Lower	Upper
112	100		
64	54.8	44.1	66.1
63	30.0	23.5	35.3

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#6
Aniline
Concen: 20.05 ng
RT: 7.005 min Scan# 666
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 93 Resp: 217646

Ion	Ratio	Lower	Upper
93	100		
66	38.8	31.0	46.6
65	19.5	15.5	23.3





#7
Phenol-d6
Concen: 39.83 ng
RT: 6.858 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 99 Resp: 33481

Ion	Ratio	Lower	Upper
99	100		
42	17.4	14.4	21.6
71	35.3	29.2	43.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#8
2-Chlorophenol
Concen: 19.76 ng
RT: 7.240 min Scan# 706
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:128 Resp: 147726

Ion	Ratio	Lower	Upper
128	100		
130	33.5	13.0	53.0
64	43.1	22.3	62.3





#9
Benzaldehyde
Concen: 21.69 ng m
RT: 6.822 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 77 Resp: 11782

Ion	Ratio	Lower	Upper
77	100		
105	98.5	83.3	123.3
106	94.4	78.0	118.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#10
Phenol
Concen: 19.82 ng
RT: 6.881 min Scan# 645
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 94 Resp: 171949

Ion	Ratio	Lower	Upper
94	100		
65	28.9	8.7	48.7
66	40.3	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 19.86 ng
RT: 7.099 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 93 Resp: 14159

Ion	Ratio	Lower	Upper
93	100		
63	69.4	50.6	90.6
95	32.0	12.9	52.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#12
1,3-Dichlorobenzene
Concen: 19.91 ng
RT: 7.552 min Scan# 759
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:146 Resp: 171270

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.4	77.2
75	32.4	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 19.85 ng
RT: 7.699 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:146 Resp: 17175

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.0	76.6
111	43.7	33.0	49.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#14
1,2-Dichlorobenzene
Concen: 20.02 ng
RT: 8.011 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:146 Resp: 168795

Ion	Ratio	Lower	Upper
146	100		
148	62.2	51.0	76.4
111	43.9	35.2	52.8





#15
Benzyl Alcohol
Concen: 19.89 ng
RT: 7.911 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 79 Resp: 12707

Ion	Ratio	Lower	Upper
79	100		
108	74.8	60.8	91.2
77	65.1	51.9	77.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 20.06 ng
RT: 8.187 min Scan# 867
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 45 Resp: 171274

Ion	Ratio	Lower	Upper
45	100		
77	17.5	0.0	37.7
79	14.1	0.0	34.0





#17
2-Methylphenol
Concen: 20.01 ng
RT: 8.111 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:107 Resp: 12450

Ion	Ratio	Lower	Upper
107	100		
108	110.8	88.9	133.3
77	47.9	39.0	58.6
79	48.4	38.4	57.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#18
Hexachloroethane
Concen: 19.76 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:117 Resp: 65527

Ion	Ratio	Lower	Upper
117	100		
119	102.4	79.4	119.0
201	100.8	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 21.05 ng
RT: 8.458 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 70 Resp: 12238

Ion	Ratio	Lower	Upper
70	100		
42	52.1	43.0	64.6
101	10.5	8.7	13.1
130	23.4	19.4	29.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#20
3+4-Methylphenols
Concen: 20.00 ng
RT: 8.440 min Scan# 910
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:107 Resp: 169264

Ion	Ratio	Lower	Upper
107	100		
108	88.3	66.7	106.7
77	35.9	13.1	53.1
79	30.9	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:136 Resp: 48872

Ion	Ratio	Lower	Upper
136	100		
137	10.3	8.8	13.2
54	6.7	5.5	8.3
68	4.6	4.2	6.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#22
Acetophenone
Concen: 20.45 ng
RT: 8.481 min Scan# 917
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:105 Resp: 249692

Ion	Ratio	Lower	Upper
105	100		
71	1.2	2.0	3.0#
51	25.5	20.4	30.6
120	22.7	18.1	27.1





#23
Nitrobenzene-d5
Concen: 40.50 ng
RT: 8.811 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 82 Resp: 39173

Ion	Ratio	Lower	Upper
82	100		
128	45.4	37.0	55.4
54	49.1	38.7	58.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#24
Nitrobenzene
Concen: 20.40 ng
RT: 8.858 min Scan# 981
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 77 Resp: 173686

Ion	Ratio	Lower	Upper
77	100		
123	47.6	37.4	56.0
65	14.1	11.3	16.9





#25
Isophorone
Concen: 20.25 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 82 Resp: 33969

Ion	Ratio	Lower	Upper
82	100		
95	7.9	6.3	9.5
138	19.3	15.3	22.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#26
2-Nitrophenol
Concen: 19.73 ng
RT: 9.563 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:139 Resp: 82278

Ion	Ratio	Lower	Upper
139	100		
109	32.8	25.9	38.9
65	45.6	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 19.96 ng
RT: 9.622 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:122 Resp: 14681

Ion	Ratio	Lower	Upper
122	100		
107	124.1	96.2	144.2
121	60.0	47.6	71.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#28
bis(2-Chloroethoxy)methane
Concen: 20.04 ng
RT: 9.852 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 93 Resp: 200872

Ion	Ratio	Lower	Upper
93	100		
95	32.3	25.5	38.3
123	12.4	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 19.90 ng
RT: 10.099 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:162 Resp: 14178

Ion	Ratio	Lower	Upper
162	100		
164	64.0	42.3	82.3
98	36.8	18.4	58.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#30
1,2,4-Trichlorobenzene
Concen: 19.92 ng
RT: 10.299 min Scan# 1226
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:180 Resp: 159998

Ion	Ratio	Lower	Upper
180	100		
182	93.7	79.0	118.4
145	31.5	26.6	39.8





#31
Naphthalene
Concen: 20.25 ng
RT: 10.481 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:128 Resp: 51288

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.8	13.2
127	13.5	10.7	16.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#32
Benzoic acid
Concen: 18.73 ng m
RT: 9.746 min Scan# 1132
Delta R.T. -0.029 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:122 Resp: 71538

Ion	Ratio	Lower	Upper
122	100		
105	127.1	106.9	146.9
77	91.3	75.1	115.1





#33
4-Chloroaniline
Concen: 20.10 ng
RT: 10.605 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:127 Resp: 20543

Ion	Ratio	Lower	Upper
127	100		
129	32.3	25.8	38.6
65	28.6	23.9	35.9
92	19.6	16.1	24.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#34
Hexachlorobutadiene
Concen: 19.87 ng
RT: 10.763 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:225 Resp: 103364

Ion	Ratio	Lower	Upper
225	100		
223	63.0	51.0	76.6
227	65.8	53.1	79.7





#35
Caprolactam
Concen: 19.68 ng
RT: 11.381 min Scan# 14
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:113 Resp: 5072

Ion	Ratio	Lower	Upper
113	100		
55	136.8	119.0	159.0
56	111.1	90.3	130.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#36
4-Chloro-3-methylphenol
Concen: 20.12 ng
RT: 11.734 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:107 Resp: 174840

Ion	Ratio	Lower	Upper
107	100		
144	24.3	20.1	30.1
142	76.6	62.1	93.1





#37
2-Methylnaphthalene
Concen: 19.97 ng
RT: 12.093 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:142 Resp: 32744

Ion	Ratio	Lower	Upper
142	100		
141	89.7	71.2	106.8
115	38.3	30.6	46.0

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#38
1-Methylnaphthalene
Concen: 20.07 ng
RT: 12.316 min Scan# 1569
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:142 Resp: 352091

Ion	Ratio	Lower	Upper
142	100		
141	93.1	74.7	112.1
116	4.4	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.299 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:164 Resp: 32392

Ion	Ratio	Lower	Upper
164	100		
162	103.8	81.8	122.6
160	47.6	36.3	54.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 19.77 ng
RT: 12.469 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:216 Resp: 185982

Ion	Ratio	Lower	Upper
216	100		
214	79.1	62.3	93.5
179	22.9	17.7	26.5
108	20.2	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 19.90 ng
RT: 12.440 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:237 Resp: 11349

Ion	Ratio	Lower	Upper
237	100		
235	61.1	39.9	79.9
272	14.0	0.0	33.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#42
2,4,6-Tribromophenol
Concen: 39.53 ng
RT: 15.810 min Scan# 2163
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:330 Resp: 217092

Ion	Ratio	Lower	Upper
330	100		
332	97.8	78.2	117.4
141	30.5	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 19.95 ng
RT: 12.722 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:196 Resp: 12152

Ion	Ratio	Lower	Upper
196	100		
198	95.8	76.9	115.3
200	30.8	25.2	37.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#44
2,4,5-Trichlorophenol
Concen: 19.95 ng
RT: 12.799 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:196 Resp: 135306

Ion	Ratio	Lower	Upper
196	100		
198	95.2	79.2	118.8
97	53.8	42.3	63.5
132	29.0	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 40.69 ng
RT: 12.904 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:172 Resp: 100615

Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.8	43.2
170	23.8	18.8	28.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#46
1,1'-Biphenyl
Concen: 20.19 ng
RT: 13.116 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:154 Resp: 483352

Ion	Ratio	Lower	Upper
154	100		
153	41.2	21.0	61.0
76	11.4	0.0	31.5





#47
2-Chloronaphthalene
Concen: 20.06 ng
RT: 13.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:162 Resp: 36624

Ion	Ratio	Lower	Upper
162	100		
127	42.3	33.8	50.6
164	32.8	26.0	39.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#48
2-Nitroaniline
Concen: 19.81 ng
RT: 13.375 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 65 Resp: 107062

Ion	Ratio	Lower	Upper
65	100		
92	69.8	58.7	88.1
138	111.5	90.2	135.2





#49
Acenaphthylene
Concen: 20.36 ng
RT: 14.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:152 Resp: 62199

Ion	Ratio	Lower	Upper
152	100		
151	20.5	16.2	24.4
153	12.8	10.2	15.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#50
Dimethylphthalate
Concen: 19.91 ng
RT: 13.751 min Scan# 1813
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:163 Resp: 491623

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	9.8	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 20.23 ng
RT: 13.869 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:165 Resp: 10547

Ion	Ratio	Lower	Upper
165	100		
63	55.6	42.4	63.6
89	58.1	44.8	67.2

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#52
Acenaphthene
Concen: 20.19 ng
RT: 14.357 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:154 Resp: 354519

Ion	Ratio	Lower	Upper
154	100		
153	112.7	91.0	136.4
152	55.1	42.9	64.3





#53
3-Nitroaniline
Concen: 20.01 ng
RT: 14.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:138 Resp: 10652

Ion	Ratio	Lower	Upper
138	100		
108	11.3	9.0	13.4
92	114.5	90.1	135.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#54
2,4-Dinitrophenol
Concen: 19.37 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:184 Resp: 46351

Ion	Ratio	Lower	Upper
184	100		
63	72.5	51.3	76.9
154	72.4	53.5	80.3





#55
Dibenzofuran
Concen: 20.12 ng
RT: 14.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:168 Resp: 56371

Ion	Ratio	Lower	Upper
168	100		
139	41.8	34.2	51.4
169	13.1	10.8	16.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#56
4-Nitrophenol
Concen: 19.02 ng
RT: 14.551 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:139 Resp: 71975

Ion	Ratio	Lower	Upper
139	100		
109	85.7	61.4	101.4
65	101.4	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 20.09 ng
RT: 14.675 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:165 Resp: 14767

Ion	Ratio	Lower	Upper
165	100		
63	42.7	33.8	50.6
89	70.4	55.6	83.4
182	3.1	2.4	3.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#58
Fluorene
Concen: 20.00 ng
RT: 15.357 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:166 Resp: 456792

Ion	Ratio	Lower	Upper
166	100		
165	98.9	78.5	117.7
167	13.8	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 19.59 ng
RT: 14.934 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:232 Resp: 12326

Ion	Ratio	Lower	Upper
232	100		
131	50.2	39.5	59.3
130	2.7	2.2	3.4
166	32.0	25.7	38.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#60
Diethylphthalate
Concen: 19.95 ng
RT: 15.128 min Scan# 2047
Delta R.T. -0.012 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:149 Resp: 497916

Ion	Ratio	Lower	Upper
149	100		
177	21.4	16.7	25.1
150	12.5	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 19.91 ng
RT: 15.351 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:204 Resp: 22679

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.0	39.0
141	68.0	52.9	79.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#62
4-Nitroaniline
Concen: 20.37 ng
RT: 15.393 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:138 Resp: 105617

Ion	Ratio	Lower	Upper
138	100		
92	50.4	29.2	69.2
108	89.0	69.0	109.0





#63
Azobenzene
Concen: 20.57 ng
RT: 15.651 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion: 77 Resp: 43918

Ion	Ratio	Lower	Upper
77	100		
182	25.4	5.7	45.7
105	21.1	1.3	41.3
51	30.3	10.6	50.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.087 min Scan# 2380
Delta R.T. -0.012 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:188 Resp: 644367

Ion	Ratio	Lower	Upper
188	100		
94	8.3	6.5	9.7
80	9.2	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 18.53 ng
RT: 15.451 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:198 Resp: 7713

Ion	Ratio	Lower	Upper
198	100		
51	43.2	21.1	61.1
105	46.8	23.6	63.6

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#66
n-Nitrosodiphenylamine
Concen: 20.42 ng
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:169 Resp: 404085

Ion	Ratio	Lower	Upper
169	100		
168	67.4	53.2	79.8
167	36.5	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 20.07 ng
RT: 16.263 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:248 Resp: 15239

Ion	Ratio	Lower	Upper
248	100		
250	96.8	78.2	117.4
141	74.2	63.0	94.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#68
Hexachlorobenzene
Concen: 19.93 ng
RT: 16.381 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:284 Resp: 192290

Ion	Ratio	Lower	Upper
284	100		
142	31.9	25.9	38.9
249	30.7	24.1	36.1





#69
Atrazine
Concen: 20.19 ng
RT: 16.545 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:200 Resp: 14646

Ion	Ratio	Lower	Upper
200	100		
173	26.7	5.9	45.9
215	47.5	28.3	68.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#70
Pentachlorophenol
Concen: 19.36 ng
RT: 16.740 min Scan# 2321
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:266 Resp: 104916

Ion	Ratio	Lower	Upper
266	100		
268	62.1	52.2	78.4
264	63.5	51.1	76.7





#71
Phenanthrene
Concen: 20.24 ng
RT: 17.128 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:178 Resp: 72492

Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.2
179	15.6	12.2	18.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#72
Anthracene
Concen: 20.42 ng
RT: 17.222 min Scan# 2403
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:178 Resp: 735787

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	14.9	12.2	18.2





#73
Carbazole
Concen: 20.53 ng
RT: 17.510 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:167 Resp: 66973

Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.7	26.5
139	14.4	11.3	16.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#74
Di-n-butylphthalate
Concen: 20.46 ng
RT: 18.092 min Scan# 2551
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:149 Resp: 858522

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.2	10.8
104	5.5	4.3	6.5





#75
Fluoranthene
Concen: 20.22 ng
RT: 19.210 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:202 Resp: 84668

Ion	Ratio	Lower	Upper
202	100		
101	10.4	0.0	30.3
203	17.3	0.0	37.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.522 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:240 Resp: 745224

Ion	Ratio	Lower	Upper
240	100		
120	8.8	7.8	11.6
236	25.5	20.6	31.0





#77
Benzidine
Concen: 19.79 ng
RT: 19.422 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:184 Resp: 40489

Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.2	16.8
183	11.8	9.8	14.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#78
Pyrene
Concen: 19.98 ng
RT: 19.592 min Scan# 2806
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:202 Resp: 892037

Ion	Ratio	Lower	Upper
202	100		
200	21.4	16.9	25.3
203	17.2	13.9	20.9





#79
Terphenyl-d14
Concen: 39.96 ng
RT: 19.810 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:244 Resp: 173687

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.5	8.3
122	9.9	7.4	11.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#80
Butylbenzylphthalate
Concen: 20.14 ng
RT: 20.539 min Scan# 2967
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:149 Resp: 400050

Ion	Ratio	Lower	Upper
149	100		
91	66.6	55.4	83.0
206	21.3	16.2	24.2





#81
Benzo(a)anthracene
Concen: 19.95 ng
RT: 21.498 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:228 Resp: 93361

Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.4	32.0
229	19.7	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#82
3,3'-Dichlorobenzidine
Concen: 20.31 ng
RT: 21.428 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:252 Resp: 353000

Ion	Ratio	Lower	Upper
252	100		
254	64.9	51.8	77.6
126	10.1	7.7	11.5





#83
Chrysene
Concen: 19.91 ng
RT: 21.569 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:228 Resp: 88300

Ion	Ratio	Lower	Upper
228	100		
226	29.5	23.9	35.9
229	20.0	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 20.45 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:149 Resp: 597012

Ion	Ratio	Lower	Upper
149	100		
167	26.6	22.1	33.1
279	5.0	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 19.85 ng
RT: 22.686 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:149 Resp: 94757

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	8.3	6.3	9.5

Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#86
Perylene-d12
Concen: 20.00 ng
RT: 24.792 min Scan# 3690
Delta R.T. -0.012 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:264 Resp: 850264

Ion	Ratio	Lower	Upper
264	100		
260	23.2	18.7	28.1
265	21.2	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 19.81 ng m
RT: 28.486 min Scan# 4
Delta R.T. -0.029 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:276 Resp: 123012

Ion	Ratio	Lower	Upper
276	100		
138	15.2	11.2	16.8
277	20.9	15.7	23.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#88
Benzo(b)fluoranthene
Concen: 20.13 ng
RT: 23.757 min Scan# 3514
Delta R.T. 0.000 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:252 Resp: 979873

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	8.7	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 20.05 ng
RT: 23.822 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:252 Resp: 100534

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.6	26.4
125	8.5	7.0	10.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#90
Benzo(a)pyrene
Concen: 20.23 ng
RT: 24.627 min Scan# 3662
Delta R.T. -0.017 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:252 Resp: 945764

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	9.8	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 19.97 ng
RT: 28.568 min Scan# 4
Delta R.T. -0.029 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:278 Resp: 100828

Ion	Ratio	Lower	Upper
278	100		
139	13.5	10.2	15.4
279	23.6	19.0	28.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC020

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#92
Benzo(g,h,i)perylene
Concen: 19.75 ng
RT: 29.627 min Scan# 4512
Delta R.T. -0.035 min
Lab File: BP024617.D
Acq: 13 May 2025 12:43

Tgt Ion:276 Resp: 991852

Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.4
138	18.4	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024618.D
Acq On : 13 May 2025 13:24
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:06:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	120950	20.00	ng	0.00
21) Naphthalene-d8	10.434	136	540452	20.00	ng	0.00
39) Acenaphthene-d10	14.304	164	362458	20.00	ng	0.00
64) Phenanthrene-d10	17.098	188	717362	20.00	ng	0.00
76) Chrysene-d12	21.521	240	810498	20.00	ng	0.00
86) Perylene-d12	24.804	264	929890	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.287	112	534477	75.58	ng	0.00
7) Phenol-d6	6.852	99	712342	78.93	ng	0.00
23) Nitrobenzene-d5	8.810	82	826567	77.28	ng	0.00
42) 2,4,6-Tribromophenol	15.810	330	467808	76.13	ng	0.00
45) 2-Fluorobiphenyl	12.904	172	2076354	75.05	ng	0.00
79) Terphenyl-d14	19.810	244	3647961	77.18	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.228	88	122181	36.79	ng	100
3) Pyridine	3.622	79	289809	39.36	ng	100
4) n-Nitrosodimethylamine	3.534	42	125087	38.48	ng	100
6) Aniline	7.005	93	463675	39.79	ng	100
8) 2-Chlorophenol	7.240	128	315079	39.25	ng	100
9) Benzaldehyde	6.822	77	231076m	39.62	ng	
10) Phenol	6.881	94	364948	39.18	ng	100
11) bis(2-Chloroethyl)ether	7.099	93	293331	38.32	ng	100
12) 1,3-Dichlorobenzene	7.552	146	348242	37.71	ng	100
13) 1,4-Dichlorobenzene	7.699	146	355533	38.27	ng	100
14) 1,2-Dichlorobenzene	8.010	146	344980	38.11	ng	100
15) Benzyl Alcohol	7.905	79	277892	40.52	ng	100
16) 2,2'-oxybis(1-Chloropr...	8.187	45	354872	38.71	ng	100
17) 2-Methylphenol	8.110	107	265962	39.81	ng	100
18) Hexachloroethane	8.728	117	133224	37.43	ng	100
19) n-Nitroso-di-n-propyla...	8.457	70	253553	40.61	ng	100
20) 3+4-Methylphenols	8.434	107	366564	40.33	ng	100
22) Acetophenone	8.475	105	521603	38.64	ng	100
24) Nitrobenzene	8.852	77	363767	38.65	ng	100
25) Isophorone	9.375	82	727081	39.19	ng	100
26) 2-Nitrophenol	9.563	139	185587	40.24	ng	100
27) 2,4-Dimethylphenol	9.628	122	322227	39.62	ng	100
28) bis(2-Chloroethoxy)met...	9.851	93	430958	38.88	ng	100
29) 2,4-Dichlorophenol	10.093	162	314830	39.95	ng	100
30) 1,2,4-Trichlorobenzene	10.299	180	330659	37.23	ng	100
31) Naphthalene	10.481	128	1059316	37.82	ng	100
32) Benzoic acid	9.775	122	206829m	38.01	ng	
33) 4-Chloroaniline	10.599	127	452150	40.01	ng	100
34) Hexachlorobutadiene	10.763	225	213978	37.19	ng	100
35) Caprolactam	11.387	113	118047	41.43	ng	100
36) 4-Chloro-3-methylphenol	11.734	107	385077	40.07	ng	100
37) 2-Methylnaphthalene	12.093	142	692133	38.17	ng	100
38) 1-Methylnaphthalene	12.316	142	745849	38.44	ng	100
40) 1,2,4,5-Tetrachloroben...	12.469	216	398230	37.84	ng	100
41) Hexachlorocyclopentadiene	12.445	237	258329	40.48	ng	100
43) 2,4,6-Trichlorophenol	12.716	196	268675	39.41	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024618.D
Acq On : 13 May 2025 13:24
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:06:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.798	196	293079	38.62	ng	100
46) 1,1'-Biphenyl	13.116	154	1002344	37.42	ng	100
47) 2-Chloronaphthalene	13.157	162	771952	37.79	ng	100
48) 2-Nitroaniline	13.375	65	242485	40.10	ng	100
49) Acenaphthylene	14.016	152	1307740	38.26	ng	100
50) Dimethylphthalate	13.757	163	1051056	38.04	ng	100
51) 2,6-Dinitrotoluene	13.881	165	226086	38.75	ng	100
52) Acenaphthene	14.363	154	745335	37.93	ng	100
53) 3-Nitroaniline	14.216	138	246561	41.39	ng	100
54) 2,4-Dinitrophenol	14.428	184	124517	37.54	ng	100
55) Dibenzofuran	14.698	168	1183806	37.77	ng	100
56) 4-Nitrophenol	14.551	139	186112	38.41	ng	100
57) 2,4-Dinitrotoluene	14.681	165	329156	40.02	ng	100
58) Fluorene	15.357	166	968487	37.89	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.939	232	274384	38.98	ng	100
60) Diethylphthalate	15.139	149	1046478	37.48	ng	100
61) 4-Chlorophenyl-phenyle...	15.357	204	475327	37.29	ng	100
62) 4-Nitroaniline	15.392	138	239212	41.24	ng	100
63) Azobenzene	15.651	77	906274	37.94	ng	100
65) 4,6-Dinitro-2-methylph...	15.457	198	184593	39.84	ng	100
66) n-Nitrosodiphenylamine	15.581	169	851188	38.63	ng	100
67) 4-Bromophenyl-phenylether	16.263	248	323558	38.28	ng	100
68) Hexachlorobenzene	16.381	284	403916	37.60	ng	100
69) Atrazine	16.557	200	312226	38.67	ng	100
70) Pentachlorophenol	16.745	266	246828	40.91	ng	100
71) Phenanthrene	17.133	178	1507602	37.80	ng	100
72) Anthracene	17.228	178	1542275	38.44	ng	100
73) Carbazole	17.516	167	1436294	39.55	ng	100
74) Di-n-butylphthalate	18.098	149	1807097	38.69	ng	100
75) Fluoranthene	19.216	202	1794046	38.49	ng	100
77) Benzidine	19.416	184	907109	40.77	ng	100
78) Pyrene	19.592	202	1869519	38.50	ng	100
80) Butylbenzylphthalate	20.539	149	854913	39.58	ng	100
81) Benzo(a)anthracene	21.504	228	1957632	38.47	ng	100
82) 3,3'-Dichlorobenzidine	21.427	252	773866	40.94	ng	100
83) Chrysene	21.569	228	1833959	38.01	ng	100
84) Bis(2-ethylhexyl)phtha...	21.439	149	1240956	39.09	ng	100
85) Di-n-octyl phthalate	22.692	149	2074694	39.97	ng	100
87) Indeno(1,2,3-cd)pyrene	28.515	276	2648648	39.00	ng	100
88) Benzo(b)fluoranthene	23.757	252	2050806	38.53	ng	100
89) Benzo(k)fluoranthene	23.827	252	2120721	38.67	ng	100
90) Benzo(a)pyrene	24.645	252	1979017	38.71	ng	100
91) Dibenzo(a,h)anthracene	28.598	278	2170712	39.30	ng	100
92) Benzo(g,h,i)perylene	29.662	276	2132921	38.84	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024618.D
Acq On : 13 May 2025 13:24
Operator : RC/JU
Sample : SSTDICCC040
Misc :
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 13 16:06:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.663 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:152 Resp: 12095

Ion	Ratio	Lower	Upper
152	100		
150	157.4	125.9	188.9
115	62.6	50.1	75.1



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#2
1,4-Dioxane
Concen: 36.79 ng
RT: 3.228 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 88 Resp: 122181

Ion	Ratio	Lower	Upper
88	100		
58	67.5	54.0	81.0
43	26.7	21.4	32.0





#3
Pyridine
Concen: 39.36 ng
RT: 3.622 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 79 Resp: 28980

Ion	Ratio	Lower	Upper
79	100		
52	62.7	50.2	75.2
51	33.5	26.8	40.2



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#4
n-Nitrosodimethylamine
Concen: 38.48 ng
RT: 3.534 min Scan# 76
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 42 Resp: 125087

Ion	Ratio	Lower	Upper
42	100		
74	132.7	106.2	159.2
44	6.8	5.4	8.2





#5
2-Fluorophenol
Concen: 75.58 ng
RT: 5.287 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:112 Resp: 53447

Ion	Ratio	Lower	Upper
112	100		
64	55.1	44.1	66.1
63	29.4	23.5	35.3



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#6
Aniline
Concen: 39.79 ng
RT: 7.005 min Scan# 666
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 93 Resp: 463675

Ion	Ratio	Lower	Upper
93	100		
66	38.8	31.0	46.6
65	19.4	15.5	23.3





#7
Phenol-d6
Concen: 78.93 ng
RT: 6.852 min Scan# 64
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 99 Resp: 71234

Ion	Ratio	Lower	Upper
99	100		
42	18.0	14.4	21.6
71	36.5	29.2	43.8



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#8
2-Chlorophenol
Concen: 39.25 ng
RT: 7.240 min Scan# 706
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:128 Resp: 315079

Ion	Ratio	Lower	Upper
128	100		
130	33.0	13.0	53.0
64	42.3	22.3	62.3





#9
Benzaldehyde
Concen: 39.62 ng m
RT: 6.822 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 77 Resp: 23107

Ion	Ratio	Lower	Upper
77	100		
105	103.3	83.3	123.3
106	98.0	78.0	118.0



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#10
Phenol
Concen: 39.18 ng
RT: 6.881 min Scan# 645
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 94 Resp: 364948

Ion	Ratio	Lower	Upper
94	100		
65	28.7	8.7	48.7
66	39.4	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 38.32 ng
RT: 7.099 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 93 Resp: 29333

Ion	Ratio	Lower	Upper
93	100		
63	70.6	50.6	90.6
95	32.9	12.9	52.9



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#12
1,3-Dichlorobenzene
Concen: 37.71 ng
RT: 7.552 min Scan# 759
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:146 Resp: 348242

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.2
75	31.5	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 38.27 ng
RT: 7.699 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:146 Resp: 35553

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.0	76.6
111	41.3	33.0	49.6



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#14
1,2-Dichlorobenzene
Concen: 38.11 ng
RT: 8.010 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:146 Resp: 344980

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.4
111	44.0	35.2	52.8





#15
Benzyl Alcohol
Concen: 40.52 ng
RT: 7.905 min Scan# 81
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 79 Resp: 27789

Ion	Ratio	Lower	Upper
79	100		
108	76.0	60.8	91.2
77	64.9	51.9	77.9

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.71 ng
RT: 8.187 min Scan# 867
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 45 Resp: 354872

Ion	Ratio	Lower	Upper
45	100		
77	17.7	0.0	37.7
79	14.0	0.0	34.0





#17
2-Methylphenol
Concen: 39.81 ng
RT: 8.110 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:107 Resp: 26596

Ion	Ratio	Lower	Upper
107	100		
108	111.1	88.9	133.3
77	48.8	39.0	58.6
79	48.0	38.4	57.6



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#18
Hexachloroethane
Concen: 37.43 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:117 Resp: 133224

Ion	Ratio	Lower	Upper
117	100		
119	99.2	79.4	119.0
201	103.4	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 40.61 ng
RT: 8.457 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 70 Resp: 25355

Ion	Ratio	Lower	Upper
70	100		
42	53.8	43.0	64.6
101	10.9	8.7	13.1
130	24.2	19.4	29.0



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#20
3+4-Methylphenols
Concen: 40.33 ng
RT: 8.434 min Scan# 909
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:107 Resp: 366564

Ion	Ratio	Lower	Upper
107	100		
108	86.7	66.7	106.7
77	33.1	13.1	53.1
79	28.9	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:136 Resp: 54045

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	6.9	5.5	8.3
68	5.2	4.2	6.2



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#22
Acetophenone
Concen: 38.64 ng
RT: 8.475 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:105 Resp: 521603

Ion	Ratio	Lower	Upper
105	100		
71	2.5	2.0	3.0
51	25.5	20.4	30.6
120	22.6	18.1	27.1





#23
Nitrobenzene-d5
Concen: 77.28 ng
RT: 8.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 82 Resp: 82656

Ion	Ratio	Lower	Upper
82	100		
128	46.2	37.0	55.4
54	48.4	38.7	58.1



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#24
Nitrobenzene
Concen: 38.65 ng
RT: 8.852 min Scan# 980
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 77 Resp: 363767

Ion	Ratio	Lower	Upper
77	100		
123	46.7	37.4	56.0
65	14.1	11.3	16.9





#25
Isophorone
Concen: 39.19 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 82 Resp: 72708

Ion	Ratio	Lower	Upper
82	100		
95	7.9	6.3	9.5
138	19.1	15.3	22.9



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#26
2-Nitrophenol
Concen: 40.24 ng
RT: 9.563 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:139 Resp: 185587

Ion	Ratio	Lower	Upper
139	100		
109	32.4	25.9	38.9
65	47.4	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 39.62 ng
RT: 9.628 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:122 Resp: 32222

Ion	Ratio	Lower	Upper
122	100		
107	120.2	96.2	144.2
121	59.5	47.6	71.4



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#28
bis(2-Chloroethoxy)methane
Concen: 38.88 ng
RT: 9.851 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 93 Resp: 430958

Ion	Ratio	Lower	Upper
93	100		
95	31.9	25.5	38.3
123	12.1	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 39.95 ng
RT: 10.093 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:162 Resp: 31483

Ion	Ratio	Lower	Upper
162	100		
164	62.3	42.3	82.3
98	38.4	18.4	58.4



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#30
1,2,4-Trichlorobenzene
Concen: 37.23 ng
RT: 10.299 min Scan# 1226
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:180 Resp: 330659

Ion	Ratio	Lower	Upper
180	100		
182	98.7	79.0	118.4
145	33.2	26.6	39.8





#31
Naphthalene
Concen: 37.82 ng
RT: 10.481 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:128 Resp: 105931

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.4	10.7	16.1



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#32
Benzoic acid
Concen: 38.01 ng m
RT: 9.775 min Scan# 1137
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:122 Resp: 206829

Ion	Ratio	Lower	Upper
122	100		
105	126.9	106.9	146.9
77	95.1	75.1	115.1





#33
4-Chloroaniline
Concen: 40.01 ng
RT: 10.599 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:127 Resp: 45215

Ion	Ratio	Lower	Upper
127	100		
129	32.2	25.8	38.6
65	29.9	23.9	35.9
92	20.1	16.1	24.1



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#34
Hexachlorobutadiene
Concen: 37.19 ng
RT: 10.763 min Scan# 1305
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:225 Resp: 213978

Ion	Ratio	Lower	Upper
225	100		
223	63.8	51.0	76.6
227	66.4	53.1	79.7





#35
Caprolactam
Concen: 41.43 ng
RT: 11.387 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:113 Resp: 11804

Ion	Ratio	Lower	Upper
113	100		
55	139.0	119.0	159.0
56	110.3	90.3	130.3

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#36
4-Chloro-3-methylphenol
Concen: 40.07 ng
RT: 11.734 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:107 Resp: 385077

Ion	Ratio	Lower	Upper
107	100		
144	25.1	20.1	30.1
142	77.6	62.1	93.1





#37
2-Methylnaphthalene
Concen: 38.17 ng
RT: 12.093 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:142 Resp: 69213

Ion	Ratio	Lower	Upper
142	100		
141	89.0	71.2	106.8
115	38.3	30.6	46.0



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#38
1-Methylnaphthalene
Concen: 38.44 ng
RT: 12.316 min Scan# 1569
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:142 Resp: 745849

Ion	Ratio	Lower	Upper
142	100		
141	93.4	74.7	112.1
116	4.3	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.304 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:164 Resp: 36245

Ion	Ratio	Lower	Upper
164	100		
162	102.2	81.8	122.6
160	45.4	36.3	54.5



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.84 ng
RT: 12.469 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:216 Resp: 398230

Ion	Ratio	Lower	Upper
216	100		
214	77.9	62.3	93.5
179	22.1	17.7	26.5
108	20.0	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 40.48 ng
RT: 12.445 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:237 Resp: 25832

Ion	Ratio	Lower	Upper
237	100		
235	59.9	39.9	79.9
272	13.2	0.0	33.2



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#42
2,4,6-Tribromophenol
Concen: 76.13 ng
RT: 15.810 min Scan# 2163
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:330 Resp: 467808

Ion	Ratio	Lower	Upper
330	100		
332	97.8	78.2	117.4
141	30.4	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 39.41 ng
RT: 12.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:196 Resp: 26867

Ion	Ratio	Lower	Upper
196	100		
198	96.1	76.9	115.3
200	31.5	25.2	37.8



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#44
2,4,5-Trichlorophenol
Concen: 38.62 ng
RT: 12.798 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:196 Resp: 293079

Ion	Ratio	Lower	Upper
196	100		
198	99.0	79.2	118.8
97	52.9	42.3	63.5
132	29.1	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 75.05 ng
RT: 12.904 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:172 Resp: 207635

Ion	Ratio	Lower	Upper
172	100		
171	36.0	28.8	43.2
170	23.5	18.8	28.2



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#46
1,1'-Biphenyl
Concen: 37.42 ng
RT: 13.116 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:154 Resp: 1002344

Ion	Ratio	Lower	Upper
154	100		
153	41.0	21.0	61.0
76	11.5	0.0	31.5





#47
2-Chloronaphthalene
Concen: 37.79 ng
RT: 13.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:162 Resp: 77195

Ion	Ratio	Lower	Upper
162	100		
127	42.2	33.8	50.6
164	32.5	26.0	39.0



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#48
2-Nitroaniline
Concen: 40.10 ng
RT: 13.375 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 65 Resp: 242485

Ion	Ratio	Lower	Upper
65	100		
92	73.4	58.7	88.1
138	112.7	90.2	135.2





#49
Acenaphthylene
Concen: 38.26 ng
RT: 14.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:152 Resp: 130774

Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.2	24.4
153	12.8	10.2	15.4



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#50
Dimethylphthalate
Concen: 38.04 ng
RT: 13.757 min Scan# 1814
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:163 Resp: 1051056

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.1	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 38.75 ng
RT: 13.881 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:165 Resp: 22608

Ion	Ratio	Lower	Upper
165	100		
63	53.0	42.4	63.6
89	56.0	44.8	67.2



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#52
Acenaphthene
Concen: 37.93 ng
RT: 14.363 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:154 Resp: 745335

Ion	Ratio	Lower	Upper
154	100		
153	113.7	91.0	136.4
152	53.6	42.9	64.3





#53
3-Nitroaniline
Concen: 41.39 ng
RT: 14.216 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:138 Resp: 24656

Ion	Ratio	Lower	Upper
138	100		
108	11.2	9.0	13.4
92	112.6	90.1	135.1

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#54
2,4-Dinitrophenol
Concen: 37.54 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:184 Resp: 124517

Ion	Ratio	Lower	Upper
184	100		
63	64.1	51.3	76.9
154	66.9	53.5	80.3





#55
Dibenzofuran
Concen: 37.77 ng
RT: 14.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:168 Resp: 118380

Ion	Ratio	Lower	Upper
168	100		
139	42.8	34.2	51.4
169	13.5	10.8	16.2



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#56
4-Nitrophenol
Concen: 38.41 ng
RT: 14.551 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:139 Resp: 186112

Ion	Ratio	Lower	Upper
139	100		
109	81.4	61.4	101.4
65	92.2	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 40.02 ng
RT: 14.681 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:165 Resp: 32915

Ion	Ratio	Lower	Upper
165	100		
63	42.2	33.8	50.6
89	69.5	55.6	83.4
182	3.0	2.4	3.6



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#58
Fluorene
Concen: 37.89 ng
RT: 15.357 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:166 Resp: 968487

Ion	Ratio	Lower	Upper
166	100		
165	98.1	78.5	117.7
167	13.4	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 38.98 ng
RT: 14.939 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:232 Resp: 27438

Ion	Ratio	Lower	Upper
232	100		
131	49.4	39.5	59.3
130	2.8	2.2	3.4
166	32.1	25.7	38.5



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#60
Diethylphthalate
Concen: 37.48 ng
RT: 15.139 min Scan# 2049
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:149 Resp: 1046478

Ion	Ratio	Lower	Upper
149	100		
177	20.9	16.7	25.1
150	12.7	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 37.29 ng
RT: 15.357 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:204 Resp: 47532

Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.0	39.0
141	66.1	52.9	79.3



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#62
4-Nitroaniline
Concen: 41.24 ng
RT: 15.392 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:138 Resp: 239212

Ion	Ratio	Lower	Upper
138	100		
92	49.2	29.2	69.2
108	89.0	69.0	109.0





#63
Azobenzene
Concen: 37.94 ng
RT: 15.651 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion: 77 Resp: 90627

Ion	Ratio	Lower	Upper
77	100		
182	25.7	5.7	45.7
105	21.3	1.3	41.3
51	30.6	10.6	50.6



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.098 min Scan# 2382
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:188 Resp: 717362

Ion	Ratio	Lower	Upper
188	100		
94	8.1	6.5	9.7
80	9.1	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 39.84 ng
RT: 15.457 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:198 Resp: 18459

Ion	Ratio	Lower	Upper
198	100		
51	41.1	21.1	61.1
105	43.6	23.6	63.6

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#66
n-Nitrosodiphenylamine
Concen: 38.63 ng
RT: 15.581 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:169 Resp: 851188

Ion	Ratio	Lower	Upper
169	100		
168	66.5	53.2	79.8
167	35.8	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 38.28 ng
RT: 16.263 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:248 Resp: 32355

Ion	Ratio	Lower	Upper
248	100		
250	97.8	78.2	117.4
141	78.8	63.0	94.6



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#68
Hexachlorobenzene
Concen: 37.60 ng
RT: 16.381 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:284 Resp: 403916

Ion	Ratio	Lower	Upper
284	100		
142	32.4	25.9	38.9
249	30.1	24.1	36.1





#69
Atrazine
Concen: 38.67 ng
RT: 16.557 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:200 Resp: 31222

Ion	Ratio	Lower	Upper
200	100		
173	25.9	5.9	45.9
215	48.3	28.3	68.3



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#70
Pentachlorophenol
Concen: 40.91 ng
RT: 16.745 min Scan# 2322
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:266 Resp: 246828

Ion	Ratio	Lower	Upper
266	100		
268	65.3	52.2	78.4
264	63.9	51.1	76.7





#71
Phenanthrene
Concen: 37.80 ng
RT: 17.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:178 Resp: 150760

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.4	23.2
179	15.3	12.2	18.4



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#72
Anthracene
Concen: 38.44 ng
RT: 17.228 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:178 Resp: 1542275

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.2	12.2	18.2





#73
Carbazole
Concen: 39.55 ng
RT: 17.516 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:167 Resp: 143629

Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.7	26.5
139	14.1	11.3	16.9



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#74
Di-n-butylphthalate
Concen: 38.69 ng
RT: 18.098 min Scan# 2552
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:149 Resp: 1807097

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.2	10.8
104	5.4	4.3	6.5





#75
Fluoranthene
Concen: 38.49 ng
RT: 19.216 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:202 Resp: 179404

Ion	Ratio	Lower	Upper
202	100		
101	10.3	0.0	30.3
203	17.5	0.0	37.5



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.521 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:240 Resp: 810498

Ion	Ratio	Lower	Upper
240	100		
120	9.7	7.8	11.6
236	25.8	20.6	31.0





#77
Benzidine
Concen: 40.77 ng
RT: 19.416 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:184 Resp: 90710

Ion	Ratio	Lower	Upper
184	100		
185	14.0	11.2	16.8
183	12.3	9.8	14.8



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#78
Pyrene
Concen: 38.50 ng
RT: 19.592 min Scan# 2806
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:202 Resp: 1869519

Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.9	25.3
203	17.4	13.9	20.9





#79
Terphenyl-d14
Concen: 77.18 ng
RT: 19.810 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:244 Resp: 364796

Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.5	8.3
122	9.2	7.4	11.0



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#80
Butylbenzylphthalate
Concen: 39.58 ng
RT: 20.539 min Scan# 2967
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:149 Resp: 854913

Ion	Ratio	Lower	Upper
149	100		
91	69.2	55.4	83.0
206	20.2	16.2	24.2





#81
Benzo(a)anthracene
Concen: 38.47 ng
RT: 21.504 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:228 Resp: 195763

Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.4	32.0
229	19.4	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#82
3,3'-Dichlorobenzidine
Concen: 40.94 ng
RT: 21.427 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:252 Resp: 773866

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.8	77.6
126	9.6	7.7	11.5





#83
Chrysene
Concen: 38.01 ng
RT: 21.569 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:228 Resp: 183395

Ion	Ratio	Lower	Upper
228	100		
226	29.9	23.9	35.9
229	19.4	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 39.09 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:149 Resp: 1240956

Ion	Ratio	Lower	Upper
149	100		
167	27.6	22.1	33.1
279	5.7	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 39.97 ng
RT: 22.692 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:149 Resp: 207469

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.9	6.3	9.5

Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#86
Perylene-d12
Concen: 20.00 ng
RT: 24.804 min Scan# 3692
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:264 Resp: 929890

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.7	28.1
265	21.3	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 39.00 ng
RT: 28.515 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:276 Resp: 264864

Ion	Ratio	Lower	Upper
276	100		
138	14.0	11.2	16.8
277	19.6	15.7	23.5



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#88
Benzo(b)fluoranthene
Concen: 38.53 ng
RT: 23.757 min Scan# 3514
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:252 Resp: 2050806

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	9.4	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 38.67 ng
RT: 23.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:252 Resp: 212072

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	8.7	7.0	10.4



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#90
Benzo(a)pyrene
Concen: 38.71 ng
RT: 24.645 min Scan# 3665
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:252 Resp: 1979017

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	10.1	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 39.30 ng
RT: 28.598 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:278 Resp: 217071

Ion	Ratio	Lower	Upper
278	100		
139	12.8	10.2	15.4
279	23.7	19.0	28.4



Instrument :
BNA_P
ClientSampleId :
SSTDICCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#92
Benzo(g,h,i)perylene
Concen: 38.84 ng
RT: 29.662 min Scan# 4518
Delta R.T. 0.000 min
Lab File: BP024618.D
Acq: 13 May 2025 13:24

Tgt Ion:276 Resp: 2132921

Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.0	28.4
138	18.0	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024619.D
Acq On : 13 May 2025 14:05
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: May 13 16:07:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.663	152	111395	20.00	ng		0.00
21) Naphthalene-d8	10.434	136	489806	20.00	ng		0.00
39) Acenaphthene-d10	14.292	164	318760	20.00	ng		-0.01
64) Phenanthrene-d10	17.086	188	628781	20.00	ng		-0.01
76) Chrysene-d12	21.522	240	717340	20.00	ng		0.00
86) Perylene-d12	24.810	264	811421	20.00	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.287	112	708626	108.80	ng		0.00
7) Phenol-d6	6.852	99	930484	111.94	ng		0.00
23) Nitrobenzene-d5	8.810	82	1049445	108.26	ng		0.00
42) 2,4,6-Tribromophenol	15.810	330	587523	108.72	ng		0.00
45) 2-Fluorobiphenyl	12.904	172	2553765	104.96	ng		0.00
79) Terphenyl-d14	19.822	244	4398872	105.15	ng		0.01
Target Compounds							Qvalue
2) 1,4-Dioxane	3.228	88	156571	51.20	ng		98
3) Pyridine	3.622	79	379522	55.97	ng		99
4) n-Nitrosodimethylamine	3.534	42	161654	54.00	ng		99
6) Aniline	7.005	93	593141	55.27	ng		99
8) 2-Chlorophenol	7.240	128	408016	55.19	ng		99
9) Benzaldehyde	6.822	77	290405	54.06	ng		100
10) Phenol	6.881	94	475413	55.41	ng		99
11) bis(2-Chloroethyl)ether	7.093	93	377787	53.59	ng		100
12) 1,3-Dichlorobenzene	7.552	146	451750	53.11	ng		99
13) 1,4-Dichlorobenzene	7.699	146	455796	53.28	ng		99
14) 1,2-Dichlorobenzene	8.010	146	439707	52.74	ng		100
15) Benzyl Alcohol	7.910	79	363689	57.58	ng		98
16) 2,2'-oxybis(1-Chloropr...	8.181	45	450215	53.33	ng		98
17) 2-Methylphenol	8.110	107	347065	56.40	ng		98
18) Hexachloroethane	8.728	117	173580	52.95	ng		98
19) n-Nitroso-di-n-propyla...	8.463	70	318058	55.32	ng		99
20) 3+4-Methylphenols	8.440	107	474048	56.64	ng		98
22) Acetophenone	8.481	105	653878	53.44	ng	#	98
24) Nitrobenzene	8.852	77	456929	53.56	ng		99
25) Isophorone	9.375	82	921892	54.82	ng		99
26) 2-Nitrophenol	9.557	139	239844	57.38	ng		98
27) 2,4-Dimethylphenol	9.628	122	402585	54.62	ng		99
28) bis(2-Chloroethoxy)met...	9.852	93	542581	54.01	ng		100
29) 2,4-Dichlorophenol	10.099	162	400983	56.14	ng		98
30) 1,2,4-Trichlorobenzene	10.299	180	424561	52.75	ng		97
31) Naphthalene	10.481	128	1336265	52.65	ng		100
32) Benzoic acid	9.787	122	263657	50.71	ng		99
33) 4-Chloroaniline	10.599	127	569812	55.63	ng		99
34) Hexachlorobutadiene	10.769	225	271578	52.09	ng		99
35) Caprolactam	11.393	113	148586	57.53	ng		98
36) 4-Chloro-3-methylphenol	11.734	107	478787	54.97	ng		99
37) 2-Methylnaphthalene	12.093	142	873274	53.13	ng		99
38) 1-Methylnaphthalene	12.316	142	919827	52.31	ng		98
40) 1,2,4,5-Tetrachloroben...	12.469	216	497848	53.78	ng		100
41) Hexachlorocyclopentadiene	12.446	237	334752	59.64	ng		96
43) 2,4,6-Trichlorophenol	12.716	196	341059	56.89	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024619.D
Acq On : 13 May 2025 14:05
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: May 13 16:07:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.793	196	375330	56.24	ng	98
46) 1,1'-Biphenyl	13.116	154	1263640	53.64	ng	100
47) 2-Chloronaphthalene	13.157	162	952857	53.04	ng	99
48) 2-Nitroaniline	13.369	65	303475	57.06	ng	99
49) Acenaphthylene	14.010	152	1616938	53.79	ng	100
50) Dimethylphthalate	13.751	163	1283491	52.83	ng	99
51) 2,6-Dinitrotoluene	13.875	165	281728	54.91	ng	99
52) Acenaphthene	14.357	154	923238	53.42	ng	100
53) 3-Nitroaniline	14.210	138	302341	57.71	ng	99
54) 2,4-Dinitrophenol	14.422	184	159888	51.86	ng	98
55) Dibenzofuran	14.698	168	1447333	52.50	ng	99
56) 4-Nitrophenol	14.545	139	238294	53.99	ng	99
57) 2,4-Dinitrotoluene	14.669	165	404684	55.94	ng	98
58) Fluorene	15.357	166	1190632	52.97	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.934	232	341204	55.11	ng	100
60) Diethylphthalate	15.134	149	1305791	53.18	ng	99
61) 4-Chlorophenyl-phenyle...	15.351	204	590421	52.67	ng	97
62) 4-Nitroaniline	15.392	138	298681	58.55	ng	99
63) Azobenzene	15.651	77	1121556	53.39	ng	98
65) 4,6-Dinitro-2-methylph...	15.451	198	234095	57.64	ng	98
66) n-Nitrosodiphenylamine	15.569	169	1048396	54.28	ng	99
67) 4-Bromophenyl-phenylether	16.263	248	396385	53.50	ng	99
68) Hexachlorobenzene	16.381	284	504987	53.64	ng	99
69) Atrazine	16.551	200	389749	55.07	ng	100
70) Pentachlorophenol	16.739	266	310477	58.71	ng	98
71) Phenanthrene	17.128	178	1838891	52.61	ng	100
72) Anthracene	17.228	178	1894573	53.87	ng	100
73) Carbazole	17.510	167	1750681	55.00	ng	100
74) Di-n-butylphthalate	18.104	149	2275163	55.58	ng	100
75) Fluoranthene	19.222	202	2199818	53.84	ng	100
77) Benzidine	19.422	184	1178505	59.85	ng	99
78) Pyrene	19.604	202	2282702	53.11	ng	100
80) Butylbenzylphthalate	20.551	149	1083395	56.68	ng	97
81) Benzo(a)anthracene	21.504	228	2396398	53.20	ng	99
82) 3,3'-Dichlorobenzidine	21.439	252	966906	57.80	ng	100
83) Chrysene	21.574	228	2250800	52.71	ng	100
84) Bis(2-ethylhexyl)phtha...	21.451	149	1585070	56.42	ng	100
85) Di-n-octyl phthalate	22.698	149	2614856	56.91	ng	100
87) Indeno(1,2,3-cd)pyrene	28.515	276	3262830	55.06	ng	99
88) Benzo(b)fluoranthene	23.763	252	2612438	56.25	ng	99
89) Benzo(k)fluoranthene	23.827	252	2552663	53.34	ng	99
90) Benzo(a)pyrene	24.645	252	2475976	55.51	ng	100
91) Dibenzo(a,h)anthracene	28.586	278	2679367	55.60	ng	99
92) Benzo(g,h,i)perylene	29.656	276	2625625	54.79	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024619.D
Acq On : 13 May 2025 14:05
Operator : RC/JU
Sample : SSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Quant Time: May 13 16:07:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.663 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 111395

Ion	Ratio	Lower	Upper
152	100		
150	156.2	125.9	188.9
115	63.6	50.1	75.1





#2
1,4-Dioxane
Concen: 51.20 ng
RT: 3.228 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 88 Resp: 156571

Ion	Ratio	Lower	Upper
88	100		
58	70.0	54.0	81.0
43	26.5	21.4	32.0





#3
Pyridine
Concen: 55.97 ng
RT: 3.622 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 379522

Ion	Ratio	Lower	Upper
79	100		
52	63.2	50.2	75.2
51	34.8	26.8	40.2





#4
n-Nitrosodimethylamine
Concen: 54.00 ng
RT: 3.534 min Scan# 76
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 42 Resp: 161654

Ion	Ratio	Lower	Upper
42	100		
74	132.2	106.2	159.2
44	6.0	5.4	8.2





#5
2-Fluorophenol
Concen: 108.80 ng
RT: 5.287 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:112 Resp: 708626

Ion	Ratio	Lower	Upper
112	100		
64	54.4	44.1	66.1
63	29.4	23.5	35.3





#6
Aniline
Concen: 55.27 ng
RT: 7.005 min Scan# 666
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 93 Resp: 593141

Ion	Ratio	Lower	Upper
93	100		
66	39.2	31.0	46.6
65	19.7	15.5	23.3





#7
Phenol-d6
Concen: 111.94 ng
RT: 6.852 min Scan# 64
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 99 Resp: 930484

Ion	Ratio	Lower	Upper
99	100		
42	17.7	14.4	21.6
71	36.2	29.2	43.8





#8
2-Chlorophenol
Concen: 55.19 ng
RT: 7.240 min Scan# 706
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:128 Resp: 408016

Ion	Ratio	Lower	Upper
128	100		
130	32.4	13.0	53.0
64	42.8	22.3	62.3





#9
Benzaldehyde
Concen: 54.06 ng
RT: 6.822 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 290405

Ion	Ratio	Lower	Upper
77	100		
105	103.0	83.3	123.3
106	98.4	78.0	118.0





#10
Phenol
Concen: 55.41 ng
RT: 6.881 min Scan# 645
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 94 Resp: 475413

Ion	Ratio	Lower	Upper
94	100		
65	28.3	8.7	48.7
66	39.9	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 53.59 ng
RT: 7.093 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 93 Resp: 377787

Ion	Ratio	Lower	Upper
93	100		
63	70.5	50.6	90.6
95	32.5	12.9	52.9





#12
1,3-Dichlorobenzene
Concen: 53.11 ng
RT: 7.552 min Scan# 759
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:146 Resp: 451750

Ion	Ratio	Lower	Upper
146	100		
148	63.4	51.4	77.2
75	31.9	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 53.28 ng
RT: 7.699 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:146 Resp: 455796

Ion	Ratio	Lower	Upper
146	100		
148	64.5	51.0	76.6
111	42.2	33.0	49.6





#14
1,2-Dichlorobenzene
Concen: 52.74 ng
RT: 8.010 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:146 Resp: 439707

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.0	76.4
111	43.9	35.2	52.8





#15
Benzyl Alcohol
Concen: 57.58 ng
RT: 7.910 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 79 Resp: 363689

Ion	Ratio	Lower	Upper
79	100		
108	78.2	60.8	91.2
77	65.6	51.9	77.9





#16
2,2'-oxybis(1-Chloropropane)
Concen: 53.33 ng
RT: 8.181 min Scan# 866
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 45 Resp: 450215

Ion	Ratio	Lower	Upper
45	100		
77	17.1	0.0	37.7
79	13.2	0.0	34.0





#17
2-Methylphenol
Concen: 56.40 ng
RT: 8.110 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:107 Resp: 347065

Ion	Ratio	Lower	Upper
107	100		
108	109.2	88.9	133.3
77	47.2	39.0	58.6
79	46.6	38.4	57.6





#18
Hexachloroethane
Concen: 52.95 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:117 Resp: 173580

Ion	Ratio	Lower	Upper
117	100		
119	97.5	79.4	119.0
201	101.3	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 55.32 ng
RT: 8.463 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 70 Resp: 318058

Ion	Ratio	Lower	Upper
70	100		
42	54.3	43.0	64.6
101	11.5	8.7	13.1
130	25.1	19.4	29.0





#20
3+4-Methylphenols
Concen: 56.64 ng
RT: 8.440 min Scan# 910
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:107 Resp: 474048

Ion	Ratio	Lower	Upper
107	100		
108	89.3	66.7	106.7
77	33.4	13.1	53.1
79	30.7	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:136 Resp: 489806

Ion	Ratio	Lower	Upper
136	100		
137	10.7	8.8	13.2
54	7.0	5.5	8.3
68	5.0	4.2	6.2





#22
Acetophenone
Concen: 53.44 ng
RT: 8.481 min Scan# 917
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:105 Resp: 653878

Ion	Ratio	Lower	Upper
105	100		
71	1.9	2.0	3.0#
51	24.6	20.4	30.6
120	23.2	18.1	27.1





#23
Nitrobenzene-d5
Concen: 108.26 ng
RT: 8.810 min Scan# 9
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 1049445

Ion	Ratio	Lower	Upper
82	100		
128	46.1	37.0	55.4
54	48.0	38.7	58.1





#24
Nitrobenzene
Concen: 53.56 ng
RT: 8.852 min Scan# 980
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 77 Resp: 456929

Ion	Ratio	Lower	Upper
77	100		
123	47.4	37.4	56.0
65	13.7	11.3	16.9





#25
Isophorone
Concen: 54.82 ng
RT: 9.375 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 921892

Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.3	9.5
138	18.3	15.3	22.9





#26
2-Nitrophenol
Concen: 57.38 ng
RT: 9.557 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:139 Resp: 239844

Ion	Ratio	Lower	Upper
139	100		
109	33.6	25.9	38.9
65	49.0	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 54.62 ng
RT: 9.628 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:122 Resp: 402585

Ion	Ratio	Lower	Upper
122	100		
107	121.6	96.2	144.2
121	60.2	47.6	71.4





#28
bis(2-Chloroethoxy)methane
Concen: 54.01 ng
RT: 9.852 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 93 Resp: 542581

Ion	Ratio	Lower	Upper
93	100		
95	31.9	25.5	38.3
123	12.2	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 56.14 ng
RT: 10.099 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 400983

Ion	Ratio	Lower	Upper
162	100		
164	64.1	42.3	82.3
98	37.0	18.4	58.4





#30
1,2,4-Trichlorobenzene
Concen: 52.75 ng
RT: 10.299 min Scan# 1226
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:180 Resp: 424561

Ion	Ratio	Lower	Upper
180	100		
182	95.5	79.0	118.4
145	32.6	26.6	39.8





#31
Naphthalene
Concen: 52.65 ng
RT: 10.481 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:128 Resp: 1336265

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.8	13.2
127	13.4	10.7	16.1





#32
Benzoic acid
Concen: 50.71 ng
RT: 9.787 min Scan# 1139
Delta R.T. 0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:122 Resp: 263657

Ion	Ratio	Lower	Upper
122	100		
105	127.2	106.9	146.9
77	93.3	75.1	115.1





#33
4-Chloroaniline
Concen: 55.63 ng
RT: 10.599 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:127 Resp: 569812

Ion	Ratio	Lower	Upper
127	100		
129	32.6	25.8	38.6
65	29.6	23.9	35.9
92	19.7	16.1	24.1





#34
Hexachlorobutadiene
Concen: 52.09 ng
RT: 10.769 min Scan# 1306
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:225 Resp: 271578

Ion	Ratio	Lower	Upper
225	100		
223	64.4	51.0	76.6
227	64.9	53.1	79.7





#35
Caprolactam
Concen: 57.53 ng
RT: 11.393 min Scan# 14
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:113 Resp: 148586

Ion	Ratio	Lower	Upper
113	100		
55	136.6	119.0	159.0
56	112.0	90.3	130.3





#36
4-Chloro-3-methylphenol
Concen: 54.97 ng
RT: 11.734 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:107 Resp: 478787

Ion	Ratio	Lower	Upper
107	100		
144	25.5	20.1	30.1
142	78.6	62.1	93.1





#37
2-Methylnaphthalene
Concen: 53.13 ng
RT: 12.093 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:142 Resp: 873274

Ion	Ratio	Lower	Upper
142	100		
141	89.6	71.2	106.8
115	37.9	30.6	46.0





#38
1-Methylnaphthalene
Concen: 52.31 ng
RT: 12.316 min Scan# 1569
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:142 Resp: 919827

Ion	Ratio	Lower	Upper
142	100		
141	91.9	74.7	112.1
116	4.1	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.292 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:164 Resp: 318760

Ion	Ratio	Lower	Upper
164	100		
162	101.9	81.8	122.6
160	45.1	36.3	54.5





#40
1,2,4,5-Tetrachlorobenzene
Concen: 53.78 ng
RT: 12.469 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:216 Resp: 497848

Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.3	93.5
179	21.9	17.7	26.5
108	19.7	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 59.64 ng
RT: 12.446 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:237 Resp: 334752

Ion	Ratio	Lower	Upper
237	100		
235	63.6	39.9	79.9
272	13.8	0.0	33.2





#42
2,4,6-Tribromophenol
Concen: 108.72 ng
RT: 15.810 min Scan# 2163
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:330 Resp: 587523

Ion	Ratio	Lower	Upper
330	100		
332	96.9	78.2	117.4
141	29.7	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 56.89 ng
RT: 12.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:196 Resp: 341059

Ion	Ratio	Lower	Upper
196	100		
198	94.5	76.9	115.3
200	30.9	25.2	37.8





#44
2,4,5-Trichlorophenol
Concen: 56.24 ng
RT: 12.793 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:196 Resp: 375330

Ion	Ratio	Lower	Upper
196	100		
198	96.5	79.2	118.8
97	54.0	42.3	63.5
132	30.8	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 104.96 ng
RT: 12.904 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:172 Resp: 2553765

Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.8	43.2
170	23.7	18.8	28.2





#46
1,1'-Biphenyl
Concen: 53.64 ng
RT: 13.116 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:154 Resp: 1263640

Ion	Ratio	Lower	Upper
154	100		
153	41.2	21.0	61.0
76	11.6	0.0	31.5





#47
2-Chloronaphthalene
Concen: 53.04 ng
RT: 13.157 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:162 Resp: 952857

Ion	Ratio	Lower	Upper
162	100		
127	41.4	33.8	50.6
164	32.3	26.0	39.0





#48
2-Nitroaniline
Concen: 57.06 ng
RT: 13.369 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion: 65 Resp: 303475

Ion	Ratio	Lower	Upper
65	100		
92	72.5	58.7	88.1
138	114.3	90.2	135.2





#49
Acenaphthylene
Concen: 53.79 ng
RT: 14.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:152 Resp: 1616938

Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.2	24.4
153	12.9	10.2	15.4





#50
Dimethylphthalate
Concen: 52.83 ng
RT: 13.751 min Scan# 1813
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:163 Resp: 1283491

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	10.4	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 54.91 ng
RT: 13.875 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 281728

Ion	Ratio	Lower	Upper
165	100		
63	53.2	42.4	63.6
89	54.5	44.8	67.2





#52
Acenaphthene
Concen: 53.42 ng
RT: 14.357 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:154 Resp: 923238

Ion	Ratio	Lower	Upper
154	100		
153	113.5	91.0	136.4
152	53.6	42.9	64.3





#53
3-Nitroaniline
Concen: 57.71 ng
RT: 14.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:138 Resp: 302341

Ion	Ratio	Lower	Upper
138	100		
108	11.2	9.0	13.4
92	111.5	90.1	135.1





#54
2,4-Dinitrophenol
Concen: 51.86 ng
RT: 14.422 min Scan# 1927
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:184 Resp: 159888

Ion	Ratio	Lower	Upper
184	100		
63	63.3	51.3	76.9
154	69.3	53.5	80.3





#55
Dibenzofuran
Concen: 52.50 ng
RT: 14.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:168 Resp: 1447333

Ion	Ratio	Lower	Upper
168	100		
139	42.3	34.2	51.4
169	13.2	10.8	16.2





#56
4-Nitrophenol
Concen: 53.99 ng
RT: 14.545 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:139 Resp: 238294

Ion	Ratio	Lower	Upper
139	100		
109	82.0	61.4	101.4
65	90.8	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 55.94 ng
RT: 14.669 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:165 Resp: 404684

Ion	Ratio	Lower	Upper
165	100		
63	43.0	33.8	50.6
89	71.7	55.6	83.4
182	3.0	2.4	3.6





#58
Fluorene
Concen: 52.97 ng
RT: 15.357 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:166 Resp: 1190632

Ion	Ratio	Lower	Upper
166	100		
165	97.5	78.5	117.7
167	13.4	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 55.11 ng
RT: 14.934 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:232 Resp: 341204

Ion	Ratio	Lower	Upper
232	100		
131	49.3	39.5	59.3
130	2.6	2.2	3.4
166	31.8	25.7	38.5





#60
Diethylphthalate
Concen: 53.18 ng
RT: 15.134 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:149 Resp: 1305791

Ion	Ratio	Lower	Upper
149	100		
177	21.3	16.7	25.1
150	12.3	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 52.67 ng
RT: 15.351 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:204 Resp: 590421

Ion	Ratio	Lower	Upper
204	100		
206	33.4	26.0	39.0
141	68.7	52.9	79.3





#62
4-Nitroaniline
Concen: 58.55 ng
RT: 15.392 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:138 Resp: 298681

Ion	Ratio	Lower	Upper
138	100		
92	50.3	29.2	69.2
108	89.6	69.0	109.0





#63
Azobenzene
Concen: 53.39 ng
RT: 15.651 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion: 77 Resp: 1121556

Ion	Ratio	Lower	Upper
77	100		
182	27.6	5.7	45.7
105	21.6	1.3	41.3
51	29.5	10.6	50.6





#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.086 min Scan# 2380
Delta R.T. -0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:188 Resp: 628781

Ion	Ratio	Lower	Upper
188	100		
94	9.2	6.5	9.7
80	10.3	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 57.64 ng
RT: 15.451 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:198 Resp: 234095

Ion	Ratio	Lower	Upper
198	100		
51	40.5	21.1	61.1
105	42.1	23.6	63.6





#66
n-Nitrosodiphenylamine
Concen: 54.28 ng
RT: 15.569 min Scan# 2122
Delta R.T. -0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:169 Resp: 1048396

Ion	Ratio	Lower	Upper
169	100		
168	67.3	53.2	79.8
167	36.1	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 53.50 ng
RT: 16.263 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:248 Resp: 396385

Ion	Ratio	Lower	Upper
248	100		
250	99.0	78.2	117.4
141	78.2	63.0	94.6





#68
Hexachlorobenzene
Concen: 53.64 ng
RT: 16.381 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:284 Resp: 504987

Ion	Ratio	Lower	Upper
284	100		
142	32.1	25.9	38.9
249	30.4	24.1	36.1





#69
Atrazine
Concen: 55.07 ng
RT: 16.551 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:200 Resp: 389749

Ion	Ratio	Lower	Upper
200	100		
173	26.0	5.9	45.9
215	47.9	28.3	68.3





#70
Pentachlorophenol
Concen: 58.71 ng
RT: 16.739 min Scan# 2321
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:266 Resp: 310477

Ion	Ratio	Lower	Upper
266	100		
268	63.5	52.2	78.4
264	62.6	51.1	76.7





#71
Phenanthrene
Concen: 52.61 ng
RT: 17.128 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:178 Resp: 1838891

Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.4	23.2
179	15.5	12.2	18.4





#72
Anthracene
Concen: 53.87 ng
RT: 17.228 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:178 Resp: 1894573

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.1	12.2	18.2





#73
Carbazole
Concen: 55.00 ng
RT: 17.510 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:167 Resp: 1750681

Ion	Ratio	Lower	Upper
167	100		
166	22.1	17.7	26.5
139	14.0	11.3	16.9





#74
Di-n-butylphthalate
Concen: 55.58 ng
RT: 18.104 min Scan# 2553
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:149 Resp: 2275163

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.2	10.8
104	5.4	4.3	6.5





#75
Fluoranthene
Concen: 53.84 ng
RT: 19.222 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:202 Resp: 2199818

Ion	Ratio	Lower	Upper
202	100		
101	10.5	0.0	30.3
203	17.4	0.0	37.5





#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.522 min Scan# 3134
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:240 Resp: 717340

Ion	Ratio	Lower	Upper
240	100		
120	10.4	7.8	11.6
236	25.1	20.6	31.0





#77
Benzidine
Concen: 59.85 ng
RT: 19.422 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:184 Resp: 1178505

Ion	Ratio	Lower	Upper
184	100		
185	14.7	11.2	16.8
183	12.2	9.8	14.8





#78
Pyrene
Concen: 53.11 ng
RT: 19.604 min Scan# 2808
Delta R.T. 0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:202 Resp: 2282702

Ion	Ratio	Lower	Upper
202	100		
200	21.3	16.9	25.3
203	17.4	13.9	20.9





#79
Terphenyl-d14
Concen: 105.15 ng
RT: 19.822 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:244 Resp: 4398872

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.5	8.3
122	9.2	7.4	11.0





#80
Butylbenzylphthalate
Concen: 56.68 ng
RT: 20.551 min Scan# 2969
Delta R.T. 0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:149 Resp: 1083395

Ion	Ratio	Lower	Upper
149	100		
91	66.6	55.4	83.0
206	20.8	16.2	24.2





#81
Benzo(a)anthracene
Concen: 53.20 ng
RT: 21.504 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 2396398

Ion	Ratio	Lower	Upper
228	100		
226	27.5	21.4	32.0
229	19.9	15.5	23.3





#82
3,3'-Dichlorobenzidine
Concen: 57.80 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:252 Resp: 966906

Ion	Ratio	Lower	Upper
252	100		
254	65.0	51.8	77.6
126	9.7	7.7	11.5





#83
Chrysene
Concen: 52.71 ng
RT: 21.574 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:228 Resp: 2250800

Ion	Ratio	Lower	Upper
228	100		
226	30.2	23.9	35.9
229	19.3	15.5	23.3





#84
Bis(2-ethylhexyl)phthalate
Concen: 56.42 ng
RT: 21.451 min Scan# 3122
Delta R.T. 0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:149 Resp: 1585070

Ion	Ratio	Lower	Upper
149	100		
167	27.8	22.1	33.1
279	5.9	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 56.91 ng
RT: 22.698 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:149 Resp: 2614856

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	8.0	6.3	9.5





#86
Perylene-d12
Concen: 20.00 ng
RT: 24.810 min Scan# 3693
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:264 Resp: 811421

Ion	Ratio	Lower	Upper
264	100		
260	22.9	18.7	28.1
265	21.9	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 55.06 ng
RT: 28.515 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:276 Resp: 3262830

Ion	Ratio	Lower	Upper
276	100		
138	13.6	11.2	16.8
277	19.3	15.7	23.5





#88
Benzo(b)fluoranthene
Concen: 56.25 ng
RT: 23.763 min Scan# 3515
Delta R.T. 0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:252 Resp: 2612438

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	8.2	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 53.34 ng
RT: 23.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:252 Resp: 2552663

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.6	26.4
125	9.1	7.0	10.4





#90
Benzo(a)pyrene
Concen: 55.51 ng
RT: 24.645 min Scan# 3665
Delta R.T. 0.000 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:252 Resp: 2475976

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	10.2	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 55.60 ng
RT: 28.586 min Scan# 4
Delta R.T. -0.012 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:278 Resp: 2679367

Ion	Ratio	Lower	Upper
278	100		
139	13.2	10.2	15.4
279	23.5	19.0	28.4





#92
Benzo(g,h,i)perylene
Concen: 54.79 ng
RT: 29.656 min Scan# 4517
Delta R.T. -0.006 min
Lab File: BP024619.D
Acq: 13 May 2025 14:05

Tgt Ion:276 Resp: 2625625

Ion	Ratio	Lower	Upper
276	100		
277	23.4	19.0	28.4
138	18.2	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024620.D
Acq On : 13 May 2025 14:45
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: May 13 16:08:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.663	152	106955	20.00	ng		0.00
21) Naphthalene-d8	10.434	136	484921	20.00	ng		0.00
39) Acenaphthene-d10	14.292	164	325327	20.00	ng		-0.01
64) Phenanthrene-d10	17.086	188	646756	20.00	ng		-0.01
76) Chrysene-d12	21.521	240	726281	20.00	ng		0.00
86) Perylene-d12	24.810	264	822943	20.00	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.281	112	784577	125.47	ng		0.00
7) Phenol-d6	6.858	99	1039562	130.26	ng		0.00
23) Nitrobenzene-d5	8.816	82	1188192	123.80	ng		0.00
42) 2,4,6-Tribromophenol	15.816	330	696428	126.27	ng		0.00
45) 2-Fluorobiphenyl	12.910	172	2956939	119.08	ng		0.00
79) Terphenyl-d14	19.822	244	5119872	120.88	ng		0.01
Target Compounds							Qvalue
2) 1,4-Dioxane	3.222	88	170174	57.95	ng		99
3) Pyridine	3.617	79	419426	64.42	ng		98
4) n-Nitrosodimethylamine	3.528	42	178836	62.22	ng		99
6) Aniline	7.005	93	670773	65.10	ng		100
8) 2-Chlorophenol	7.240	128	459952	64.79	ng		99
9) Benzaldehyde	6.816	77	305987	59.33	ng		98
10) Phenol	6.881	94	535743	65.04	ng		99
11) bis(2-Chloroethyl)ether	7.099	93	416957	61.60	ng		99
12) 1,3-Dichlorobenzene	7.552	146	496139	60.75	ng		99
13) 1,4-Dichlorobenzene	7.699	146	502743	61.20	ng		99
14) 1,2-Dichlorobenzene	8.010	146	490458	61.27	ng		99
15) Benzyl Alcohol	7.910	79	413146	68.12	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.181	45	501972	61.93	ng		98
17) 2-Methylphenol	8.110	107	389272	65.89	ng		98
18) Hexachloroethane	8.728	117	191530	60.85	ng		97
19) n-Nitroso-di-n-propyla...	8.463	70	363059	65.76	ng		99
20) 3+4-Methylphenols	8.434	107	543003	67.57	ng		99
22) Acetophenone	8.481	105	744359	61.45	ng	#	98
24) Nitrobenzene	8.857	77	518098	61.34	ng		97
25) Isophorone	9.381	82	1048461	62.98	ng		99
26) 2-Nitrophenol	9.557	139	278764	67.36	ng		99
27) 2,4-Dimethylphenol	9.622	122	462148	63.34	ng		100
28) bis(2-Chloroethoxy)met...	9.851	93	610221	61.35	ng		99
29) 2,4-Dichlorophenol	10.093	162	459806	65.03	ng		98
30) 1,2,4-Trichlorobenzene	10.299	180	478505	60.05	ng		97
31) Naphthalene	10.487	128	1508460	60.03	ng		100
32) Benzoic acid	9.799	122	319375	60.53	ng		99
33) 4-Chloroaniline	10.599	127	671888	66.26	ng		99
34) Hexachlorobutadiene	10.769	225	303458	58.79	ng		98
35) Caprolactam	11.404	113	175653	68.70	ng		97
36) 4-Chloro-3-methylphenol	11.740	107	559138	64.84	ng		100
37) 2-Methylnaphthalene	12.098	142	1008171	61.96	ng		99
38) 1-Methylnaphthalene	12.316	142	1066290	61.25	ng		99
40) 1,2,4,5-Tetrachloroben...	12.469	216	566530	59.97	ng		100
41) Hexachlorocyclopentadiene	12.445	237	381621	66.62	ng		97
43) 2,4,6-Trichlorophenol	12.716	196	400552	65.46	ng		99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024620.D
Acq On : 13 May 2025 14:45
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: May 13 16:08:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.798	196	436361	64.06	ng		99
46) 1,1'-Biphenyl	13.122	154	1437786	59.80	ng		100
47) 2-Chloronaphthalene	13.163	162	1105298	60.29	ng		100
48) 2-Nitroaniline	13.369	65	354977	65.40	ng		98
49) Acenaphthylene	14.016	152	1884018	61.41	ng		100
50) Dimethylphthalate	13.757	163	1498318	60.42	ng		100
51) 2,6-Dinitrotoluene	13.875	165	328682	62.76	ng		99
52) Acenaphthene	14.363	154	1072483	60.80	ng		100
53) 3-Nitroaniline	14.210	138	361174	67.54	ng		99
54) 2,4-Dinitrophenol	14.428	184	196350	61.10	ng		96
55) Dibenzofuran	14.698	168	1691790	60.13	ng		99
56) 4-Nitrophenol	14.545	139	281293	61.78	ng		98
57) 2,4-Dinitrotoluene	14.675	165	471525	63.87	ng		99
58) Fluorene	15.363	166	1376412	60.00	ng		99
59) 2,3,4,6-Tetrachlorophenol	14.934	232	396678	62.78	ng		99
60) Diethylphthalate	15.134	149	1525628	60.87	ng		99
61) 4-Chlorophenyl-phenyle...	15.357	204	692230	60.51	ng		99
62) 4-Nitroaniline	15.404	138	354287	68.05	ng		99
63) Azobenzene	15.657	77	1310800	61.14	ng		97
65) 4,6-Dinitro-2-methylph...	15.457	198	278024	66.55	ng		95
66) n-Nitrosodiphenylamine	15.575	169	1205525	60.68	ng		100
67) 4-Bromophenyl-phenylether	16.269	248	469127	61.56	ng		99
68) Hexachlorobenzene	16.386	284	590615	60.99	ng		99
69) Atrazine	16.557	200	460742	63.29	ng		98
70) Pentachlorophenol	16.739	266	371465	68.29	ng		99
71) Phenanthrene	17.133	178	2183329	60.73	ng		100
72) Anthracene	17.233	178	2210412	61.11	ng		99
73) Carbazole	17.510	167	2042777	62.39	ng		100
74) Di-n-butylphthalate	18.098	149	2644959	62.81	ng		99
75) Fluoranthene	19.222	202	2594532	61.73	ng		100
77) Benzidine	19.422	184	1304259	65.42	ng		99
78) Pyrene	19.598	202	2661811	61.17	ng		100
80) Butylbenzylphthalate	20.551	149	1243403	64.25	ng		100
81) Benzo(a)anthracene	21.504	228	2778689	60.93	ng		99
82) 3,3'-Dichlorobenzidine	21.433	252	1114122	65.77	ng		100
83) Chrysene	21.574	228	2588513	59.88	ng		100
84) Bis(2-ethylhexyl)phtha...	21.445	149	1822019	64.05	ng		99
85) Di-n-octyl phthalate	22.692	149	2998832	64.47	ng		100
87) Indeno(1,2,3-cd)pyrene	28.533	276	3824922	63.64	ng	#	92
88) Benzo(b)fluoranthene	23.768	252	2922103	62.04	ng		99
89) Benzo(k)fluoranthene	23.845	252	3055252	62.95	ng		100
90) Benzo(a)pyrene	24.657	252	2881508	63.69	ng		100
91) Dibenzo(a,h)anthracene	28.603	278	3137905	64.20	ng		99
92) Benzo(g,h,i)perylene	29.692	276	3065416	63.07	ng		100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024620.D
Acq On : 13 May 2025 14:45
Operator : RC/JU
Sample : SSTDICC060
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Quant Time: May 13 16:08:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.663 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 106955

Ion	Ratio	Lower	Upper
152	100		
150	156.6	125.9	188.9
115	62.4	50.1	75.1





#2
1,4-Dioxane
Concen: 57.95 ng
RT: 3.222 min Scan# 23
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 88 Resp: 170174

Ion	Ratio	Lower	Upper
88	100		
58	68.9	54.0	81.0
43	26.5	21.4	32.0





#3
Pyridine
Concen: 64.42 ng
RT: 3.617 min Scan# 9
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 419426

Ion	Ratio	Lower	Upper
79	100		
52	64.6	50.2	75.2
51	34.9	26.8	40.2





#4
n-Nitrosodimethylamine
Concen: 62.22 ng
RT: 3.528 min Scan# 75
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 42 Resp: 178836

Ion	Ratio	Lower	Upper
42	100		
74	131.7	106.2	159.2
44	6.3	5.4	8.2





#5
2-Fluorophenol
Concen: 125.47 ng
RT: 5.281 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:112 Resp: 784577

Ion	Ratio	Lower	Upper
112	100		
64	54.8	44.1	66.1
63	29.8	23.5	35.3





#6
Aniline
Concen: 65.10 ng
RT: 7.005 min Scan# 666
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 93 Resp: 670773

Ion	Ratio	Lower	Upper
93	100		
66	38.8	31.0	46.6
65	19.8	15.5	23.3





#7
Phenol-d6
Concen: 130.26 ng
RT: 6.858 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 99 Resp: 1039562

Ion	Ratio	Lower	Upper
99	100		
42	17.1	14.4	21.6
71	36.1	29.2	43.8





#8
2-Chlorophenol
Concen: 64.79 ng
RT: 7.240 min Scan# 706
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:128 Resp: 459952

Ion	Ratio	Lower	Upper
128	100		
130	32.3	13.0	53.0
64	41.9	22.3	62.3





#9
Benzaldehyde
Concen: 59.33 ng
RT: 6.816 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 305987

Ion	Ratio	Lower	Upper
77	100		
105	101.6	83.3	123.3
106	95.5	78.0	118.0





#10
Phenol
Concen: 65.04 ng
RT: 6.881 min Scan# 645
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 94 Resp: 535743

Ion	Ratio	Lower	Upper
94	100		
65	27.8	8.7	48.7
66	39.4	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 61.60 ng
RT: 7.099 min Scan# 6
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 93 Resp: 416957

Ion	Ratio	Lower	Upper
93	100		
63	71.6	50.6	90.6
95	32.8	12.9	52.9





#12
1,3-Dichlorobenzene
Concen: 60.75 ng
RT: 7.552 min Scan# 759
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:146 Resp: 496139

Ion	Ratio	Lower	Upper
146	100		
148	63.8	51.4	77.2
75	30.8	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 61.20 ng
RT: 7.699 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:146 Resp: 502743

Ion	Ratio	Lower	Upper
146	100		
148	64.4	51.0	76.6
111	41.9	33.0	49.6





#14
1,2-Dichlorobenzene
Concen: 61.27 ng
RT: 8.010 min Scan# 837
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:146 Resp: 490458

Ion	Ratio	Lower	Upper
146	100		
148	62.9	51.0	76.4
111	42.9	35.2	52.8





#15
Benzyl Alcohol
Concen: 68.12 ng
RT: 7.910 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 79 Resp: 413146

Ion	Ratio	Lower	Upper
79	100		
108	77.5	60.8	91.2
77	64.4	51.9	77.9





#16
2,2'-oxybis(1-Chloropropane)
Concen: 61.93 ng
RT: 8.181 min Scan# 866
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 45 Resp: 501972

Ion	Ratio	Lower	Upper
45	100		
77	16.7	0.0	37.7
79	13.1	0.0	34.0





#17
2-Methylphenol
Concen: 65.89 ng
RT: 8.110 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:107 Resp: 389272

Ion	Ratio	Lower	Upper
107	100		
108	109.0	88.9	133.3
77	47.6	39.0	58.6
79	46.6	38.4	57.6





#18
Hexachloroethane
Concen: 60.85 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:117 Resp: 191530

Ion	Ratio	Lower	Upper
117	100		
119	95.3	79.4	119.0
201	101.2	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 65.76 ng
RT: 8.463 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 70 Resp: 363059

Ion	Ratio	Lower	Upper
70	100		
42	52.7	43.0	64.6
101	11.0	8.7	13.1
130	24.4	19.4	29.0





#20
3+4-Methylphenols
Concen: 67.57 ng
RT: 8.434 min Scan# 909
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:107 Resp: 543003

Ion	Ratio	Lower	Upper
107	100		
108	86.0	66.7	106.7
77	34.0	13.1	53.1
79	28.2	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:136 Resp: 484921

Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.8	13.2
54	6.8	5.5	8.3
68	5.2	4.2	6.2





#22
Acetophenone
Concen: 61.45 ng
RT: 8.481 min Scan# 917
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:105 Resp: 744359

Ion	Ratio	Lower	Upper
105	100		
71	1.9	2.0	3.0#
51	24.1	20.4	30.6
120	23.0	18.1	27.1





#23
Nitrobenzene-d5
Concen: 123.80 ng
RT: 8.816 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 1188192

Ion	Ratio	Lower	Upper
82	100		
128	45.8	37.0	55.4
54	48.4	38.7	58.1





#24
Nitrobenzene
Concen: 61.34 ng
RT: 8.857 min Scan# 981
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 77 Resp: 518098

Ion	Ratio	Lower	Upper
77	100		
123	48.7	37.4	56.0
65	13.7	11.3	16.9





#25
Isophorone
Concen: 62.98 ng
RT: 9.381 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 1048461

Ion	Ratio	Lower	Upper
82	100		
95	8.1	6.3	9.5
138	19.7	15.3	22.9





#26
2-Nitrophenol
Concen: 67.36 ng
RT: 9.557 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:139 Resp: 278764

Ion	Ratio	Lower	Upper
139	100		
109	33.4	25.9	38.9
65	47.3	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 63.34 ng
RT: 9.622 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:122 Resp: 462148

Ion	Ratio	Lower	Upper
122	100		
107	120.2	96.2	144.2
121	59.3	47.6	71.4





#28
bis(2-Chloroethoxy)methane
Concen: 61.35 ng
RT: 9.851 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 93 Resp: 610221

Ion	Ratio	Lower	Upper
93	100		
95	32.2	25.5	38.3
123	11.9	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 65.03 ng
RT: 10.093 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 459806

Ion	Ratio	Lower	Upper
162	100		
164	62.9	42.3	82.3
98	36.7	18.4	58.4





#30
1,2,4-Trichlorobenzene
Concen: 60.05 ng
RT: 10.299 min Scan# 1226
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:180 Resp: 478505

Ion	Ratio	Lower	Upper
180	100		
182	95.9	79.0	118.4
145	32.2	26.6	39.8





#31
Naphthalene
Concen: 60.03 ng
RT: 10.487 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:128 Resp: 1508460

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.8	13.2
127	13.3	10.7	16.1





#32
Benzoic acid
Concen: 60.53 ng
RT: 9.799 min Scan# 1141
Delta R.T. 0.024 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:122 Resp: 319375

Ion	Ratio	Lower	Upper
122	100		
105	126.3	106.9	146.9
77	93.0	75.1	115.1





#33
4-Chloroaniline
Concen: 66.26 ng
RT: 10.599 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:127 Resp: 671888

Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.8	38.6
65	29.4	23.9	35.9
92	19.6	16.1	24.1





#34
Hexachlorobutadiene
Concen: 58.79 ng
RT: 10.769 min Scan# 1306
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:225 Resp: 303458

Ion	Ratio	Lower	Upper
225	100		
223	62.5	51.0	76.6
227	64.5	53.1	79.7





#35
Caprolactam
Concen: 68.70 ng
RT: 11.404 min Scan# 14
Delta R.T. 0.018 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:113 Resp: 175653

Ion	Ratio	Lower	Upper
113	100		
55	134.5	119.0	159.0
56	107.3	90.3	130.3





#36
4-Chloro-3-methylphenol
Concen: 64.84 ng
RT: 11.740 min Scan# 1471
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:107 Resp: 559138

Ion	Ratio	Lower	Upper
107	100		
144	24.9	20.1	30.1
142	77.8	62.1	93.1





#37
2-Methylnaphthalene
Concen: 61.96 ng
RT: 12.098 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:142 Resp: 1008171

Ion	Ratio	Lower	Upper
142	100		
141	90.4	71.2	106.8
115	38.1	30.6	46.0





#38
1-Methylnaphthalene
Concen: 61.25 ng
RT: 12.316 min Scan# 1569
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:142 Resp: 1066290

Ion	Ratio	Lower	Upper
142	100		
141	92.9	74.7	112.1
116	4.2	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.292 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:164 Resp: 325327

Ion	Ratio	Lower	Upper
164	100		
162	100.9	81.8	122.6
160	44.6	36.3	54.5





#40
1,2,4,5-Tetrachlorobenzene
Concen: 59.97 ng
RT: 12.469 min Scan# 1595
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:216 Resp: 566530

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.3	93.5
179	22.4	17.7	26.5
108	20.3	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 66.62 ng
RT: 12.445 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 381621

Ion	Ratio	Lower	Upper
237	100		
235	62.4	39.9	79.9
272	13.6	0.0	33.2





#42
2,4,6-Tribromophenol
Concen: 126.27 ng
RT: 15.816 min Scan# 2164
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:330 Resp: 696428

Ion	Ratio	Lower	Upper
330	100		
332	97.5	78.2	117.4
141	29.3	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 65.46 ng
RT: 12.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:196 Resp: 400552

Ion	Ratio	Lower	Upper
196	100		
198	97.0	76.9	115.3
200	31.0	25.2	37.8





#44
2,4,5-Trichlorophenol
Concen: 64.06 ng
RT: 12.798 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:196 Resp: 436361

Ion	Ratio	Lower	Upper
196	100		
198	97.5	79.2	118.8
97	51.4	42.3	63.5
132	29.2	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 119.08 ng
RT: 12.910 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:172 Resp: 2956939

Ion	Ratio	Lower	Upper
172	100		
171	36.1	28.8	43.2
170	23.7	18.8	28.2





#46
1,1'-Biphenyl
Concen: 59.80 ng
RT: 13.122 min Scan# 1706
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:154 Resp: 1437786

Ion	Ratio	Lower	Upper
154	100		
153	41.0	21.0	61.0
76	11.3	0.0	31.5





#47
2-Chloronaphthalene
Concen: 60.29 ng
RT: 13.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:162 Resp: 1105298

Ion	Ratio	Lower	Upper
162	100		
127	42.3	33.8	50.6
164	32.8	26.0	39.0





#48
2-Nitroaniline
Concen: 65.40 ng
RT: 13.369 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion: 65 Resp: 354977

Ion	Ratio	Lower	Upper
65	100		
92	70.9	58.7	88.1
138	114.3	90.2	135.2





#49
Acenaphthylene
Concen: 61.41 ng
RT: 14.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:152 Resp: 1884018

Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.2	24.4
153	13.0	10.2	15.4





#50
Dimethylphthalate
Concen: 60.42 ng
RT: 13.757 min Scan# 1814
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:163 Resp: 1498318

Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.0
164	10.2	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 62.76 ng
RT: 13.875 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 328682

Ion	Ratio	Lower	Upper
165	100		
63	53.7	42.4	63.6
89	56.3	44.8	67.2





#52
Acenaphthene
Concen: 60.80 ng
RT: 14.363 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:154 Resp: 1072483

Ion	Ratio	Lower	Upper
154	100		
153	113.7	91.0	136.4
152	54.4	42.9	64.3





#53
3-Nitroaniline
Concen: 67.54 ng
RT: 14.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:138 Resp: 361174

Ion	Ratio	Lower	Upper
138	100		
108	11.1	9.0	13.4
92	113.7	90.1	135.1





#54
2,4-Dinitrophenol
Concen: 61.10 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:184 Resp: 196350

Ion	Ratio	Lower	Upper
184	100		
63	59.3	51.3	76.9
154	65.7	53.5	80.3





#55
Dibenzofuran
Concen: 60.13 ng
RT: 14.698 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:168 Resp: 1691790

Ion	Ratio	Lower	Upper
168	100		
139	41.8	34.2	51.4
169	13.3	10.8	16.2





#56
4-Nitrophenol
Concen: 61.78 ng
RT: 14.545 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:139 Resp: 281293

Ion	Ratio	Lower	Upper
139	100		
109	80.9	61.4	101.4
65	88.2	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 63.87 ng
RT: 14.675 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 471525

Ion	Ratio	Lower	Upper
165	100		
63	42.1	33.8	50.6
89	70.6	55.6	83.4
182	2.7	2.4	3.6





#58
Fluorene
Concen: 60.00 ng
RT: 15.363 min Scan# 2087
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:166 Resp: 1376412

Ion	Ratio	Lower	Upper
166	100		
165	98.8	78.5	117.7
167	13.6	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 62.78 ng
RT: 14.934 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:232 Resp: 396678

Ion	Ratio	Lower	Upper
232	100		
131	49.0	39.5	59.3
130	2.7	2.2	3.4
166	31.9	25.7	38.5





#60
Diethylphthalate
Concen: 60.87 ng
RT: 15.134 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:149 Resp: 1525628

Ion	Ratio	Lower	Upper
149	100		
177	21.2	16.7	25.1
150	12.6	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 60.51 ng
RT: 15.357 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:204 Resp: 692230

Ion	Ratio	Lower	Upper
204	100		
206	33.5	26.0	39.0
141	65.2	52.9	79.3





#62
4-Nitroaniline
Concen: 68.05 ng
RT: 15.404 min Scan# 2094
Delta R.T. 0.012 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:138 Resp: 354287

Ion	Ratio	Lower	Upper
138	100		
92	49.0	29.2	69.2
108	90.5	69.0	109.0





#63
Azobenzene
Concen: 61.14 ng
RT: 15.657 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion: 77 Resp: 1310800

Ion	Ratio	Lower	Upper
77	100		
182	27.9	5.7	45.7
105	21.8	1.3	41.3
51	29.6	10.6	50.6





#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.086 min Scan# 2380
Delta R.T. -0.012 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:188 Resp: 646756

Ion	Ratio	Lower	Upper
188	100		
94	8.6	6.5	9.7
80	9.6	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 66.55 ng
RT: 15.457 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 278024

Ion	Ratio	Lower	Upper
198	100		
51	38.1	21.1	61.1
105	40.6	23.6	63.6





#66
n-Nitrosodiphenylamine
Concen: 60.68 ng
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:169 Resp: 1205525

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.2	79.8
167	36.3	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 61.56 ng
RT: 16.269 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:248 Resp: 469127

Ion	Ratio	Lower	Upper
248	100		
250	98.6	78.2	117.4
141	76.7	63.0	94.6





#68
Hexachlorobenzene
Concen: 60.99 ng
RT: 16.386 min Scan# 2261
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:284 Resp: 590615

Ion	Ratio	Lower	Upper
284	100		
142	31.9	25.9	38.9
249	30.3	24.1	36.1





#69
Atrazine
Concen: 63.29 ng
RT: 16.557 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:200 Resp: 460742

Ion	Ratio	Lower	Upper
200	100		
173	26.4	5.9	45.9
215	46.1	28.3	68.3





#70
Pentachlorophenol
Concen: 68.29 ng
RT: 16.739 min Scan# 2321
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:266 Resp: 371465

Ion	Ratio	Lower	Upper
266	100		
268	64.5	52.2	78.4
264	62.9	51.1	76.7





#71
Phenanthrene
Concen: 60.73 ng
RT: 17.133 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:178 Resp: 2183329

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.4	23.2
179	15.1	12.2	18.4





#72
Anthracene
Concen: 61.11 ng
RT: 17.233 min Scan# 2405
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:178 Resp: 2210412

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.0	22.4
179	15.3	12.2	18.2





#73
Carbazole
Concen: 62.39 ng
RT: 17.510 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:167 Resp: 2042777

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.7	26.5
139	14.1	11.3	16.9





#74
Di-n-butylphthalate
Concen: 62.81 ng
RT: 18.098 min Scan# 2552
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:149 Resp: 2644959

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.2	10.8
104	5.3	4.3	6.5





#75
Fluoranthene
Concen: 61.73 ng
RT: 19.222 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:202 Resp: 2594532

Ion	Ratio	Lower	Upper
202	100		
101	10.2	0.0	30.3
203	17.8	0.0	37.5





#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.521 min Scan# 3134
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:240 Resp: 726281

Ion	Ratio	Lower	Upper
240	100		
120	9.5	7.8	11.6
236	25.6	20.6	31.0





#77
Benzidine
Concen: 65.42 ng
RT: 19.422 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:184 Resp: 1304259

Ion	Ratio	Lower	Upper
184	100		
185	14.5	11.2	16.8
183	12.4	9.8	14.8





#78
Pyrene
Concen: 61.17 ng
RT: 19.598 min Scan# 2807
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:202 Resp: 2661811

Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.9	25.3
203	17.4	13.9	20.9





#79
Terphenyl-d14
Concen: 120.88 ng
RT: 19.822 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:244 Resp: 5119872

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.5	8.3
122	8.6	7.4	11.0





#80
Butylbenzylphthalate
Concen: 64.25 ng
RT: 20.551 min Scan# 2969
Delta R.T. 0.012 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:149 Resp: 1243403

Ion	Ratio	Lower	Upper
149	100		
91	69.5	55.4	83.0
206	20.3	16.2	24.2





#81
Benzo(a)anthracene
Concen: 60.93 ng
RT: 21.504 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:228 Resp: 2778689

Ion	Ratio	Lower	Upper
228	100		
226	27.5	21.4	32.0
229	20.0	15.5	23.3





#82
3,3'-Dichlorobenzidine
Concen: 65.77 ng
RT: 21.433 min Scan# 3119
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:252 Resp: 1114122

Ion	Ratio	Lower	Upper
252	100		
254	64.9	51.8	77.6
126	10.0	7.7	11.5





#83
Chrysene
Concen: 59.88 ng
RT: 21.574 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:228 Resp: 2588513

Ion	Ratio	Lower	Upper
228	100		
226	30.1	23.9	35.9
229	19.5	15.5	23.3





#84
Bis(2-ethylhexyl)phthalate
Concen: 64.05 ng
RT: 21.445 min Scan# 3121
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:149 Resp: 1822019

Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.1	33.1
279	5.6	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 64.47 ng
RT: 22.692 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:149 Resp: 2998832

Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.9	6.3	9.5





#86
Perylene-d12
Concen: 20.00 ng
RT: 24.810 min Scan# 3693
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:264 Resp: 822943

Ion	Ratio	Lower	Upper
264	100		
260	24.6	18.7	28.1
265	22.1	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 63.64 ng
RT: 28.533 min Scan# 4
Delta R.T. 0.018 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:276 Resp: 3824922

Ion	Ratio	Lower	Upper
276	100		
138	14.4	11.2	16.8
277	25.5	15.7	23.5#





#88
Benzo(b)fluoranthene
Concen: 62.04 ng
RT: 23.768 min Scan# 3516
Delta R.T. 0.012 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:252 Resp: 2922103

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	8.7	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 62.95 ng
RT: 23.845 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:252 Resp: 3055252

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	8.6	7.0	10.4





#90
Benzo(a)pyrene
Concen: 63.69 ng
RT: 24.657 min Scan# 3667
Delta R.T. 0.012 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:252 Resp: 2881508

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	9.7	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 64.20 ng
RT: 28.603 min Scan# 4
Delta R.T. 0.006 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDICC060

Tgt Ion:278 Resp: 3137905

Ion	Ratio	Lower	Upper
278	100		
139	13.3	10.2	15.4
279	23.8	19.0	28.4





#92
Benzo(g,h,i)perylene
Concen: 63.07 ng
RT: 29.692 min Scan# 4523
Delta R.T. 0.029 min
Lab File: BP024620.D
Acq: 13 May 2025 14:45

Tgt Ion:276 Resp: 3065416

Ion	Ratio	Lower	Upper
276	100		
277	23.5	19.0	28.4
138	17.8	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024621.D
Acq On : 13 May 2025 15:26
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 16:10:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.663	152	134481	20.00	ng		0.00
21) Naphthalene-d8	10.434	136	584325	20.00	ng		0.00
39) Acenaphthene-d10	14.298	164	373575	20.00	ng		0.00
64) Phenanthrene-d10	17.104	188	713479	20.00	ng		0.00
76) Chrysene-d12	21.516	240	765954	20.00	ng		0.00
86) Perylene-d12	24.804	264	877352	20.00	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.287	112	1283582	163.25	ng		0.00
7) Phenol-d6	6.858	99	1670035	166.42	ng		0.00
23) Nitrobenzene-d5	8.822	82	1840183	159.12	ng		0.01
42) 2,4,6-Tribromophenol	15.828	330	1032638	163.05	ng		0.02
45) 2-Fluorobiphenyl	12.910	172	4438179	155.65	ng		0.00
79) Terphenyl-d14	19.816	244	7152255	160.11	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.228	88	275790	74.70	ng		99
3) Pyridine	3.622	79	679364	82.98	ng		98
4) n-Nitrosodimethylamine	3.534	42	289012	79.96	ng		100
6) Aniline	7.005	93	1075238	82.99	ng		100
8) 2-Chlorophenol	7.240	128	740005	82.91	ng		99
9) Benzaldehyde	6.822	77	412028	63.54	ng		100
10) Phenol	6.887	94	859122	82.95	ng		100
11) bis(2-Chloroethyl)ether	7.099	93	670572	78.79	ng		100
12) 1,3-Dichlorobenzene	7.552	146	795953	77.51	ng		100
13) 1,4-Dichlorobenzene	7.699	146	802182	77.67	ng		98
14) 1,2-Dichlorobenzene	8.010	146	779316	77.43	ng		99
15) Benzyl Alcohol	7.910	79	664684	87.16	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.187	45	782240	76.75	ng		98
17) 2-Methylphenol	8.110	107	620901	83.58	ng		99
18) Hexachloroethane	8.728	117	306812	77.52	ng		97
19) n-Nitroso-di-n-propyla...	8.469	70	551845	79.50	ng		98
20) 3+4-Methylphenols	8.440	107	852078	84.32	ng		99
22) Acetophenone	8.481	105	1147610	78.62	ng	#	99
24) Nitrobenzene	8.857	77	804153	79.02	ng		98
25) Isophorone	9.381	82	1593235	79.42	ng		100
26) 2-Nitrophenol	9.557	139	442185	88.67	ng		98
27) 2,4-Dimethylphenol	9.628	122	722439	82.17	ng		100
28) bis(2-Chloroethoxy)met...	9.857	93	946615	78.98	ng		99
29) 2,4-Dichlorophenol	10.099	162	721701	84.70	ng		97
30) 1,2,4-Trichlorobenzene	10.299	180	755529	78.69	ng		98
31) Naphthalene	10.487	128	2354789	77.77	ng		100
32) Benzoic acid	9.828	122	525308	80.15	ng		99
33) 4-Chloroaniline	10.599	127	1033865	84.61	ng		99
34) Hexachlorobutadiene	10.769	225	481140	77.35	ng		97
35) Caprolactam	11.416	113	257252	83.50	ng		96
36) 4-Chloro-3-methylphenol	11.734	107	838983	80.74	ng		99
37) 2-Methylnaphthalene	12.093	142	1529445	78.01	ng		100
38) 1-Methylnaphthalene	12.322	142	1613775	76.92	ng		99
40) 1,2,4,5-Tetrachloroben...	12.469	216	875750	80.73	ng		100
41) Hexachlorocyclopentadiene	12.446	237	628436	95.54	ng		97
43) 2,4,6-Trichlorophenol	12.716	196	601332	85.58	ng		99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024621.D
Acq On : 13 May 2025 15:26
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 13 16:10:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.798	196	657399	84.05	ng	97
46) 1,1'-Biphenyl	13.122	154	2175624	78.79	ng	100
47) 2-Chloronaphthalene	13.163	162	1662979	78.99	ng	99
48) 2-Nitroaniline	13.375	65	512458	82.22	ng	97
49) Acenaphthylene	14.016	152	2773742	78.73	ng	99
50) Dimethylphthalate	13.763	163	2182359	76.64	ng	100
51) 2,6-Dinitrotoluene	13.887	165	476606	79.26	ng	97
52) Acenaphthene	14.363	154	1572103	77.62	ng	99
53) 3-Nitroaniline	14.222	138	522849	85.15	ng	100
54) 2,4-Dinitrophenol	14.428	184	300458	79.29	ng	98
55) Dibenzofuran	14.704	168	2477374	76.68	ng	99
56) 4-Nitrophenol	14.557	139	409954	77.27	ng	98
57) 2,4-Dinitrotoluene	14.681	165	678287	80.01	ng	97
58) Fluorene	15.363	166	2003709	76.06	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.945	232	588947	81.17	ng	98
60) Diethylphthalate	15.145	149	2193488	76.22	ng	99
61) 4-Chlorophenyl-phenyle...	15.363	204	997863	75.96	ng	99
62) 4-Nitroaniline	15.404	138	509582	85.24	ng	99
63) Azobenzene	15.663	77	1858146	75.47	ng	99
65) 4,6-Dinitro-2-methylph...	15.469	198	401011	87.01	ng	98
66) n-Nitrosodiphenylamine	15.587	169	1711679	78.10	ng	99
67) 4-Bromophenyl-phenylether	16.281	248	684619	81.43	ng	98
68) Hexachlorobenzene	16.398	284	856547	80.18	ng	99
69) Atrazine	16.575	200	645877	80.42	ng	99
70) Pentachlorophenol	16.751	266	548492	91.41	ng	99
71) Phenanthrene	17.145	178	3091665	77.95	ng	99
72) Anthracene	17.239	178	3161669	79.23	ng	99
73) Carbazole	17.522	167	2892236	80.08	ng	99
74) Di-n-butylphthalate	18.110	149	3747927	80.68	ng	99
75) Fluoranthene	19.228	202	3629085	78.27	ng	99
77) Benzidine	19.422	184	1678507	79.83	ng	100
78) Pyrene	19.604	202	3747276	81.65	ng	99
80) Butylbenzylphthalate	20.545	149	1741954	85.34	ng	98
81) Benzo(a)anthracene	21.498	228	3829787	79.63	ng	99
82) 3,3'-Dichlorobenzidine	21.422	252	1575691	88.21	ng	100
83) Chrysene	21.569	228	3615014	79.29	ng	99
84) Bis(2-ethylhexyl)phtha...	21.439	149	2577445	85.91	ng	99
85) Di-n-octyl phthalate	22.686	149	4341653	88.50	ng	99
87) Indeno(1,2,3-cd)pyrene	28.515	276	5243999m	81.85	ng	
88) Benzo(b)fluoranthene	23.768	252	4173798	83.11	ng	99
89) Benzo(k)fluoranthene	23.833	252	4160268	80.40	ng	100
90) Benzo(a)pyrene	24.645	252	3988360	82.69	ng	99
91) Dibenzo(a,h)anthracene	28.598	278	4338518	83.26	ng	99
92) Benzo(g,h,i)perylene	29.680	276	4253459	82.08	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024621.D
Acq On : 13 May 2025 15:26
Operator : RC/JU
Sample : SSTDICC080
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 13 16:10:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 15:55:25 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.663 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:152 Resp: 13448

Ion	Ratio	Lower	Upper
152	100		
150	153.6	125.9	188.9
115	63.0	50.1	75.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#2
1,4-Dioxane
Concen: 74.70 ng
RT: 3.228 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 88 Resp: 275790

Ion	Ratio	Lower	Upper
88	100		
58	68.0	54.0	81.0
43	26.1	21.4	32.0





#3
Pyridine
Concen: 82.98 ng
RT: 3.622 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 79 Resp: 67936

Ion	Ratio	Lower	Upper
79	100		
52	64.7	50.2	75.2
51	34.6	26.8	40.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#4
n-Nitrosodimethylamine
Concen: 79.96 ng
RT: 3.534 min Scan# 76
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 42 Resp: 289012

Ion	Ratio	Lower	Upper
42	100		
74	132.2	106.2	159.2
44	6.0	5.4	8.2





#5
2-Fluorophenol
Concen: 163.25 ng
RT: 5.287 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:112 Resp: 128358

Ion	Ratio	Lower	Upper
112	100		
64	54.8	44.1	66.1
63	29.4	23.5	35.3

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#6
Aniline
Concen: 82.99 ng
RT: 7.005 min Scan# 666
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 93 Resp: 1075238

Ion	Ratio	Lower	Upper
93	100		
66	38.8	31.0	46.6
65	19.5	15.5	23.3





#7
Phenol-d6
Concen: 166.42 ng
RT: 6.858 min Scan# 64
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 99 Resp: 167003

Ion	Ratio	Lower	Upper
99	100		
42	17.4	14.4	21.6
71	36.6	29.2	43.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#8
2-Chlorophenol
Concen: 82.91 ng
RT: 7.240 min Scan# 706
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:128 Resp: 740005

Ion	Ratio	Lower	Upper
128	100		
130	32.2	13.0	53.0
64	42.1	22.3	62.3





#9
Benzaldehyde
Concen: 63.54 ng
RT: 6.822 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 77 Resp: 41202

Ion	Ratio	Lower	Upper
77	100		
105	103.1	83.3	123.3
106	98.7	78.0	118.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#10
Phenol
Concen: 82.95 ng
RT: 6.887 min Scan# 646
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 94 Resp: 859122

Ion	Ratio	Lower	Upper
94	100		
65	28.6	8.7	48.7
66	39.3	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 78.79 ng
RT: 7.099 min Scan# 6
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 93 Resp: 67057

Ion	Ratio	Lower	Upper
93	100		
63	70.6	50.6	90.6
95	32.7	12.9	52.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#12
1,3-Dichlorobenzene
Concen: 77.51 ng
RT: 7.552 min Scan# 759
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:146 Resp: 795953

Ion	Ratio	Lower	Upper
146	100		
148	64.0	51.4	77.2
75	31.8	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 77.67 ng
RT: 7.699 min Scan# 7
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:146 Resp: 80218

Ion	Ratio	Lower	Upper
146	100		
148	65.1	51.0	76.6
111	43.0	33.0	49.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#14
1,2-Dichlorobenzene
Concen: 77.43 ng
RT: 8.010 min Scan# 837
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:146 Resp: 779316

Ion	Ratio	Lower	Upper
146	100		
148	63.2	51.0	76.4
111	43.5	35.2	52.8





#15
Benzyl Alcohol
Concen: 87.16 ng
RT: 7.910 min Scan# 8
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 79 Resp: 66468

Ion	Ratio	Lower	Upper
79	100		
108	76.7	60.8	91.2
77	63.9	51.9	77.9

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#16
2,2'-oxybis(1-Chloropropane)
Concen: 76.75 ng
RT: 8.187 min Scan# 867
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 45 Resp: 782240

Ion	Ratio	Lower	Upper
45	100		
77	17.6	0.0	37.7
79	12.6	0.0	34.0





#17
2-Methylphenol
Concen: 83.58 ng
RT: 8.110 min Scan# 8
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:107 Resp: 62090

Ion	Ratio	Lower	Upper
107	100		
108	110.9	88.9	133.3
77	47.6	39.0	58.6
79	46.7	38.4	57.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#18
Hexachloroethane
Concen: 77.52 ng
RT: 8.728 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:117 Resp: 306812

Ion	Ratio	Lower	Upper
117	100		
119	96.9	79.4	119.0
201	100.5	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 79.50 ng
RT: 8.469 min Scan# 9
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 70 Resp: 55184

Ion	Ratio	Lower	Upper
70	100		
42	54.6	43.0	64.6
101	11.2	8.7	13.1
130	25.6	19.4	29.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#20
3+4-Methylphenols
Concen: 84.32 ng
RT: 8.440 min Scan# 910
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:107 Resp: 852078

Ion	Ratio	Lower	Upper
107	100		
108	88.1	66.7	106.7
77	33.7	13.1	53.1
79	28.8	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:136 Resp: 58432

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	7.1	5.5	8.3
68	5.1	4.2	6.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#22
Acetophenone
Concen: 78.62 ng
RT: 8.481 min Scan# 917
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:105 Resp: 1147610

Ion	Ratio	Lower	Upper
105	100		
71	4.0	2.0	3.0#
51	25.0	20.4	30.6
120	22.6	18.1	27.1





#23
Nitrobenzene-d5
Concen: 159.12 ng
RT: 8.822 min Scan# 9
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 82 Resp: 184018

Ion	Ratio	Lower	Upper
82	100		
128	47.8	37.0	55.4
54	48.0	38.7	58.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#24
Nitrobenzene
Concen: 79.02 ng
RT: 8.857 min Scan# 981
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 77 Resp: 804153

Ion	Ratio	Lower	Upper
77	100		
123	48.2	37.4	56.0
65	14.0	11.3	16.9





#25
Isophorone
Concen: 79.42 ng
RT: 9.381 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 82 Resp: 159323

Ion	Ratio	Lower	Upper
82	100		
95	8.0	6.3	9.5
138	18.9	15.3	22.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#26
2-Nitrophenol
Concen: 88.67 ng
RT: 9.557 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:139 Resp: 442185

Ion	Ratio	Lower	Upper
139	100		
109	33.1	25.9	38.9
65	48.8	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 82.17 ng
RT: 9.628 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:122 Resp: 72243

Ion	Ratio	Lower	Upper
122	100		
107	120.6	96.2	144.2
121	59.7	47.6	71.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#28
bis(2-Chloroethoxy)methane
Concen: 78.98 ng
RT: 9.857 min Scan# 1151
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 93 Resp: 946615

Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.5	38.3
123	12.2	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 84.70 ng
RT: 10.099 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:162 Resp: 72170

Ion	Ratio	Lower	Upper
162	100		
164	65.1	42.3	82.3
98	36.5	18.4	58.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#30
1,2,4-Trichlorobenzene
Concen: 78.69 ng
RT: 10.299 min Scan# 1226
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:180 Resp: 755529

Ion	Ratio	Lower	Upper
180	100		
182	96.5	79.0	118.4
145	33.1	26.6	39.8





#31
Naphthalene
Concen: 77.77 ng
RT: 10.487 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:128 Resp: 235478

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.5	10.7	16.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#32
Benzoic acid
Concen: 80.15 ng
RT: 9.828 min Scan# 1146
Delta R.T. 0.053 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:122 Resp: 525308

Ion	Ratio	Lower	Upper
122	100		
105	127.9	106.9	146.9
77	94.2	75.1	115.1





#33
4-Chloroaniline
Concen: 84.61 ng
RT: 10.599 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:127 Resp: 103386

Ion	Ratio	Lower	Upper
127	100		
129	31.9	25.8	38.6
65	29.1	23.9	35.9
92	19.5	16.1	24.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#34
Hexachlorobutadiene
Concen: 77.35 ng
RT: 10.769 min Scan# 1306
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:225 Resp: 481140

Ion	Ratio	Lower	Upper
225	100		
223	61.9	51.0	76.6
227	63.8	53.1	79.7





#35
Caprolactam
Concen: 83.50 ng
RT: 11.416 min Scan# 14
Delta R.T. 0.029 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:113 Resp: 25725

Ion	Ratio	Lower	Upper
113	100		
55	133.3	119.0	159.0
56	107.8	90.3	130.3

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#36
4-Chloro-3-methylphenol
Concen: 80.74 ng
RT: 11.734 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:107 Resp: 838983

Ion	Ratio	Lower	Upper
107	100		
144	24.8	20.1	30.1
142	77.1	62.1	93.1





#37
2-Methylnaphthalene
Concen: 78.01 ng
RT: 12.093 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:142 Resp: 152944

Ion	Ratio	Lower	Upper
142	100		
141	89.3	71.2	106.8
115	38.5	30.6	46.0

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#38
1-Methylnaphthalene
Concen: 76.92 ng
RT: 12.322 min Scan# 1570
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:142 Resp: 1613775

Ion	Ratio	Lower	Upper
142	100		
141	92.6	74.7	112.1
116	4.1	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.298 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:164 Resp: 37357

Ion	Ratio	Lower	Upper
164	100		
162	101.5	81.8	122.6
160	45.5	36.3	54.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 80.73 ng
RT: 12.469 min Scan# 1595
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:216 Resp: 875750

Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.3	93.5
179	22.4	17.7	26.5
108	19.6	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 95.54 ng
RT: 12.446 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:237 Resp: 62843

Ion	Ratio	Lower	Upper
237	100		
235	61.9	39.9	79.9
272	14.1	0.0	33.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#42
2,4,6-Tribromophenol
Concen: 163.05 ng
RT: 15.828 min Scan# 2166
Delta R.T. 0.018 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:330 Resp: 1032638

Ion	Ratio	Lower	Upper
330	100		
332	96.7	78.2	117.4
141	28.8	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 85.58 ng
RT: 12.716 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:196 Resp: 60133

Ion	Ratio	Lower	Upper
196	100		
198	94.9	76.9	115.3
200	30.6	25.2	37.8

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#44
2,4,5-Trichlorophenol
Concen: 84.05 ng
RT: 12.798 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:196 Resp: 657399

Ion	Ratio	Lower	Upper
196	100		
198	95.7	79.2	118.8
97	51.1	42.3	63.5
132	29.5	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 155.65 ng
RT: 12.910 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:172 Resp: 443817

Ion	Ratio	Lower	Upper
172	100		
171	36.3	28.8	43.2
170	24.1	18.8	28.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#46
1,1'-Biphenyl
Concen: 78.79 ng
RT: 13.122 min Scan# 1706
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:154 Resp: 2175624

Ion	Ratio	Lower	Upper
154	100		
153	41.1	21.0	61.0
76	11.2	0.0	31.5





#47
2-Chloronaphthalene
Concen: 78.99 ng
RT: 13.163 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:162 Resp: 166297

Ion	Ratio	Lower	Upper
162	100		
127	42.8	33.8	50.6
164	32.6	26.0	39.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#48
2-Nitroaniline
Concen: 82.22 ng
RT: 13.375 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 65 Resp: 512458

Ion	Ratio	Lower	Upper
65	100		
92	72.7	58.7	88.1
138	117.2	90.2	135.2





#49
Acenaphthylene
Concen: 78.73 ng
RT: 14.016 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:152 Resp: 277374

Ion	Ratio	Lower	Upper
152	100		
151	20.6	16.2	24.4
153	13.1	10.2	15.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#50
Dimethylphthalate
Concen: 76.64 ng
RT: 13.763 min Scan# 1815
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:163 Resp: 2182359

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.0
164	10.0	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 79.26 ng
RT: 13.887 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:165 Resp: 47660

Ion	Ratio	Lower	Upper
165	100		
63	50.8	42.4	63.6
89	54.4	44.8	67.2

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#52
Acenaphthene
Concen: 77.62 ng
RT: 14.363 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:154 Resp: 1572103

Ion	Ratio	Lower	Upper
154	100		
153	112.6	91.0	136.4
152	53.6	42.9	64.3





#53
3-Nitroaniline
Concen: 85.15 ng
RT: 14.222 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:138 Resp: 52284

Ion	Ratio	Lower	Upper
138	100		
108	10.8	9.0	13.4
92	112.3	90.1	135.1



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#54
2,4-Dinitrophenol
Concen: 79.29 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:184 Resp: 300458

Ion	Ratio	Lower	Upper
184	100		
63	62.1	51.3	76.9
154	67.9	53.5	80.3





#55
Dibenzofuran
Concen: 76.68 ng
RT: 14.704 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:168 Resp: 247737

Ion	Ratio	Lower	Upper
168	100		
139	41.7	34.2	51.4
169	13.4	10.8	16.2



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#56
4-Nitrophenol
Concen: 77.27 ng
RT: 14.557 min Scan# 1950
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:139 Resp: 409954

Ion	Ratio	Lower	Upper
139	100		
109	82.6	61.4	101.4
65	89.8	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 80.01 ng
RT: 14.681 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:165 Resp: 67828

Ion	Ratio	Lower	Upper
165	100		
63	44.4	33.8	50.6
89	71.6	55.6	83.4
182	3.1	2.4	3.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#58
Fluorene
Concen: 76.06 ng
RT: 15.363 min Scan# 2087
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:166 Resp: 2003709

Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.5	117.7
167	13.7	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 81.17 ng
RT: 14.945 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:232 Resp: 58894

Ion	Ratio	Lower	Upper
232	100		
131	48.1	39.5	59.3
130	2.6	2.2	3.4
166	31.4	25.7	38.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#60
Diethylphthalate
Concen: 76.22 ng
RT: 15.145 min Scan# 2050
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:149 Resp: 2193488

Ion	Ratio	Lower	Upper
149	100		
177	21.9	16.7	25.1
150	12.9	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 75.96 ng
RT: 15.363 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:204 Resp: 99786

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.0	39.0
141	65.6	52.9	79.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#62
4-Nitroaniline
Concen: 85.24 ng
RT: 15.404 min Scan# 2094
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:138 Resp: 509582

Ion	Ratio	Lower	Upper
138	100		
92	50.6	29.2	69.2
108	88.2	69.0	109.0





#63
Azobenzene
Concen: 75.47 ng
RT: 15.663 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion: 77 Resp: 185814

Ion	Ratio	Lower	Upper
77	100		
182	26.7	5.7	45.7
105	21.6	1.3	41.3
51	30.9	10.6	50.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.104 min Scan# 2383
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:188 Resp: 713479

Ion	Ratio	Lower	Upper
188	100		
94	8.5	6.5	9.7
80	9.0	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 87.01 ng
RT: 15.469 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:198 Resp: 40101

Ion	Ratio	Lower	Upper
198	100		
51	39.8	21.1	61.1
105	42.9	23.6	63.6

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#66
n-Nitrosodiphenylamine
Concen: 78.10 ng
RT: 15.587 min Scan# 2125
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:169 Resp: 1711679

Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.2	79.8
167	36.7	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 81.43 ng
RT: 16.281 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:248 Resp: 68461

Ion	Ratio	Lower	Upper
248	100		
250	97.8	78.2	117.4
141	74.7	63.0	94.6



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#68
Hexachlorobenzene
Concen: 80.18 ng
RT: 16.398 min Scan# 2263
Delta R.T. 0.018 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:284 Resp: 856547

Ion	Ratio	Lower	Upper
284	100		
142	31.4	25.9	38.9
249	29.7	24.1	36.1





#69
Atrazine
Concen: 80.42 ng
RT: 16.575 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:200 Resp: 64587

Ion	Ratio	Lower	Upper
200	100		
173	26.3	5.9	45.9
215	47.1	28.3	68.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#70
Pentachlorophenol
Concen: 91.41 ng
RT: 16.751 min Scan# 2323
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:266 Resp: 548492

Ion	Ratio	Lower	Upper
266	100		
268	64.2	52.2	78.4
264	62.9	51.1	76.7





#71
Phenanthrene
Concen: 77.95 ng
RT: 17.145 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:178 Resp: 309166

Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.4	23.2
179	15.4	12.2	18.4

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#72
Anthracene
Concen: 79.23 ng
RT: 17.239 min Scan# 2406
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:178 Resp: 3161669

Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.0	22.4
179	15.5	12.2	18.2





#73
Carbazole
Concen: 80.08 ng
RT: 17.522 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:167 Resp: 289223

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.7	26.5
139	14.5	11.3	16.9



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#74
Di-n-butylphthalate
Concen: 80.68 ng
RT: 18.110 min Scan# 2554
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:149 Resp: 3747927

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.2	10.8
104	5.5	4.3	6.5





#75
Fluoranthene
Concen: 78.27 ng
RT: 19.228 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:202 Resp: 362908

Ion	Ratio	Lower	Upper
202	100		
101	10.5	0.0	30.3
203	17.8	0.0	37.5

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.516 min Scan# 3133
Delta R.T. -0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:240 Resp: 765954

Ion	Ratio	Lower	Upper
240	100		
120	10.3	7.8	11.6
236	26.2	20.6	31.0





#77
Benzidine
Concen: 79.83 ng
RT: 19.422 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:184 Resp: 167850

Ion	Ratio	Lower	Upper
184	100		
185	14.1	11.2	16.8
183	12.4	9.8	14.8



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#78
Pyrene
Concen: 81.65 ng
RT: 19.604 min Scan# 2808
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:202 Resp: 3747276

Ion	Ratio	Lower	Upper
202	100		
200	21.7	16.9	25.3
203	17.6	13.9	20.9





#79
Terphenyl-d14
Concen: 160.11 ng
RT: 19.816 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:244 Resp: 715225

Ion	Ratio	Lower	Upper
244	100		
212	7.4	5.5	8.3
122	9.1	7.4	11.0



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#80
Butylbenzylphthalate
Concen: 85.34 ng
RT: 20.545 min Scan# 2968
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:149 Resp: 1741954

Ion	Ratio	Lower	Upper
149	100		
91	67.3	55.4	83.0
206	20.9	16.2	24.2





#81
Benzo(a)anthracene
Concen: 79.63 ng
RT: 21.498 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:228 Resp: 382978

Ion	Ratio	Lower	Upper
228	100		
226	27.6	21.4	32.0
229	19.9	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#82
3,3'-Dichlorobenzidine
Concen: 88.21 ng
RT: 21.422 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:252 Resp: 1575691

Ion	Ratio	Lower	Upper
252	100		
254	64.8	51.8	77.6
126	10.2	7.7	11.5





#83
Chrysene
Concen: 79.29 ng
RT: 21.569 min Scan# 3
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:228 Resp: 361501

Ion	Ratio	Lower	Upper
228	100		
226	30.2	23.9	35.9
229	19.7	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 85.91 ng
RT: 21.439 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:149 Resp: 2577445

Ion	Ratio	Lower	Upper
149	100		
167	28.2	22.1	33.1
279	5.7	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 88.50 ng
RT: 22.686 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:149 Resp: 434165

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.6	6.3	9.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#86
Perylene-d12
Concen: 20.00 ng
RT: 24.804 min Scan# 3692
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:264 Resp: 877352

Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.7	28.1
265	21.9	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 81.85 ng m
RT: 28.515 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:276 Resp: 524399

Ion	Ratio	Lower	Upper
276	100		
138	13.7	11.2	16.8
277	20.5	15.7	23.5



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#88
Benzo(b)fluoranthene
Concen: 83.11 ng
RT: 23.768 min Scan# 3516
Delta R.T. 0.012 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:252 Resp: 4173798

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	8.2	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 80.40 ng
RT: 23.833 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:252 Resp: 416026

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.6	26.4
125	8.5	7.0	10.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#90
Benzo(a)pyrene
Concen: 82.69 ng
RT: 24.645 min Scan# 3665
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:252 Resp: 3988360

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	9.5	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 83.26 ng
RT: 28.598 min Scan# 4
Delta R.T. 0.000 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:278 Resp: 433851

Ion	Ratio	Lower	Upper
278	100		
139	13.3	10.2	15.4
279	23.9	19.0	28.4



Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#92
Benzo(g,h,i)perylene
Concen: 82.08 ng
RT: 29.680 min Scan# 4521
Delta R.T. 0.018 min
Lab File: BP024621.D
Acq: 13 May 2025 15:26

Tgt Ion:276 Resp: 4253459

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.0	28.4
138	17.3	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.657	152	144423	20.00	ng	0.00
21) Naphthalene-d8	10.434	136	621777	20.00	ng	0.00
39) Acenaphthene-d10	14.292	164	391112	20.00	ng	-0.01
64) Phenanthrene-d10	17.092	188	780511	20.00	ng	0.00
76) Chrysene-d12	21.521	240	868822	20.00	ng	0.00
86) Perylene-d12	24.815	264	1015341	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.287	112	658249	77.95	ng	0.00
7) Phenol-d6	6.852	99	852330	79.09	ng	0.00
23) Nitrobenzene-d5	8.810	82	962660	78.23	ng	0.00
42) 2,4,6-Tribromophenol	15.810	330	520660	78.53	ng	0.00
45) 2-Fluorobiphenyl	12.904	172	2290715	76.73	ng	0.00
79) Terphenyl-d14	19.822	244	3889002	76.75	ng	0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.222	88	145871	36.79	ng	99
3) Pyridine	3.622	79	347515	39.53	ng	97
4) n-Nitrosodimethylamine	3.528	42	147412	37.98	ng	98
6) Aniline	6.999	93	545556	39.21	ng	100
8) 2-Chlorophenol	7.234	128	377358	39.37	ng	97
9) Benzaldehyde	6.816	77	272303	39.03	ng	99
10) Phenol	6.881	94	434788	39.09	ng	99
11) bis(2-Chloroethyl)ether	7.093	93	346744	37.94	ng	99
12) 1,3-Dichlorobenzene	7.552	146	417899	37.89	ng	98
13) 1,4-Dichlorobenzene	7.699	146	420432	37.90	ng	99
14) 1,2-Dichlorobenzene	8.010	146	409587	37.90	ng	99
15) Benzyl Alcohol	7.904	79	332015	40.54	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.181	45	416339	38.04	ng	99
17) 2-Methylphenol	8.110	107	316066	39.62	ng	99
18) Hexachloroethane	8.728	117	158167	37.21	ng	99
19) n-Nitroso-di-n-propyla...	8.463	70	291961	39.16	ng	99
20) 3+4-Methylphenols	8.434	107	429326	39.56	ng	99
22) Acetophenone	8.475	105	597464	38.47	ng	99
24) Nitrobenzene	8.851	77	416985	38.50	ng	96
25) Isophorone	9.375	82	821637	38.49	ng	99
26) 2-Nitrophenol	9.557	139	218211	41.12	ng	95
27) 2,4-Dimethylphenol	9.622	122	369418	39.48	ng	100
28) bis(2-Chloroethoxy)met...	9.851	93	493722	38.71	ng	100
29) 2,4-Dichlorophenol	10.093	162	364373	40.19	ng	97
30) 1,2,4-Trichlorobenzene	10.298	180	389362	38.11	ng	98
31) Naphthalene	10.481	128	1226624	38.07	ng	100
32) Benzoic acid	9.781	122	223660	36.46	ng	98
33) 4-Chloroaniline	10.598	127	524873	40.37	ng	99
34) Hexachlorobutadiene	10.763	225	246776	37.29	ng	98
35) Caprolactam	11.387	113	130399	39.75	ng	99
36) 4-Chloro-3-methylphenol	11.740	107	431344	39.01	ng	100
37) 2-Methylnaphthalene	12.098	142	789724	37.85	ng	99
38) 1-Methylnaphthalene	12.316	142	833521	37.34	ng	100
40) 1,2,4,5-Tetrachloroben...	12.469	216	444420	39.13	ng	100
41) Hexachlorocyclopentadiene	12.445	237	305575	44.37	ng	98
43) 2,4,6-Trichlorophenol	12.716	196	305651	41.55	ng	99

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.798	196	336252	41.06	ng	97
46) 1,1'-Biphenyl	13.116	154	1124669	38.91	ng	99
47) 2-Chloronaphthalene	13.157	162	864853	39.24	ng	98
48) 2-Nitroaniline	13.369	65	264965	40.60	ng	97
49) Acenaphthylene	14.010	152	1419821	38.49	ng	100
50) Dimethylphthalate	13.751	163	1155211	38.75	ng	99
51) 2,6-Dinitrotoluene	13.875	165	247537	39.32	ng	99
52) Acenaphthene	14.357	154	803755	37.90	ng	99
53) 3-Nitroaniline	14.210	138	262737	40.87	ng	100
54) 2,4-Dinitrophenol	14.428	184	135624	37.83	ng	96
55) Dibenzofuran	14.698	168	1282228	37.91	ng	99
56) 4-Nitrophenol	14.545	139	202528	38.70	ng	97
57) 2,4-Dinitrotoluene	14.675	165	354817	39.97	ng	99
58) Fluorene	15.357	166	1059642	38.42	ng	100
59) 2,3,4,6-Tetrachlorophenol	14.934	232	298327	39.27	ng	99
60) Diethylphthalate	15.134	149	1150091	38.17	ng	100
61) 4-Chlorophenyl-phenyle...	15.357	204	526066	38.25	ng	99
62) 4-Nitroaniline	15.398	138	265185	42.72	ng	98
63) Azobenzene	15.657	77	1000653	38.82	ng	98
65) 4,6-Dinitro-2-methylph...	15.451	198	203200	40.30	ng	98
66) n-Nitrosodiphenylamine	15.575	169	920187	38.38	ng	100
67) 4-Bromophenyl-phenylether	16.269	248	357846	38.91	ng	97
68) Hexachlorobenzene	16.386	284	446142	38.18	ng	99
69) Atrazine	16.551	200	347661	39.57	ng	99
70) Pentachlorophenol	16.739	266	269781	41.12	ng	99
71) Phenanthrene	17.133	178	1629118	37.55	ng	99
72) Anthracene	17.228	178	1669395	38.24	ng	100
73) Carbazole	17.510	167	1563685	39.57	ng	99
74) Di-n-butylphthalate	18.098	149	1978741	38.94	ng	100
75) Fluoranthene	19.222	202	1956043	38.57	ng	100
77) Benzidine	19.422	184	1023514	43.25	ng	99
78) Pyrene	19.604	202	1992845	38.28	ng	100
80) Butylbenzylphthalate	20.557	149	938210	40.52	ng	96
81) Benzo(a)anthracene	21.510	228	2121561	38.89	ng	99
82) 3,3'-Dichlorobenzidine	21.433	252	861323	42.51	ng	100
83) Chrysene	21.574	228	1976940	38.23	ng	99
84) Bis(2-ethylhexyl)phtha...	21.445	149	1375919	40.43	ng	100
85) Di-n-octyl phthalate	22.698	149	2321673	41.72	ng	100
87) Indeno(1,2,3-cd)pyrene	28.515	276	2879864m	38.85	ng	
88) Benzo(b)fluoranthene	23.768	252	2283322	39.29	ng	100
89) Benzo(k)fluoranthene	23.833	252	2289000	38.22	ng	100
90) Benzo(a)pyrene	24.668	252	2177020	39.00	ng	99
91) Dibenzo(a,h)anthracene	28.615	278	2374558	39.38	ng	100
92) Benzo(g,h,i)perylene	29.686	276	2333151	38.91	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.657 min Scan# 7
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:152 Resp: 14442

Ion	Ratio	Lower	Upper
152	100		
150	157.1	125.9	188.9
115	64.8	50.1	75.1



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#2
1,4-Dioxane
Concen: 36.79 ng
RT: 3.222 min Scan# 23
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 88 Resp: 145871

Ion	Ratio	Lower	Upper
88	100		
58	68.6	54.0	81.0
43	25.9	21.4	32.0





#3
Pyridine
Concen: 39.53 ng
RT: 3.622 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 79 Resp: 34751

Ion	Ratio	Lower	Upper
79	100		
52	65.0	50.2	75.2
51	34.5	26.8	40.2



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#4
n-Nitrosodimethylamine
Concen: 37.98 ng
RT: 3.528 min Scan# 75
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 42 Resp: 147412

Ion	Ratio	Lower	Upper
42	100		
74	130.6	106.2	159.2
44	6.4	5.4	8.2





#5
2-Fluorophenol
Concen: 77.95 ng
RT: 5.287 min Scan# 3
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:112 Resp: 65824

Ion	Ratio	Lower	Upper
112	100		
64	53.7	44.1	66.1
63	28.7	23.5	35.3



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#6
Aniline
Concen: 39.21 ng
RT: 6.999 min Scan# 665
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 93 Resp: 545556

Ion	Ratio	Lower	Upper
93	100		
66	39.0	31.0	46.6
65	19.6	15.5	23.3





#7
Phenol-d6
Concen: 79.09 ng
RT: 6.852 min Scan# 64
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 99 Resp: 85233

Ion	Ratio	Lower	Upper
99	100		
42	17.0	14.4	21.6
71	35.9	29.2	43.8



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#8
2-Chlorophenol
Concen: 39.37 ng
RT: 7.234 min Scan# 705
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:128 Resp: 377358

Ion	Ratio	Lower	Upper
128	100		
130	31.4	13.0	53.0
64	43.9	22.3	62.3





#9
Benzaldehyde
Concen: 39.03 ng
RT: 6.816 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 77 Resp: 27230

Ion	Ratio	Lower	Upper
77	100		
105	102.4	83.3	123.3
106	98.5	78.0	118.0



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#10
Phenol
Concen: 39.09 ng
RT: 6.881 min Scan# 645
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 94 Resp: 434788

Ion	Ratio	Lower	Upper
94	100		
65	28.6	8.7	48.7
66	40.5	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 37.94 ng
RT: 7.093 min Scan# 6
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 93 Resp: 34674

Ion	Ratio	Lower	Upper
93	100		
63	71.9	50.6	90.6
95	32.1	12.9	52.9



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#12
1,3-Dichlorobenzene
Concen: 37.89 ng
RT: 7.552 min Scan# 759
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:146 Resp: 417899

Ion	Ratio	Lower	Upper
146	100		
148	63.3	51.4	77.2
75	30.1	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 37.90 ng
RT: 7.699 min Scan# 7
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:146 Resp: 42043

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.0	76.6
111	41.8	33.0	49.6



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#14
1,2-Dichlorobenzene
Concen: 37.90 ng
RT: 8.010 min Scan# 837
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:146 Resp: 409587

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.0	76.4
111	43.6	35.2	52.8





#15
Benzyl Alcohol
Concen: 40.54 ng
RT: 7.904 min Scan# 81
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 79 Resp: 33201

Ion	Ratio	Lower	Upper
79	100		
108	76.4	60.8	91.2
77	63.7	51.9	77.9

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.04 ng
RT: 8.181 min Scan# 866
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 45 Resp: 416339

Ion	Ratio	Lower	Upper
45	100		
77	17.8	0.0	37.7
79	13.2	0.0	34.0





#17
2-Methylphenol
Concen: 39.62 ng
RT: 8.110 min Scan# 8
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:107 Resp: 31606

Ion	Ratio	Lower	Upper
107	100		
108	111.5	88.9	133.3
77	46.8	39.0	58.6
79	47.0	38.4	57.6



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#18
Hexachloroethane
Concen: 37.21 ng
RT: 8.728 min Scan# 959
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:117 Resp: 158167

Ion	Ratio	Lower	Upper
117	100		
119	97.9	79.4	119.0
201	103.0	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 39.16 ng
RT: 8.463 min Scan# 9
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 70 Resp: 29196

Ion	Ratio	Lower	Upper
70	100		
42	53.7	43.0	64.6
101	10.8	8.7	13.1
130	25.2	19.4	29.0



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#20
3+4-Methylphenols
Concen: 39.56 ng
RT: 8.434 min Scan# 909
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:107 Resp: 429326

Ion	Ratio	Lower	Upper
107	100		
108	88.1	66.7	106.7
77	33.0	13.1	53.1
79	29.8	8.9	48.9





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:136 Resp: 62177

Ion	Ratio	Lower	Upper
136	100		
137	10.4	8.8	13.2
54	6.7	5.5	8.3
68	4.9	4.2	6.2



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#22
Acetophenone
Concen: 38.47 ng
RT: 8.475 min Scan# 916
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:105 Resp: 597464

Ion	Ratio	Lower	Upper
105	100		
71	2.6	2.0	3.0
51	24.9	20.4	30.6
120	22.7	18.1	27.1





#23
Nitrobenzene-d5
Concen: 78.23 ng
RT: 8.810 min Scan# 9
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 82 Resp: 96266

Ion	Ratio	Lower	Upper
82	100		
128	46.4	37.0	55.4
54	48.3	38.7	58.1



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#24
Nitrobenzene
Concen: 38.50 ng
RT: 8.851 min Scan# 980
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 77 Resp: 416985

Ion	Ratio	Lower	Upper
77	100		
123	49.5	37.4	56.0
65	13.5	11.3	16.9





#25
Isophorone
Concen: 38.49 ng
RT: 9.375 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 82 Resp: 82163

Ion	Ratio	Lower	Upper
82	100		
95	8.2	6.3	9.5
138	18.9	15.3	22.9



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#26
2-Nitrophenol
Concen: 41.12 ng
RT: 9.557 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:139 Resp: 218211

Ion	Ratio	Lower	Upper
139	100		
109	34.0	25.9	38.9
65	51.3	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 39.48 ng
RT: 9.622 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:122 Resp: 369418

Ion	Ratio	Lower	Upper
122	100		
107	120.3	96.2	144.2
121	58.9	47.6	71.4



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#28
bis(2-Chloroethoxy)methane
Concen: 38.71 ng
RT: 9.851 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 93 Resp: 493722

Ion	Ratio	Lower	Upper
93	100		
95	32.1	25.5	38.3
123	12.0	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 40.19 ng
RT: 10.093 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:162 Resp: 36437

Ion	Ratio	Lower	Upper
162	100		
164	64.1	42.3	82.3
98	36.5	18.4	58.4



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#30
1,2,4-Trichlorobenzene
Concen: 38.11 ng
RT: 10.298 min Scan# 1226
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:180 Resp: 389362

Ion	Ratio	Lower	Upper
180	100		
182	96.8	79.0	118.4
145	32.9	26.6	39.8





#31
Naphthalene
Concen: 38.07 ng
RT: 10.481 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:128 Resp: 122662

Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.8	13.2
127	13.4	10.7	16.1



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#32
Benzoic acid
Concen: 36.46 ng
RT: 9.781 min Scan# 1138
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:122 Resp: 223660

Ion	Ratio	Lower	Upper
122	100		
105	124.2	106.9	146.9
77	92.5	75.1	115.1





#33
4-Chloroaniline
Concen: 40.37 ng
RT: 10.598 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:127 Resp: 52487

Ion	Ratio	Lower	Upper
127	100		
129	31.6	25.8	38.6
65	30.3	23.9	35.9
92	19.9	16.1	24.1



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#34
Hexachlorobutadiene
Concen: 37.29 ng
RT: 10.763 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:225 Resp: 246776

Ion	Ratio	Lower	Upper
225	100		
223	62.5	51.0	76.6
227	64.0	53.1	79.7





#35
Caprolactam
Concen: 39.75 ng
RT: 11.387 min Scan# 14
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:113 Resp: 13039

Ion	Ratio	Lower	Upper
113	100		
55	138.6	119.0	159.0
56	108.8	90.3	130.3



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#36
4-Chloro-3-methylphenol
Concen: 39.01 ng
RT: 11.740 min Scan# 1471
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:107 Resp: 431344

Ion	Ratio	Lower	Upper
107	100		
144	25.7	20.1	30.1
142	77.8	62.1	93.1





#37
2-Methylnaphthalene
Concen: 37.85 ng
RT: 12.098 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:142 Resp: 78972

Ion	Ratio	Lower	Upper
142	100		
141	88.5	71.2	106.8
115	37.2	30.6	46.0



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#38
1-Methylnaphthalene
Concen: 37.34 ng
RT: 12.316 min Scan# 1569
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:142 Resp: 833521

Ion	Ratio	Lower	Upper
142	100		
141	93.2	74.7	112.1
116	4.2	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.292 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:164 Resp: 39111

Ion	Ratio	Lower	Upper
164	100		
162	102.8	81.8	122.6
160	45.2	36.3	54.5



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.13 ng
RT: 12.469 min Scan# 1595
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:216 Resp: 444420

Ion	Ratio	Lower	Upper
216	100		
214	78.0	62.3	93.5
179	22.2	17.7	26.5
108	20.3	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 44.37 ng
RT: 12.445 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:237 Resp: 30557

Ion	Ratio	Lower	Upper
237	100		
235	61.2	39.9	79.9
272	14.3	0.0	33.2



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#42
2,4,6-Tribromophenol
Concen: 78.53 ng
RT: 15.810 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:330 Resp: 520660

Ion	Ratio	Lower	Upper
330	100		
332	98.2	78.2	117.4
141	29.8	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 41.55 ng
RT: 12.716 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:196 Resp: 30565

Ion	Ratio	Lower	Upper
196	100		
198	95.5	76.9	115.3
200	32.0	25.2	37.8



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#44
2,4,5-Trichlorophenol
Concen: 41.06 ng
RT: 12.798 min Scan# 1651
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:196 Resp: 336252

Ion	Ratio	Lower	Upper
196	100		
198	95.6	79.2	118.8
97	50.2	42.3	63.5
132	28.5	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 76.73 ng
RT: 12.904 min Scan# 1(
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:172 Resp: 229071

Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.8	43.2
170	23.9	18.8	28.2



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#46
1,1'-Biphenyl
Concen: 38.91 ng
RT: 13.116 min Scan# 1705
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:154 Resp: 1124669

Ion	Ratio	Lower	Upper
154	100		
153	40.5	21.0	61.0
76	11.5	0.0	31.5





#47
2-Chloronaphthalene
Concen: 39.24 ng
RT: 13.157 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:162 Resp: 86485

Ion	Ratio	Lower	Upper
162	100		
127	43.4	33.8	50.6
164	33.5	26.0	39.0



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#48
2-Nitroaniline
Concen: 40.60 ng
RT: 13.369 min Scan# 1748
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 65 Resp: 264965

Ion	Ratio	Lower	Upper
65	100		
92	71.2	58.7	88.1
138	108.5	90.2	135.2





#49
Acenaphthylene
Concen: 38.49 ng
RT: 14.010 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:152 Resp: 141982

Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.2	24.4
153	13.0	10.2	15.4



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#50
Dimethylphthalate
Concen: 38.75 ng
RT: 13.751 min Scan# 1813
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:163 Resp: 1155211

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.4	5.0
164	10.4	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 39.32 ng
RT: 13.875 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:165 Resp: 24753

Ion	Ratio	Lower	Upper
165	100		
63	53.5	42.4	63.6
89	56.6	44.8	67.2



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#52
Acenaphthene
Concen: 37.90 ng
RT: 14.357 min Scan# 1916
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:154 Resp: 803755

Ion	Ratio	Lower	Upper
154	100		
153	114.5	91.0	136.4
152	54.0	42.9	64.3





#53
3-Nitroaniline
Concen: 40.87 ng
RT: 14.210 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:138 Resp: 26273

Ion	Ratio	Lower	Upper
138	100		
108	11.4	9.0	13.4
92	112.7	90.1	135.1



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#54
2,4-Dinitrophenol
Concen: 37.83 ng
RT: 14.428 min Scan# 1928
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:184 Resp: 135624

Ion	Ratio	Lower	Upper
184	100		
63	59.3	51.3	76.9
154	68.5	53.5	80.3





#55
Dibenzofuran
Concen: 37.91 ng
RT: 14.698 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:168 Resp: 128222

Ion	Ratio	Lower	Upper
168	100		
139	42.1	34.2	51.4
169	13.5	10.8	16.2



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#56
4-Nitrophenol
Concen: 38.70 ng
RT: 14.545 min Scan# 1948
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:139 Resp: 202528

Ion	Ratio	Lower	Upper
139	100		
109	83.2	61.4	101.4
65	96.3	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 39.97 ng
RT: 14.675 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:165 Resp: 35481

Ion	Ratio	Lower	Upper
165	100		
63	40.7	33.8	50.6
89	68.7	55.6	83.4
182	3.0	2.4	3.6



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#58
Fluorene
Concen: 38.42 ng
RT: 15.357 min Scan# 2086
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:166 Resp: 1059642

Ion	Ratio	Lower	Upper
166	100		
165	98.2	78.5	117.7
167	13.8	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 39.27 ng
RT: 14.934 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:232 Resp: 29832

Ion	Ratio	Lower	Upper
232	100		
131	48.9	39.5	59.3
130	2.8	2.2	3.4
166	32.5	25.7	38.5



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#60
Diethylphthalate
Concen: 38.17 ng
RT: 15.134 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:149 Resp: 1150091

Ion	Ratio	Lower	Upper
149	100		
177	21.0	16.7	25.1
150	12.9	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 38.25 ng
RT: 15.357 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:204 Resp: 52606

Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.0	39.0
141	66.6	52.9	79.3



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#62
4-Nitroaniline
Concen: 42.72 ng
RT: 15.398 min Scan# 2093
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:138 Resp: 265185

Ion	Ratio	Lower	Upper
138	100		
92	48.8	29.2	69.2
108	87.1	69.0	109.0





#63
Azobenzene
Concen: 38.82 ng
RT: 15.657 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion: 77 Resp: 100065

Ion	Ratio	Lower	Upper
77	100		
182	27.9	5.7	45.7
105	21.9	1.3	41.3
51	29.7	10.6	50.6



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.092 min Scan# 2381
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:188 Resp: 780511

Ion	Ratio	Lower	Upper
188	100		
94	8.0	6.5	9.7
80	9.2	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 40.30 ng
RT: 15.451 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:198 Resp: 20320

Ion	Ratio	Lower	Upper
198	100		
51	43.2	21.1	61.1
105	43.8	23.6	63.6

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025





#66
n-Nitrosodiphenylamine
Concen: 38.38 ng
RT: 15.575 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:169 Resp: 920187

Ion	Ratio	Lower	Upper
169	100		
168	66.7	53.2	79.8
167	36.1	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 38.91 ng
RT: 16.269 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:248 Resp: 35784

Ion	Ratio	Lower	Upper
248	100		
250	99.3	78.2	117.4
141	74.1	63.0	94.6



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#68
Hexachlorobenzene
Concen: 38.18 ng
RT: 16.386 min Scan# 2261
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:284 Resp: 446142

Ion	Ratio	Lower	Upper
284	100		
142	31.6	25.9	38.9
249	29.3	24.1	36.1





#69
Atrazine
Concen: 39.57 ng
RT: 16.551 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:200 Resp: 34766

Ion	Ratio	Lower	Upper
200	100		
173	26.4	5.9	45.9
215	47.6	28.3	68.3



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#70
Pentachlorophenol
Concen: 41.12 ng
RT: 16.739 min Scan# 2321
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:266 Resp: 269781

Ion	Ratio	Lower	Upper
266	100		
268	63.7	52.2	78.4
264	63.6	51.1	76.7





#71
Phenanthrene
Concen: 37.55 ng
RT: 17.133 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:178 Resp: 162911

Ion	Ratio	Lower	Upper
178	100		
176	19.7	15.4	23.2
179	15.4	12.2	18.4



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#72
Anthracene
Concen: 38.24 ng
RT: 17.228 min Scan# 2404
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:178 Resp: 1669395

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	15.5	12.2	18.2





#73
Carbazole
Concen: 39.57 ng
RT: 17.510 min Scan# 24
Delta R.T. -0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:167 Resp: 156368

Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.7	26.5
139	14.2	11.3	16.9



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#74
Di-n-butylphthalate
Concen: 38.94 ng
RT: 18.098 min Scan# 2552
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:149 Resp: 1978741

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.2	10.8
104	5.4	4.3	6.5





#75
Fluoranthene
Concen: 38.57 ng
RT: 19.222 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:202 Resp: 195604

Ion	Ratio	Lower	Upper
202	100		
101	10.0	0.0	30.3
203	17.4	0.0	37.5



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.521 min Scan# 3134
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:240 Resp: 868822

Ion	Ratio	Lower	Upper
240	100		
120	9.7	7.8	11.6
236	25.2	20.6	31.0





#77
Benzidine
Concen: 43.25 ng
RT: 19.422 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:184 Resp: 102351

Ion	Ratio	Lower	Upper
184	100		
185	14.9	11.2	16.8
183	12.3	9.8	14.8



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#78
Pyrene
Concen: 38.28 ng
RT: 19.604 min Scan# 2808
Delta R.T. 0.012 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:202 Resp: 1992845

Ion	Ratio	Lower	Upper
202	100		
200	21.3	16.9	25.3
203	17.5	13.9	20.9





#79
Terphenyl-d14
Concen: 76.75 ng
RT: 19.822 min Scan# 2
Delta R.T. 0.012 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:244 Resp: 388900

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.5	8.3
122	9.0	7.4	11.0



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#80
Butylbenzylphthalate
Concen: 40.52 ng
RT: 20.557 min Scan# 2970
Delta R.T. 0.018 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:149 Resp: 938210

Ion	Ratio	Lower	Upper
149	100		
91	65.7	55.4	83.0
206	21.5	16.2	24.2





#81
Benzo(a)anthracene
Concen: 38.89 ng
RT: 21.510 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:228 Resp: 212156

Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.4	32.0
229	19.6	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#82
3,3'-Dichlorobenzidine
Concen: 42.51 ng
RT: 21.433 min Scan# 3119
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:252 Resp: 861323

Ion	Ratio	Lower	Upper
252	100		
254	64.4	51.8	77.6
126	10.1	7.7	11.5





#83
Chrysene
Concen: 38.23 ng
RT: 21.574 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:228 Resp: 197694

Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.9	35.9
229	19.9	15.5	23.3



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.43 ng
RT: 21.445 min Scan# 3121
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:149 Resp: 1375919

Ion	Ratio	Lower	Upper
149	100		
167	27.6	22.1	33.1
279	5.7	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 41.72 ng
RT: 22.698 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:149 Resp: 232167

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.8	6.3	9.5



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#86
Perylene-d12
Concen: 20.00 ng
RT: 24.815 min Scan# 3694
Delta R.T. 0.012 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:264 Resp: 1015341

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.7	28.1
265	22.0	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 38.85 ng m
RT: 28.515 min Scan# 4
Delta R.T. -0.000 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:276 Resp: 287986

Ion	Ratio	Lower	Upper
276	100		
138	14.2	11.2	16.8
277	10.6	15.7	23.5



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#88
Benzo(b)fluoranthene
Concen: 39.29 ng
RT: 23.768 min Scan# 3516
Delta R.T. 0.012 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:252 Resp: 2283322

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	9.0	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 38.22 ng
RT: 23.833 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:252 Resp: 228900

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	8.9	7.0	10.4



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#90
Benzo(a)pyrene
Concen: 39.00 ng
RT: 24.668 min Scan# 3669
Delta R.T. 0.023 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:252 Resp: 2177020

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	9.7	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 39.38 ng
RT: 28.615 min Scan# 4
Delta R.T. 0.018 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:278 Resp: 237455

Ion	Ratio	Lower	Upper
278	100		
139	12.4	10.2	15.4
279	23.8	19.0	28.4



Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/14/2025
Supervised By :Jagrut Upadhyay 05/14/2025



#92
Benzo(g,h,i)perylene
Concen: 38.91 ng
RT: 29.686 min Scan# 4522
Delta R.T. 0.023 min
Lab File: BP024622.D
Acq: 13 May 2025 17:01

Tgt Ion:276 Resp: 2333151

Ion	Ratio	Lower	Upper
276	100		
277	23.6	19.0	28.4
138	18.4	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2		1,4-Dioxane	0.549	0.505	8.0	119	0.00
3		Pyridine	1.218	1.203	1.2	120	0.00
4		n-Nitrosodimethylamine	0.538	0.510	5.2	118	0.00
5	S	2-Fluorophenol	1.169	1.139	2.6	123	0.00
6		Aniline	1.927	1.889	2.0	118	0.00
7	S	Phenol-d6	1.492	1.475	1.1	120	0.00
8		2-Chlorophenol	1.327	1.306	1.6	120	0.00
9		Benzaldehyde	0.966	0.943	2.4	118	0.00
10	C	Phenol	1.540	1.505	2.3	119	0.00
11		bis(2-Chloroethyl)ether	1.266	1.200	5.2	118	0.00
12		1,3-Dichlorobenzene	1.527	1.447	5.2	120	0.00
13	C	1,4-Dichlorobenzene	1.536	1.456	5.2	118	0.00
14		1,2-Dichlorobenzene	1.497	1.418	5.3	119	0.00
15		Benzyl Alcohol	1.134	1.149	-1.3	119	0.00
16		2,2'-oxybis(1-Chloropropane	1.516	1.441	4.9	117	0.00
17		2-Methylphenol	1.105	1.094	1.0	119	0.00
18		Hexachloroethane	0.589	0.548	7.0	119	0.00
19	P	n-Nitroso-di-n-propylamine	1.032	1.011	2.0	115	0.00
20		3+4-Methylphenols	1.503	1.486	1.1	117	0.00
21	I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
22		Acetophenone	0.500	0.480	4.0	115	0.00
23	S	Nitrobenzene-d5	0.396	0.387	2.3	116	0.00
24		Nitrobenzene	0.348	0.335	3.7	115	0.00
25		Isophorone	0.687	0.661	3.8	113	0.00
26	C	2-Nitrophenol	0.171	0.175	-2.3	118	0.00
27		2,4-Dimethylphenol	0.301	0.297	1.3	115	0.00
28		bis(2-Chloroethoxy)methane	0.410	0.397	3.2	115	0.00
29	C	2,4-Dichlorophenol	0.292	0.293	-0.3	116	0.00
30		1,2,4-Trichlorobenzene	0.329	0.313	4.9	118	0.00
31		Naphthalene	1.036	0.986	4.8	116	0.00
32		Benzoic acid	0.181	0.180	0.6	108	0.00
33		4-Chloroaniline	0.418	0.422	-1.0	116	0.00
34	C	Hexachlorobutadiene	0.213	0.198	7.0	115	0.00
35		Caprolactam	0.106	0.105	0.9	110	0.00
36	C	4-Chloro-3-methylphenol	0.356	0.347	2.5	112	0.00
37		2-Methylnaphthalene	0.671	0.635	5.4	114	0.00
38		1-Methylnaphthalene	0.718	0.670	6.7	112	0.00
39	I	Acenaphthene-d10	1.000	1.000	0.0	108	-0.01
40		1,2,4,5-Tetrachlorobenzene	0.581	0.568	2.2	112	0.00
41	P	Hexachlorocyclopentadiene	0.352	0.391	-11.1	118	0.00
42	S	2,4,6-Tribromophenol	0.339	0.333	1.8	111	0.00
43	C	2,4,6-Trichlorophenol	0.376	0.391	-4.0	114	0.00
44		2,4,5-Trichlorophenol	0.419	0.430	-2.6	115	0.00
45	S	2-Fluorobiphenyl	1.527	1.464	4.1	110	0.00
46		1,1'-Biphenyl	1.478	1.438	2.7	112	0.00
47		2-Chloronaphthalene	1.127	1.106	1.9	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.334	0.339	-1.5	109	0.00
49		Acenaphthylene	1.886	1.815	3.8	109	0.00
50		Dimethylphthalate	1.524	1.477	3.1	110	0.00
51		2,6-Dinitrotoluene	0.322	0.316	1.9	109	0.00
52	C	Acenaphthene	1.084	1.028	5.2	108	0.00
53		3-Nitroaniline	0.329	0.336	-2.1	107	0.00
54	P	2,4-Dinitrophenol	0.169	0.173	-2.4	109	0.00
55		Dibenzofuran	1.730	1.639	5.3	108	0.00
56	P	4-Nitrophenol	0.250	0.259	-3.6	109	0.00
57		2,4-Dinitrotoluene	0.454	0.454	0.0	108	0.00
58		Fluorene	1.410	1.355	3.9	109	0.00
59		2,3,4,6-Tetrachlorophenol	0.388	0.381	1.8	109	0.00
60		Diethylphthalate	1.541	1.470	4.6	110	0.00
61		4-Chlorophenyl-phenylether	0.703	0.673	4.3	111	0.00
62		4-Nitroaniline	0.317	0.339	-6.9	111	0.00
63		Azobenzene	1.318	1.279	3.0	110	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
65		4,6-Dinitro-2-methylphenol	0.129	0.130	-0.8	110	0.00
66	c	n-Nitrosodiphenylamine	0.614	0.589	4.1	108	0.00
67		4-Bromophenyl-phenylether	0.236	0.229	3.0	111	0.00
68		Hexachlorobenzene	0.299	0.286	4.3	110	0.00
69		Atrazine	0.225	0.223	0.9	111	0.00
70	C	Pentachlorophenol	0.168	0.173	-3.0	109	0.00
71		Phenanthrene	1.112	1.044	6.1	108	0.00
72		Anthracene	1.119	1.069	4.5	108	0.00
73		Carbazole	1.012	1.002	1.0	109	0.00
74		Di-n-butylphthalate	1.302	1.268	2.6	109	0.00
75	C	Fluoranthene	1.300	1.253	3.6	109	0.00
76	I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77		Benzidine	0.545	0.589	-8.1	113	0.00
78		Pyrene	1.198	1.147	4.3	107	0.01
79	S	Terphenyl-d14	1.166	1.119	4.0	107	0.01
80		Butylbenzylphthalate	0.533	0.540	-1.3	110	0.02
81		Benzo(a)anthracene	1.256	1.221	2.8	108	0.00
82		3,3'-Dichlorobenzidine	0.466	0.496	-6.4	111	0.00
83		Chrysene	1.190	1.138	4.4	108	0.00
84		Bis(2-ethylhexyl)phthalate	0.783	0.792	-1.1	111	0.00
85	c	Di-n-octyl phthalate	1.281	1.336	-4.3	112	0.00
86	I	Perylene-d12	1.000	1.000	0.0	109	0.01
87		Indeno(1,2,3-cd)pyrene	1.460	1.418	2.9	109	0.00
88		Benzo(b)fluoranthene	1.145	1.124	1.8	111	0.01
89		Benzo(k)fluoranthene	1.180	1.127	4.5	108	0.00
90	C	Benzo(a)pyrene	1.099	1.072	2.5	110	0.02
91		Dibenzo(a,h)anthracene	1.188	1.169	1.6	109	0.02
92		Benzo(g,h,i)perylene	1.181	1.149	2.7	109	0.02

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	119	0.00
2		1,4-Dioxane	40.000	36.789	8.0	119	0.00
3		Pyridine	40.000	39.526	1.2	120	0.00
4		n-Nitrosodimethylamine	40.000	37.979	5.1	118	0.00
5	S	2-Fluorophenol	80.000	77.955	2.6	123	0.00
6		Aniline	40.000	39.210	2.0	118	0.00
7	S	Phenol-d6	80.000	79.089	1.1	120	0.00
8		2-Chlorophenol	40.000	39.367	1.6	120	0.00
9		Benzaldehyde	40.000	39.033	2.4	118	0.00
10	C	Phenol	40.000	39.089	2.3	119	0.00
11		bis(2-Chloroethyl)ether	40.000	37.938	5.2	118	0.00
12		1,3-Dichlorobenzene	40.000	37.893	5.3	120	0.00
13	C	1,4-Dichlorobenzene	40.000	37.904	5.2	118	0.00
14		1,2-Dichlorobenzene	40.000	37.896	5.3	119	0.00
15		Benzyl Alcohol	40.000	40.541	-1.4	119	0.00
16		2,2'-oxybis(1-Chloropropane	40.000	38.037	4.9	117	0.00
17		2-Methylphenol	40.000	39.618	1.0	119	0.00
18		Hexachloroethane	40.000	37.211	7.0	119	0.00
19	P	n-Nitroso-di-n-propylamine	40.000	39.165	2.1	115	0.00
20		3+4-Methylphenols	40.000	39.562	1.1	117	0.00
21	I	Naphthalene-d8	20.000	20.000	0.0	115	0.00
22		Acetophenone	40.000	38.468	3.8	115	0.00
23	S	Nitrobenzene-d5	80.000	78.227	2.2	116	0.00
24		Nitrobenzene	40.000	38.505	3.7	115	0.00
25		Isophorone	40.000	38.491	3.8	113	0.00
26	C	2-Nitrophenol	40.000	41.121	-2.8	118	0.00
27		2,4-Dimethylphenol	40.000	39.484	1.3	115	0.00
28		bis(2-Chloroethoxy)methane	40.000	38.714	3.2	115	0.00
29	C	2,4-Dichlorophenol	40.000	40.189	-0.5	116	0.00
30		1,2,4-Trichlorobenzene	40.000	38.109	4.7	118	0.00
31		Naphthalene	40.000	38.069	4.8	116	0.00
32		Benzoic acid	40.000	36.464	8.8	108	0.00
33		4-Chloroaniline	40.000	40.368	-0.9	116	0.00
34	C	Hexachlorobutadiene	40.000	37.285	6.8	115	0.00
35		Caprolactam	40.000	39.752	0.6	110	0.00
36	C	4-Chloro-3-methylphenol	40.000	39.013	2.5	112	0.00
37		2-Methylnaphthalene	40.000	37.852	5.4	114	0.00
38		1-Methylnaphthalene	40.000	37.339	6.7	112	0.00
39	I	Acenaphthene-d10	20.000	20.000	0.0	108	-0.01
40		1,2,4,5-Tetrachlorobenzene	40.000	39.130	2.2	112	0.00
41	P	Hexachlorocyclopentadiene	40.000	44.371	-10.9	118	0.00
42	S	2,4,6-Tribromophenol	80.000	78.525	1.8	111	0.00
43	C	2,4,6-Trichlorophenol	40.000	41.549	-3.9	114	0.00
44		2,4,5-Trichlorophenol	40.000	41.064	-2.7	115	0.00
45	S	2-Fluorobiphenyl	80.000	76.733	4.1	110	0.00
46		1,1'-Biphenyl	40.000	38.906	2.7	112	0.00
47		2-Chloronaphthalene	40.000	39.238	1.9	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	40.604	-1.5	109	0.00
49		Acenaphthylene	40.000	38.493	3.8	109	0.00
50		Dimethylphthalate	40.000	38.751	3.1	110	0.00
51		2,6-Dinitrotoluene	40.000	39.317	1.7	109	0.00
52	C	Acenaphthene	40.000	37.904	5.2	108	0.00
53		3-Nitroaniline	40.000	40.871	-2.2	107	0.00
54	P	2,4-Dinitrophenol	40.000	37.829	5.4	109	0.00
55		Dibenzofuran	40.000	37.910	5.2	108	0.00
56	P	4-Nitrophenol	40.000	38.698	3.3	109	0.00
57		2,4-Dinitrotoluene	40.000	39.975	0.1	108	0.00
58		Fluorene	40.000	38.422	3.9	109	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	39.274	1.8	109	0.00
60		Diethylphthalate	40.000	38.171	4.6	110	0.00
61		4-Chlorophenyl-phenylether	40.000	38.249	4.4	111	0.00
62		4-Nitroaniline	40.000	42.722	-6.8	111	0.00
63		Azobenzene	40.000	38.822	2.9	110	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	109	0.00
65		4,6-Dinitro-2-methylphenol	40.000	40.305	-0.8	110	0.00
66	c	n-Nitrosodiphenylamine	40.000	38.382	4.0	108	0.00
67		4-Bromophenyl-phenylether	40.000	38.909	2.7	111	0.00
68		Hexachlorobenzene	40.000	38.175	4.6	110	0.00
69		Atrazine	40.000	39.570	1.1	111	0.00
70	C	Pentachlorophenol	40.000	41.124	-2.8	109	0.00
71		Phenanthrene	40.000	37.546	6.1	108	0.00
72		Anthracene	40.000	38.241	4.4	108	0.00
73		Carbazole	40.000	39.575	1.1	109	0.00
74		Di-n-butylphthalate	40.000	38.939	2.7	109	0.00
75	C	Fluoranthene	40.000	38.566	3.6	109	0.00
76	I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77		Benzidine	40.000	43.253	-8.1	113	0.00
78		Pyrene	40.000	38.281	4.3	107	0.01
79	S	Terphenyl-d14	80.000	76.753	4.1	107	0.01
80		Butylbenzylphthalate	40.000	40.523	-1.3	110	0.02
81		Benzo(a)anthracene	40.000	38.889	2.8	108	0.00
82		3,3'-Dichlorobenzidine	40.000	42.508	-6.3	111	0.00
83		Chrysene	40.000	38.227	4.4	108	0.00
84		Bis(2-ethylhexyl)phthalate	40.000	40.433	-1.1	111	0.00
85	c	Di-n-octyl phthalate	40.000	41.721	-4.3	112	0.00
86	I	Perylene-d12	20.000	20.000	0.0	109	0.01
87		Indeno(1,2,3-cd)pyrene	40.000	38.851	2.9	109	0.00
88		Benzo(b)fluoranthene	40.000	39.289	1.8	111	0.01
89		Benzo(k)fluoranthene	40.000	38.223	4.4	108	0.00
90	C	Benzo(a)pyrene	40.000	39.004	2.5	110	0.02
91		Dibenzo(a,h)anthracene	40.000	39.377	1.6	109	0.02
92		Benzo(g,h,i)perylene	40.000	38.906	2.7	109	0.02

ICAL Form

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024622.D
Acq On : 13 May 2025 17:01
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP051325

Quant Time: May 13 18:01:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG No.: Q1872

Instrument ID: BNA_M Calibration Date/Time: 05/08/2025 10:41

Lab File ID: BM050140.D Init. Calib. Date(s): 04/28/2025 04/28/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:30 17:04

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.142	1.109		-2.9	
Phenol-d6	1.436	1.425		-0.8	
Nitrobenzene-d5	0.406	0.390		-3.9	
2-Fluorobiphenyl	1.691	1.640		-3.0	
2,4,6-Tribromophenol	0.296	0.286		-3.4	
Atrazine	0.222	0.212		-4.5	
Terphenyl-d14	1.352	1.555		15.0	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.740	152	370324	20.000	ng	0.00
21) Naphthalene-d8	10.534	136	1368857	20.000	ng	0.00
39) Acenaphthene-d10	14.392	164	879874	20.000	ng	0.00
64) Phenanthrene-d10	17.139	188	1654930	20.000	ng	0.00
76) Chrysene-d12	21.386	240	1237713	20.000	ng	0.00
86) Perylene-d12	24.380	264	1026252	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.340	112	1642415	77.661	ng	0.00
7) Phenol-d6	6.934	99	2111558	79.439	ng	0.00
23) Nitrobenzene-d5	8.904	82	2134237	76.829	ng	0.00
42) 2,4,6-Tribromophenol	15.892	330	1005758	77.306	ng	0.00
45) 2-Fluorobiphenyl	13.010	172	5771543	77.598	ng	0.00
79) Terphenyl-d14	19.768	244	7699588	92.052	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.234	88	335516	37.001	ng	99
3) Pyridine	3.634	79	904531	38.824	ng	96
4) n-Nitrosodimethylamine	3.546	42	359788	39.374	ng	98
6) Aniline	7.075	93	1314457	39.162	ng	99
8) 2-Chlorophenol	7.316	128	887499	38.813	ng	98
9) Benzaldehyde	6.887	77	690756	38.833	ng	98
10) Phenol	6.963	94	1065717	39.605	ng	98
11) bis(2-Chloroethyl)ether	7.169	93	863881	38.447	ng	99
12) 1,3-Dichlorobenzene	7.634	146	1038233	37.588	ng	99
13) 1,4-Dichlorobenzene	7.775	146	1040843	37.509	ng	99
14) 1,2-Dichlorobenzene	8.093	146	1005092	37.497	ng	99
15) Benzyl Alcohol	7.993	79	751209	40.752	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.263	45	1076308	39.406	ng	99
17) 2-Methylphenol	8.204	107	693712	39.133	ng	98
18) Hexachloroethane	8.816	117	369976	37.507	ng	100
19) n-Nitroso-di-n-propyla...	8.551	70	674550	39.568	ng	100
20) 3+4-Methylphenols	8.534	107	947336	39.573	ng	98
22) Acetophenone	8.569	105	1298428	38.153	ng	98
24) Nitrobenzene	8.945	77	922534	38.408	ng	100
25) Isophorone	9.475	82	1815169	38.333	ng	100
26) 2-Nitrophenol	9.657	139	500100	40.755	ng	99
27) 2,4-Dimethylphenol	9.728	122	800845	39.354	ng	100
28) bis(2-Chloroethoxy)met...	9.951	93	1125441	38.454	ng	100
29) 2,4-Dichlorophenol	10.210	162	882444	38.708	ng	99
30) 1,2,4-Trichlorobenzene	10.404	180	1043351	37.563	ng	100
31) Naphthalene	10.586	128	2598647	37.486	ng	100
32) Benzoic acid	9.916	122	496992	38.186	ng	98
33) 4-Chloroaniline	10.704	127	1093472	38.483	ng	99
34) Hexachlorobutadiene	10.869	225	666762	37.817	ng	99
35) Caprolactam	11.516	113	246749	37.121	ng	94
36) 4-Chloro-3-methylphenol	11.857	107	803329	38.140	ng	98
37) 2-Methylnaphthalene	12.204	142	1761314	37.896	ng	100
38) 1-Methylnaphthalene	12.422	142	1840027	37.410	ng	100
40) 1,2,4,5-Tetrachloroben...	12.580	216	1207652	39.649	ng	100
41) Hexachlorocyclopentadiene	12.551	237	692945	37.227	ng	100
43) 2,4,6-Trichlorophenol	12.833	196	781494	40.452	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.922	196	827607	39.727	ng	99
46) 1,1'-Biphenyl	13.222	154	2502138	38.256	ng	99
47) 2-Chloronaphthalene	13.263	162	1978827	38.200	ng	99
48) 2-Nitroaniline	13.480	65	487178	40.071	ng	100
49) Acenaphthylene	14.116	152	3100974	37.967	ng	100
50) Dimethylphthalate	13.851	163	2380206	37.016	ng	99
51) 2,6-Dinitrotoluene	13.975	165	516023	38.760	ng	98
52) Acenaphthene	14.457	154	1780304	37.658	ng	99
53) 3-Nitroaniline	14.316	138	495431	38.978	ng	99
54) 2,4-Dinitrophenol	14.539	184	272567	35.437	ng	98
55) Dibenzofuran	14.792	168	2956841	37.591	ng	99
56) 4-Nitrophenol	14.669	139	319245	33.657	ng	98
57) 2,4-Dinitrotoluene	14.774	165	719782	39.126	ng	98
58) Fluorene	15.445	166	2401802	37.305	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.033	232	677686	37.581	ng	99
60) Diethylphthalate	15.216	149	2205320	36.397	ng	99
61) 4-Chlorophenyl-phenyle...	15.433	204	1351541	38.219	ng	99
62) 4-Nitroaniline	15.480	138	432649	35.280	ng	99
63) Azobenzene	15.727	77	1882944	37.081	ng	99
65) 4,6-Dinitro-2-methylph...	15.545	198	400737	38.684	ng	96
66) n-Nitrosodiphenylamine	15.651	169	1995726	39.272	ng	100
67) 4-Bromophenyl-phenylether	16.327	248	793425	39.646	ng	98
68) Hexachlorobenzene	16.451	284	896115	38.846	ng	98
69) Atrazine	16.610	200	701944	38.293	ng	99
70) Pentachlorophenol	16.810	266	489963	37.733	ng	99
71) Phenanthrene	17.186	178	3506508	37.567	ng	100
72) Anthracene	17.274	178	3566867	37.761	ng	100
73) Carbazole	17.551	167	3040432	37.093	ng	99
74) Di-n-butylphthalate	18.104	149	3579619	36.672	ng	99
75) Fluoranthene	19.204	202	3923349	35.801	ng	100
77) Benzidine	19.392	184	1769429	40.247	ng	99
78) Pyrene	19.568	202	3921227	45.115	ng	100
80) Butylbenzylphthalate	20.462	149	1331850	40.913	ng	99
81) Benzo(a)anthracene	21.368	228	3251627	38.150	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	1193314	36.574	ng	100
83) Chrysene	21.427	228	2971758	37.672	ng	100
84) Bis(2-ethylhexyl)phtha...	21.280	149	1827791	37.590	ng	100
85) Di-n-octyl phthalate	22.403	149	2757430	34.387	ng	100
87) Indeno(1,2,3-cd)pyrene	27.762	276	2903047	37.964	ng	100
88) Benzo(b)fluoranthene	23.433	252	2548858	38.191	ng	100
89) Benzo(k)fluoranthene	23.497	252	2625481	38.890	ng	99
90) Benzo(a)pyrene	24.239	252	2381402	38.097	ng	99
91) Dibenzo(a,h)anthracene	27.809	278	2367754	37.985	ng	100
92) Benzo(g,h,i)perylene	28.826	276	2251454	37.727	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.740 min Scan# 80
Delta R.T. -0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 370324

Ion	Ratio	Lower	Upper
152	100		
150	154.3	124.1	186.1
115	53.2	41.2	61.8





#2
1,4-Dioxane
Concen: 37.001 ng
RT: 3.234 min Scan# 42
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 88 Resp: 335516

Ion	Ratio	Lower	Upper
88	100		
58	69.9	55.1	82.7
43	25.9	20.9	31.3





#3
Pyridine
Concen: 38.824 ng
RT: 3.634 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 904531

Ion	Ratio	Lower	Upper
79	100		
52	63.6	48.6	72.8
51	32.9	24.6	37.0





#4
n-Nitrosodimethylamine
Concen: 39.374 ng
RT: 3.546 min Scan# 95
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 42 Resp: 359788

Ion	Ratio	Lower	Upper
42	100		
74	148.1	120.2	180.2
44	7.0	6.2	9.2





#5
2-Fluorophenol
Concen: 77.661 ng
RT: 5.340 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 1642415

Ion	Ratio	Lower	Upper
112	100		
64	52.6	42.6	64.0
63	27.4	22.2	33.4





#6
Aniline
Concen: 39.162 ng
RT: 7.075 min Scan# 695
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 93 Resp: 1314457

Ion	Ratio	Lower	Upper
93	100		
66	38.0	29.9	44.9
65	19.0	14.5	21.7





#7
Phenol-d6
Concen: 79.439 ng
RT: 6.934 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 2111558

Ion	Ratio	Lower	Upper
99	100		
42	15.9	13.0	19.4
71	32.7	25.6	38.4





#8
2-Chlorophenol
Concen: 38.813 ng
RT: 7.316 min Scan# 736
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:128 Resp: 887499

Ion	Ratio	Lower	Upper
128	100		
130	33.0	12.8	52.8
64	42.8	21.2	61.2





#9
Benzaldehyde
Concen: 38.833 ng
RT: 6.887 min Scan# 6
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 690756

Ion	Ratio	Lower	Upper
77	100		
105	98.8	80.5	120.5
106	94.3	76.1	116.1





#10
Phenol
Concen: 39.605 ng
RT: 6.963 min Scan# 676
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 94 Resp: 1065717

Ion	Ratio	Lower	Upper
94	100		
65	26.9	7.6	47.6
66	33.9	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 38.447 ng
RT: 7.169 min Scan# 7
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 863881

Ion	Ratio	Lower	Upper
93	100		
63	73.6	52.3	92.3
95	33.0	12.8	52.8





#12
1,3-Dichlorobenzene
Concen: 37.588 ng
RT: 7.634 min Scan# 790
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:146 Resp: 1038233

Ion	Ratio	Lower	Upper
146	100		
148	65.9	52.5	78.7
75	27.3	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 37.509 ng
RT: 7.775 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 1040843

Ion	Ratio	Lower	Upper
146	100		
148	65.3	52.2	78.2
111	37.7	29.3	43.9





#14
1,2-Dichlorobenzene
Concen: 37.497 ng
RT: 8.093 min Scan# 868
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:146 Resp: 1005092

Ion	Ratio	Lower	Upper
146	100		
148	64.5	52.4	78.6
111	38.8	31.0	46.6





#15
Benzyl Alcohol
Concen: 40.752 ng
RT: 7.993 min Scan# 8
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 751209

Ion	Ratio	Lower	Upper
79	100		
108	78.9	66.0	99.0
77	64.7	51.6	77.4





#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.406 ng
RT: 8.263 min Scan# 897
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 45 Resp: 1076308

Ion	Ratio	Lower	Upper
45	100		
77	16.4	0.0	36.4
79	12.1	0.0	32.7





#17
2-Methylphenol
Concen: 39.133 ng
RT: 8.204 min Scan# 8
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 693712

Ion	Ratio	Lower	Upper
107	100		
108	116.0	91.1	136.7
77	46.6	36.2	54.4
79	45.2	35.8	53.8





#18
Hexachloroethane
Concen: 37.507 ng
RT: 8.816 min Scan# 991
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:117 Resp: 369976

Ion	Ratio	Lower	Upper
117	100		
119	96.7	77.3	115.9
201	128.0	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 39.568 ng
RT: 8.551 min Scan# 94
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 674550

Ion	Ratio	Lower	Upper
70	100		
42	50.4	40.5	60.7
101	9.8	7.8	11.6
130	21.8	17.4	26.2





#20
3+4-Methylphenols
Concen: 39.573 ng
RT: 8.534 min Scan# 943
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:107 Resp: 947336

Ion	Ratio	Lower	Upper
107	100		
108	90.3	72.4	112.4
77	33.7	13.6	53.6
79	27.6	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.534 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 1368857

Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	6.8	5.1	7.7
68	5.7	4.4	6.6





#22
Acetophenone
Concen: 38.153 ng
RT: 8.569 min Scan# 949
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:105 Resp: 1298428

Ion	Ratio	Lower	Upper
105	100		
71	2.4	1.7	2.5
51	24.4	18.6	27.8
120	23.5	18.3	27.5





#23
Nitrobenzene-d5
Concen: 76.829 ng
RT: 8.904 min Scan# 1(
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 2134237

Ion	Ratio	Lower	Upper
82	100		
128	45.2	35.7	53.5
54	44.7	35.4	53.2





#24
Nitrobenzene
Concen: 38.408 ng
RT: 8.945 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 77 Resp: 922534

Ion	Ratio	Lower	Upper
77	100		
123	47.6	37.9	56.9
65	12.9	10.1	15.1





#25
Isophorone
Concen: 38.333 ng
RT: 9.475 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 1815169

Ion	Ratio	Lower	Upper
82	100		
95	7.0	5.7	8.5
138	19.3	15.5	23.3





#26
2-Nitrophenol
Concen: 40.755 ng
RT: 9.657 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:139 Resp: 500100

Ion	Ratio	Lower	Upper
139	100		
109	26.6	21.9	32.9
65	44.1	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 39.354 ng
RT: 9.728 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 800845

Ion	Ratio	Lower	Upper
122	100		
107	111.6	89.0	133.6
121	56.9	45.6	68.4





#28
bis(2-Chloroethoxy)methane
Concen: 38.454 ng
RT: 9.951 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 93 Resp: 1125441

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.1	39.1
123	11.9	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 38.708 ng
RT: 10.210 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 882444

Ion	Ratio	Lower	Upper
162	100		
164	65.4	45.4	85.4
98	30.2	11.3	51.3





#30
1,2,4-Trichlorobenzene
Concen: 37.563 ng
RT: 10.404 min Scan# 1261
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:180 Resp: 1043351

Ion	Ratio	Lower	Upper
180	100		
182	96.0	77.0	115.4
145	25.3	20.6	30.8





#31
Naphthalene
Concen: 37.486 ng
RT: 10.586 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 2598647

Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.8	13.2
127	12.9	10.4	15.6





#32
Benzoic acid
Concen: 38.186 ng
RT: 9.916 min Scan# 1178
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:122 Resp: 496992

Ion	Ratio	Lower	Upper
122	100		
105	120.6	103.6	143.6
77	93.4	73.7	113.7





#33
4-Chloroaniline
Concen: 38.483 ng
RT: 10.704 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 1093472

Ion	Ratio	Lower	Upper
127	100		
129	32.1	26.3	39.5
65	28.4	22.5	33.7
92	17.8	14.5	21.7





#34
Hexachlorobutadiene
Concen: 37.817 ng
RT: 10.869 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:225 Resp: 666762

Ion	Ratio	Lower	Upper
225	100		
223	63.3	50.2	75.2
227	64.0	50.6	76.0





#35
Caprolactam
Concen: 37.121 ng
RT: 11.516 min Scan# 14
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 246749

Ion	Ratio	Lower	Upper
113	100		
55	150.2	122.9	162.9
56	115.2	88.1	128.1





#36
4-Chloro-3-methylphenol
Concen: 38.140 ng
RT: 11.857 min Scan# 1508
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:107 Resp: 803329

Ion	Ratio	Lower	Upper
107	100		
144	29.5	23.8	35.6
142	92.1	71.9	107.9





#37
2-Methylnaphthalene
Concen: 37.896 ng
RT: 12.204 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 1761314

Ion	Ratio	Lower	Upper
142	100		
141	87.6	70.4	105.6
115	31.8	25.4	38.2





#38
1-Methylnaphthalene
Concen: 37.410 ng
RT: 12.422 min Scan# 1604
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:142 Resp: 1840027

Ion	Ratio	Lower	Upper
142	100		
141	89.9	72.1	108.1
116	3.6	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.392 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 879874

Ion	Ratio	Lower	Upper
164	100		
162	99.9	79.4	119.0
160	44.6	36.2	54.4





#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.649 ng
RT: 12.580 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:216 Resp: 1207652

Ion	Ratio	Lower	Upper
216	100		
214	78.3	62.3	93.5
179	19.2	15.4	23.0
108	14.3	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 37.227 ng
RT: 12.551 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 692945

Ion	Ratio	Lower	Upper
237	100		
235	62.5	42.9	82.9
272	13.0	0.0	32.9





#42
2,4,6-Tribromophenol
Concen: 77.306 ng
RT: 15.892 min Scan# 2194
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:330 Resp: 1005758

Ion	Ratio	Lower	Upper
330	100		
332	97.1	77.3	115.9
141	26.5	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 40.452 ng
RT: 12.833 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 781494

Ion	Ratio	Lower	Upper
196	100		
198	96.4	79.0	118.6
200	31.4	25.1	37.7





#44
2,4,5-Trichlorophenol
Concen: 39.727 ng
RT: 12.922 min Scan# 1689
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:196 Resp: 827607

Ion	Ratio	Lower	Upper
196	100		
198	97.6	76.7	115.1
97	38.2	30.1	45.1
132	21.1	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 77.598 ng
RT: 13.010 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 5771543

Ion	Ratio	Lower	Upper
172	100		
171	34.8	27.8	41.6
170	23.3	18.6	28.0





#46
1,1'-Biphenyl
Concen: 38.256 ng
RT: 13.222 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:154 Resp: 2502138

Ion	Ratio	Lower	Upper
154	100		
153	39.3	20.0	60.0
76	10.4	0.0	30.3





#47
2-Chloronaphthalene
Concen: 38.200 ng
RT: 13.263 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 1978827

Ion	Ratio	Lower	Upper
162	100		
127	33.2	25.9	38.9
164	33.5	26.6	40.0





#48
2-Nitroaniline
Concen: 40.071 ng
RT: 13.480 min Scan# 1784
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion: 65 Resp: 487178

Ion	Ratio	Lower	Upper
65	100		
92	71.0	57.1	85.7
138	121.5	97.0	145.4





#49
Acenaphthylene
Concen: 37.967 ng
RT: 14.116 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 3100974

Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.2	10.6	15.8





#50
Dimethylphthalate
Concen: 37.016 ng
RT: 13.851 min Scan# 1847
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:163 Resp: 2380206

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.2
164	10.4	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 38.760 ng
RT: 13.975 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 516023

Ion	Ratio	Lower	Upper
165	100		
63	45.9	35.1	52.7
89	47.8	38.4	57.6





#52
Acenaphthene
Concen: 37.658 ng
RT: 14.457 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:154 Resp: 1780304

Ion	Ratio	Lower	Upper
154	100		
153	112.6	89.0	133.4
152	53.3	43.0	64.6





#53
3-Nitroaniline
Concen: 38.978 ng
RT: 14.316 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 495431

Ion	Ratio	Lower	Upper
138	100		
108	9.4	7.4	11.2
92	103.5	84.0	126.0





#54
2,4-Dinitrophenol
Concen: 35.437 ng
RT: 14.539 min Scan# 1964
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:184 Resp: 272567

Ion	Ratio	Lower	Upper
184	100		
63	51.8	39.8	59.6
154	56.2	46.1	69.1





#55
Dibenzofuran
Concen: 37.591 ng
RT: 14.792 min Scan# 2(
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 2956841

Ion	Ratio	Lower	Upper
168	100		
139	32.3	26.1	39.1
169	13.3	10.8	16.2





#56
4-Nitrophenol
Concen: 33.657 ng
RT: 14.669 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:139 Resp: 319245

Ion	Ratio	Lower	Upper
139	100		
109	68.6	47.0	87.0
65	86.7	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 39.126 ng
RT: 14.774 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 719782

Ion	Ratio	Lower	Upper
165	100		
63	38.8	29.0	43.6
89	62.7	49.4	74.2
182	2.7	2.2	3.2





#58
Fluorene
Concen: 37.305 ng
RT: 15.445 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:166 Resp: 2401802

Ion	Ratio	Lower	Upper
166	100		
165	96.7	76.6	115.0
167	13.5	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 37.581 ng
RT: 15.033 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 677686

Ion	Ratio	Lower	Upper
232	100		
131	35.4	28.2	42.4
130	1.9	1.5	2.3
166	27.3	22.6	34.0





#60
Diethylphthalate
Concen: 36.397 ng
RT: 15.216 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:149 Resp: 2205320

Ion	Ratio	Lower	Upper
149	100		
177	24.6	19.1	28.7
150	12.4	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 38.219 ng
RT: 15.433 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 1351541

Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.3	39.5
141	43.3	34.0	51.0





#62
4-Nitroaniline
Concen: 35.280 ng
RT: 15.480 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:138 Resp: 432649

Ion	Ratio	Lower	Upper
138	100		
92	46.5	25.9	65.9
108	73.4	52.4	92.4





#63
Azobenzene
Concen: 37.081 ng
RT: 15.727 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 1882944

Ion	Ratio	Lower	Upper
77	100		
182	30.7	10.5	50.5
105	21.6	1.1	41.1
51	26.2	6.3	46.3





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2406
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:188 Resp: 1654930

Ion	Ratio	Lower	Upper
188	100		
94	7.3	5.7	8.5
80	7.9	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 38.684 ng
RT: 15.545 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 400737

Ion	Ratio	Lower	Upper
198	100		
51	33.6	10.4	50.4
105	36.0	14.6	54.6





#66
n-Nitrosodiphenylamine
Concen: 39.272 ng
RT: 15.651 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:169 Resp: 1995726

Ion	Ratio	Lower	Upper
169	100		
168	66.6	53.4	80.2
167	36.3	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 39.646 ng
RT: 16.327 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 793425

Ion	Ratio	Lower	Upper
248	100		
250	96.8	76.2	114.2
141	54.6	45.9	68.9





#68
Hexachlorobenzene
Concen: 38.846 ng
RT: 16.451 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:284 Resp: 896115

Ion	Ratio	Lower	Upper
284	100		
142	27.9	23.9	35.9
249	31.0	24.5	36.7





#69
Atrazine
Concen: 38.293 ng
RT: 16.610 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 701944

Ion	Ratio	Lower	Upper
200	100		
173	24.8	4.3	44.3
215	50.9	30.6	70.6





#70
Pentachlorophenol
Concen: 37.733 ng
RT: 16.810 min Scan# 2350
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:266 Resp: 489963

Ion	Ratio	Lower	Upper
266	100		
268	62.9	51.2	76.8
264	60.9	49.6	74.4





#71
Phenanthrene
Concen: 37.567 ng
RT: 17.186 min Scan# 24
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 3506508

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.3	22.9
179	15.4	12.4	18.6





#72
Anthracene
Concen: 37.761 ng
RT: 17.274 min Scan# 2429
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:178 Resp: 3566867

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.7	12.2	18.4





#73
Carbazole
Concen: 37.093 ng
RT: 17.551 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 3040432

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.2	25.8
139	11.1	9.0	13.6





#74
Di-n-butylphthalate
Concen: 36.672 ng
RT: 18.104 min Scan# 2570
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:149 Resp: 3579619

Ion	Ratio	Lower	Upper
149	100		
150	9.5	7.3	10.9
104	4.6	3.8	5.6





#75
Fluoranthene
Concen: 35.801 ng
RT: 19.204 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 3923349

Ion	Ratio	Lower	Upper
202	100		
101	7.6	0.0	28.0
203	17.7	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:240 Resp: 1237713

Ion	Ratio	Lower	Upper
240	100		
120	8.2	6.5	9.7
236	26.1	20.9	31.3





#77
Benzidine
Concen: 40.247 ng
RT: 19.392 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 1769429

Ion	Ratio	Lower	Upper
184	100		
185	15.0	11.5	17.3
183	11.3	8.9	13.3





#78
Pyrene
Concen: 45.115 ng
RT: 19.568 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:202 Resp: 3921227

Ion	Ratio	Lower	Upper
202	100		
200	21.2	16.7	25.1
203	17.5	14.1	21.1





#79
Terphenyl-d14
Concen: 92.052 ng
RT: 19.768 min Scan# 2
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 7699588

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.6	8.4
122	6.5	5.6	8.4





#80
Butylbenzylphthalate
Concen: 40.913 ng
RT: 20.462 min Scan# 2971
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:149 Resp: 1331850

Ion	Ratio	Lower	Upper
149	100		
91	60.0	47.4	71.2
206	24.3	19.4	29.0





#81
Benzo(a)anthracene
Concen: 38.150 ng
RT: 21.368 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 3251627

Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.1	33.1
229	19.9	15.7	23.5





#82
3,3'-Dichlorobenzidine
Concen: 36.574 ng
RT: 21.292 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:252 Resp: 1193314

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.8	77.6
126	8.0	6.0	9.0





#83
Chrysene
Concen: 37.672 ng
RT: 21.427 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 2971758

Ion	Ratio	Lower	Upper
228	100		
226	30.5	24.2	36.2
229	19.8	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 37.590 ng
RT: 21.280 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:149 Resp: 1827791

Ion	Ratio	Lower	Upper
149	100		
167	30.2	24.1	36.1
279	5.0	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 34.387 ng
RT: 22.403 min Scan# 3
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 2757430

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.4	5.8	8.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.380 min Scan# 3637
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:264 Resp: 1026252

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.8	28.2
265	22.1	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 37.964 ng
RT: 27.762 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 2903047

Ion	Ratio	Lower	Upper
276	100		
138	20.6	16.4	24.6
277	24.9	19.8	29.8





#88
Benzo(b)fluoranthene
Concen: 38.191 ng
RT: 23.433 min Scan# 3476
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:252 Resp: 2548858

Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.7	26.5
125	7.6	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 38.890 ng
RT: 23.497 min Scan# 34
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 2625481

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.4	26.2
125	7.5	5.8	8.8





#90
Benzo(a)pyrene
Concen: 38.097 ng
RT: 24.239 min Scan# 3613
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:252 Resp: 2381402

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	8.5	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 37.985 ng
RT: 27.809 min Scan# 4
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2367754

Ion	Ratio	Lower	Upper
278	100		
139	13.3	10.5	15.7
279	24.0	19.4	29.0





#92
Benzo(g,h,i)perylene
Concen: 37.727 ng
RT: 28.826 min Scan# 4393
Delta R.T. 0.000 min
Lab File: BM050140.D
Acq: 08 May 2025 10:41

Tgt Ion:276 Resp: 2251454

Ion	Ratio	Lower	Upper
276	100		
277	23.5	19.1	28.7
138	17.7	14.3	21.5



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	-0.03
2		1,4-Dioxane	0.490	0.453	7.6	136	-0.02
3		Pyridine	1.258	1.221	2.9	141	-0.02
4		n-Nitrosodimethylamine	0.493	0.486	1.4	143	-0.02
5	S	2-Fluorophenol	1.142	1.109	2.9	141	-0.02
6		Aniline	1.813	1.775	2.1	140	-0.02
7	S	Phenol-d6	1.436	1.425	0.8	142	-0.02
8		2-Chlorophenol	1.235	1.198	3.0	140	-0.02
9		Benzaldehyde	0.961	0.933	2.9	139	-0.02
10	C	Phenol	1.453	1.439	1.0	144	-0.02
11		bis(2-Chloroethyl)ether	1.213	1.166	3.9	140	-0.02
12		1,3-Dichlorobenzene	1.492	1.402	6.0	138	-0.02
13	C	1,4-Dichlorobenzene	1.499	1.405	6.3	138	-0.03
14		1,2-Dichlorobenzene	1.448	1.357	6.3	137	-0.03
15		Benzyl Alcohol	0.996	1.014	-1.8	144	-0.02
16		2,2'-oxybis(1-Chloropropane	1.475	1.453	1.5	144	-0.03
17		2-Methylphenol	0.957	0.937	2.1	139	-0.02
18		Hexachloroethane	0.533	0.500	6.2	138	-0.03
19	P	n-Nitroso-di-n-propylamine	0.921	0.911	1.1	139	-0.02
20		3+4-Methylphenols	1.293	1.279	1.1	141	-0.02
21	I	Naphthalene-d8	1.000	1.000	0.0	154#	-0.03
22		Acetophenone	0.497	0.474	4.6	137	-0.02
23	S	Nitrobenzene-d5	0.406	0.390	3.9	138	-0.02
24		Nitrobenzene	0.351	0.337	4.0	139	-0.02
25		Isophorone	0.692	0.663	4.2	140	-0.02
26	C	2-Nitrophenol	0.179	0.183	-2.2	143	-0.02
27		2,4-Dimethylphenol	0.297	0.293	1.3	139	-0.02
28		bis(2-Chloroethoxy)methane	0.428	0.411	4.0	139	-0.02
29	C	2,4-Dichlorophenol	0.333	0.322	3.3	137	-0.02
30		1,2,4-Trichlorobenzene	0.406	0.381	6.2	137	-0.02
31		Naphthalene	1.013	0.949	6.3	137	-0.02
32		Benzoic acid	0.190	0.182	4.2	141	0.00
33		4-Chloroaniline	0.415	0.399	3.9	139	-0.02
34	C	Hexachlorobutadiene	0.258	0.244	5.4	138	-0.03
35		Caprolactam	0.097	0.090	7.2	133	0.00
36	C	4-Chloro-3-methylphenol	0.308	0.293	4.9	139	-0.01
37		2-Methylnaphthalene	0.679	0.643	5.3	139	-0.02
38		1-Methylnaphthalene	0.719	0.672	6.5	137	-0.02
39	I	Acenaphthene-d10	1.000	1.000	0.0	149	-0.02
40		1,2,4,5-Tetrachlorobenzene	0.692	0.686	0.9	137	-0.02
41	P	Hexachlorocyclopentadiene	0.423	0.394	6.9	129	-0.03
42	S	2,4,6-Tribromophenol	0.296	0.286	3.4	134	-0.01
43	C	2,4,6-Trichlorophenol	0.439	0.444	-1.1	139	-0.02
44		2,4,5-Trichlorophenol	0.474	0.470	0.8	135	-0.01
45	S	2-Fluorobiphenyl	1.691	1.640	3.0	135	-0.02
46		1,1'-Biphenyl	1.487	1.422	4.4	134	-0.02
47		2-Chloronaphthalene	1.177	1.124	4.5	134	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.276	0.277	-0.4	136	-0.02
49		Acenaphthylene	1.857	1.762	5.1	133	-0.02
50		Dimethylphthalate	1.462	1.353	7.5	131	-0.02
51		2,6-Dinitrotoluene	0.303	0.293	3.3	133	-0.02
52	C	Acenaphthene	1.075	1.012	5.9	133	-0.02
53		3-Nitroaniline	0.289	0.282	2.4	131	-0.01
54	P	2,4-Dinitrophenol	0.164	0.155	5.5	130	0.00
55		Dibenzofuran	1.788	1.680	6.0	133	-0.02
56	P	4-Nitrophenol	0.205	0.181	11.7	120	0.00
57		2,4-Dinitrotoluene	0.418	0.409	2.2	133	-0.01
58		Fluorene	1.463	1.365	6.7	130	-0.02
59		2,3,4,6-Tetrachlorophenol	0.410	0.385	6.1	132	-0.01
60		Diethylphthalate	1.377	1.253	9.0	130	-0.02
61		4-Chlorophenyl-phenylether	0.804	0.768	4.5	133	-0.02
62		4-Nitroaniline	0.279	0.246	11.8	117	-0.01
63		Azobenzene	1.154	1.070	7.3	131	-0.02
64	I	Phenanthrene-d10	1.000	1.000	0.0	142	-0.02
65		4,6-Dinitro-2-methylphenol	0.125	0.121	3.2	127	-0.01
66	c	n-Nitrosodiphenylamine	0.614	0.603	1.8	131	-0.02
67		4-Bromophenyl-phenylether	0.242	0.240	0.8	132	-0.02
68		Hexachlorobenzene	0.279	0.271	2.9	131	-0.02
69		Atrazine	0.222	0.212	4.5	128	-0.02
70	C	Pentachlorophenol	0.157	0.148	5.7	126	-0.01
71		Phenanthrene	1.128	1.059	6.1	127	-0.01
72		Anthracene	1.142	1.078	5.6	126	-0.02
73		Carbazole	0.991	0.919	7.3	124	-0.02
74		Di-n-butylphthalate	1.180	1.082	8.3	123	-0.02
75	C	Fluoranthene	1.324	1.185	10.5	120	-0.01
76	I	Chrysene-d12	1.000	1.000	0.0	111	-0.01
77		Benzidine	0.710	0.715	-0.7	103	-0.01
78		Pyrene	1.404	1.584	-12.8	116	-0.01
79	S	Terphenyl-d14	1.352	1.555	-15.0	113	-0.01
80		Butylbenzylphthalate	0.526	0.538	-2.3	106	-0.02
81		Benzo(a)anthracene	1.377	1.314	4.6	99	-0.01
82		3,3'-Dichlorobenzidine	0.527	0.482	8.5	92	-0.01
83		Chrysene	1.275	1.201	5.8	99	-0.02
84		Bis(2-ethylhexyl)phthalate	0.786	0.738	6.1	97	-0.02
85	c	Di-n-octyl phthalate	1.296	1.114	14.0	91	-0.02
86	I	Perylene-d12	1.000	1.000	0.0	96	-0.02
87		Indeno(1,2,3-cd)pyrene	1.490	1.414	5.1	84	-0.04
88		Benzo(b)fluoranthene	1.301	1.242	4.5	85	-0.02
89		Benzo(k)fluoranthene	1.316	1.279	2.8	88	-0.02
90	C	Benzo(a)pyrene	1.218	1.160	4.8	86	-0.02
91		Dibenzo(a,h)anthracene	1.215	1.154	5.0	85	-0.04
92		Benzo(g,h,i)perylene	1.163	1.097	5.7	85	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	154	-0.03
2		1,4-Dioxane	40.000	37.001	7.5	136	-0.02
3		Pyridine	40.000	38.824	2.9	141	-0.02
4		n-Nitrosodimethylamine	40.000	39.374	1.6	143	-0.02
5	S	2-Fluorophenol	80.000	77.661	2.9	141	-0.02
6		Aniline	40.000	39.162	2.1	140	-0.02
7	S	Phenol-d6	80.000	79.439	0.7	142	-0.02
8		2-Chlorophenol	40.000	38.813	3.0	140	-0.02
9		Benzaldehyde	40.000	38.833	2.9	139	-0.02
10	C	Phenol	40.000	39.605	1.0	144	-0.02
11		bis(2-Chloroethyl)ether	40.000	38.447	3.9	140	-0.02
12		1,3-Dichlorobenzene	40.000	37.588	6.0	138	-0.02
13	C	1,4-Dichlorobenzene	40.000	37.509	6.2	138	-0.03
14		1,2-Dichlorobenzene	40.000	37.497	6.3	137	-0.03
15		Benzyl Alcohol	40.000	40.752	-1.9	144	-0.02
16		2,2'-oxybis(1-Chloropropane	40.000	39.406	1.5	144	-0.03
17		2-Methylphenol	40.000	39.133	2.2	139	-0.02
18		Hexachloroethane	40.000	37.507	6.2	138	-0.03
19	P	n-Nitroso-di-n-propylamine	40.000	39.568	1.1	139	-0.02
20		3+4-Methylphenols	40.000	39.573	1.1	141	-0.02
21	I	Naphthalene-d8	20.000	20.000	0.0	154	-0.03
22		Acetophenone	40.000	38.153	4.6	137	-0.02
23	S	Nitrobenzene-d5	80.000	76.829	4.0	138	-0.02
24		Nitrobenzene	40.000	38.408	4.0	139	-0.02
25		Isophorone	40.000	38.333	4.2	140	-0.02
26	C	2-Nitrophenol	40.000	40.755	-1.9	143	-0.02
27		2,4-Dimethylphenol	40.000	39.354	1.6	139	-0.02
28		bis(2-Chloroethoxy)methane	40.000	38.454	3.9	139	-0.02
29	C	2,4-Dichlorophenol	40.000	38.708	3.2	137	-0.02
30		1,2,4-Trichlorobenzene	40.000	37.563	6.1	137	-0.02
31		Naphthalene	40.000	37.486	6.3	137	-0.02
32		Benzoic acid	40.000	38.186	4.5	141	0.00
33		4-Chloroaniline	40.000	38.483	3.8	139	-0.02
34	C	Hexachlorobutadiene	40.000	37.817	5.5	138	-0.03
35		Caprolactam	40.000	37.121	7.2	133	0.00
36	C	4-Chloro-3-methylphenol	40.000	38.140	4.6	139	-0.01
37		2-Methylnaphthalene	40.000	37.896	5.3	139	-0.02
38		1-Methylnaphthalene	40.000	37.410	6.5	137	-0.02
39	I	Acenaphthene-d10	20.000	20.000	0.0	149	-0.02
40		1,2,4,5-Tetrachlorobenzene	40.000	39.649	0.9	137	-0.02
41	P	Hexachlorocyclopentadiene	40.000	37.227	6.9	129	-0.03
42	S	2,4,6-Tribromophenol	80.000	77.306	3.4	134	-0.01
43	C	2,4,6-Trichlorophenol	40.000	40.452	-1.1	139	-0.02
44		2,4,5-Trichlorophenol	40.000	39.727	0.7	135	-0.01
45	S	2-Fluorobiphenyl	80.000	77.598	3.0	135	-0.02
46		1,1'-Biphenyl	40.000	38.256	4.4	134	-0.02
47		2-Chloronaphthalene	40.000	38.200	4.5	134	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	40.071	-0.2	136	-0.02
49		Acenaphthylene	40.000	37.967	5.1	133	-0.02
50		Dimethylphthalate	40.000	37.016	7.5	131	-0.02
51		2,6-Dinitrotoluene	40.000	38.760	3.1	133	-0.02
52	C	Acenaphthene	40.000	37.658	5.9	133	-0.02
53		3-Nitroaniline	40.000	38.978	2.6	131	-0.01
54	P	2,4-Dinitrophenol	40.000	35.437	11.4	130	0.00
55		Dibenzofuran	40.000	37.591	6.0	133	-0.02
56	P	4-Nitrophenol	40.000	33.657	15.9	120	0.00
57		2,4-Dinitrotoluene	40.000	39.126	2.2	133	-0.01
58		Fluorene	40.000	37.305	6.7	130	-0.02
59		2,3,4,6-Tetrachlorophenol	40.000	37.581	6.0	132	-0.01
60		Diethylphthalate	40.000	36.397	9.0	130	-0.02
61		4-Chlorophenyl-phenylether	40.000	38.219	4.5	133	-0.02
62		4-Nitroaniline	40.000	35.280	11.8	117	-0.01
63		Azobenzene	40.000	37.081	7.3	131	-0.02
64	I	Phenanthrene-d10	20.000	20.000	0.0	142	-0.02
65		4,6-Dinitro-2-methylphenol	40.000	38.684	3.3	127	-0.01
66	c	n-Nitrosodiphenylamine	40.000	39.272	1.8	131	-0.02
67		4-Bromophenyl-phenylether	40.000	39.646	0.9	132	-0.02
68		Hexachlorobenzene	40.000	38.846	2.9	131	-0.02
69		Atrazine	40.000	38.293	4.3	128	-0.02
70	C	Pentachlorophenol	40.000	37.733	5.7	126	-0.01
71		Phenanthrene	40.000	37.567	6.1	127	-0.01
72		Anthracene	40.000	37.761	5.6	126	-0.02
73		Carbazole	40.000	37.093	7.3	124	-0.02
74		Di-n-butylphthalate	40.000	36.672	8.3	123	-0.02
75	C	Fluoranthene	40.000	35.801	10.5	120	-0.01
76	I	Chrysene-d12	20.000	20.000	0.0	111	-0.01
77		Benzidine	40.000	40.247	-0.6	103	-0.01
78		Pyrene	40.000	45.115	-12.8	116	-0.01
79	S	Terphenyl-d14	80.000	92.052	-15.1	113	-0.01
80		Butylbenzylphthalate	40.000	40.913	-2.3	106	-0.02
81		Benzo(a)anthracene	40.000	38.150	4.6	99	-0.01
82		3,3'-Dichlorobenzidine	40.000	36.574	8.6	92	-0.01
83		Chrysene	40.000	37.672	5.8	99	-0.02
84		Bis(2-ethylhexyl)phthalate	40.000	37.590	6.0	97	-0.02
85	c	Di-n-octyl phthalate	40.000	34.387	14.0	91	-0.02
86	I	Perylene-d12	20.000	20.000	0.0	96	-0.02
87		Indeno(1,2,3-cd)pyrene	40.000	37.964	5.1	84	-0.04
88		Benzo(b)fluoranthene	40.000	38.191	4.5	85	-0.02
89		Benzo(k)fluoranthene	40.000	38.890	2.8	88	-0.02
90	C	Benzo(a)pyrene	40.000	38.097	4.8	86	-0.02
91		Dibenzo(a,h)anthracene	40.000	37.985	5.0	85	-0.04
92		Benzo(g,h,i)perylene	40.000	37.727	5.7	85	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050140.D
Acq On : 08 May 2025 10:41
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: May 08 12:03:45 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG No.: Q1872

Instrument ID: BNA_P Calibration Date/Time: 05/08/2025 14:45

Lab File ID: BP024568.D Init. Calib. Date(s): 04/14/2025 04/14/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:06 17:13

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.210	1.238		2.3	
Phenol-d6	1.656	1.642		-0.8	
Nitrobenzene-d5	0.351	0.357		1.7	
2-Fluorobiphenyl	1.306	1.381		5.7	
2,4,6-Tribromophenol	0.277	0.353		27.4	
Atrazine	0.152	0.143		-5.9	
Terphenyl-d14	0.992	1.078		8.7	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	179226	20.000	ng	-0.02
21) Naphthalene-d8	10.475	136	763529	20.000	ng	#-0.02
39) Acenaphthene-d10	14.334	164	492115	20.000	ng	-0.02
64) Phenanthrene-d10	17.139	188	999951	20.000	ng	0.00
76) Chrysene-d12	21.580	240	1084299	20.000	ng	-0.01
86) Perylene-d12	24.898	264	1119935	20.000	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.328	112	887341	81.862	ng	-0.02
7) Phenol-d6	6.893	99	1177489	79.333	ng	-0.02
23) Nitrobenzene-d5	8.852	82	1089270	81.348	ng	-0.02
42) 2,4,6-Tribromophenol	15.863	330	695426	102.138	ng	0.00
45) 2-Fluorobiphenyl	12.945	172	2718677	84.586	ng	-0.02
79) Terphenyl-d14	19.875	244	4676990	86.942	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.281	88	172053	37.526	ng	# 95
3) Pyridine	3.675	79	437577	36.143	ng	92
4) n-Nitrosodimethylamine	3.581	42	185073	46.061	ng	# 74
6) Aniline	7.040	93	499516	37.663	ng	97
8) 2-Chlorophenol	7.275	128	492030	40.656	ng	99
9) Benzaldehyde	6.858	77	329330	44.856	ng	95
10) Phenol	6.922	94	579749	38.938	ng	92
11) bis(2-Chloroethyl)ether	7.134	93	444400	37.045	ng	98
12) 1,3-Dichlorobenzene	7.593	146	527229	40.466	ng	99
13) 1,4-Dichlorobenzene	7.734	146	531653	40.218	ng	99
14) 1,2-Dichlorobenzene	8.052	146	508023	39.799	ng	98
15) Benzyl Alcohol	7.946	79	417641	43.511	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.228	45	494171	38.111	ng	99
17) 2-Methylphenol	8.152	107	399259	41.452	ng	97
18) Hexachloroethane	8.769	117	195493	40.921	ng	95
19) n-Nitroso-di-n-propyla...	8.499	70	371703	40.480	ng	95
20) 3+4-Methylphenols	8.481	107	560739	41.957	ng	98
22) Acetophenone	8.516	105	773287	40.919	ng	# 96
24) Nitrobenzene	8.893	77	528646	39.909	ng	99
25) Isophorone	9.416	82	999725	41.247	ng	96
26) 2-Nitrophenol	9.599	139	297932	45.960	ng	97
27) 2,4-Dimethylphenol	9.663	122	342518	42.081	ng	98
28) bis(2-Chloroethoxy)met...	9.893	93	640326	39.108	ng	98
29) 2,4-Dichlorophenol	10.140	162	488604	44.025	ng	99
30) 1,2,4-Trichlorobenzene	10.340	180	497055	41.801	ng	99
31) Naphthalene	10.528	128	1612593	40.094	ng	99
32) Benzoic acid	9.846	122	388873	41.002	ng	99
33) 4-Chloroaniline	10.640	127	600480	42.395	ng	98
34) Hexachlorobutadiene	10.810	225	315349	45.131	ng	99
35) Caprolactam	11.440	113	176675	43.144	ng	97
36) 4-Chloro-3-methylphenol	11.781	107	578324	44.738	ng	99
37) 2-Methylnaphthalene	12.140	142	1151688	41.289	ng	97
38) 1-Methylnaphthalene	12.357	142	1113391	40.803	ng	98
40) 1,2,4,5-Tetrachloroben...	12.510	216	593239	43.836	ng	100
41) Hexachlorocyclopentadiene	12.487	237	195571	40.860	ng	99
43) 2,4,6-Trichlorophenol	12.757	196	410368	45.608	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.845	196	452088	45.373	ng		99
46) 1,1'-Biphenyl	13.157	154	1506006	41.633	ng		98
47) 2-Chloronaphthalene	13.204	162	1141685	42.230	ng		98
48) 2-Nitroaniline	13.410	65	341991	45.097	ng		96
49) Acenaphthylene	14.051	152	1758429	41.312	ng		99
50) Dimethylphthalate	13.798	163	1552572	43.383	ng		100
51) 2,6-Dinitrotoluene	13.916	165	330872	43.539	ng		94
52) Acenaphthene	14.404	154	1089034	40.357	ng		98
53) 3-Nitroaniline	14.257	138	333536	41.760	ng		97
54) 2,4-Dinitrophenol	14.469	184	193261	43.173	ng		92
55) Dibenzofuran	14.740	168	1788471	40.547	ng		96
56) 4-Nitrophenol	14.598	139	329176	47.720	ng		88
57) 2,4-Dinitrotoluene	14.710	165	471679	45.500	ng		97
58) Fluorene	15.398	166	1421604	41.633	ng		100
59) 2,3,4,6-Tetrachlorophenol	14.975	232	409796	44.593	ng		99
60) Diethylphthalate	15.175	149	1596391	43.286	ng		99
61) 4-Chlorophenyl-phenyle...	15.398	204	710661	42.859	ng		97
62) 4-Nitroaniline	15.434	138	363252	42.962	ng		87
63) Azobenzene	15.698	77	1356150	40.008	ng		96
65) 4,6-Dinitro-2-methylph...	15.492	198	273977	43.459	ng		95
66) n-Nitrosodiphenylamine	15.622	169	1219495	40.859	ng		100
67) 4-Bromophenyl-phenylether	16.316	248	486848	44.517	ng		100
68) Hexachlorobenzene	16.434	284	599108	46.102	ng		98
69) Atrazine	16.604	200	286974	37.752	ng		97
70) Pentachlorophenol	16.792	266	399667	44.085	ng		99
71) Phenanthrene	17.186	178	2151569	39.442	ng		100
72) Anthracene	17.281	178	2137635	40.659	ng		99
73) Carbazole	17.569	167	2068524	40.267	ng		99
74) Di-n-butylphthalate	18.151	149	2817521	44.024	ng		100
75) Fluoranthene	19.269	202	2696306	41.559	ng		97
77) Benzidine	19.480	184	527701	38.394	ng		98
78) Pyrene	19.651	202	2736185	39.708	ng		99
80) Butylbenzylphthalate	20.598	149	1305824	44.243	ng		97
81) Benzo(a)anthracene	21.563	228	2714351	40.275	ng		100
82) 3,3'-Dichlorobenzidine	21.492	252	965428	40.813	ng		100
83) Chrysene	21.627	228	2571062	39.819	ng		100
84) Bis(2-ethylhexyl)phtha...	21.498	149	1863398	43.066	ng		100
85) Di-n-octyl phthalate	22.763	149	3046434	43.309	ng		100
87) Indeno(1,2,3-cd)pyrene	28.674	276	2894688	37.230	ng		98
88) Benzo(b)fluoranthene	23.863	252	2837774	42.273	ng	#	97
89) Benzo(k)fluoranthene	23.933	252	2710069m	41.794	ng		
90) Benzo(a)pyrene	24.751	252	2377084	41.051	ng		98
91) Dibenzo(a,h)anthracene	28.756	278	2378097	36.849	ng		97
92) Benzo(g,h,i)perylene	29.862	276	2352397	35.812	ng	#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.699 min Scan# 7
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:152 Resp: 17922

Ion	Ratio	Lower	Upper
152	100		
150	158.0	124.9	187.3
115	63.6	47.7	71.5



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#2
1,4-Dioxane
Concen: 37.526 ng
RT: 3.281 min Scan# 33
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 88 Resp: 172053

Ion	Ratio	Lower	Upper
88	100		
58	67.4	51.4	77.2
43	26.0	17.1	25.7#





#3
Pyridine
Concen: 36.143 ng
RT: 3.675 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 79 Resp: 43757

Ion	Ratio	Lower	Upper
79	100		
52	63.1	45.4	68.2
51	32.5	22.6	33.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#4
n-Nitrosodimethylamine
Concen: 46.061 ng
RT: 3.581 min Scan# 84
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 42 Resp: 185073

Ion	Ratio	Lower	Upper
42	100		
74	136.5	139.5	209.3#
44	6.0	5.8	8.8





#5
2-Fluorophenol
Concen: 81.862 ng
RT: 5.328 min Scan# 3
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:112 Resp: 88734

Ion	Ratio	Lower	Upper
112	100		
64	53.8	41.1	61.7
63	28.6	20.6	30.8

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#6
Aniline
Concen: 37.663 ng
RT: 7.040 min Scan# 672
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 93 Resp: 499516

Ion	Ratio	Lower	Upper
93	100		
66	38.5	29.1	43.7
65	18.7	14.7	22.1





#7
Phenol-d6
Concen: 79.333 ng
RT: 6.893 min Scan# 64
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 99 Resp: 117748

Ion	Ratio	Lower	Upper
99	100		
42	16.5	10.6	16.0
71	34.6	23.7	35.5



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#8
2-Chlorophenol
Concen: 40.656 ng
RT: 7.275 min Scan# 712
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:128 Resp: 492030

Ion	Ratio	Lower	Upper
128	100		
130	32.0	12.9	52.9
64	42.7	22.7	62.7





#9
Benzaldehyde
Concen: 44.856 ng
RT: 6.858 min Scan# 64
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 77 Resp: 32933

Ion	Ratio	Lower	Upper
77	100		
105	105.5	79.2	119.2
106	99.7	76.2	116.2



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#10
Phenol
Concen: 38.938 ng
RT: 6.922 min Scan# 652
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 94 Resp: 579749

Ion	Ratio	Lower	Upper
94	100		
65	26.8	4.5	44.5
66	38.5	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 37.045 ng
RT: 7.134 min Scan# 6
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 93 Resp: 44440

Ion	Ratio	Lower	Upper
93	100		
63	70.5	48.6	88.6
95	32.3	12.3	52.3

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#12
1,3-Dichlorobenzene
Concen: 40.466 ng
RT: 7.593 min Scan# 766
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:146 Resp: 527229

Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.4	77.0
75	30.0	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 40.218 ng
RT: 7.734 min Scan# 7
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:146 Resp: 53165

Ion	Ratio	Lower	Upper
146	100		
148	63.6	51.1	76.7
111	41.9	32.6	48.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#14
1,2-Dichlorobenzene
Concen: 39.799 ng
RT: 8.052 min Scan# 844
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:146 Resp: 508023

Ion	Ratio	Lower	Upper
146	100		
148	64.9	51.0	76.4
111	43.6	33.8	50.8





#15
Benzyl Alcohol
Concen: 43.511 ng
RT: 7.946 min Scan# 8
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 79 Resp: 41764

Ion	Ratio	Lower	Upper
79	100		
108	77.1	65.1	97.7
77	62.8	50.7	76.1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.111 ng
RT: 8.228 min Scan# 874
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 45 Resp: 494171

Ion	Ratio	Lower	Upper
45	100		
77	18.1	0.0	37.7
79	13.7	0.0	33.5





#17
2-Methylphenol
Concen: 41.452 ng
RT: 8.152 min Scan# 8
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:107 Resp: 39925

Ion	Ratio	Lower	Upper
107	100		
108	110.6	89.8	134.6
77	45.7	34.2	51.4
79	45.9	33.8	50.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#18
Hexachloroethane
Concen: 40.921 ng
RT: 8.769 min Scan# 966
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:117 Resp: 195493

Ion	Ratio	Lower	Upper
117	100		
119	96.7	77.8	116.6
201	104.2	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 40.480 ng
RT: 8.499 min Scan# 9
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 70 Resp: 37170

Ion	Ratio	Lower	Upper
70	100		
42	51.6	37.3	55.9
101	10.8	8.2	12.4
130	24.1	18.6	28.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#20
3+4-Methylphenols
Concen: 41.957 ng
RT: 8.481 min Scan# 917
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:107 Resp: 560739

Ion	Ratio	Lower	Upper
107	100		
108	90.1	69.8	109.8
77	34.3	11.5	51.5
79	30.3	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.475 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:136 Resp: 76352

Ion	Ratio	Lower	Upper
136	100		
137	10.7	9.2	13.8
54	6.6	6.0	9.0
68	5.3	5.5	8.3#



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#22
Acetophenone
Concen: 40.919 ng
RT: 8.516 min Scan# 923
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:105 Resp: 773287

Ion	Ratio	Lower	Upper
105	100		
71	2.8	1.7	2.5#
51	24.2	16.9	25.3
120	22.8	18.6	28.0





#23
Nitrobenzene-d5
Concen: 81.348 ng
RT: 8.852 min Scan# 9
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 82 Resp: 108927

Ion	Ratio	Lower	Upper
82	100		
128	46.9	36.5	54.7
54	47.0	33.4	50.2



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#24
Nitrobenzene
Concen: 39.909 ng
RT: 8.893 min Scan# 987
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 77 Resp: 528646

Ion	Ratio	Lower	Upper
77	100		
123	48.5	39.2	58.8
65	13.3	9.8	14.6





#25
Isophorone
Concen: 41.247 ng
RT: 9.416 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 82 Resp: 99972

Ion	Ratio	Lower	Upper
82	100		
95	8.0	5.4	8.2
138	19.6	14.4	21.6

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#26
2-Nitrophenol
Concen: 45.960 ng
RT: 9.599 min Scan# 1107
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:139 Resp: 297932

Ion	Ratio	Lower	Upper
139	100		
109	30.9	21.8	32.6
65	45.3	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 42.081 ng
RT: 9.663 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:122 Resp: 34251

Ion	Ratio	Lower	Upper
122	100		
107	119.6	93.7	140.5
121	60.5	48.2	72.2



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#28
bis(2-Chloroethoxy)methane
Concen: 39.108 ng
RT: 9.893 min Scan# 1157
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 93 Resp: 640326

Ion	Ratio	Lower	Upper
93	100		
95	33.2	26.0	39.0
123	12.7	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 44.025 ng
RT: 10.140 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:162 Resp: 48860

Ion	Ratio	Lower	Upper
162	100		
164	62.8	43.3	83.3
98	36.9	16.6	56.6

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#30
1,2,4-Trichlorobenzene
Concen: 41.801 ng
RT: 10.340 min Scan# 1233
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:180 Resp: 497055

Ion	Ratio	Lower	Upper
180	100		
182	97.6	77.3	115.9
145	32.8	25.1	37.7





#31
Naphthalene
Concen: 40.094 ng
RT: 10.528 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:128 Resp: 161259

Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.6	13.0
127	13.3	10.3	15.5

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#32
Benzoic acid
Concen: 41.002 ng
RT: 9.846 min Scan# 1149
Delta R.T. 0.024 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:122 Resp: 388873

Ion	Ratio	Lower	Upper
122	100		
105	122.6	104.8	144.8
77	88.6	69.2	109.2





#33
4-Chloroaniline
Concen: 42.395 ng
RT: 10.640 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:127 Resp: 60048

Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.1	39.1
65	29.0	22.1	33.1
92	19.9	15.0	22.4



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#34
Hexachlorobutadiene
Concen: 45.131 ng
RT: 10.810 min Scan# 1313
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:225 Resp: 315349

Ion	Ratio	Lower	Upper
225	100		
223	61.4	49.5	74.3
227	64.8	50.9	76.3





#35
Caprolactam
Concen: 43.144 ng
RT: 11.440 min Scan# 14
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:113 Resp: 17667

Ion	Ratio	Lower	Upper
113	100		
55	135.3	115.9	155.9
56	107.6	81.1	121.1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#36
4-Chloro-3-methylphenol
Concen: 44.738 ng
RT: 11.781 min Scan# 1478
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:107 Resp: 578324

Ion	Ratio	Lower	Upper
107	100		
144	25.5	20.5	30.7
142	77.9	63.4	95.2





#37
2-Methylnaphthalene
Concen: 41.289 ng
RT: 12.140 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:142 Resp: 115168

Ion	Ratio	Lower	Upper
142	100		
141	89.9	70.3	105.5
115	36.8	26.9	40.3



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#38
1-Methylnaphthalene
Concen: 40.803 ng
RT: 12.357 min Scan# 1576
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:142 Resp: 1113391

Ion	Ratio	Lower	Upper
142	100		
141	93.7	73.6	110.4
116	4.1	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.334 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:164 Resp: 49211

Ion	Ratio	Lower	Upper
164	100		
162	103.1	80.6	120.8
160	45.2	35.3	52.9



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 43.836 ng
RT: 12.510 min Scan# 1602
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:216 Resp: 593239

Ion	Ratio	Lower	Upper
216	100		
214	78.2	62.6	93.8
179	21.8	17.2	25.8
108	18.2	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 40.860 ng
RT: 12.487 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:237 Resp: 19557

Ion	Ratio	Lower	Upper
237	100		
235	62.9	43.9	83.9
272	13.5	0.0	33.6

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#42
2,4,6-Tribromophenol
Concen: 102.138 ng
RT: 15.863 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:330 Resp: 695426

Ion	Ratio	Lower	Upper
330	100		
332	98.1	77.1	115.7
141	28.6	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 45.608 ng
RT: 12.757 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:196 Resp: 41036

Ion	Ratio	Lower	Upper
196	100		
198	95.9	74.3	111.5
200	31.5	24.0	36.0

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#44
2,4,5-Trichlorophenol
Concen: 45.373 ng
RT: 12.845 min Scan# 1659
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:196 Resp: 452088

Ion	Ratio	Lower	Upper
196	100		
198	96.5	77.3	115.9
97	48.1	40.4	60.6
132	28.5	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 84.586 ng
RT: 12.945 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:172 Resp: 271867

Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.2	42.2
170	23.5	18.6	27.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#46
1,1'-Biphenyl
Concen: 41.633 ng
RT: 13.157 min Scan# 1712
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:154 Resp: 1506006

Ion	Ratio	Lower	Upper
154	100		
153	40.9	20.3	60.3
76	11.7	0.0	33.5





#47
2-Chloronaphthalene
Concen: 42.230 ng
RT: 13.204 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:162 Resp: 114168

Ion	Ratio	Lower	Upper
162	100		
127	42.4	31.9	47.9
164	32.8	26.2	39.4



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#48
2-Nitroaniline
Concen: 45.097 ng
RT: 13.410 min Scan# 1755
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 65 Resp: 341991

Ion	Ratio	Lower	Upper
65	100		
92	73.9	62.3	93.5
138	118.2	97.2	145.8





#49
Acenaphthylene
Concen: 41.312 ng
RT: 14.051 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:152 Resp: 175842

Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.9	23.9
153	12.9	10.6	15.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#50
Dimethylphthalate
Concen: 43.383 ng
RT: 13.798 min Scan# 1821
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:163 Resp: 1552572

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	10.1	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 43.539 ng
RT: 13.916 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:165 Resp: 33087

Ion	Ratio	Lower	Upper
165	100		
63	52.1	38.0	57.0
89	55.8	41.7	62.5

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#52
Acenaphthene
Concen: 40.357 ng
RT: 14.404 min Scan# 1924
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:154 Resp: 1089034

Ion	Ratio	Lower	Upper
154	100		
153	114.8	89.8	134.6
152	54.4	42.6	63.8





#53
3-Nitroaniline
Concen: 41.760 ng
RT: 14.257 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 33353

Ion	Ratio	Lower	Upper
138	100		
108	11.2	8.3	12.5
92	108.5	89.8	134.8

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#54
2,4-Dinitrophenol
Concen: 43.173 ng
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:184 Resp: 193261

Ion	Ratio	Lower	Upper
184	100		
63	59.8	42.6	64.0
154	64.7	47.2	70.8





#55
Dibenzofuran
Concen: 40.547 ng
RT: 14.740 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:168 Resp: 178847

Ion	Ratio	Lower	Upper
168	100		
139	41.4	30.3	45.5
169	13.1	10.6	16.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#56
4-Nitrophenol
Concen: 47.720 ng
RT: 14.598 min Scan# 1957
Delta R.T. 0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:139 Resp: 329176

Ion	Ratio	Lower	Upper
139	100		
109	80.9	46.1	86.1
65	91.5	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 45.500 ng
RT: 14.710 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:165 Resp: 47167

Ion	Ratio	Lower	Upper
165	100		
63	42.0	30.4	45.6
89	69.6	54.6	82.0
182	3.1	2.2	3.2



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#58
Fluorene
Concen: 41.633 ng
RT: 15.398 min Scan# 2093
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:166 Resp: 1421604

Ion	Ratio	Lower	Upper
166	100		
165	97.7	78.1	117.1
167	13.5	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 44.593 ng
RT: 14.975 min Scan# 2
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:232 Resp: 40979

Ion	Ratio	Lower	Upper
232	100		
131	46.4	36.8	55.2
130	2.5	1.9	2.9
166	30.5	25.0	37.6



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#60
Diethylphthalate
Concen: 43.286 ng
RT: 15.175 min Scan# 2055
Delta R.T. -0.018 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:149 Resp: 1596391

Ion	Ratio	Lower	Upper
149	100		
177	21.5	17.4	26.2
150	12.8	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 42.859 ng
RT: 15.398 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:204 Resp: 71066

Ion	Ratio	Lower	Upper
204	100		
206	32.4	26.5	39.7
141	64.2	49.4	74.0

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#62
4-Nitroaniline
Concen: 42.962 ng
RT: 15.434 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:138 Resp: 363252

Ion	Ratio	Lower	Upper
138	100		
92	49.5	27.8	67.8
108	85.4	48.9	88.9





#63
Azobenzene
Concen: 40.008 ng
RT: 15.698 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion: 77 Resp: 135615

Ion	Ratio	Lower	Upper
77	100		
182	28.2	7.3	47.3
105	22.1	2.3	42.3
51	29.0	3.4	43.4



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2389
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:188 Resp: 999951

Ion	Ratio	Lower	Upper
188	100		
94	8.9	8.6	13.0
80	9.8	9.8	14.6





#65
4,6-Dinitro-2-methylphenol
Concen: 43.459 ng
RT: 15.492 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:198 Resp: 27397

Ion	Ratio	Lower	Upper
198	100		
51	40.9	18.4	58.4
105	43.1	26.5	66.5

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#66
n-Nitrosodiphenylamine
Concen: 40.859 ng
RT: 15.622 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:169 Resp: 1219495

Ion	Ratio	Lower	Upper
169	100		
168	66.2	53.0	79.4
167	36.0	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 44.517 ng
RT: 16.316 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:248 Resp: 48684

Ion	Ratio	Lower	Upper
248	100		
250	97.5	78.2	117.4
141	72.6	57.7	86.5



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#68
Hexachlorobenzene
Concen: 46.102 ng
RT: 16.434 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:284 Resp: 599108

Ion	Ratio	Lower	Upper
284	100		
142	29.4	24.6	37.0
249	29.4	23.8	35.6





#69
Atrazine
Concen: 37.752 ng
RT: 16.604 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:200 Resp: 28697

Ion	Ratio	Lower	Upper
200	100		
173	25.5	7.5	47.5
215	47.1	28.8	68.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#70
Pentachlorophenol
Concen: 44.085 ng
RT: 16.792 min Scan# 2330
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:266 Resp: 399667

Ion	Ratio	Lower	Upper
266	100		
268	65.3	51.0	76.6
264	62.4	49.8	74.6





#71
Phenanthrene
Concen: 39.442 ng
RT: 17.186 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:178 Resp: 215156

Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.4	23.0
179	15.5	12.5	18.7



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#72
Anthracene
Concen: 40.659 ng
RT: 17.281 min Scan# 2413
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:178 Resp: 2137635

Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.8	12.4	18.6





#73
Carbazole
Concen: 40.267 ng
RT: 17.569 min Scan# 24
Delta R.T. 0.000 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:167 Resp: 206852

Ion	Ratio	Lower	Upper
167	100		
166	22.2	17.1	25.7
139	13.8	10.6	16.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#74
Di-n-butylphthalate
Concen: 44.024 ng
RT: 18.151 min Scan# 2561
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:149 Resp: 2817521

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.4	11.2
104	5.2	3.9	5.9





#75
Fluoranthene
Concen: 41.559 ng
RT: 19.269 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:202 Resp: 269630

Ion	Ratio	Lower	Upper
202	100		
101	10.4	0.0	33.2
203	17.7	0.0	37.7



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.580 min Scan# 3144
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:240 Resp: 1084299

Ion	Ratio	Lower	Upper
240	100		
120	9.9	9.8	14.8
236	25.6	20.9	31.3





#77
Benzidine
Concen: 38.394 ng
RT: 19.480 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:184 Resp: 52770

Ion	Ratio	Lower	Upper
184	100		
185	13.8	11.6	17.4
183	12.3	9.2	13.8



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#78
Pyrene
Concen: 39.708 ng
RT: 19.651 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:202 Resp: 2736185

Ion	Ratio	Lower	Upper
202	100		
200	21.0	17.2	25.8
203	17.5	14.4	21.6





#79
Terphenyl-d14
Concen: 86.942 ng
RT: 19.875 min Scan# 2
Delta R.T. 0.000 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:244 Resp: 467699

Ion	Ratio	Lower	Upper
244	100		
212	7.0	5.6	8.4
122	8.9	9.4	14.0



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#80
Butylbenzylphthalate
Concen: 44.243 ng
RT: 20.598 min Scan# 2977
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:149 Resp: 1305824

Ion	Ratio	Lower	Upper
149	100		
91	67.4	56.4	84.6
206	21.4	16.3	24.5





#81
Benzo(a)anthracene
Concen: 40.275 ng
RT: 21.563 min Scan# 3
Delta R.T. -0.006 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:228 Resp: 271435

Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.8	32.8
229	19.7	15.9	23.9

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#82
3,3'-Dichlorobenzidine
Concen: 40.813 ng
RT: 21.492 min Scan# 3129
Delta R.T. 0.000 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:252 Resp: 965428

Ion	Ratio	Lower	Upper
252	100		
254	64.7	51.7	77.5
126	10.1	8.8	13.2





#83
Chrysene
Concen: 39.819 ng
RT: 21.627 min Scan# 3
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:228 Resp: 257106

Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.2	36.2
229	19.7	15.6	23.4



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 43.066 ng
RT: 21.498 min Scan# 3130
Delta R.T. -0.012 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:149 Resp: 1863398

Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.4	33.6
279	5.3	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 43.309 ng
RT: 22.763 min Scan# 3
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:149 Resp: 304643

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.6	6.0	9.0

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.898 min Scan# 3708
Delta R.T. -0.035 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:264 Resp: 1119935

Ion	Ratio	Lower	Upper
264	100		
260	23.8	18.9	28.3
265	21.8	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 37.230 ng
RT: 28.674 min Scan# 4
Delta R.T. -0.065 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:276 Resp: 289468

Ion	Ratio	Lower	Upper
276	100		
138	16.0	14.5	21.7
277	25.5	20.2	30.2



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#88
Benzo(b)fluoranthene
Concen: 42.273 ng
RT: 23.863 min Scan# 3532
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:252 Resp: 2837774

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	9.3	9.4	14.0#





#89
Benzo(k)fluoranthene
Concen: 41.794 ng m
RT: 23.933 min Scan# 3
Delta R.T. -0.023 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:252 Resp: 271006

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	9.0	9.1	13.7



Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#90
Benzo(a)pyrene
Concen: 41.051 ng
RT: 24.751 min Scan# 3683
Delta R.T. -0.029 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:252 Resp: 2377084

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.6	26.4
125	10.1	10.1	15.1





#91
Dibenzo(a,h)anthracene
Concen: 36.849 ng
RT: 28.756 min Scan# 4
Delta R.T. -0.047 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:278 Resp: 237809

Ion	Ratio	Lower	Upper
278	100		
139	13.9	12.6	19.0
279	24.2	18.5	27.7

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#92
Benzo(g,h,i)perylene
Concen: 35.812 ng
RT: 29.862 min Scan# 4552
Delta R.T. -0.035 min
Lab File: BP024568.D
Acq: 08 May 2025 14:45

Tgt Ion:276 Resp: 2352397

Ion	Ratio	Lower	Upper
276	100		
277	24.2	18.9	28.3
138	18.5	18.7	28.1#



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	-0.02
2		1,4-Dioxane	0.512	0.480	6.3	68	0.00
3		Pyridine	1.351	1.221	9.6	64	-0.01
4		n-Nitrosodimethylamine	0.448	0.516	-15.2	82	-0.01
5	S	2-Fluorophenol	1.210	1.238	-2.3	72	-0.02
6		Aniline	1.480	1.394	5.8	64	-0.02
7	S	Phenol-d6	1.656	1.642	0.8	68	-0.02
8		2-Chlorophenol	1.350	1.373	-1.7	71	-0.02
9		Benzaldehyde	0.819	0.919	-12.2	80	-0.02
10	C	Phenol	1.661	1.617	2.6	67	-0.02
11		bis(2-Chloroethyl)ether	1.339	1.240	7.4	65	-0.02
12		1,3-Dichlorobenzene	1.454	1.471	-1.2	72	-0.02
13	C	1,4-Dichlorobenzene	1.475	1.483	-0.5	72	-0.02
14		1,2-Dichlorobenzene	1.424	1.417	0.5	71	-0.02
15		Benzyl Alcohol	1.071	1.165	-8.8	73	-0.02
16		2,2'-oxybis(1-Chloropropane	1.447	1.379	4.7	66	-0.02
17		2-Methylphenol	1.075	1.114	-3.6	71	-0.02
18		Hexachloroethane	0.533	0.545	-2.3	73	-0.02
19	P	n-Nitroso-di-n-propylamine	1.025	1.037	-1.2	70	-0.02
20		3+4-Methylphenols	1.491	1.564	-4.9	71	-0.01
21	I	Naphthalene-d8	1.000	1.000	0.0	73	-0.02
22		Acetophenone	0.495	0.506	-2.2	71	-0.02
23	S	Nitrobenzene-d5	0.351	0.357	-1.7	71	-0.02
24		Nitrobenzene	0.347	0.346	0.3	69	-0.02
25		Isophorone	0.635	0.655	-3.1	71	-0.01
26	C	2-Nitrophenol	0.170	0.195	-14.7	78	-0.02
27		2,4-Dimethylphenol	0.213	0.224	-5.2	72	-0.02
28		bis(2-Chloroethoxy)methane	0.429	0.419	2.3	68	-0.02
29	C	2,4-Dichlorophenol	0.291	0.320	-10.0	76	-0.01
30		1,2,4-Trichlorobenzene	0.311	0.325	-4.5	74	-0.02
31		Naphthalene	1.054	1.056	-0.2	71	-0.02
32		Benzoic acid	0.248	0.255	-2.8	77	0.02
33		4-Chloroaniline	0.371	0.393	-5.9	72	-0.02
34	C	Hexachlorobutadiene	0.183	0.207	-13.1	80	-0.02
35		Caprolactam	0.107	0.116	-8.4	75	0.00
36	C	4-Chloro-3-methylphenol	0.339	0.379	-11.8	77	-0.01
37		2-Methylnaphthalene	0.731	0.754	-3.1	73	-0.02
38		1-Methylnaphthalene	0.715	0.729	-2.0	73	-0.02
39	I	Acenaphthene-d10	1.000	1.000	0.0	75	-0.02
40		1,2,4,5-Tetrachlorobenzene	0.550	0.603	-9.6	78	-0.02
41	P	Hexachlorocyclopentadiene	0.195	0.199	-2.1	68	-0.02
42	S	2,4,6-Tribromophenol	0.277	0.353	-27.4#	92	0.00
43	C	2,4,6-Trichlorophenol	0.366	0.417	-13.9	80	-0.02
44		2,4,5-Trichlorophenol	0.405	0.459	-13.3	80	-0.01
45	S	2-Fluorobiphenyl	1.306	1.381	-5.7	76	-0.02
46		1,1'-Biphenyl	1.470	1.530	-4.1	75	-0.02
47		2-Chloronaphthalene	1.099	1.160	-5.6	76	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.308	0.347	-12.7	79	-0.01
49		Acenaphthylene	1.730	1.787	-3.3	74	-0.02
50		Dimethylphthalate	1.454	1.577	-8.5	79	-0.01
51		2,6-Dinitrotoluene	0.309	0.336	-8.7	77	-0.01
52	C	Acenaphthene	1.097	1.106	-0.8	74	-0.02
53		3-Nitroaniline	0.325	0.339	-4.3	72	0.00
54	P	2,4-Dinitrophenol	0.182	0.196	-7.7	79	0.00
55		Dibenzofuran	1.793	1.817	-1.3	74	-0.02
56	P	4-Nitrophenol	0.280	0.334	-19.3	82	0.00
57		2,4-Dinitrotoluene	0.421	0.479	-13.8	80	-0.01
58		Fluorene	1.388	1.444	-4.0	77	-0.02
59		2,3,4,6-Tetrachlorophenol	0.373	0.416	-11.5	80	-0.02
60		Diethylphthalate	1.499	1.622	-8.2	78	-0.02
61		4-Chlorophenyl-phenylether	0.674	0.722	-7.1	79	-0.01
62		4-Nitroaniline	0.344	0.369	-7.3	75	0.00
63		Azobenzene	1.378	1.378	0.0	72	-0.01
64	I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
65		4,6-Dinitro-2-methylphenol	0.126	0.137	-8.7	82	0.00
66	c	n-Nitrosodiphenylamine	0.597	0.610	-2.2	76	0.00
67		4-Bromophenyl-phenylether	0.219	0.243	-11.0	83	0.00
68		Hexachlorobenzene	0.260	0.300	-15.4	87	0.00
69		Atrazine	0.152	0.143	5.9	93	0.00
70	C	Pentachlorophenol	0.181	0.200	-10.5	80	0.00
71		Phenanthrene	1.091	1.076	1.4	75	0.00
72		Anthracene	1.052	1.069	-1.6	75	0.00
73		Carbazole	1.027	1.034	-0.7	75	0.00
74		Di-n-butylphthalate	1.280	1.409	-10.1	79	0.00
75	C	Fluoranthene	1.298	1.348	-3.9	78	-0.01
76	I	Chrysene-d12	1.000	1.000	0.0	80	-0.01
77		Benzidine	0.254	0.243	4.3	45#	0.00
78		Pyrene	1.271	1.262	0.7	78	0.00
79	S	Terphenyl-d14	0.992	1.078	-8.7	84	0.00
80		Butylbenzylphthalate	0.544	0.602	-10.7	81	0.00
81		Benzo(a)anthracene	1.243	1.252	-0.7	78	0.00
82		3,3'-Dichlorobenzidine	0.436	0.445	-2.1	74	0.00
83		Chrysene	1.191	1.186	0.4	79	-0.01
84		Bis(2-ethylhexyl)phthalate	0.798	0.859	-7.6	77	-0.01
85	c	Di-n-octyl phthalate	1.297	1.405	-8.3	77	-0.02
86	I	Perylene-d12	1.000	1.000	0.0	75	-0.04
87		Indeno(1,2,3-cd)pyrene	1.388	1.292	6.9	67	-0.06
88		Benzo(b)fluoranthene	1.199	1.267	-5.7	76	-0.02
89		Benzo(k)fluoranthene	1.158	1.210	-4.5	75	-0.02
90	C	Benzo(a)pyrene	1.034	1.061	-2.6	73	-0.03
91		Dibenzo(a,h)anthracene	1.152	1.062	7.8	67	-0.05
92		Benzo(g,h,i)perylene	1.173	1.050	10.5	65	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	73	-0.02
2		1,4-Dioxane	40.000	37.526	6.2	68	0.00
3		Pyridine	40.000	36.143	9.6	64	-0.01
4		n-Nitrosodimethylamine	40.000	46.061	-15.2	82	-0.01
5	S	2-Fluorophenol	80.000	81.862	-2.3	72	-0.02
6		Aniline	40.000	37.663	5.8	64	-0.02
7	S	Phenol-d6	80.000	79.333	0.8	68	-0.02
8		2-Chlorophenol	40.000	40.656	-1.6	71	-0.02
9		Benzaldehyde	40.000	44.856	-12.1	80	-0.02
10	C	Phenol	40.000	38.938	2.7	67	-0.02
11		bis(2-Chloroethyl)ether	40.000	37.045	7.4	65	-0.02
12		1,3-Dichlorobenzene	40.000	40.466	-1.2	72	-0.02
13	C	1,4-Dichlorobenzene	40.000	40.218	-0.5	72	-0.02
14		1,2-Dichlorobenzene	40.000	39.799	0.5	71	-0.02
15		Benzyl Alcohol	40.000	43.511	-8.8	73	-0.02
16		2,2'-oxybis(1-Chloropropane	40.000	38.111	4.7	66	-0.02
17		2-Methylphenol	40.000	41.452	-3.6	71	-0.02
18		Hexachloroethane	40.000	40.921	-2.3	73	-0.02
19	P	n-Nitroso-di-n-propylamine	40.000	40.480	-1.2	70	-0.02
20		3+4-Methylphenols	40.000	41.957	-4.9	71	-0.01
21	I	Naphthalene-d8	20.000	20.000	0.0	73	-0.02
22		Acetophenone	40.000	40.919	-2.3	71	-0.02
23	S	Nitrobenzene-d5	80.000	81.348	-1.7	71	-0.02
24		Nitrobenzene	40.000	39.909	0.2	69	-0.02
25		Isophorone	40.000	41.247	-3.1	71	-0.01
26	C	2-Nitrophenol	40.000	45.960	-14.9	78	-0.02
27		2,4-Dimethylphenol	40.000	42.081	-5.2	72	-0.02
28		bis(2-Chloroethoxy)methane	40.000	39.108	2.2	68	-0.02
29	C	2,4-Dichlorophenol	40.000	44.025	-10.1	76	-0.01
30		1,2,4-Trichlorobenzene	40.000	41.801	-4.5	74	-0.02
31		Naphthalene	40.000	40.094	-0.2	71	-0.02
32		Benzoic acid	40.000	41.002	-2.5	77	0.02
33		4-Chloroaniline	40.000	42.395	-6.0	72	-0.02
34	C	Hexachlorobutadiene	40.000	45.131	-12.8	80	-0.02
35		Caprolactam	40.000	43.144	-7.9	75	0.00
36	C	4-Chloro-3-methylphenol	40.000	44.738	-11.8	77	-0.01
37		2-Methylnaphthalene	40.000	41.289	-3.2	73	-0.02
38		1-Methylnaphthalene	40.000	40.803	-2.0	73	-0.02
39	I	Acenaphthene-d10	20.000	20.000	0.0	75	-0.02
40		1,2,4,5-Tetrachlorobenzene	40.000	43.836	-9.6	78	-0.02
41	P	Hexachlorocyclopentadiene	40.000	40.860	-2.1	68	-0.02
42	S	2,4,6-Tribromophenol	80.000	102.138	-27.7#	92	0.00
43	C	2,4,6-Trichlorophenol	40.000	45.608	-14.0	80	-0.02
44		2,4,5-Trichlorophenol	40.000	45.373	-13.4	80	-0.01
45	S	2-Fluorobiphenyl	80.000	84.586	-5.7	76	-0.02
46		1,1'-Biphenyl	40.000	41.633	-4.1	75	-0.02
47		2-Chloronaphthalene	40.000	42.230	-5.6	76	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	45.097	-12.7	79	-0.01
49		Acenaphthylene	40.000	41.312	-3.3	74	-0.02
50		Dimethylphthalate	40.000	43.383	-8.5	79	-0.01
51		2,6-Dinitrotoluene	40.000	43.539	-8.8	77	-0.01
52	C	Acenaphthene	40.000	40.357	-0.9	74	-0.02
53		3-Nitroaniline	40.000	41.760	-4.4	72	0.00
54	P	2,4-Dinitrophenol	40.000	43.173	-7.9	79	0.00
55		Dibenzofuran	40.000	40.547	-1.4	74	-0.02
56	P	4-Nitrophenol	40.000	47.720	-19.3	82	0.00
57		2,4-Dinitrotoluene	40.000	45.500	-13.8	80	-0.01
58		Fluorene	40.000	41.633	-4.1	77	-0.02
59		2,3,4,6-Tetrachlorophenol	40.000	44.593	-11.5	80	-0.02
60		Diethylphthalate	40.000	43.286	-8.2	78	-0.02
61		4-Chlorophenyl-phenylether	40.000	42.859	-7.1	79	-0.01
62		4-Nitroaniline	40.000	42.962	-7.4	75	0.00
63		Azobenzene	40.000	40.008	-0.0	72	-0.01
64	I	Phenanthrene-d10	20.000	20.000	0.0	78	0.00
65		4,6-Dinitro-2-methylphenol	40.000	43.459	-8.6	82	0.00
66	c	n-Nitrosodiphenylamine	40.000	40.859	-2.1	76	0.00
67		4-Bromophenyl-phenylether	40.000	44.517	-11.3	83	0.00
68		Hexachlorobenzene	40.000	46.102	-15.3	87	0.00
69		Atrazine	40.000	37.752	5.6	93	0.00
70	C	Pentachlorophenol	40.000	44.085	-10.2	80	0.00
71		Phenanthrene	40.000	39.442	1.4	75	0.00
72		Anthracene	40.000	40.659	-1.6	75	0.00
73		Carbazole	40.000	40.267	-0.7	75	0.00
74		Di-n-butylphthalate	40.000	44.024	-10.1	79	0.00
75	C	Fluoranthene	40.000	41.559	-3.9	78	-0.01
76	I	Chrysene-d12	20.000	20.000	0.0	80	-0.01
77		Benzidine	40.000	38.394	4.0	45	0.00
78		Pyrene	40.000	39.708	0.7	78	0.00
79	S	Terphenyl-d14	80.000	86.942	-8.7	84	0.00
80		Butylbenzylphthalate	40.000	44.243	-10.6	81	0.00
81		Benzo(a)anthracene	40.000	40.275	-0.7	78	0.00
82		3,3'-Dichlorobenzidine	40.000	40.813	-2.0	74	0.00
83		Chrysene	40.000	39.819	0.5	79	-0.01
84		Bis(2-ethylhexyl)phthalate	40.000	43.066	-7.7	77	-0.01
85	c	Di-n-octyl phthalate	40.000	43.309	-8.3	77	-0.02
86	I	Perylene-d12	20.000	20.000	0.0	75	-0.04
87		Indeno(1,2,3-cd)pyrene	40.000	37.230	6.9	67	-0.06
88		Benzo(b)fluoranthene	40.000	42.273	-5.7	76	-0.02
89		Benzo(k)fluoranthene	40.000	41.794	-4.5	75	-0.02
90	C	Benzo(a)pyrene	40.000	41.051	-2.6	73	-0.03
91		Dibenzo(a,h)anthracene	40.000	36.849	7.9	67	-0.05
92		Benzo(g,h,i)perylene	40.000	35.812	10.5	65	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024568.D
Acq On : 08 May 2025 14:45
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 08 15:12:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ALLI03

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG No.: Q1872

Instrument ID: BNA_P Calibration Date/Time: 05/14/2025 11:35

Lab File ID: BP024625.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:41 15:26

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.169	1.128		-3.5	
Phenol-d6	1.492	1.457		-2.3	
Nitrobenzene-d5	0.396	0.391		-1.3	
2-Fluorobiphenyl	1.527	1.435		-6.0	
2,4,6-Tribromophenol	0.339	0.349		3.0	
Atrazine	0.225	0.219		-2.7	
Terphenyl-d14	1.166	1.115		-4.4	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.628	152	115714	20.000	ng		-0.04
21) Naphthalene-d8	10.410	136	507416	20.000	ng		-0.02
39) Acenaphthene-d10	14.286	164	336535	20.000	ng		-0.02
64) Phenanthrene-d10	17.092	188	705050	20.000	ng		0.00
76) Chrysene-d12	21.533	240	813969	20.000	ng		0.01
86) Perylene-d12	24.821	264	940710	20.000	ng		0.02
System Monitoring Compounds							
5) 2-Fluorophenol	5.258	112	521980	77.154	ng		-0.03
7) Phenol-d6	6.822	99	674475	78.114	ng		-0.03
23) Nitrobenzene-d5	8.781	82	793572	79.021	ng		-0.03
42) 2,4,6-Tribromophenol	15.810	330	469524	82.297	ng		0.00
45) 2-Fluorobiphenyl	12.892	172	1931765	75.203	ng		-0.01
79) Terphenyl-d14	19.833	244	3628929	76.446	ng		0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.211	88	121989	38.399	ng		99
3) Pyridine	3.605	79	274040	38.903	ng		98
4) n-Nitrosodimethylamine	3.511	42	127064	40.858	ng		92
6) Aniline	6.975	93	428908	38.474	ng		100
8) 2-Chlorophenol	7.204	128	301635	39.274	ng		100
9) Benzaldehyde	6.787	77	216909m	38.807	ng		
10) Phenol	6.846	94	345991	38.823	ng		98
11) bis(2-Chloroethyl)ether	7.063	93	274967	37.549	ng		100
12) 1,3-Dichlorobenzene	7.522	146	331540	37.521	ng		98
13) 1,4-Dichlorobenzene	7.663	146	337195	37.942	ng		99
14) 1,2-Dichlorobenzene	7.981	146	330466	38.161	ng		99
15) Benzyl Alcohol	7.875	79	270028	41.153	ng		98
16) 2,2'-oxybis(1-Chloropr...	8.152	45	331737	37.827	ng		99
17) 2-Methylphenol	8.075	107	254105	39.754	ng		99
18) Hexachloroethane	8.699	117	130316	38.265	ng		97
19) n-Nitroso-di-n-propyla...	8.434	70	237966	39.841	ng		99
20) 3+4-Methylphenols	8.404	107	345708	39.761	ng		98
22) Acetophenone	8.451	105	487619	38.471	ng	#	99
24) Nitrobenzene	8.822	77	347674	39.340	ng		99
25) Isophorone	9.346	82	673166	38.643	ng		98
26) 2-Nitrophenol	9.528	139	174274	40.243	ng		95
27) 2,4-Dimethylphenol	9.598	122	301398	39.475	ng		98
28) bis(2-Chloroethoxy)met...	9.828	93	398711	38.310	ng		99
29) 2,4-Dichlorophenol	10.075	162	295603	39.952	ng		98
30) 1,2,4-Trichlorobenzene	10.269	180	322652	38.698	ng		98
31) Naphthalene	10.463	128	1002824	38.138	ng		99
32) Benzoic acid	9.757	122	199091m	39.096	ng		
33) 4-Chloroaniline	10.575	127	427253	40.266	ng		100
34) Hexachlorobutadiene	10.745	225	212669	39.374	ng		97
35) Caprolactam	11.375	113	105084	39.255	ng		99
36) 4-Chloro-3-methylphenol	11.716	107	350402	38.834	ng		98
37) 2-Methylnaphthalene	12.081	142	646503	37.971	ng		99
38) 1-Methylnaphthalene	12.298	142	678827	37.262	ng		99
40) 1,2,4,5-Tetrachloroben...	12.457	216	363473	37.193	ng		99
41) Hexachlorocyclopentadiene	12.428	237	251504	42.442	ng		96
43) 2,4,6-Trichlorophenol	12.704	196	250626	39.595	ng		100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.787	196	278502	39.527	ng		98
46) 1,1'-Biphenyl	13.110	154	930967	37.428	ng		100
47) 2-Chloronaphthalene	13.151	162	710183	37.446	ng		98
48) 2-Nitroaniline	13.357	65	221555	39.458	ng		97
49) Acenaphthylene	14.004	152	1199079	37.780	ng		100
50) Dimethylphthalate	13.751	163	959581	37.409	ng		100
51) 2,6-Dinitrotoluene	13.863	165	210286	38.817	ng		91
52) Acenaphthene	14.351	154	695773	38.133	ng		100
53) 3-Nitroaniline	14.198	138	223399	40.387	ng		97
54) 2,4-Dinitrophenol	14.422	184	118947	38.435	ng		99
55) Dibenzofuran	14.692	168	1127191	38.731	ng		100
56) 4-Nitrophenol	14.539	139	171551	38.161	ng		96
57) 2,4-Dinitrotoluene	14.669	165	310445	40.648	ng		95
58) Fluorene	15.357	166	937420	39.503	ng		99
59) 2,3,4,6-Tetrachlorophenol	14.928	232	259540	39.708	ng		99
60) Diethylphthalate	15.133	149	1020334	39.356	ng		99
61) 4-Chlorophenyl-phenyle...	15.351	204	467584	39.510	ng		97
62) 4-Nitroaniline	15.398	138	234140	43.838	ng		99
63) Azobenzene	15.651	77	886547	39.973	ng		99
65) 4,6-Dinitro-2-methylph...	15.457	198	178715	39.242	ng		93
66) n-Nitrosodiphenylamine	15.569	169	820871	37.904	ng		99
67) 4-Bromophenyl-phenylether	16.269	248	318213	38.302	ng		97
68) Hexachlorobenzene	16.386	284	395863	37.499	ng		98
69) Atrazine	16.557	200	309117	38.949	ng		98
70) Pentachlorophenol	16.745	266	250707	42.306	ng		98
71) Phenanthrene	17.139	178	1460711	37.268	ng		100
72) Anthracene	17.233	178	1509621	38.282	ng		100
73) Carbazole	17.522	167	1394376	39.067	ng		100
74) Di-n-butylphthalate	18.116	149	1838243	40.046	ng		100
75) Fluoranthene	19.233	202	1782976	38.916	ng		99
77) Benzidine	19.439	184	911891	41.133	ng		100
78) Pyrene	19.610	202	1857977	38.096	ng		100
80) Butylbenzylphthalate	20.563	149	868834	40.056	ng		97
81) Benzo(a)anthracene	21.510	228	1961745	38.383	ng		99
82) 3,3'-Dichlorobenzidine	21.445	252	778902	41.031	ng		99
83) Chrysene	21.586	228	1835723	37.889	ng		99
84) Bis(2-ethylhexyl)phtha...	21.451	149	1281105	40.184	ng		99
85) Di-n-octyl phthalate	22.715	149	2125149	40.763	ng		100
87) Indeno(1,2,3-cd)pyrene	28.545	276	2703892	39.371	ng	#	93
88) Benzo(b)fluoranthene	23.792	252	2090102	38.817	ng		99
89) Benzo(k)fluoranthene	23.851	252	2149476	38.741	ng		99
90) Benzo(a)pyrene	24.668	252	2019628	39.054	ng		99
91) Dibenzo(a,h)anthracene	28.627	278	2207636m	39.513	ng		
92) Benzo(g,h,i)perylene	29.692	276	2182397	39.279	ng		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.628 min Scan# 7
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 115714

Ion	Ratio	Lower	Upper
152	100		
150	150.9	125.9	188.9
115	63.0	50.1	75.1





#2
1,4-Dioxane
Concen: 38.399 ng
RT: 3.211 min Scan# 21
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 88 Resp: 121989

Ion	Ratio	Lower	Upper
88	100		
58	67.7	54.0	81.0
43	27.7	21.4	32.0





#3
Pyridine
Concen: 38.903 ng
RT: 3.605 min Scan# 8
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 274040

Ion	Ratio	Lower	Upper
79	100		
52	63.5	50.2	75.2
51	35.5	26.8	40.2





#4
n-Nitrosodimethylamine
Concen: 40.858 ng
RT: 3.511 min Scan# 72
Delta R.T. -0.024 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 42 Resp: 127064

Ion	Ratio	Lower	Upper
42	100		
74	122.9	106.2	159.2
44	5.9	5.4	8.2





#5
2-Fluorophenol
Concen: 77.154 ng
RT: 5.258 min Scan# 3
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:112 Resp: 521980

Ion	Ratio	Lower	Upper
112	100		
64	54.6	44.1	66.1
63	30.4	23.5	35.3





#6
Aniline
Concen: 38.474 ng
RT: 6.975 min Scan# 661
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 93 Resp: 428908

Ion	Ratio	Lower	Upper
93	100		
66	39.0	31.0	46.6
65	19.6	15.5	23.3





#7
Phenol-d6
Concen: 78.114 ng
RT: 6.822 min Scan# 6
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 674475

Ion	Ratio	Lower	Upper
99	100		
42	18.5	14.4	21.6
71	37.7	29.2	43.8





#8
2-Chlorophenol
Concen: 39.274 ng
RT: 7.204 min Scan# 700
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:128 Resp: 301635

Ion	Ratio	Lower	Upper
128	100		
130	32.8	13.0	53.0
64	42.1	22.3	62.3





#9
Benzaldehyde
Concen: 38.807 ng m
RT: 6.787 min Scan# 6
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 216909

Ion	Ratio	Lower	Upper
77	100		
105	99.5	83.3	123.3
106	97.0	78.0	118.0





#10
Phenol
Concen: 38.823 ng
RT: 6.846 min Scan# 639
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 94 Resp: 345991

Ion	Ratio	Lower	Upper
94	100		
65	29.4	8.7	48.7
66	41.4	19.4	59.4





#11
bis(2-Chloroethyl)ether
Concen: 37.549 ng
RT: 7.063 min Scan# 6
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 274967

Ion	Ratio	Lower	Upper
93	100		
63	70.7	50.6	90.6
95	32.8	12.9	52.9





#12
1,3-Dichlorobenzene
Concen: 37.521 ng
RT: 7.522 min Scan# 754
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:146 Resp: 331540

Ion	Ratio	Lower	Upper
146	100		
148	62.2	51.4	77.2
75	31.3	25.2	37.8





#13
1,4-Dichlorobenzene
Concen: 37.942 ng
RT: 7.663 min Scan# 7
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 337195

Ion	Ratio	Lower	Upper
146	100		
148	63.7	51.0	76.6
111	42.4	33.0	49.6





#14
1,2-Dichlorobenzene
Concen: 38.161 ng
RT: 7.981 min Scan# 832
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:146 Resp: 330466

Ion	Ratio	Lower	Upper
146	100		
148	64.2	51.0	76.4
111	44.8	35.2	52.8





#15
Benzyl Alcohol
Concen: 41.153 ng
RT: 7.875 min Scan# 8
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 270028

Ion	Ratio	Lower	Upper
79	100		
108	73.0	60.8	91.2
77	65.4	51.9	77.9





#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.827 ng
RT: 8.152 min Scan# 861
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 45 Resp: 331737

Ion	Ratio	Lower	Upper
45	100		
77	17.2	0.0	37.7
79	13.5	0.0	34.0





#17
2-Methylphenol
Concen: 39.754 ng
RT: 8.075 min Scan# 84
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 254105

Ion	Ratio	Lower	Upper
107	100		
108	111.9	88.9	133.3
77	50.1	39.0	58.6
79	48.5	38.4	57.6





#18
Hexachloroethane
Concen: 38.265 ng
RT: 8.699 min Scan# 954
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:117 Resp: 130316

Ion	Ratio	Lower	Upper
117	100		
119	97.5	79.4	119.0
201	98.9	82.7	124.1





#19
n-Nitroso-di-n-propylamine
Concen: 39.841 ng
RT: 8.434 min Scan# 9
Delta R.T. -0.024 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 237966

Ion	Ratio	Lower	Upper
70	100		
42	54.0	43.0	64.6
101	10.5	8.7	13.1
130	23.5	19.4	29.0





#20
3+4-Methylphenols
Concen: 39.761 ng
RT: 8.404 min Scan# 904
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:107 Resp: 345708

Ion	Ratio	Lower	Upper
107	100		
108	87.5	66.7	106.7
77	36.5	13.1	53.1
79	30.6	8.9	48.9





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.410 min Scan# 1
Delta R.T. -0.024 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:136 Resp: 507416

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.8	13.2
54	7.2	5.5	8.3
68	5.0	4.2	6.2





#22
Acetophenone
Concen: 38.471 ng
RT: 8.451 min Scan# 912
Delta R.T. -0.024 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:105 Resp: 487619

Ion	Ratio	Lower	Upper
105	100		
71	1.6	2.0	3.0#
51	26.6	20.4	30.6
120	22.6	18.1	27.1





#23
Nitrobenzene-d5
Concen: 79.021 ng
RT: 8.781 min Scan# 9
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 793572

Ion	Ratio	Lower	Upper
82	100		
128	45.3	37.0	55.4
54	49.1	38.7	58.1





#24
Nitrobenzene
Concen: 39.340 ng
RT: 8.822 min Scan# 975
Delta R.T. -0.029 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 77 Resp: 347674

Ion	Ratio	Lower	Upper
77	100		
123	45.7	37.4	56.0
65	14.0	11.3	16.9





#25
Isophorone
Concen: 38.643 ng
RT: 9.346 min Scan# 1
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 673166

Ion	Ratio	Lower	Upper
82	100		
95	8.3	6.3	9.5
138	18.3	15.3	22.9





#26
2-Nitrophenol
Concen: 40.243 ng
RT: 9.528 min Scan# 1095
Delta R.T. -0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:139 Resp: 174274

Ion	Ratio	Lower	Upper
139	100		
109	34.7	25.9	38.9
65	51.3	37.9	56.9





#27
2,4-Dimethylphenol
Concen: 39.475 ng
RT: 9.598 min Scan# 11
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:122 Resp: 301398

Ion	Ratio	Lower	Upper
122	100		
107	123.6	96.2	144.2
121	60.4	47.6	71.4





#28
bis(2-Chloroethoxy)methane
Concen: 38.310 ng
RT: 9.828 min Scan# 1146
Delta R.T. -0.024 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 93 Resp: 398711

Ion	Ratio	Lower	Upper
93	100		
95	32.2	25.5	38.3
123	12.4	9.7	14.5





#29
2,4-Dichlorophenol
Concen: 39.952 ng
RT: 10.075 min Scan# 11
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 295603

Ion	Ratio	Lower	Upper
162	100		
164	64.3	42.3	82.3
98	37.6	18.4	58.4





#30
1,2,4-Trichlorobenzene
Concen: 38.698 ng
RT: 10.269 min Scan# 1221
Delta R.T. -0.030 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:180 Resp: 322652

Ion	Ratio	Lower	Upper
180	100		
182	95.8	79.0	118.4
145	33.0	26.6	39.8





#31
Naphthalene
Concen: 38.138 ng
RT: 10.463 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:128 Resp: 1002824

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.0	10.7	16.1





#32
Benzoic acid
Concen: 39.096 ng m
RT: 9.757 min Scan# 1134
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:122 Resp: 199091

Ion	Ratio	Lower	Upper
122	100		
105	126.4	106.9	146.9
77	93.9	75.1	115.1





#33
4-Chloroaniline
Concen: 40.266 ng
RT: 10.575 min Scan# 1
Delta R.T. -0.024 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:127 Resp: 427253

Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.8	38.6
65	29.8	23.9	35.9
92	19.7	16.1	24.1





#34
Hexachlorobutadiene
Concen: 39.374 ng
RT: 10.745 min Scan# 1302
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:225 Resp: 212669

Ion	Ratio	Lower	Upper
225	100		
223	61.8	51.0	76.6
227	63.8	53.1	79.7





#35
Caprolactam
Concen: 39.255 ng
RT: 11.375 min Scan# 14
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 105084

Ion	Ratio	Lower	Upper
113	100		
55	140.3	119.0	159.0
56	109.7	90.3	130.3





#36
4-Chloro-3-methylphenol
Concen: 38.834 ng
RT: 11.716 min Scan# 1467
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:107 Resp: 350402

Ion	Ratio	Lower	Upper
107	100		
144	24.8	20.1	30.1
142	79.9	62.1	93.1





#37
2-Methylnaphthalene
Concen: 37.971 ng
RT: 12.081 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 646503

Ion	Ratio	Lower	Upper
142	100		
141	89.3	71.2	106.8
115	40.3	30.6	46.0





#38
1-Methylnaphthalene
Concen: 37.262 ng
RT: 12.298 min Scan# 1566
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:142 Resp: 678827

Ion	Ratio	Lower	Upper
142	100		
141	92.0	74.7	112.1
116	4.4	3.4	5.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.286 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:164 Resp: 336535

Ion	Ratio	Lower	Upper
164	100		
162	103.5	81.8	122.6
160	44.7	36.3	54.5





#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.193 ng
RT: 12.457 min Scan# 1593
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:216 Resp: 363473

Ion	Ratio	Lower	Upper
216	100		
214	79.0	62.3	93.5
179	23.0	17.7	26.5
108	20.7	16.0	24.0





#41
Hexachlorocyclopentadiene
Concen: 42.442 ng
RT: 12.428 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 251504

Ion	Ratio	Lower	Upper
237	100		
235	63.4	39.9	79.9
272	13.5	0.0	33.2





#42
2,4,6-Tribromophenol
Concen: 82.297 ng
RT: 15.810 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:330 Resp: 469524

Ion	Ratio	Lower	Upper
330	100		
332	97.2	78.2	117.4
141	30.2	24.3	36.5





#43
2,4,6-Trichlorophenol
Concen: 39.595 ng
RT: 12.704 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:196 Resp: 250626

Ion	Ratio	Lower	Upper
196	100		
198	96.1	76.9	115.3
200	31.0	25.2	37.8





#44
2,4,5-Trichlorophenol
Concen: 39.527 ng
RT: 12.787 min Scan# 1649
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:196 Resp: 278502

Ion	Ratio	Lower	Upper
196	100		
198	96.3	79.2	118.8
97	52.0	42.3	63.5
132	30.1	23.3	34.9





#45
2-Fluorobiphenyl
Concen: 75.203 ng
RT: 12.892 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:172 Resp: 1931765

Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.8	43.2
170	23.4	18.8	28.2





#46
1,1'-Biphenyl
Concen: 37.428 ng
RT: 13.110 min Scan# 1704
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:154 Resp: 930967

Ion	Ratio	Lower	Upper
154	100		
153	41.0	21.0	61.0
76	10.7	0.0	31.5





#47
2-Chloronaphthalene
Concen: 37.446 ng
RT: 13.151 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:162 Resp: 710183

Ion	Ratio	Lower	Upper
162	100		
127	43.8	33.8	50.6
164	32.9	26.0	39.0





#48
2-Nitroaniline
Concen: 39.458 ng
RT: 13.357 min Scan# 1746
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion: 65 Resp: 221555

Ion	Ratio	Lower	Upper
65	100		
92	71.0	58.7	88.1
138	108.9	90.2	135.2





#49
Acenaphthylene
Concen: 37.780 ng
RT: 14.004 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:152 Resp: 1199079

Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.2	24.4
153	13.0	10.2	15.4





#50
Dimethylphthalate
Concen: 37.409 ng
RT: 13.751 min Scan# 1813
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:163 Resp: 959581

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.0
164	10.3	8.1	12.1





#51
2,6-Dinitrotoluene
Concen: 38.817 ng
RT: 13.863 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 210286

Ion	Ratio	Lower	Upper
165	100		
63	59.8	42.4	63.6
89	61.9	44.8	67.2





#52
Acenaphthene
Concen: 38.133 ng
RT: 14.351 min Scan# 1915
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:154 Resp: 695773

Ion	Ratio	Lower	Upper
154	100		
153	114.4	91.0	136.4
152	53.5	42.9	64.3





#53
3-Nitroaniline
Concen: 40.387 ng
RT: 14.198 min Scan# 1
Delta R.T. -0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:138 Resp: 223399

Ion	Ratio	Lower	Upper
138	100		
108	11.5	9.0	13.4
92	115.5	90.1	135.1





#54
2,4-Dinitrophenol
Concen: 38.435 ng
RT: 14.422 min Scan# 1927
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:184 Resp: 118947

Ion	Ratio	Lower	Upper
184	100		
63	64.6	51.3	76.9
154	66.4	53.5	80.3





#55
Dibenzofuran
Concen: 38.731 ng
RT: 14.692 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 1127191

Ion	Ratio	Lower	Upper
168	100		
139	43.0	34.2	51.4
169	13.2	10.8	16.2





#56
4-Nitrophenol
Concen: 38.161 ng
RT: 14.539 min Scan# 1947
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:139 Resp: 171551

Ion	Ratio	Lower	Upper
139	100		
109	85.0	61.4	101.4
65	95.2	72.2	112.2





#57
2,4-Dinitrotoluene
Concen: 40.648 ng
RT: 14.669 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:165 Resp: 310445

Ion	Ratio	Lower	Upper
165	100		
63	45.8	33.8	50.6
89	73.1	55.6	83.4
182	3.0	2.4	3.6





#58
Fluorene
Concen: 39.503 ng
RT: 15.357 min Scan# 2086
Delta R.T. -0.000 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:166 Resp: 937420

Ion	Ratio	Lower	Upper
166	100		
165	97.4	78.5	117.7
167	13.8	10.7	16.1





#59
2,3,4,6-Tetrachlorophenol
Concen: 39.708 ng
RT: 14.928 min Scan# 2
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:232 Resp: 259540

Ion	Ratio	Lower	Upper
232	100		
131	50.0	39.5	59.3
130	2.9	2.2	3.4
166	32.4	25.7	38.5





#60
Diethylphthalate
Concen: 39.356 ng
RT: 15.133 min Scan# 2048
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:149 Resp: 1020334

Ion	Ratio	Lower	Upper
149	100		
177	21.4	16.7	25.1
150	12.5	10.2	15.2





#61
4-Chlorophenyl-phenylether
Concen: 39.510 ng
RT: 15.351 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 467584

Ion	Ratio	Lower	Upper
204	100		
206	33.1	26.0	39.0
141	69.5	52.9	79.3





#62
4-Nitroaniline
Concen: 43.838 ng
RT: 15.398 min Scan# 2093
Delta R.T. 0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:138 Resp: 234140

Ion	Ratio	Lower	Upper
138	100		
92	49.2	29.2	69.2
108	90.7	69.0	109.0





#63
Azobenzene
Concen: 39.973 ng
RT: 15.651 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 886547

Ion	Ratio	Lower	Upper
77	100		
182	26.9	5.7	45.7
105	21.5	1.3	41.3
51	30.8	10.6	50.6





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.092 min Scan# 2381
Delta R.T. -0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:188 Resp: 705050

Ion	Ratio	Lower	Upper
188	100		
94	8.6	6.5	9.7
80	9.6	7.3	10.9





#65
4,6-Dinitro-2-methylphenol
Concen: 39.242 ng
RT: 15.457 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 178715

Ion	Ratio	Lower	Upper
198	100		
51	47.0	21.1	61.1
105	46.4	23.6	63.6





#66
n-Nitrosodiphenylamine
Concen: 37.904 ng
RT: 15.569 min Scan# 2122
Delta R.T. -0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:169 Resp: 820871

Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.2	79.8
167	36.7	28.6	43.0





#67
4-Bromophenyl-phenylether
Concen: 38.302 ng
RT: 16.269 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:248 Resp: 318213

Ion	Ratio	Lower	Upper
248	100		
250	99.7	78.2	117.4
141	75.0	63.0	94.6





#68
Hexachlorobenzene
Concen: 37.499 ng
RT: 16.386 min Scan# 2261
Delta R.T. 0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:284 Resp: 395863

Ion	Ratio	Lower	Upper
284	100		
142	34.3	25.9	38.9
249	30.8	24.1	36.1





#69
Atrazine
Concen: 38.949 ng
RT: 16.557 min Scan# 2
Delta R.T. -0.000 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:200 Resp: 309117

Ion	Ratio	Lower	Upper
200	100		
173	26.8	5.9	45.9
215	46.5	28.3	68.3





#70
Pentachlorophenol
Concen: 42.306 ng
RT: 16.745 min Scan# 2322
Delta R.T. -0.000 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:266 Resp: 250707

Ion	Ratio	Lower	Upper
266	100		
268	63.8	52.2	78.4
264	61.8	51.1	76.7





#71
Phenanthrene
Concen: 37.268 ng
RT: 17.139 min Scan# 2
Delta R.T. 0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 1460711

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.4	23.2
179	15.7	12.2	18.4





#72
Anthracene
Concen: 38.282 ng
RT: 17.233 min Scan# 2405
Delta R.T. 0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:178 Resp: 1509621

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.4	12.2	18.2





#73
Carbazole
Concen: 39.067 ng
RT: 17.522 min Scan# 24
Delta R.T. 0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:167 Resp: 1394376

Ion	Ratio	Lower	Upper
167	100		
166	22.0	17.7	26.5
139	14.0	11.3	16.9





#74
Di-n-butylphthalate
Concen: 40.046 ng
RT: 18.116 min Scan# 2555
Delta R.T. 0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:149 Resp: 1838243

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.2	10.8
104	5.7	4.3	6.5





#75
Fluoranthene
Concen: 38.916 ng
RT: 19.233 min Scan# 2
Delta R.T. 0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:202 Resp: 1782976

Ion	Ratio	Lower	Upper
202	100		
101	10.1	0.0	30.3
203	17.2	0.0	37.5





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.533 min Scan# 3136
Delta R.T. 0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:240 Resp: 813969

Ion	Ratio	Lower	Upper
240	100		
120	9.5	7.8	11.6
236	25.4	20.6	31.0





#77
Benzidine
Concen: 41.133 ng
RT: 19.439 min Scan# 2
Delta R.T. 0.023 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:184 Resp: 911891

Ion	Ratio	Lower	Upper
184	100		
185	14.1	11.2	16.8
183	12.0	9.8	14.8





#78
Pyrene
Concen: 38.096 ng
RT: 19.610 min Scan# 2809
Delta R.T. 0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:202 Resp: 1857977

Ion	Ratio	Lower	Upper
202	100		
200	21.3	16.9	25.3
203	17.5	13.9	20.9





#79
Terphenyl-d14
Concen: 76.446 ng
RT: 19.833 min Scan# 2
Delta R.T. 0.023 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:244 Resp: 3628929

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.5	8.3
122	9.2	7.4	11.0





#80
Butylbenzylphthalate
Concen: 40.056 ng
RT: 20.563 min Scan# 2971
Delta R.T. 0.023 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:149 Resp: 868834

Ion	Ratio	Lower	Upper
149	100		
91	66.6	55.4	83.0
206	21.1	16.2	24.2





#81
Benzo(a)anthracene
Concen: 38.383 ng
RT: 21.510 min Scan# 3
Delta R.T. 0.006 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 1961745

Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.4	32.0
229	19.9	15.5	23.3





#82
3,3'-Dichlorobenzidine
Concen: 41.031 ng
RT: 21.445 min Scan# 3121
Delta R.T. 0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:252 Resp: 778902

Ion	Ratio	Lower	Upper
252	100		
254	65.4	51.8	77.6
126	10.3	7.7	11.5





#83
Chrysene
Concen: 37.889 ng
RT: 21.586 min Scan# 3
Delta R.T. 0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:228 Resp: 1835723

Ion	Ratio	Lower	Upper
228	100		
226	30.3	23.9	35.9
229	19.3	15.5	23.3





#84
Bis(2-ethylhexyl)phthalate
Concen: 40.184 ng
RT: 21.451 min Scan# 3122
Delta R.T. 0.012 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:149 Resp: 1281105

Ion	Ratio	Lower	Upper
149	100		
167	28.0	22.1	33.1
279	5.2	4.6	6.8





#85
Di-n-octyl phthalate
Concen: 40.763 ng
RT: 22.715 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 2125149

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	8.0	6.3	9.5





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.821 min Scan# 3695
Delta R.T. 0.018 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:264 Resp: 940710

Ion	Ratio	Lower	Upper
264	100		
260	23.4	18.7	28.1
265	22.0	17.0	25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 39.371 ng
RT: 28.545 min Scan# 4
Delta R.T. 0.029 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:276 Resp: 2703892

Ion	Ratio	Lower	Upper
276	100		
138	14.2	11.2	16.8
277	25.3	15.7	23.5#





#88
Benzo(b)fluoranthene
Concen: 38.817 ng
RT: 23.792 min Scan# 3520
Delta R.T. 0.035 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:252 Resp: 2090102

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.7	26.5
125	8.5	7.5	11.3





#89
Benzo(k)fluoranthene
Concen: 38.741 ng
RT: 23.851 min Scan# 3
Delta R.T. 0.023 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:252 Resp: 2149476

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.6	26.4
125	8.3	7.0	10.4





#90
Benzo(a)pyrene
Concen: 39.054 ng
RT: 24.668 min Scan# 3669
Delta R.T. 0.023 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:252 Resp: 2019628

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	9.8	8.1	12.1





#91
Dibenzo(a,h)anthracene
Concen: 39.513 ng m
RT: 28.627 min Scan# 4
Delta R.T. 0.029 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2207636

Ion	Ratio	Lower	Upper
278	100		
139	13.3	10.2	15.4
279	23.4	19.0	28.4





#92
Benzo(g,h,i)perylene
Concen: 39.279 ng
RT: 29.692 min Scan# 4523
Delta R.T. 0.029 min
Lab File: BP024625.D
Acq: 14 May 2025 11:35

Tgt Ion:276 Resp: 2182397

Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.0	28.4
138	17.8	14.4	21.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	-0.04
2		1,4-Dioxane	0.549	0.527	4.0	100	-0.02
3		Pyridine	1.218	1.184	2.8	95	-0.02
4		n-Nitrosodimethylamine	0.538	0.549	-2.0	102	-0.02
5	S	2-Fluorophenol	1.169	1.128	3.5	98	-0.03
6		Aniline	1.927	1.853	3.8	93	-0.03
7	S	Phenol-d6	1.492	1.457	2.3	95	-0.03
8		2-Chlorophenol	1.327	1.303	1.8	96	-0.04
9		Benzaldehyde	0.966	0.937	3.0	94	-0.04
10	C	Phenol	1.540	1.495	2.9	95	-0.04
11		bis(2-Chloroethyl)ether	1.266	1.188	6.2	94	-0.04
12		1,3-Dichlorobenzene	1.527	1.433	6.2	95	-0.03
13	C	1,4-Dichlorobenzene	1.536	1.457	5.1	95	-0.04
14		1,2-Dichlorobenzene	1.497	1.428	4.6	96	-0.03
15		Benzyl Alcohol	1.134	1.167	-2.9	97	-0.03
16		2,2'-oxybis(1-Chloropropane	1.516	1.433	5.5	93	-0.04
17		2-Methylphenol	1.105	1.098	0.6	96	-0.04
18		Hexachloroethane	0.589	0.563	4.4	98	-0.03
19	P	n-Nitroso-di-n-propylamine	1.032	1.028	0.4	94	-0.02
20		3+4-Methylphenols	1.503	1.494	0.6	94	-0.03
21	I	Naphthalene-d8	1.000	1.000	0.0	94	-0.02
22		Acetophenone	0.500	0.480	4.0	93	-0.02
23	S	Nitrobenzene-d5	0.396	0.391	1.3	96	-0.03
24		Nitrobenzene	0.348	0.343	1.4	96	-0.03
25		Isophorone	0.687	0.663	3.5	93	-0.03
26	C	2-Nitrophenol	0.171	0.172	-0.6	94	-0.04
27		2,4-Dimethylphenol	0.301	0.297	1.3	94	-0.03
28		bis(2-Chloroethoxy)methane	0.410	0.393	4.1	93	-0.02
29	C	2,4-Dichlorophenol	0.292	0.291	0.3	94	-0.02
30		1,2,4-Trichlorobenzene	0.329	0.318	3.3	98	-0.03
31		Naphthalene	1.036	0.988	4.6	95	-0.02
32		Benzoic acid	0.181	0.196	-8.3	96	-0.02
33		4-Chloroaniline	0.418	0.421	-0.7	94	-0.02
34	C	Hexachlorobutadiene	0.213	0.210	1.4	99	-0.02
35		Caprolactam	0.106	0.104	1.9	89	-0.01
36	C	4-Chloro-3-methylphenol	0.356	0.345	3.1	91	-0.02
37		2-Methylnaphthalene	0.671	0.637	5.1	93	-0.01
38		1-Methylnaphthalene	0.718	0.669	6.8	91	-0.02
39	I	Acenaphthene-d10	1.000	1.000	0.0	93	-0.02
40		1,2,4,5-Tetrachlorobenzene	0.581	0.540	7.1	91	-0.01
41	P	Hexachlorocyclopentadiene	0.352	0.374	-6.3	97	-0.02
42	S	2,4,6-Tribromophenol	0.339	0.349	-2.9	100	0.00
43	C	2,4,6-Trichlorophenol	0.376	0.372	1.1	93	-0.01
44		2,4,5-Trichlorophenol	0.419	0.414	1.2	95	-0.01
45	S	2-Fluorobiphenyl	1.527	1.435	6.0	93	-0.01
46		1,1'-Biphenyl	1.478	1.383	6.4	93	0.00
47		2-Chloronaphthalene	1.127	1.055	6.4	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48		2-Nitroaniline	0.334	0.329	1.5	91	-0.02
49		Acenaphthylene	1.886	1.782	5.5	92	-0.01
50		Dimethylphthalate	1.524	1.426	6.4	91	0.00
51		2,6-Dinitrotoluene	0.322	0.312	3.1	93	-0.02
52	C	Acenaphthene	1.084	1.034	4.6	93	-0.01
53		3-Nitroaniline	0.329	0.332	-0.9	91	-0.02
54	P	2,4-Dinitrophenol	0.169	0.177	-4.7	96	0.00
55		Dibenzofuran	1.730	1.675	3.2	95	0.00
56	P	4-Nitrophenol	0.250	0.255	-2.0	92	-0.01
57		2,4-Dinitrotoluene	0.454	0.461	-1.5	94	-0.01
58		Fluorene	1.410	1.393	1.2	97	0.00
59		2,3,4,6-Tetrachlorophenol	0.388	0.386	0.5	95	-0.01
60		Diethylphthalate	1.541	1.516	1.6	98	0.00
61		4-Chlorophenyl-phenylether	0.703	0.695	1.1	98	0.00
62		4-Nitroaniline	0.317	0.348	-9.8	98	0.00
63		Azobenzene	1.318	1.317	0.1	98	0.00
64	I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
65		4,6-Dinitro-2-methylphenol	0.129	0.127	1.6	97	0.00
66	c	n-Nitrosodiphenylamine	0.614	0.582	5.2	96	-0.01
67		4-Bromophenyl-phenylether	0.236	0.226	4.2	98	0.00
68		Hexachlorobenzene	0.299	0.281	6.0	98	0.00
69		Atrazine	0.225	0.219	2.7	99	0.00
70	C	Pentachlorophenol	0.168	0.178	-6.0	102	0.00
71		Phenanthrene	1.112	1.036	6.8	97	0.00
72		Anthracene	1.119	1.071	4.3	98	0.00
73		Carbazole	1.012	0.989	2.3	97	0.00
74		Di-n-butylphthalate	1.302	1.304	-0.2	102	0.02
75	C	Fluoranthene	1.300	1.264	2.8	99	0.02
76	I	Chrysene-d12	1.000	1.000	0.0	100	0.01
77		Benzidine	0.545	0.560	-2.8	101	0.02
78		Pyrene	1.198	1.141	4.8	99	0.02
79	S	Terphenyl-d14	1.166	1.115	4.4	99	0.02
80		Butylbenzylphthalate	0.533	0.534	-0.2	102	0.02
81		Benzo(a)anthracene	1.256	1.205	4.1	100	0.00
82		3,3'-Dichlorobenzidine	0.466	0.478	-2.6	101	0.02
83		Chrysene	1.190	1.128	5.2	100	0.02
84		Bis(2-ethylhexyl)phthalate	0.783	0.787	-0.5	103	0.01
85	c	Di-n-octyl phthalate	1.281	1.305	-1.9	102	0.02
86	I	Perylene-d12	1.000	1.000	0.0	101	0.02
87		Indeno(1,2,3-cd)pyrene	1.460	1.437	1.6	102	0.03
88		Benzo(b)fluoranthene	1.145	1.111	3.0	102	0.04
89		Benzo(k)fluoranthene	1.180	1.142	3.2	101	0.02
90	C	Benzo(a)pyrene	1.099	1.073	2.4	102	0.02
91		Dibenzo(a,h)anthracene	1.188	1.173	1.3	102	0.03
92		Benzo(g,h,i)perylene	1.181	1.160	1.8	102	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	-0.04
2		1,4-Dioxane	40.000	38.399	4.0	100	-0.02
3		Pyridine	40.000	38.903	2.7	95	-0.02
4		n-Nitrosodimethylamine	40.000	40.858	-2.1	102	-0.02
5	S	2-Fluorophenol	80.000	77.154	3.6	98	-0.03
6		Aniline	40.000	38.474	3.8	93	-0.03
7	S	Phenol-d6	80.000	78.114	2.4	95	-0.03
8		2-Chlorophenol	40.000	39.274	1.8	96	-0.04
9		Benzaldehyde	40.000	38.807	3.0	94	-0.04
10	C	Phenol	40.000	38.823	2.9	95	-0.04
11		bis(2-Chloroethyl)ether	40.000	37.549	6.1	94	-0.04
12		1,3-Dichlorobenzene	40.000	37.521	6.2	95	-0.03
13	C	1,4-Dichlorobenzene	40.000	37.942	5.1	95	-0.04
14		1,2-Dichlorobenzene	40.000	38.161	4.6	96	-0.03
15		Benzyl Alcohol	40.000	41.153	-2.9	97	-0.03
16		2,2'-oxybis(1-Chloropropane	40.000	37.827	5.4	93	-0.04
17		2-Methylphenol	40.000	39.754	0.6	96	-0.04
18		Hexachloroethane	40.000	38.265	4.3	98	-0.03
19	P	n-Nitroso-di-n-propylamine	40.000	39.841	0.4	94	-0.02
20		3+4-Methylphenols	40.000	39.761	0.6	94	-0.03
21	I	Naphthalene-d8	20.000	20.000	0.0	94	-0.02
22		Acetophenone	40.000	38.471	3.8	93	-0.02
23	S	Nitrobenzene-d5	80.000	79.021	1.2	96	-0.03
24		Nitrobenzene	40.000	39.340	1.6	96	-0.03
25		Isophorone	40.000	38.643	3.4	93	-0.03
26	C	2-Nitrophenol	40.000	40.243	-0.6	94	-0.04
27		2,4-Dimethylphenol	40.000	39.475	1.3	94	-0.03
28		bis(2-Chloroethoxy)methane	40.000	38.310	4.2	93	-0.02
29	C	2,4-Dichlorophenol	40.000	39.952	0.1	94	-0.02
30		1,2,4-Trichlorobenzene	40.000	38.698	3.3	98	-0.03
31		Naphthalene	40.000	38.138	4.7	95	-0.02
32		Benzoic acid	40.000	39.096	2.3	96	-0.02
33		4-Chloroaniline	40.000	40.266	-0.7	94	-0.02
34	C	Hexachlorobutadiene	40.000	39.374	1.6	99	-0.02
35		Caprolactam	40.000	39.255	1.9	89	-0.01
36	C	4-Chloro-3-methylphenol	40.000	38.834	2.9	91	-0.02
37		2-Methylnaphthalene	40.000	37.971	5.1	93	-0.01
38		1-Methylnaphthalene	40.000	37.262	6.8	91	-0.02
39	I	Acenaphthene-d10	20.000	20.000	0.0	93	-0.02
40		1,2,4,5-Tetrachlorobenzene	40.000	37.193	7.0	91	-0.01
41	P	Hexachlorocyclopentadiene	40.000	42.442	-6.1	97	-0.02
42	S	2,4,6-Tribromophenol	80.000	82.297	-2.9	100	0.00
43	C	2,4,6-Trichlorophenol	40.000	39.595	1.0	93	-0.01
44		2,4,5-Trichlorophenol	40.000	39.527	1.2	95	-0.01
45	S	2-Fluorobiphenyl	80.000	75.203	6.0	93	-0.01
46		1,1'-Biphenyl	40.000	37.428	6.4	93	0.00
47		2-Chloronaphthalene	40.000	37.446	6.4	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48		2-Nitroaniline	40.000	39.458	1.4	91	-0.02
49		Acenaphthylene	40.000	37.780	5.5	92	-0.01
50		Dimethylphthalate	40.000	37.409	6.5	91	0.00
51		2,6-Dinitrotoluene	40.000	38.817	3.0	93	-0.02
52	C	Acenaphthene	40.000	38.133	4.7	93	-0.01
53		3-Nitroaniline	40.000	40.387	-1.0	91	-0.02
54	P	2,4-Dinitrophenol	40.000	38.435	3.9	96	0.00
55		Dibenzofuran	40.000	38.731	3.2	95	0.00
56	P	4-Nitrophenol	40.000	38.161	4.6	92	-0.01
57		2,4-Dinitrotoluene	40.000	40.648	-1.6	94	-0.01
58		Fluorene	40.000	39.503	1.2	97	0.00
59		2,3,4,6-Tetrachlorophenol	40.000	39.708	0.7	95	-0.01
60		Diethylphthalate	40.000	39.356	1.6	98	0.00
61		4-Chlorophenyl-phenylether	40.000	39.510	1.2	98	0.00
62		4-Nitroaniline	40.000	43.838	-9.6	98	0.00
63		Azobenzene	40.000	39.973	0.1	98	0.00
64	I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65		4,6-Dinitro-2-methylphenol	40.000	39.242	1.9	97	0.00
66	c	n-Nitrosodiphenylamine	40.000	37.904	5.2	96	-0.01
67		4-Bromophenyl-phenylether	40.000	38.302	4.2	98	0.00
68		Hexachlorobenzene	40.000	37.499	6.3	98	0.00
69		Atrazine	40.000	38.949	2.6	99	0.00
70	C	Pentachlorophenol	40.000	42.306	-5.8	102	0.00
71		Phenanthrene	40.000	37.268	6.8	97	0.00
72		Anthracene	40.000	38.282	4.3	98	0.00
73		Carbazole	40.000	39.067	2.3	97	0.00
74		Di-n-butylphthalate	40.000	40.046	-0.1	102	0.02
75	C	Fluoranthene	40.000	38.916	2.7	99	0.02
76	I	Chrysene-d12	20.000	20.000	0.0	100	0.01
77		Benzidine	40.000	41.133	-2.8	101	0.02
78		Pyrene	40.000	38.096	4.8	99	0.02
79	S	Terphenyl-d14	80.000	76.446	4.4	99	0.02
80		Butylbenzylphthalate	40.000	40.056	-0.1	102	0.02
81		Benzo(a)anthracene	40.000	38.383	4.0	100	0.00
82		3,3'-Dichlorobenzidine	40.000	41.031	-2.6	101	0.02
83		Chrysene	40.000	37.889	5.3	100	0.02
84		Bis(2-ethylhexyl)phthalate	40.000	40.184	-0.5	103	0.01
85	c	Di-n-octyl phthalate	40.000	40.763	-1.9	102	0.02
86	I	Perylene-d12	20.000	20.000	0.0	101	0.02
87		Indeno(1,2,3-cd)pyrene	40.000	39.371	1.6	102	0.03
88		Benzo(b)fluoranthene	40.000	38.817	3.0	102	0.04
89		Benzo(k)fluoranthene	40.000	38.741	3.1	101	0.02
90	C	Benzo(a)pyrene	40.000	39.054	2.4	102	0.02
91		Dibenzo(a,h)anthracene	40.000	39.513	1.2	102	0.03
92		Benzo(g,h,i)perylene	40.000	39.279	1.8	102	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024625.D
Acq On : 14 May 2025 11:35
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 14 13:09:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0

QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050023.D
Acq On : 28 Apr 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Apr 28 18:09:16 2025





AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.8	131521	PASS
68	69	0.00	2	1.3	2026	PASS
69	198	0.00	100	28.5	157569	PASS
70	69	0.00	2	0.4	660	PASS
127	198	10	80	35.7	197099	PASS
197	198	0.00	2	0.4	2228	PASS
198	198	100	100	100.0	552704	PASS
199	198	5	9	7.0	38461	PASS
275	198	10	60	26.2	144571	PASS
365	198	1	100	3.5	19537	PASS
441	198	0.01	100	11.6	64259	PASS
442	442	50	100	100.0	413099	PASS
443	442	15	24	19.3	79557	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050023.D
Acq On : 28 Apr 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Apr 28 15:37:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 09 04:00:55 2025
Response via : Initial Calibration







TIC: BM050023.D\data.ms

(70) Pentachlorophenol (C)

16.821min (+ 0.006) 652247.26 ng

response 613835

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.30	64.13
264.00	62.30	63.38
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050023.D
Acq On : 28 Apr 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Apr 28 15:37:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 09 04:00:55 2025
Response via : Initial Calibration







TIC: BM050023.D\data.ms

(77) Benzidine

19.404min (+ 0.000) 1761877.84 ng

response 3412576

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.30	14.06
183.00	11.00	11.29
0.00	0.00	0.00

Instrument :
BNA_M
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/28/2025	BNA_M	BM050023.D
Compound Name	**Response**	**Retention Time**
DDT	1978095	20.639
DDD	17600	20.251
DDE	1351	19.692
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
18951	1997046	0.95

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050139.D
Acq On : 08 May 2025 09:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu May 08 12:03:21 2025



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.7	118511	PASS
68	69	0.00	2	1.4	2018	PASS
69	198	0.00	100	25.8	140810	PASS
70	69	0.00	2	0.4	628	PASS
127	198	10	80	33.5	182635	PASS
197	198	0.00	2	0.4	2160	PASS
198	198	100	100	100.0	545024	PASS
199	198	5	9	6.7	36635	PASS
275	198	10	60	27.3	149061	PASS
365	198	1	100	3.9	21040	PASS
441	198	0.01	100	13.6	74341	PASS
442	442	50	100	100.0	468224	PASS
443	442	15	24	19.5	91352	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050139.D
Acq On : 08 May 2025 09:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: May 08 11:50:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration







TIC: BM050139.D\data.ms

(70) Pentachlorophenol (C)

16.810min (-0.011) 902047.65 ng

response 580377

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.00	62.90
264.00	62.00	62.46
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050139.D
Acq On : 08 May 2025 09:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: May 08 11:50:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration







TIC: BM050139.D\data.ms

(77) Benzidine

19.392min (-0.012) 27949.85 ng

response 2854293

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	15.07
183.00	11.10	11.46
0.00	0.00	0.00

Instrument :
BNA_M
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/8/2025	BNA_M	BM050139.D
Compound Name	**Response**	**Retention Time**
DDT	1691185	20.621
DDD	32020	20.186
DDE	2862	19.674
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
34882	1726067	2.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024274.D
Acq On : 14 Apr 2025 10:25
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 15 04:48:42 2025



AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2463

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.4	129796	PASS
68	69	0.00	2	1.8	2815	PASS
69	198	0.00	100	36.0	153442	PASS
70	69	0.00	2	0.5	836	PASS
127	198	10	80	48.5	206858	PASS
197	198	0.00	2	0.5	2122	PASS
198	198	100	100	100.0	426527	PASS
199	198	5	9	6.9	29549	PASS
275	198	10	60	29.1	124199	PASS
365	198	1	100	4.1	17663	PASS
441	198	0.01	100	15.2	64862	PASS
442	442	100	100	100.0	422528	PASS
443	442	15	24	19.4	82090	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024274.D
Acq On : 14 Apr 2025 10:25
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Apr 14 15:07:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 15:04:09 2025
Response via : Initial Calibration







TIC: BP024274.D\data.ms

(70) Pentachlorophenol (C)

16.822min (+ 0.012) 3117222.42 ng

response 611847

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	65.61
264.00	62.20	62.81
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041425\
Data File : BP024274.D
Acq On : 14 Apr 2025 10:25
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Apr 14 15:07:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 14 15:04:09 2025
Response via : Initial Calibration







TIC: BP024274.D\data.ms

(77) Benzidine

19.504min (+ 0.006) 7415471.13 ng

response 2354373

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.60
183.00	11.50	11.64
0.00	0.00	0.00

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/14/2025	BNA_P	BP024274.D
Compound Name	**Response**	**Retention Time**
DDT	1715151	20.798
DDD	26002	20.386
DDE	1572	19.81
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
27574	1742725	1.58

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024567.D
Acq On : 08 May 2025 13:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Apr 18 12:04:48 2025





AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.9	120114	PASS
68	69	0.00	2	1.2	1591	PASS
69	198	0.00	100	30.1	129294	PASS
70	69	0.00	2	0.6	754	PASS
127	198	10	80	45.5	195716	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	430133	PASS
199	198	5	9	6.8	29109	PASS
275	198	10	60	33.5	143952	PASS
365	198	1	100	5.2	22422	PASS
441	198	0.01	100	20.7	89080	PASS
442	442	100	100	100.0	570069	PASS
443	442	15	24	19.3	110223	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024567.D
Acq On : 08 May 2025 13:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 08 15:11:37 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration







TIC: BP024567.D\data.ms

(70) Pentachlorophenol (C)

16.786min (-0.012) 35045.60 ng

response 573194

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	64.70
264.00	62.20	62.72
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024567.D
Acq On : 08 May 2025 13:24
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 08 15:11:37 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration







TIC: BP024567.D\data.ms

(77) Benzidine

19.468min (-0.012) 33720.75 ng

response 2634734

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	13.96
183.00	11.50	12.24
0.00	0.00	0.00

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/8/2025	BNA_P	BP024567.D
Compound Name	**Response**	**Retention Time**
DDT	1299893	20.756
DDD	30641	20.298
DDE	2361	19.768
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
33002	1332895	2.48

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024613.D
Acq On : 13 May 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue May 13 16:21:39 2025



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2449

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.5	63968	PASS
68	69	0.00	2	0.6	438	PASS
69	198	0.00	100	36.2	69152	PASS
70	69	0.00	2	0.6	419	PASS
127	198	10	80	49.3	94192	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	191223	PASS
199	198	5	9	6.6	12661	PASS
275	198	10	60	33.1	63293	PASS
365	198	1	100	5.4	10286	PASS
441	198	0.01	100	19.4	37080	PASS
442	442	100	100	100.0	238569	PASS
443	442	15	24	19.4	46397	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024613.D
Acq On : 13 May 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 13 12:28:43 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 12:28:24 2025
Response via : Initial Calibration







TIC: BP024613.D\data.ms

(70) Pentachlorophenol (C)

16.745min (-0.000) 42670.46 ng

response 398426

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.60	63.59
264.00	64.00	62.12
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051325\
Data File : BP024613.D
Acq On : 13 May 2025 10:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 13 12:28:43 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 12:28:24 2025
Response via : Initial Calibration







TIC: BP024613.D\data.ms

(77) Benzidine

19.439min (+ 0.006) 5637.89 ng

response 2074486

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.10	14.46
183.00	12.10	12.23
0.00	0.00	0.00

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/13/2025	BNA_P	BP024613.D
Compound Name	**Response**	**Retention Time**
DDT	750509	20.716
DDD	13774	20.269
DDE	1248	19.733
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
15022	765531	1.96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024624.D
Acq On : 14 May 2025 10:29
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue May 13 16:21:39 2025



AutoFind: Scans 2460, 2461, 2462; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.4	53822	PASS
68	69	0.00	2	1.9	1106	PASS
69	198	0.00	100	34.4	57285	PASS
70	69	0.00	2	0.6	340	PASS
127	198	10	80	48.6	80786	PASS
197	198	0.00	2	0.2	309	PASS
198	198	100	100	100.0	166319	PASS
199	198	5	9	6.8	11274	PASS
275	198	10	60	33.4	55576	PASS
365	198	1	100	5.8	9657	PASS
441	198	0.01	100	21.2	35184	PASS
442	442	100	100	100.0	226240	PASS
443	442	15	24	19.5	44145	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024624.D
Acq On : 14 May 2025 10:29
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 14 11:18:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration







TIC: BP024624.D\data.ms

(70) Pentachlorophenol (C)

16.739min (-0.006) 4865870.98 ng

response 327184

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.30	63.09
264.00	63.90	61.98
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051425\
Data File : BP024624.D
Acq On : 14 May 2025 10:29
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: May 14 11:18:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP051325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 13 16:21:39 2025
Response via : Initial Calibration







TIC: BP024624.D\data.ms

(77) Benzidine

19.428min (+ 0.012) 72527.19 ng

response 2010903

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.07
183.00	12.30	12.28
0.00	0.00	0.00

Instrument :
BNA_P
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
5/14/2025	BNA_P	BP024624.D
Compound Name	**Response**	**Retention Time**
DDT	712469	20.71
DDD	10959	20.304
DDE	1350	19.728
SUM(DDD+DDE)	**SUM(DDT+DDD+DDE)**	**% Breakdown Of DDT**
12309	724778	1.70

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark			Date Collected:	
Project:	NJ Soil PT			Date Received:	
Client Sample ID:	PB167904BL			SDG No.:	Q1872
Lab Sample ID:	PB167904BL			Matrix:	SOIL
Analytical Method:	8270E			% Solid:	100
Sample Wt/Vol:	30.01	Units:	g	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	SVOCMS Group4
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024569.D	1	05/08/25 09:15	05/08/25 15:26	PB167904

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1912-24-9	Atrazine	34.0	U	34.0	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	88.3		18 - 107	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	95.4		20 - 109	95%	SPK: 100
1718-51-0	Terphenyl-d14	95.5		10 - 105	96%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	198000	7.699			
1146-65-2	Naphthalene-d8	737000	10.475			
15067-26-2	Acenaphthene-d10	410000	14.334			
1517-22-2	Phenanthrene-d10	729000	17.139			
1719-03-5	Chrysene-d12	698000	21.58			
1520-96-3	Perylene-d12	845000	24.921			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024569.D
Acq On : 08 May 2025 15:26
Operator : RC/JU
Sample : PB167904BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Quant Time: May 08 15:46:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	197835	20.000	ng	-0.02
21) Naphthalene-d8	10.475	136	736878	20.000	ng	#-0.02
39) Acenaphthene-d10	14.334	164	410163	20.000	ng	-0.02
64) Phenanthrene-d10	17.139	188	728812	20.000	ng	0.00
76) Chrysene-d12	21.580	240	697724	20.000	ng	-0.01
86) Perylene-d12	24.921	264	845260	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.334	112	1541430	128.829	ng	-0.01
7) Phenol-d6	6.893	99	1806733	110.279	ng	-0.02
23) Nitrobenzene-d5	8.852	82	1140879	88.284	ng	-0.02
42) 2,4,6-Tribromophenol	15.851	330	827418	145.805	ng	-0.01
45) 2-Fluorobiphenyl	12.940	172	2556924	95.448	ng	-0.02
79) Terphenyl-d14	19.857	244	3306800	95.529	ng	-0.02
Target Compounds						Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024569.D
Acq On : 08 May 2025 15:26
Operator : RC/JU
Sample : PB167904BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Quant Time: May 08 15:46:33 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.699 min Scan# 7
Delta R.T. -0.023 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Tgt Ion:152 Resp: 197835

Ion	Ratio	Lower	Upper
152	100		
150	158.6	124.9	187.3
115	61.6	47.7	71.5





#5
2-Fluorophenol
Concen: 128.829 ng
RT: 5.334 min Scan# 382
Delta R.T. -0.012 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Tgt Ion:112 Resp: 1541430

Ion	Ratio	Lower	Upper
112	100		
64	52.3	41.1	61.7
63	28.0	20.6	30.8





#7
Phenol-d6
Concen: 110.279 ng
RT: 6.893 min Scan# 64
Delta R.T. -0.017 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Tgt Ion: 99 Resp: 1806733

Ion	Ratio	Lower	Upper
99	100		
42	16.6	10.6	16.0#
71	35.5	23.7	35.5





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.475 min Scan# 1256
Delta R.T. -0.023 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Tgt Ion:136 Resp: 736878

Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.2	13.8
54	6.8	6.0	9.0
68	5.0	5.5	8.3#





#23
Nitrobenzene-d5
Concen: 88.284 ng
RT: 8.852 min Scan# 9
Delta R.T. -0.017 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Tgt Ion: 82 Resp: 1140879

Ion	Ratio	Lower	Upper
82	100		
128	47.4	36.5	54.7
54	46.7	33.4	50.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.334 min Scan# 1912
Delta R.T. -0.017 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Tgt Ion:164 Resp: 410163

Ion	Ratio	Lower	Upper
164	100		
162	102.2	80.6	120.8
160	45.8	35.3	52.9





#42
2,4,6-Tribromophenol
Concen: 145.805 ng
RT: 15.851 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Tgt Ion:330 Resp: 827418

Ion	Ratio	Lower	Upper
330	100		
332	97.7	77.1	115.7
141	28.9	24.7	37.1





#45
2-Fluorobiphenyl
Concen: 95.448 ng
RT: 12.940 min Scan# 1675
Delta R.T. -0.023 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Tgt Ion:172 Resp: 2556924

Ion	Ratio	Lower	Upper
172	100		
171	36.2	28.2	42.2
170	23.6	18.6	27.8





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Tgt Ion:188 Resp: 728812

Ion	Ratio	Lower	Upper
188	100		
94	9.2	8.6	13.0
80	10.5	9.8	14.6





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.580 min Scan# 3144
Delta R.T. -0.012 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Tgt Ion:240 Resp: 697724

Ion	Ratio	Lower	Upper
240	100		
120	10.8	9.8	14.8
236	25.8	20.9	31.3





#79
Terphenyl-d14
Concen: 95.529 ng
RT: 19.857 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Instrument :
BNA_P
ClientSampleId :
PB167904BL

Tgt Ion:244 Resp: 3306800

Ion	Ratio	Lower	Upper
244	100		
212	6.9	5.6	8.4
122	9.6	9.4	14.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.921 min Scan# 3712
Delta R.T. -0.012 min
Lab File: BP024569.D
Acq: 08 May 2025 15:26

Tgt Ion:264 Resp: 845260

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.9	28.3
265	21.4	17.8	26.6



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark			Date Collected:	
Project:	NJ Soil PT			Date Received:	
Client Sample ID:	PB167904BS			SDG No.:	Q1872
Lab Sample ID:	PB167904BS			Matrix:	SOIL
Analytical Method:	8270E			% Solid:	100
Sample Wt/Vol:	30.02	Units:	g	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	SVOCMS Group4
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024570.D	1	05/08/25 09:15	05/08/25 16:07	PB167904

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1912-24-9	Atrazine	2500		34.0	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	88.0		18 - 107	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	91.1		20 - 109	91%	SPK: 100
1718-51-0	Terphenyl-d14	102		10 - 105	102%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	175000	7.705			
1146-65-2	Naphthalene-d8	733000	10.475			
15067-26-2	Acenaphthene-d10	498000	14.328			
1517-22-2	Phenanthrene-d10	980000	17.134			
1719-03-5	Chrysene-d12	1020000	21.575			
1520-96-3	Perylene-d12	1020000	24.898			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024570.D
Acq On : 08 May 2025 16:07
Operator : RC/JU
Sample : PB167904BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 08 16:42:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.705	152	175167	20.000	ng		-0.02
21) Naphthalene-d8	10.475	136	732511	20.000	ng		#-0.02
39) Acenaphthene-d10	14.328	164	497695	20.000	ng		-0.02
64) Phenanthrene-d10	17.134	188	979666	20.000	ng		#-0.01
76) Chrysene-d12	21.575	240	1019336	20.000	ng		#-0.02
86) Perylene-d12	24.898	264	1024841	20.000	ng		-0.04
System Monitoring Compounds							
5) 2-Fluorophenol	5.328	112	1467895	138.559	ng		-0.02
7) Phenol-d6	6.899	99	1846598	127.298	ng		-0.01
23) Nitrobenzene-d5	8.852	82	1130867	88.031	ng		-0.02
42) 2,4,6-Tribromophenol	15.851	330	1208690	175.532	ng		-0.01
45) 2-Fluorobiphenyl	12.946	172	2961395	91.104	ng		-0.02
79) Terphenyl-d14	19.857	244	5174977	102.330	ng		-0.02
Target Compounds							Qvalue
2) 1,4-Dioxane	3.276	88	155062	34.604	ng	#	92
3) Pyridine	3.664	79	392229	33.148	ng	#	89
4) n-Nitrosodimethylamine	3.581	42	197341	50.253	ng	#	71
6) Aniline	7.046	93	584331	45.078	ng		98
8) 2-Chlorophenol	7.281	128	569451	48.144	ng		98
9) Benzaldehyde	6.858	77	215548	30.039	ng		99
10) Phenol	6.922	94	668130	45.914	ng		89
11) bis(2-Chloroethyl)ether	7.134	93	472724	40.319	ng		99
12) 1,3-Dichlorobenzene	7.593	146	571193	44.856	ng		99
13) 1,4-Dichlorobenzene	7.740	146	577199	44.675	ng		99
14) 1,2-Dichlorobenzene	8.052	146	560913	44.960	ng		99
15) Benzyl Alcohol	7.946	79	465865	49.660	ng		97
16) 2,2'-oxybis(1-Chloropr...	8.222	45	548968	43.318	ng		100
17) 2-Methylphenol	8.152	107	475468	50.508	ng		96
18) Hexachloroethane	8.769	117	218631	46.825	ng		95
19) n-Nitroso-di-n-propyla...	8.505	70	384521	42.847	ng		94
20) 3+4-Methylphenols	8.475	107	653624	50.041	ng		98
22) Acetophenone	8.516	105	794605	43.828	ng	#	96
24) Nitrobenzene	8.893	77	562229	44.241	ng		98
25) Isophorone	9.416	82	1102296	47.405	ng		97
26) 2-Nitrophenol	9.599	139	315406	50.716	ng	#	93
27) 2,4-Dimethylphenol	9.663	122	552984	70.815	ng		97
28) bis(2-Chloroethoxy)met...	9.893	93	675879	43.027	ng		99
29) 2,4-Dichlorophenol	10.134	162	561902	52.774	ng		99
30) 1,2,4-Trichlorobenzene	10.340	180	536581	47.036	ng		100
31) Naphthalene	10.522	128	1677067	43.463	ng		99
32) Benzoic acid	9.846	122	374414	41.149	ng		99
33) 4-Chloroaniline	10.646	127	356521	26.237	ng		99
34) Hexachlorobutadiene	10.805	225	344271	51.356	ng		99
35) Caprolactam	11.446	113	196242m	49.951	ng		
36) 4-Chloro-3-methylphenol	11.775	107	658827	53.124	ng		99
37) 2-Methylnaphthalene	12.134	142	1125224	42.049	ng		97
38) 1-Methylnaphthalene	12.357	142	1190789	45.487	ng		98
40) 1,2,4,5-Tetrachloroben...	12.510	216	641022	46.835	ng		98
41) Hexachlorocyclopentadiene	12.487	237	889691	183.797	ng		98
43) 2,4,6-Trichlorophenol	12.757	196	475847	52.293	ng		97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024570.D
Acq On : 08 May 2025 16:07
Operator : RC/JU
Sample : PB167904BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 08 16:42:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
44) 2,4,5-Trichlorophenol	12.840	196	524586	52.059	ng		98
46) 1,1'-Biphenyl	13.157	154	1632780	44.632	ng		98
47) 2-Chloronaphthalene	13.199	162	1234859	45.164	ng		98
48) 2-Nitroaniline	13.410	65	402707	52.508	ng		94
49) Acenaphthylene	14.051	152	2132979	49.549	ng		99
50) Dimethylphthalate	13.793	163	1752656	48.425	ng		100
51) 2,6-Dinitrotoluene	13.910	165	383012	49.835	ng		91
52) Acenaphthene	14.398	154	1225957	44.922	ng		99
53) 3-Nitroaniline	14.251	138	318498	39.430	ng		99
54) 2,4-Dinitrophenol	14.463	184	493492	109.007	ng		91
55) Dibenzofuran	14.740	168	1968680	44.132	ng		96
56) 4-Nitrophenol	14.593	139	673353	96.520	ng		90
57) 2,4-Dinitrotoluene	14.710	165	569881	54.357	ng		98
58) Fluorene	15.393	166	1621926	46.967	ng		100
59) 2,3,4,6-Tetrachlorophenol	14.975	232	470855	50.663	ng		99
60) Diethylphthalate	15.175	149	1770392	47.466	ng		100
61) 4-Chlorophenyl-phenyle...	15.393	204	813013	48.482	ng		98
62) 4-Nitroaniline	15.434	138	409802	47.924	ng		85
63) Azobenzene	15.693	77	1492279	43.531	ng		94
65) 4,6-Dinitro-2-methylph...	15.493	198	321014	51.974	ng		94
66) n-Nitrosodiphenylamine	15.616	169	1421507	48.614	ng		99
67) 4-Bromophenyl-phenylether	16.304	248	570417	53.238	ng		98
68) Hexachlorobenzene	16.428	284	699484	54.940	ng		99
69) Atrazine	16.592	200	550524	73.922	ng		97
70) Pentachlorophenol	16.787	266	946297	106.541	ng		100
71) Phenanthrene	17.175	178	2499026	46.760	ng		100
72) Anthracene	17.269	178	2594917	50.379	ng		100
73) Carbazole	17.557	167	2428994	48.263	ng		99
74) Di-n-butylphthalate	18.139	149	3175548	50.645	ng		99
75) Fluoranthene	19.263	202	2999313	47.187	ng		96
77) Benzidine	19.463	184	1434382	111.012	ng		99
78) Pyrene	19.633	202	3062088	47.269	ng		99
80) Butylbenzylphthalate	20.586	149	1408621	50.767	ng		96
81) Benzo(a)anthracene	21.545	228	3060293	48.302	ng		100
82) 3,3'-Dichlorobenzidine	21.480	252	753351	33.877	ng		99
83) Chrysene	21.622	228	2811892	46.325	ng		100
84) Bis(2-ethylhexyl)phtha...	21.480	149	1999200	49.150	ng		100
85) Di-n-octyl phthalate	22.751	149	3196061	48.332	ng		100
87) Indeno(1,2,3-cd)pyrene	28.674	276	3290167	46.243	ng	#	96
88) Benzo(b)fluoranthene	23.845	252	2918757	47.514	ng	#	97
89) Benzo(k)fluoranthene	23.921	252	2931273	49.400	ng	#	97
90) Benzo(a)pyrene	24.751	252	2750615	51.909	ng	#	97
91) Dibenzo(a,h)anthracene	28.751	278	2694984	45.635	ng		97
92) Benzo(g,h,i)perylene	29.845	276	2599244	43.241	ng	#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050825\
Data File : BP024570.D
Acq On : 08 May 2025 16:07
Operator : RC/JU
Sample : PB167904BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 08 16:42:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.705 min Scan# 7
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:152 Resp: 17516

Ion	Ratio	Lower	Upper
152	100		
150	153.7	124.9	187.3
115	61.7	47.7	71.5



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#2
1,4-Dioxane
Concen: 34.604 ng
RT: 3.276 min Scan# 32
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 88 Resp: 155062

Ion	Ratio	Lower	Upper
88	100		
58	68.7	51.4	77.2
43	28.3	17.1	25.7#





#3
Pyridine
Concen: 33.148 ng
RT: 3.664 min Scan# 9
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 79 Resp: 39222

Ion	Ratio	Lower	Upper
79	100		
52	65.0	45.4	68.2
51	34.6	22.6	33.8#



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#4
n-Nitrosodimethylamine
Concen: 50.253 ng
RT: 3.581 min Scan# 84
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 42 Resp: 197341

Ion	Ratio	Lower	Upper
42	100		
74	132.7	139.5	209.3#
44	6.3	5.8	8.8





#5
2-Fluorophenol
Concen: 138.559 ng
RT: 5.328 min Scan# 3
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:112 Resp: 146789

Ion	Ratio	Lower	Upper
112	100		
64	53.7	41.1	61.7
63	29.0	20.6	30.8

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#6
Aniline
Concen: 45.078 ng
RT: 7.046 min Scan# 673
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 93 Resp: 584331

Ion	Ratio	Lower	Upper
93	100		
66	37.8	29.1	43.7
65	18.7	14.7	22.1





#7
Phenol-d6
Concen: 127.298 ng
RT: 6.899 min Scan# 64
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 99 Resp: 184659

Ion	Ratio	Lower	Upper
99	100		
42	17.0	10.6	16.0#
71	35.9	23.7	35.5#



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#8
2-Chlorophenol
Concen: 48.144 ng
RT: 7.281 min Scan# 713
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:128 Resp: 569451

Ion	Ratio	Lower	Upper
128	100		
130	33.0	12.9	52.9
64	41.0	22.7	62.7





#9
Benzaldehyde
Concen: 30.039 ng
RT: 6.858 min Scan# 64
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 77 Resp: 21554

Ion	Ratio	Lower	Upper
77	100		
105	100.7	79.2	119.2
106	96.9	76.2	116.2



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#10
Phenol
Concen: 45.914 ng
RT: 6.922 min Scan# 652
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 94 Resp: 668130

Ion	Ratio	Lower	Upper
94	100		
65	28.1	4.5	44.5
66	40.2	12.4	52.4





#11
bis(2-Chloroethyl)ether
Concen: 40.319 ng
RT: 7.134 min Scan# 6
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 93 Resp: 47272

Ion	Ratio	Lower	Upper
93	100		
63	70.2	48.6	88.6
95	32.2	12.3	52.3

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#12
1,3-Dichlorobenzene
Concen: 44.856 ng
RT: 7.593 min Scan# 766
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:146 Resp: 571193

Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.4	77.0
75	29.5	22.9	34.3





#13
1,4-Dichlorobenzene
Concen: 44.675 ng
RT: 7.740 min Scan# 7
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:146 Resp: 57719

Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.1	76.7
111	41.4	32.6	48.8



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#14
1,2-Dichlorobenzene
Concen: 44.960 ng
RT: 8.052 min Scan# 844
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:146 Resp: 560913

Ion	Ratio	Lower	Upper
146	100		
148	63.5	51.0	76.4
111	43.3	33.8	50.8





#15
Benzyl Alcohol
Concen: 49.660 ng
RT: 7.946 min Scan# 8
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 79 Resp: 46586

Ion	Ratio	Lower	Upper
79	100		
108	76.0	65.1	97.7
77	63.3	50.7	76.1

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#16
2,2'-oxybis(1-Chloropropane)
Concen: 43.318 ng
RT: 8.222 min Scan# 873
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 45 Resp: 548968

Ion	Ratio	Lower	Upper
45	100		
77	17.5	0.0	37.7
79	13.3	0.0	33.5





#17
2-Methylphenol
Concen: 50.508 ng
RT: 8.152 min Scan# 8
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:107 Resp: 47546

Ion	Ratio	Lower	Upper
107	100		
108	109.7	89.8	134.6
77	45.8	34.2	51.4
79	46.0	33.8	50.8



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#18
Hexachloroethane
Concen: 46.825 ng
RT: 8.769 min Scan# 966
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:117 Resp: 218631

Ion	Ratio	Lower	Upper
117	100		
119	95.4	77.8	116.6
201	102.7	75.9	113.9





#19
n-Nitroso-di-n-propylamine
Concen: 42.847 ng
RT: 8.505 min Scan# 9
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 70 Resp: 38452

Ion	Ratio	Lower	Upper
70	100		
42	51.6	37.3	55.9
101	11.0	8.2	12.4
130	25.0	18.6	28.0



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#20
3+4-Methylphenols
Concen: 50.041 ng
RT: 8.475 min Scan# 916
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:107 Resp: 653624

Ion	Ratio	Lower	Upper
107	100		
108	88.1	69.8	109.8
77	33.7	11.5	51.5
79	28.3	7.8	47.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.475 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:136 Resp: 73251

Ion	Ratio	Lower	Upper
136	100		
137	10.8	9.2	13.8
54	6.5	6.0	9.0
68	5.1	5.5	8.3#



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#22
Acetophenone
Concen: 43.828 ng
RT: 8.516 min Scan# 923
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:105 Resp: 794605

Ion	Ratio	Lower	Upper
105	100		
71	2.9	1.7	2.5#
51	24.7	16.9	25.3
120	23.3	18.6	28.0





#23
Nitrobenzene-d5
Concen: 88.031 ng
RT: 8.852 min Scan# 9
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 82 Resp: 113086

Ion	Ratio	Lower	Upper
82	100		
128	46.5	36.5	54.7
54	48.2	33.4	50.2



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#24
Nitrobenzene
Concen: 44.241 ng
RT: 8.893 min Scan# 987
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 77 Resp: 562229

Ion	Ratio	Lower	Upper
77	100		
123	50.0	39.2	58.8
65	13.6	9.8	14.6





#25
Isophorone
Concen: 47.405 ng
RT: 9.416 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 82 Resp: 110229

Ion	Ratio	Lower	Upper
82	100		
95	8.0	5.4	8.2
138	19.3	14.4	21.6

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#26
2-Nitrophenol
Concen: 50.716 ng
RT: 9.599 min Scan# 1107
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:139 Resp: 315406

Ion	Ratio	Lower	Upper
139	100		
109	33.9	21.8	32.6#
65	47.6	36.2	54.2





#27
2,4-Dimethylphenol
Concen: 70.815 ng
RT: 9.663 min Scan# 11
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:122 Resp: 55298

Ion	Ratio	Lower	Upper
122	100		
107	120.7	93.7	140.5
121	59.2	48.2	72.2



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#28
bis(2-Chloroethoxy)methane
Concen: 43.027 ng
RT: 9.893 min Scan# 1157
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 93 Resp: 675879

Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.0	39.0
123	12.5	9.1	13.7





#29
2,4-Dichlorophenol
Concen: 52.774 ng
RT: 10.134 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:162 Resp: 56190

Ion	Ratio	Lower	Upper
162	100		
164	64.6	43.3	83.3
98	36.4	16.6	56.6



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#30
1,2,4-Trichlorobenzene
Concen: 47.036 ng
RT: 10.340 min Scan# 1233
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:180 Resp: 536581

Ion	Ratio	Lower	Upper
180	100		
182	96.3	77.3	115.9
145	31.5	25.1	37.7





#31
Naphthalene
Concen: 43.463 ng
RT: 10.522 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:128 Resp: 167706

Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.6	13.0
127	13.6	10.3	15.5



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#32
Benzoic acid
Concen: 41.149 ng
RT: 9.846 min Scan# 1149
Delta R.T. 0.024 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:122 Resp: 374414

Ion	Ratio	Lower	Upper
122	100		
105	125.8	104.8	144.8
77	91.0	69.2	109.2





#33
4-Chloroaniline
Concen: 26.237 ng
RT: 10.646 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:127 Resp: 35652

Ion	Ratio	Lower	Upper
127	100		
129	32.0	26.1	39.1
65	28.1	22.1	33.1
92	18.9	15.0	22.4



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#34
Hexachlorobutadiene
Concen: 51.356 ng
RT: 10.805 min Scan# 1312
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:225 Resp: 344271

Ion	Ratio	Lower	Upper
225	100		
223	62.8	49.5	74.3
227	64.3	50.9	76.3





#35
Caprolactam
Concen: 49.951 ng m
RT: 11.446 min Scan# 14
Delta R.T. 0.000 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:113 Resp: 19624

Ion	Ratio	Lower	Upper
113	100		
55	133.2	115.9	155.9
56	107.8	81.1	121.1

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#36
4-Chloro-3-methylphenol
Concen: 53.124 ng
RT: 11.775 min Scan# 1477
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:107 Resp: 658827

Ion	Ratio	Lower	Upper
107	100		
144	25.5	20.5	30.7
142	77.9	63.4	95.2





#37
2-Methylnaphthalene
Concen: 42.049 ng
RT: 12.134 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:142 Resp: 112522

Ion	Ratio	Lower	Upper
142	100		
141	88.8	70.3	105.5
115	37.5	26.9	40.3



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#38
1-Methylnaphthalene
Concen: 45.487 ng
RT: 12.357 min Scan# 1576
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:142 Resp: 1190789

Ion	Ratio	Lower	Upper
142	100		
141	93.5	73.6	110.4
116	4.1	3.0	4.6





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.328 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:164 Resp: 49769

Ion	Ratio	Lower	Upper
164	100		
162	103.1	80.6	120.8
160	45.9	35.3	52.9



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 46.835 ng
RT: 12.510 min Scan# 1602
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:216 Resp: 641022

Ion	Ratio	Lower	Upper
216	100		
214	78.6	62.6	93.8
179	22.0	17.2	25.8
108	21.9	15.0	22.6





#41
Hexachlorocyclopentadiene
Concen: 183.797 ng
RT: 12.487 min Scan# 1
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:237 Resp: 88969

Ion	Ratio	Lower	Upper
237	100		
235	61.9	43.9	83.9
272	14.3	0.0	33.6

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#42
2,4,6-Tribromophenol
Concen: 175.532 ng
RT: 15.851 min Scan# 2170
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:330 Resp: 1208690

Ion	Ratio	Lower	Upper
330	100		
332	96.9	77.1	115.7
141	27.8	24.7	37.1





#43
2,4,6-Trichlorophenol
Concen: 52.293 ng
RT: 12.757 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:196 Resp: 47584

Ion	Ratio	Lower	Upper
196	100		
198	96.2	74.3	111.5
200	31.4	24.0	36.0



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#44
2,4,5-Trichlorophenol
Concen: 52.059 ng
RT: 12.840 min Scan# 1658
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:196 Resp: 524586

Ion	Ratio	Lower	Upper
196	100		
198	98.2	77.3	115.9
97	52.6	40.4	60.6
132	29.4	23.4	35.0





#45
2-Fluorobiphenyl
Concen: 91.104 ng
RT: 12.946 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:172 Resp: 296139

Ion	Ratio	Lower	Upper
172	100		
171	36.1	28.2	42.2
170	23.6	18.6	27.8



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#46
1,1'-Biphenyl
Concen: 44.632 ng
RT: 13.157 min Scan# 1712
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:154 Resp: 1632780

Ion	Ratio	Lower	Upper
154	100		
153	40.7	20.3	60.3
76	10.8	0.0	33.5





#47
2-Chloronaphthalene
Concen: 45.164 ng
RT: 13.199 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:162 Resp: 123485

Ion	Ratio	Lower	Upper
162	100		
127	41.6	31.9	47.9
164	32.6	26.2	39.4

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#48
2-Nitroaniline
Concen: 52.508 ng
RT: 13.410 min Scan# 1755
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 65 Resp: 402707

Ion	Ratio	Lower	Upper
65	100		
92	71.6	62.3	93.5
138	115.0	97.2	145.8





#49
Acenaphthylene
Concen: 49.549 ng
RT: 14.051 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:152 Resp: 213297

Ion	Ratio	Lower	Upper
152	100		
151	20.3	15.9	23.9
153	13.3	10.6	15.8



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#50
Dimethylphthalate
Concen: 48.425 ng
RT: 13.793 min Scan# 1820
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:163 Resp: 1752656

Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	10.0	8.0	12.0





#51
2,6-Dinitrotoluene
Concen: 49.835 ng
RT: 13.910 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:165 Resp: 38301

Ion	Ratio	Lower	Upper
165	100		
63	54.3	38.0	57.0
89	57.4	41.7	62.5



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#52
Acenaphthene
Concen: 44.922 ng
RT: 14.398 min Scan# 1923
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:154 Resp: 1225957

Ion	Ratio	Lower	Upper
154	100		
153	114.0	89.8	134.6
152	54.0	42.6	63.8





#53
3-Nitroaniline
Concen: 39.430 ng
RT: 14.251 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:138 Resp: 31849

Ion	Ratio	Lower	Upper
138	100		
108	11.2	8.3	12.5
92	113.7	89.8	134.8



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#54
2,4-Dinitrophenol
Concen: 109.007 ng
RT: 14.463 min Scan# 1934
Delta R.T. -0.006 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:184 Resp: 493492

Ion	Ratio	Lower	Upper
184	100		
63	59.9	42.6	64.0
154	65.7	47.2	70.8





#55
Dibenzofuran
Concen: 44.132 ng
RT: 14.740 min Scan# 1
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:168 Resp: 196868

Ion	Ratio	Lower	Upper
168	100		
139	40.8	30.3	45.5
169	13.0	10.6	16.0



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#56
4-Nitrophenol
Concen: 96.520 ng
RT: 14.593 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:139 Resp: 673353

Ion	Ratio	Lower	Upper
139	100		
109	79.8	46.1	86.1
65	89.1	65.0	105.0





#57
2,4-Dinitrotoluene
Concen: 54.357 ng
RT: 14.710 min Scan# 1
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:165 Resp: 56988

Ion	Ratio	Lower	Upper
165	100		
63	40.7	30.4	45.6
89	68.1	54.6	82.0
182	3.1	2.2	3.2



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#58
Fluorene
Concen: 46.967 ng
RT: 15.393 min Scan# 2092
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:166 Resp: 1621926

Ion	Ratio	Lower	Upper
166	100		
165	98.1	78.1	117.1
167	13.7	10.9	16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 50.663 ng
RT: 14.975 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:232 Resp: 47085

Ion	Ratio	Lower	Upper
232	100		
131	47.3	36.8	55.2
130	2.7	1.9	2.9
166	31.1	25.0	37.6



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#60
Diethylphthalate
Concen: 47.466 ng
RT: 15.175 min Scan# 2055
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:149 Resp: 1770392

Ion	Ratio	Lower	Upper
149	100		
177	21.7	17.4	26.2
150	12.6	10.1	15.1





#61
4-Chlorophenyl-phenylether
Concen: 48.482 ng
RT: 15.393 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:204 Resp: 81301

Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.5	39.7
141	64.3	49.4	74.0

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#62
4-Nitroaniline
Concen: 47.924 ng
RT: 15.434 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:138 Resp: 409802

Ion	Ratio	Lower	Upper
138	100		
92	50.9	27.8	67.8
108	86.9	48.9	88.9





#63
Azobenzene
Concen: 43.531 ng
RT: 15.693 min Scan# 21
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion: 77 Resp: 149227

Ion	Ratio	Lower	Upper
77	100		
182	29.8	7.3	47.3
105	22.6	2.3	42.3
51	29.2	3.4	43.4



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.134 min Scan# 2388
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:188 Resp: 979666

Ion	Ratio	Lower	Upper
188	100		
94	8.5	8.6	13.0#
80	9.6	9.8	14.6#





#65
4,6-Dinitro-2-methylphenol
Concen: 51.974 ng
RT: 15.493 min Scan# 2
Delta R.T. -0.006 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:198 Resp: 32101

Ion	Ratio	Lower	Upper
198	100		
51	37.0	18.4	58.4
105	41.0	26.5	66.5

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#66
n-Nitrosodiphenylamine
Concen: 48.614 ng
RT: 15.616 min Scan# 2130
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:169 Resp: 1421507

Ion	Ratio	Lower	Upper
169	100		
168	67.2	53.0	79.4
167	36.5	28.9	43.3





#67
4-Bromophenyl-phenylether
Concen: 53.238 ng
RT: 16.304 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:248 Resp: 57041

Ion	Ratio	Lower	Upper
248	100		
250	99.2	78.2	117.4
141	75.3	57.7	86.5



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#68
Hexachlorobenzene
Concen: 54.940 ng
RT: 16.428 min Scan# 2268
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:284 Resp: 699484

Ion	Ratio	Lower	Upper
284	100		
142	30.0	24.6	37.0
249	29.4	23.8	35.6





#69
Atrazine
Concen: 73.922 ng
RT: 16.592 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:200 Resp: 55052

Ion	Ratio	Lower	Upper
200	100		
173	25.8	7.5	47.5
215	46.8	28.8	68.8



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#70
Pentachlorophenol
Concen: 106.541 ng
RT: 16.787 min Scan# 2329
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:266 Resp: 946297

Ion	Ratio	Lower	Upper
266	100		
268	64.0	51.0	76.6
264	61.8	49.8	74.6





#71
Phenanthrene
Concen: 46.760 ng
RT: 17.175 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:178 Resp: 249902

Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.4	23.0
179	15.5	12.5	18.7



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#72
Anthracene
Concen: 50.379 ng
RT: 17.269 min Scan# 2411
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:178 Resp: 2594917

Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.6	12.4	18.6





#73
Carbazole
Concen: 48.263 ng
RT: 17.557 min Scan# 24
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:167 Resp: 242899

Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.1	25.7
139	13.7	10.6	16.0



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#74
Di-n-butylphthalate
Concen: 50.645 ng
RT: 18.139 min Scan# 2559
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:149 Resp: 3175548

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.2
104	5.3	3.9	5.9





#75
Fluoranthene
Concen: 47.187 ng
RT: 19.263 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:202 Resp: 299931

Ion	Ratio	Lower	Upper
202	100		
101	9.9	0.0	33.2
203	17.5	0.0	37.7



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.575 min Scan# 3143
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:240 Resp: 1019336

Ion	Ratio	Lower	Upper
240	100		
120	9.0	9.8	14.8#
236	26.2	20.9	31.3





#77
Benzidine
Concen: 111.012 ng
RT: 19.463 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:184 Resp: 143438

Ion	Ratio	Lower	Upper
184	100		
185	14.0	11.6	17.4
183	12.2	9.2	13.8



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#78
Pyrene
Concen: 47.269 ng
RT: 19.633 min Scan# 2813
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:202 Resp: 3062088

Ion	Ratio	Lower	Upper
202	100		
200	21.2	17.2	25.8
203	17.6	14.4	21.6





#79
Terphenyl-d14
Concen: 102.330 ng
RT: 19.857 min Scan# 2
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:244 Resp: 517497

Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.6	8.4
122	9.5	9.4	14.0

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#80
Butylbenzylphthalate
Concen: 50.767 ng
RT: 20.586 min Scan# 2975
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:149 Resp: 1408621

Ion	Ratio	Lower	Upper
149	100		
91	67.2	56.4	84.6
206	21.5	16.3	24.5





#81
Benzo(a)anthracene
Concen: 48.302 ng
RT: 21.545 min Scan# 3
Delta R.T. -0.023 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:228 Resp: 306029

Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.8	32.8
229	20.0	15.9	23.9



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#82
3,3'-Dichlorobenzidine
Concen: 33.877 ng
RT: 21.480 min Scan# 3127
Delta R.T. -0.012 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:252 Resp: 753351

Ion	Ratio	Lower	Upper
252	100		
254	65.1	51.7	77.5
126	9.4	8.8	13.2





#83
Chrysene
Concen: 46.325 ng
RT: 21.622 min Scan# 3
Delta R.T. -0.017 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:228 Resp: 281189

Ion	Ratio	Lower	Upper
228	100		
226	30.0	24.2	36.2
229	19.6	15.6	23.4



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 49.150 ng
RT: 21.480 min Scan# 3127
Delta R.T. -0.029 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:149 Resp: 1999200

Ion	Ratio	Lower	Upper
149	100		
167	27.9	22.4	33.6
279	5.4	4.0	6.0





#85
Di-n-octyl phthalate
Concen: 48.332 ng
RT: 22.751 min Scan# 3
Delta R.T. -0.035 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:149 Resp: 319606

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.2	1.8
43	7.7	6.0	9.0

Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.898 min Scan# 3708
Delta R.T. -0.035 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:264 Resp: 1024841

Ion	Ratio	Lower	Upper
264	100		
260	23.3	18.9	28.3
265	21.6	17.8	26.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 46.243 ng
RT: 28.674 min Scan# 4
Delta R.T. -0.065 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:276 Resp: 329016

Ion	Ratio	Lower	Upper
276	100		
138	14.3	14.5	21.7#
277	25.4	20.2	30.2



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#88
Benzo(b)fluoranthene
Concen: 47.514 ng
RT: 23.845 min Scan# 3529
Delta R.T. -0.041 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:252 Resp: 2918757

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	8.9	9.4	14.0#





#89
Benzo(k)fluoranthene
Concen: 49.400 ng
RT: 23.921 min Scan# 3
Delta R.T. -0.035 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:252 Resp: 293127

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.7	26.5
125	8.4	9.1	13.7#



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#90
Benzo(a)pyrene
Concen: 51.909 ng
RT: 24.751 min Scan# 3683
Delta R.T. -0.029 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:252 Resp: 2750615

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.6	26.4
125	9.8	10.1	15.1#





#91
Dibenzo(a,h)anthracene
Concen: 45.635 ng
RT: 28.751 min Scan# 4
Delta R.T. -0.053 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:278 Resp: 269498

Ion	Ratio	Lower	Upper
278	100		
139	12.8	12.6	19.0
279	23.4	18.5	27.7



Instrument :
BNA_P
ClientSampleId :
PB167904BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#92
Benzo(g,h,i)perylene
Concen: 43.241 ng
RT: 29.845 min Scan# 4549
Delta R.T. -0.053 min
Lab File: BP024570.D
Acq: 08 May 2025 16:07

Tgt Ion:276 Resp: 2599244

Ion	Ratio	Lower	Upper
276	100		
277	24.3	18.9	28.3
138	18.3	18.7	28.1#



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	05/06/25
Project:	NJ Soil PT	Date Received:	05/06/25
Client Sample ID:	MH-LLMS	SDG No.:	Q1872
Lab Sample ID:	Q1964-01MS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	80.8
Sample Wt/Vol:	50.06 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group4
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050146.D	1	05/08/25 09:15	05/08/25 14:42	PB167904

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1912-24-9	Atrazine	1100		25.2	130	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	62.7		18 - 107	63%	SPK: 100
321-60-8	2-Fluorobiphenyl	55.9		20 - 109	56%	SPK: 100
1718-51-0	Terphenyl-d14	51.3		10 - 105	51%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	264000	7.74			
1146-65-2	Naphthalene-d8	907000	10.534			
15067-26-2	Acenaphthene-d10	563000	14.386			
1517-22-2	Phenanthrene-d10	1040000	17.139			
1719-03-5	Chrysene-d12	1000000	21.386			
1520-96-3	Perylene-d12	1080000	24.38			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050146.D
Acq On : 08 May 2025 14:42
Operator : RC/JU
Sample : Q1964-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Quant Time: May 08 15:42:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.740	152	264341	20.000	ng		0.00
21) Naphthalene-d8	10.534	136	906955	20.000	ng		0.00
39) Acenaphthene-d10	14.386	164	562871	20.000	ng		0.00
64) Phenanthrene-d10	17.139	188	1043026	20.000	ng		0.00
76) Chrysene-d12	21.386	240	1002002	20.000	ng		0.00
86) Perylene-d12	24.380	264	1081327	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.334	112	1593314	105.546	ng		0.00
7) Phenol-d6	6.934	99	1968415	103.745	ng		0.00
23) Nitrobenzene-d5	8.898	82	1153669	62.681	ng		0.00
42) 2,4,6-Tribromophenol	15.886	330	838438	100.740	ng		0.00
45) 2-Fluorobiphenyl	13.004	172	2660755	55.921	ng		0.00
79) Terphenyl-d14	19.762	244	3471536	51.267	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.228	88	246843	38.136	ng		98
3) Pyridine	3.634	79	669714	40.270	ng		98
4) n-Nitrosodimethylamine	3.540	42	276875	42.449	ng		97
6) Aniline	7.075	93	527166	22.003	ng		98
8) 2-Chlorophenol	7.316	128	716311	43.886	ng		99
9) Benzaldehyde	6.887	77	299377	23.578	ng		99
10) Phenol	6.957	94	847839	44.140	ng		99
11) bis(2-Chloroethyl)ether	7.163	93	667998	41.649	ng		99
12) 1,3-Dichlorobenzene	7.628	146	821925	41.687	ng		99
13) 1,4-Dichlorobenzene	7.775	146	828063	41.805	ng		99
14) 1,2-Dichlorobenzene	8.092	146	796060	41.606	ng		99
15) Benzyl Alcohol	7.992	79	566884	43.082	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.263	45	812019	41.650	ng		99
17) 2-Methylphenol	8.204	107	540586	42.722	ng		99
18) Hexachloroethane	8.810	117	294157	41.777	ng		97
19) n-Nitroso-di-n-propyla...	8.551	70	490922	40.342	ng		99
20) 3+4-Methylphenols	8.534	107	729232	42.675	ng		96
22) Acetophenone	8.563	105	967046	42.888	ng	#	98
24) Nitrobenzene	8.945	77	699191	43.934	ng		97
25) Isophorone	9.469	82	1301869	41.495	ng		99
26) 2-Nitrophenol	9.657	139	373396	45.926	ng		98
27) 2,4-Dimethylphenol	9.728	122	603294	44.745	ng		99
28) bis(2-Chloroethoxy)met...	9.945	93	825737	42.582	ng		100
29) 2,4-Dichlorophenol	10.210	162	681099	45.091	ng		98
30) 1,2,4-Trichlorobenzene	10.398	180	795869	43.246	ng		100
31) Naphthalene	10.581	128	1966145	42.807	ng		100
32) Benzoic acid	9.910	122	407313	47.234	ng		99
33) 4-Chloroaniline	10.716	127	276078	14.665	ng		99
34) Hexachlorobutadiene	10.863	225	507937	43.481	ng		100
35) Caprolactam	11.510	113	175814	39.920	ng		95
36) 4-Chloro-3-methylphenol	11.857	107	588875	42.198	ng		99
37) 2-Methylnaphthalene	12.198	142	1278304	41.511	ng		99
38) 1-Methylnaphthalene	12.422	142	1344615	41.260	ng		100
40) 1,2,4,5-Tetrachloroben...	12.575	216	885191	45.429	ng		99
41) Hexachlorocyclopentadiene	12.551	237	930801	78.168	ng		99
43) 2,4,6-Trichlorophenol	12.827	196	567896	45.951	ng		98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050146.D
Acq On : 08 May 2025 14:42
Operator : RC/JU
Sample : Q1964-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Quant Time: May 08 15:42:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.922	196	612924	45.991	ng	99
46) 1,1'-Biphenyl	13.216	154	1847009	44.144	ng	99
47) 2-Chloronaphthalene	13.263	162	1461752	44.111	ng	100
48) 2-Nitroaniline	13.480	65	352501	45.323	ng	100
49) Acenaphthylene	14.110	152	2248257	43.030	ng	100
50) Dimethylphthalate	13.851	163	1731408	42.091	ng	100
51) 2,6-Dinitrotoluene	13.974	165	373814	43.891	ng	98
52) Acenaphthene	14.457	154	1304787	43.143	ng	99
53) 3-Nitroaniline	14.316	138	209802	25.802	ng	98
54) 2,4-Dinitrophenol	14.533	184	460854	84.226	ng	98
55) Dibenzofuran	14.792	168	2124279	42.217	ng	100
56) 4-Nitrophenol	14.663	139	549471	82.890	ng	98
57) 2,4-Dinitrotoluene	14.774	165	515870	43.835	ng	97
58) Fluorene	15.439	166	1747395	42.426	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.033	232	486575	42.179	ng	100
60) Diethylphthalate	15.216	149	1573625	40.599	ng	99
61) 4-Chlorophenyl-phenyle...	15.433	204	955055	42.217	ng	99
62) 4-Nitroaniline	15.480	138	299821	38.218	ng	97
63) Azobenzene	15.727	77	1379896	42.479	ng	99
65) 4,6-Dinitro-2-methylph...	15.545	198	289940	44.408	ng	99
66) n-Nitrosodiphenylamine	15.651	169	1432677	44.731	ng	100
67) 4-Bromophenyl-phenylether	16.327	248	560826	44.464	ng	98
68) Hexachlorobenzene	16.451	284	636602	43.785	ng	97
69) Atrazine	16.610	200	512133	44.329	ng	99
70) Pentachlorophenol	16.804	266	806258	98.517	ng	98
71) Phenanthrene	17.180	178	2580538	43.866	ng	100
72) Anthracene	17.274	178	2607919	43.806	ng	100
73) Carbazole	17.551	167	2283295	44.199	ng	100
74) Di-n-butylphthalate	18.104	149	2646232	43.014	ng	100
75) Fluoranthene	19.204	202	2966465	42.950	ng	100
77) Benzidine	19.392	184	1135378	31.900	ng	99
78) Pyrene	19.568	202	3052252	43.378	ng	100
80) Butylbenzylphthalate	20.456	149	1127542	42.785	ng	94
81) Benzo(a)anthracene	21.368	228	3058099	44.320	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	682545	25.840	ng	99
83) Chrysene	21.427	228	2786599	43.635	ng	99
84) Bis(2-ethylhexyl)phtha...	21.280	149	1626197	41.312	ng	100
85) Di-n-octyl phthalate	22.403	149	2786511	42.924	ng	100
87) Indeno(1,2,3-cd)pyrene	27.768	276	3829687	47.531	ng	100
88) Benzo(b)fluoranthene	23.433	252	3038089	43.203	ng	100
89) Benzo(k)fluoranthene	23.491	252	3044885	42.806	ng	100
90) Benzo(a)pyrene	24.238	252	2922477	44.371	ng	99
91) Dibenzo(a,h)anthracene	27.809	278	3123024	47.549	ng	99
92) Benzo(g,h,i)perylene	28.821	276	2974655	47.307	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050146.D
Acq On : 08 May 2025 14:42
Operator : RC/JU
Sample : Q1964-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Quant Time: May 08 15:42:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050146.D
Acq On : 08 May 2025 14:42
Operator : RC/JU
Sample : Q1964-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Quant Time: May 08 15:42:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.740 min Scan# 80
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:152 Resp: 264341

Ion	Ratio	Lower	Upper
152	100		
150	153.0	124.1	186.1
115	52.7	41.2	61.8





#2
1,4-Dioxane
Concen: 38.136 ng
RT: 3.228 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 88 Resp: 246843

Ion	Ratio	Lower	Upper
88	100		
58	69.6	55.1	82.7
43	29.2	20.9	31.3





#3
Pyridine
Concen: 40.270 ng
RT: 3.634 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 79 Resp: 669714

Ion	Ratio	Lower	Upper
79	100		
52	62.9	48.6	72.8
51	31.2	24.6	37.0





#4
n-Nitrosodimethylamine
Concen: 42.449 ng
RT: 3.540 min Scan# 94
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 42 Resp: 276875

Ion	Ratio	Lower	Upper
42	100		
74	147.0	120.2	180.2
44	7.1	6.2	9.2





#5
2-Fluorophenol
Concen: 105.546 ng
RT: 5.334 min Scan# 3
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:112 Resp: 1593314

Ion	Ratio	Lower	Upper
112	100		
64	54.3	42.6	64.0
63	28.2	22.2	33.4





#6
Aniline
Concen: 22.003 ng
RT: 7.075 min Scan# 695
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 93 Resp: 527166

Ion	Ratio	Lower	Upper
93	100		
66	36.5	29.9	44.9
65	19.4	14.5	21.7





#7
Phenol-d6
Concen: 103.745 ng
RT: 6.934 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 99 Resp: 1968415

Ion	Ratio	Lower	Upper
99	100		
42	15.6	13.0	19.4
71	32.1	25.6	38.4





#8
2-Chlorophenol
Concen: 43.886 ng
RT: 7.316 min Scan# 736
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:128 Resp: 716311

Ion	Ratio	Lower	Upper
128	100		
130	32.4	12.8	52.8
64	42.5	21.2	61.2





#9
Benzaldehyde
Concen: 23.578 ng
RT: 6.887 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 77 Resp: 299377

Ion	Ratio	Lower	Upper
77	100		
105	99.6	80.5	120.5
106	95.7	76.1	116.1





#10
Phenol
Concen: 44.140 ng
RT: 6.957 min Scan# 675
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 94 Resp: 847839

Ion	Ratio	Lower	Upper
94	100		
65	27.4	7.6	47.6
66	35.4	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 41.649 ng
RT: 7.163 min Scan# 7
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 93 Resp: 667998

Ion	Ratio	Lower	Upper
93	100		
63	72.6	52.3	92.3
95	32.0	12.8	52.8





#12
1,3-Dichlorobenzene
Concen: 41.687 ng
RT: 7.628 min Scan# 789
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:146 Resp: 821925

Ion	Ratio	Lower	Upper
146	100		
148	64.2	52.5	78.7
75	28.0	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 41.805 ng
RT: 7.775 min Scan# 81
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:146 Resp: 828063

Ion	Ratio	Lower	Upper
146	100		
148	66.1	52.2	78.2
111	36.8	29.3	43.9





#14
1,2-Dichlorobenzene
Concen: 41.606 ng
RT: 8.092 min Scan# 868
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:146 Resp: 796060

Ion	Ratio	Lower	Upper
146	100		
148	65.1	52.4	78.6
111	39.1	31.0	46.6





#15
Benzyl Alcohol
Concen: 43.082 ng
RT: 7.992 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 79 Resp: 566884

Ion	Ratio	Lower	Upper
79	100		
108	80.8	66.0	99.0
77	64.3	51.6	77.4





#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.650 ng
RT: 8.263 min Scan# 897
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 45 Resp: 812019

Ion	Ratio	Lower	Upper
45	100		
77	17.0	0.0	36.4
79	12.7	0.0	32.7





#17
2-Methylphenol
Concen: 42.722 ng
RT: 8.204 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:107 Resp: 540586

Ion	Ratio	Lower	Upper
107	100		
108	114.3	91.1	136.7
77	47.0	36.2	54.4
79	44.3	35.8	53.8





#18
Hexachloroethane
Concen: 41.777 ng
RT: 8.810 min Scan# 990
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:117 Resp: 294157

Ion	Ratio	Lower	Upper
117	100		
119	96.9	77.3	115.9
201	123.4	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 40.342 ng
RT: 8.551 min Scan# 94
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 70 Resp: 490922

Ion	Ratio	Lower	Upper
70	100		
42	49.9	40.5	60.7
101	9.5	7.8	11.6
130	22.6	17.4	26.2





#20
3+4-Methylphenols
Concen: 42.675 ng
RT: 8.534 min Scan# 943
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:107 Resp: 729232

Ion	Ratio	Lower	Upper
107	100		
108	88.4	72.4	112.4
77	33.8	13.6	53.6
79	26.5	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.534 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:136 Resp: 906955

Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.9	13.3
54	6.6	5.1	7.7
68	5.5	4.4	6.6





#22
Acetophenone
Concen: 42.888 ng
RT: 8.563 min Scan# 948
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:105 Resp: 967046

Ion	Ratio	Lower	Upper
105	100		
71	3.5	1.7	2.5#
51	25.3	18.6	27.8
120	23.0	18.3	27.5





#23
Nitrobenzene-d5
Concen: 62.681 ng
RT: 8.898 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 82 Resp: 1153669

Ion	Ratio	Lower	Upper
82	100		
128	43.8	35.7	53.5
54	45.8	35.4	53.2





#24
Nitrobenzene
Concen: 43.934 ng
RT: 8.945 min Scan# 1013
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 77 Resp: 699191

Ion	Ratio	Lower	Upper
77	100		
123	49.8	37.9	56.9
65	13.0	10.1	15.1





#25
Isophorone
Concen: 41.495 ng
RT: 9.469 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 82 Resp: 1301869

Ion	Ratio	Lower	Upper
82	100		
95	6.8	5.7	8.5
138	19.0	15.5	23.3





#26
2-Nitrophenol
Concen: 45.926 ng
RT: 9.657 min Scan# 1134
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:139 Resp: 373396

Ion	Ratio	Lower	Upper
139	100		
109	27.7	21.9	32.9
65	45.9	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 44.745 ng
RT: 9.728 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:122 Resp: 603294

Ion	Ratio	Lower	Upper
122	100		
107	112.1	89.0	133.6
121	56.7	45.6	68.4





#28
bis(2-Chloroethoxy)methane
Concen: 42.582 ng
RT: 9.945 min Scan# 1183
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 93 Resp: 825737

Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.1	39.1
123	11.9	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 45.091 ng
RT: 10.210 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:162 Resp: 681099

Ion	Ratio	Lower	Upper
162	100		
164	64.8	45.4	85.4
98	29.6	11.3	51.3





#30
1,2,4-Trichlorobenzene
Concen: 43.246 ng
RT: 10.398 min Scan# 1260
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:180 Resp: 795869

Ion	Ratio	Lower	Upper
180	100		
182	96.2	77.0	115.4
145	25.8	20.6	30.8





#31
Naphthalene
Concen: 42.807 ng
RT: 10.581 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:128 Resp: 1966145

Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	12.7	10.4	15.6





#32
Benzoic acid
Concen: 47.234 ng
RT: 9.910 min Scan# 1177
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:122 Resp: 407313

Ion	Ratio	Lower	Upper
122	100		
105	122.5	103.6	143.6
77	92.3	73.7	113.7





#33
4-Chloroaniline
Concen: 14.665 ng
RT: 10.716 min Scan# 1
Delta R.T. 0.012 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:127 Resp: 276078

Ion	Ratio	Lower	Upper
127	100		
129	33.6	26.3	39.5
65	28.6	22.5	33.7
92	17.7	14.5	21.7





#34
Hexachlorobutadiene
Concen: 43.481 ng
RT: 10.863 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:225 Resp: 507937

Ion	Ratio	Lower	Upper
225	100		
223	62.9	50.2	75.2
227	63.8	50.6	76.0





#35
Caprolactam
Concen: 39.920 ng
RT: 11.510 min Scan# 14
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:113 Resp: 175814

Ion	Ratio	Lower	Upper
113	100		
55	148.1	122.9	162.9
56	113.9	88.1	128.1





#36
4-Chloro-3-methylphenol
Concen: 42.198 ng
RT: 11.857 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:107 Resp: 588875

Ion	Ratio	Lower	Upper
107	100		
144	29.7	23.8	35.6
142	91.4	71.9	107.9





#37
2-Methylnaphthalene
Concen: 41.511 ng
RT: 12.198 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:142 Resp: 1278304

Ion	Ratio	Lower	Upper
142	100		
141	87.3	70.4	105.6
115	31.6	25.4	38.2





#38
1-Methylnaphthalene
Concen: 41.260 ng
RT: 12.422 min Scan# 1604
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:142 Resp: 1344615

Ion	Ratio	Lower	Upper
142	100		
141	90.6	72.1	108.1
116	3.5	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.386 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:164 Resp: 562871

Ion	Ratio	Lower	Upper
164	100		
162	97.8	79.4	119.0
160	44.3	36.2	54.4





#40
1,2,4,5-Tetrachlorobenzene
Concen: 45.429 ng
RT: 12.575 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:216 Resp: 885191

Ion	Ratio	Lower	Upper
216	100		
214	78.4	62.3	93.5
179	19.2	15.4	23.0
108	15.6	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 78.168 ng
RT: 12.551 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:237 Resp: 930801

Ion	Ratio	Lower	Upper
237	100		
235	63.8	42.9	82.9
272	13.3	0.0	32.9





#42
2,4,6-Tribromophenol
Concen: 100.740 ng
RT: 15.886 min Scan# 2193
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:330 Resp: 838438

Ion	Ratio	Lower	Upper
330	100		
332	96.3	77.3	115.9
141	27.3	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 45.951 ng
RT: 12.827 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:196 Resp: 567896

Ion	Ratio	Lower	Upper
196	100		
198	97.1	79.0	118.6
200	30.9	25.1	37.7





#44
2,4,5-Trichlorophenol
Concen: 45.991 ng
RT: 12.922 min Scan# 1689
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:196 Resp: 612924

Ion	Ratio	Lower	Upper
196	100		
198	96.8	76.7	115.1
97	38.1	30.1	45.1
132	21.3	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 55.921 ng
RT: 13.004 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:172 Resp: 2660755

Ion	Ratio	Lower	Upper
172	100		
171	34.5	27.8	41.6
170	23.3	18.6	28.0





#46
1,1'-Biphenyl
Concen: 44.144 ng
RT: 13.216 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:154 Resp: 1847009

Ion	Ratio	Lower	Upper
154	100		
153	39.3	20.0	60.0
76	10.9	0.0	30.3





#47
2-Chloronaphthalene
Concen: 44.111 ng
RT: 13.263 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:162 Resp: 1461752

Ion	Ratio	Lower	Upper
162	100		
127	32.5	25.9	38.9
164	33.1	26.6	40.0





#48
2-Nitroaniline
Concen: 45.323 ng
RT: 13.480 min Scan# 1784
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion: 65 Resp: 352501

Ion	Ratio	Lower	Upper
65	100		
92	71.7	57.1	85.7
138	121.3	97.0	145.4





#49
Acenaphthylene
Concen: 43.030 ng
RT: 14.110 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:152 Resp: 2248257

Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.8	23.8
153	13.2	10.6	15.8





#50
Dimethylphthalate
Concen: 42.091 ng
RT: 13.851 min Scan# 1847
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:163 Resp: 1731408

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.2
164	10.1	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 43.891 ng
RT: 13.974 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:165 Resp: 373814

Ion	Ratio	Lower	Upper
165	100		
63	43.3	35.1	52.7
89	45.7	38.4	57.6





#52
Acenaphthene
Concen: 43.143 ng
RT: 14.457 min Scan# 1950
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:154 Resp: 1304787

Ion	Ratio	Lower	Upper
154	100		
153	113.4	89.0	133.4
152	53.7	43.0	64.6





#53
3-Nitroaniline
Concen: 25.802 ng
RT: 14.316 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:138 Resp: 209802

Ion	Ratio	Lower	Upper
138	100		
108	9.5	7.4	11.2
92	102.8	84.0	126.0





#54
2,4-Dinitrophenol
Concen: 84.226 ng
RT: 14.533 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:184 Resp: 460854

Ion	Ratio	Lower	Upper
184	100		
63	48.6	39.8	59.6
154	56.4	46.1	69.1





#55
Dibenzofuran
Concen: 42.217 ng
RT: 14.792 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:168 Resp: 2124279

Ion	Ratio	Lower	Upper
168	100		
139	32.9	26.1	39.1
169	13.5	10.8	16.2





#56
4-Nitrophenol
Concen: 82.890 ng
RT: 14.663 min Scan# 1985
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:139 Resp: 549471

Ion	Ratio	Lower	Upper
139	100		
109	68.4	47.0	87.0
65	85.9	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 43.835 ng
RT: 14.774 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:165 Resp: 515870

Ion	Ratio	Lower	Upper
165	100		
63	40.0	29.0	43.6
89	62.2	49.4	74.2
182	2.9	2.2	3.2





#58
Fluorene
Concen: 42.426 ng
RT: 15.439 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:166 Resp: 1747395

Ion	Ratio	Lower	Upper
166	100		
165	96.4	76.6	115.0
167	13.7	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 42.179 ng
RT: 15.033 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:232 Resp: 486575

Ion	Ratio	Lower	Upper
232	100		
131	35.5	28.2	42.4
130	2.1	1.5	2.3
166	28.3	22.6	34.0





#60
Diethylphthalate
Concen: 40.599 ng
RT: 15.216 min Scan# 2079
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:149 Resp: 1573625

Ion	Ratio	Lower	Upper
149	100		
177	24.3	19.1	28.7
150	12.4	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 42.217 ng
RT: 15.433 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:204 Resp: 955055

Ion	Ratio	Lower	Upper
204	100		
206	33.0	26.3	39.5
141	43.5	34.0	51.0





#62
4-Nitroaniline
Concen: 38.218 ng
RT: 15.480 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:138 Resp: 299821

Ion	Ratio	Lower	Upper
138	100		
92	47.5	25.9	65.9
108	75.5	52.4	92.4





#63
Azobenzene
Concen: 42.479 ng
RT: 15.727 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion: 77 Resp: 1379896

Ion	Ratio	Lower	Upper
77	100		
182	31.5	10.5	50.5
105	21.6	1.1	41.1
51	26.0	6.3	46.3





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2406
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:188 Resp: 1043026

Ion	Ratio	Lower	Upper
188	100		
94	7.3	5.7	8.5
80	7.7	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 44.408 ng
RT: 15.545 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:198 Resp: 289940

Ion	Ratio	Lower	Upper
198	100		
51	31.1	10.4	50.4
105	34.2	14.6	54.6





#66
n-Nitrosodiphenylamine
Concen: 44.731 ng
RT: 15.651 min Scan# 2153
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:169 Resp: 1432677

Ion	Ratio	Lower	Upper
169	100		
168	67.1	53.4	80.2
167	36.3	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 44.464 ng
RT: 16.327 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:248 Resp: 560826

Ion	Ratio	Lower	Upper
248	100		
250	96.8	76.2	114.2
141	55.5	45.9	68.9





#68
Hexachlorobenzene
Concen: 43.785 ng
RT: 16.451 min Scan# 2289
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:284 Resp: 636602

Ion	Ratio	Lower	Upper
284	100		
142	27.4	23.9	35.9
249	29.8	24.5	36.7





#69
Atrazine
Concen: 44.329 ng
RT: 16.610 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:200 Resp: 512133

Ion	Ratio	Lower	Upper
200	100		
173	24.2	4.3	44.3
215	51.6	30.6	70.6





#70
Pentachlorophenol
Concen: 98.517 ng
RT: 16.804 min Scan# 2349
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:266 Resp: 806258

Ion	Ratio	Lower	Upper
266	100		
268	65.2	51.2	76.8
264	63.9	49.6	74.4





#71
Phenanthrene
Concen: 43.866 ng
RT: 17.180 min Scan# 24
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:178 Resp: 2580538

Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.4	12.4	18.6





#72
Anthracene
Concen: 43.806 ng
RT: 17.274 min Scan# 2429
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:178 Resp: 2607919

Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.4	12.2	18.4





#73
Carbazole
Concen: 44.199 ng
RT: 17.551 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:167 Resp: 2283295

Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.2	25.8
139	11.4	9.0	13.6





#74
Di-n-butylphthalate
Concen: 43.014 ng
RT: 18.104 min Scan# 2570
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:149 Resp: 2646232

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.3	10.9
104	4.5	3.8	5.6





#75
Fluoranthene
Concen: 42.950 ng
RT: 19.204 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:202 Resp: 2966465

Ion	Ratio	Lower	Upper
202	100		
101	8.1	0.0	28.0
203	17.7	0.0	37.7





#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:240 Resp: 1002002

Ion	Ratio	Lower	Upper
240	100		
120	8.3	6.5	9.7
236	26.0	20.9	31.3





#77
Benzidine
Concen: 31.900 ng
RT: 19.392 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:184 Resp: 1135378

Ion	Ratio	Lower	Upper
184	100		
185	14.8	11.5	17.3
183	11.2	8.9	13.3





#78
Pyrene
Concen: 43.378 ng
RT: 19.568 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:202 Resp: 3052252

Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.7	25.1
203	17.5	14.1	21.1





#79
Terphenyl-d14
Concen: 51.267 ng
RT: 19.762 min Scan# 2
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:244 Resp: 3471536

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.6	8.4
122	7.5	5.6	8.4





#80
Butylbenzylphthalate
Concen: 42.785 ng
RT: 20.456 min Scan# 2970
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:149 Resp: 1127542

Ion	Ratio	Lower	Upper
149	100		
91	65.0	47.4	71.2
206	22.5	19.4	29.0





#81
Benzo(a)anthracene
Concen: 44.320 ng
RT: 21.368 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:228 Resp: 3058099

Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.1	33.1
229	19.7	15.7	23.5





#82
3,3'-Dichlorobenzidine
Concen: 25.840 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:252 Resp: 682545

Ion	Ratio	Lower	Upper
252	100		
254	64.3	51.8	77.6
126	8.7	6.0	9.0





#83
Chrysene
Concen: 43.635 ng
RT: 21.427 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:228 Resp: 2786599

Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.2	36.2
229	19.9	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 41.312 ng
RT: 21.280 min Scan# 3110
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:149 Resp: 1626197

Ion	Ratio	Lower	Upper
149	100		
167	30.3	24.1	36.1
279	5.2	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 42.924 ng
RT: 22.403 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:149 Resp: 2786511

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.4	5.8	8.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 24.380 min Scan# 3637
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:264 Resp: 1081327

Ion	Ratio	Lower	Upper
264	100		
260	23.9	18.8	28.2
265	22.0	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 47.531 ng
RT: 27.768 min Scan# 4
Delta R.T. 0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:276 Resp: 3829687

Ion	Ratio	Lower	Upper
276	100		
138	20.4	16.4	24.6
277	25.1	19.8	29.8





#88
Benzo(b)fluoranthene
Concen: 43.203 ng
RT: 23.433 min Scan# 3476
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:252 Resp: 3038089

Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.7	26.5
125	7.4	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 42.806 ng
RT: 23.491 min Scan# 34
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:252 Resp: 3044885

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.2
125	7.4	5.8	8.8





#90
Benzo(a)pyrene
Concen: 44.371 ng
RT: 24.238 min Scan# 3613
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:252 Resp: 2922477

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.4	26.0
125	8.3	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 47.549 ng
RT: 27.809 min Scan# 4
Delta R.T. -0.000 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Instrument :
BNA_M
ClientSampleId :
MH-LLMS

Tgt Ion:278 Resp: 3123024

Ion	Ratio	Lower	Upper
278	100		
139	12.8	10.5	15.7
279	24.0	19.4	29.0





#92
Benzo(g,h,i)perylene
Concen: 47.307 ng
RT: 28.821 min Scan# 4392
Delta R.T. -0.006 min
Lab File: BM050146.D
Acq: 08 May 2025 14:42

Tgt Ion:276 Resp: 2974655

Ion	Ratio	Lower	Upper
276	100		
277	24.5	19.1	28.7
138	17.8	14.3	21.5



Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	05/06/25
Project:	NJ Soil PT	Date Received:	05/06/25
Client Sample ID:	MH-LLMSD	SDG No.:	Q1872
Lab Sample ID:	Q1964-01MSD	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	80.8
Sample Wt/Vol:	50.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group4
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050147.D	1	05/08/25 09:15	05/08/25 15:21	PB167904

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1912-24-9	Atrazine	1100		25.2	130	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	60.5		18 - 107	61%	SPK: 100
321-60-8	2-Fluorobiphenyl	53.4		20 - 109	53%	SPK: 100
1718-51-0	Terphenyl-d14	47.9		10 - 105	48%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	253000	7.74			
1146-65-2	Naphthalene-d8	884000	10.533			
15067-26-2	Acenaphthene-d10	557000	14.386			
1517-22-2	Phenanthrene-d10	1070000	17.139			
1719-03-5	Chrysene-d12	1110000	21.386			
1520-96-3	Perylene-d12	1160000	24.38			

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements
J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050147.D
Acq On : 08 May 2025 15:21
Operator : RC/JU
Sample : Q1964-01MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 08 15:55:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) 1,4-Dichlorobenzene-d4	7.740	152	253439	20.000	ng		0.00
21) Naphthalene-d8	10.533	136	884473	20.000	ng		0.00
39) Acenaphthene-d10	14.386	164	556753	20.000	ng		0.00
64) Phenanthrene-d10	17.139	188	1069463	20.000	ng		0.00
76) Chrysene-d12	21.386	240	1112266	20.000	ng		0.00
86) Perylene-d12	24.380	264	1161580	20.000	ng		0.00
System Monitoring Compounds							
5) 2-Fluorophenol	5.334	112	1484585	102.574	ng		0.00
7) Phenol-d6	6.934	99	1847635	101.568	ng		0.00
23) Nitrobenzene-d5	8.904	82	1086064	60.508	ng		0.00
42) 2,4,6-Tribromophenol	15.886	330	821030	99.732	ng		0.00
45) 2-Fluorobiphenyl	13.004	172	2514924	53.437	ng		0.00
79) Terphenyl-d14	19.768	244	3600484	47.900	ng		0.00
Target Compounds							Qvalue
2) 1,4-Dioxane	3.228	88	230959	37.217	ng		97
3) Pyridine	3.634	79	631083	39.580	ng		99
4) n-Nitrosodimethylamine	3.540	42	258544	41.343	ng		99
6) Aniline	7.069	93	484984	21.113	ng		99
8) 2-Chlorophenol	7.316	128	666662	42.601	ng		99
9) Benzaldehyde	6.887	77	281630	23.135	ng		98
10) Phenol	6.957	94	797888	43.327	ng		98
11) bis(2-Chloroethyl)ether	7.163	93	625782	40.695	ng		99
12) 1,3-Dichlorobenzene	7.628	146	768371	40.647	ng		99
13) 1,4-Dichlorobenzene	7.775	146	773451	40.728	ng		99
14) 1,2-Dichlorobenzene	8.092	146	747198	40.732	ng		98
15) Benzyl Alcohol	7.992	79	532761	42.231	ng		99
16) 2,2'-oxybis(1-Chloropr...	8.263	45	752772	40.272	ng		99
17) 2-Methylphenol	8.204	107	511655	42.175	ng		99
18) Hexachloroethane	8.810	117	273774	40.554	ng		98
19) n-Nitroso-di-n-propyla...	8.551	70	460973	39.511	ng		100
20) 3+4-Methylphenols	8.534	107	682817	41.678	ng		97
22) Acetophenone	8.569	105	912791	41.510	ng	#	98
24) Nitrobenzene	8.945	77	654098	42.146	ng		98
25) Isophorone	9.469	82	1220307	39.884	ng		99
26) 2-Nitrophenol	9.657	139	347403	43.816	ng		99
27) 2,4-Dimethylphenol	9.728	122	560067	42.595	ng		99
28) bis(2-Chloroethoxy)met...	9.945	93	774120	40.935	ng		100
29) 2,4-Dichlorophenol	10.210	162	640382	43.473	ng		99
30) 1,2,4-Trichlorobenzene	10.398	180	750507	41.818	ng		100
31) Naphthalene	10.581	128	1846618	41.226	ng		99
32) Benzoic acid	9.904	122	406451m	48.332	ng		
33) 4-Chloroaniline	10.710	127	286757	15.619	ng		99
34) Hexachlorobutadiene	10.863	225	477099	41.880	ng		99
35) Caprolactam	11.504	113	169493	39.463	ng		100
36) 4-Chloro-3-methylphenol	11.857	107	560083	41.154	ng		99
37) 2-Methylnaphthalene	12.198	142	1213064	40.394	ng		100
38) 1-Methylnaphthalene	12.422	142	1269611	39.949	ng		100
40) 1,2,4,5-Tetrachloroben...	12.575	216	833595	43.251	ng		99
41) Hexachlorocyclopentadiene	12.551	237	810639	68.825	ng		100
43) 2,4,6-Trichlorophenol	12.827	196	544147	44.513	ng		96

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050147.D
Acq On : 08 May 2025 15:21
Operator : RC/JU
Sample : Q1964-01MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 08 15:55:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.922	196	585065	44.383	ng	99
46) 1,1'-Biphenyl	13.216	154	1748393	42.246	ng	99
47) 2-Chloronaphthalene	13.263	162	1394532	42.545	ng	100
48) 2-Nitroaniline	13.480	65	340250	44.229	ng	100
49) Acenaphthylene	14.110	152	2162339	41.840	ng	100
50) Dimethylphthalate	13.845	163	1655423	40.686	ng	100
51) 2,6-Dinitrotoluene	13.974	165	359914	42.724	ng	99
52) Acenaphthene	14.451	154	1253654	41.908	ng	99
53) 3-Nitroaniline	14.316	138	206422	25.665	ng	99
54) 2,4-Dinitrophenol	14.533	184	435912	80.794	ng	99
55) Dibenzofuran	14.792	168	2043574	41.059	ng	99
56) 4-Nitrophenol	14.663	139	548863	83.663	ng	98
57) 2,4-Dinitrotoluene	14.774	165	506950	43.550	ng	97
58) Fluorene	15.439	166	1699192	41.709	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.033	232	481053	42.159	ng	99
60) Diethylphthalate	15.216	149	1540274	40.175	ng	100
61) 4-Chlorophenyl-phenyle...	15.433	204	925658	41.368	ng	99
62) 4-Nitroaniline	15.480	138	299512	38.598	ng	97
63) Azobenzene	15.727	77	1340591	41.722	ng	99
65) 4,6-Dinitro-2-methylph...	15.545	198	279835	41.801	ng	98
66) n-Nitrosodiphenylamine	15.651	169	1403891	42.749	ng	99
67) 4-Bromophenyl-phenylether	16.327	248	548181	42.387	ng	97
68) Hexachlorobenzene	16.451	284	635267	42.614	ng	98
69) Atrazine	16.610	200	506484	42.756	ng	99
70) Pentachlorophenol	16.804	266	814874	97.109	ng	99
71) Phenanthrene	17.180	178	2579206	42.759	ng	99
72) Anthracene	17.274	178	2589878	42.428	ng	100
73) Carbazole	17.551	167	2301037	43.441	ng	100
74) Di-n-butylphthalate	18.104	149	2665191	42.251	ng	99
75) Fluoranthene	19.203	202	3053515	43.118	ng	100
77) Benzidine	19.392	184	1267141	32.073	ng	100
78) Pyrene	19.568	202	3204476	41.027	ng	100
80) Butylbenzylphthalate	20.462	149	1193455	40.796	ng	98
81) Benzo(a)anthracene	21.368	228	3314921	43.279	ng	100
82) 3,3'-Dichlorobenzidine	21.292	252	775949	26.464	ng	100
83) Chrysene	21.427	228	3006106	42.406	ng	100
84) Bis(2-ethylhexyl)phtha...	21.280	149	1748323	40.011	ng	99
85) Di-n-octyl phthalate	22.403	149	2992138	41.523	ng	100
87) Indeno(1,2,3-cd)pyrene	27.774	276	3912545	45.204	ng	99
88) Benzo(b)fluoranthene	23.433	252	3221230	42.642	ng	100
89) Benzo(k)fluoranthene	23.491	252	3194428	41.805	ng	99
90) Benzo(a)pyrene	24.238	252	3046262	43.055	ng	100
91) Dibenzo(a,h)anthracene	27.815	278	3203191	45.400	ng	99
92) Benzo(g,h,i)perylene	28.820	276	3027834	44.826	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050147.D
Acq On : 08 May 2025 15:21
Operator : RC/JU
Sample : Q1964-01MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 08 15:55:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM050825\
Data File : BM050147.D
Acq On : 08 May 2025 15:21
Operator : RC/JU
Sample : Q1964-01MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 08 15:55:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 08 12:03:21 2025
Response via : Initial Calibration

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.740 min Scan# 80
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:152 Resp: 25343

Ion	Ratio	Lower	Upper
152	100		
150	155.0	124.1	186.1
115	51.4	41.2	61.8



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#2
1,4-Dioxane
Concen: 37.217 ng
RT: 3.228 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 88 Resp: 230959

Ion	Ratio	Lower	Upper
88	100		
58	70.5	55.1	82.7
43	28.7	20.9	31.3





#3
Pyridine
Concen: 39.580 ng
RT: 3.634 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 79 Resp: 6310

Ion	Ratio	Lower	Upper
79	100		
52	61.9	48.6	72.8
51	31.0	24.6	37.0



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#4
n-Nitrosodimethylamine
Concen: 41.343 ng
RT: 3.540 min Scan# 94
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 42 Resp: 258544

Ion	Ratio	Lower	Upper
42	100		
74	148.4	120.2	180.2
44	7.7	6.2	9.2





#5
2-Fluorophenol
Concen: 102.574 ng
RT: 5.334 min Scan# 3
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:112 Resp: 148458

Ion	Ratio	Lower	Upper
112	100		
64	54.2	42.6	64.0
63	28.1	22.2	33.4

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#6
Aniline
Concen: 21.113 ng
RT: 7.069 min Scan# 694
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 93 Resp: 484984

Ion	Ratio	Lower	Upper
93	100		
66	37.5	29.9	44.9
65	18.9	14.5	21.7





#7
Phenol-d6
Concen: 101.568 ng
RT: 6.934 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 99 Resp: 184763

Ion	Ratio	Lower	Upper
99	100		
42	15.6	13.0	19.4
71	31.4	25.6	38.4



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#8
2-Chlorophenol
Concen: 42.601 ng
RT: 7.316 min Scan# 736
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:128 Resp: 666662

Ion	Ratio	Lower	Upper
128	100		
130	33.0	12.8	52.8
64	42.0	21.2	61.2





#9
Benzaldehyde
Concen: 23.135 ng
RT: 6.887 min Scan# 6
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 77 Resp: 28163

Ion	Ratio	Lower	Upper
77	100		
105	97.7	80.5	120.5
106	94.8	76.1	116.1



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#10
Phenol
Concen: 43.327 ng
RT: 6.957 min Scan# 675
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 94 Resp: 797888

Ion	Ratio	Lower	Upper
94	100		
65	26.6	7.6	47.6
66	35.1	15.8	55.8





#11
bis(2-Chloroethyl)ether
Concen: 40.695 ng
RT: 7.163 min Scan# 7
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 93 Resp: 62578

Ion	Ratio	Lower	Upper
93	100		
63	71.9	52.3	92.3
95	31.9	12.8	52.8



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#12
1,3-Dichlorobenzene
Concen: 40.647 ng
RT: 7.628 min Scan# 789
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:146 Resp: 768371

Ion	Ratio	Lower	Upper
146	100		
148	64.8	52.5	78.7
75	28.1	22.6	34.0





#13
1,4-Dichlorobenzene
Concen: 40.728 ng
RT: 7.775 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:146 Resp: 77345

Ion	Ratio	Lower	Upper
146	100		
148	64.4	52.2	78.2
111	36.9	29.3	43.9



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#14
1,2-Dichlorobenzene
Concen: 40.732 ng
RT: 8.092 min Scan# 868
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:146 Resp: 747198

Ion	Ratio	Lower	Upper
146	100		
148	64.1	52.4	78.6
111	37.6	31.0	46.6





#15
Benzyl Alcohol
Concen: 42.231 ng
RT: 7.992 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 79 Resp: 53276

Ion	Ratio	Lower	Upper
79	100		
108	81.3	66.0	99.0
77	65.1	51.6	77.4



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.272 ng
RT: 8.263 min Scan# 897
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 45 Resp: 752772

Ion	Ratio	Lower	Upper
45	100		
77	16.9	0.0	36.4
79	12.6	0.0	32.7





#17
2-Methylphenol
Concen: 42.175 ng
RT: 8.204 min Scan# 8
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:107 Resp: 51165

Ion	Ratio	Lower	Upper
107	100		
108	113.2	91.1	136.7
77	46.0	36.2	54.4
79	44.2	35.8	53.8



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#18
Hexachloroethane
Concen: 40.554 ng
RT: 8.810 min Scan# 990
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:117 Resp: 273774

Ion	Ratio	Lower	Upper
117	100		
119	96.4	77.3	115.9
201	124.9	102.8	154.2





#19
n-Nitroso-di-n-propylamine
Concen: 39.511 ng
RT: 8.551 min Scan# 94
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 70 Resp: 46097

Ion	Ratio	Lower	Upper
70	100		
42	50.6	40.5	60.7
101	9.9	7.8	11.6
130	22.1	17.4	26.2



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#20
3+4-Methylphenols
Concen: 41.678 ng
RT: 8.534 min Scan# 943
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:107 Resp: 682817

Ion	Ratio	Lower	Upper
107	100		
108	88.8	72.4	112.4
77	33.1	13.6	53.6
79	26.9	9.2	49.2





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.533 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:136 Resp: 88447

Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.9	13.3
54	6.9	5.1	7.7
68	5.6	4.4	6.6



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#22
Acetophenone
Concen: 41.510 ng
RT: 8.569 min Scan# 949
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:105 Resp: 912791

Ion	Ratio	Lower	Upper
105	100		
71	1.2	1.7	2.5#
51	24.1	18.6	27.8
120	23.6	18.3	27.5





#23
Nitrobenzene-d5
Concen: 60.508 ng
RT: 8.904 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 82 Resp: 108606

Ion	Ratio	Lower	Upper
82	100		
128	45.4	35.7	53.5
54	45.0	35.4	53.2



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#24
Nitrobenzene
Concen: 42.146 ng
RT: 8.945 min Scan# 1013
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 77 Resp: 654098

Ion	Ratio	Lower	Upper
77	100		
123	48.8	37.9	56.9
65	12.7	10.1	15.1





#25
Isophorone
Concen: 39.884 ng
RT: 9.469 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 82 Resp: 122030

Ion	Ratio	Lower	Upper
82	100		
95	6.9	5.7	8.5
138	18.9	15.5	23.3



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#26
2-Nitrophenol
Concen: 43.816 ng
RT: 9.657 min Scan# 1134
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:139 Resp: 347403

Ion	Ratio	Lower	Upper
139	100		
109	27.9	21.9	32.9
65	44.9	35.4	53.2





#27
2,4-Dimethylphenol
Concen: 42.595 ng
RT: 9.728 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:122 Resp: 56006

Ion	Ratio	Lower	Upper
122	100		
107	113.3	89.0	133.6
121	57.2	45.6	68.4



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#28
bis(2-Chloroethoxy)methane
Concen: 40.935 ng
RT: 9.945 min Scan# 1183
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 93 Resp: 774120

Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.1	39.1
123	11.9	9.5	14.3





#29
2,4-Dichlorophenol
Concen: 43.473 ng
RT: 10.210 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:162 Resp: 64038

Ion	Ratio	Lower	Upper
162	100		
164	64.7	45.4	85.4
98	31.5	11.3	51.3



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#30
1,2,4-Trichlorobenzene
Concen: 41.818 ng
RT: 10.398 min Scan# 1260
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:180 Resp: 750507

Ion	Ratio	Lower	Upper
180	100		
182	96.1	77.0	115.4
145	25.8	20.6	30.8





#31
Naphthalene
Concen: 41.226 ng
RT: 10.581 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:128 Resp: 184661

Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.8	13.2
127	12.8	10.4	15.6



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#32
Benzoic acid
Concen: 48.332 ng m
RT: 9.904 min Scan# 1176
Delta R.T. -0.012 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:122 Resp: 406451

Ion	Ratio	Lower	Upper
122	100		
105	119.8	103.6	143.6
77	94.0	73.7	113.7





#33
4-Chloroaniline
Concen: 15.619 ng
RT: 10.710 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:127 Resp: 28675

Ion	Ratio	Lower	Upper
127	100		
129	32.8	26.3	39.5
65	29.0	22.5	33.7
92	18.1	14.5	21.7



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#34
Hexachlorobutadiene
Concen: 41.880 ng
RT: 10.863 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:225 Resp: 477099

Ion	Ratio	Lower	Upper
225	100		
223	63.5	50.2	75.2
227	64.1	50.6	76.0





#35
Caprolactam
Concen: 39.463 ng
RT: 11.504 min Scan# 14
Delta R.T. -0.012 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:113 Resp: 16949

Ion	Ratio	Lower	Upper
113	100		
55	143.3	122.9	162.9
56	108.6	88.1	128.1



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#36
4-Chloro-3-methylphenol
Concen: 41.154 ng
RT: 11.857 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:107 Resp: 560083

Ion	Ratio	Lower	Upper
107	100		
144	30.3	23.8	35.6
142	91.1	71.9	107.9





#37
2-Methylnaphthalene
Concen: 40.394 ng
RT: 12.198 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:142 Resp: 121306

Ion	Ratio	Lower	Upper
142	100		
141	87.8	70.4	105.6
115	32.2	25.4	38.2



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#38
1-Methylnaphthalene
Concen: 39.949 ng
RT: 12.422 min Scan# 1604
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:142 Resp: 1269611

Ion	Ratio	Lower	Upper
142	100		
141	90.5	72.1	108.1
116	3.6	2.7	4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.386 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:164 Resp: 55675

Ion	Ratio	Lower	Upper
164	100		
162	98.5	79.4	119.0
160	43.6	36.2	54.4



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 43.251 ng
RT: 12.575 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:216 Resp: 833595

Ion	Ratio	Lower	Upper
216	100		
214	78.5	62.3	93.5
179	19.2	15.4	23.0
108	15.7	11.7	17.5





#41
Hexachlorocyclopentadiene
Concen: 68.825 ng
RT: 12.551 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:237 Resp: 81063

Ion	Ratio	Lower	Upper
237	100		
235	62.6	42.9	82.9
272	13.2	0.0	32.9



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#42
2,4,6-Tribromophenol
Concen: 99.732 ng
RT: 15.886 min Scan# 2193
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:330 Resp: 821030

Ion	Ratio	Lower	Upper
330	100		
332	96.5	77.3	115.9
141	27.4	22.5	33.7





#43
2,4,6-Trichlorophenol
Concen: 44.513 ng
RT: 12.827 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:196 Resp: 54414

Ion	Ratio	Lower	Upper
196	100		
198	94.1	79.0	118.6
200	29.9	25.1	37.7



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#44
2,4,5-Trichlorophenol
Concen: 44.383 ng
RT: 12.922 min Scan# 1689
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:196 Resp: 585065

Ion	Ratio	Lower	Upper
196	100		
198	96.3	76.7	115.1
97	38.2	30.1	45.1
132	21.0	16.6	25.0





#45
2-Fluorobiphenyl
Concen: 53.437 ng
RT: 13.004 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:172 Resp: 251492

Ion	Ratio	Lower	Upper
172	100		
171	34.2	27.8	41.6
170	23.0	18.6	28.0



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#46
1,1'-Biphenyl
Concen: 42.246 ng
RT: 13.216 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:154 Resp: 1748393

Ion	Ratio	Lower	Upper
154	100		
153	39.9	20.0	60.0
76	11.3	0.0	30.3





#47
2-Chloronaphthalene
Concen: 42.545 ng
RT: 13.263 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:162 Resp: 139453

Ion	Ratio	Lower	Upper
162	100		
127	32.2	25.9	38.9
164	33.2	26.6	40.0



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#48
2-Nitroaniline
Concen: 44.229 ng
RT: 13.480 min Scan# 1784
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 65 Resp: 340250

Ion	Ratio	Lower	Upper
65	100		
92	71.4	57.1	85.7
138	121.6	97.0	145.4





#49
Acenaphthylene
Concen: 41.840 ng
RT: 14.110 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:152 Resp: 216233

Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.8
153	13.4	10.6	15.8



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#50
Dimethylphthalate
Concen: 40.686 ng
RT: 13.845 min Scan# 1846
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:163 Resp: 1655423

Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.4	5.2
164	10.3	8.2	12.2





#51
2,6-Dinitrotoluene
Concen: 42.724 ng
RT: 13.974 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:165 Resp: 35991

Ion	Ratio	Lower	Upper
165	100		
63	44.2	35.1	52.7
89	46.4	38.4	57.6



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#52
Acenaphthene
Concen: 41.908 ng
RT: 14.451 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:154 Resp: 1253654

Ion	Ratio	Lower	Upper
154	100		
153	111.5	89.0	133.4
152	53.1	43.0	64.6





#53
3-Nitroaniline
Concen: 25.665 ng
RT: 14.316 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:138 Resp: 20642

Ion	Ratio	Lower	Upper
138	100		
108	9.8	7.4	11.2
92	103.6	84.0	126.0



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#54
2,4-Dinitrophenol
Concen: 80.794 ng
RT: 14.533 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:184 Resp: 435912

Ion	Ratio	Lower	Upper
184	100		
63	49.9	39.8	59.6
154	58.2	46.1	69.1





#55
Dibenzofuran
Concen: 41.059 ng
RT: 14.792 min Scan# 2(
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:168 Resp: 204357

Ion	Ratio	Lower	Upper
168	100		
139	33.2	26.1	39.1
169	13.2	10.8	16.2



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#56
4-Nitrophenol
Concen: 83.663 ng
RT: 14.663 min Scan# 1985
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:139 Resp: 548863

Ion	Ratio	Lower	Upper
139	100		
109	68.2	47.0	87.0
65	86.6	64.3	104.3





#57
2,4-Dinitrotoluene
Concen: 43.550 ng
RT: 14.774 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:165 Resp: 50695

Ion	Ratio	Lower	Upper
165	100		
63	40.7	29.0	43.6
89	62.1	49.4	74.2
182	2.6	2.2	3.2



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#58
Fluorene
Concen: 41.709 ng
RT: 15.439 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:166 Resp: 1699192

Ion	Ratio	Lower	Upper
166	100		
165	95.9	76.6	115.0
167	13.5	10.6	16.0





#59
2,3,4,6-Tetrachlorophenol
Concen: 42.159 ng
RT: 15.033 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:232 Resp: 48105

Ion	Ratio	Lower	Upper
232	100		
131	35.3	28.2	42.4
130	2.0	1.5	2.3
166	27.3	22.6	34.0



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#60
Diethylphthalate
Concen: 40.175 ng
RT: 15.216 min Scan# 2079
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:149 Resp: 1540274

Ion	Ratio	Lower	Upper
149	100		
177	24.1	19.1	28.7
150	12.0	9.6	14.4





#61
4-Chlorophenyl-phenylether
Concen: 41.368 ng
RT: 15.433 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:204 Resp: 92565

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.3	39.5
141	43.5	34.0	51.0



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#62
4-Nitroaniline
Concen: 38.598 ng
RT: 15.480 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:138 Resp: 299512

Ion	Ratio	Lower	Upper
138	100		
92	47.2	25.9	65.9
108	75.0	52.4	92.4





#63
Azobenzene
Concen: 41.722 ng
RT: 15.727 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion: 77 Resp: 134059

Ion	Ratio	Lower	Upper
77	100		
182	31.1	10.5	50.5
105	21.4	1.1	41.1
51	26.4	6.3	46.3



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED
Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2406
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:188 Resp: 1069463

Ion	Ratio	Lower	Upper
188	100		
94	7.3	5.7	8.5
80	7.8	6.2	9.4





#65
4,6-Dinitro-2-methylphenol
Concen: 41.801 ng
RT: 15.545 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:198 Resp: 27983

Ion	Ratio	Lower	Upper
198	100		
51	31.6	10.4	50.4
105	35.6	14.6	54.6

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#66
n-Nitrosodiphenylamine
Concen: 42.749 ng
RT: 15.651 min Scan# 2153
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:169 Resp: 1403891

Ion	Ratio	Lower	Upper
169	100		
168	66.5	53.4	80.2
167	35.9	29.4	44.0





#67
4-Bromophenyl-phenylether
Concen: 42.387 ng
RT: 16.327 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:248 Resp: 54818

Ion	Ratio	Lower	Upper
248	100		
250	97.8	76.2	114.2
141	55.3	45.9	68.9



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#68
Hexachlorobenzene
Concen: 42.614 ng
RT: 16.451 min Scan# 2289
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:284 Resp: 635267

Ion	Ratio	Lower	Upper
284	100		
142	27.5	23.9	35.9
249	30.5	24.5	36.7





#69
Atrazine
Concen: 42.756 ng
RT: 16.610 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:200 Resp: 50648

Ion	Ratio	Lower	Upper
200	100		
173	25.4	4.3	44.3
215	50.7	30.6	70.6



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#70
Pentachlorophenol
Concen: 97.109 ng
RT: 16.804 min Scan# 2349
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:266 Resp: 814874

Ion	Ratio	Lower	Upper
266	100		
268	64.6	51.2	76.8
264	63.4	49.6	74.4





#71
Phenanthrene
Concen: 42.759 ng
RT: 17.180 min Scan# 2
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:178 Resp: 257920

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.3	22.9
179	15.2	12.4	18.6



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#72
Anthracene
Concen: 42.428 ng
RT: 17.274 min Scan# 2429
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:178 Resp: 2589878

Ion	Ratio	Lower	Upper
178	100		
176	18.8	15.0	22.4
179	15.4	12.2	18.4





#73
Carbazole
Concen: 43.441 ng
RT: 17.551 min Scan# 24
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:167 Resp: 230103

Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.2	25.8
139	11.1	9.0	13.6

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#74
Di-n-butylphthalate
Concen: 42.251 ng
RT: 18.104 min Scan# 2570
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:149 Resp: 2665191

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.3	10.9
104	4.4	3.8	5.6





#75
Fluoranthene
Concen: 43.118 ng
RT: 19.203 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:202 Resp: 305351

Ion	Ratio	Lower	Upper
202	100		
101	7.9	0.0	28.0
203	17.5	0.0	37.7



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.386 min Scan# 3128
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:240 Resp: 1112266

Ion	Ratio	Lower	Upper
240	100		
120	8.4	6.5	9.7
236	26.1	20.9	31.3





#77
Benzidine
Concen: 32.073 ng
RT: 19.392 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:184 Resp: 126714

Ion	Ratio	Lower	Upper
184	100		
185	14.2	11.5	17.3
183	11.1	8.9	13.3



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#78
Pyrene
Concen: 41.027 ng
RT: 19.568 min Scan# 2819
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:202 Resp: 3204476

Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.7	25.1
203	17.5	14.1	21.1





#79
Terphenyl-d14
Concen: 47.900 ng
RT: 19.768 min Scan# 2
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:244 Resp: 360048

Ion	Ratio	Lower	Upper
244	100		
212	6.8	5.6	8.4
122	6.9	5.6	8.4

Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025





#80
Butylbenzylphthalate
Concen: 40.796 ng
RT: 20.462 min Scan# 2971
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:149 Resp: 1193455

Ion	Ratio	Lower	Upper
149	100		
91	61.4	47.4	71.2
206	24.2	19.4	29.0





#81
Benzo(a)anthracene
Concen: 43.279 ng
RT: 21.368 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:228 Resp: 331492

Ion	Ratio	Lower	Upper
228	100		
226	27.4	22.1	33.1
229	19.7	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#82
3,3'-Dichlorobenzidine
Concen: 26.464 ng
RT: 21.292 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:252 Resp: 775949

Ion	Ratio	Lower	Upper
252	100		
254	64.5	51.8	77.6
126	7.8	6.0	9.0





#83
Chrysene
Concen: 42.406 ng
RT: 21.427 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:228 Resp: 300610

Ion	Ratio	Lower	Upper
228	100		
226	30.4	24.2	36.2
229	19.9	15.7	23.5



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.011 ng
RT: 21.280 min Scan# 3110
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:149 Resp: 1748323

Ion	Ratio	Lower	Upper
149	100		
167	29.8	24.1	36.1
279	5.2	3.9	5.9





#85
Di-n-octyl phthalate
Concen: 41.523 ng
RT: 22.403 min Scan# 3
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:149 Resp: 299213

Ion	Ratio	Lower	Upper
149	100		
167	1.6	1.3	1.9
43	7.4	5.8	8.8



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.380 min Scan# 3637
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:264 Resp: 1161580

Ion	Ratio	Lower	Upper
264	100		
260	23.8	18.8	28.2
265	22.3	17.0	25.4





#87
Indeno(1,2,3-cd)pyrene
Concen: 45.204 ng
RT: 27.774 min Scan# 4
Delta R.T. 0.012 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:276 Resp: 391254

Ion	Ratio	Lower	Upper
276	100		
138	20.1	16.4	24.6
277	25.1	19.8	29.8



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#88
Benzo(b)fluoranthene
Concen: 42.642 ng
RT: 23.433 min Scan# 3476
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:252 Resp: 3221230

Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.7	26.5
125	7.4	6.1	9.1





#89
Benzo(k)fluoranthene
Concen: 41.805 ng
RT: 23.491 min Scan# 3
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:252 Resp: 319442

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.4	26.2
125	7.7	5.8	8.8



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#90
Benzo(a)pyrene
Concen: 43.055 ng
RT: 24.238 min Scan# 3613
Delta R.T. -0.000 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:252 Resp: 3046262

Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.0
125	8.2	6.6	9.8





#91
Dibenzo(a,h)anthracene
Concen: 45.400 ng
RT: 27.815 min Scan# 4
Delta R.T. 0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:278 Resp: 320319

Ion	Ratio	Lower	Upper
278	100		
139	12.6	10.5	15.7
279	23.7	19.4	29.0



Instrument :
BNA_M
ClientSampleId :
MH-LLMSD

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/09/2025
Supervised By :Jagrut Upadhyay 05/09/2025



#92
Benzo(g,h,i)perylene
Concen: 44.826 ng
RT: 28.820 min Scan# 4392
Delta R.T. -0.006 min
Lab File: BM050147.D
Acq: 08 May 2025 15:21

Tgt Ion:276 Resp: 3027834

Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.1	28.7
138	17.5	14.3	21.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BM042825	Instrument	BNA_m

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BM050025.D	4-Nitroaniline	Rahul	4/29/2025 8:55:04 AM	Jagrut	4/29/2025 11:57:10 AM	Peak Integrated by Software
SSTDICC005	BM050025.D	Benzaldehyde	Rahul	4/29/2025 8:55:04 AM	Jagrut	4/29/2025 11:57:10 AM	Peak Integrated by Software
SSTDICC005	BM050025.D	Benzo(k)fluoranthene	Rahul	4/29/2025 8:55:04 AM	Jagrut	4/29/2025 11:57:10 AM	Peak Integrated by Software
SSTDICC005	BM050025.D	Caprolactam	Rahul	4/29/2025 8:55:04 AM	Jagrut	4/29/2025 11:57:10 AM	Peak Integrated by Software
SSTDICC010	BM050026.D	4-Nitroaniline	Rahul	4/29/2025 8:55:07 AM	Jagrut	4/29/2025 11:57:13 AM	Peak Integrated by Software
SSTDICC010	BM050026.D	Benzaldehyde	Rahul	4/29/2025 8:55:07 AM	Jagrut	4/29/2025 11:57:13 AM	Peak Integrated by Software
SSTDICC010	BM050026.D	Benzoic acid	Rahul	4/29/2025 8:55:07 AM	Jagrut	4/29/2025 11:57:13 AM	Peak Integrated by Software
SSTDICC020	BM050027.D	Benzoic acid	Rahul	4/29/2025 8:55:10 AM	Jagrut	4/29/2025 11:57:15 AM	Peak Integrated by Software
SSTDICV040	BM050032.D	Benzaldehyde	Rahul	4/29/2025 8:55:13 AM	Jagrut	4/29/2025 11:57:18 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	bm050825	Instrument	BNA_m

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1964-01MSD	BM050147.D	Benzoic acid	Rahul	5/9/2025 9:43:12 AM	Jagrut	5/9/2025 3:39:56 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	bp041425	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BP024276.D	Benzaldehyde	Rahul	4/15/2025 9:45:41 AM	Jagrut	4/15/2025 10:52:38 AM	Peak Integrated by Software
SSTDICC005	BP024276.D	Benzidine	Rahul	4/15/2025 9:45:41 AM	Jagrut	4/15/2025 10:52:38 AM	Peak Integrated by Software
SSTDICC010	BP024277.D	Benzaldehyde	Rahul	4/15/2025 9:46:12 AM	Jagrut	4/15/2025 10:52:41 AM	Peak Integrated by Software
SSTDICC010	BP024277.D	Benzoic acid	Rahul	4/15/2025 9:46:12 AM	Jagrut	4/15/2025 10:52:41 AM	Peak Integrated by Software
SSTDICC020	BP024278.D	Benzoic acid	Rahul	4/15/2025 9:46:28 AM	Jagrut	4/15/2025 10:52:44 AM	Peak Integrated by Software
SSTDICC050	BP024280.D	Indeno(1,2,3-cd)pyrene	Rahul	4/15/2025 9:47:02 AM	Jagrut	4/15/2025 10:52:46 AM	Peak Integrated by Software
SSTDICC050	BP024280.D	Pyridine	Rahul	4/15/2025 9:47:02 AM	Jagrut	4/15/2025 10:52:46 AM	Peak Integrated by Software
SSTDICC060	BP024281.D	Benzoic acid	Rahul	4/15/2025 9:47:31 AM	Jagrut	4/15/2025 10:52:48 AM	Peak Integrated by Software
SSTDICC060	BP024281.D	Indeno(1,2,3-cd)pyrene	Rahul	4/15/2025 9:47:31 AM	Jagrut	4/15/2025 10:52:48 AM	Peak Integrated by Software
SSTDICC060	BP024281.D	Pyridine	Rahul	4/15/2025 9:47:31 AM	Jagrut	4/15/2025 10:52:48 AM	Peak Integrated by Software
SSTDICC080	BP024282.D	Benzidine	Rahul	4/15/2025 9:48:04 AM	Jagrut	4/15/2025 10:52:51 AM	Peak Integrated by Software
SSTDICC080	BP024282.D	Benzoic acid	Rahul	4/15/2025 9:48:04 AM	Jagrut	4/15/2025 10:52:51 AM	Peak Integrated by Software
SSTDICC080	BP024282.D	Pyridine	Rahul	4/15/2025 9:48:04 AM	Jagrut	4/15/2025 10:52:51 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	bp041425	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICV040	BP024283.D	Indeno(1,2,3-cd)pyrene	Rahul	4/15/2025 9:50:13 AM	Jagrut	4/15/2025 10:52:54 AM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BP050825	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BP024568.D	Benzo(k)fluoranthene	Rahul	5/9/2025 9:46:56 AM	Jagrut	5/9/2025 3:43:06 PM	Peak Integrated by Software
PB167904BS	BP024570.D	Caprolactam	Rahul	5/9/2025 9:47:00 AM	Jagrut	5/9/2025 3:43:08 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BP051325	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BP024615.D	4-Nitroaniline	Rahul	5/14/2025 2:19:40 PM	Jagrut	5/14/2025 5:15:55 PM	Peak Integrated by Software
SSTDICC005	BP024615.D	Benzaldehyde	Rahul	5/14/2025 2:19:40 PM	Jagrut	5/14/2025 5:15:55 PM	Peak Integrated by Software
SSTDICC005	BP024615.D	Caprolactam	Rahul	5/14/2025 2:19:40 PM	Jagrut	5/14/2025 5:15:55 PM	Peak Integrated by Software
SSTDICC010	BP024616.D	4-Nitroaniline	Rahul	5/14/2025 2:19:43 PM	Jagrut	5/14/2025 5:16:07 PM	Peak Integrated by Software
SSTDICC010	BP024616.D	Benzaldehyde	Rahul	5/14/2025 2:19:43 PM	Jagrut	5/14/2025 5:16:07 PM	Peak Integrated by Software
SSTDICC010	BP024616.D	Benzidine	Rahul	5/14/2025 2:19:43 PM	Jagrut	5/14/2025 5:16:07 PM	Peak Integrated by Software
SSTDICC010	BP024616.D	Benzoic acid	Rahul	5/14/2025 2:19:43 PM	Jagrut	5/14/2025 5:16:07 PM	Peak Integrated by Software
SSTDICC010	BP024616.D	Caprolactam	Rahul	5/14/2025 2:19:43 PM	Jagrut	5/14/2025 5:16:07 PM	Peak Integrated by Software
SSTDICC010	BP024616.D	Pentachlorophenol	Rahul	5/14/2025 2:19:43 PM	Jagrut	5/14/2025 5:16:07 PM	Peak Integrated by Software
SSTDICC020	BP024617.D	Benzaldehyde	Rahul	5/14/2025 2:19:45 PM	Jagrut	5/14/2025 5:16:09 PM	Peak Integrated by Software
SSTDICC020	BP024617.D	Benzoic acid	Rahul	5/14/2025 2:19:45 PM	Jagrut	5/14/2025 5:16:09 PM	Peak Integrated by Software
SSTDICC020	BP024617.D	Indeno(1,2,3-cd)pyrene	Rahul	5/14/2025 2:19:45 PM	Jagrut	5/14/2025 5:16:09 PM	Peak Integrated by Software
SSTDICCC040	BP024618.D	Benzaldehyde	Rahul	5/14/2025 2:19:48 PM	Jagrut	5/14/2025 5:16:12 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BP051325	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICCC040	BP024618.D	Benzoic acid	Rahul	5/14/2025 2:19:48 PM	Jagrut	5/14/2025 5:16:12 PM	Peak Integrated by Software
SSTDICC080	BP024621.D	Indeno(1,2,3-cd)pyrene	Rahul	5/14/2025 2:19:51 PM	Jagrut	5/14/2025 5:16:14 PM	Peak Integrated by Software
SSTDICV040	BP024622.D	Indeno(1,2,3-cd)pyrene	Rahul	5/14/2025 2:19:57 PM	Jagrut	5/14/2025 5:16:17 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BP051425	Instrument	BNA_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BP024625.D	Benzaldehyde	Rahul	5/15/2025 9:34:58 AM	Jagrut	5/15/2025 3:51:50 PM	Peak Integrated by Software
SSTDCCC040	BP024625.D	Benzoic acid	Rahul	5/15/2025 9:34:58 AM	Jagrut	5/15/2025 3:51:50 PM	Peak Integrated by Software
SSTDCCC040	BP024625.D	Dibenzo(a,h)anthracene	Rahul	5/15/2025 9:34:58 AM	Jagrut	5/15/2025 3:51:50 PM	Peak Integrated by Software
SSTDCCC040	BP024632.D	Benzoic acid	Rahul	5/15/2025 9:35:02 AM	Jagrut	5/15/2025 3:51:53 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QCBatch ID # BM042825

Review By	Rahul	Review On	4/29/2025 8:56:26 AM		
Supervise By	Jagrut	Supervise On	4/29/2025 11:57:30 AM		
SubDirectory	BM042825	HP Acquire Method	BNA_M	HP Processing Method	BM042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM050023.D	28 Apr 2025 11:46	RC/JU	Ok
2	SSTDICC2.5	BM050024.D	28 Apr 2025 12:30	RC/JU	Ok
3	SSTDICC005	BM050025.D	28 Apr 2025 13:09	RC/JU	Ok,M
4	SSTDICC010	BM050026.D	28 Apr 2025 13:48	RC/JU	Ok,M
5	SSTDICC020	BM050027.D	28 Apr 2025 14:27	RC/JU	Ok,M
6	SSTDICCC040	BM050028.D	28 Apr 2025 15:06	RC/JU	Ok
7	SSTDICC050	BM050029.D	28 Apr 2025 15:45	RC/JU	Ok
8	SSTDICC060	BM050030.D	28 Apr 2025 16:24	RC/JU	Ok
9	SSTDICC080	BM050031.D	28 Apr 2025 17:04	RC/JU	Ok
10	SSTDICV040	BM050032.D	28 Apr 2025 17:43	RC/JU	Ok,M
11	PB167739BL	BM050033.D	28 Apr 2025 19:40	RC/JU	Not Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QCBatch ID # BM050825

Review By	Rahul	Review On	5/9/2025 9:43:48 AM				
Supervise By	Jagrut	Supervise On	5/9/2025 3:41:05 PM				
SubDirectory	BM050825	HP Acquire Method	BNA_M		HP Processing Method	BM042825	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12663,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM050139.D	08 May 2025 09:59	RC/JU	Ok
2	SSTDCCC040	BM050140.D	08 May 2025 10:41	RC/JU	Ok
3	PB167889BL	BM050141.D	08 May 2025 11:21	RC/JU	Ok
4	PB167889BS	BM050142.D	08 May 2025 12:00	RC/JU	Ok
5	PB167889BSD	BM050143.D	08 May 2025 12:39	RC/JU	Ok,M
6	PB167851TB	BM050144.D	08 May 2025 13:18	RC/JU	Ok
7	Q1964-01	BM050145.D	08 May 2025 14:02	RC/JU	Ok
8	Q1964-01MS	BM050146.D	08 May 2025 14:42	RC/JU	Ok
9	Q1964-01MSD	BM050147.D	08 May 2025 15:21	RC/JU	Ok,M
10	Q1966-02	BM050148.D	08 May 2025 16:00	RC/JU	Ok
11	Q1968-01	BM050149.D	08 May 2025 16:39	RC/JU	Ok
12	Q1972-13	BM050150.D	08 May 2025 17:19	RC/JU	Ok
13	Q1978-01	BM050151.D	08 May 2025 17:58	RC/JU	Ok
14	Q1976-01	BM050152.D	08 May 2025 18:37	RC/JU	Ok
15	Q1972-21	BM050153.D	08 May 2025 19:17	RC/JU	Ok,M
16	Q1968-05	BM050154.D	08 May 2025 19:56	RC/JU	Ok,M
17	Q1969-01	BM050155.D	08 May 2025 20:35	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM		
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM		
SubDirectory	BP041425	HP Acquire Method	BNA_P	HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024274.D	14 Apr 2025 10:25	RC/JU	Ok
2	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06	RC/JU	Ok
3	SSTDICC005	BP024276.D	14 Apr 2025 11:47	RC/JU	Ok,M
4	SSTDICC010	BP024277.D	14 Apr 2025 12:27	RC/JU	Ok,M
5	SSTDICC020	BP024278.D	14 Apr 2025 13:08	RC/JU	Ok,M
6	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	RC/JU	Ok
7	SSTDICC050	BP024280.D	14 Apr 2025 15:10	RC/JU	Ok,M
8	SSTDICC060	BP024281.D	14 Apr 2025 16:32	RC/JU	Ok,M
9	SSTDICC080	BP024282.D	14 Apr 2025 17:13	RC/JU	Ok,M
10	SSTDICV040	BP024283.D	14 Apr 2025 18:35	RC/JU	Ok,M
11	PB167488TB	BP024284.D	14 Apr 2025 19:16	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP050825

Review By	Rahul	Review On	5/9/2025 12:08:37 PM				
Supervise By	Jagrut	Supervise On	5/9/2025 4:03:53 PM				
SubDirectory	BP050825	HP Acquire Method	BNA_P		HP Processing Method	bp041425	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12663,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024567.D	08 May 2025 13:24	RC/JU	Ok
2	SSTDCCC040	BP024568.D	08 May 2025 14:45	RC/JU	Ok,M
3	PB167904BL	BP024569.D	08 May 2025 15:26	RC/JU	Ok
4	PB167904BS	BP024570.D	08 May 2025 16:07	RC/JU	Ok,M
5	Q1947-04RE	BP024571.D	08 May 2025 16:55	RC/JU	Confirms
6	Q1966-01	BP024572.D	08 May 2025 17:36	RC/JU	Ok
7	Q1972-01	BP024573.D	08 May 2025 18:17	RC/JU	Ok
8	Q1972-05	BP024574.D	08 May 2025 18:58	RC/JU	Ok
9	Q1972-09	BP024575.D	08 May 2025 19:39	RC/JU	Ok
10	Q1972-17	BP024576.D	08 May 2025 20:19	RC/JU	Ok
11	Q1968-08	BP024577.D	08 May 2025 21:00	RC/JU	ReRun
12	Q1972-04	BP024578.D	08 May 2025 21:41	RC/JU	Ok,M
13	Q1972-08	BP024579.D	08 May 2025 22:22	RC/JU	ReRun
14	Q1972-12	BP024580.D	08 May 2025 23:02	RC/JU	Ok
15	Q1972-16	BP024581.D	08 May 2025 23:43	RC/JU	ReRun
16	Q1972-20	BP024582.D	09 May 2025 00:24	RC/JU	Ok
17	Q1972-24	BP024583.D	09 May 2025 01:04	RC/JU	ReRun

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP051325

Review By	Rahul	Review On	5/14/2025 2:20:20 PM		
Supervise By	Jagrut	Supervise On	5/14/2025 5:16:42 PM		
SubDirectory	BP051325	HP Acquire Method	BNA_P	HP Processing Method	bp051325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12664,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024613.D	13 May 2025 10:00	RC/JU	Ok
2	SSTDICC2.5	BP024614.D	13 May 2025 10:41	RC/JU	Ok
3	SSTDICC005	BP024615.D	13 May 2025 11:22	RC/JU	Ok,M
4	SSTDICC010	BP024616.D	13 May 2025 12:03	RC/JU	Ok,M
5	SSTDICC020	BP024617.D	13 May 2025 12:43	RC/JU	Ok,M
6	SSTDICCC040	BP024618.D	13 May 2025 13:24	RC/JU	Ok,M
7	SSTDICC050	BP024619.D	13 May 2025 14:05	RC/JU	Ok
8	SSTDICC060	BP024620.D	13 May 2025 14:45	RC/JU	Ok
9	SSTDICC080	BP024621.D	13 May 2025 15:26	RC/JU	Ok,M
10	SSTDICV040	BP024622.D	13 May 2025 17:01	RC/JU	Ok,M
11	PB167917TB	BP024623.D	13 May 2025 18:23	RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP051425

Review By	Rahul	Review On	5/15/2025 9:37:56 AM				
Supervise By	Jagrut	Supervise On	5/15/2025 3:53:11 PM				
SubDirectory	BP051425	HP Acquire Method	BNA_P		HP Processing Method	BP051325	

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12664,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024624.D	14 May 2025 10:29	RC/JU	Ok
2	SSTDCCC040	BP024625.D	14 May 2025 11:35	RC/JU	Ok,M
3	PB167951BL	BP024626.D	14 May 2025 13:17	RC/JU	Ok
4	Q2000-02	BP024627.D	14 May 2025 14:51	RC/JU	Dilution
5	Q2000-02DL	BP024628.D	14 May 2025 15:32	RC/JU	Ok
6	Q1872-11	BP024629.D	14 May 2025 17:15	RC/JU	Dilution
7	Q1872-11DL	BP024630.D	14 May 2025 17:56	RC/JU	Ok
8	Q1872-12	BP024631.D	14 May 2025 18:37	RC/JU	Ok
9	SSTDCCC040	BP024632.D	14 May 2025 19:17	RC/JU	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QCBatch ID # BM042825

Review By	Rahul	Review On	4/29/2025 8:56:26 AM		
Supervise By	Jagrut	Supervise On	4/29/2025 11:57:30 AM		
SubDirectory	BM042825	HP Acquire Method	BNA_M	HP Processing Method	BM042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050023.D	28 Apr 2025 11:46		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM050024.D	28 Apr 2025 12:30		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM050025.D	28 Apr 2025 13:09	Compound#32-41-54-56-65-70-77 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM050026.D	28 Apr 2025 13:48		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM050027.D	28 Apr 2025 14:27		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM050028.D	28 Apr 2025 15:06	Compound#54 & 56 Kept on LR	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BM050029.D	28 Apr 2025 15:45		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BM050030.D	28 Apr 2025 16:24		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BM050031.D	28 Apr 2025 17:04		RC/JU	Ok
10	SSTDICV040	ICVBM042825	BM050032.D	28 Apr 2025 17:43		RC/JU	Ok,M
11	PB167739BL	PB167739BL	BM050033.D	28 Apr 2025 19:40	Analyzed for contamination check	RC/JU	Not Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QCBatch ID # BM050825

Review By	Rahul	Review On	5/9/2025 9:43:48 AM		
Supervise By	Jagrut	Supervise On	5/9/2025 3:41:05 PM		
SubDirectory	BM050825	HP Acquire Method	BNA_M	HP Processing Method	BM042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12663,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050139.D	08 May 2025 09:59		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050140.D	08 May 2025 10:41		RC/JU	Ok
3	PB167889BL	PB167889BL	BM050141.D	08 May 2025 11:21		RC/JU	Ok
4	PB167889BS	PB167889BS	BM050142.D	08 May 2025 12:00		RC/JU	Ok
5	PB167889BSD	PB167889BSD	BM050143.D	08 May 2025 12:39		RC/JU	Ok,M
6	PB167851TB	PB167851TB	BM050144.D	08 May 2025 13:18		RC/JU	Ok
7	Q1964-01	MH-LL	BM050145.D	08 May 2025 14:02		RC/JU	Ok
8	Q1964-01MS	MH-LLMS	BM050146.D	08 May 2025 14:42		RC/JU	Ok
9	Q1964-01MSD	MH-LLMSD	BM050147.D	08 May 2025 15:21		RC/JU	Ok,M
10	Q1966-02	CONCRETE-2	BM050148.D	08 May 2025 16:00		RC/JU	Ok
11	Q1968-01	MH-N	BM050149.D	08 May 2025 16:39		RC/JU	Ok
12	Q1972-13	WC-B-7	BM050150.D	08 May 2025 17:19		RC/JU	Ok
13	Q1978-01	72-12016	BM050151.D	08 May 2025 17:58		RC/JU	Ok
14	Q1976-01	VNJ-207	BM050152.D	08 May 2025 18:37		RC/JU	Ok
15	Q1972-21	WC-B-1	BM050153.D	08 May 2025 19:17		RC/JU	Ok,M
16	Q1968-05	MH-M	BM050154.D	08 May 2025 19:56		RC/JU	Ok,M
17	Q1969-01	TR-06-05072025	BM050155.D	08 May 2025 20:35		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM		
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM		
SubDirectory	BP041425	HP Acquire Method	BNA_P	HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024274.D	14 Apr 2025 10:25		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024276.D	14 Apr 2025 11:47	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024277.D	14 Apr 2025 12:27		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024278.D	14 Apr 2025 13:08		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	This calibration is fail for Com#77	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024280.D	14 Apr 2025 15:10	This calibration is good for 8270E, 8270 DOD and 625.1 methods except Compound#77	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024281.D	14 Apr 2025 16:32	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024282.D	14 Apr 2025 17:13	Compound #69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP041425	BP024283.D	14 Apr 2025 18:35		RC/JU	Ok,M
11	PB167488TB	PB167488TB	BP024284.D	14 Apr 2025 19:16		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP050825

Review By	Rahul	Review On	5/9/2025 12:08:37 PM		
Supervise By	Jagrut	Supervise On	5/9/2025 4:03:53 PM		
SubDirectory	BP050825	HP Acquire Method	BNA_P	HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12663,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024567.D	08 May 2025 13:24		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024568.D	08 May 2025 14:45		RC/JU	Ok,M
3	PB167904BL	PB167904BL	BP024569.D	08 May 2025 15:26		RC/JU	Ok
4	PB167904BS	PB167904BS	BP024570.D	08 May 2025 16:07		RC/JU	Ok,M
5	Q1947-04RE	MH-PRE	BP024571.D	08 May 2025 16:55	Internal Standard and surrogate Fail	RC/JU	Confirms
6	Q1966-01	CONCRETE-1	BP024572.D	08 May 2025 17:36		RC/JU	Ok
7	Q1972-01	SUB-WC	BP024573.D	08 May 2025 18:17		RC/JU	Ok
8	Q1972-05	WC-B-10	BP024574.D	08 May 2025 18:58		RC/JU	Ok
9	Q1972-09	WC-B-2	BP024575.D	08 May 2025 19:39		RC/JU	Ok
10	Q1972-17	SUB-WC-2	BP024576.D	08 May 2025 20:19		RC/JU	Ok
11	Q1968-08	MH-M	BP024577.D	08 May 2025 21:00	Internal Standard Fail	RC/JU	ReRun
12	Q1972-04	SUB-WC	BP024578.D	08 May 2025 21:41		RC/JU	Ok,M
13	Q1972-08	WC-B-10	BP024579.D	08 May 2025 22:22	Internal Standard Fail	RC/JU	ReRun
14	Q1972-12	WC-B-2	BP024580.D	08 May 2025 23:02		RC/JU	Ok
15	Q1972-16	WC-B-7	BP024581.D	08 May 2025 23:43	Internal Standard Fail	RC/JU	ReRun
16	Q1972-20	SUB-WC-2	BP024582.D	09 May 2025 00:24		RC/JU	Ok
17	Q1972-24	WC-B-1	BP024583.D	09 May 2025 01:04	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP051325

Review By	Rahul	Review On	5/14/2025 2:20:20 PM		
Supervise By	Jagrut	Supervise On	5/14/2025 5:16:42 PM		
SubDirectory	BP051325	HP Acquire Method	BNA_P	HP Processing Method	bp051325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12664,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024613.D	13 May 2025 10:00		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024614.D	13 May 2025 10:41		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024615.D	13 May 2025 11:22	Compound#32-54-56-65-77 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024616.D	13 May 2025 12:03		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024617.D	13 May 2025 12:43		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024618.D	13 May 2025 13:24	Compound#32-54-56 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP024619.D	13 May 2025 14:05		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BP024620.D	13 May 2025 14:45		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BP024621.D	13 May 2025 15:26		RC/JU	Ok,M
10	SSTDICV040	ICVBP051325	BP024622.D	13 May 2025 17:01		RC/JU	Ok,M
11	PB167917TB	PB167917TB	BP024623.D	13 May 2025 18:23		RC/JU	Ok

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QCBatch ID # BP051425

Review By	Rahul	Review On	5/15/2025 9:37:56 AM		
Supervise By	Jagrut	Supervise On	5/15/2025 3:53:11 PM		
SubDirectory	BP051425	HP Acquire Method	BNA_P	HP Processing Method	BP051325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6784,SP6785,SP6786,SP6787,SP6788,SP6790,SP6789,SP6791
CCC	SP6787
Internal Standard/PEM	S12664,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024624.D	14 May 2025 10:29		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024625.D	14 May 2025 11:35		RC/JU	Ok,M
3	PB167951BL	PB167951BL	BP024626.D	14 May 2025 13:17		RC/JU	Ok
4	Q2000-02	38073-040323	BP024627.D	14 May 2025 14:51	PT Sample, Need 5X Dilution	RC/JU	Dilution
5	Q2000-02DL	38073-040323DL	BP024628.D	14 May 2025 15:32		RC/JU	Ok
6	Q1872-11	HW0425-PT-BNA-SOIL	BP024629.D	14 May 2025 17:15	PT Sample, Need 5X Dilution	RC/JU	Dilution
7	Q1872-11DL	HW0425-PT-BNA-SOIL	BP024630.D	14 May 2025 17:56		RC/JU	Ok
8	Q1872-12	HW0425-PT-TRIAZINE	BP024631.D	14 May 2025 18:37	PT Sample	RC/JU	Ok
9	SSTDCCC040	SSTDCCC040EC	BP024632.D	14 May 2025 19:17		RC/JU	Ok,M

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/25/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC: LB135545

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1869-01	MH-F	1	1.14	10.43	11.57	10.47	89.5	
Q1869-02	MH-F-EPH	2	1.18	9.96	11.14	10.12	89.8	
Q1869-03	MH-F-VOC	3	1.16	10.28	11.44	10.4	89.9	
Q1871-01	MH-A	4	1.14	9.59	10.73	9.86	90.9	
Q1871-02	MH-A-EPH	5	1.18	9.97	11.15	10.25	91.0	
Q1871-03	MH-A-VOC	6	1.15	10.22	11.37	10.47	91.2	
Q1871-05	MH-B	7	1.18	10.31	11.49	10.58	91.2	
Q1871-06	MH-B-EPH	8	1.16	9.63	10.79	10.05	92.3	
Q1871-07	MH-B-VOC	9	1.18	10.35	11.53	10.75	92.5	
Q1872-01	HW0425-PT-AN-SOIL	31	1.00	1.00	2.00	2.00	100.0	
Q1872-02	HW0425-PT-CORR-SOIL	32	1.00	1.00	2.00	2.00	100.0	
Q1872-03	HW0425-PT-CN-SOIL	33	1.00	1.00	2.00	2.00	100.0	
Q1872-04	HW0425-PT-CN-SOIL	34	1.00	1.00	2.00	2.00	100.0	
Q1872-05	HW0425-PT-FP-SOIL	35	1.00	1.00	2.00	2.00	100.0	
Q1872-06	HW0425-PT-CR6-SOIL	36	1.00	1.00	2.00	2.00	100.0	
Q1872-07	HW0425-PT-NUT-SOIL	37	1.00	1.00	2.00	2.00	100.0	
Q1872-08	HW0425-PT-NUT-SOIL	38	1.00	1.00	2.00	2.00	100.0	
Q1872-09	HW0425-PT-OGR-SOIL	39	1.00	1.00	2.00	2.00	100.0	
Q1872-10	HW0425-PT-MET-SOIL	40	1.00	1.00	2.00	2.00	100.0	
Q1872-11	HW0425-PT-BNA-SOIL	41	1.00	1.00	2.00	2.00	100.0	
Q1872-12	HW0425-PT-TRIAZINE-SOIL	42	1.00	1.00	2.00	2.00	100.0	
Q1872-13	HW0425-PT-PAH-SOIL	43	1.00	1.00	2.00	2.00	100.0	
Q1872-14	HW0425-PT-DIES-SOIL	44	1.00	1.00	2.00	2.00	100.0	
Q1872-15	HW0425-PT-GAS-SOIL	45	1.00	1.00	2.00	2.00	100.0	
Q1872-16	HW0425-PT-NJEPH-SOIL	46	1.00	1.00	2.00	2.00	100.0	
Q1872-17	HW0425-PT-HERB-SOIL	47	1.00	1.00	2.00	2.00	100.0	
Q1872-18	HW0425-PT-PCB-SOIL	48	1.00	1.00	2.00	2.00	100.0	
Q1872-19	HW0425-PT-PCBO-SOIL	49	1.00	1.00	2.00	2.00	100.0	

Alliance TECHNICAL GROUP

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/25/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC: LB135545

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1872-20	HW0425-PT-PEST-SOIL	50	1.00	1.00	2.00	2.00	100.0	
Q1872-21	HW0425-PT-CHLR-SOIL	51	1.00	1.00	2.00	2.00	100.0	
Q1872-22	HW0425-PT-TXP-SOIL	52	1.00	1.00	2.00	2.00	100.0	
Q1872-23	HW0425-PT-VOA-SOIL	53	1.00	1.00	2.00	2.00	100.0	
Q1872-25	HW0425-PT-NO2-SOIL	54	1.00	1.00	2.00	2.00	100.0	
Q1873-01	CAM-40619	10	1.14	10.70	11.84	4.97	35.8	
Q1873-02	CAM-40620	11	1.15	10.42	11.57	6.19	48.4	
Q1873-03	CAM-40619-20	12	1.18	10.21	11.39	4.77	35.2	
Q1874-01	VNJ-236	13	1.19	10.45	11.64	10.89	92.8	
Q1874-03	RT1491	14	1.19	11.16	12.35	11.43	91.8	
Q1874-05	HT3727	15	1.16	10.63	11.79	11.06	93.1	
Q1875-01	AUD-25-0053	16	1.14	10.75	11.89	11.19	93.5	
Q1875-03	AUD-25-0054	17	1.14	10.02	11.16	10.52	93.6	
Q1875-04	AUD-25-0024	18	1.14	10.03	11.17	10.77	96.0	
Q1876-01	AUD-25-0058	19	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-02	AUD-25-0059	20	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-03	AUD-25-0060	21	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-04	AUD-25-0061	22	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-05	AUD-25-0062	23	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-06	AUD-25-0063	24	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-07	AUD-25-0064	25	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-08	AUD-25-0065	26	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1876-09	AUD-25-0066	27	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1877-01	AU-6-042425	55	1.14	10.25	11.39	10.72	93.5	
Q1877-02	AU-6-042425	28	1.14	10.21	11.35	10.54	92.1	
Q1878-01	TR-4-042425	29	1.14	10.17	11.31	11.2	98.9	
Q1878-02	TR-4-042425-E2	30	1.19	10.28	11.47	10.92	94.6	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/25/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC: LB135545

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

LB 135545

WorkList Name : %1-042425 **WorkList ID :** 189122 **Department :** Wet-Chemistry **Date :** 04-24-2025 08:52:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1869-01	MH-F	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1869-02	MH-F-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1869-03	MH-F-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-01	MH-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-02	MH-A-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-03	MH-A-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-05	MH-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-06	MH-B-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1871-07	MH-B-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1872-01	HW0425-PT-AN-SOIL	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1872-02	HW0425-PT-CORR-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-03	HW0425-PT-CN-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-04	HW0425-PT-CN-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-05	HW0425-PT-FP-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-06	HW0425-PT-CR6-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-07	HW0425-PT-NUT-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-08	HW0425-PT-NUT-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-09	HW0425-PT-OGR-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-10	HW0425-PT-MET-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-11	HW0425-PT-BNA-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-12	HW0425-PT-TRIAZINE-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO

Date/Time 04/24/25 1530

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time 04/24/25 1725

Raw Sample Received by:

Raw Sample Relinquished by:

WORKLIST(Hardcopy Internal Chain)

135545

WorkList Name : %1-042425 **WorkList ID :** 189122 **Department :** Wet-Chemistry **Date :** 04-24-2025 08:52:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1872-13	HW0425-PT-PAH-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-14	HW0425-PT-DIES-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-15	HW0425-PT-GAS-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-16	HW0425-PT-NJEPH-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-17	HW0425-PT-HERB-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-18	HW0425-PT-PCB-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-19	HW0425-PT-PCBO-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-20	HW0425-PT-PEST-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-21	HW0425-PT-CHLR-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-22	HW0425-PT-TXP-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-23	HW0425-PT-VOA-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1872-25	HW0425-PT-NO2-SOIL	Solid	Percent Solids	Cool 4 deg C	ALLI03	QA Of	04/21/2025	Chemtech -SO
Q1873-01	CAM-40619	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1873-02	CAM-40620	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1873-03	CAM-40619-20	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1874-01	VNJ-236	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1874-03	RT1491	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1874-05	HT3727	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1875-01	AUD-25-0053	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1875-03	AUD-25-0054	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO
Q1875-04	AUD-25-0024	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/24/2025	Chemtech -SO

Date/Time 04/24/25 15:30

Raw Sample Received by: JD WDC

Raw Sample Relinquished by:

Date/Time 04/24/25 17:25

Raw Sample Received by:

Raw Sample Relinquished by: JD WDC

WORKLIST(Hardcopy Internal Chain)

135545

WorkList Name : %1-042425 **WorkList ID :** 189122 **Department :** Wet-Chemistry **Date :** 04-24-2025 08:52:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1876-01	AUD-25-0058	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-02	AUD-25-0059	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-03	AUD-25-0060	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-04	AUD-25-0061	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-05	AUD-25-0062	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-06	AUD-25-0063	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-07	AUD-25-0064	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-08	AUD-25-0065	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1876-09	AUD-25-0066	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1877-01	AU-6-042425	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/24/2025	Chemtech -SO
Q1877-02	AU-6-042425	Solid	Percent Solids	Cool 4 deg C	PSEG05	L41	04/24/2025	Chemtech -SO
Q1878-01	TR-4-042425	Solid	Percent Solids	Cool 4 deg C	PSEG05	L41	04/24/2025	Chemtech -SO
Q1878-02	TR-4-042425-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L41	04/24/2025	Chemtech -SO

Date/Time 04/24/25 15:30

Raw Sample Received by: [signature]

Raw Sample Relinquished by: [signature] SM

Date/Time 04/24/25 17:25

Raw Sample Received by: [signature] SM

Raw Sample Relinquished by: [signature]

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PB167904

SOP ID:	M3541-ASE Extraction-14				
Clean Up SOP #:	N/A			**Extraction Start Date :**	05/08/2025
Matrix :	Solid			**Extraction Start Time :**	09:15
Weigh By:	EH	**Extraction By:**	RJ	**Extraction End Date :**	05/08/2025
Balance check:	RJ	**Filter By:**	RJ	**Extraction End Time :**	12:30
Balance ID:	EX-SC-2	**pH Meter ID:**	N/A	**Concentration By:**	EH
pH Strip Lot#:	N/A	**Hood ID:**	3,7	**Supervisor By :**	RUPESH

Extraction Method: ☐ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☑ Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6752
Surrogate	1.0ML	100/150 PPM	SP6754
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
MeCl2/Acetone/1:1	N/A	EP2600
Baked Na2SO4	N/A	EP2607
Sand	N/A	EP2865
Methylene Chloride	N/A	E3930
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5ML Vial Lot # 2210443.

KD Bath ID:	N/A	**Envap ID:**	NEVAP-02
KD Bath Temperature:	N/A	**Envap Temperature:**	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/8/25	RS (Ext lab)	RC/SVOC
12:35	**Preparation Group**	**Analysis Group**

Alliance TECHNICAL GROUP

EXTRACTION LOGPAGE

PrepBatch ID : PB167904

Analytical Method: M3541-ASE Extraction-14 **Concentration Date:** 05/08/2025

Sample ID	Client Sample ID	Test	(g)/ mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167904BL	SBLK904	SVOC-TCL BNA -20	30.01	N/A	ritesh	Evelyn	1			U1-1
PB167904BS	SLCS904	SVOC-TCL BNA -20	30.02	N/A	ritesh	Evelyn	1			2
Q1872-11	HW0425-PT-BNA-SOIL	SVOCMS Group1	30.21	N/A	ritesh	Evelyn	1	A		3
Q1872-12	HW0425-PT-TRIAZINE-SOIL	SVOCMS Group4	30.37	N/A	ritesh	Evelyn	1	A		4
Q1964-01	MH-LL	SVOC-TCL BNA -20	50.07	N/A	ritesh	Evelyn	1	A		5
Q1964-01MS	MH-LLMS	SVOC-TCL BNA -20	50.06	N/A	ritesh	Evelyn	1	A		6
Q1964-01MSD	MH-LLMSD	SVOC-TCL BNA -20	50.03	N/A	ritesh	Evelyn	1	A		U6-1
Q1966-01	CONCRETE-1	SVOC-PAH	50.05	N/A	ritesh	Evelyn	1	D	Concrete	2
Q1966-02	CONCRETE-2	SVOC-PAH	50.02	N/A	ritesh	Evelyn	1	D	Concrete	3
Q1968-01	MH-N	SVOC-TCL BNA -20	50.03	N/A	ritesh	Evelyn	1	E		4
Q1968-05	MH-M	SVOC-TCL BNA -20	50.08	N/A	ritesh	Evelyn	1	E		5
Q1969-01	TR-06-05072025	SVOC-TCL BNA -20	50.01	N/A	ritesh	Evelyn	1	E		6
Q1972-01	SUB-WC	SVOC-TCL BNA -20	50.07	N/A	ritesh	Evelyn	1	E		U2-1
Q1972-05	WC-B-10	SVOC-TCL BNA -20	50.06	N/A	ritesh	Evelyn	1	E		2
Q1972-09	WC-B-2	SVOC-TCL BNA -20	50.03	N/A	ritesh	Evelyn	1	E		3
Q1972-13	WC-B-7	SVOC-TCL BNA -20	50.01	N/A	ritesh	Evelyn	1	E		4
Q1972-17	SUB-WC-2	SVOC-TCL BNA -20	50.04	N/A	ritesh	Evelyn	1	E		5
Q1972-21	WC-B-1	SVOC-TCL BNA -20	50.08	N/A	ritesh	Evelyn	1	E		6
Q1976-01	VNJ-207	SVOC-TCL BNA -20	50.03	N/A	ritesh	Evelyn	1	E		U3-1
Q1978-01	72-12016	SVOC-TCL BNA -20	50.02	N/A	ritesh	Evelyn	1	D		2

5/8

* **Extracts relinquished on the same date as received.**

Q1872-SVOCMS Group4

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1969 **WorkList ID :** 189383 **Department :** Extraction **Date :** 05-08-2025 09:08:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1872-11	HW0425-PT-BNA-SOIL	Solid	SVOCMS Group1	Cool 4 deg C	ALLI03	QA Of	04/21/2025	8270E
Q1872-12	HW0425-PT-TRIAZINE-SOIL	Solid	SVOCMS Group4	Cool 4 deg C	ALLI03	QA Of	04/21/2025	8270E
Q1964-01	MH-LL	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L31	05/06/2025	8270E
Q1966-01	CONCRETE-1	Solid	SVOC-PAH	Cool 4 deg C	PSEG03	L41	05/06/2025	8270E
Q1966-02	CONCRETE-2	Solid	SVOC-PAH	Cool 4 deg C	PSEG03	L41	05/06/2025	8270E
Q1968-01	MH-N	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L41	05/06/2025	8270E
Q1968-05	MH-M	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L41	05/06/2025	8270E
Q1969-01	TR-06-05072025	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG05	L41	05/07/2025	8270E
Q1972-01	SUB-WC	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L51	05/07/2025	8270E
Q1972-05	WC-B-10	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L51	05/07/2025	8270E
Q1972-09	WC-B-2	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L51	05/07/2025	8270E
Q1972-13	WC-B-7	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L51	05/07/2025	8270E
Q1972-17	SUB-WC-2	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L51	05/07/2025	8270E
Q1972-21	WC-B-1	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L51	05/07/2025	8270E
Q1976-01	VNJ-207	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L41	05/07/2025	8270E
Q1978-01	72-12016	Solid	SVOC-TCL BNA -20	Cool 4 deg C	PSEG03	L41	05/07/2025	8270E

Date/Time 05/08/25 9:10
Raw Sample Received by: RJ (EXT-LOB)
Raw Sample Relinquished by:

Date/Time 05/08/25 9:40
Raw Sample Received by:
Raw Sample Relinquished by: RJ (EXT-LOB)

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q1872

Test : SVOCMS Group4

Prepbatch ID : PB167904,

Sequence ID/Qc Batch ID: BM050825,bp050825,BP051425,

Standard ID :
EP2600,EP2607,EP2865,SP6721,SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729,SP6752,SP6754,SP6757,SP6769,SP6770,SP6783,SP6784,SP6785,SP6786,SP6787,SP6788,SP6789,SP6790,SP6791,

Chemical ID :
10ul/1000ul sample,E3551,E3874,E3876,E3902,E3904,E3917,E3926,E3930,S10104,S10397,S10584,S11087,S11143,S11161,S11162,S11487,S11495,S11496,S11650,S11785,S11786,S11787,S11788,S12114,S12115,S12190,S12191,S12192,S12193,S12194,S12195,S12209,S12210,S12211,S12212,S12213,S12214,S12215,S12216,S12270,S12271,S12276,S12277,S12478,S12479,S12480,S12481,S12482,S12483,S12484,S12485,S12486,S12525,S12526,S12527,S12528,S12529,S12530,S12531,S12532,S12533,S12534,S12577,S12649,S12658,S12662,S12663,S12664,S12791,S12793,S12966,S12967,S12968,S12969,S12970,S12971,S12972,S12973,S12974,S13078,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Extractions STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
2017	1:1 ACETONE/METHYLENE CHLORIDE	EP2600	04/07/2025	10/03/2025	Rajesh Parikh	None	None	Riteshkumar Patel 04/07/2025

FROM 8000.00000ml of E3904 + 8000.00000ml of E3917 = Final Quantity: 16000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3923	Baked Sodium Sulfate	EP2607	04/25/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 04/25/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3764	8270/625 Stock solution 100 ng	SP6721	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.26700ml of S10104 + 0.40000ml of S11495 + 0.50000ml of S12114 + 1.00000ml of S11087 + 1.00000ml of S11161 + 1.00000ml of S12270 + 1.00000ml of S12276 + 1.00000ml of S12791 + 3.83300ml of E3874 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
413	80 ng BNA ICC, 80 PPM	SP6722	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.20000ml of E3874 + 0.80000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
412	60 ng BNA ICC, 60 PPM	SP6723	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.40000ml of E3874 + 0.60000ml of SP6721 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
411	50 ng BNA ICC, 50 PPM	SP6724	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.50000ml of E3874 + 0.50000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
410	40 ng BNA ICC, 40 PPM	SP6725	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.60000ml of E3874 + 0.40000ml of SP6721 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3678	20 ng BNA ICC, 20 PPM	SP6726	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.80000ml of E3874 + 0.20000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
408	10 ng BNA ICC, 10 PPM	SP6727	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.90000ml of E3874 + 0.10000ml of SP6721 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
407	5 ng BNA ICC, 5 PPM	SP6728	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.95000ml of E3874 + 0.05000ml of SP6721 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
175	2.5 ng BNA ICC, 2.5 PPM	SP6729	01/30/2025	05/12/2025	Jagrut Upadhyay	None	None	Yogesh Patel 02/07/2025

FROM 0.01000ml of S12649 + 0.50000ml of E3874 + 0.50000ml of SP6728 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
171	8270/625 Spike Solution, 50/100 PPM	SP6752	03/10/2025	05/31/2025	Rahul Chavli	None	None	Jagrut Upadhyay 04/03/2025

FROM 0.10000ml of S12478 + 0.30000ml of S12525 + 0.40000ml of S10397 + 0.40000ml of S10584 + 0.40000ml of S11143 + 0.40000ml of S11487 + 0.40000ml of S11650 + 0.40000ml of S12533 + 0.40000ml of S12974 + 0.60000ml of S12486 + 0.80000ml of S12966 + 1.10000ml of S11788 + 1.20000ml of S11785 + 1.20000ml of S12483 + 1.20000ml of S12526 + 1.20000ml of S12967 + 1.20000ml of S12968 + 1.20000ml of S12970 + 1.20000ml of S12972 + 1.30000ml of S11786 + 1.30000ml of S12481 + 1.30000ml of S12482 + 1.30000ml of S12484 + 1.30000ml of S12528 + 1.30000ml of S12529 + 1.30000ml of S12531 + 1.30000ml of S12969 + 1.30000ml of S12973 + 1.40000ml of S11787 + 1.40000ml of S12479 + 1.40000ml of S12480 + 1.40000ml of S12485 + 1.40000ml of S12527 + 1.40000ml of S12530 + 1.40000ml of S12532 + 1.40000ml of S12971 + 163.00000ml of E3876 = Final Quantity: 200.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
19	8270/CLP Surrogate Solution, 100 PPM BN/150 PPM ACID	SP6754	03/18/2025	09/18/2025	Rahul Chavli	None	None	Jagrut Upadhyay 04/03/2025

FROM 1930.00000ml of E3902 + 2.60000ml of S12216 + 2.70000ml of S12195 + 5.20000ml of S12210 + 5.30000ml of S12192 + 5.30000ml of S12194 + 5.30000ml of S12209 + 5.30000ml of S12211 + 5.30000ml of S12212 + 5.30000ml of S12213 + 5.40000ml of S12190 + 5.40000ml of S12214 + 5.60000ml of S12191 + 5.60000ml of S12215 + 5.70000ml of S12193 = Final Quantity: 2000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3895	50 ug/ml DFTPP 8270E	SP6757	03/31/2025	09/30/2025	Rahul Chavli	None	None	Jagrut Upadhyay 04/01/2025

FROM 1.00000ml of S12577 + 19.00000ml of E3904 = Final Quantity: 20.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
18	Second Source Calibration Stock Standard, 100 PPM, (8270/625/CLP)	SP6769	04/10/2025	09/10/2025	Jagrut Upadhyay	None	None	Sohil Jodhani 04/16/2025

FROM 0.04000ml of S12195 + 0.08000ml of S12216 + 0.10000ml of S11788 + 0.20000ml of S12486 + 0.20000ml of S12534 + 0.20000ml of S12974 + 1.18000ml of E3926 = Final Quantity: 2.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
416	40 ng BNA ICV, 40 PPM	SP6770	04/10/2025	09/10/2025	Jagrut Upadhyay	None	None	Sohil Jodhani 04/16/2025

FROM 0.01000ml of S12658 + 0.60000ml of E3926 + 0.40000ml of SP6769 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3764	8270/625 Stock solution 100 ng	SP6783	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.26700ml of S10104 + 0.40000ml of S11496 + 0.50000ml of S12115 + 1.00000ml of S11162 + 1.00000ml of S12271 + 1.00000ml of S12277 + 1.00000ml of S12793 + 1.00000ml of S13078 + 3.83300ml of E3930 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
413	80 ng BNA ICC, 80 PPM	SP6784	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.20000ml of E3930 + 0.80000ml of SP6783 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
412	60 ng BNA ICC, 60 PPM	SP6785	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.40000ml of E3930 + 0.60000ml of SP6783 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
411	50 ng BNA ICC, 50 PPM	SP6786	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.50000ml of E3930 + 0.50000ml of SP6783 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
410	40 ng BNA ICC, 40 PPM	SP6787	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.60000ml of E3930 + 0.40000ml of SP6783 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3678	20 ng BNA ICC, 20 PPM	SP6788	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.80000ml of E3930 + 0.20000ml of SP6783 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
407	5 ng BNA ICC, 5 PPM	SP6789	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.95000ml of E3930 + 0.05000ml of SP6783 = Final Quantity: 1.010 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
408	10 ng BNA ICC, 10 PPM	SP6790	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.90000ml of E3930 + 0.10000ml of SP6783 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SVOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
175	2.5 ng BNA ICC, 2.5 PPM	SP6791	05/05/2025	06/21/2025	Jagrut Upadhyay	None	None	Rahul Chavli 05/16/2025

FROM 0.01000ml of S12662 + 0.50000ml of E3930 + 0.50000ml of SP6789 = Final Quantity: 1.010 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	07/30/2025	01/30/2025 / Rajesh	01/20/2025 / Rajesh	E3874

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	08/25/2025	02/25/2025 / Rajesh	02/12/2025 / Rajesh	E3876

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	09/18/2025	03/18/2025 / RUPESH	02/12/2025 / RUPESH	E3902

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	01/07/2026	03/13/2025 / RUPESH	12/27/2024 / RUPESH	E3904

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	10/08/2025	04/08/2025 / Rajesh	02/07/2025 / Rajesh	E3926

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	02/20/2026	05/02/2025 / RUPESH	03/09/2025 / RUPESH	E3930

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	07/30/2025	01/30/2025 / anahy	12/09/2021 / Christian	S10104

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555871 / Custom Standard, 4-nitrophenol Std [CS 5238-4]	A0185300	05/31/2025	01/29/2025 / anahy	05/18/2022 / Christian	S10397

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555868 / Custom Standard, Benzidine Std [CS 5328-1]	A0186373	06/30/2025	01/29/2025 / anahy	07/05/2022 / Christian	S10584

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days)	406703	07/30/2025	01/30/2025 / anahy	02/07/2023 / Christian	S11087

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555869 / Custom Standard, hexachlorocyclopentadiene Std [CS 5328-2]	A0194702	07/29/2025	01/29/2025 / anahy	02/20/2023 / Christian	S11143

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110817-01 / Custom 8270 Mix, 4-55, 1000 mg/L, 1 ml, (Maximum Expiration: 90 Days)	414125	06/21/2025	01/30/2025 / anahy	03/06/2023 / Christian	S11161

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110817-01 / Custom 8270 Mix, 4-55, 1000 mg/L, 1 ml, (Maximum Expiration: 90 Days)	414125	06/21/2025	05/05/2025 / Jagrut	03/06/2023 / Christian	S11162

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555870 / Custom Standard, 2,4-dinitrophenol Std [CS 5328-3]	A0200549	08/31/2026	01/29/2025 / anahy	08/10/2023 / yogesh	S11487

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	05/12/2025	11/12/2024 / Jagrut	08/11/2023 / Yogesh	S11495

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	10/28/2025	04/28/2025 / Jagrut	08/11/2023 / Yogesh	S11496

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	07/29/2025	01/29/2025 / anahy	11/09/2023 / Yogesh	S11650

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	07/29/2025	01/29/2025 / anahy	11/21/2023 / Rahul	S11785

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	09/10/2025	03/10/2025 / anahy	11/21/2023 / Rahul	S11786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	09/10/2025	03/10/2025 / anahy	11/21/2023 / Rahul	S11787

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	09/10/2025	03/10/2025 / anahy	11/21/2023 / Rahul	S11788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	05/12/2025	11/12/2024 / Jagrut	03/08/2024 / Rahul	S12114

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	10/28/2025	04/28/2025 / Jagrut	03/08/2024 / Rahul	S12115

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12190

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12191

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12192

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12193

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12194

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12195

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12209

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12210

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12211

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12212

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12213

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12214

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12215

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH2Cl2,5ml	A0206381	09/18/2025	03/18/2025 / anahy	03/15/2024 / Rahul	S12216

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	07/30/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12270

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	10/28/2025	04/28/2025 / Jagrut	05/24/2024 / Rahul	S12271

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days)	495833	05/12/2025	11/12/2024 / Jagrut	05/24/2024 / Rahul	S12276

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010442-07 / Benzaldehyde Solution, 1000 mg/L, 1.3 ml, (Maximum Expiration: 90 Days)	495833	11/05/2025	05/05/2025 / Jagrut	05/24/2024 / Rahul	S12277

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12478

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12479

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12480

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12481

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12482

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12483

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12484

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12485

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at $100 per ampul if requested - contact ARM with Request] [CS 4978-1]	A0214021	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12486

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12525

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12526

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12527

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12528

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12529

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12530

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12531

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12532

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	09/10/2025	03/10/2025 / anahy	07/23/2024 / RAHUL	S12533

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at $85 per ampul if requested - contact ARM with Request] [CS 4978-2]	A0214017	10/10/2025	04/10/2025 / Jagrut	07/23/2024 / RAHUL	S12534

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0212955	06/30/2027	03/31/2025 / Rahul	08/01/2024 / Rahul	S12577

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0212266	07/21/2025	01/21/2025 / anahy	09/20/2024 / anahy	S12649

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0212266	04/30/2030	04/07/2025 / anahy	09/20/2024 / anahy	S12658

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0212266	11/01/2025	05/01/2025 / Rahul	09/20/2024 / anahy	S12662

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0212266	11/07/2025	05/07/2025 / Rahul	09/20/2024 / anahy	S12663

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH2Cl2, 1mL	A0212266	11/12/2025	05/12/2025 / Rahul	09/20/2024 / anahy	S12664

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	06/21/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12791

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	06/21/2025	05/05/2025 / Jagrut	05/24/2024 / Rahul	S12793

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	07/29/2025	01/29/2025 / anahy	12/11/2024 / anahy	S12966

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12967

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12968

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12969

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12970

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12973

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	09/10/2025	03/10/2025 / anahy	12/11/2024 / anahy	S12974

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-010074-07 / 3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 ml, (Maximum Expiration: 180 days)	525551	11/05/2025	05/05/2025 / Jagrut	03/10/2025 / anahy	S13078

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010074-07	406703	≤ -10 °C	Methylene Chloride	3/30/2025	3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
3,3'-dichlorobenzidine	91-94-1	99.5	74.3.26P	989 ± 7.53

Received on
02/07/23
by
CG

S11084
to
S11088

*Not a certified value

Certified By: _____
Jacob Mulloy
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricllly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0 **Page 1 of 1**

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110817-01	414125	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-55, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acetophenone	98-86-2	99.2	85.8.1P	998 ± 11.5
benzoic acid	65-85-0	100	123.7.1P	1010 ± 5.88
biphenyl	92-52-4	99.9	366.29.1P	999 ± 5.82
1,2,4,5-tetrachlorobenzene	95-94-3	99.7	53.7.2P	993 ± 5.79

Received on
02/07/23
by
CG
S11089
to
S11093

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0 **Page 1 of 1**

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110817-01	414125	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-55, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acetophenone	98-86-2	99.2	85.8.1P	998 ± 11.5
benzoic acid	65-85-0	100	123.7.1P	1010 ± 5.88
biphenyl	92-52-4	99.9	366.29.1P	999 ± 5.82
1,2,4,5-tetrachlorobenzene	95-94-3	99.7	53.7.2P	993 ± 5.79

Received on
02/07/23
by
CG
S11089
to
S11093

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-112090-04	440246	≤ -10 °C	Methylene Chloride	2/16/2026	CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d_4	93951-73-6	99.3	248.12.7P	7487 ± 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 ± 17.26
phenol-d6	13127-88-3	99.9	949.120.8P	7481 ± 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 ± 17.17

Received on
02/25/21
by
CG

S9236
to
S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: Erica Castiglione

Erica Castiglione
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Gravimetric Certificate

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 05/19/22
S10393 to S10402

Catalog No. : 555871 **Lot No.:** A0185300

Description : Custom 4-Nitrophenol Standard

Custom 4-Nitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2025 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	4-Nitrophenol	25,060.0 µg/mL	+/- 231.9100 µg/mL	Gravimetric
	CAS # 100-02-7 (Lot MKCN1089)		+/- 753.2622 µg/mL	Unstressed
	Purity 99%		+/- 905.6020 µg/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Katelyn McGinnis - Operations Tech I

Date Mixed: 16-May-2022 **Balance:** 1128342314

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Gravimetric Certificate

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received by CG on 07/05/22 S10583 to S10592

Catalog No. : 555868 **Lot No.:** A0186373

Description : Custom Benzidine Standard

Custom Benzidine Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2025 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Benzidine	25,200.0 µg/mL	+/- 233.2055 µg/mL	Gravimetric
	CAS # 92-87-5 (Lot 220511RSR)		+/- 351.6606 µg/mL	Unstressed
	Purity 99%		+/- 512.6054 µg/mL	Stressed

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckar - Mix Technician **Date Mixed:** 16-Jun-2022 **Balance:** 1122030677

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555869 **Lot No.:** A0194702

Description : Custom Hexachlorocyclopentadiene Standard

Custom Hexachlorocyclopentadiene Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 28, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)
1	Hexachlorocyclopentadiene	77-47-4	0012019	99%	25,008.0 µg/mL

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Russ Bookhamer - Operations Technician I

Date Mixed: 15-Feb-2023 **Balance:** B442140311

Manufactured under Restek Registered Quality Certificate #FM

tified Reference Material Notes

tes:

n date valid for unopened ampul stored in compliance with the recommended conditions.

nty, concentration, and expiration of the CRM are based on the unopened product being stored according to the ended condition found in the storage field.

d/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, LC/MS, RI, and/or melting point.

nds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A n factor is used to calculate the amount of compound necessary to achieve the desired concentration of the mpound in solution.

isomeric compounds is reported as the sum of the isomers.

lues are rounded to the nearest whole number.

rtainty Value Notes:

rtainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded ty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability ty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

erage factor of 2, which gives a level of confidence of approximately 95%.

ged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure nimum packaged amount can be sufficiently transferred.

Notes:

tion is based upon gravimetric preparation using either a balance whose calibration has been verified daily traceable weights, and/or dilutions with Class A glassware.

:

the unopened product, when stored in compliance with the recommended conditions, is guaranteed through ion displayed on the product label and certificate. Contact Restek for additional opened product stability ı, with the knowledge/understanding that open product stability is subject to the specific handling and ntal conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with ards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom m. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, des complete instructions.

ssolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely

PRODUCTOS QUIMICOS MONTERREY, S.A. DE C.V.

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na_2SO_4
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na_2SO_4)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO_4)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreing matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RG-02-01, Ed. 3

Recd. by RB on 7/24/23 E3551

PO: PO2-1178.2 PRODUCT CODE: SHIP DATE: 1/20/2025

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date:2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	99.9 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.8 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3874

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

PO: PO2-798 PRODUCT CODE: SHIP DATE: 12/9/2024

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor | J.T.Baker

Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date:2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3902

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™ | J.T.Baker

Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date:2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ($(CH_3)_2CO$) (by GC, corrected forwater)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titrable Acid (µeq/g)	<= 0.3	0.2
Titrable Base (µeq/g)	<= 0.6	<0.1
Water (H_2O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

PO: PO2-1308 PRODUCT CODE: SHIP DATE: 2/7/25

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date:2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	99.9 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.8 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3926

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

PO: PO2-1178.2 PRODUCT CODE: SHIP DATE: 1/20/2025

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor | J.T.Baker

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date:2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol)Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	99.9 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.8 ppm
Titrable Acid (µeq/g)	<= 0.3	<0.1
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	<0.01 %

For Laboratory,Research,or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3930

JCroak
Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials,LLC
100 Matsonford Rd, Suite 200,Radnor,PA,19087.U.S.A.Phone 610.386.1700

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555870 **Lot No.:** A0200549

Description : Custom 2,4-Dinitrophenol Standard

Custom 2,4-Dinitrophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

S11484 ↓ S11493 Y.P. 08/10/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	25,008.0 µg/mL	+/- 777.3323

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Tom Suckar - Mix Technician **Date Mixed:** 02-Aug-2023 **Balance:** 1128342314

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d_4	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 to 511498 } Y.P. 08/11/2023

*Not a certified value

Certified By: _____

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d_4	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 to 511498 } Y.P. 08/11/2023

*Not a certified value

Certified By: _____

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0201728

Description : Custom Pentachlorophenol Standard

Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

511649 ↓ 511658 Y.P. 11/13/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023 **Balance:** B251644995

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : March 31, 2028 **Storage:** 0°C or colder

Ship: Ambient

S11749 ↓ S11794 RC 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant flow 1.8 mL/min.

Temp. Program:
80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:
250°C

Det. Temp:
340°C

Det. Type:
FID

Split Vent:
100 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

S12112 ↓ S12116 } RC/ 03/08/24

*Not a certified value

Certified By: Melissa Workoff

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

S12112 ↓ S12116 } RC/ 03/08/24

*Not a certified value

Certified By: Melissa Workoff

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

S12187 ↓ S12206 } RC/ 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024 **Balance Serial #** 1128360905

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381

Description : B/N Surrogate Mix (4/89 SOW)

Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul

S12207 ↓ S12221 RC/ 03/18/24

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonicate prior to use. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
2 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024 **Balance Serial #** 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

S12270 ↓ S12274 } RC 05/24/24

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

S12270 ↓ S12274 } RC 05/24/24

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

Certificate of Analysis

Catalog No.: Z-110381-01 **Lot No.:** 520963 **Expiration Date:** 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value

Certified By: Kerry E Kane

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010442-07	495833	≤ -10 °C	Methylene Chloride	1/16/2028	Benzaldehyde Solution, 1000 mg/L, 1.3 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
benzaldehyde	100-52-7	98.3	442.421.1P	996.8 ± 11.49

S12275 ↓ S12279 } RC/ 05/24/24

*Not a certified value

Certified By: S. Hunter

Scott Hunter
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010442-07	495833	≤ -10 °C	Methylene Chloride	1/16/2028	Benzaldehyde Solution, 1000 mg/L, 1.3 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
benzaldehyde	100-52-7	98.3	442.421.1P	996.8 ± 11.49

S12275 ↓ S12279 } RC/ 05/24/24

*Not a certified value

Certified By: S. Hunter

Scott Hunter
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 ↓ S12508 } RC 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024 **Balance:** 1128353505

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 ↓ S12568 } RC/ 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024 **Balance:** 1128360905

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31615 **Lot No.:** A0212955

Description : GC/MS Tuning Mixture

GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2027 **Storage:** 10°C or colder

Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,004.5 µg/mL	+/- 44.8902
2	DFTPP (Decafluorotriphenylphosphine)	5074-71-5	Q117-147	99%	1,004.5 µg/mL	+/- 44.8902
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 44.9572
4	4,4'-DDT	50-29-3	S240530RSR	97%	1,000.1 µg/mL	+/- 44.6922

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12577 ↓ S12579 RC 8/2/24

Quality Confirmation Test

Column:
30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:
hydrogen-constant pressure 10 psi.

Temp. Program:
75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:
250°C

Det. Temp:
330°C

Det. Type:
FID

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Ethan Winiarski - Operations Tech I

Date Mixed: 19-Jun-2024 **Balance Serial #** 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 26-Jun-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

chromatographic plus

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0212266

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,000.6 µg/mL	+/- 90.1075
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,000.3 µg/mL	+/- 90.0925
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,000.4 µg/mL	+/- 90.1000
4	Phenanthrene-d10	1517-22-2	PR-34099	99%	2,000.5 µg/mL	+/- 90.1037
5	Chrysene-d12	1719-03-5	PR-33506	99%	2,000.7 µg/mL	+/- 90.1112
6	Perylene-d12	1520-96-3	PR-33205	99%	2,000.6 µg/mL	+/- 90.1075

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12645 ↓ S12674 AC 10/1/24

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤-10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

S12280 → S12284 RCI 05/24/24

New numbers Generated.

S12790 → S12794 RCI 11/12/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤-10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

S12280 ↓ S12284 RCI 05/24/24

New numbers Generated.

S12790 ↓ S12794 RCI 11/12/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____
Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963 ↓ S12992 AC 12/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302

54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCCK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCCK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:
N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

CPI INTERNATIONAL

5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received:_____

Certificate of Analysis

Rev 0

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-010074-07	525551	-18°C +/- 4°C	Methylene Chloride	9/3/2029	3,3'-Dichlorobenzidine Solution, 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
3,3'-dichlorobenzidine	91-94-1	97.9	74.421.2P	994.7 ± 30.56

This RM is intended for use as a calibration standard or a quality control standard for chromatography equipment such as GC, GC/MS, HPLC, and HPLC/MS. It may also be used for various USEPA, NIOSH and ASTM methods.

Recommended storage container for ampuled products after opening is a 12 mm x 32 mm amber vial with screw cap Teflon lined silicon septum. The modeled % change per day can be calculated using the following:

$$\% \text{ Change} = 116192x^{-2.578} + 40.383e^{-0.03y}$$

where x = boiling point of the most volatile analyte in the mix (in degrees K)
y = boiling point of the solvent (in degrees K)

This model assumes the container is stored at -10 °C and is unopened during storage. The user should determine what the acceptable error for their process is and calculate the maximum number of days the opened ampule should be stored. The minimum sample size recommended for use is 1µL.

513078
↓
513087
AC
3/10/25

Manufactured by

Raevyn Steele
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00. Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetricly.

*Not a certified valu

SHIPPING DOCUMENTS

phenova® | A Phenomenex® Company
Certified Reference Materials

Packing List

Date	Order #
04/21/2025	333293

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

Received by: SJ
4/23/2025 15:50

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
PO2-1668	Net 30	ZCM-100	1500470	FedEx Collect 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	PT-MET-SOIL	SOIL/HW Trace Metals	HW0425	7100-04
1	1	0	PT-CR6-SOIL	SOIL/HW Hexavalent Chromium	HW0425	7100-05B
1	1	0	PT-CN-SOIL	SOIL/HW Cyanide	HW0425	7100-06
1	1	0	PT-CORR-SOIL	SOIL/HW Corrosivity/pH	HW0425	7100-11
1	1	0	PT-FP-SOIL	SOIL/HW Flash Point	HW0425	7100-10
1	1	0	PT-AN-SOIL	SOIL/HW Anions	HW0425	7100-08
1	1	0	PT-NUT-SOIL	SOIL/HW Nutrients	HW0425	7100-09B
1	1	0	PT-SOL-SOIL	SOIL/HW Solids	HW0425	7100-31
1	1	0	PT-NO2-SOIL	SOIL/HW Nitrite as N	HW0425	7100-71
1	1	0	PT-GAS-SOIL	SOIL/HW Gasoline	HW0425	7100-96
1	1	0	PT-OGR-SOIL	SOIL/HW Oil and Grease	HW0425	7100-94
1	1	0	PT-VOA-SOIL	SOIL/HW Volatiles	HW0425	7100-12
1	1	0	PT-BNA-SOIL	SOIL/HW BNAs	HW0425	7100-13
1	1	0	PT-PEST-SOIL	SOIL/HW Pesticides	HW0425	7100-14
1	1	0	PT-CHLR-SOIL	SOIL/HW Chlordane	HW0425	7100-15
1	1	0	PT-TXP-SOIL	SOIL/HW Toxaphene	HW0425	7100-16
1	1	0	PT-PCB-SOIL	SOIL/HW PCBs	HW0425	7100-17
1	1	0	PT-PCBO-SOIL	SOIL/HW PCBs in Oil	HW0425	7100-88
1	1	0	PT-HERB-SOIL	SOIL/HW Herbicides	HW0425	7100-18
1	1	0	PT-PAH-SOIL	SOIL/HW PAHs	HW0425	7100-22
1	1	0	PT-TRIAZINE-SOIL	SOIL/HW Triazine Pesticides	HW0425	7100-106
1	1	0	PT-NJEPH-SOIL	NJ EPH in SOIL	HW0425	7100-105

phenova® Certified Reference Materials | A Phenomenex® Company

Packing List

Date	Order #
04/25/2025	337220

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: SJ
4/28/2025 9:40

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
CPR	Net 30	ZCM-100	1500470	FedEx Next Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
1	1	0	PT-DIES-SOIL	SOIL/HW Diesel in Soil	HW0425	7100-100

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488