

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
 Data File : BM050032.D
 Acq On : 28 Apr 2025 17:43
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM042825

Quant Time: Apr 28 18:14:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 28 18:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	95	0.00
2	1,4-Dioxane	40.000	44.976	-12.4	103	0.00
3	Pyridine	40.000	46.534	-16.3	105	0.00
4	n-Nitrosodimethylamine	40.000	45.956	-14.9	104	0.00
5 S	2-Fluorophenol	80.000	87.928	-9.9	99	0.00
6	Aniline	40.000	46.744	-16.9	104	0.00
7 S	Phenol-d6	80.000	90.140	-12.7	100	0.00
8	2-Chlorophenol	40.000	45.724	-14.3	103	0.00
9	Benzaldehyde	40.000	43.937	-9.8	98	0.00
10 C	Phenol	40.000	46.913	-17.3	106	0.00
11	bis(2-Chloroethyl)ether	40.000	45.521	-13.8	103	0.00
12	1,3-Dichlorobenzene	40.000	44.672	-11.7	102	0.00
13 C	1,4-Dichlorobenzene	40.000	44.941	-12.4	103	0.00
14	1,2-Dichlorobenzene	40.000	45.174	-12.9	103	0.00
15	Benzyl Alcohol	40.000	48.167	-20.4	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	44.929	-12.3	102	0.00
17	2-Methylphenol	40.000	47.153	-17.9	104	0.00
18	Hexachloroethane	40.000	44.738	-11.8	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	47.892	-19.7	105	0.00
20	3+4-Methylphenols	40.000	47.974	-19.9	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	45.153	-12.9	104	0.00
23 S	Nitrobenzene-d5	80.000	77.798	2.8	90	0.00
24	Nitrobenzene	40.000	45.104	-12.8	104	0.00
25	Isophorone	40.000	45.596	-14.0	107	0.00
26 C	2-Nitrophenol	40.000	47.087	-17.7	106	0.00
27	2,4-Dimethylphenol	40.000	46.404	-16.0	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	44.920	-12.3	104	0.00
29 C	2,4-Dichlorophenol	40.000	46.375	-15.9	105	0.00
30	1,2,4-Trichlorobenzene	40.000	44.348	-10.9	104	0.00
31	Naphthalene	40.000	44.651	-11.6	105	0.00
32	Benzoic acid	40.000	46.063	-15.2	109	0.00
33	4-Chloroaniline	40.000	47.375	-18.4	110	0.00
34 C	Hexachlorobutadiene	40.000	44.034	-10.1	103	0.00
35	Caprolactam	40.000	46.481	-16.2	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	46.825	-17.1	109	0.00
37	2-Methylnaphthalene	40.000	45.374	-13.4	107	0.00
38	1-Methylnaphthalene	40.000	45.198	-13.0	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.140	-10.4	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.374	-10.9	106	0.00
42 S	2,4,6-Tribromophenol	80.000	90.503	-13.1	108	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.094	-15.2	109	0.00
44	2,4,5-Trichlorophenol	40.000	46.271	-15.7	108	0.00
45 S	2-Fluorobiphenyl	80.000	76.259	4.7	91	0.00
46	1,1'-Biphenyl	40.000	44.207	-10.5	107	0.00
47	2-Chloronaphthalene	40.000	44.684	-11.7	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
 Data File : BM050032.D
 Acq On : 28 Apr 2025 17:43
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM042825

Quant Time: Apr 28 18:14:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 28 18:09:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	47.621	-19.1	112	0.00
49	Acenaphthylene	40.000	45.315	-13.3	109	0.00
50	Dimethylphthalate	40.000	45.338	-13.3	111	0.00
51	2,6-Dinitrotoluene	40.000	47.554	-18.9	113	0.00
52 C	Acenaphthene	40.000	45.210	-13.0	110	0.00
53	3-Nitroaniline	40.000	48.843	-22.1	113	0.00
54 P	2,4-Dinitrophenol	40.000	45.189	-13.0	119	0.00
55	Dibenzofuran	40.000	44.828	-12.1	109	0.00
56 P	4-Nitrophenol	40.000	46.012	-15.0	118	0.00
57	2,4-Dinitrotoluene	40.000	48.311	-20.8	113	0.00
58	Fluorene	40.000	46.115	-15.3	111	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	46.970	-17.4	113	0.00
60	Diethylphthalate	40.000	45.299	-13.2	111	0.00
61	4-Chlorophenyl-phenylether	40.000	45.950	-14.9	110	0.00
62	4-Nitroaniline	40.000	50.610	-26.5#	115	0.00
63	Azobenzene	40.000	45.747	-14.4	111	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.920	-17.3	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	44.485	-11.2	112	0.00
67	4-Bromophenyl-phenylether	40.000	43.944	-9.9	111	0.00
68	Hexachlorobenzene	40.000	44.174	-10.4	113	0.00
69	Atrazine	40.000	44.483	-11.2	112	0.00
70 C	Pentachlorophenol	40.000	46.059	-15.1	117	0.00
71	Phenanthrene	40.000	44.895	-12.2	115	0.00
72	Anthracene	40.000	45.381	-13.5	115	0.00
73	Carbazole	40.000	46.001	-15.0	116	0.00
74	Di-n-butylphthalate	40.000	45.654	-14.1	116	0.00
75 C	Fluoranthene	40.000	46.998	-17.5	119	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	115	0.00
77	Benzydine	40.000	44.204	-10.5	117	0.00
78	Pyrene	40.000	44.209	-10.5	118	0.00
79 S	Terphenyl-d14	80.000	78.970	1.3	101	0.00
80	Butylbenzylphthalate	40.000	44.718	-11.8	121	0.00
81	Benzo(a)anthracene	40.000	45.497	-13.7	122	0.00
82	3,3'-Dichlorobenzidine	40.000	43.871	-9.7	115	0.00
83	Chrysene	40.000	44.967	-12.4	123	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.224	-13.1	122	0.00
85 c	Di-n-octyl phthalate	40.000	44.974	-12.4	123	0.00
86 I	Perylene-d12	20.000	20.000	0.0	118	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	46.249	-15.6	127	0.00
88	Benzo(b)fluoranthene	40.000	45.066	-12.7	123	0.00
89	Benzo(k)fluoranthene	40.000	45.780	-14.5	128	0.00
90 C	Benzo(a)pyrene	40.000	45.577	-13.9	127	0.00
91	Dibenzo(a,h)anthracene	40.000	46.091	-15.2	127	-0.01
92	Benzo(g,h,i)perylene	40.000	45.437	-13.6	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042825\
Data File : BM050032.D
Acq On : 28 Apr 2025 17:43
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM042825

Quant Time: Apr 28 18:14:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 28 18:09:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0