

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf071725

SequenceID :	<u>bf071725</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)		_____	_____	<u>✓</u>
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)		_____	_____	<u>✓</u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series		_____	_____	<u>✓</u>
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series		_____	_____	<u>✓</u>
5. GC/MS Calibration Met:		_____	_____	<u>✓</u>
a. Initial calibration Meet Criteria		_____	_____	<u>✓</u>
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____
The %RSD is greater than 20% for certain compounds in the Initial Calibration (Method 8270-BF071725.M) for Benzoic acid, 2,4-Dinitrophenol, 4,6-Dinitro-2-methylphenol this Compound is passing on Linear Regression . Compound#77 removed from 80				
b. Continuous Calibration(CCC) Meet Criteria		_____	_____	<u>✓</u>
If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:		_____	<u>✓</u>	_____
a. B/N Fraction				
d. Acid Fraction				

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

____✓

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

____✓

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

____✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

____✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

____✓

ADDITIONAL COMMENTS:

Pratik

Analyst

07/18/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf071725

SequenceID : bf071725 NA NO YES