

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BP051425

SequenceID : BP051425

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

\_\_\_\_

\_\_\_\_

\_\_\_\_✓

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

\_\_\_\_✓

\_\_\_\_

\_\_\_\_

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

\_\_\_\_

\_\_\_\_

\_\_\_\_✓

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_

\_\_\_\_

\_\_\_\_✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_

\_\_\_\_

\_\_\_\_✓

ADDITIONAL COMMENTS:

\_\_\_\_\_  
krunal  
Analyst

\_\_\_\_\_  
05/15/2025  
Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BP051425

SequenceID : BP051425 NA NO YES