

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN042825

Review By	Rahul	Review On	4/29/2025 8:54:47 AM
Supervise By	Jagrut	Supervise On	4/29/2025 11:57:54 AM
SubDirectory	BN042825	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN042825
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036922.D	28 Apr 2025 10:56	RC/JU	Ok
2	SSTDICC0.1	BN036923.D	28 Apr 2025 11:35	RC/JU	Ok,M
3	SSTDICC0.2	BN036924.D	28 Apr 2025 12:11	RC/JU	Ok,M
4	SSTDICCC0.4	BN036925.D	28 Apr 2025 12:47	RC/JU	Ok
5	SSTDICC0.8	BN036926.D	28 Apr 2025 13:24	RC/JU	Ok
6	SSTDICC1.6	BN036927.D	28 Apr 2025 14:00	RC/JU	Ok
7	SSTDICC3.2	BN036928.D	28 Apr 2025 14:36	RC/JU	Ok
8	SSTDICC5.0	BN036929.D	28 Apr 2025 15:12	RC/JU	Ok
9	SSTDICV0.4	BN036930.D	28 Apr 2025 15:51	RC/JU	Ok
10	PB167430BL	BN036931.D	28 Apr 2025 17:39	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN050825

Review By	Rahul	Review On	5/9/2025 9:44:58 AM
Supervise By	Jagrut	Supervise On	5/9/2025 3:46:05 PM
SubDirectory	BN050825	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN042825
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036971.D	08 May 2025 13:49	RC/JU	Ok
2	SSTDCCC0.4	BN036972.D	08 May 2025 14:28	RC/JU	Ok
3	PB167915BL	BN036973.D	08 May 2025 15:04	RC/JU	Ok
4	Q1939-04	BN036974.D	08 May 2025 15:40	RC/JU	Ok
5	Q1939-04MS	BN036975.D	08 May 2025 16:16	RC/JU	Ok
6	Q1939-04MSD	BN036976.D	08 May 2025 16:53	RC/JU	Ok
7	Q1939-02	BN036977.D	08 May 2025 17:29	RC/JU	Ok
8	Q1939-06	BN036978.D	08 May 2025 18:05	RC/JU	Ok,M
9	PB167915BS	BN036979.D	08 May 2025 18:41	RC/JU	Ok
10	SSTDCCC0.4	BN036980.D	08 May 2025 19:18	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN051225

Review By	Rahul	Review On	5/13/2025 4:20:20 PM
Supervise By	Jagrut	Supervise On	5/13/2025 5:37:31 PM
SubDirectory	BN051225	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn051225
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036981.D	12 May 2025 13:04	RC/JU	Ok
2	SSTDICC0.1	BN036982.D	12 May 2025 13:43	RC/JU	Ok,M
3	SSTDICC0.2	BN036983.D	12 May 2025 14:19	RC/JU	Ok,M
4	SSTDICCC0.4	BN036984.D	12 May 2025 14:55	RC/JU	Ok
5	SSTDICC0.8	BN036985.D	12 May 2025 15:31	RC/JU	Ok
6	SSTDICC1.6	BN036986.D	12 May 2025 16:07	RC/JU	Ok
7	SSTDICC3.2	BN036987.D	12 May 2025 16:43	RC/JU	Ok
8	SSTDICC5.0	BN036988.D	12 May 2025 17:19	RC/JU	Ok
9	SSTDICV0.4	BN036989.D	12 May 2025 17:55	RC/JU	Ok
10	PB167888BL	BN036990.D	12 May 2025 19:44	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # BN051325

Review By	Rahul	Review On	5/14/2025 2:19:26 PM
Supervise By	Jagrut	Supervise On	5/14/2025 5:15:40 PM
SubDirectory	BN051325	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn051225
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036991.D	13 May 2025 09:55	RC/JU	Ok
2	SSTDCCC0.4	BN036992.D	13 May 2025 10:39	RC/JU	Ok
3	PB167915BL	BN036993.D	13 May 2025 12:27	RC/JU	Not Ok
4	Q1872-11	BN036994.D	13 May 2025 13:23	RC/JU	Dilution
5	Q1872-11DL	BN036995.D	13 May 2025 14:20	RC/JU	Ok
6	Q1872-12	BN036996.D	13 May 2025 15:14	RC/JU	Ok
7	SSTDCCC0.4	BN036997.D	13 May 2025 15:58	RC/JU	Ok

M : Manual Integration