

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1872 SAS No.: Q1872 SDG NO.: Q1872

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/14/2025

Lab File ID: BN037009.D Time Analyzed: 10:31

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4780	17.009	4137	21.215	4043	23.421
UPPER LIMIT	9560	17.509	8274	21.715	8086	23.921
LOWER LIMIT	2390	16.509	2068.5	20.715	2021.5	22.921
EPA SAMPLE NO.						
01 HW0425-PT-PAH-SOIL	7619	17.01	7847	21.21	7344	23.41
02 HW0425-PT-PAH-SOILDL	7555	17.01	6828	21.21	6318	23.42

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.